

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

Agency: FDLE PENS

Date/Time: 03/20/2009 10:50:21

Data Download Date/Time: 03/20/2009 14:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

Agency: FDLE FMROC

Date/Time: 08/11/2009 11:15:40

Data Download Date/Time: 08/11/2009 16:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000221

Agency: FDLE TROC TRAINING

Date/Time: 12/08/2009 16:23:31

Data Download Date/Time: 12/09/2009 10:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

Agency: FDLE / JROC

Date/Time: 01/05/2009 15:47:16

Data Download Date/Time: 01/13/2009 10:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FDLE / JROC

Date/Time: 01/05/2009 15:48:12

Data Download Date/Time: 01/13/2009 10:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

Agency: FDLE / JROC

Date/Time: 01/05/2009 15:56:04

Data Download Date/Time: 01/13/2009 10:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FDLE / JROC

Date/Time: 01/05/2009 15:57:23

Data Download Date/Time: 01/13/2009 10:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

Agency: FDLE / JROC

Date/Time: 08/11/2009 08:40:30

Data Download Date/Time: 08/11/2009 12:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE / JROC

Date/Time: 08/11/2009 09:43:46

Data Download Date/Time: 08/11/2009 12:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

Agency: FDLE / JROC

Date/Time: 08/25/2009 08:43:20

Data Download Date/Time: 09/25/2009 09:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FDLE / JROC

Date/Time: 08/25/2009 08:50:55

Data Download Date/Time: 09/25/2009 09:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

Agency: FDLE / JROC

Date/Time: 08/25/2009 09:00:47

Data Download Date/Time: 09/25/2009 09:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

Agency: MT DORA P.D.

Date/Time: 03/24/2009 14:05:39

Data Download Date/Time: 03/30/2009 22:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MT DORA P.D.

Date/Time: 11/16/2009 13:29:11

Data Download Date/Time: 11/16/2009 17:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

Agency: FDLE

Date/Time: 03/06/2009 14:02:56

Data Download Date/Time: 03/09/2009 09:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE

Date/Time: 04/06/2009 14:49:54

Data Download Date/Time: 04/08/2009 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

Agency: FDLE

Date/Time: 09/04/2009 13:39:18

Data Download Date/Time: 09/04/2009 15:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

Agency: ESCAMBIA COUNTY SO

Date/Time: 07/16/2009 13:52:14

Data Download Date/Time: 07/17/2009 09:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

Agency: ESCAMBIA COUNTY SO

Date/Time: 03/26/2009 15:47:34

Data Download Date/Time: 03/26/2009 17:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: ESCAMBIA COUNTY SO

Date/Time: 06/03/2009 15:12:59

Data Download Date/Time: 06/04/2009 09:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

Agency: ESCAMBIA COUNTY SO

Date/Time: 06/18/2009 14:49:51

Data Download Date/Time: 06/18/2009 19:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

Agency: MARION COUNTY SO

Date/Time: 04/21/2009 13:05:19

Data Download Date/Time: 04/21/2009 17:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 05/16/2009 17:44:54

Data Download Date/Time: 06/23/2009 09:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

Agency: MARION COUNTY SO

Date/Time: 08/06/2009 16:11:08

Data Download Date/Time: 08/06/2009 19:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 09/08/2009 10:40:20

Data Download Date/Time: 09/09/2009 10:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

Agency: JACKSON COUNTY SO

Date/Time: 10/01/2009 08:08:20

Data Download Date/Time: 10/30/2009 14:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: JACKSON COUNTY SO

Date/Time: 10/30/2009 08:55:33

Data Download Date/Time: 10/30/2009 14:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

Agency: ALACHUA COUNTY SO

Date/Time: 02/10/2009 13:32:26

Data Download Date/Time: 02/10/2009 17:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 04/30/2009 12:36:31

Data Download Date/Time: 05/26/2009 22:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/02/2009 19:55:18

Data Download Date/Time: 01/02/2009 21:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 02/13/2009 18:57:41

Data Download Date/Time: 02/13/2009 20:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

Agency: SANTA ROSA COUNTY SO

Date/Time: 03/20/2009 19:50:31

Data Download Date/Time: 03/20/2009 23:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 04/24/2009 18:03:06

Data Download Date/Time: 04/24/2009 19:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

Agency: SANTA ROSA COUNTY SO

Date/Time: 05/08/2009 18:48:30

Data Download Date/Time: 05/08/2009 20:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 07/29/2009 18:35:29

Data Download Date/Time: 07/29/2009 20:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

Agency: SANTA ROSA COUNTY SO

Date/Time: 08/31/2009 12:33:15

Data Download Date/Time: 08/31/2009 14:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/29/2009 09:02:39

Data Download Date/Time: 09/29/2009 10:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/29/2009 09:19:50

Data Download Date/Time: 09/29/2009 10:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 10/20/2009 12:17:25

Data Download Date/Time: 10/20/2009 14:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000765

Agency: SANTA ROSA COUNTY SO

Date/Time: 11/30/2009 10:46:49

Data Download Date/Time: 11/30/2009 12:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 12/28/2009 13:05:47

Data Download Date/Time: 12/28/2009 14:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

Agency: HAMILTON COUNTY SO

Date/Time: 01/29/2009 01:44:53

Data Download Date/Time: 01/29/2009 03:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HAMILTON COUNTY SO

Date/Time: 03/31/2009 01:24:05

Data Download Date/Time: 03/31/2009 02:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

Agency: HAMILTON COUNTY SO

Date/Time: 04/28/2009 21:10:28

Data Download Date/Time: 04/28/2009 21:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HAMILTON COUNTY SO

Date/Time: 05/29/2009 11:58:22

Data Download Date/Time: 05/29/2009 12:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/02/2009 19:24:49

Data Download Date/Time: 01/02/2009 21:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 02/13/2009 17:44:24

Data Download Date/Time: 02/13/2009 19:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

Agency: SANTA ROSA COUNTY SO

Date/Time: 03/20/2009 18:42:54

Data Download Date/Time: 03/20/2009 20:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 04/24/2009 17:12:09

Data Download Date/Time: 04/24/2009 18:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

Agency: SANTA ROSA COUNTY SO

Date/Time: 05/08/2009 18:30:45

Data Download Date/Time: 05/08/2009 20:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 06/19/2009 17:21:25

Data Download Date/Time: 06/19/2009 18:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

Agency: SANTA ROSA COUNTY SO

Date/Time: 07/29/2009 18:09:42

Data Download Date/Time: 08/18/2009 09:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SANTA ROSA COUNTY SO

Date/Time: 07/29/2009 18:10:19

Data Download Date/Time: 08/18/2009 09:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

Agency: SANTA ROSA COUNTY SO

Date/Time: 07/29/2009 18:11:47

Data Download Date/Time: 08/18/2009 09:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SANTA ROSA COUNTY SO

Date/Time: 12/28/2009 12:10:14

Data Download Date/Time: 12/28/2009 14:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

Agency: LAFAYETTE CO SO

Date/Time: 01/28/2009 09:26:32

Data Download Date/Time: 01/28/2009 13:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

Agency: COLUMBIA COUNTY SO

Date/Time: 04/03/2009 14:02:00

Data Download Date/Time: 04/29/2009 16:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency:

Date/Time: 06/04/2009 11:51:50

Data Download Date/Time: 06/04/2009 14:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

Agency: LAFAYETTE CO SO

Date/Time: 02/25/2009 16:05:19

Data Download Date/Time: 03/25/2009 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

Agency: COLUMBIA COUNTY SO

Date/Time: 05/29/2009 14:24:27

Data Download Date/Time: 05/29/2009 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

Agency: MADISON COUNTY SO

Date/Time: 08/06/2009 14:36:26

Data Download Date/Time: 08/06/2009 18:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

Agency: SUWANNEE CO SO

Date/Time: 05/20/2009 16:02:53

Data Download Date/Time: 05/30/2009 12:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

Agency: HOLMES COUNTY SO

Date/Time: 10/12/2009 15:57:00

Data Download Date/Time: 10/13/2009 14:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH PD

Date/Time: 02/20/2009 14:50:34

Data Download Date/Time: 03/05/2009 14:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

Agency: HILLSBOROUGH CO SO

Date/Time: 11/24/2009 13:15:54

Data Download Date/Time: 11/24/2009 14:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

Agency: PERRY POLICE DEPT

Date/Time: 11/10/2009 15:40:09

Data Download Date/Time: 11/27/2009 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

Agency: HERNANDO COUNTY SO

Date/Time: 06/25/2009 10:05:32

Data Download Date/Time: 06/25/2009 14:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

Agency: MARION COUNTY SO

Date/Time: 03/19/2009 09:51:24

Data Download Date/Time: 03/19/2009 10:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

Agency: SUMTER COUNTY SO

Date/Time: 11/16/2009 10:37:52

Data Download Date/Time: 11/16/2009 14:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SUMTER COUNTY SO

Date/Time: 11/17/2009 12:45:26

Data Download Date/Time: 11/17/2009 14:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency: CITRUS COUNTY SO

Date/Time: 09/17/2009 08:46:47

Data Download Date/Time: 09/17/2009 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 01/31/2009 12:17:46

Data Download Date/Time: 01/31/2009 14:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CITRUS COUNTY SO

Date/Time: 02/18/2009 12:39:04

Data Download Date/Time: 02/18/2009 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 03/16/2009 17:56:43

Data Download Date/Time: 04/09/2009 08:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CITRUS COUNTY SO

Date/Time: 04/23/2009 14:41:11

Data Download Date/Time: 04/23/2009 17:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 06/08/2009 08:31:27

Data Download Date/Time: 06/19/2009 13:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CITRUS COUNTY SO

Date/Time: 06/12/2009 13:20:19

Data Download Date/Time: 06/19/2009 13:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 06/19/2009 11:07:38

Data Download Date/Time: 06/19/2009 13:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency: LAKE COUNTY S.O.

Date/Time: 08/22/2009 20:43:47

Data Download Date/Time: 08/29/2009 11:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE COUNTY S.O.

Date/Time: 08/22/2009 21:01:58

Data Download Date/Time: 08/29/2009 11:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency: LAKE COUNTY S.O.

Date/Time: 08/22/2009 21:00:20

Data Download Date/Time: 08/29/2009 10:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 09/04/2009 22:02:46

Data Download Date/Time: 09/30/2009 10:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEESBURG P.D.

Date/Time: 09/30/2009 09:46:25

Data Download Date/Time: 09/30/2009 10:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency: MOUNT DORA P.D.

Date/Time: 04/22/2009 13:51:42

Data Download Date/Time: 04/23/2009 16:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency: MARION COUNTY SO

Date/Time: 03/29/2009 10:09:09

Data Download Date/Time: 04/18/2009 09:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: MARION COUNTY SO

Date/Time: 03/29/2009 10:29:55

Data Download Date/Time: 04/18/2009 09:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency: MARION COUNTY SO

Date/Time: 03/29/2009 10:41:51

Data Download Date/Time: 04/18/2009 09:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 03/31/2009 09:40:12

Data Download Date/Time: 04/18/2009 09:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency: MARION COUNTY SO

Date/Time: 05/16/2009 18:19:27

Data Download Date/Time: 06/23/2009 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 05/16/2009 18:20:46

Data Download Date/Time: 06/23/2009 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency: MARION COUNTY SO

Date/Time: 05/16/2009 18:22:39

Data Download Date/Time: 06/23/2009 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 05/16/2009 18:33:37

Data Download Date/Time: 06/23/2009 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency: MARION COUNTY SO

Date/Time: 05/16/2009 18:35:01

Data Download Date/Time: 06/23/2009 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 07/13/2009 01:32:38

Data Download Date/Time: 07/19/2009 08:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000829

Agency: HILLSBOROUGH CO S0

Date/Time: 03/25/2009 13:42:34

Data Download Date/Time: 03/26/2009 08:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HILLSBOROUGH CO S0

Date/Time: 11/24/2009 15:25:44

Data Download Date/Time: 11/25/2009 11:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency: HILLSBOROUGH CO SO

Date/Time: 03/30/2009 16:36:44

Data Download Date/Time: 03/31/2009 14:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 04/02/2009 21:10:53

Data Download Date/Time: 04/14/2009 11:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: MARTIN COUNTY SO

Date/Time: 07/09/2009 12:27:53

Data Download Date/Time: 07/09/2009 18:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency: HILLSBOROUGH CO SO

Date/Time: 02/17/2009 14:18:35

Data Download Date/Time: 02/17/2009 17:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HILLSBOROUGH CO SO

Date/Time: 04/29/2009 12:51:08

Data Download Date/Time: 06/03/2009 15:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency: HILLSBOROUGH CO SO

Date/Time: 06/03/2009 10:51:45

Data Download Date/Time: 06/03/2009 15:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HILLSBOROUGH CO SO

Date/Time: 06/18/2009 10:02:30

Data Download Date/Time: 06/18/2009 12:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency: HILLSBOROUGH CO SO

Date/Time: 09/19/2009 23:00:48

Data Download Date/Time: 09/25/2009 11:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HILLSBOROUGH CO SO

Date/Time: 11/01/2009 02:54:39

Data Download Date/Time: 11/24/2009 10:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency: HILLSBOROUGH CO SO

Date/Time: 11/09/2009 08:38:22

Data Download Date/Time: 11/24/2009 10:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HILLSBOROUGH CO SO

Date/Time: 11/09/2009 08:48:47

Data Download Date/Time: 11/24/2009 10:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency: HILLSBOROUGH CO SO

Date/Time: 11/09/2009 08:49:33

Data Download Date/Time: 11/24/2009 10:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH PD

Date/Time: 07/23/2009 14:07:18

Data Download Date/Time: 07/24/2009 15:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency: LYNN HAVEN PD

Date/Time: 04/11/2009 07:33:25

Data Download Date/Time: 04/29/2009 11:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency: CAPE CORAL PD

Date/Time: 05/07/2009 14:38:02

Data Download Date/Time: 05/12/2009 09:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CAPE CORAL PD

Date/Time: 10/21/2009 16:16:07

Data Download Date/Time: 10/22/2009 09:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

Agency: MONROE COUNTY SO

Date/Time: 08/19/2009 21:01:27

Data Download Date/Time: 09/25/2009 21:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency: HOMESTEAD PD

Date/Time: 01/27/2009 10:24:33

Data Download Date/Time: 01/27/2009 15:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency: MIAMI-DADE PD

Date/Time: 10/14/2009 08:59:30

Data Download Date/Time: 10/22/2009 12:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI-DADE PD

Date/Time: 10/14/2009 09:02:29

Data Download Date/Time: 10/22/2009 12:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

Agency: PINELLAS COUNTY SO

Date/Time: 01/11/2009 04:27:57

Data Download Date/Time: 01/13/2009 18:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PINELLAS COUNTY SO

Date/Time: 09/24/2009 01:47:49

Data Download Date/Time: 09/24/2009 22:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency: FWC JACKSONVILLE

Date/Time: 06/05/2009 15:20:54

Data Download Date/Time: 06/05/2009 17:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FWC JACKSONVILLE

Date/Time: 06/18/2009 12:50:23

Data Download Date/Time: 06/18/2009 15:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency: FFWCC TAMPA

Date/Time: 11/12/2009 15:48:43

Data Download Date/Time: 11/24/2009 14:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FFWCC TAMPA

Date/Time: 11/24/2009 08:36:01

Data Download Date/Time: 11/24/2009 14:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

Agency: FWC - NORTHEAST

Date/Time: 06/24/2009 10:10:27

Data Download Date/Time: 06/24/2009 14:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

Agency: FISH & WILDLIFE CC

Date/Time: 07/16/2009 15:59:33

Data Download Date/Time: 07/17/2009 11:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FISH & WILDLIFE CC

Date/Time: 07/17/2009 08:38:10

Data Download Date/Time: 07/17/2009 11:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 06/02/2009 13:27:30

Data Download Date/Time: 06/26/2009 23:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

Agency: PINECREST PD

Date/Time: 10/22/2009 11:33:18

Data Download Date/Time: 10/22/2009 16:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000910

Agency: NAS WHITING FIELD

Date/Time: 02/27/2009 02:47:38

Data Download Date/Time: 03/25/2009 02:09

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	Fail
Temperature Regulation Test	Fail

Agency: NAS WHITING FIELD

Date/Time: 02/27/2009 03:15:30

Data Download Date/Time: 03/25/2009 02:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000910

Agency: NAS WHITING FIELD

Date/Time: 05/18/2009 10:19:57

Data Download Date/Time: 05/18/2009 14:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

Agency: FORT MYERS PD

Date/Time: 05/08/2009 11:05:51

Data Download Date/Time: 05/12/2009 13:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

Agency: SANIBEL PD

Date/Time: 11/25/2009 09:15:27

Data Download Date/Time: 11/25/2009 13:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

Agency: LEE COUNTY SO

Date/Time: 05/05/2009 08:10:19

Data Download Date/Time: 05/12/2009 09:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LEE COUNTY SO

Date/Time: 05/05/2009 08:11:22

Data Download Date/Time: 05/12/2009 09:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

Agency: LEE COUNTY SO

Date/Time: 05/12/2009 08:30:20

Data Download Date/Time: 05/12/2009 09:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEE COUNTY SO

Date/Time: 05/12/2009 08:31:06

Data Download Date/Time: 05/12/2009 09:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

Agency: COLLIER COUNTY SO

Date/Time: 06/19/2009 13:20:17

Data Download Date/Time: 06/19/2009 17:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

Agency: COLLIER COUNTY SO

Date/Time: 04/13/2009 13:18:42

Data Download Date/Time: 04/14/2009 08:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

Agency: CHARLOTTE COUNTY SO

Date/Time: 12/16/2009 11:30:11

Data Download Date/Time: 12/16/2009 14:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

Agency: CHARLOTTE COUNTY SO

Date/Time: 01/27/2009 16:07:13

Data Download Date/Time: 01/28/2009 11:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CHARLOTTE COUNTY SO

Date/Time: 12/16/2009 11:27:07

Data Download Date/Time: 12/16/2009 13:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

Agency: GLADES COUNTY S.O.

Date/Time: 06/04/2009 11:00:23

Data Download Date/Time: 06/04/2009 14:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

Agency: TALLAHASSEE PD

Date/Time: 03/09/2009 10:23:44

Data Download Date/Time: 03/14/2009 13:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TALLAHASSEE PD

Date/Time: 10/13/2009 15:04:44

Data Download Date/Time: 10/13/2009 18:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

Agency: HAVANA PD

Date/Time: 05/21/2009 11:15:49

Data Download Date/Time: 05/26/2009 14:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HAVANA PD

Date/Time: 06/17/2009 14:26:23

Data Download Date/Time: 06/17/2009 17:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

Agency: HENDRY COUNTY S.O.

Date/Time: 12/14/2009 13:34:59

Data Download Date/Time: 12/29/2009 18:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HENDRY COUNTY S.O.

Date/Time: 12/14/2009 14:33:54

Data Download Date/Time: 12/29/2009 18:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

Agency: CLEWISTON P.D.

Date/Time: 09/15/2009 09:42:23

Data Download Date/Time: 09/15/2009 10:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CLEWISTON P.D.

Date/Time: 10/02/2009 23:29:34

Data Download Date/Time: 10/26/2009 16:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

Agency: CLEWISTON P.D.

Date/Time: 10/02/2009 23:31:32

Data Download Date/Time: 10/26/2009 16:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CLEWISTON P.D.

Date/Time: 10/02/2009 23:34:23

Data Download Date/Time: 10/26/2009 16:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

Agency: CLEWISTON P.D.

Date/Time: 10/02/2009 23:35:25

Data Download Date/Time: 10/26/2009 16:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

Agency: GADSDEN COUNTY SO

Date/Time: 09/16/2009 14:19:23

Data Download Date/Time: 10/02/2009 14:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

Agency: WAKULLA COUNTY SO

Date/Time: 05/18/2009 15:14:54

Data Download Date/Time: 05/19/2009 09:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

Agency: OSCEOLA CORRECTIONS

Date/Time: 02/16/2009 09:38:26

Data Download Date/Time: 02/16/2009 14:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OSCEOLA CORRECTIONS

Date/Time: 02/23/2009 09:53:40

Data Download Date/Time: 02/23/2009 11:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

Agency: OSCEOLA CORRECTIONS

Date/Time: 11/11/2009 13:13:36

Data Download Date/Time: 11/13/2009 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

Agency: OSCEOLA CORRECTIONS

Date/Time: 11/11/2009 13:13:11

Data Download Date/Time: 11/13/2009 10:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 07/31/2009 14:08:16

Data Download Date/Time: 07/31/2009 14:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

Agency: SEMINOLE COUNTY SO

Date/Time: 04/29/2009 08:42:24

Data Download Date/Time: 04/29/2009 12:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SEMINOLE COUNTY S.O.

Date/Time: 09/08/2009 17:37:21

Data Download Date/Time: 09/10/2009 09:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

Agency: SATELLITE BEACH P.D.

Date/Time: 05/21/2009 08:11:17

Data Download Date/Time: 06/04/2009 22:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

Agency: PINELLAS COUNTY SO

Date/Time: 08/05/2009 13:56:33

Data Download Date/Time: 08/06/2009 11:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PINELLAS COUNTY SO

Date/Time: 11/24/2009 13:18:03

Data Download Date/Time: 11/24/2009 16:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

Agency: BREVARD COUNTY S.O.

Date/Time: 08/01/2009 18:56:05

Data Download Date/Time: 08/19/2009 16:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: BREVARD COUNTY S.O.

Date/Time: 08/01/2009 18:58:14

Data Download Date/Time: 08/19/2009 16:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

Agency: BREVARD COUNTY S.O.

Date/Time: 08/01/2009 18:58:57

Data Download Date/Time: 08/19/2009 16:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 09/16/2009 08:25:23

Data Download Date/Time: 09/17/2009 10:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 05/04/2009 16:42:53

Data Download Date/Time: 05/05/2009 10:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HAINES CITY PD

Date/Time: 05/04/2009 17:10:19

Data Download Date/Time: 05/05/2009 10:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 09/10/2009 03:12:26

Data Download Date/Time: 10/07/2009 19:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

Agency: PALM BAY P.D.

Date/Time: 01/07/2009 06:16:16

Data Download Date/Time: 01/07/2009 07:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PALM BAY P.D.

Date/Time: 02/08/2009 08:03:37

Data Download Date/Time: 03/03/2009 06:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

Agency: BREVARD COUNTY S.O.

Date/Time: 02/16/2009 15:55:50

Data Download Date/Time: 02/17/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BREVARD COUNTY S.O.

Date/Time: 02/23/2009 09:56:15

Data Download Date/Time: 02/23/2009 12:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

Agency: BREVARD COUNTY S.O.

Date/Time: 11/17/2009 15:04:22

Data Download Date/Time: 11/18/2009 09:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

Agency: JACKSONVILLE SO

Date/Time: 02/20/2009 06:52:07

Data Download Date/Time: 02/26/2009 13:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: JACKSONVILLE SO

Date/Time: 07/30/2009 07:41:51

Data Download Date/Time: 09/30/2009 14:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

Agency: JACKSONVILLE SO

Date/Time: 07/30/2009 07:49:09

Data Download Date/Time: 09/30/2009 14:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: JACKSONVILLE SO

Date/Time: 07/30/2009 07:50:33

Data Download Date/Time: 09/30/2009 14:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

Agency: JACKSONVILLE SO

Date/Time: 10/06/2009 09:31:56

Data Download Date/Time: 11/05/2009 09:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: JACKSONVILLE SO

Date/Time: 11/05/2009 14:38:40

Data Download Date/Time: 11/13/2009 13:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

Agency: LAKE ALFRED PD

Date/Time: 07/08/2009 14:11:00

Data Download Date/Time: 07/09/2009 09:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE ALFRED PD

Date/Time: 07/23/2009 11:02:52

Data Download Date/Time: 07/23/2009 15:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

Agency: LAKE ALFRED PD

Date/Time: 08/06/2009 11:35:05

Data Download Date/Time: 08/06/2009 15:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 06/02/2009 16:12:09

Data Download Date/Time: 06/03/2009 10:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 06/18/2009 15:23:28

Data Download Date/Time: 06/19/2009 10:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency: CORAL SPRINGS PD

Date/Time: 07/31/2009 13:02:44

Data Download Date/Time: 07/31/2009 16:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CORAL SPRINGS PD

Date/Time: 09/15/2009 09:00:39

Data Download Date/Time: 09/25/2009 09:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Agency: ST PETERSBURG PD

Date/Time: 07/09/2009 09:23:01

Data Download Date/Time: 07/09/2009 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST PETERSBURG PD

Date/Time: 08/21/2009 03:46:20

Data Download Date/Time: 08/27/2009 14:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Agency: ST PETERSBURG PD

Date/Time: 10/19/2009 13:50:46

Data Download Date/Time: 11/04/2009 20:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001052

Agency: JACKSONVILLE SO

Date/Time: 03/23/2009 13:45:44

Data Download Date/Time: 03/31/2009 10:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 03/30/2009 12:10:35

Data Download Date/Time: 03/30/2009 15:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKELAND PD

Date/Time: 07/31/2009 14:07:57

Data Download Date/Time: 07/31/2009 17:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 12/07/2009 17:09:33

Data Download Date/Time: 12/08/2009 14:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency: WINTER HAVEN PD

Date/Time: 01/29/2009 11:04:04

Data Download Date/Time: 01/29/2009 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WINTER HAVEN PD

Date/Time: 02/09/2009 14:36:51

Data Download Date/Time: 02/10/2009 12:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency: GULFPORT PD

Date/Time: 02/05/2009 18:21:38

Data Download Date/Time: 02/23/2009 23:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

Agency: TREASURE ISLAND PD

Date/Time: 04/14/2009 08:38:59

Data Download Date/Time: 04/20/2009 16:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TREASURE ISLAND PD

Date/Time: 06/11/2009 10:04:55

Data Download Date/Time: 06/12/2009 10:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

Agency: TREASURE ISLAND PD

Date/Time: 11/06/2009 09:06:34

Data Download Date/Time: 11/06/2009 10:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency: HOLLYWOOD PD

Date/Time: 03/01/2009 14:26:27

Data Download Date/Time: 03/01/2009 15:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency: NEW PORT RICHEY PD

Date/Time: 04/11/2009 04:22:46

Data Download Date/Time: 04/30/2009 22:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: NEW PORT RICHEY PD

Date/Time: 04/30/2009 22:42:00

Data Download Date/Time: 06/03/2009 22:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency: BROWARD COUNTY SO

Date/Time: 11/25/2009 11:01:56

Data Download Date/Time: 11/25/2009 15:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 06/06/2009 19:01:51

Data Download Date/Time: 06/30/2009 13:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CLEARWATER PD

Date/Time: 12/17/2009 20:21:25

Data Download Date/Time: 12/17/2009 22:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency: DADE CITY PD

Date/Time: 08/03/2009 10:56:52

Data Download Date/Time: 08/19/2009 12:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 09/01/2009 11:31:08

Data Download Date/Time: 09/02/2009 11:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency: ZEPHYRHILLS PD

Date/Time: 09/02/2009 09:39:58

Data Download Date/Time: 09/02/2009 14:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency: ST PETERSBURG PD

Date/Time: 07/07/2009 14:25:05

Data Download Date/Time: 07/07/2009 16:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST PETERSBURG PD

Date/Time: 07/08/2009 15:47:31

Data Download Date/Time: 07/09/2009 12:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency: ST PETERSBURG PD

Date/Time: 10/29/2009 14:48:36

Data Download Date/Time: 10/30/2009 09:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 02/20/2009 15:28:54

Data Download Date/Time: 02/20/2009 18:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LARGO PD

Date/Time: 02/23/2009 12:38:18

Data Download Date/Time: 03/18/2009 11:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 05/26/2009 09:36:39

Data Download Date/Time: 05/26/2009 12:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 06/24/2009 14:34:25

Data Download Date/Time: 06/25/2009 10:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency: FL HIGHWAY PATROL

Date/Time: 03/25/2009 13:18:38

Data Download Date/Time: 03/31/2009 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

Agency: FHP TROOP E MIAMI

Date/Time: 08/31/2009 15:29:06

Data Download Date/Time: 09/01/2009 11:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency: FHP TROOP D - COCOA

Date/Time: 09/21/2009 12:53:17

Data Download Date/Time: 09/24/2009 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency: FL HIGHWAY PATROL

Date/Time: 02/27/2009 10:16:49

Data Download Date/Time: 03/05/2009 16:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/27/2009 10:18:18

Data Download Date/Time: 03/05/2009 16:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	Fail
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency: FL HIGHWAY PATROL

Date/Time: 02/27/2009 10:19:26

Data Download Date/Time: 03/05/2009 16:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 06/12/2009 14:56:39

Data Download Date/Time: 06/24/2009 16:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency: FL HIGHWAY PATROL

Date/Time: 06/24/2009 12:15:27

Data Download Date/Time: 06/24/2009 16:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 08/05/2009 18:03:04

Data Download Date/Time: 08/06/2009 13:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency: DELAND P.D.

Date/Time: 05/20/2009 08:33:17

Data Download Date/Time: 05/20/2009 15:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency: EGLIN AFB

Date/Time: 06/21/2009 02:52:34

Data Download Date/Time: 07/01/2009 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: EGLIN AFB

Date/Time: 06/21/2009 02:53:21

Data Download Date/Time: 07/01/2009 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency: EGLIN AFB

Date/Time: 06/21/2009 02:56:28

Data Download Date/Time: 07/01/2009 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency: PASCO COUNTY SO

Date/Time: 05/20/2009 16:06:51

Data Download Date/Time: 05/21/2009 10:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 06/02/2009 12:27:41

Data Download Date/Time: 06/03/2009 08:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency: PASCO COUNTY SO

Date/Time: 06/19/2009 08:53:50

Data Download Date/Time: 06/19/2009 16:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 06/19/2009 12:14:55

Data Download Date/Time: 06/19/2009 16:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FHP TROOP D - DELAND

Date/Time: 07/16/2009 11:21:05

Data Download Date/Time: 07/16/2009 15:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency: FHP

Date/Time: 04/04/2009 02:10:06

Data Download Date/Time: 04/26/2009 20:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 08/11/2009 16:13:40

Data Download Date/Time: 10/21/2009 10:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency: FHP

Date/Time: 08/12/2009 10:45:24

Data Download Date/Time: 10/21/2009 10:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 10/19/2009 11:01:45

Data Download Date/Time: 10/21/2009 10:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency: FHP

Date/Time: 10/19/2009 12:46:55

Data Download Date/Time: 10/21/2009 10:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 10/20/2009 15:40:58

Data Download Date/Time: 10/21/2009 10:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP F FT MYERS

Date/Time: 07/03/2009 13:44:15

Data Download Date/Time: 08/08/2009 12:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP F FT MYERS

Date/Time: 09/17/2009 10:25:18

Data Download Date/Time: 09/17/2009 12:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 12/04/2009 16:12:26

Data Download Date/Time: 12/08/2009 09:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 12/07/2009 16:41:24

Data Download Date/Time: 12/08/2009 09:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency: PUTNAM COUNTY SO

Date/Time: 08/10/2009 15:49:00

Data Download Date/Time: 08/11/2009 09:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency: PUTNAM COUNTY SO

Date/Time: 12/04/2009 10:55:22

Data Download Date/Time: 12/04/2009 15:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency: VOLUSIA COUNTY S.O.

Date/Time: 11/19/2009 17:03:47

Data Download Date/Time: 11/20/2009 00:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: VOLUSIA COUNTY S.O.

Date/Time: 11/19/2009 17:04:57

Data Download Date/Time: 11/20/2009 00:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

Agency: ST JOHN'S COUNTY SO

Date/Time: 01/26/2009 14:13:37

Data Download Date/Time: 01/26/2009 16:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

Agency: DELAND P.D.

Date/Time: 02/19/2009 10:06:28

Data Download Date/Time: 02/19/2009 15:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001137

Agency: FLAGLER COUNTY SO

Date/Time: 06/29/2009 09:25:26

Data Download Date/Time: 06/29/2009 17:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency: FLAGLER COUNTY SO

Date/Time: 06/30/2009 09:23:47

Data Download Date/Time: 06/29/2009 18:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

Agency: LAKE HELEN P.D.

Date/Time: 10/28/2009 11:46:59

Data Download Date/Time: 10/29/2009 10:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE HELEN P.D.

Date/Time: 10/28/2009 16:53:14

Data Download Date/Time: 10/29/2009 10:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

Agency: MANATEE COUNTY SO

Date/Time: 08/22/2009 19:29:14

Data Download Date/Time: 08/26/2009 10:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 03/28/2009 15:42:40

Data Download Date/Time: 03/30/2009 09:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

Agency: LEON COUNTY

Date/Time: 10/21/2009 14:46:43

Data Download Date/Time: 10/21/2009 17:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

Agency: ORMOND BEACH P.D.

Date/Time: 08/03/2009 09:27:09

Data Download Date/Time: 08/03/2009 14:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 01/20/2009 00:38:36

Data Download Date/Time: 01/20/2009 01:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ORANGE CITY P.D.

Date/Time: 02/16/2009 20:14:10

Data Download Date/Time: 02/16/2009 20:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 02/25/2009 11:39:23

Data Download Date/Time: 03/13/2009 00:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ORANGE CITY P.D.

Date/Time: 03/13/2009 00:10:03

Data Download Date/Time: 03/13/2009 00:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 04/04/2009 14:41:35

Data Download Date/Time: 04/04/2009 15:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ORANGE CITY P.D.

Date/Time: 05/06/2009 20:33:53

Data Download Date/Time: 05/06/2009 21:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 06/03/2009 14:39:50

Data Download Date/Time: 06/03/2009 15:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ORANGE CITY P.D.

Date/Time: 07/17/2009 09:58:46

Data Download Date/Time: 07/17/2009 14:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 07/25/2009 00:18:08

Data Download Date/Time: 07/25/2009 00:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ORANGE CITY P.D.

Date/Time: 08/23/2009 15:52:56

Data Download Date/Time: 08/23/2009 16:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 09/07/2009 01:02:44

Data Download Date/Time: 09/07/2009 01:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ORANGE CITY P.D.

Date/Time: 10/12/2009 00:14:47

Data Download Date/Time: 10/12/2009 00:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 11/03/2009 04:41:55

Data Download Date/Time: 11/03/2009 05:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ORANGE CITY P.D.

Date/Time: 12/15/2009 00:09:46

Data Download Date/Time: 12/15/2009 00:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

Agency: BREVARD COUNTY S.O.

Date/Time: 04/15/2009 08:41:20

Data Download Date/Time: 04/15/2009 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BREVARD COUNTY S.O.

Date/Time: 04/15/2009 08:53:09

Data Download Date/Time: 04/15/2009 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

Agency: BREVARD COUNTY S.O.

Date/Time: 04/15/2009 08:59:22

Data Download Date/Time: 04/15/2009 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BREVARD COUNTY S.O.

Date/Time: 04/15/2009 09:12:33

Data Download Date/Time: 04/15/2009 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

Agency: BREVARD COUNTY S.O.

Date/Time: 04/15/2009 09:21:09

Data Download Date/Time: 04/15/2009 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BREVARD COUNTY S.O.

Date/Time: 12/10/2009 16:22:01

Data Download Date/Time: 12/14/2009 14:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: LADY LAKE P.D.

Date/Time: 04/22/2009 09:06:57

Data Download Date/Time: 04/22/2009 13:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LADY LAKE P.D.

Date/Time: 09/21/2009 10:55:56

Data Download Date/Time: 09/25/2009 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: LADY LAKE P.D.

Date/Time: 09/21/2009 14:12:40

Data Download Date/Time: 09/25/2009 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

Agency: MIAMI PD

Date/Time: 05/07/2009 10:06:50

Data Download Date/Time: 05/07/2009 18:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

Agency: MIAMI PD

Date/Time: 07/24/2009 20:02:47

Data Download Date/Time: 10/14/2009 10:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: MIAMI PD

Date/Time: 07/24/2009 20:03:43

Data Download Date/Time: 10/14/2009 10:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

Agency: BOCA RATON PD

Date/Time: 04/10/2009 11:51:32

Data Download Date/Time: 04/15/2009 14:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: BOCA RATON PD

Date/Time: 04/10/2009 11:52:49

Data Download Date/Time: 04/15/2009 14:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 04/02/2009 21:16:03

Data Download Date/Time: 04/14/2009 11:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

Agency: BOYNTON BEACH PD

Date/Time: 11/16/2009 15:21:44

Data Download Date/Time: 11/17/2009 17:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 02/19/2009 16:22:28

Data Download Date/Time: 02/27/2009 14:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COLLIER COUNTY SO

Date/Time: 02/19/2009 16:38:47

Data Download Date/Time: 02/27/2009 14:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 02/20/2009 14:26:25

Data Download Date/Time: 02/27/2009 14:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COLLIER COUNTY SO

Date/Time: 02/20/2009 14:27:20

Data Download Date/Time: 02/27/2009 14:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 02/20/2009 14:28:25

Data Download Date/Time: 02/27/2009 14:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COLLIER COUNTY SO

Date/Time: 02/20/2009 15:34:49

Data Download Date/Time: 02/27/2009 14:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 10/22/2009 08:00:38

Data Download Date/Time: 10/22/2009 13:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COLLIER COUNTY SO

Date/Time: 11/23/2009 09:18:32

Data Download Date/Time: 11/23/2009 15:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 12/04/2009 09:30:02

Data Download Date/Time: 12/04/2009 14:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency: COLLIER COUNTY SO

Date/Time: 09/02/2009 12:30:33

Data Download Date/Time: 09/02/2009 16:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency: PANAMA CITY PD

Date/Time: 12/05/2009 15:01:06

Data Download Date/Time: 12/16/2009 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency: NAS PENSACOLA PD

Date/Time: 08/01/2009 04:39:40

Data Download Date/Time: 08/22/2009 10:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: NAS PENSACOLA PD

Date/Time: 08/01/2009 04:40:26

Data Download Date/Time: 08/22/2009 10:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency: LEE COUNTY SO

Date/Time: 07/07/2009 14:45:03

Data Download Date/Time: 08/12/2009 07:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency: TYNDALL AFB

Date/Time: 04/22/2009 10:17:20

Data Download Date/Time: 04/23/2009 11:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TYNDALL AFB

Date/Time: 08/14/2009 11:13:54

Data Download Date/Time: 09/02/2009 09:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency: TYNDALL AFB

Date/Time: 08/14/2009 11:14:47

Data Download Date/Time: 09/02/2009 09:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 12/09/2009 15:39:47

Data Download Date/Time: 12/09/2009 18:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001233

Agency: PALM BEACH CO SO

Date/Time: 12/25/2009 04:39:33

Data Download Date/Time: 01/22/2010 11:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency: GROVELAND P.D.

Date/Time: 07/31/2009 18:23:05

Data Download Date/Time: 08/03/2009 16:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GROVELAND P.D.

Date/Time: 08/03/2009 12:17:47

Data Download Date/Time: 08/03/2009 16:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency: EDGEWATER P.D.

Date/Time: 11/24/2009 22:17:02

Data Download Date/Time: 11/30/2009 12:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency: PUTNAM COUNTY SO

Date/Time: 12/04/2009 14:41:56

Data Download Date/Time: 12/07/2009 09:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PUTNAM COUNTY SO

Date/Time: 12/04/2009 14:43:13

Data Download Date/Time: 12/07/2009 09:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency: PORT ORANGE P.D.

Date/Time: 08/05/2009 10:53:14

Data Download Date/Time: 08/05/2009 16:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ORANGE P.D.

Date/Time: 08/20/2009 12:39:24

Data Download Date/Time: 08/20/2009 13:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency: PORT ORANGE P.D.

Date/Time: 08/20/2009 12:44:51

Data Download Date/Time: 08/20/2009 13:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency: SOUTH DAYTONA P.D.

Date/Time: 07/31/2009 08:10:46

Data Download Date/Time: 07/31/2009 11:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SOUTH DAYTONA P.D.

Date/Time: 09/17/2009 12:36:05

Data Download Date/Time: 09/17/2009 15:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency: SOUTH DAYTONA P.D.

Date/Time: 09/30/2009 16:33:16

Data Download Date/Time: 10/14/2009 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency: BREVARD S.O.

Date/Time: 03/09/2009 13:19:42

Data Download Date/Time: 04/02/2009 17:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency: ROCKLEDGE P.D.

Date/Time: 11/23/2009 16:14:02

Data Download Date/Time: 11/24/2009 12:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency: PALM BAY P.D.

Date/Time: 01/07/2009 06:15:31

Data Download Date/Time: 01/07/2009 07:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY P.D.

Date/Time: 02/08/2009 08:05:35

Data Download Date/Time: 02/08/2009 10:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency: PALM BAY P.D.

Date/Time: 04/29/2009 06:27:17

Data Download Date/Time: 04/29/2009 07:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency: PALM BAY P.D.

Date/Time: 01/07/2009 06:13:16

Data Download Date/Time: 01/07/2009 07:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY P.D.

Date/Time: 02/08/2009 08:04:37

Data Download Date/Time: 02/08/2009 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency: PALM BAY P.D.

Date/Time: 04/29/2009 06:25:05

Data Download Date/Time: 04/29/2009 07:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PALM BAY P.D.

Date/Time: 04/29/2009 06:33:06

Data Download Date/Time: 04/29/2009 07:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency: PALM BAY P.D.

Date/Time: 10/14/2009 22:32:47

Data Download Date/Time: 09/17/2010 16:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	Fail
Temperature Regulation Test	OK

Agency: PALM BAY P.D.

Date/Time: 10/14/2009 22:33:27

Data Download Date/Time: 09/17/2010 16:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	Fail
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001269

Agency: PALM BEACH CO SO

Date/Time: 04/30/2009 15:03:30

Data Download Date/Time: 05/15/2009 12:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PALM BEACH CO SO

Date/Time: 04/30/2009 15:16:20

Data Download Date/Time: 05/15/2009 12:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001269

Agency: PALM BEACH CO SO

Date/Time: 09/01/2009 08:47:30

Data Download Date/Time: 09/02/2009 08:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency: SEMINOLE COUNTY S.O.

Date/Time: 03/31/2009 16:31:33

Data Download Date/Time: 04/01/2009 10:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SEMINOLE COUNTY S.O.

Date/Time: 06/25/2009 14:23:20

Data Download Date/Time: 06/25/2009 17:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency: SEMINOLE COUNTY S.O.

Date/Time: 06/25/2009 14:34:48

Data Download Date/Time: 06/25/2009 17:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency: PINELLAS COUNTY SO

Date/Time: 06/04/2009 15:18:31

Data Download Date/Time: 06/04/2009 17:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PINELLAS COUNTY SO

Date/Time: 06/18/2009 11:28:32

Data Download Date/Time: 06/22/2009 12:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency: PINELLAS COUNTY SO

Date/Time: 06/22/2009 12:35:39

Data Download Date/Time: 06/22/2009 12:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PINELLAS COUNTY SO

Date/Time: 07/16/2009 09:53:27

Data Download Date/Time: 07/16/2009 15:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency: PINELLAS COUNTY SO

Date/Time: 09/10/2009 10:31:18

Data Download Date/Time: 09/10/2009 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency: HIALEAH GARDENS PD

Date/Time: 02/18/2009 08:12:39

Data Download Date/Time: 02/18/2009 11:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency: MIAMI-DADE PD

Date/Time: 03/06/2009 01:03:22

Data Download Date/Time: 03/19/2009 17:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MIAMI-DADE PD

Date/Time: 10/26/2009 13:17:59

Data Download Date/Time: 11/10/2009 13:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency: JACKSONVILLE SO

Date/Time: 01/21/2009 09:37:58

Data Download Date/Time: 02/26/2009 12:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: JACKSONVILLE SO

Date/Time: 01/21/2009 09:39:04

Data Download Date/Time: 02/26/2009 12:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

Agency: NASSAU COUNTY SO

Date/Time: 06/08/2009 05:29:08

Data Download Date/Time: 06/12/2009 15:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: NASSAU COUNTY SO

Date/Time: 06/09/2009 10:46:04

Data Download Date/Time: 06/12/2009 15:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

Agency: UNIVERSITY OF FL PD

Date/Time: 02/18/2009 16:09:51

Data Download Date/Time: 02/19/2009 09:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: UNIVERSITY OF FL PD

Date/Time: 11/15/2009 08:08:18

Data Download Date/Time: 11/15/2009 09:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

Agency: HIGH SPRINGS PD

Date/Time: 11/23/2009 13:05:44

Data Download Date/Time: 11/23/2009 17:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/02/2009 21:00:31

Data Download Date/Time: 01/02/2009 22:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/02/2009 21:01:50

Data Download Date/Time: 01/02/2009 22:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 02/13/2009 20:49:49

Data Download Date/Time: 02/13/2009 22:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 03/20/2009 21:35:20

Data Download Date/Time: 03/20/2009 23:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 04/24/2009 20:08:30

Data Download Date/Time: 04/24/2009 21:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 05/08/2009 20:32:31

Data Download Date/Time: 05/08/2009 22:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 06/19/2009 19:00:39

Data Download Date/Time: 06/19/2009 20:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 07/29/2009 20:22:59

Data Download Date/Time: 07/29/2009 21:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 08/31/2009 14:20:59

Data Download Date/Time: 08/31/2009 15:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/29/2009 10:54:36

Data Download Date/Time: 09/29/2009 12:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 10/20/2009 13:35:37

Data Download Date/Time: 10/20/2009 15:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 11/30/2009 12:41:46

Data Download Date/Time: 11/30/2009 14:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 12/28/2009 12:10:15

Data Download Date/Time: 12/28/2009 14:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/02/2009 19:15:44

Data Download Date/Time: 01/02/2009 20:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 02/13/2009 17:54:01

Data Download Date/Time: 02/13/2009 19:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 03/20/2009 19:13:09

Data Download Date/Time: 03/20/2009 20:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 04/24/2009 17:22:30

Data Download Date/Time: 04/24/2009 18:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 05/08/2009 17:59:14

Data Download Date/Time: 05/08/2009 19:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 06/19/2009 17:16:30

Data Download Date/Time: 06/19/2009 18:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 07/29/2009 18:19:49

Data Download Date/Time: 07/29/2009 19:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 08/31/2009 12:41:12

Data Download Date/Time: 08/31/2009 14:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/18/2009 12:48:45

Data Download Date/Time: 09/25/2009 10:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/25/2009 09:42:45

Data Download Date/Time: 09/25/2009 10:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	Fail
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/25/2009 09:48:40

Data Download Date/Time: 09/29/2009 10:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	Fail
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/25/2009 09:59:51

Data Download Date/Time: 09/29/2009 10:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 09/29/2009 09:03:13

Data Download Date/Time: 09/29/2009 10:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 10/20/2009 12:16:15

Data Download Date/Time: 10/20/2009 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 11/30/2009 10:44:23

Data Download Date/Time: 11/30/2009 12:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SANTA ROSA COUNTY SO

Date/Time: 11/30/2009 11:25:27

Data Download Date/Time: 12/28/2009 14:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 07/31/2009 16:21:23

Data Download Date/Time: 07/31/2009 19:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

Agency: BAY COUNTY SO

Date/Time: 03/23/2009 09:15:56

Data Download Date/Time: 03/23/2009 14:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 12/08/2009 10:05:05

Data Download Date/Time: 12/08/2009 14:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

Agency: OKEECHOBEE CO SO

Date/Time: 10/11/2009 04:01:24

Data Download Date/Time: 10/15/2009 15:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

Agency: OKEECHOBEE COUNTY SO

Date/Time: 10/11/2009 04:18:43

Data Download Date/Time: 10/15/2009 15:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OKEECHOBEE COUNTY SO

Date/Time: 10/11/2009 04:24:38

Data Download Date/Time: 10/15/2009 15:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

Agency: OKEECHOBEE PD

Date/Time: 04/27/2009 15:41:11

Data Download Date/Time: 04/27/2009 16:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

Agency: DESOTO COUNTY SO

Date/Time: 03/28/2009 00:28:52

Data Download Date/Time: 04/29/2009 00:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

Agency: DESOTO COUNTY SO

Date/Time: 03/12/2009 23:27:49

Data Download Date/Time: 04/28/2009 17:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: DESOTO COUNTY SO

Date/Time: 04/28/2009 13:42:11

Data Download Date/Time: 04/28/2009 17:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

Agency: SARASOTA COUNTY SO

Date/Time: 03/04/2009 13:58:57

Data Download Date/Time: 03/21/2009 00:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SARASOTA COUNTY SO

Date/Time: 06/03/2009 15:52:47

Data Download Date/Time: 06/15/2009 10:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

Agency: SARASOTA COUNTY SO

Date/Time: 01/28/2009 12:39:49

Data Download Date/Time: 01/28/2009 15:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency: SARASOTA COUNTY SO

Date/Time: 06/25/2009 11:41:13

Data Download Date/Time: 06/25/2009 16:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency: EVERGLADES NAT PARK

Date/Time: 07/25/2009 15:20:55

Data Download Date/Time: 07/25/2009 16:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: EVERGLADES NAT PARK

Date/Time: 07/25/2009 15:21:45

Data Download Date/Time: 07/25/2009 16:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency: EVERGLADES NAT PARK

Date/Time: 11/08/2009 15:43:03

Data Download Date/Time: 12/28/2009 16:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/11/2009 07:17:03

Data Download Date/Time: 03/11/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CHARLOTTE COUNTY SO

Date/Time: 12/16/2009 11:29:52

Data Download Date/Time: 12/16/2009 13:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency: ORANGE COUNTY S.O.

Date/Time: 10/15/2009 08:53:19

Data Download Date/Time: 10/15/2009 11:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001441

Agency: KENNEDY SPACE CENTER

Date/Time: 03/27/2009 09:40:41

Data Download Date/Time: 03/27/2009 13:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNEDY SPACE CENTER

Date/Time: 06/05/2009 08:55:40

Data Download Date/Time: 06/05/2009 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001441

Agency: KENNEDY SPACE CENTER

Date/Time: 06/18/2009 13:57:29

Data Download Date/Time: 06/19/2009 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001469

Agency: HURLBURT FIELD PD

Date/Time: 03/01/2009 08:37:19

Data Download Date/Time: 03/01/2009 11:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency: CAPE CORAL PD

Date/Time: 07/07/2009 13:13:20

Data Download Date/Time: 07/07/2009 15:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Agency: KENNETH CITY PD

Date/Time: 01/30/2009 10:57:12

Data Download Date/Time: 01/30/2009 16:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNETH CITY PD

Date/Time: 03/21/2009 18:15:55

Data Download Date/Time: 03/23/2009 08:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Agency: KENNETH CITY PD

Date/Time: 04/19/2009 23:45:50

Data Download Date/Time: 05/05/2009 16:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNETH CITY PD

Date/Time: 06/17/2009 11:56:45

Data Download Date/Time: 06/17/2009 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Agency: KENNETH CITY PD

Date/Time: 08/17/2009 14:40:58

Data Download Date/Time: 08/31/2009 20:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001650

Agency: KEY BISCAYNE PD

Date/Time: 10/12/2009 13:04:30

Data Download Date/Time: 10/12/2009 17:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency: ST PETERSBURG PD

Date/Time: 08/03/2009 14:47:08

Data Download Date/Time: 08/05/2009 10:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST PETERSBURG PD

Date/Time: 10/13/2009 08:15:49

Data Download Date/Time: 10/13/2009 16:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency: OSCEOLA CORRECTIONS

Date/Time: 07/04/2009 17:21:45

Data Download Date/Time: 07/20/2009 15:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OSCEOLA CORRECTIONS

Date/Time: 07/20/2009 09:24:58

Data Download Date/Time: 07/20/2009 15:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency: OSCEOLA CORRECTIONS

Date/Time: 07/20/2009 09:27:11

Data Download Date/Time: 07/20/2009 15:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OSCEOLA CORRECTIONS

Date/Time: 07/20/2009 12:15:34

Data Download Date/Time: 07/20/2009 15:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency: PALM BAY P.D.

Date/Time: 01/07/2009 06:14:50

Data Download Date/Time: 01/07/2009 07:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY P.D.

Date/Time: 02/08/2009 08:03:29

Data Download Date/Time: 02/08/2009 10:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency: LEE COUNTY SO

Date/Time: 05/12/2009 13:45:47

Data Download Date/Time: 06/11/2009 13:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEE COUNTY SO

Date/Time: 10/12/2009 10:45:43

Data Download Date/Time: 10/12/2009 15:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: KENNETH CITY P.D.

Date/Time: 06/17/2009 09:13:41

Data Download Date/Time: 06/17/2009 14:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNETH CITY P.D.

Date/Time: 12/01/2009 16:37:36

Data Download Date/Time: 12/09/2009 17:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: KENNETH CITY P.D.

Date/Time: 12/09/2009 14:06:13

Data Download Date/Time: 12/09/2009 17:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY SO

Date/Time: 02/04/2009 09:45:08

Data Download Date/Time: 02/04/2009 14:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:33:28

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:34:33

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:35:09

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:35:40

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:36:09

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:36:38

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:37:07

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:37:37

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:38:06

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY SO

Date/Time: 03/21/2009 07:38:34

Data Download Date/Time: 04/08/2009 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CHARLOTTE COUNTY SO

Date/Time: 04/23/2009 17:07:30

Data Download Date/Time: 05/21/2009 11:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY SO

Date/Time: 12/16/2009 11:29:12

Data Download Date/Time: 12/16/2009 13:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 04/01/2009 13:17:03

Data Download Date/Time: 04/02/2009 08:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 04/29/2009 13:13:08

Data Download Date/Time: 04/29/2009 17:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 01/16/2009 07:20:26

Data Download Date/Time: 01/16/2009 16:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 02/02/2009 09:17:40

Data Download Date/Time: 02/06/2009 17:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 02/02/2009 09:19:39

Data Download Date/Time: 02/06/2009 17:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 02/06/2009 08:13:53

Data Download Date/Time: 02/06/2009 17:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 02/06/2009 15:29:06

Data Download Date/Time: 02/06/2009 17:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 03/06/2009 17:36:38

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 03/06/2009 17:37:10

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:02:49

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:03:19

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:04:27

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:07:09

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:36:34

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:38:27

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:39:08

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:39:48

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:43:16

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 03/16/2009 13:58:02

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 03/19/2009 08:22:51

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 03/19/2009 08:25:27

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 04/21/2009 08:39:57

Data Download Date/Time: 04/21/2009 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 06/18/2009 08:29:35

Data Download Date/Time: 06/18/2009 11:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 06/24/2009 10:24:24

Data Download Date/Time: 06/24/2009 15:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 07/29/2009 09:13:41

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 08/03/2009 12:55:13

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 08/03/2009 16:44:38

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 08/11/2009 16:43:07

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 08/11/2009 16:43:50

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 08/11/2009 16:45:13

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 08/12/2009 07:53:23

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 08/12/2009 08:48:39

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 08/12/2009 08:51:25

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 08/12/2009 08:52:12

Data Download Date/Time: 08/12/2009 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 09/03/2009 15:50:32

Data Download Date/Time: 09/04/2009 10:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 09/04/2009 06:40:30

Data Download Date/Time: 09/04/2009 10:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 09/04/2009 06:46:58

Data Download Date/Time: 09/04/2009 10:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 09/04/2009 08:35:17

Data Download Date/Time: 09/04/2009 10:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 10/15/2009 08:24:38

Data Download Date/Time: 10/15/2009 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 10/15/2009 08:25:27

Data Download Date/Time: 10/15/2009 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 10/15/2009 08:26:29

Data Download Date/Time: 10/15/2009 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: US AIR FORCE MACDILL

Date/Time: 10/15/2009 08:29:43

Data Download Date/Time: 10/15/2009 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 10/15/2009 08:35:25

Data Download Date/Time: 10/15/2009 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: US AIR FORCE MACDILL

Date/Time: 10/15/2009 08:37:26

Data Download Date/Time: 10/15/2009 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001866

Agency: US AIR FORCE MACDILL

Date/Time: 10/29/2009 08:22:30

Data Download Date/Time: 04/05/2010 14:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency: NAS KEY WEST

Date/Time: 12/09/2009 04:22:37

Data Download Date/Time: 12/21/2009 12:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency: NEW SYMRNA BEACH PD

Date/Time: 02/07/2009 20:47:09

Data Download Date/Time: 02/27/2009 19:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: NEW SYMRNA BEACH PD

Date/Time: 08/11/2009 13:29:16

Data Download Date/Time: 08/30/2009 09:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency: NEW SYMRNA BEACH PD

Date/Time: 10/09/2009 13:30:08

Data Download Date/Time: 05/21/2010 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: NEW SYMRNA BEACH PD

Date/Time: 10/09/2009 13:30:59

Data Download Date/Time: 05/21/2010 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency: NEW SYMRNA BEACH PD

Date/Time: 10/09/2009 13:31:48

Data Download Date/Time: 05/21/2010 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency: HIALEAH PD

Date/Time: 06/09/2009 15:36:45

Data Download Date/Time: 06/29/2009 11:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HIALEAH PD

Date/Time: 06/09/2009 15:52:15

Data Download Date/Time: 06/29/2009 11:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency: HIALEAH PD

Date/Time: 06/12/2009 20:05:10

Data Download Date/Time: 06/29/2009 11:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HIALEAH PD

Date/Time: 06/12/2009 20:07:01

Data Download Date/Time: 06/29/2009 11:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency: HIALEAH PD

Date/Time: 06/12/2009 22:39:25

Data Download Date/Time: 06/29/2009 11:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency: MIAMI GARDENS PD

Date/Time: 09/30/2009 12:11:49

Data Download Date/Time: 09/30/2009 15:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

Agency: MIAMI GARDENS PD

Date/Time: 09/30/2009 14:24:46

Data Download Date/Time: 09/30/2009 17:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency: MIAMI BEACH PD

Date/Time: 10/12/2009 13:53:48

Data Download Date/Time: 10/28/2009 08:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 04/11/2009 03:47:38

Data Download Date/Time: 05/11/2009 08:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency: HILLSBOROUGH CO SO

Date/Time: 09/15/2009 15:00:22

Data Download Date/Time: 09/16/2009 09:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency: HILLSBOROUGH CO SO

Date/Time: 09/16/2009 11:42:39

Data Download Date/Time: 09/17/2009 09:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HILLSBOROUGH CO SO

Date/Time: 12/08/2009 14:43:20

Data Download Date/Time: 12/08/2009 18:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency: HILLSBOROUGH CO SO

Date/Time: 12/08/2009 15:16:52

Data Download Date/Time: 12/08/2009 18:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency: OSCEOLA COUNTY CORR

Date/Time: 10/29/2009 13:14:46

Data Download Date/Time: 10/29/2009 16:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency: CMI INC

Date/Time: 10/28/2009 14:09:44

Data Download Date/Time: 11/02/2009 09:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CMI INC

Date/Time: 10/28/2009 14:35:10

Data Download Date/Time: 11/02/2009 09:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK
