

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: FREY, ROBERT, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: LAVANTURE, KEVIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 01:10

Gender: Female

Observation Period Began: 01:28

Subject's Last Name: BANKS

Date/Time of Test: 08/03/2008 02:00

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Subject Sample #1	0.190	01:55
Breath Volume	1.523	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:57
Subject Sample #2	0.191	01:58
Breath Volume	1.839	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MASSOLINI, CHAD, E

Breath Test Operator Agency: Brevard County SO

Arresting Officer: MASSOLINI, CHAD

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 22:01

Gender: Male

Observation Period Began: 22:01

Subject's Last Name: KENNEDY

Date/Time of Test: 08/04/2008 23:13

---

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:08	
Control Test	0.080	23:08	
Air Blank	0.000	23:09	
Subject Sample #1		23:11	Subject Test Refused
Breath Volume	0.152	23:11	
Air Blank	0.000	23:12	
Control Test	0.082	23:12	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: ZIMMERMAN, SCOTT, M

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: RING, URSULA

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 23:32

Gender: Male

Observation Period Began: 23:33

Subject's Last Name: BURRIS

Date/Time of Test: 08/05/2008 00:12

---

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Subject Sample #1	0.209	00:08
Breath Volume	1.398	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample #2	0.223	00:11
Breath Volume	1.574	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MASSOLINI, CHAD, E

Breath Test Operator Agency: Brevard County SO

Arresting Officer: MASSOLINI, CHAD

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 02:32

Gender: Male

Observation Period Began: 02:32

Subject's Last Name: WILLIAMS

Date/Time of Test: 08/07/2008 03:04

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Subject Sample #1	0.159	02:59
Breath Volume	1.445	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample #2	0.151	03:03
Breath Volume	1.640	03:03
Air Blank	0.000	03:03
Control Test	0.078	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, D

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 01:40

Gender: Female

Observation Period Began: 01:40

Subject's Last Name: LEBLANC-HAWKINS

Date/Time of Test: 08/08/2008 02:15

---

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:05	
Subject Sample #1	0.132	02:08	Volume Not Met
Breath Volume	0.121	02:08	
Air Blank	0.000	02:09	
Air Blank	0.000	02:11	
Subject Sample #2	0.171	02:14	Slope Not Level
Breath Volume	2.093	02:14	
Air Blank	0.000	02:14	
Control Test	0.079	02:15	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, D

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 01:40

Gender: Female

Observation Period Began: 01:40

Subject's Last Name: LEBLANC-HAWKINS

Date/Time of Test: 08/08/2008 02:27

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:19
Subject Sample #1	0.172	02:20
Breath Volume	1.410	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample #2	0.176	02:25
Breath Volume	1.410	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: HARRISON, DEP.

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 17:39

Gender: Male

Observation Period Began: 17:39

Subject's Last Name: QUINN

Date/Time of Test: 08/11/2008 18:41

---

Diagnostics Check	OK	18:34
Air Blank	0.000	18:35
Control Test	0.080	18:35
Air Blank	0.000	18:35
Subject Sample #1	0.110	18:36
Breath Volume	1.781	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample #2	0.103	18:40
Breath Volume	1.902	18:40
Air Blank	0.000	18:41
Control Test	0.079	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: ZIMMERMAN, SCOTT, M

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: RING, URSULA

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 22:10

Gender: Female

Observation Period Began: 22:10

Subject's Last Name: FRANCK

Date/Time of Test: 08/22/2008 23:18

---

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.080	23:05	
Air Blank	0.000	23:05	
Subject Sample #1	0.338	23:06	No .020 agreement
Breath Volume	1.878	23:06	
Air Blank	0.000	23:07	
Air Blank	0.000	23:08	
Subject Sample #2	0.269	23:12	Volume Not Met
Breath Volume	0.152	23:12	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample #3	0.290	23:16	No .020 agreement
Breath Volume	1.230	23:16	
Air Blank	0.000	23:17	
Control Test	0.078	23:18	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:18	

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: ZIMMERMAN, SCOTT, M

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: VALENTI, NEAL

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 17:33

Gender: Male

Observation Period Began: 17:40

Subject's Last Name: FISHER

Date/Time of Test: 08/23/2008 18:13

---

Diagnostics Check	OK	18:07
Air Blank	0.000	18:08
Control Test	0.080	18:08
Air Blank	0.000	18:08
Subject Sample #1	0.238	18:09
Breath Volume	2.363	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:11
Subject Sample #2	0.246	18:12
Breath Volume	2.113	18:12
Air Blank	0.000	18:12
Control Test	0.079	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MASSOLINI, CHAD, E

Breath Test Operator Agency: Brevard County SO

Arresting Officer: MASSOLINI, CHAD

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 07/14/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 00:50

Gender: Female

Observation Period Began: 00:50

Subject's Last Name: WEBSTER

Date/Time of Test: 08/27/2008 01:55

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.169	01:48
Breath Volume	1.695	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample #2	0.164	01:53
Breath Volume	1.867	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

---

Data Download Date/Time: 08/29/2008 01:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEE, SELENA, A

Breath Test Operator Agency: Lee County SO

Arresting Officer: LEE, SELENA

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: Yes

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 01:00

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: ALONSO

Date/Time of Test: 08/01/2008 02:18

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample #1	0.186	02:12
Breath Volume	2.410	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.192	02:16
Breath Volume	2.527	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SINIBALDI, ANTHONY, R

Breath Test Operator Agency: Lee County SO

Arresting Officer: JUSTHAM, RYAN

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 03:42

Gender: Male

Observation Period Began: 04:05

Subject's Last Name: WHITE

Date/Time of Test: 08/01/2008 04:36

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample #1	0.139	04:30
Breath Volume	1.402	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample #2	0.133	04:34
Breath Volume	2.574	04:34
Air Blank	0.000	04:35
Control Test	0.076	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: RONGISH, SHERRY

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 00:47

Gender: Male

Observation Period Began: 01:45

Subject's Last Name: GONZALEZ

Date/Time of Test: 08/02/2008 02:20

---

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Subject Sample #1	0.221	02:15	
Breath Volume	2.246	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:17	
Subject Sample #2		02:19	Subject Test Refused
Breath Volume	0.972	02:19	
Air Blank	0.000	02:19	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: MCDANIEL, JAMES

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 01:29

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: JOSE

Date/Time of Test: 08/02/2008 03:13

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Subject Sample #1	0.224	03:08
Breath Volume	2.433	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample #2	0.215	03:11
Breath Volume	1.570	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: YATES, SHAWN

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 02:01

Gender: Male

Observation Period Began: 02:34

Subject's Last Name: CAMPBELL

Date/Time of Test: 08/02/2008 03:23

---

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Subject Sample #1	0.032	03:18
Breath Volume	1.375	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample #2	0.035	03:21
Breath Volume	2.171	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: GUMM, BRIAN

Arresting Officer Agency: Cape Coral PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 23:29

Gender: Male

Observation Period Began: 23:59

Subject's Last Name: BARRESI

Date/Time of Test: 08/03/2008 00:33

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample #1	0.042	00:28
Breath Volume	1.996	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample #2	0.043	00:31
Breath Volume	2.339	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: MARGRANDER, ROBERT

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 23:01

Gender: Male

Observation Period Began: 00:01

Subject's Last Name: GACCIONE

Date/Time of Test: 08/03/2008 00:41

---

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample #1	0.194	00:37
Breath Volume	2.917	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample #2	0.197	00:40
Breath Volume	2.546	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: THORPE, ASHLEY

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 01:00

Gender: Male

Observation Period Began: 01:45

Subject's Last Name: RESENDIZ GALLEGOS

Date/Time of Test: 08/03/2008 02:28

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample #1	0.169	02:23
Breath Volume	2.074	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample #2	0.158	02:26
Breath Volume	1.367	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: WILLIAMS, ALLEN

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 19:18

Gender: Female

Observation Period Began: 20:25

Subject's Last Name: PRICE

Date/Time of Test: 08/03/2008 21:06

---

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	20:59
Subject Sample #1	0.198	21:01
Breath Volume	2.894	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample #2	0.191	21:04
Breath Volume	2.625	21:04
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:06

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: SPRAGUE, JASON

Arresting Officer Agency: Cape Coral PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 20:07

Gender: Male

Observation Period Began: 20:40

Subject's Last Name: MCCARTHY

Date/Time of Test: 08/03/2008 21:14

---

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.080	21:09
Air Blank	0.000	21:09
Subject Sample #1	0.057	21:10
Breath Volume	4.312	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:12
Subject Sample #2	0.053	21:13
Breath Volume	2.937	21:13
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, ROBERT, T

Breath Test Operator Agency: Lee County SO

Arresting Officer: SHAW, ROBERT

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 22:15

Gender: Female

Observation Period Began: 23:12

Subject's Last Name: ETEO

Date/Time of Test: 08/03/2008 23:53

---

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.079	23:45	
Air Blank	0.000	23:45	
Subject Sample #1	0.181	23:48	Volume Not Met
Breath Volume	0.066	23:48	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample #2		23:52	Subject Test Refused
Breath Volume	1.058	23:52	
Air Blank	0.000	23:52	
Control Test	0.078	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SINIBALDI, ANTHONY, R

Breath Test Operator Agency: Lee County SO

Arresting Officer: NAVA, JERRY

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 19:54

Gender: Male

Observation Period Began: 20:05

Subject's Last Name: DAVIS

Date/Time of Test: 08/04/2008 21:23

---

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Subject Sample #1	0.037	21:18
Breath Volume	2.144	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample #2	0.036	21:21
Breath Volume	1.558	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: WAYLAND, JOSEPH, T

Breath Test Operator Agency: Cape Coral PD

Arresting Officer: HURD, ERIK

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 01:50

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: BROCKINGTON

Date/Time of Test: 08/05/2008 03:33

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Subject Sample #1		03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:33	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: SANDT, JASON

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 01:10

Gender: Male

Observation Period Began: 01:45

Subject's Last Name: ORAVETZ

Date/Time of Test: 08/07/2008 02:25

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample #1	0.220	02:20
Breath Volume	2.132	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample #2	0.221	02:23
Breath Volume	2.230	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: DOWNEY, ERIC, A

Breath Test Operator Agency: Lee County SO

Arresting Officer: MOORE, JASON

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 05:49

Gender: Male

Observation Period Began: 06:09

Subject's Last Name: BEAL

Date/Time of Test: 08/07/2008 06:47

---

Diagnostics Check	OK	06:34	
Air Blank	0.000	06:35	
Control Test	0.080	06:35	
Air Blank	0.000	06:36	
Subject Sample #1	0.073	06:39	Volume Not Met
Breath Volume	0.281	06:39	
Air Blank	0.000	06:40	
Air Blank	0.000	06:41	
Subject Sample #2	0.082	06:42	
Breath Volume	1.125	06:42	
Air Blank	0.000	06:43	
Air Blank	0.000	06:45	
Subject Sample #3	0.079	06:45	
Breath Volume	1.191	06:45	
Air Blank	0.000	06:46	
Control Test	0.079	06:46	
Air Blank	0.000	06:47	
Diagnostics Check	OK	06:47	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: SPRAGUE, JASON

Arresting Officer Agency: Cape Coral PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 02:43

Gender: Female

Observation Period Began: 03:05

Subject's Last Name: BICKERSTAFF

Date/Time of Test: 08/08/2008 03:33

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample #1	0.229	03:28
Breath Volume	1.941	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample #2	0.234	03:32
Breath Volume	2.644	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SINIBALDI, ANTHONY, R

Breath Test Operator Agency: Lee County SO

Arresting Officer: BOCK, ANDREW

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 19:35

Gender: Male

Observation Period Began: 19:35

Subject's Last Name: GUTIERREZ

Date/Time of Test: 08/08/2008 20:01

---

Diagnostics Check	OK	19:56	
Air Blank	0.000	19:56	
Control Test	0.079	19:56	
Air Blank	0.000	19:57	
Subject Sample #1		20:00	Subject Test Refused
Breath Volume	0.019	20:00	
Air Blank		20:01	Purge Fail
Air Blank	0.000	20:01	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SINIBALDI, ANTHONY, R

Breath Test Operator Agency: Lee County SO

Arresting Officer: RABITT, ROBERT

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: DUI

Arrest Time: 20:01

Gender: Male

Observation Period Began: 20:35

Subject's Last Name: DELAPORTE

Date/Time of Test: 08/08/2008 21:20

---

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:12	
Control Test	0.079	21:12	
Air Blank	0.000	21:12	
Subject Sample #1	0.194	21:14	
Breath Volume	1.609	21:14	
Air Blank	0.000	21:14	
Air Blank	0.000	21:16	
Subject Sample #2	0.181	21:19	
Breath Volume	1.207	21:19	
Air Blank		21:20	Purge Fail
Air Blank	0.000	21:20	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: TESTER, T

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8016011

Dry Gas Expiry Date: 01/18/2010

Violation Code: SYS CHECK

Arrest Time: 07:00

Gender: Male

Observation Period Began: 07:00

Subject's Last Name: TESTEE

Date/Time of Test: 08/11/2008 07:39

---

Diagnostics Check	OK	07:33
Air Blank	0.000	07:34
Control Test	0.080	07:34
Air Blank	0.000	07:34
Subject Sample #1	0.000	07:35
Breath Volume	1.839	07:35
Air Blank	0.000	07:35
Air Blank	0.000	07:37
Subject Sample #2	0.000	07:38
Breath Volume	1.417	07:38
Air Blank	0.000	07:38
Control Test	0.079	07:38
Air Blank	0.000	07:39
Diagnostics Check	OK	07:39

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: EHLERS, DAVID

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 00:53

Gender: Male

Observation Period Began: 01:39

Subject's Last Name: DURST

Date/Time of Test: 08/12/2008 02:08

---

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.081	02:04	
Air Blank	0.000	02:04	
Subject Sample #1		02:06	Subject Test Refused
Breath Volume	1.078	02:06	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SCOTT, MICHAEL, T

Breath Test Operator Agency: FHP Troop F

Arresting Officer: SCOTT, MICHAEL

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 16:09

Gender: Male

Observation Period Began: 17:00

Subject's Last Name: THIBAUT

Date/Time of Test: 08/12/2008 18:47

---

Diagnostics Check	OK	18:40
Air Blank	0.000	18:40
Control Test	0.080	18:41
Air Blank	0.000	18:41
Subject Sample #1	0.136	18:42
Breath Volume	1.578	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:44
Subject Sample #2	0.141	18:45
Breath Volume	4.046	18:45
Air Blank	0.000	18:46
Control Test	0.079	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MILLER, MATTHEW,

Breath Test Operator Agency: Lee County SO

Arresting Officer: ZECTMAN, R

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 15:15

Gender: Female

Observation Period Began: 16:20

Subject's Last Name: FORD

Date/Time of Test: 08/13/2008 16:52

---

Diagnostics Check	OK	16:44
Air Blank	0.000	16:45
Control Test	0.080	16:45
Air Blank	0.000	16:45
Subject Sample #1	0.274	16:46
Breath Volume	1.613	16:46
Air Blank	0.000	16:47
Air Blank	0.000	16:49
Subject Sample #2	0.258	16:50
Breath Volume	1.460	16:50
Air Blank	0.000	16:51
Control Test	0.080	16:51
Air Blank	0.000	16:52
Diagnostics Check	OK	16:52

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEE, SELENA, A

Breath Test Operator Agency: Lee County SO

Arresting Officer: LEE, SELENA

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 02:04

Gender: Male

Observation Period Began: 03:06

Subject's Last Name: MCCULLEY

Date/Time of Test: 08/14/2008 03:34

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample #1	0.177	03:29
Breath Volume	2.140	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.182	03:32
Breath Volume	1.878	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: BRUNO, THOMAS, M

Breath Test Operator Agency: Lee County SO

Arresting Officer: ZELTMAN, RICHARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 15:29

Gender: Male

Observation Period Began: 16:05

Subject's Last Name: ALFARO

Date/Time of Test: 08/14/2008 16:40

---

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.081	16:31
Air Blank	0.000	16:31
Subject Sample #1	0.312	16:32
Breath Volume	4.765	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:34
Subject Sample #2	0.284	16:35
Breath Volume	3.414	16:35
Air Blank	0.000	16:36
Air Blank	0.000	16:37
Subject Sample #3	0.315	16:38
Breath Volume	5.460	16:38
Air Blank	0.000	16:39
Control Test	0.078	16:39
Air Blank	0.000	16:40
Diagnostics Check	OK	16:40

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, ROBERT, T

Breath Test Operator Agency: Lee County SO

Arresting Officer: SHAW, ROBERT

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 22:38

Gender: Male

Observation Period Began: 23:05

Subject's Last Name: LOPEZ

Date/Time of Test: 08/15/2008 23:47

---

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:42
Subject Sample #1	0.078	23:43
Breath Volume	2.253	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample #2	0.086	23:46
Breath Volume	2.609	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEE, SELENA, A

Breath Test Operator Agency: Lee County SO

Arresting Officer: GALLOWAY, N

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 01:24

Gender: Male

Observation Period Began: 02:20

Subject's Last Name: VELASQUEZ

Date/Time of Test: 08/16/2008 02:50

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:44
Subject Sample #1	0.129	02:46
Breath Volume	1.761	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample #2	0.122	02:49
Breath Volume	2.027	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEE, SELENA, A

Breath Test Operator Agency: Lee County SO

Arresting Officer: JERZ, D

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 03:03

Gender: Male

Observation Period Began: 03:30

Subject's Last Name: CASTILLO

Date/Time of Test: 08/16/2008 04:06

---

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample #1	0.121	04:02
Breath Volume	1.765	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample #2	0.127	04:05
Breath Volume	1.910	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CHAPPELL, THOMAS, E

Breath Test Operator Agency: Lee County SO

Arresting Officer: CHAPPELL, THOMAS

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 02:27

Gender: Male

Observation Period Began: 03:52

Subject's Last Name: MATEO

Date/Time of Test: 08/16/2008 04:23

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample #1	0.222	04:18
Breath Volume	2.128	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample #2	0.204	04:22
Breath Volume	1.308	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: HARRIS, WILLIAM, H

Breath Test Operator Agency: FHP Troop F

Arresting Officer: ELLIS, ROBERT

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 10:02

Gender: Male

Observation Period Began: 12:07

Subject's Last Name: RODRIGUEZ

Date/Time of Test: 08/16/2008 12:44

---

Diagnostics Check	OK	12:29	
Air Blank	0.000	12:30	
Control Test	0.080	12:30	
Air Blank	0.000	12:31	
Subject Sample #1	0.128	12:34	Volume Not Met
Breath Volume	0.300	12:34	
Air Blank	0.000	12:35	
Air Blank	0.000	12:36	
Subject Sample #2	0.129	12:37	
Breath Volume	1.589	12:37	
Air Blank	0.000	12:38	
Air Blank	0.000	12:40	
Subject Sample #3	0.133	12:42	
Breath Volume	3.042	12:42	
Air Blank	0.000	12:43	
Control Test	0.078	12:43	
Air Blank	0.000	12:44	
Diagnostics Check	OK	12:44	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, ROBERT, T

Breath Test Operator Agency: Lee County SO

Arresting Officer: SHAW, ROBERT

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 23:05

Gender: Male

Observation Period Began: 00:00

Subject's Last Name: PEREZ-AGUILAR

Date/Time of Test: 08/17/2008 00:33

---

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample #1	0.153	00:28
Breath Volume	2.179	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample #2	0.154	00:31
Breath Volume	1.863	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MILOSEVIC, PATRICK, H

Breath Test Operator Agency: Lee County SO

Arresting Officer: MINO-RAMIREZ, ALEX

Arresting Officer Agency: Cape Coral PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 01:42

Gender: Male

Observation Period Began: 01:47

Subject's Last Name: GOVANUS

Date/Time of Test: 08/17/2008 02:35

---

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.156	02:29
Breath Volume	2.468	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample #2	0.149	02:33
Breath Volume	1.886	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SARGIS, ANDREW, J

Breath Test Operator Agency: Lee County SO

Arresting Officer: STESKAL, CHRIS

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 08:47

Gender: Male

Observation Period Began: 10:20

Subject's Last Name: MENDEZ

Date/Time of Test: 08/17/2008 10:51

---

Diagnostics Check	OK	10:44
Air Blank	0.000	10:45
Control Test	0.080	10:45
Air Blank	0.000	10:46
Subject Sample #1	0.018	10:46
Breath Volume	1.886	10:46
Air Blank	0.000	10:47
Air Blank	0.000	10:49
Subject Sample #2	0.017	10:49
Breath Volume	3.183	10:49
Air Blank	0.000	10:50
Control Test	0.080	10:50
Air Blank	0.000	10:50
Diagnostics Check	OK	10:51

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: MILOSEVIC, PATRICK, H

Breath Test Operator Agency: Lee County SO

Arresting Officer: SANDT, J

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 01:06

Gender: Male

Observation Period Began: 01:28

Subject's Last Name: SUGERMAN

Date/Time of Test: 08/18/2008 02:22

---

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.081	02:12	
Air Blank	0.000	02:12	
Subject Sample #1	0.054	02:15	Volume Not Met
Breath Volume	0.910	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:18	
Subject Sample #2		02:20	Subject Test Refused
Breath Volume	0.425	02:20	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SINIBALDI, ANTHONY,

Breath Test Operator Agency: Lee County SO

Arresting Officer: NAYLOR, MAURICE

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: Yes

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 22:45

Gender: Male

Observation Period Began: 23:20

Subject's Last Name: VALLIA

Date/Time of Test: 08/20/2008 00:10

---

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample #1	0.209	00:06
Breath Volume	2.882	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample #2	0.200	00:09
Breath Volume	1.375	00:09
Air Blank	0.000	00:10
Control Test	0.077	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SINIBALDI, ANTHONY, R

Breath Test Operator Agency: Lee County SO

Arresting Officer: ZIMMERMAN, MICHEAL

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 03:37

Gender: Male

Observation Period Began: 04:04

Subject's Last Name: PISANESCHI

Date/Time of Test: 08/20/2008 04:50

---

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Subject Sample #1	0.131	04:45
Breath Volume	3.761	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample #2	0.128	04:48
Breath Volume	3.820	04:48
Air Blank	0.000	04:49
Control Test	0.077	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: GOULD, LEONARD,

Breath Test Operator Agency: Lee County SO

Arresting Officer: GOULD, LEONARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 21:38

Gender: Male

Observation Period Began: 22:15

Subject's Last Name: HERBERT

Date/Time of Test: 08/20/2008 22:45

---

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.081	22:39
Air Blank	0.000	22:40
Subject Sample #1	0.216	22:41
Breath Volume	1.656	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample #2	0.208	22:43
Breath Volume	1.890	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, ROBERT, T

Breath Test Operator Agency: Lee County SO

Arresting Officer: SHAW, ROBERT

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 00:10

Gender: Male

Observation Period Began: 00:35

Subject's Last Name: RODRIGO

Date/Time of Test: 08/21/2008 01:10

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.079	01:05
Air Blank	0.000	01:05
Subject Sample #1	0.083	01:06
Breath Volume	1.417	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample #2	0.083	01:09
Breath Volume	1.152	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: GOULD, LEONARD,

Breath Test Operator Agency: Lee County SO

Arresting Officer: GOULD, LEONARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 02:14

Gender: Female

Observation Period Began: 02:55

Subject's Last Name: GASDOW

Date/Time of Test: 08/22/2008 03:24

---

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:18
Air Blank	0.000	03:18
Subject Sample #1	0.118	03:19
Breath Volume	3.085	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample #2	0.118	03:22
Breath Volume	3.164	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, ROBERT, T

Breath Test Operator Agency: Lee County SO

Arresting Officer: SHAW, ROBERT

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 04:00

Gender: Male

Observation Period Began: 04:25

Subject's Last Name: GONZALEZ

Date/Time of Test: 08/22/2008 05:09

---

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.081	05:01
Air Blank	0.000	05:02
Subject Sample #1	0.171	05:04
Breath Volume	2.187	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample #2	0.167	05:07
Breath Volume	1.816	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: GOULD, LEONARD,

Breath Test Operator Agency: Lee County SO

Arresting Officer: GOULD, LEONARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 18:11

Gender: Male

Observation Period Began: 18:50

Subject's Last Name: SHANNON

Date/Time of Test: 08/22/2008 19:19

---

Diagnostics Check	OK	19:12
Air Blank	0.000	19:13
Control Test	0.080	19:13
Air Blank	0.000	19:14
Subject Sample #1	0.240	19:14
Breath Volume	3.488	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample #2	0.242	19:17
Breath Volume	3.046	19:17
Air Blank	0.000	19:18
Control Test	0.080	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: HUTCHINSON, MICHAEL, J

Breath Test Operator Agency: Lee County SO

Arresting Officer: SPEAS, RANA

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 12:04

Gender: Male

Observation Period Began: 15:16

Subject's Last Name: SCHUMACHER

Date/Time of Test: 08/23/2008 15:48

---

Diagnostics Check	OK	15:44	
Air Blank	0.000	15:44	
Control Test	0.080	15:45	
Air Blank	0.000	15:45	
Subject Sample #1		15:46	Interferent Detect
Breath Volume	1.292	15:46	
Air Blank	0.000	15:47	
Control Test	0.081	15:47	
Air Blank	0.000	15:48	
Diagnostics Check	OK	15:48	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: HUTCHINSON, MICHAEL, J

Breath Test Operator Agency: Lee County SO

Arresting Officer: SPEAS, RANA

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 12:04

Gender: Male

Observation Period Began: 15:16

Subject's Last Name: SCHUMACHER

Date/Time of Test: 08/23/2008 15:54

---

Diagnostics Check	OK	15:50	
Air Blank	0.000	15:50	
Control Test	0.080	15:50	
Air Blank	0.000	15:51	
Subject Sample #1		15:53	Sequence Aborted
Breath Volume	0.695	15:53	
Air Blank	0.000	15:54	

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SINIBALDI, ANTHONY, R

Breath Test Operator Agency: Lee County SO

Arresting Officer: BOCK, ANDREW

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 16:43

Gender: Male

Observation Period Began: 17:14

Subject's Last Name: BROCK

Date/Time of Test: 08/23/2008 18:07

---

Diagnostics Check	OK	18:00
Air Blank	0.000	18:01
Control Test	0.080	18:01
Air Blank	0.000	18:01
Subject Sample #1	0.070	18:02
Breath Volume	1.484	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample #2	0.068	18:05
Breath Volume	1.734	18:05
Air Blank	0.000	18:06
Control Test	0.079	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: BARRINGER, BRIAN, M

Breath Test Operator Agency: Lee County SO

Arresting Officer: DESROSIERS, JOHN

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 22:33

Gender: Female

Observation Period Began: 23:20

Subject's Last Name: SCHUST-FINAN

Date/Time of Test: 08/24/2008 00:08

---

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample #1	0.000	00:04
Breath Volume	2.507	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample #2	0.000	00:07
Breath Volume	2.265	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEE, SELENA, A

Breath Test Operator Agency: Lee County SO

Arresting Officer: NAYLOR, M

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 00:32

Gender: Male

Observation Period Began: 01:11

Subject's Last Name: SANCHEZ

Date/Time of Test: 08/24/2008 01:45

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:39
Subject Sample #1	0.132	01:40
Breath Volume	1.183	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample #2	0.152	01:44
Breath Volume	3.101	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: BARRINGER, BRIAN, M

Breath Test Operator Agency: Lee County SO

Arresting Officer: SANDT, J

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 01:45

Subject's Last Name: UHLIG

Date/Time of Test: 08/24/2008 02:19

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample #1	0.200	02:14
Breath Volume	1.210	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample #2	0.197	02:17
Breath Volume	1.378	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: HUTCHINSON, MICHAEL, J

Breath Test Operator Agency: Lee County SO

Arresting Officer: HURD, ERIK

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 03:17

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: ENDSLEY

Date/Time of Test: 08/24/2008 05:20

---

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.080	05:15
Air Blank	0.000	05:15
Subject Sample #1	0.161	05:16
Breath Volume	2.660	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample #2	0.162	05:19
Breath Volume	2.699	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: WAYLAND, JOSEPH, T

Breath Test Operator Agency: Cape Coral PD

Arresting Officer: WAYLAND, JOSEPH

Arresting Officer Agency: Cape Coral PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8042011

Dry Gas Expiry Date: 02/14/2010

Violation Code: DUI

Arrest Time: 20:49

Gender: Male

Observation Period Began: 20:50

Subject's Last Name: PATERSON

Date/Time of Test: 08/24/2008 22:07

---

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Subject Sample #1	0.072	22:02
Breath Volume	1.386	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample #2	0.068	22:05
Breath Volume	1.335	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

---

Data Download Date/Time: 08/25/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: REYES, GUSTAVO

Arresting Officer Agency: FHP Troop F

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 20:32

Gender: Male

Observation Period Began: 21:15

Subject's Last Name: HERNANDEZ

Date/Time of Test: 08/26/2008 22:08

---

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.081	21:58
Air Blank	0.000	21:59
Subject Sample #1	0.255	22:00
Breath Volume	2.679	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample #2	0.233	22:03
Breath Volume	1.371	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample #3	0.225	22:07
Breath Volume	2.027	22:07
Air Blank	0.000	22:07
Control Test	0.079	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: MCNALLY, DONNA

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 03:07

Gender: Male

Observation Period Began: 03:47

Subject's Last Name: RICHTER

Date/Time of Test: 08/27/2008 04:25

---

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Subject Sample #1	0.000	04:20
Breath Volume	3.238	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.000	04:23
Breath Volume	2.535	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: GOULD, LEONARD,

Breath Test Operator Agency: Lee County SO

Arresting Officer: GOULD, LEONARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 01:45

Subject's Last Name: O'BRIEN

Date/Time of Test: 08/28/2008 02:12

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Subject Sample #1	0.193	02:07
Breath Volume	2.777	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.180	02:10
Breath Volume	2.429	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SINIBALDI, ANTHONY, R

Breath Test Operator Agency: Lee County SO

Arresting Officer: LOUGHREN, CHRISTOPHER

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 00:52

Gender: Male

Observation Period Began: 01:20

Subject's Last Name: GORDON

Date/Time of Test: 08/29/2008 01:54

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Subject Sample #1	0.164	01:49
Breath Volume	2.394	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample #2	0.165	01:52
Breath Volume	3.191	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SARGIS, ANDREW, J

Breath Test Operator Agency: Lee County SO

Arresting Officer: ITZKO, MARK

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 14:05

Gender: Male

Observation Period Began: 15:15

Subject's Last Name: RONDON

Date/Time of Test: 08/29/2008 15:56

---

Diagnostics Check	OK	15:49
Air Blank	0.000	15:50
Control Test	0.080	15:50
Air Blank	0.000	15:51
Subject Sample #1	0.188	15:52
Breath Volume	1.917	15:52
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample #2	0.192	15:55
Breath Volume	3.343	15:55
Air Blank	0.000	15:55
Control Test	0.079	15:56
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: MOORE, JASON

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 02:29

Gender: Female

Observation Period Began: 02:44

Subject's Last Name: HAY

Date/Time of Test: 08/30/2008 03:15

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Subject Sample #1	0.170	03:10
Breath Volume	2.023	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.163	03:13
Breath Volume	1.332	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SHAW, ROBERT, T

Breath Test Operator Agency: Lee County SO

Arresting Officer: SHAW, ROBERT

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 02:20

Gender: Male

Observation Period Began: 03:02

Subject's Last Name: KOPP

Date/Time of Test: 08/30/2008 03:30

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample #1	0.104	03:26
Breath Volume	1.464	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample #2	0.104	03:29
Breath Volume	1.355	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SUTPHIN, MARSHA, L

Breath Test Operator Agency: Lee County SO

Arresting Officer: SANDT, JASON

Arresting Officer Agency: Fort Myers PD

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 03:08

Gender: Male

Observation Period Began: 03:15

Subject's Last Name: ALANIZ

Date/Time of Test: 08/30/2008 04:18

---

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample #1	0.174	04:13
Breath Volume	1.875	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample #2	0.159	04:16
Breath Volume	1.546	04:16
Air Blank	0.000	04:17
Control Test	0.078	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CHAPPELL, THOMAS, E

Breath Test Operator Agency: Lee County SO

Arresting Officer: CHAPPELL, THOMAS

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 02:43

Gender: Male

Observation Period Began: 04:06

Subject's Last Name: GRAJALES RENDON

Date/Time of Test: 08/30/2008 04:33

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample #1	0.088	04:29
Breath Volume	1.902	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample #2	0.089	04:31
Breath Volume	1.667	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: FARESE, JOHN, V

Breath Test Operator Agency: Lee County SO

Arresting Officer: FARESE, JOHN

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 18:35

Gender: Male

Observation Period Began: 19:52

Subject's Last Name: STEWART

Date/Time of Test: 08/30/2008 20:28

---

Diagnostics Check	OK	20:16	
Air Blank	0.000	20:16	
Control Test	0.081	20:17	
Air Blank	0.000	20:17	
Subject Sample #1	0.169	20:20	Volume Not Met
Breath Volume	1.093	20:20	
Air Blank	0.000	20:21	
Air Blank	0.000	20:23	
Subject Sample #2	0.170	20:26	Volume Not Met
Breath Volume	0.902	20:26	
Air Blank	0.000	20:27	
Control Test	0.078	20:27	
Air Blank	0.000	20:28	
Diagnostics Check	OK	20:28	

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CHAPPELL, THOMAS, E

Breath Test Operator Agency: Lee County SO

Arresting Officer: THORPE, ASHLEY

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 21:35

Gender: Male

Observation Period Began: 23:09

Subject's Last Name: QUINONES-GUZMAN

Date/Time of Test: 08/30/2008 23:46

---

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Subject Sample #1	0.215	23:40
Breath Volume	1.734	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample #2	0.205	23:44
Breath Volume	1.328	23:44
Air Blank	0.000	23:45
Control Test	0.078	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SARGIS, ANDREW, J

Breath Test Operator Agency: Lee County SO

Arresting Officer: HAMER, ROB

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 04:17

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: CONTRERAS

Date/Time of Test: 08/31/2008 05:45

---

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:42	
Subject Sample #1	0.207	05:43	
Breath Volume	3.359	05:43	
Air Blank		05:44	Purge Fail
Air Blank	0.000	05:45	

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: SARGIS, ANDREW, J

Breath Test Operator Agency: Lee County SO

Arresting Officer: HAMER, ROB

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 04:17

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: CONTRERAS

Date/Time of Test: 08/31/2008 05:48

---

Diagnostics Check	OK	05:47	
Air Blank	0.000	05:47	
Control Test	0.071	05:48	Control Outside Tolerance
Air Blank	0.000	05:48	

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: LEE, SELENA, A

Breath Test Operator Agency: Lee County SO

Arresting Officer: LEE, SELENA

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 08/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8221011

Dry Gas Expiry Date: 08/11/2010

Violation Code: DUI

Arrest Time: 18:39

Gender: Female

Observation Period Began: 19:30

Subject's Last Name: ELLENBARGER

Date/Time of Test: 08/31/2008 20:02

---

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.080	19:54
Air Blank	0.000	19:54
Subject Sample #1	0.163	19:57
Breath Volume	1.203	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample #2	0.171	20:00
Breath Volume	1.507	20:00
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

---

Data Download Date/Time: 09/15/2008 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO  
Software Version: 8100.27  
Breath Test Operator: SNYDER, MELONIE, E  
Breath Test Operator Agency: Hillsborough Co SO  
Arresting Officer: SEALE, J  
Arresting Officer Agency: Hillsborough Co SO  
Last Agency Inspection: 07/24/2008

Date or Time Changed: No  
Dry Gas Standard Lot #: 717102I  
Dry Gas Expiry Date: 06/21/2009  
Violation Code: DUI  
Arrest Time: 20:33  
Gender: Male  
Observation Period Began: 21:03

Subject's Last Name: HINGSON

Date/Time of Test: 08/15/2008 21:30

---

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.078	21:24
Air Blank	0.000	21:24
Subject Sample #1	0.050	21:25
Breath Volume	2.617	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample #2	0.048	21:28
Breath Volume	1.453	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

---

Data Download Date/Time: 08/22/2008 12:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: SNYDER, MELONIE, E

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: MCMURPHY, M

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 07/24/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 717102I

Dry Gas Expiry Date: 06/21/2009

Violation Code: DUI

Arrest Time: 02:00

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: DIAZ

Date/Time of Test: 08/16/2008 03:03

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:58
Air Blank	0.000	02:58
Subject Sample #1	0.115	02:59
Breath Volume	2.410	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample #2	0.117	03:02
Breath Volume	1.199	03:02
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 08/22/2008 12:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO  
Software Version: 8100.27  
Breath Test Operator: SNYDER, MELONIE, E  
Breath Test Operator Agency: Hillsborough Co SO  
Arresting Officer: MORET, F  
Arresting Officer Agency: Hillsborough Co SO  
Last Agency Inspection: 07/24/2008

Date or Time Changed: No  
Dry Gas Standard Lot #: 717102I  
Dry Gas Expiry Date: 06/21/2009  
Violation Code: DUI  
Arrest Time: 02:06  
Gender: Male  
Observation Period Began: 02:37

Subject's Last Name: SIMPSON

Date/Time of Test: 08/16/2008 03:12

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:05
Subject Sample #1	0.187	03:07
Breath Volume	1.855	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.183	03:11
Breath Volume	1.867	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 08/22/2008 12:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: SNYDER, MELONIE, E

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: PORTMAN, A

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 08/22/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 717102I

Dry Gas Expiry Date: 06/21/2009

Violation Code: DUI

Arrest Time: 01:29

Gender: Male

Observation Period Began: 01:48

Subject's Last Name: LAVENDER

Date/Time of Test: 08/29/2008 02:16

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Subject Sample #1	0.126	02:12
Breath Volume	1.605	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample #2	0.120	02:15
Breath Volume	1.289	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

---

Data Download Date/Time: 09/11/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: SNYDER, MELONIE, E

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: BARBOSA, F

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 08/22/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 717102I

Dry Gas Expiry Date: 06/21/2009

Violation Code: DUI

Arrest Time: 01:23

Gender: Male

Observation Period Began: 01:48

Subject's Last Name: PANCIOLI

Date/Time of Test: 08/29/2008 02:25

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.076	02:19
Air Blank	0.000	02:19
Subject Sample #1	0.135	02:20
Breath Volume	1.878	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.138	02:23
Breath Volume	2.156	02:23
Air Blank	0.000	02:24
Control Test	0.077	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time: 09/11/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: SNYDER, MELONIE, E

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: CRAGG, S

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 08/22/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 717102I

Dry Gas Expiry Date: 06/21/2009

Violation Code: DUI

Arrest Time: 02:30

Gender: Female

Observation Period Began: 02:48

Subject's Last Name: RANDOLPH

Date/Time of Test: 08/29/2008 03:15

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.077	03:10
Air Blank	0.000	03:10
Subject Sample #1	0.127	03:11
Breath Volume	2.425	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.128	03:14
Breath Volume	2.636	03:14
Air Blank	0.000	03:14
Control Test	0.077	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time: 09/11/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: SNYDER, MELONIE, E

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: MORREY, S

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 08/22/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 717102I

Dry Gas Expiry Date: 06/21/2009

Violation Code: DUI

Arrest Time: 00:48

Gender: Female

Observation Period Began: 01:09

Subject's Last Name: VANDOORN

Date/Time of Test: 08/30/2008 01:35

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.076	01:30
Air Blank	0.000	01:30
Subject Sample #1	0.114	01:31
Breath Volume	1.523	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample #2	0.110	01:34
Breath Volume	1.468	01:34
Air Blank	0.000	01:35
Control Test	0.076	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

---

Data Download Date/Time: 09/11/2008 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: BYRNE, CATHY, S

Breath Test Operator Agency: FHP Troop E

Arresting Officer: BYRNE, CATHY

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 04:15

Gender: Male

Observation Period Began: 04:02

Subject's Last Name: ARAMBULA

Date/Time of Test: 08/02/2008 04:53

---

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:49	
Control Test	0.078	04:49	
Air Blank	0.000	04:50	
Subject Sample #1		04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:52	
Control Test	0.077	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M, W

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 05:20

Gender: Male

Observation Period Began: 05:32

Subject's Last Name: ROSALES

Date/Time of Test: 08/02/2008 05:56

---

Diagnostics Check	OK	05:53	
Air Blank	0.000	05:53	
Control Test	0.077	05:54	
Air Blank	0.000	05:54	
Subject Sample #1		05:54	Subject Test Refused
Breath Volume	0.000	05:54	
Air Blank	0.000	05:55	
Control Test	0.077	05:55	
Air Blank	0.000	05:56	
Diagnostics Check	OK	05:56	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M, W

Breath Test Operator Agency: South Miami PD

Arresting Officer: EISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 03:30

Gender: Male

Observation Period Began: 03:46

Subject's Last Name: CIVALERO

Date/Time of Test: 08/07/2008 04:16

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.076	04:09
Air Blank	0.000	04:10
Subject Sample #1	0.158	04:11
Breath Volume	1.398	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample #2	0.157	04:15
Breath Volume	3.277	04:15
Air Blank	0.000	04:15
Control Test	0.075	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: BYRNE, CATHY, S

Breath Test Operator Agency: FHP Troop E

Arresting Officer: BYRNE, CATHY

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 06:00

Gender: Male

Observation Period Began: 05:40

Subject's Last Name: BRAVO

Date/Time of Test: 08/09/2008 07:23

---

Diagnostics Check	OK	07:16
Air Blank	0.000	07:17
Control Test	0.077	07:17
Air Blank	0.000	07:17
Subject Sample #1	0.090	07:19
Breath Volume	2.585	07:19
Air Blank	0.000	07:19
Air Blank	0.000	07:21
Subject Sample #2	0.091	07:22
Breath Volume	3.117	07:22
Air Blank	0.000	07:23
Control Test	0.075	07:23
Air Blank	0.000	07:23
Diagnostics Check	OK	07:23

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: BYRNE, CATHY, S

Breath Test Operator Agency: FHP Troop E

Arresting Officer: MILNES, KENNETH

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 02:15

Gender: Male

Observation Period Began: 03:40

Subject's Last Name: SANTOS-ALBORNA

Date/Time of Test: 08/10/2008 04:08

---

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.077	04:01
Air Blank	0.000	04:02
Subject Sample #1	0.035	04:03
Breath Volume	2.082	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.034	04:06
Breath Volume	3.128	04:06
Air Blank	0.000	04:07
Control Test	0.076	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: BYRNE, CATHY, S

Breath Test Operator Agency: FHP Troop E

Arresting Officer: GRIEGO, CHARLENE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 06:15

Gender: Female

Observation Period Began: 06:15

Subject's Last Name: RODRIGUEZ

Date/Time of Test: 08/10/2008 07:24

---

Diagnostics Check	OK	07:11	
Air Blank	0.000	07:11	
Control Test	0.077	07:11	
Air Blank	0.000	07:12	
Subject Sample #1	0.117	07:15	Volume Not Met
Breath Volume	0.218	07:15	
Air Blank	0.000	07:16	
Air Blank	0.000	07:18	
Subject Sample #2	0.116	07:19	
Breath Volume	1.937	07:19	
Air Blank	0.000	07:20	
Air Blank	0.000	07:21	
Subject Sample #3	0.106	07:22	
Breath Volume	1.957	07:22	
Air Blank	0.000	07:23	
Control Test	0.073	07:23	Control Outside Tolerance
Air Blank	0.000	07:24	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M, W

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 01:01

Gender: Female

Observation Period Began: 01:12

Subject's Last Name: GARRIDO

Date/Time of Test: 08/14/2008 01:46

---

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.077	01:44	
Air Blank	0.000	01:44	
Subject Sample #1		01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.077	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: CORTES, HECTOR, E

Breath Test Operator Agency: FHP Troop E

Arresting Officer: CORTES, HECTOR

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 07/31/2008

Date or Time Changed: Yes

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 17:57

Gender: Male

Observation Period Began: 18:28

Subject's Last Name: COSTA SOSA

Date/Time of Test: 08/16/2008 20:19

---

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:15	
Control Test	0.077	20:15	
Air Blank	0.000	20:16	
Subject Sample #1		20:18	Subject Test Refused
Breath Volume	0.000	20:18	
Air Blank	0.000	20:18	
Control Test	0.077	20:18	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M, W

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 04:01

Gender: Female

Observation Period Began: 04:21

Subject's Last Name: JANKOWSKI

Date/Time of Test: 08/21/2008 04:46

---

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:44	
Control Test	0.077	04:44	
Air Blank	0.000	04:44	
Subject Sample #1		04:45	Subject Test Refused
Breath Volume	0.000	04:45	
Air Blank	0.000	04:45	
Control Test	0.077	04:45	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: NADAL, JOSE, V

Breath Test Operator Agency: FHP Troop E

Arresting Officer: NADAL, JOSE

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 03:35

Gender: Male

Observation Period Began: 03:40

Subject's Last Name: AYRA III

Date/Time of Test: 08/23/2008 04:30

---

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.076	04:26	
Air Blank	0.000	04:26	
Subject Sample #1		04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.076	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: BYRNE, CATHY, S

Breath Test Operator Agency: FHP Troop E

Arresting Officer: THOMAS, RICHARD

Arresting Officer Agency: FHP Troop E

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 07:55

Gender: Male

Observation Period Began: 07:22

Subject's Last Name: PRENT

Date/Time of Test: 08/23/2008 08:43

---

Diagnostics Check	OK	08:36
Air Blank	0.000	08:37
Control Test	0.076	08:37
Air Blank	0.000	08:38
Subject Sample #1	0.193	08:38
Breath Volume	2.621	08:38
Air Blank	0.000	08:39
Air Blank	0.000	08:41
Subject Sample #2	0.176	08:42
Breath Volume	1.472	08:42
Air Blank	0.000	08:43
Control Test	0.075	08:43
Air Blank	0.000	08:43
Diagnostics Check	OK	08:43

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M, W

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 04:41

Gender: Male

Observation Period Began: 04:50

Subject's Last Name: TORRES

Date/Time of Test: 08/24/2008 05:29

---

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.076	05:23
Air Blank	0.000	05:24
Subject Sample #1	0.110	05:24
Breath Volume	2.167	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample #2	0.108	05:27
Breath Volume	1.875	05:27
Air Blank	0.000	05:28
Control Test	0.076	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M, W

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 03:00

Gender: Female

Observation Period Began: 03:10

Subject's Last Name: LOPEZ

Date/Time of Test: 08/25/2008 03:33

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.077	03:31	
Air Blank	0.000	03:31	
Subject Sample #1		03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.076	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M, W

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 04:00

Gender: Male

Observation Period Began: 04:08

Subject's Last Name: PERRERO

Date/Time of Test: 08/25/2008 04:40

---

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:37	
Control Test	0.076	04:38	
Air Blank	0.000	04:38	
Subject Sample #1		04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:39	
Control Test	0.077	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

---

Data Download Date/Time: 08/30/2008 22:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001737

Instrument Registered To: SOUTH MIAMI PD

Software Version: 8100.27

Breath Test Operator: WEISSBERG, M, W

Breath Test Operator Agency: South Miami PD

Arresting Officer: WEISSBERG, M

Arresting Officer Agency: South Miami PD

Last Agency Inspection: 08/30/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 626102I

Dry Gas Expiry Date: 09/20/2008

Violation Code: DUI

Arrest Time: 05:35

Gender: Female

Observation Period Began: 05:43

Subject's Last Name: SOSA

Date/Time of Test: 08/31/2008 06:13

---

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.076	06:06
Air Blank	0.000	06:06
Subject Sample #1	0.110	06:08
Breath Volume	2.714	06:08
Air Blank	0.000	06:09
Air Blank	0.000	06:10
Subject Sample #2	0.107	06:11
Breath Volume	1.433	06:11
Air Blank	0.000	06:12
Control Test	0.076	06:12
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

---

Data Download Date/Time: 09/25/2008 00:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: HATCH, OFC, M

Breath Test Operator Agency: Clearwater PD

Arresting Officer: LYONS, C

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 07/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7290011

Dry Gas Expiry Date: 10/23/2009

Violation Code: DUI

Arrest Time: 03:16

Gender: Male

Observation Period Began: 04:10

Subject's Last Name: BAXCAJAY

Date/Time of Test: 08/01/2008 04:41

---

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:35
Air Blank	0.000	04:35
Subject Sample #1	0.157	04:36
Breath Volume	2.960	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample #2	0.155	04:39
Breath Volume	2.625	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:41

---

Data Download Date/Time: 08/06/2008 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 07/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7290011

Dry Gas Expiry Date: 10/23/2009

Violation Code: DUI

Arrest Time: 00:50

Gender: Male

Observation Period Began: 01:26

Subject's Last Name: HERNANDEZ ROCHA

Date/Time of Test: 08/02/2008 02:02

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Subject Sample #1	0.184	01:57
Breath Volume	1.406	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample #2	0.175	02:01
Breath Volume	2.425	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 08/06/2008 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: LAMBE, JASON

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 07/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7290011

Dry Gas Expiry Date: 10/23/2009

Violation Code: DUI

Arrest Time: 01:03

Gender: Male

Observation Period Began: 02:03

Subject's Last Name: LAWLESS

Date/Time of Test: 08/02/2008 02:34

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Subject Sample #1	0.226	02:30
Breath Volume	1.351	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample #2	0.234	02:33
Breath Volume	3.511	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 08/06/2008 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 07/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7290011

Dry Gas Expiry Date: 10/23/2009

Violation Code: DUI

Arrest Time: 03:19

Gender: Male

Observation Period Began: 04:11

Subject's Last Name: LOPEZ CASTILLO

Date/Time of Test: 08/02/2008 04:47

---

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.080	04:41	
Air Blank	0.000	04:42	
Subject Sample #1	0.201	04:43	
Breath Volume	2.421	04:43	
Air Blank	0.000	04:44	
Air Blank	0.000	04:45	
Subject Sample #2	0.187	04:46	
Breath Volume	2.109	04:46	
Air Blank	0.000	04:46	
Control Test	0.074	04:47	Control Outside Tolerance
Air Blank	0.000	04:47	

---

Data Download Date/Time: 08/06/2008 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 07/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7290011

Dry Gas Expiry Date: 10/23/2009

Violation Code: DUI

Arrest Time: 03:19

Gender: Male

Observation Period Began: 04:48

Subject's Last Name: LOPEZ CASTILLO

Date/Time of Test: 08/02/2008 05:21

---

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:15	
Control Test	0.078	05:15	
Air Blank	0.000	05:15	
Subject Sample #1	0.181	05:16	
Breath Volume	2.179	05:16	
Air Blank	0.000	05:17	
Air Blank	0.000	05:19	
Subject Sample #2	0.185	05:19	
Breath Volume	2.339	05:19	
Air Blank	0.000	05:20	
Control Test	0.078	05:20	
Air Blank	0.000	05:21	
Diagnostics Check	Fail	05:21	Temperature Regulation Test
Air Blank	0.000	05:21	

---

Data Download Date/Time: 08/06/2008 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: SUERETH, DONALD, P

Breath Test Operator Agency: FDLE

Arresting Officer: OFFICER, ARRESTING

Arresting Officer Agency: FDLE

Last Agency Inspection: 07/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7096011

Dry Gas Expiry Date: 04/08/2009

Violation Code: SYS CHECK

Arrest Time: 12:00

Gender: Female

Observation Period Began: 12:15

Subject's Last Name: LICENSE

Date/Time of Test: 08/06/2008 12:50

---

Diagnostics Check	OK	12:43
Air Blank	0.000	12:43
Control Test	0.081	12:44
Air Blank	0.000	12:44
Subject Sample #1	0.000	12:45
Breath Volume	2.542	12:45
Air Blank	0.000	12:46
Air Blank	0.000	12:48
Subject Sample #2	0.000	12:48
Breath Volume	2.542	12:48
Air Blank	0.000	12:49
Control Test	0.080	12:49
Air Blank	0.000	12:49
Diagnostics Check	OK	12:50

---

Data Download Date/Time: 08/06/2008 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 08/06/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8094011

Dry Gas Expiry Date: 03/21/2010

Violation Code: DUI

Arrest Time: 23:30

Gender: Male

Observation Period Began: 00:08

Subject's Last Name: STARRETT

Date/Time of Test: 08/09/2008 00:39

---

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.082	00:37	
Air Blank	0.000	00:37	
Subject Sample #1		00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Control Test	0.083	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

---

Data Download Date/Time: 09/19/2008 04:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: WALSER, JONATHAN, B

Breath Test Operator Agency: Clearwater PD

Arresting Officer: LULEK, T

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 08/06/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8094011

Dry Gas Expiry Date: 03/21/2010

Violation Code: DUI

Arrest Time: 12:19

Gender: Male

Observation Period Began: 13:12

Subject's Last Name: THOMPSON

Date/Time of Test: 08/14/2008 13:43

---

Diagnostics Check	OK	13:37
Air Blank	0.000	13:37
Control Test	0.081	13:38
Air Blank	0.000	13:38
Subject Sample #1	0.050	13:39
Breath Volume	2.914	13:39
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample #2	0.047	13:41
Breath Volume	2.398	13:41
Air Blank	0.000	13:42
Control Test	0.080	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

---

Data Download Date/Time: 09/19/2008 04:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 08/06/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8094011

Dry Gas Expiry Date: 03/21/2010

Violation Code: DUI

Arrest Time: 00:42

Gender: Male

Observation Period Began: 01:32

Subject's Last Name: HEINE

Date/Time of Test: 08/17/2008 02:04

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.082	01:54
Air Blank	0.000	01:54
Subject Sample #1	0.246	01:55
Breath Volume	1.160	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample #2	0.224	01:59
Breath Volume	1.703	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample #3	0.210	02:03
Breath Volume	1.648	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time: 09/19/2008 04:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: BODLE, RANDY, S

Breath Test Operator Agency: Clearwater PD

Arresting Officer: NUGENT, D

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 08/06/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8094011

Dry Gas Expiry Date: 03/21/2010

Violation Code: DUI

Arrest Time: 23:03

Gender: Male

Observation Period Began: 00:28

Subject's Last Name: DUNN

Date/Time of Test: 08/20/2008 00:52

---

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.081	00:49	
Air Blank	0.000	00:49	
Subject Sample #1	0.200	00:50	
Breath Volume	1.308	00:50	
Air Blank		00:50	RFI Detect
Air Blank	0.000	00:51	
Control Test	0.082	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 09/19/2008 04:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: BODLE, RANDY, S

Breath Test Operator Agency: Clearwater PD

Arresting Officer: NUGENT, DAVID

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 08/06/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8094011

Dry Gas Expiry Date: 03/21/2010

Violation Code: DUI

Arrest Time: 23:03

Gender: Male

Observation Period Began: 00:20

Subject's Last Name: DUNN

Date/Time of Test: 08/20/2008 01:02

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:55
Subject Sample #1	0.179	00:57
Breath Volume	4.000	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample #2	0.186	01:01
Breath Volume	2.902	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

---

Data Download Date/Time: 09/19/2008 04:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 08/06/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8094011

Dry Gas Expiry Date: 03/21/2010

Violation Code: DUI

Arrest Time: 02:36

Gender: Male

Observation Period Began: 03:05

Subject's Last Name: ROMINE

Date/Time of Test: 08/23/2008 03:39

---

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.081	03:36	
Air Blank	0.000	03:36	
Subject Sample #1		03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.082	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:39	

---

Data Download Date/Time: 09/19/2008 04:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: WIERZBA, ROBERT, J

Breath Test Operator Agency: Clearwater PD

Arresting Officer: WIERZBA, ROBERT

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 08/06/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8094011

Dry Gas Expiry Date: 03/21/2010

Violation Code: DUI

Arrest Time: 02:46

Gender: Male

Observation Period Began: 03:02

Subject's Last Name: JELKS

Date/Time of Test: 08/28/2008 03:34

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.081	03:28
Air Blank	0.000	03:28
Subject Sample #1	0.111	03:29
Breath Volume	2.671	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample #2	0.115	03:32
Breath Volume	1.898	03:32
Air Blank	0.000	03:33
Control Test	0.082	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

---

Data Download Date/Time: 09/19/2008 04:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: MILNE, THOMAS, P

Breath Test Operator Agency: Clearwater PD

Arresting Officer: MILNE, THOMAS

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 08/06/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8094011

Dry Gas Expiry Date: 03/21/2010

Violation Code: DUI

Arrest Time: 02:48

Gender: Male

Observation Period Began: 03:15

Subject's Last Name: MAY

Date/Time of Test: 08/30/2008 03:49

---

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:37	
Subject Sample #1	0.212	03:38	
Breath Volume	1.101	03:38	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample #2	0.199	03:44	Volume Not Met
Breath Volume	0.550	03:44	
Air Blank	0.000	03:45	
Air Blank	0.000	03:46	
Subject Sample #3	0.213	03:47	
Breath Volume	1.167	03:47	
Air Blank	0.000	03:48	
Control Test	0.077	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:49	

---

Data Download Date/Time: 09/19/2008 04:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001865

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: GOMEZ, ALEX,

Breath Test Operator Agency: Lakeland PD

Arresting Officer: GOMEZ, ALEX

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 07/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 02:12

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: SHERMAN

Date/Time of Test: 08/02/2008 03:17

---

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample #1	0.183	03:12
Breath Volume	1.886	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.167	03:15
Breath Volume	1.273	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time: 08/27/2008 12:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001865

Instrument Registered To: LAKELAND PD

Software Version: 8100.27

Breath Test Operator: LEHMAN, HANS,

Breath Test Operator Agency: Lakeland PD

Arresting Officer: LEHMAN, HANS

Arresting Officer Agency: Lakeland PD

Last Agency Inspection: 07/25/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 706102I

Dry Gas Expiry Date: 03/05/2009

Violation Code: DUI

Arrest Time: 23:41

Gender: Male

Observation Period Began: 23:58

Subject's Last Name: ALGER

Date/Time of Test: 08/23/2008 00:29

---

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:24
Subject Sample #1	0.214	00:24
Breath Volume	3.742	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample #2	0.211	00:28
Breath Volume	3.871	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time: 08/27/2008 12:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRINGTON, GERARD, J

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: MOORE-LECLERC, SUZANNE

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 03:03

Gender: Male

Observation Period Began: 04:05

Subject's Last Name: BROWN

Date/Time of Test: 08/04/2008 04:32

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample #1	0.118	04:27
Breath Volume	2.933	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample #2	0.104	04:30
Breath Volume	1.601	04:30
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

---

Data Download Date/Time: 08/31/2008 21:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD  
Software Version: 8100.27  
Breath Test Operator: ELLIOTT, TONY, P  
Breath Test Operator Agency: Port St. Lucie PD  
Arresting Officer: THOMPSON, BRIAN  
Arresting Officer Agency: Port St. Lucie PD  
Last Agency Inspection: 07/31/2008

Date or Time Changed: No  
Dry Gas Standard Lot #: 701201G  
Dry Gas Expiry Date: 01/15/2009  
Violation Code: DUI  
Arrest Time: 18:47  
Gender: Female  
Observation Period Began: 18:50

Subject's Last Name: BRIANTE

Date/Time of Test: 08/15/2008 19:27

---

Diagnostics Check	OK	19:24	
Air Blank	0.000	19:24	
Control Test	0.079	19:24	
Air Blank	0.000	19:25	
Subject Sample #1		19:26	Subject Test Refused
Breath Volume	0.000	19:26	
Air Blank	0.000	19:26	
Control Test	0.079	19:26	
Air Blank	0.000	19:27	
Diagnostics Check	OK	19:27	

---

Data Download Date/Time: 08/31/2008 21:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BRADY, KEVIN, E

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: MOORE LECLERC, SUZANNE

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 01:27

Gender: Male

Observation Period Began: 01:40

Subject's Last Name: HERNANDEZ

Date/Time of Test: 08/16/2008 02:12

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Subject Sample #1	0.172	02:08
Breath Volume	2.726	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample #2	0.169	02:11
Breath Volume	2.019	02:11
Air Blank	0.000	02:11
Control Test	0.075	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 08/31/2008 21:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: FAZIO, JOHN, A

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: O'HARA, ROBERT

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 14:51

Gender: Female

Observation Period Began: 14:51

Subject's Last Name: WORDEN

Date/Time of Test: 08/19/2008 15:40

---

Diagnostics Check	OK	15:33
Air Blank	0.000	15:33
Control Test	0.079	15:34
Air Blank	0.000	15:34
Subject Sample #1	0.111	15:35
Breath Volume	1.269	15:35
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample #2	0.099	15:38
Breath Volume	1.425	15:38
Air Blank	0.000	15:39
Control Test	0.075	15:39
Air Blank	0.000	15:39
Diagnostics Check	OK	15:40

---

Data Download Date/Time: 08/31/2008 21:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BUTTERWORTH, DAVID, S

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: BUTTERWORTH, DAVID

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 07/31/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 701201G

Dry Gas Expiry Date: 01/15/2009

Violation Code: DUI

Arrest Time: 02:30

Gender: Female

Observation Period Began: 02:44

Subject's Last Name: WARREN

Date/Time of Test: 08/23/2008 03:12

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample #1	0.179	03:08
Breath Volume	2.601	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample #2	0.174	03:11
Breath Volume	3.304	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 08/31/2008 21:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE PD

Software Version: 8100.27

Breath Test Operator: ADAMS, KENNETH, D

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: CLINE, PATRICK

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 07/21/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 23:20

Gender: Male

Observation Period Began: 23:53

Subject's Last Name: BRASWELL

Date/Time of Test: 08/03/2008 00:30

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Subject Sample #1	0.222	00:25
Breath Volume	4.382	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample #2	0.210	00:28
Breath Volume	1.878	00:28
Air Blank	0.000	00:29
Control Test	0.076	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:30

---

Data Download Date/Time: 08/18/2008 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE PD

Software Version: 8100.27

Breath Test Operator: BAYOT, DAVID, C

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: BAYOT, DAVID

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 07/21/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: ZERO TOL

Arrest Time: 09:30

Gender: Male

Observation Period Began: 09:45

Subject's Last Name: CRUMP

Date/Time of Test: 08/04/2008 10:21

---

Diagnostics Check	OK	10:15
Air Blank	0.000	10:15
Control Test	0.080	10:16
Air Blank	0.000	10:16
Subject Sample #1	0.018	10:17
Breath Volume	6.484	10:17
Air Blank	0.000	10:17
Air Blank	0.000	10:19
Subject Sample #2	0.018	10:20
Breath Volume	5.113	10:20
Air Blank	0.000	10:20
Control Test	0.080	10:21
Air Blank	0.000	10:21
Diagnostics Check	OK	10:21

---

Data Download Date/Time: 08/18/2008 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE PD

Software Version: 8100.27

Breath Test Operator: SYREK, JERRY,

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: ADAMS, KENNETH

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 07/21/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 02:35

Gender: Male

Observation Period Began: 02:35

Subject's Last Name: TRAVIS

Date/Time of Test: 08/15/2008 03:15

---

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:11	
Subject Sample #1	0.184	03:12	
Breath Volume	2.238	03:12	
Air Blank	0.000	03:13	
Air Blank		03:14	RFI Detect
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

---

Data Download Date/Time: 08/18/2008 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE PD

Software Version: 8100.27

Breath Test Operator: SYREK, JERRY, L

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: HODGE, GREG

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 07/21/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: ADMIN

Arrest Time: 03:00

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: BROADNAX

Date/Time of Test: 08/15/2008 03:28

---

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	RFI Detect
Air Blank	0.000	03:28	

---

Data Download Date/Time: 08/18/2008 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE PD

Software Version: 8100.27

Breath Test Operator: SYREK, JERRY, L

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: HODGE, GREG

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 07/21/2008

Date or Time Changed: Yes

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: ADMIN

Arrest Time: 03:00

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: BROADNAX

Date/Time of Test: 08/15/2008 03:39

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample #1	0.142	03:35
Breath Volume	2.671	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample #2	0.124	03:38
Breath Volume	2.937	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 08/18/2008 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002171

Instrument Registered To: FL STATE UNIV PD  
Software Version: 8100.27  
Breath Test Operator: CLARK, ALAN, J  
Breath Test Operator Agency: FL State University  
Arresting Officer: SCHULZE, JON  
Arresting Officer Agency: Tallahassee PD  
Last Agency Inspection: 07/01/2008

Date or Time Changed: No  
Dry Gas Standard Lot #: 812101G  
Dry Gas Expiry Date: 05/06/2010  
Violation Code: DUI  
Arrest Time: 01:28  
Gender: Male  
Observation Period Began: 01:31

Subject's Last Name: SQUIRES

Date/Time of Test: 08/20/2008 02:08

---

Diagnostics Check	OK	02:07	
Air Blank		02:08	RFI Detect
Air Blank		02:08	Ambient Fail

---

Data Download Date/Time: 08/29/2008 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MILLAN, KEVIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 07/24/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: SYS CHECK

Arrest Time: 06:50

Gender: Male

Observation Period Began: 07:00

Subject's Last Name: TEST

Date/Time of Test: 08/01/2008 07:48

---

Diagnostics Check	OK	07:42
Air Blank	0.000	07:42
Control Test	0.080	07:43
Air Blank	0.000	07:43
Subject Sample #1	0.000	07:43
Breath Volume	1.535	07:43
Air Blank	0.000	07:44
Air Blank	0.000	07:46
Subject Sample #2	0.000	07:46
Breath Volume	1.222	07:46
Air Blank	0.000	07:47
Control Test	0.079	07:47
Air Blank	0.000	07:47
Diagnostics Check	OK	07:48

---

Data Download Date/Time: 08/01/2008 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 02:49

Gender: Female

Observation Period Began: 02:59

Subject's Last Name: DE LA FUENTE

Date/Time of Test: 08/02/2008 03:28

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample #1	0.203	03:22
Breath Volume	1.445	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample #2	0.205	03:26
Breath Volume	1.585	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:20

Gender: Female

Observation Period Began: 03:58

Subject's Last Name: BRIDON

Date/Time of Test: 08/02/2008 04:22

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.079	04:19	
Air Blank	0.000	04:20	
Subject Sample #1		04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:10

Gender: Male

Observation Period Began: 05:15

Subject's Last Name: DUFANAL

Date/Time of Test: 08/02/2008 05:48

---

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.079	05:43
Air Blank	0.000	05:43
Subject Sample #1	0.159	05:44
Breath Volume	2.210	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample #2	0.156	05:47
Breath Volume	2.203	05:47
Air Blank	0.000	05:47
Control Test	0.078	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 06:40

Gender: Female

Observation Period Began: 06:46

Subject's Last Name: SEBASTIANI

Date/Time of Test: 08/02/2008 07:09

---

Diagnostics Check	OK	07:07	
Air Blank	0.000	07:07	
Control Test	0.079	07:07	
Air Blank	0.000	07:08	
Subject Sample #1		07:08	Subject Test Refused
Breath Volume	0.000	07:08	
Air Blank	0.000	07:09	
Control Test	0.077	07:09	
Air Blank	0.000	07:09	
Diagnostics Check	OK	07:09	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, FREDERICK, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SAZONOV, VLADISLAV

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 18:30

Gender: Male

Observation Period Began: 18:59

Subject's Last Name: JIMENO

Date/Time of Test: 08/02/2008 19:22

---

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:20	
Control Test	0.080	19:20	
Air Blank	0.000	19:21	
Subject Sample #1		19:21	Subject Test Refused
Breath Volume	0.000	19:21	
Air Blank	0.000	19:21	
Control Test	0.080	19:22	
Air Blank	0.000	19:22	
Diagnostics Check	OK	19:22	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, FREDERICK, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, FREDERICK

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 21:35

Gender: Male

Observation Period Began: 21:44

Subject's Last Name: REYNOLDS

Date/Time of Test: 08/02/2008 22:12

---

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Subject Sample #1	0.183	22:07
Breath Volume	1.363	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample #2	0.180	22:10
Breath Volume	2.273	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: HUMPHRY, K

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 01:08

Gender: Male

Observation Period Began: 01:19

Subject's Last Name: MARTIN

Date/Time of Test: 08/03/2008 01:56

---

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample #1	0.150	01:51
Breath Volume	1.992	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample #2	0.153	01:54
Breath Volume	1.582	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BARRETO-VO, LIMURA

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:56

Gender: Female

Observation Period Began: 06:04

Subject's Last Name: CONLEY

Date/Time of Test: 08/03/2008 06:52

---

Diagnostics Check	OK	06:39	
Air Blank	0.000	06:39	
Control Test	0.079	06:39	
Air Blank	0.000	06:40	
Subject Sample #1	0.271	06:43	
Breath Volume	1.285	06:43	
Air Blank	0.000	06:44	
Air Blank	0.000	06:45	
Subject Sample #2	0.296	06:46	
Breath Volume	1.988	06:46	
Air Blank	0.000	06:47	
Air Blank	0.000	06:48	
Subject Sample #3	0.276	06:50	
Breath Volume	1.613	06:50	
Air Blank	0.000	06:51	
Control Test	0.074	06:52	Control Outside Tolerance
Air Blank	0.000	06:52	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BARRETO-VO, LIMURA

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:56

Gender: Female

Observation Period Began: 06:04

Subject's Last Name: CONLEY

Date/Time of Test: 08/03/2008 07:01

---

Diagnostics Check	OK	06:54
Air Blank	0.000	06:55
Control Test	0.076	06:55
Air Blank	0.000	06:55
Subject Sample #1	0.270	06:57
Breath Volume	1.531	06:57
Air Blank	0.000	06:58
Air Blank	0.000	06:59
Subject Sample #2	0.287	06:59
Breath Volume	1.156	06:59
Air Blank	0.000	07:00
Control Test	0.075	07:01
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WILDER, NEWELL

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: Yes

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:54

Gender: Female

Observation Period Began: 04:48

Subject's Last Name: DANS

Date/Time of Test: 08/04/2008 05:15

---

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:12	
Control Test	0.079	05:13	
Air Blank	0.000	05:13	
Subject Sample #1		05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.080	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: HUMPHRY, KERRY, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BROOKS, W

Arresting Officer Agency: Surfside PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:00

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: PAULETTI

Date/Time of Test: 08/05/2008 03:53

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample #1	0.013	03:49
Breath Volume	1.656	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample #2	0.013	03:52
Breath Volume	1.386	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:09

Gender: Male

Observation Period Began: 04:00

Subject's Last Name: FIELDS

Date/Time of Test: 08/06/2008 05:02

---

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.080	05:00	
Air Blank	0.000	05:01	
Subject Sample #1		05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Control Test	0.080	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 04:07

Gender: Male

Observation Period Began: 04:14

Subject's Last Name: MATSAFI

Date/Time of Test: 08/07/2008 04:37

---

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:35	
Control Test	0.079	04:35	
Air Blank	0.000	04:36	
Subject Sample #1		04:36	Subject Test Refused
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Control Test	0.078	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERTO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 04:50

Gender: Female

Observation Period Began: 05:15

Subject's Last Name: TILLMAN

Date/Time of Test: 08/07/2008 05:45

---

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.079	05:38
Air Blank	0.000	05:39
Subject Sample #1	0.143	05:40
Breath Volume	1.375	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:42
Subject Sample #2	0.144	05:43
Breath Volume	1.757	05:43
Air Blank	0.000	05:44
Control Test	0.078	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:45

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FORTE, FRANKLY

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 04:09

Gender: Male

Observation Period Began: 04:24

Subject's Last Name: MITAT

Date/Time of Test: 08/08/2008 04:51

---

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:48	
Control Test	0.080	04:49	
Air Blank	0.000	04:49	
Subject Sample #1		04:49	Subject Test Refused
Breath Volume	0.000	04:49	
Air Blank	0.000	04:50	
Control Test	0.079	04:50	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERTO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 04:54

Gender: Female

Observation Period Began: 05:06

Subject's Last Name: PEREZ

Date/Time of Test: 08/08/2008 05:41

---

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.079	05:39	
Air Blank	0.000	05:40	
Subject Sample #1		05:40	Subject Test Refused
Breath Volume	0.000	05:40	
Air Blank	0.000	05:40	
Control Test	0.080	05:41	
Air Blank	0.000	05:41	
Diagnostics Check	OK	05:41	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:55

Gender: Male

Observation Period Began: 04:02

Subject's Last Name: ROACH

Date/Time of Test: 08/09/2008 04:39

---

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.079	04:37	
Air Blank	0.000	04:37	
Subject Sample #1		04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, P

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:25

Gender: Male

Observation Period Began: 05:37

Subject's Last Name: MEJIA

Date/Time of Test: 08/09/2008 06:00

---

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:58	
Control Test	0.080	05:58	
Air Blank	0.000	05:59	
Subject Sample #1		05:59	Subject Test Refused
Breath Volume	0.000	05:59	
Air Blank	0.000	06:00	
Control Test	0.079	06:00	
Air Blank	0.000	06:00	
Diagnostics Check	OK	06:00	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 04:02

Gender: Male

Observation Period Began: 04:16

Subject's Last Name: PORTILLO-HERNANDEZ

Date/Time of Test: 08/10/2008 04:43

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample #1	0.100	04:38
Breath Volume	2.097	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample #2	0.093	04:41
Breath Volume	2.132	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BROOKS, W

Arresting Officer Agency: Sunrise PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 00:59

Gender: Male

Observation Period Began: 01:23

Subject's Last Name: PORTELLA

Date/Time of Test: 08/11/2008 01:53

---

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Subject Sample #1	0.095	01:46
Breath Volume	2.472	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample #2	0.089	01:51
Breath Volume	1.816	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:45

Gender: Male

Observation Period Began: 04:05

Subject's Last Name: LEE-LUM

Date/Time of Test: 08/11/2008 04:30

---

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:28	
Control Test	0.079	04:28	
Air Blank	0.000	04:29	
Subject Sample #1		04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.080	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WILDER, NEWELL

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:57

Gender: Female

Observation Period Began: 06:47

Subject's Last Name: CASANOVA RODRIGUEZ

Date/Time of Test: 08/12/2008 07:25

---

Diagnostics Check	OK	07:17
Air Blank	0.000	07:18
Control Test	0.079	07:18
Air Blank	0.000	07:19
Subject Sample #1	0.121	07:20
Breath Volume	1.511	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample #2	0.126	07:23
Breath Volume	1.621	07:23
Air Blank	0.000	07:24
Control Test	0.076	07:24
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WILDER, NEWELL

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 02:18

Gender: Female

Observation Period Began: 02:41

Subject's Last Name: ALEXANDER

Date/Time of Test: 08/14/2008 03:12

---

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Subject Sample #1		03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.080	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 01:47

Gender: Male

Observation Period Began: 02:18

Subject's Last Name: SINIBALDI-PALACIOS

Date/Time of Test: 08/15/2008 02:42

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.079	02:39	
Air Blank	0.000	02:39	
Subject Sample #1		02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BARRETO-VO, LIMURA

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:37

Gender: Male

Observation Period Began: 03:52

Subject's Last Name: BRILL

Date/Time of Test: 08/15/2008 04:27

---

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Subject Sample #1		04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.080	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERTO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:31

Gender: Male

Observation Period Began: 05:38

Subject's Last Name: LAINETTE

Date/Time of Test: 08/15/2008 06:19

---

Diagnostics Check	OK	06:16	
Air Blank	0.000	06:17	
Control Test	0.079	06:17	
Air Blank	0.000	06:17	
Subject Sample #1		06:18	Subject Test Refused
Breath Volume	0.000	06:18	
Air Blank	0.000	06:18	
Control Test	0.079	06:18	
Air Blank	0.000	06:19	
Diagnostics Check	OK	06:19	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 01:03

Gender: Male

Observation Period Began: 01:23

Subject's Last Name: BELGRANO

Date/Time of Test: 08/16/2008 01:54

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Subject Sample #1		01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.079	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FORTE, F

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 01:22

Gender: Female

Observation Period Began: 01:55

Subject's Last Name: AMUCHASTEGUI

Date/Time of Test: 08/16/2008 03:17

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Subject Sample #1		03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BROOKS, WOODWARD

Arresting Officer Agency: Surfside PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:10

Gender: Female

Observation Period Began: 05:40

Subject's Last Name: PAREJA

Date/Time of Test: 08/16/2008 06:23

---

Diagnostics Check	OK	06:20	
Air Blank	0.000	06:21	
Control Test	0.080	06:21	
Air Blank	0.000	06:21	
Subject Sample #1		06:22	Subject Test Refused
Breath Volume	0.000	06:22	
Air Blank	0.000	06:22	
Control Test	0.080	06:22	
Air Blank	0.000	06:23	
Diagnostics Check	OK	06:23	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FORTE, F

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:39

Gender: Female

Observation Period Began: 06:30

Subject's Last Name: CASTRO

Date/Time of Test: 08/16/2008 07:09

---

Diagnostics Check	OK	07:06	
Air Blank	0.000	07:07	
Control Test	0.079	07:07	
Air Blank	0.000	07:07	
Subject Sample #1		07:08	Subject Test Refused
Breath Volume	0.000	07:08	
Air Blank	0.000	07:08	
Control Test	0.079	07:08	
Air Blank	0.000	07:09	
Diagnostics Check	OK	07:09	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:21

Gender: Female

Observation Period Began: 03:56

Subject's Last Name: KHORIATY

Date/Time of Test: 08/17/2008 04:36

---

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.079	04:34	
Air Blank	0.000	04:35	
Subject Sample #1		04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 02:18

Gender: Male

Observation Period Began: 02:30

Subject's Last Name: LOPEZ

Date/Time of Test: 08/18/2008 03:03

---

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.079	03:00	
Air Blank	0.000	03:01	
Subject Sample #1		03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:03	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:12

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: URDANETA ROSALES

Date/Time of Test: 08/20/2008 03:48

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample #1	0.172	03:43
Breath Volume	2.632	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample #2	0.160	03:46
Breath Volume	1.757	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERTO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 04:13

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: SANDOVAL

Date/Time of Test: 08/20/2008 04:47

---

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Subject Sample #1	0.102	04:43
Breath Volume	2.089	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample #2	0.101	04:46
Breath Volume	3.507	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 04:27

Gender: Male

Observation Period Began: 04:37

Subject's Last Name: DELACRUZ

Date/Time of Test: 08/20/2008 05:00

---

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.079	04:58	
Air Blank	0.000	04:58	
Subject Sample #1		04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.078	04:59	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:00	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: CAREAGA, JOSE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:52

Gender: Female

Observation Period Began: 04:05

Subject's Last Name: AROCENA

Date/Time of Test: 08/21/2008 04:29

---

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.079	04:27	
Air Blank	0.000	04:27	
Subject Sample #1		04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Control Test	0.079	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERTO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:00

Gender: Male

Observation Period Began: 05:07

Subject's Last Name: CHICLY

Date/Time of Test: 08/21/2008 05:34

---

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:28	
Control Test	0.079	05:28	
Air Blank	0.000	05:29	
Subject Sample #1	0.048	05:30	
Breath Volume	2.101	05:30	
Air Blank	0.000	05:31	
Air Blank	0.000	05:32	
Subject Sample #2		05:33	Subject Test Refused
Breath Volume	0.000	05:33	
Air Blank	0.000	05:33	
Control Test	0.079	05:33	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:33

Gender: Male

Observation Period Began: 05:47

Subject's Last Name: DEPASQUALE

Date/Time of Test: 08/21/2008 06:13

---

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:11	
Control Test	0.079	06:11	
Air Blank	0.000	06:12	
Subject Sample #1		06:12	Subject Test Refused
Breath Volume	0.000	06:12	
Air Blank	0.000	06:12	
Control Test	0.079	06:13	
Air Blank	0.000	06:13	
Diagnostics Check	OK	06:13	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 23:10

Gender: Male

Observation Period Began: 23:25

Subject's Last Name: ALONSO

Date/Time of Test: 08/22/2008 00:06

---

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.079	00:03	
Air Blank	0.000	00:04	
Subject Sample #1		00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:06	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MEDINA, A

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 02:49

Gender: Female

Observation Period Began: 02:58

Subject's Last Name: MARCHANT

Date/Time of Test: 08/22/2008 03:25

---

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Subject Sample #1		03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:25	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERTO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:10

Gender: Male

Observation Period Began: 05:35

Subject's Last Name: AVELLO

Date/Time of Test: 08/22/2008 06:01

---

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:59	
Control Test	0.079	05:59	
Air Blank	0.000	06:00	
Subject Sample #1		06:00	Subject Test Refused
Breath Volume	0.000	06:00	
Air Blank	0.000	06:00	
Control Test	0.080	06:01	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, P

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:29

Gender: Female

Observation Period Began: 03:40

Subject's Last Name: NUNEZ

Date/Time of Test: 08/23/2008 04:09

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample #1	0.148	04:04
Breath Volume	1.296	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample #2	0.149	04:07
Breath Volume	1.117	04:07
Air Blank	0.000	04:08
Control Test	0.077	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:50

Gender: Female

Observation Period Began: 04:09

Subject's Last Name: ENZINNA

Date/Time of Test: 08/23/2008 04:37

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.078	04:31
Air Blank	0.000	04:31
Subject Sample #1	0.150	04:32
Breath Volume	1.414	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample #2	0.150	04:35
Breath Volume	1.640	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERT

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:10

Gender: Male

Observation Period Began: 05:32

Subject's Last Name: PENDLETON

Date/Time of Test: 08/23/2008 05:55

---

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:53	
Subject Sample #1		05:54	Subject Test Refused
Breath Volume	0.000	05:54	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank	0.000	05:55	
Diagnostics Check	OK	05:55	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 02:44

Gender: Male

Observation Period Began: 03:04

Subject's Last Name: GARCIA-FRANCO

Date/Time of Test: 08/24/2008 03:42

---

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.077	03:35
Air Blank	0.000	03:35
Subject Sample #1	0.254	03:36
Breath Volume	2.039	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample #2	0.256	03:40
Breath Volume	1.582	03:40
Air Blank	0.000	03:41
Control Test	0.077	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BLACIO, G

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 05:11

Gender: Female

Observation Period Began: 05:23

Subject's Last Name: BLYTHE

Date/Time of Test: 08/24/2008 05:51

---

Diagnostics Check	OK	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:45
Subject Sample #1	0.178	05:46
Breath Volume	1.398	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample #2	0.171	05:50
Breath Volume	1.316	05:50
Air Blank	0.000	05:50
Control Test	0.077	05:51
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: VALDES, MANUEL,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: THOMAS, SHELTON

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 19:25

Gender: Male

Observation Period Began: 19:35

Subject's Last Name: RANCH

Date/Time of Test: 08/25/2008 20:08

---

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.079	20:02
Air Blank	0.000	20:03
Subject Sample #1	0.324	20:03
Breath Volume	2.484	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample #2	0.325	20:06
Breath Volume	2.316	20:06
Air Blank	0.000	20:07
Control Test	0.078	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 03:41

Gender: Male

Observation Period Began: 03:55

Subject's Last Name: CAMPBELL

Date/Time of Test: 08/27/2008 04:23

---

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample #1	0.123	04:18
Breath Volume	2.089	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample #2	0.123	04:21
Breath Volume	2.585	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: PEREZ, LUIS, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BROOKS, WOODWARD

Arresting Officer Agency: Surfside PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 04:38

Gender: Male

Observation Period Began: 05:02

Subject's Last Name: O'CONNELL

Date/Time of Test: 08/27/2008 05:39

---

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.079	05:33
Air Blank	0.000	05:33
Subject Sample #1	0.158	05:35
Breath Volume	1.496	05:35
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample #2	0.146	05:38
Breath Volume	1.375	05:38
Air Blank	0.000	05:38
Control Test	0.077	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, A

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 8046011

Dry Gas Expiry Date: 02/18/2010

Violation Code: DUI

Arrest Time: 06:30

Gender: Male

Observation Period Began: 06:50

Subject's Last Name: HIDROBO

Date/Time of Test: 08/28/2008 07:18

---

Diagnostics Check	OK	07:15	
Air Blank	0.000	07:15	
Control Test	0.079	07:16	
Air Blank	0.000	07:16	
Subject Sample #1		07:16	Subject Test Refused
Breath Volume	0.000	07:16	
Air Blank	0.000	07:17	
Control Test	0.079	07:17	
Air Blank	0.000	07:18	
Diagnostics Check	OK	07:18	

---

Data Download Date/Time: 08/28/2008 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002462

Instrument Registered To: HIALEAH P. D.

Software Version: 8100.27

Breath Test Operator: MORALES, DAVID, J

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MORENO, JESSENIA

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 20:59

Gender: Male

Observation Period Began: 21:56

Subject's Last Name: MARTINEZ

Date/Time of Test: 08/01/2008 22:21

---

Diagnostics Check	Fail	22:20	Temperature Regulation Test
Air Blank	0.000	22:21	

---

Data Download Date/Time: 08/27/2008 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002462

Instrument Registered To: HIALEAH P. D.

Software Version: 8100.27

Breath Test Operator: MORALES, DAVID, J

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MORENO, JESSENIA

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 20:59

Gender: Male

Observation Period Began: 21:56

Subject's Last Name: MARTINEZ

Date/Time of Test: 08/01/2008 22:41

---

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.077	22:30	
Air Blank	0.000	22:30	
Subject Sample #1	0.140	22:33	Volume Not Met
Breath Volume	0.109	22:33	
Air Blank	0.000	22:34	
Air Blank	0.000	22:36	
Subject Sample #2	0.171	22:39	Volume Not Met
Breath Volume	0.207	22:39	
Air Blank	0.000	22:40	
Control Test	0.076	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:41	

---

Data Download Date/Time: 08/27/2008 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002462

Instrument Registered To: HIALEAH P. D.

Software Version: 8100.27

Breath Test Operator: MORALES, DAVID, J

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MORENO, DARWIN

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 20:59

Gender: Male

Observation Period Began: 21:56

Subject's Last Name: MARTINEZ

Date/Time of Test: 08/01/2008 22:48

---

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.076	22:45	
Air Blank	0.000	22:46	
Subject Sample #1		22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:47	
Control Test	0.076	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:48	

---

Data Download Date/Time: 08/27/2008 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002462

Instrument Registered To: HIALEAH P. D.

Software Version: 8100.27

Breath Test Operator: FAJARDO, N,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: GARCIA, L

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 04:08

Gender: Male

Observation Period Began: 04:15

Subject's Last Name: DE ARMAS

Date/Time of Test: 08/14/2008 04:50

---

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.077	04:37	
Air Blank	0.000	04:37	
Subject Sample #1	0.255	04:40	Volume Not Met
Breath Volume	0.031	04:40	
Air Blank	0.000	04:41	
Air Blank	0.000	04:43	
Subject Sample #2	0.243	04:45	
Breath Volume	3.105	04:45	
Air Blank	0.000	04:46	
Air Blank	0.000	04:47	
Subject Sample #3	0.269	04:48	
Breath Volume	4.328	04:48	
Air Blank	0.000	04:49	
Control Test	0.071	04:49	Control Outside Tolerance
Air Blank	0.000	04:50	

---

Data Download Date/Time: 08/27/2008 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002462

Instrument Registered To: HIALEAH P. D.

Software Version: 8100.27

Breath Test Operator: FAJARDO, N,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: BANDIERA, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 05:25

Gender: Male

Observation Period Began: 05:24

Subject's Last Name: FRANCO

Date/Time of Test: 08/16/2008 06:00

---

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:57	
Control Test	0.077	05:57	
Air Blank	0.000	05:58	
Subject Sample #1		05:58	Subject Test Refused
Breath Volume	0.000	05:58	
Air Blank	0.000	05:59	
Control Test	0.077	05:59	
Air Blank	0.000	05:59	
Diagnostics Check	OK	06:00	

---

Data Download Date/Time: 08/27/2008 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MONIER, J

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 03:40

Gender: Male

Observation Period Began: 03:54

Subject's Last Name: GUERRA-ZAYAS

Date/Time of Test: 08/02/2008 04:29

---

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.078	04:16	
Air Blank	0.000	04:16	
Subject Sample #1	0.158	04:19	Volume Not Met
Breath Volume	0.085	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample #2	0.128	04:24	
Breath Volume	1.675	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:26	
Subject Sample #3	0.140	04:27	
Breath Volume	1.289	04:27	
Air Blank	0.000	04:28	
Control Test	0.073	04:28	Control Outside Tolerance
Air Blank	0.000	04:29	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MONIER, J

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 03:40

Gender: Male

Observation Period Began: 03:54

Subject's Last Name: GUERRA-ZAYAS

Date/Time of Test: 08/02/2008 04:38

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.077	04:32
Air Blank	0.000	04:32
Subject Sample #1	0.147	04:33
Breath Volume	1.316	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample #2	0.140	04:36
Breath Volume	1.519	04:36
Air Blank	0.000	04:37
Control Test	0.076	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: PALACIOS, JAVIER, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: PALACIOS, J

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 08:10

Gender: Female

Observation Period Began: 08:32

Subject's Last Name: TAMAYO ATEHORTUA

Date/Time of Test: 08/02/2008 09:19

---

Diagnostics Check	OK	09:16	
Air Blank	0.000	09:16	
Control Test	0.078	09:16	
Air Blank	0.000	09:17	
Subject Sample #1		09:17	Subject Test Refused
Breath Volume	0.000	09:17	
Air Blank	0.000	09:18	
Control Test	0.078	09:18	
Air Blank	0.000	09:18	
Diagnostics Check	OK	09:19	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CONSALVO, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 01:53

Subject's Last Name: ALFONSO GRANADOS

Date/Time of Test: 08/03/2008 02:36

---

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.078	02:29	
Air Blank	0.000	02:30	
Subject Sample #1	0.172	02:31	
Breath Volume	1.566	02:31	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample #2	0.166	02:34	
Breath Volume	1.875	02:34	
Air Blank	0.000	02:35	
Control Test	0.074	02:35	Control Outside Tolerance
Air Blank	0.000	02:36	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CONSALVO, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 01:30

Gender: Male

Observation Period Began: 01:53

Subject's Last Name: ALFONSO GRANADOS

Date/Time of Test: 08/03/2008 02:45

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.077	02:39
Air Blank	0.000	02:40
Subject Sample #1	0.166	02:40
Breath Volume	1.925	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample #2	0.164	02:43
Breath Volume	2.132	02:43
Air Blank	0.000	02:44
Control Test	0.076	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: TILLMON, D,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: TILLMON, D

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 01:00

Gender: Male

Observation Period Began: 01:38

Subject's Last Name: GIL

Date/Time of Test: 08/07/2008 02:03

---

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.077	02:00	
Air Blank	0.000	02:00	
Subject Sample #1		02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.078	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: MEDINA, ANTHONY,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MEDINA, ANTHONY

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 21:30

Gender: Male

Observation Period Began: 21:43

Subject's Last Name: RICARDO

Date/Time of Test: 08/07/2008 22:10

---

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
Control Test	0.078	22:06	
Air Blank	0.000	22:07	
Subject Sample #1		22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:09	
Control Test	0.078	22:09	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:10	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: MEDINA, ANTHONY,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MEDINA, ANTHONY

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 21:39

Gender: Male

Observation Period Began: 21:55

Subject's Last Name: SUAREZ

Date/Time of Test: 08/08/2008 22:22

---

Diagnostics Check	OK	22:22	
Air Blank		22:22	RFI Detect
Air Blank	0.000	22:22	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: MEDINA, ANTHONY,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MEDINA, ANTHONY

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 21:39

Gender: Male

Observation Period Began: 21:55

Subject's Last Name: SUAREZ

Date/Time of Test: 08/08/2008 22:28

---

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.078	22:25	
Air Blank	0.000	22:26	
Subject Sample #1		22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.078	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: TILLMON, D,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: TILLMON, D

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:37

Subject's Last Name: TARIFA

Date/Time of Test: 08/09/2008 02:03

---

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.078	02:01	
Air Blank	0.000	02:01	
Subject Sample #1		02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CORREA, C.,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CORREA, C.

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 04:20

Gender: Male

Observation Period Began: 04:35

Subject's Last Name: MARIN

Date/Time of Test: 08/09/2008 05:03

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.078	04:57
Air Blank	0.000	04:57
Subject Sample #1	0.130	04:59
Breath Volume	2.421	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample #2	0.128	05:02
Breath Volume	1.968	05:02
Air Blank	0.000	05:02
Control Test	0.076	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: PALACIOS, JAVIER, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: PALACIOS, J

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 07:15

Gender: Male

Observation Period Began: 07:42

Subject's Last Name: RODRIGUEZ

Date/Time of Test: 08/10/2008 08:21

---

Diagnostics Check	OK	08:15
Air Blank	0.000	08:15
Control Test	0.078	08:16
Air Blank	0.000	08:16
Subject Sample #1	0.176	08:17
Breath Volume	1.679	08:17
Air Blank	0.000	08:17
Air Blank	0.000	08:19
Subject Sample #2	0.186	08:20
Breath Volume	3.187	08:20
Air Blank	0.000	08:21
Control Test	0.076	08:21
Air Blank	0.000	08:21
Diagnostics Check	OK	08:21

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: JANVIER, FRITZ, G

Breath Test Operator Agency: Hialeah PD

Arresting Officer: JANVIER, FRITZ

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 00:12

Gender: Female

Observation Period Began: 01:15

Subject's Last Name: CANETE

Date/Time of Test: 08/14/2008 01:41

---

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.078	01:37	
Air Blank	0.000	01:38	
Subject Sample #1		01:39	Subject Test Refused
Breath Volume	0.183	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: ENRIQUEZ, NELSON,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: PEREZ, SERGIO

Arresting Officer Agency: Opa Locka PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 12:12

Gender: Male

Observation Period Began: 13:05

Subject's Last Name: MERIDA

Date/Time of Test: 08/19/2008 13:36

---

Diagnostics Check	OK	13:30
Air Blank	0.000	13:30
Control Test	0.077	13:30
Air Blank	0.000	13:31
Subject Sample #1	0.176	13:31
Breath Volume	1.425	13:31
Air Blank	0.000	13:32
Air Blank	0.000	13:34
Subject Sample #2	0.166	13:34
Breath Volume	2.449	13:34
Air Blank	0.000	13:35
Control Test	0.076	13:35
Air Blank	0.000	13:36
Diagnostics Check	OK	13:36

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: JORDANA, D.,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: JORDANA, D.

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 19:30

Gender: Male

Observation Period Began: 20:07

Subject's Last Name: SALOMON

Date/Time of Test: 08/20/2008 20:53

---

Diagnostics Check	OK	20:47
Air Blank	0.000	20:48
Control Test	0.079	20:48
Air Blank	0.000	20:48
Subject Sample #1	0.132	20:49
Breath Volume	1.769	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:51
Subject Sample #2	0.132	20:52
Breath Volume	1.726	20:52
Air Blank	0.000	20:52
Control Test	0.076	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: WILLIAMS, J,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: WILLIAMS, J

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 00:35

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: LEAL PEDRAZA

Date/Time of Test: 08/21/2008 01:40

---

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.078	01:38	
Air Blank	0.000	01:38	
Subject Sample #1		01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CONSALVO, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 01:50

Subject's Last Name: GUIROLA QUINTERO

Date/Time of Test: 08/22/2008 02:21

---

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.078	02:18	
Air Blank	0.000	02:18	
Subject Sample #1		02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: GONZALEZ, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 05:15

Gender: Male

Observation Period Began: 05:29

Subject's Last Name: HERNANDEZ MARTINEZ

Date/Time of Test: 08/22/2008 06:04

---

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:58	
Control Test	0.078	05:58	
Air Blank	0.000	05:59	
Subject Sample #1	0.187	06:00	
Breath Volume	3.597	06:00	
Air Blank	0.000	06:00	
Air Blank	0.000	06:02	
Subject Sample #2		06:02	RFI Detect
Breath Volume	0.000	06:02	
Air Blank	0.000	06:03	
Control Test	0.075	06:03	
Air Blank	0.000	06:04	
Diagnostics Check	OK	06:04	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: GONZALEZ, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 05:15

Gender: Male

Observation Period Began: 05:29

Subject's Last Name: HERNANDEZ MARTINEZ

Date/Time of Test: 08/22/2008 06:12

---

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.078	06:06
Air Blank	0.000	06:06
Subject Sample #1	0.176	06:07
Breath Volume	3.207	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample #2	0.182	06:11
Breath Volume	3.371	06:11
Air Blank	0.000	06:11
Control Test	0.076	06:12
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CONSALVO, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 04:50

Gender: Male

Observation Period Began: 05:30

Subject's Last Name: HERNAN

Date/Time of Test: 08/22/2008 06:23

---

Diagnostics Check	OK	06:17
Air Blank	0.000	06:17
Control Test	0.077	06:17
Air Blank	0.000	06:18
Subject Sample #1	0.248	06:19
Breath Volume	3.406	06:19
Air Blank	0.000	06:19
Air Blank	0.000	06:21
Subject Sample #2	0.234	06:22
Breath Volume	3.250	06:22
Air Blank	0.000	06:22
Control Test	0.075	06:23
Air Blank	0.000	06:23
Diagnostics Check	OK	06:23

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: GUTIERREZ, JAMES, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: ACAY, AZAEL

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 21:00

Gender: Male

Observation Period Began: 21:10

Subject's Last Name: HERNANDEZ

Date/Time of Test: 08/22/2008 21:47

---

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.077	21:45	
Air Blank	0.000	21:45	
Subject Sample #1		21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:46	
Control Test	0.078	21:46	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: LARA, P

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 03:36

Gender: Male

Observation Period Began: 05:40

Subject's Last Name: VALDEZ-VALHUERDIS

Date/Time of Test: 08/23/2008 06:11

---

Diagnostics Check	OK	06:02	
Air Blank	0.000	06:03	
Control Test	0.077	06:03	
Air Blank	0.000	06:04	
Subject Sample #1	0.120	06:05	
Breath Volume	1.156	06:05	
Air Blank	0.000	06:06	
Air Blank	0.000	06:07	
Subject Sample #2		06:09	Subject Test Refused
Breath Volume	0.378	06:09	
Air Blank	0.000	06:10	
Control Test	0.075	06:10	
Air Blank	0.000	06:11	
Diagnostics Check	OK	06:11	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: GARCIA, L

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 05:30

Gender: Male

Observation Period Began: 06:25

Subject's Last Name: ALTAMIRANO

Date/Time of Test: 08/23/2008 06:59

---

Diagnostics Check	OK	06:54	
Air Blank	0.000	06:54	
Control Test	0.077	06:54	
Air Blank	0.000	06:55	
Subject Sample #1		06:57	Subject Test Refused
Breath Volume	0.000	06:57	
Air Blank	0.000	06:58	
Control Test	0.077	06:58	
Air Blank	0.000	06:58	
Diagnostics Check	OK	06:59	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: GARCIA, L

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 05:30

Gender: Male

Observation Period Began: 06:25

Subject's Last Name: ALTAMIRANO

Date/Time of Test: 08/23/2008 07:08

---

Diagnostics Check	OK	07:00	
Air Blank	0.000	07:00	
Control Test	0.077	07:01	
Air Blank	0.000	07:01	
Subject Sample #1	0.107	07:03	
Breath Volume	1.453	07:03	
Air Blank	0.000	07:04	
Air Blank	0.000	07:06	
Subject Sample #2	0.118	07:06	
Breath Volume	1.589	07:06	
Air Blank	0.000	07:07	
Control Test	0.074	07:07	Control Outside Tolerance
Air Blank	0.000	07:08	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: GARCIA, L

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 05:30

Gender: Male

Observation Period Began: 06:25

Subject's Last Name: ALTAMIRANO

Date/Time of Test: 08/23/2008 07:19

---

Diagnostics Check	OK	07:09	
Air Blank	0.000	07:10	
Control Test	0.077	07:10	
Air Blank	0.000	07:11	
Subject Sample #1	0.121	07:13	
Breath Volume	1.238	07:13	
Air Blank	0.000	07:14	
Air Blank	0.000	07:16	
Subject Sample #2	0.112	07:17	
Breath Volume	1.691	07:17	
Air Blank	0.000	07:18	
Control Test	0.074	07:18	Control Outside Tolerance
Air Blank	0.000	07:19	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CONSALVO, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 00:30

Gender: Male

Observation Period Began: 01:11

Subject's Last Name: PEREZ CONTINO

Date/Time of Test: 08/24/2008 01:38

---

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.077	01:36	
Air Blank	0.000	01:37	
Subject Sample #1		01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

---

Data Download Date/Time: 08/27/2008 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: GARCIA, JOSE, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: RODRIGUEZ, EDWARD

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 21:35

Gender: Male

Observation Period Began: 22:19

Subject's Last Name: GOMEZ AGUILAR

Date/Time of Test: 08/24/2008 22:55

---

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.077	22:46	
Air Blank	0.000	22:46	
Subject Sample #1	0.235	22:49	
Breath Volume	1.191	22:49	
Air Blank	0.000	22:50	
Air Blank	0.000	22:51	
Subject Sample #2	0.232	22:53	
Breath Volume	1.269	22:53	
Air Blank	0.000	22:54	
Control Test	0.070	22:54	Control Outside Tolerance
Air Blank	0.000	22:55	

---

Data Download Date/Time: 08/27/2008 09:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: TILLMON, D,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: TILLMON, D

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 08/27/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: REVE AVILES

Date/Time of Test: 08/28/2008 02:11

---

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.078	02:09	
Air Blank	0.000	02:09	
Subject Sample #1		02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:10	
Control Test	0.079	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

---

Data Download Date/Time: 09/30/2008 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: CONSALVO, E, A

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CONSALVO, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 08/27/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 02:00

Subject's Last Name: ROMEROS

Date/Time of Test: 08/30/2008 02:49

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.078	02:38	
Air Blank	0.000	02:39	
Subject Sample #1		02:42	No Sample Provided
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample #2		02:47	Subject Test Refused
Breath Volume	0.230	02:47	
Air Blank	0.000	02:48	
Control Test	0.078	02:48	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

---

Data Download Date/Time: 09/30/2008 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002463

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: VECINO, YVONNE,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: REYES, ISIDRO

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 08/27/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7050011

Dry Gas Expiry Date: 02/22/2009

Violation Code: DUI

Arrest Time: 08:40

Gender: Male

Observation Period Began: 10:00

Subject's Last Name: LANZAS

Date/Time of Test: 08/30/2008 10:23

---

Diagnostics Check	OK	10:20	
Air Blank	0.000	10:20	
Control Test	0.077	10:21	
Air Blank	0.000	10:21	
Subject Sample #1		10:21	Subject Test Refused
Breath Volume	0.000	10:21	
Air Blank	0.000	10:22	
Control Test	0.079	10:22	
Air Blank	0.000	10:23	
Diagnostics Check	OK	10:23	

---

Data Download Date/Time: 09/30/2008 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002464

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: DELIMA, GENE,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CURBELO, LUIS

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7096011

Dry Gas Expiry Date: 04/08/2009

Violation Code: DUI

Arrest Time: 20:10

Gender: Male

Observation Period Began: 20:21

Subject's Last Name: JIMENEZ

Date/Time of Test: 08/03/2008 20:46

---

Diagnostics Check	OK	20:42	
Air Blank	0.000	20:43	
Control Test	0.078	20:43	
Air Blank	0.000	20:44	
Subject Sample #1		20:46	RFI Detect
Breath Volume	0.000	20:46	
Air Blank		20:46	RFI Detect
Air Blank		20:46	RFI Detect

---

Data Download Date/Time: 08/28/2008 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002464

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: DELIMA, GENE,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CURBELO, LUIS

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7096011

Dry Gas Expiry Date: 04/08/2009

Violation Code: DUI

Arrest Time: 20:10

Gender: Male

Observation Period Began: 20:20

Subject's Last Name: JIMENEZ

Date/Time of Test: 08/03/2008 20:51

---

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.077	20:49	
Air Blank	0.000	20:50	
Subject Sample #1		20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:50	
Control Test	0.077	20:51	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:51	

---

Data Download Date/Time: 08/28/2008 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002464

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: PEREZ, P, E

Breath Test Operator Agency: Hialeah PD

Arresting Officer: CLAVIJO, E

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7096011

Dry Gas Expiry Date: 04/08/2009

Violation Code: DUI

Arrest Time: 20:00

Gender: Male

Observation Period Began: 20:13

Subject's Last Name: CUELLAR ABREU

Date/Time of Test: 08/08/2008 20:44

---

Diagnostics Check	OK	20:36	
Air Blank	0.000	20:36	
Control Test	0.077	20:36	
Air Blank	0.000	20:37	
Subject Sample #1	0.328	20:40	Volume Not Met
Breath Volume	0.085	20:40	
Air Blank	0.000	20:41	
Air Blank	0.000	20:43	
Subject Sample #2		20:43	Subject Test Refused
Breath Volume	0.000	20:43	
Air Blank	0.000	20:43	
Control Test	0.075	20:44	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

---

Data Download Date/Time: 08/28/2008 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002464

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: MEDINA, ANTHONY,

Breath Test Operator Agency: Hialeah PD

Arresting Officer: MEDINA, ANTHONY

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 6311011

Dry Gas Expiry Date: 11/09/2008

Violation Code: DUI

Arrest Time: 18:29

Gender: Male

Observation Period Began: 18:48

Subject's Last Name: PRADO

Date/Time of Test: 08/14/2008 19:19

---

Diagnostics Check	OK	19:09
Air Blank	0.000	19:09
Control Test	0.077	19:09
Air Blank	0.000	19:10
Subject Sample #1	0.260	19:12
Breath Volume	1.398	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample #2	0.261	19:17
Breath Volume	1.359	19:17
Air Blank	0.000	19:18
Control Test	0.076	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

---

Data Download Date/Time: 08/28/2008 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002464

Instrument Registered To: HIALEAH P.D.

Software Version: 8100.27

Breath Test Operator: COLON, MANUEL, X

Breath Test Operator Agency: Hialeah PD

Arresting Officer: ROBINSON, K

Arresting Officer Agency: Hialeah PD

Last Agency Inspection: 07/28/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 6311011

Dry Gas Expiry Date: 11/09/2008

Violation Code: DUI

Arrest Time: 20:31

Gender: Male

Observation Period Began: 21:05

Subject's Last Name: CURBELO

Date/Time of Test: 08/20/2008 21:33

---

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.078	21:31	
Air Blank	0.000	21:32	
Subject Sample #1		21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:32	
Control Test	0.077	21:33	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

---

Data Download Date/Time: 08/28/2008 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002566

Instrument Registered To: SURFSIDE PD

Software Version: 8100.27

Breath Test Operator: HACKNEY, DWITE, N

Breath Test Operator Agency: FDLE

Arresting Officer: HACKNEY, DWITE

Arresting Officer Agency: FDLE

Last Agency Inspection: 07/29/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7096011

Dry Gas Expiry Date: 04/08/2009

Violation Code: SYS CHECK

Arrest Time: 13:00

Gender: Female

Observation Period Began: 13:00

Subject's Last Name: LICENSE

Date/Time of Test: 08/25/2008 13:37

---

Diagnostics Check	OK	13:31
Air Blank	0.000	13:32
Control Test	0.082	13:32
Air Blank	0.000	13:33
Subject Sample #1	0.000	13:33
Breath Volume	1.992	13:33
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample #2	0.000	13:36
Breath Volume	2.171	13:36
Air Blank	0.000	13:36
Control Test	0.082	13:37
Air Blank	0.000	13:37
Diagnostics Check	OK	13:37

---

Data Download Date/Time: 08/28/2008 10:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002987

Instrument Registered To: MIAMI GARDENS PD

Software Version: 8100.27

Breath Test Operator: HACKNEY, DWITE, N

Breath Test Operator Agency: FDLE

Arresting Officer: HACKNEY, DWITE

Arresting Officer Agency: FDLE

Last Agency Inspection: 08/21/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7325011

Dry Gas Expiry Date: 11/24/2009

Violation Code: SYS CHECK

Arrest Time: 12:00

Gender: Female

Observation Period Began: 12:00

Subject's Last Name: LICENSE

Date/Time of Test: 08/21/2008 12:46

---

Diagnostics Check	OK	12:40
Air Blank	0.000	12:40
Control Test	0.077	12:41
Air Blank	0.000	12:41
Subject Sample #1	0.000	12:42
Breath Volume	3.007	12:42
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample #2	0.000	12:45
Breath Volume	1.941	12:45
Air Blank	0.000	12:45
Control Test	0.077	12:46
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

---

Data Download Date/Time: 08/21/2008 12:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002988

Instrument Registered To: MIAMI GARDENS PD

Software Version: 8100.27

Breath Test Operator: HACKNEY, DWITE, N

Breath Test Operator Agency: FDLE

Arresting Officer: HACKNEY, DWITE

Arresting Officer Agency: FDLE

Last Agency Inspection: 08/21/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7325011

Dry Gas Expiry Date: 11/24/2009

Violation Code: SYS CHECK

Arrest Time: 12:00

Gender: Female

Observation Period Began: 12:00

Subject's Last Name: LICENSE

Date/Time of Test: 08/21/2008 12:44

---

Diagnostics Check	OK	12:38
Air Blank	0.000	12:38
Control Test	0.077	12:39
Air Blank	0.000	12:39
Subject Sample #1	0.000	12:40
Breath Volume	1.933	12:40
Air Blank	0.000	12:40
Air Blank	0.000	12:42
Subject Sample #2	0.000	12:43
Breath Volume	2.367	12:43
Air Blank	0.000	12:43
Control Test	0.076	12:44
Air Blank	0.000	12:44
Diagnostics Check	OK	12:44

---

Data Download Date/Time: 08/21/2008 13:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002989

Instrument Registered To: TALLAHASSEE PD

Software Version: 8100.27

Breath Test Operator: SCOTT, COURTNEY, D

Breath Test Operator Agency: Tallahassee PD

Arresting Officer: SCOTT, COURTNEY

Arresting Officer Agency: Tallahassee PD

Last Agency Inspection: 07/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 732502I

Dry Gas Expiry Date: 11/26/2009

Violation Code: DUI

Arrest Time: 03:14

Gender: Female

Observation Period Began: 02:54

Subject's Last Name: BEVIS

Date/Time of Test: 08/16/2008 03:27

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.076	03:21
Air Blank	0.000	03:22
Subject Sample #1	0.198	03:22
Breath Volume	1.890	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample #2	0.204	03:25
Breath Volume	1.570	03:25
Air Blank	0.000	03:26
Control Test	0.075	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 08/28/2008 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002989

Instrument Registered To: TALLAHASSEE PD

Software Version: 8100.27

Breath Test Operator: SCOTT, COURTNEY, D

Breath Test Operator Agency: Tallahassee PD

Arresting Officer: SCOTT, COURTNEY

Arresting Officer Agency: Tallahassee PD

Last Agency Inspection: 07/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 732502I

Dry Gas Expiry Date: 11/26/2009

Violation Code: DUI

Arrest Time: 21:02

Gender: Male

Observation Period Began: 20:42

Subject's Last Name: JOHNSON

Date/Time of Test: 08/20/2008 21:16

---

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.076	21:10
Air Blank	0.000	21:10
Subject Sample #1	0.134	21:11
Breath Volume	1.957	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample #2	0.133	21:14
Breath Volume	2.171	21:14
Air Blank	0.000	21:15
Control Test	0.076	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

---

Data Download Date/Time: 08/28/2008 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FORTE, FRANKLY

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 07/24/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 01:32

Gender: Female

Observation Period Began: 01:48

Subject's Last Name: GUTIERREZ PROKOPOVA

Date/Time of Test: 08/01/2008 02:19

---

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.079	02:17	
Air Blank	0.000	02:17	
Subject Sample #1		02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.078	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

---

Data Download Date/Time: 08/01/2008 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 07/24/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:57

Gender: Male

Observation Period Began: 04:05

Subject's Last Name: PULEO

Date/Time of Test: 08/01/2008 04:28

---

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:26	
Control Test	0.079	04:26	
Air Blank	0.000	04:27	
Subject Sample #1		04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

---

Data Download Date/Time: 08/01/2008 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 07/24/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:57

Gender: Male

Observation Period Began: 04:17

Subject's Last Name: CASTILLO

Date/Time of Test: 08/01/2008 04:41

---

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.079	04:39	
Air Blank	0.000	04:40	
Subject Sample #1		04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.079	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

---

Data Download Date/Time: 08/01/2008 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 07/24/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 04:15

Gender: Male

Observation Period Began: 04:17

Subject's Last Name: CORTAZAR

Date/Time of Test: 08/01/2008 04:58

---

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Subject Sample #1	0.161	04:54
Breath Volume	3.472	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample #2	0.156	04:57
Breath Volume	2.636	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

---

Data Download Date/Time: 08/01/2008 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 07/24/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:53

Gender: Male

Observation Period Began: 05:58

Subject's Last Name: PULGAR

Date/Time of Test: 08/01/2008 06:27

---

Diagnostics Check	OK	06:23	
Air Blank	0.000	06:24	
Control Test	0.079	06:24	
Air Blank	0.000	06:25	
Subject Sample #1		06:26	Subject Test Refused
Breath Volume	0.000	06:26	
Air Blank	0.000	06:26	
Control Test	0.079	06:26	
Air Blank	0.000	06:27	
Diagnostics Check	OK	06:27	

---

Data Download Date/Time: 08/01/2008 09:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ROBINSON, I

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 21:10

Gender: Male

Observation Period Began: 21:28

Subject's Last Name: ONEILL

Date/Time of Test: 08/01/2008 21:58

---

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.079	21:56	
Air Blank	0.000	21:56	
Subject Sample #1		21:56	Subject Test Refused
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.079	21:57	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 02:40

Gender: Female

Observation Period Began: 02:50

Subject's Last Name: PERDOMO

Date/Time of Test: 08/02/2008 03:20

---

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:18	
Subject Sample #1		03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.079	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WIRTH, PHILIP

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:05

Gender: Female

Observation Period Began: 05:22

Subject's Last Name: HERALD

Date/Time of Test: 08/02/2008 05:56

---

Diagnostics Check	OK	05:49
Air Blank	0.000	05:49
Control Test	0.079	05:50
Air Blank	0.000	05:50
Subject Sample #1	0.211	05:51
Breath Volume	1.824	05:51
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample #2	0.191	05:54
Breath Volume	1.558	05:54
Air Blank	0.000	05:55
Control Test	0.078	05:55
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: ESPADA, KENNE, J

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ALBA, EDUARD

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 10:17

Gender: Male

Observation Period Began: 10:42

Subject's Last Name: SUAREZ

Date/Time of Test: 08/03/2008 11:33

---

Diagnostics Check	OK	11:30	
Air Blank	0.000	11:30	
Control Test	0.079	11:31	
Air Blank	0.000	11:31	
Subject Sample #1		11:31	Subject Test Refused
Breath Volume	0.000	11:31	
Air Blank	0.000	11:32	
Control Test	0.080	11:32	
Air Blank	0.000	11:33	
Diagnostics Check	OK	11:33	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 04:35

Gender: Male

Observation Period Began: 04:41

Subject's Last Name: MENDOZA

Date/Time of Test: 08/07/2008 05:11

---

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:06
Air Blank	0.000	05:06
Subject Sample #1	0.088	05:07
Breath Volume	3.949	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample #2	0.092	05:10
Breath Volume	3.878	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:14

Gender: Male

Observation Period Began: 05:23

Subject's Last Name: ESTEVES-ELESPURU

Date/Time of Test: 08/08/2008 05:49

---

Diagnostics Check	OK	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:45
Subject Sample #1	0.160	05:45
Breath Volume	2.066	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:47
Subject Sample #2	0.161	05:48
Breath Volume	2.000	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:13

Gender: Female

Observation Period Began: 05:35

Subject's Last Name: RIVERA-MELENDZ

Date/Time of Test: 08/09/2008 06:03

---

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.079	05:57
Air Blank	0.000	05:57
Subject Sample #1	0.205	05:59
Breath Volume	2.273	05:59
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample #2	0.219	06:02
Breath Volume	2.351	06:02
Air Blank	0.000	06:02
Control Test	0.079	06:03
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FORTE, F

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:36

Gender: Male

Observation Period Began: 05:57

Subject's Last Name: TOLOSA

Date/Time of Test: 08/09/2008 06:24

---

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.080	06:19
Air Blank	0.000	06:19
Subject Sample #1	0.213	06:20
Breath Volume	1.140	06:20
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample #2	0.202	06:23
Breath Volume	1.203	06:23
Air Blank	0.000	06:24
Control Test	0.079	06:24
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: POMARES, M

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:10

Gender: Male

Observation Period Began: 05:31

Subject's Last Name: CASTRO

Date/Time of Test: 08/10/2008 06:02

---

Diagnostics Check	OK	05:55
Air Blank	0.000	05:56
Control Test	0.079	05:56
Air Blank	0.000	05:57
Subject Sample #1	0.151	05:57
Breath Volume	2.058	05:57
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample #2	0.149	06:00
Breath Volume	1.636	06:00
Air Blank	0.000	06:01
Control Test	0.080	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: KUNCAS, ANDREW, J

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BROOKS, WOODWARD

Arresting Officer Agency: Surfside PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:10

Gender: Female

Observation Period Began: 03:40

Subject's Last Name: CORTEZ

Date/Time of Test: 08/12/2008 04:07

---

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Subject Sample #1	0.138	04:03
Breath Volume	1.445	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.134	04:06
Breath Volume	1.515	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: ATWELL, MATT, C

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: TAVSS, ADAM

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 21:11

Gender: Male

Observation Period Began: 21:20

Subject's Last Name: CHAPPOTIN

Date/Time of Test: 08/12/2008 22:15

---

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.080	22:11	
Air Blank	0.000	22:12	
Subject Sample #1	0.000	22:12	
Breath Volume	4.714	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:15	
Subject Sample #2		22:15	Sequence Aborted
Breath Volume	0.000	22:15	
Air Blank	0.000	22:15	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: ATWELL, MATT, C

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: TAVSS, ADAM

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 21:11

Gender: Male

Observation Period Began: 21:20

Subject's Last Name: CHAPPOTIN

Date/Time of Test: 08/12/2008 22:55

---

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Subject Sample #1	0.000	22:51
Breath Volume	3.625	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample #2	0.000	22:54
Breath Volume	3.472	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 04:01

Gender: Male

Observation Period Began: 04:07

Subject's Last Name: MESA

Date/Time of Test: 08/13/2008 04:43

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Subject Sample #1	0.062	04:38
Breath Volume	1.785	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample #2	0.051	04:41
Breath Volume	1.183	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 02:03

Gender: Male

Observation Period Began: 02:07

Subject's Last Name: BARROUKH

Date/Time of Test: 08/15/2008 02:35

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:30
Subject Sample #1	0.141	02:31
Breath Volume	1.437	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample #2	0.148	02:34
Breath Volume	2.097	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERT

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:21

Gender: Male

Observation Period Began: 03:37

Subject's Last Name: FLERIANOS

Date/Time of Test: 08/15/2008 04:02

---

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	04:00	
Air Blank	0.000	04:00	
Subject Sample #1		04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:12

Gender: Female

Observation Period Began: 05:18

Subject's Last Name: HALE

Date/Time of Test: 08/15/2008 05:46

---

Diagnostics Check	OK	05:39
Air Blank	0.000	05:40
Control Test	0.080	05:40
Air Blank	0.000	05:41
Subject Sample #1	0.102	05:41
Breath Volume	2.343	05:41
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample #2	0.099	05:44
Breath Volume	2.039	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:04

Gender: Male

Observation Period Began: 05:26

Subject's Last Name: WEISBROT

Date/Time of Test: 08/16/2008 05:54

---

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.079	05:48
Air Blank	0.000	05:48
Subject Sample #1	0.110	05:48
Breath Volume	3.277	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample #2	0.112	05:52
Breath Volume	3.425	05:52
Air Blank	0.000	05:53
Control Test	0.079	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ROBINSON, I

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 19:50

Gender: Female

Observation Period Began: 20:35

Subject's Last Name: MUNOZ

Date/Time of Test: 08/16/2008 21:07

---

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Subject Sample #1	0.200	21:02
Breath Volume	1.304	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample #2	0.196	21:05
Breath Volume	1.878	21:05
Air Blank	0.000	21:06
Control Test	0.079	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WIRTH, P

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 02:50

Gender: Male

Observation Period Began: 03:01

Subject's Last Name: REYES

Date/Time of Test: 08/17/2008 03:44

---

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:43	
Subject Sample #1		03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: POLANCO, R

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:40

Gender: Male

Observation Period Began: 04:18

Subject's Last Name: VENEGAS

Date/Time of Test: 08/17/2008 04:57

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.079	04:51
Air Blank	0.000	04:51
Subject Sample #1	0.228	04:52
Breath Volume	1.390	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:54
Subject Sample #2	0.248	04:55
Breath Volume	2.578	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: FELDMAN, STEVEN, L

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BROOKS, WOODWARD

Arresting Officer Agency: Surfside PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: Yes

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:22

Gender: Male

Observation Period Began: 05:42

Subject's Last Name: LAMADRID

Date/Time of Test: 08/17/2008 06:22

---

Diagnostics Check	OK	06:19	
Air Blank	0.000	06:20	
Control Test	0.079	06:20	
Air Blank	0.000	06:21	
Subject Sample #1		06:21	Subject Test Refused
Breath Volume	0.000	06:21	
Air Blank	0.000	06:21	
Control Test	0.078	06:22	
Air Blank	0.000	06:22	
Diagnostics Check	OK	06:22	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: VALDES, MANUEL,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MAROTTA, DANIEL

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 18:16

Gender: Male

Observation Period Began: 19:00

Subject's Last Name: CLEARY

Date/Time of Test: 08/18/2008 19:27

---

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:22
Subject Sample #1	0.190	19:22
Breath Volume	1.144	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample #2	0.192	19:26
Breath Volume	1.164	19:26
Air Blank	0.000	19:26
Control Test	0.079	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: HUMPHRY, KERRY, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:55

Gender: Male

Observation Period Began: 04:18

Subject's Last Name: TREJO

Date/Time of Test: 08/20/2008 04:57

---

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Subject Sample #1	0.155	04:53
Breath Volume	1.332	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.151	04:56
Breath Volume	1.636	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WIRTH, P

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:03

Gender: Male

Observation Period Began: 03:23

Subject's Last Name: DAIZ-DEVILLEGAS

Date/Time of Test: 08/21/2008 03:55

---

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.079	03:52	
Air Blank	0.000	03:53	
Subject Sample #1		03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SILVAGNI, R,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: PEREIRA, J

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 17:45

Gender: Male

Observation Period Began: 18:01

Subject's Last Name: LINDSAY

Date/Time of Test: 08/21/2008 18:27

---

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:25	
Control Test	0.079	18:25	
Air Blank	0.000	18:26	
Subject Sample #1		18:26	Subject Test Refused
Breath Volume	0.000	18:26	
Air Blank	0.000	18:27	
Control Test	0.080	18:27	
Air Blank	0.000	18:27	
Diagnostics Check	OK	18:27	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: VASILIEVICH, MIRKO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 00:16

Gender: Male

Observation Period Began: 00:25

Subject's Last Name: GUIMET

Date/Time of Test: 08/22/2008 00:48

---

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.079	00:46	
Air Blank	0.000	00:46	
Subject Sample #1		00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MCVEY, G

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 01:46

Gender: Male

Observation Period Began: 02:17

Subject's Last Name: MARRERO

Date/Time of Test: 08/22/2008 02:47

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:39
Air Blank	0.000	02:39
Subject Sample #1	0.183	02:42
Breath Volume	1.628	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample #2	0.183	02:45
Breath Volume	1.121	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:05

Gender: Male

Observation Period Began: 03:22

Subject's Last Name: SOLOVYOV

Date/Time of Test: 08/22/2008 03:51

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample #1	0.188	03:46
Breath Volume	2.343	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample #2	0.193	03:49
Breath Volume	2.957	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SOCARRAS, PEDRO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: SOCARRAS, PEDRO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 04:31

Gender: Male

Observation Period Began: 04:37

Subject's Last Name: CASTRO

Date/Time of Test: 08/22/2008 05:04

---

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Subject Sample #1	0.153	04:59
Breath Volume	3.621	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample #2	0.167	05:02
Breath Volume	3.046	05:02
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ROBINSON, I

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 21:51

Gender: Male

Observation Period Began: 22:10

Subject's Last Name: HOEPKER

Date/Time of Test: 08/22/2008 22:48

---

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.079	22:46	
Air Blank	0.000	22:47	
Subject Sample #1		22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
Control Test	0.080	22:48	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: BARRETO-VO, LIMURA, O

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, A

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:27

Gender: Female

Observation Period Began: 03:45

Subject's Last Name: STRANNIGAN

Date/Time of Test: 08/23/2008 04:16

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample #1	0.116	04:12
Breath Volume	1.117	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample #2	0.119	04:15
Breath Volume	1.195	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: HACKWORTH, E

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 21:10

Gender: Female

Observation Period Began: 21:53

Subject's Last Name: HALL

Date/Time of Test: 08/23/2008 22:25

---

Diagnostics Check	OK	22:19
Air Blank	0.000	22:19
Control Test	0.079	22:20
Air Blank	0.000	22:20
Subject Sample #1	0.156	22:21
Breath Volume	2.183	22:21
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample #2	0.138	22:23
Breath Volume	1.515	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: EATON, M

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:40

Gender: Male

Observation Period Began: 03:49

Subject's Last Name: BOBE

Date/Time of Test: 08/24/2008 04:22

---

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Subject Sample #1	0.097	04:17
Breath Volume	1.632	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample #2	0.099	04:20
Breath Volume	2.863	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BROOKS, W

Arresting Officer Agency: Surfside PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 02:20

Gender: Male

Observation Period Began: 02:59

Subject's Last Name: GUTIERREZ

Date/Time of Test: 08/25/2008 03:25

---

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Subject Sample #1		03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, FREDERICK, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ELMORE, PHILIP

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 19:26

Gender: Male

Observation Period Began: 19:38

Subject's Last Name: SHAW

Date/Time of Test: 08/25/2008 20:07

---

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.080	19:59
Air Blank	0.000	19:59
Subject Sample #1	0.128	20:01
Breath Volume	1.343	20:01
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample #2	0.135	20:06
Breath Volume	1.160	20:06
Air Blank	0.000	20:06
Control Test	0.078	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:59

Gender: Female

Observation Period Began: 04:10

Subject's Last Name: KESHEN

Date/Time of Test: 08/27/2008 04:39

---

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample #1	0.157	04:34
Breath Volume	1.910	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample #2	0.163	04:37
Breath Volume	1.160	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: ROBINSON, IAN, M

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WIRTH, PHILLIP

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 21:26

Gender: Female

Observation Period Began: 21:34

Subject's Last Name: PETRACCHI

Date/Time of Test: 08/27/2008 22:06

---

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	22:00
Air Blank	0.000	22:00
Subject Sample #1	0.200	22:01
Breath Volume	1.105	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:03
Subject Sample #2	0.211	22:04
Breath Volume	2.015	22:04
Air Blank	0.000	22:05
Control Test	0.078	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ESTRAVIZ, ALBERT

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 03:49

Gender: Male

Observation Period Began: 04:07

Subject's Last Name: MAZZEGA

Date/Time of Test: 08/28/2008 04:34

---

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Subject Sample #1		04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:34	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 06:40

Gender: Male

Observation Period Began: 06:54

Subject's Last Name: CARRASCO

Date/Time of Test: 08/28/2008 07:17

---

Diagnostics Check	OK	07:14	
Air Blank	0.000	07:15	
Control Test	0.079	07:15	
Air Blank	0.000	07:16	
Subject Sample #1		07:16	Subject Test Refused
Breath Volume	0.000	07:16	
Air Blank	0.000	07:16	
Control Test	0.080	07:17	
Air Blank	0.000	07:17	
Diagnostics Check	OK	07:17	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MILLAN, KEVIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: COURT ORD

Arrest Time: 11:00

Gender: Female

Observation Period Began: 11:00

Subject's Last Name: DICKOW

Date/Time of Test: 08/28/2008 11:30

---

Diagnostics Check	OK	11:24
Air Blank	0.000	11:25
Control Test	0.080	11:25
Air Blank	0.000	11:26
Subject Sample #1	0.000	11:26
Breath Volume	3.144	11:26
Air Blank	0.000	11:27
Air Blank	0.000	11:29
Subject Sample #2	0.000	11:29
Breath Volume	2.382	11:29
Air Blank	0.000	11:30
Control Test	0.079	11:30
Air Blank	0.000	11:30
Diagnostics Check	OK	11:30

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 04:20

Gender: Male

Observation Period Began: 04:25

Subject's Last Name: PALAEZ

Date/Time of Test: 08/29/2008 04:58

---

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:53
Air Blank	0.000	04:53
Subject Sample #1	0.000	04:54
Breath Volume	2.039	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample #2	0.000	04:56
Breath Volume	2.414	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: ROBINSON, IAN, M

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ROBINSON, IAN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: COURT ORD

Arrest Time: 17:26

Gender: Female

Observation Period Began: 17:26

Subject's Last Name: DICKOW

Date/Time of Test: 08/29/2008 17:52

---

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.079	17:47
Air Blank	0.000	17:47
Subject Sample #1	0.000	17:48
Breath Volume	3.406	17:48
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample #2	0.000	17:50
Breath Volume	3.406	17:50
Air Blank	0.000	17:51
Control Test	0.080	17:51
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 02:35

Gender: Male

Observation Period Began: 02:50

Subject's Last Name: PERICLES

Date/Time of Test: 08/30/2008 03:18

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Subject Sample #1	0.114	03:13
Breath Volume	1.957	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample #2	0.117	03:16
Breath Volume	1.476	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:30

Gender: Female

Observation Period Began: 05:48

Subject's Last Name: HELMLY

Date/Time of Test: 08/30/2008 06:12

---

Diagnostics Check	OK	06:09	
Air Blank	0.000	06:09	
Control Test	0.079	06:10	
Air Blank	0.000	06:10	
Subject Sample #1		06:11	Subject Test Refused
Breath Volume	0.000	06:11	
Air Blank	0.000	06:11	
Control Test	0.079	06:11	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: LOPERFIDO, ANTHONY, N

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: NOT ARRESTED, NOT ARRESTED

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: PROB/PAROL

Arrest Time: 00:00

Gender: Female

Observation Period Began: 11:47

Subject's Last Name: GRICE

Date/Time of Test: 08/30/2008 12:13

---

Diagnostics Check	OK	12:07
Air Blank	0.000	12:08
Control Test	0.080	12:08
Air Blank	0.000	12:09
Subject Sample #1	0.000	12:09
Breath Volume	3.835	12:09
Air Blank	0.000	12:10
Air Blank	0.000	12:12
Subject Sample #2	0.000	12:12
Breath Volume	3.890	12:12
Air Blank	0.000	12:13
Control Test	0.079	12:13
Air Blank	0.000	12:13
Diagnostics Check	OK	12:13

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ROBINSON, I

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 18:50

Gender: Male

Observation Period Began: 19:05

Subject's Last Name: KYTE

Date/Time of Test: 08/30/2008 19:39

---

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.079	19:33
Air Blank	0.000	19:34
Subject Sample #1	0.280	19:34
Breath Volume	3.710	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample #2	0.260	19:37
Breath Volume	2.082	19:37
Air Blank	0.000	19:38
Control Test	0.078	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WIRTH, PHILIP

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: COURT ORD

Arrest Time: 20:50

Gender: Female

Observation Period Began: 20:50

Subject's Last Name: DICKOW

Date/Time of Test: 08/30/2008 21:16

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:11
Air Blank	0.000	21:11
Subject Sample #1	0.000	21:12
Breath Volume	3.285	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample #2	0.000	21:15
Breath Volume	2.984	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ROBINSON, I

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 22:56

Gender: Male

Observation Period Began: 23:12

Subject's Last Name: WRIGHT

Date/Time of Test: 08/30/2008 23:42

---

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:35	
Air Blank	0.000	23:35	
Subject Sample #1	0.256	23:38	Volume Not Met
Breath Volume	0.175	23:38	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample #2		23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Control Test	0.076	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: KUNCAS, ANDREW, J

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: KUNCAS, ANDREW

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 05:20

Gender: Female

Observation Period Began: 05:45

Subject's Last Name: CARDONA

Date/Time of Test: 08/31/2008 06:15

---

Diagnostics Check	OK	06:08
Air Blank	0.000	06:09
Control Test	0.080	06:09
Air Blank	0.000	06:09
Subject Sample #1	0.066	06:11
Breath Volume	1.523	06:11
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample #2	0.063	06:14
Breath Volume	1.390	06:14
Air Blank	0.000	06:14
Control Test	0.078	06:15
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WIRTH, PHILIP

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 12:25

Gender: Male

Observation Period Began: 12:36

Subject's Last Name: CASTILLO

Date/Time of Test: 08/31/2008 13:07

---

Diagnostics Check	OK	13:01
Air Blank	0.000	13:01
Control Test	0.080	13:02
Air Blank	0.000	13:02
Subject Sample #1	0.208	13:02
Breath Volume	1.195	13:02
Air Blank	0.000	13:03
Air Blank	0.000	13:05
Subject Sample #2	0.203	13:05
Breath Volume	1.390	13:05
Air Blank	0.000	13:06
Control Test	0.079	13:06
Air Blank	0.000	13:07
Diagnostics Check	OK	13:07

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: TEBOE, G

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 13:15

Gender: Male

Observation Period Began: 14:19

Subject's Last Name: GARAY-HERNANDEZ

Date/Time of Test: 08/31/2008 14:44

---

Diagnostics Check	OK	14:42	
Air Blank	0.000	14:43	
Control Test	0.074	14:43	Control Outside Tolerance
Air Blank	0.000	14:44	

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WIRTH, PHILIP, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: TEBOE, G

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 13:15

Gender: Male

Observation Period Began: 14:19

Subject's Last Name: GARAY-HERNANDEZ

Date/Time of Test: 08/31/2008 14:53

---

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.077	14:47
Air Blank	0.000	14:47
Subject Sample #1	0.106	14:47
Breath Volume	2.062	14:47
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample #2	0.110	14:51
Breath Volume	2.265	14:51
Air Blank	0.000	14:52
Control Test	0.077	14:52
Air Blank	0.000	14:53
Diagnostics Check	OK	14:53

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: LOPERFIDO, ANTHONY, N

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: NOT ARRESTED, NOT ARRESTED

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: PROB/PAROL

Arrest Time: 00:00

Gender: Female

Observation Period Began: 16:24

Subject's Last Name: DICKOW

Date/Time of Test: 08/31/2008 16:54

---

Diagnostics Check	OK	16:47
Air Blank	0.000	16:48
Control Test	0.079	16:48
Air Blank	0.000	16:49
Subject Sample #1	0.000	16:49
Breath Volume	3.101	16:49
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample #2	0.000	16:52
Breath Volume	2.656	16:52
Air Blank	0.000	16:53
Control Test	0.078	16:53
Air Blank	0.000	16:53
Diagnostics Check	OK	16:54

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: LOPERFIDO, ANTHONY, N

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: NOT ARRESTED, NOT ARRESTED

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: PROB/PAROL

Arrest Time: 17:16

Gender: Female

Observation Period Began: 17:16

Subject's Last Name: GRICE

Date/Time of Test: 08/31/2008 17:44

---

Diagnostics Check	OK	17:38
Air Blank	0.000	17:39
Control Test	0.079	17:39
Air Blank	0.000	17:40
Subject Sample #1	0.000	17:40
Breath Volume	3.730	17:40
Air Blank	0.000	17:41
Air Blank	0.000	17:43
Subject Sample #2	0.000	17:43
Breath Volume	3.562	17:43
Air Blank	0.000	17:44
Control Test	0.079	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

---

Data Download Date/Time: 09/24/2008 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-003224

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: SILVAGNI, R, W

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WIRTH, P

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 08/01/2008

Date or Time Changed: No

Dry Gas Standard Lot #: 7339011

Dry Gas Expiry Date: 12/07/2009

Violation Code: DUI

Arrest Time: 20:55

Gender: Male

Observation Period Began: 21:58

Subject's Last Name: ALARCON

Date/Time of Test: 08/31/2008 22:21

---

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.080	22:19	
Air Blank	0.000	22:19	
Subject Sample #1		22:20	Subject Test Refused
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

---

Data Download Date/Time: 09/24/2008 08:20