

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, DAVID

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 07:25

Gender: Male

Observation Period Began: 07:30

Subject's Last Name: KELLY

Date/Time of Test: 01/01/2007 08:07

Diagnostics Check	OK	08:00
Air Blank	0.000	08:01
Control Test	0.077	08:01
Air Blank	0.000	08:01
Subject Sample #1	0.102	08:02
Breath Volume	1.605	08:02
Air Blank	0.000	08:03
Air Blank	0.000	08:04
Subject Sample #2	0.102	08:05
Breath Volume	1.406	08:05
Air Blank	0.000	08:06
Control Test	0.077	08:06
Air Blank	0.000	08:07
Diagnostics Check	OK	08:07

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: HARRISON, JAMES, C

Breath Test Operator Agency: Brevard County SO

Arresting Officer: HARRISON, JAMES

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 02:01

Gender: Female

Observation Period Began: 02:01

Subject's Last Name: FORTHOFFER

Date/Time of Test: 01/05/2007 03:03

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:56
Subject Sample #1	0.094	02:58
Breath Volume	1.914	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample #2	0.098	03:01
Breath Volume	2.511	03:01
Air Blank	0.000	03:02
Control Test	0.076	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: BADMAN, LIN, M

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: PIDGEON, ANDREW

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 18:25

Gender: Female

Observation Period Began: 17:45

Subject's Last Name: KOLPACK

Date/Time of Test: 01/06/2007 19:01

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.078	18:55
Air Blank	0.000	18:56
Subject Sample #1	0.221	18:56
Breath Volume	2.062	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample #2	0.219	18:59
Breath Volume	1.996	18:59
Air Blank	0.000	19:00
Control Test	0.076	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: LINDSAY, JUDITH

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 02:00

Gender: Female

Observation Period Began: 02:10

Subject's Last Name: GOODMAN

Date/Time of Test: 01/07/2007 02:52

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.078	02:46	
Air Blank	0.000	02:47	
Subject Sample #1	0.154	02:47	
Breath Volume	1.394	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample #2		02:50	RFI Detect
Breath Volume	1.234	02:50	
Air Blank	0.000	02:51	
Control Test	0.076	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: LINDSAY, JUDITH

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 02:00

Gender: Female

Observation Period Began: 02:10

Subject's Last Name: GOODMAN

Date/Time of Test: 01/07/2007 03:01

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.076	02:55
Air Blank	0.000	02:55
Subject Sample #1	0.146	02:56
Breath Volume	1.257	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.138	03:00
Breath Volume	1.273	03:00
Air Blank	0.000	03:00
Control Test	0.075	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: RAMOS, MARLENA, M

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: RICHARD, MARK

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 16:30

Gender: Male

Observation Period Began: 16:46

Subject's Last Name: BULERIN

Date/Time of Test: 01/09/2007 17:44

Diagnostics Check	OK	17:38
Air Blank	0.000	17:38
Control Test	0.079	17:39
Air Blank	0.000	17:39
Subject Sample #1	0.000	17:39
Breath Volume	1.746	17:39
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample #2	0.000	17:42
Breath Volume	1.550	17:42
Air Blank	0.000	17:43
Control Test	0.079	17:43
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: RIBNICKY, THOMAS, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: RIBNICKY, THOMAS

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 18:07

Gender: Female

Observation Period Began: 18:30

Subject's Last Name: CONCEPCION

Date/Time of Test: 01/09/2007 18:54

Diagnostics Check	OK	18:50	
Air Blank	0.000	18:51	
Control Test	0.079	18:51	
Air Blank	0.000	18:52	
Subject Sample #1		18:52	Slope Not Met
Breath Volume	1.265	18:52	
Air Blank	0.000	18:53	
Control Test	0.079	18:53	
Air Blank	0.000	18:53	
Diagnostics Check	OK	18:54	

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: RIBNICKY, THOMAS, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: RIBNICKY, THOMAS

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 18:07

Gender: Female

Observation Period Began: 18:30

Subject's Last Name: CONCEPCION

Date/Time of Test: 01/09/2007 19:04

Diagnostics Check	OK	18:57
Air Blank	0.000	18:58
Control Test	0.079	18:58
Air Blank	0.000	18:58
Subject Sample #1	0.241	18:59
Breath Volume	1.390	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample #2	0.242	19:02
Breath Volume	1.898	19:02
Air Blank	0.000	19:03
Control Test	0.076	19:03
Air Blank	0.000	19:03
Diagnostics Check	OK	19:04

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, DARIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 03:20

Gender: Female

Observation Period Began: 03:10

Subject's Last Name: CHENEY

Date/Time of Test: 01/10/2007 03:46

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample #1	0.216	03:42
Breath Volume	2.082	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample #2	0.200	03:45
Breath Volume	2.285	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: FREY, ROBERT, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: BROWN, CLARK

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 23:57

Gender: Female

Observation Period Began: 23:57

Subject's Last Name: CHANCEY

Date/Time of Test: 01/11/2007 01:18

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Subject Sample #1	0.028	01:09	Volume Not Met
Breath Volume	0.171	01:09	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample #2	0.027	01:13	
Breath Volume	1.859	01:13	
Air Blank	0.000	01:14	
Air Blank	0.000	01:16	
Subject Sample #3	0.026	01:16	
Breath Volume	2.128	01:16	
Air Blank	0.000	01:17	
Control Test	0.078	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MAGEE, TONIE

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 00:20

Gender: Female

Observation Period Began: 01:00

Subject's Last Name: BISHOP

Date/Time of Test: 01/11/2007 01:34

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Subject Sample #1	0.143	01:29
Breath Volume	2.253	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample #2	0.139	01:32
Breath Volume	3.441	01:32
Air Blank	0.000	01:33
Control Test	0.078	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: VALENTI, NEAL, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: RICHARD, MARK

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 16:10

Gender: Male

Observation Period Began: 16:30

Subject's Last Name: MILLS

Date/Time of Test: 01/15/2007 17:17

Diagnostics Check	OK	17:11
Air Blank	0.000	17:11
Control Test	0.078	17:11
Air Blank	0.000	17:12
Subject Sample #1	0.030	17:13
Breath Volume	4.003	17:13
Air Blank	0.000	17:13
Air Blank	0.000	17:15
Subject Sample #2	0.030	17:16
Breath Volume	3.347	17:16
Air Blank	0.000	17:16
Control Test	0.077	17:17
Air Blank	0.000	17:17
Diagnostics Check	OK	17:17

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, DARIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 01:42

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: FEE

Date/Time of Test: 01/18/2007 02:33

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.114	02:29
Breath Volume	2.089	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample #2	0.113	02:31
Breath Volume	3.605	02:31
Air Blank	0.000	02:32
Control Test	0.077	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: CAMPBELL, JASMINE,

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: PIDGEON, ANDREW

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 19:33

Gender: Male

Observation Period Began: 19:15

Subject's Last Name: SHANNON

Date/Time of Test: 01/20/2007 20:22

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.078	20:16
Air Blank	0.000	20:17
Subject Sample #1	0.226	20:17
Breath Volume	1.902	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample #2	0.227	20:20
Breath Volume	3.050	20:20
Air Blank	0.000	20:21
Control Test	0.077	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: HARRISON, JAMES, C

Breath Test Operator Agency: Brevard County SO

Arresting Officer: HARRISON, JAMES

Arresting Officer Agency: Brevard County SO

Last Agency Inspection: 12/22/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 01:54

Gender: Male

Observation Period Began: 01:54

Subject's Last Name: JACKSON

Date/Time of Test: 01/21/2007 03:45

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.077	03:40
Air Blank	0.000	03:41
Subject Sample #1	0.057	03:41
Breath Volume	2.156	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample #2	0.055	03:44
Breath Volume	2.433	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 01/25/2007 23:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MARCINIK, DAVID

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 01/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 22:45

Gender: Male

Observation Period Began: 22:35

Subject's Last Name: WILBER

Date/Time of Test: 01/26/2007 23:10

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:05
Subject Sample #1	0.252	23:05
Breath Volume	2.207	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample #2	0.234	23:08
Breath Volume	2.328	23:08
Air Blank	0.000	23:09
Control Test	0.077	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 02/22/2007 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: STROUP, CHRISTOPHER

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 01/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 23:35

Gender: Male

Observation Period Began: 23:30

Subject's Last Name: BALRAM

Date/Time of Test: 01/27/2007 00:22

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:17
Air Blank	0.000	00:17
Subject Sample #1	0.202	00:18
Breath Volume	1.570	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample #2	0.197	00:21
Breath Volume	1.957	00:21
Air Blank	0.000	00:21
Control Test	0.078	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 02/22/2007 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MARCINIK, DAVID, E

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: STROUP, CHRISTOPHER

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 01/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 01:25

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: MCNEAL

Date/Time of Test: 01/28/2007 01:50

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Subject Sample #1	0.178	01:45
Breath Volume	1.320	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample #2	0.176	01:49
Breath Volume	1.578	01:49
Air Blank	0.000	01:49
Control Test	0.076	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 02/22/2007 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001721

Instrument Registered To: PALM BAY PD

Software Version: 8100.27

Breath Test Operator: MORGAN, DARIN, P

Breath Test Operator Agency: Palm Bay PD

Arresting Officer: MORGAN, DARIN

Arresting Officer Agency: Palm Bay PD

Last Agency Inspection: 01/25/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 502701G

Dry Gas Expiry Date: 02/02/2007

Violation Code: DUI

Arrest Time: 02:25

Gender: Male

Observation Period Began: 02:20

Subject's Last Name: HERR

Date/Time of Test: 01/30/2007 02:55

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Subject Sample #1	0.241	02:51
Breath Volume	2.468	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample #2	0.228	02:54
Breath Volume	2.281	02:54
Air Blank	0.000	02:54
Control Test	0.077	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 02/22/2007 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: GOULD, LEONARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 504502I

Dry Gas Expiry Date: 02/17/2007

Violation Code: DUI

Arrest Time: 23:26

Gender: Male

Observation Period Began: 23:30

Subject's Last Name: CINTORA

Date/Time of Test: 01/19/2007 23:57

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.075	23:51
Air Blank	0.000	23:52
Subject Sample #1	0.169	23:52
Breath Volume	2.355	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample #2	0.163	23:55
Breath Volume	2.968	23:55
Air Blank	0.000	23:56
Control Test	0.076	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 01/26/2007 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: TUCKER, THOMAS

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 504502I

Dry Gas Expiry Date: 02/17/2007

Violation Code: DUI

Arrest Time: 00:05

Gender: Male

Observation Period Began: 00:07

Subject's Last Name: MACKENZIE

Date/Time of Test: 01/20/2007 00:40

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.075	00:34
Air Blank	0.000	00:35
Subject Sample #1	0.146	00:35
Breath Volume	3.542	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample #2	0.142	00:38
Breath Volume	3.902	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 01/26/2007 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: PRZESPOLEWSKI, BRIAN

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 504502I

Dry Gas Expiry Date: 02/17/2007

Violation Code: DUI

Arrest Time: 00:20

Gender: Male

Observation Period Began: 00:20

Subject's Last Name: MATEO

Date/Time of Test: 01/20/2007 00:51

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.076	00:45
Air Blank	0.000	00:46
Subject Sample #1	0.027	00:46
Breath Volume	2.925	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample #2	0.028	00:49
Breath Volume	3.500	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 01/26/2007 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: STESKAL, CHRIS

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 504502I

Dry Gas Expiry Date: 02/17/2007

Violation Code: DUI

Arrest Time: 01:00

Gender: Male

Observation Period Began: 01:00

Subject's Last Name: GINSBERG

Date/Time of Test: 01/20/2007 01:26

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.077	01:21
Air Blank	0.000	01:21
Subject Sample #1	0.137	01:22
Breath Volume	1.730	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample #2	0.125	01:25
Breath Volume	1.136	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 01/26/2007 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001722

Instrument Registered To: LEE COUNTY SO

Software Version: 8100.27

Breath Test Operator: CLAFLIN, DAVID, S

Breath Test Operator Agency: Lee County SO

Arresting Officer: GOULD, LEONARD

Arresting Officer Agency: Lee County SO

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 504502I

Dry Gas Expiry Date: 02/17/2007

Violation Code: DUI

Arrest Time: 01:37

Gender: Male

Observation Period Began: 01:39

Subject's Last Name: KOKOSHI

Date/Time of Test: 01/20/2007 02:06

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Subject Sample #1	0.080	02:01
Breath Volume	1.566	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample #2	0.083	02:04
Breath Volume	1.585	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 01/26/2007 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: SNYDER, MELONIE, E
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: WILLIAMS, G
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 12/21/2006

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 00:19
Gender: Female
Observation Period Began: 00:41

Subject's Last Name: KELLEMS

Date/Time of Test: 01/06/2007 01:13

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.077	01:03	
Air Blank	0.000	01:04	
Subject Sample #1	0.160	01:07	Volume Not Met
Breath Volume	0.292	01:07	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample #2		01:11	Subject Test Refused
Breath Volume	0.214	01:11	
Air Blank	0.000	01:12	
Control Test	0.075	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 01/18/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: SNYDER, MELONIE, E

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: WILLIAMS, G

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 12/21/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 00:19

Gender: Female

Observation Period Began: 00:41

Subject's Last Name: KELLEMS

Date/Time of Test: 01/06/2007 01:20

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.074	01:19	Control Outside Tolerance
Air Blank	0.000	01:20	

Data Download Date/Time: 01/18/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: SNYDER, MELONIE, E
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: WILLIAMS, G
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 12/21/2006

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 00:19
Gender: Female
Observation Period Began: 00:41

Subject's Last Name: KELLEMS

Date/Time of Test: 01/06/2007 01:29

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.075	01:22
Air Blank	0.000	01:23
Subject Sample #1	0.192	01:24
Breath Volume	1.558	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample #2	0.175	01:27
Breath Volume	1.429	01:27
Air Blank	0.000	01:28
Control Test	0.075	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 01/18/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: MORRELL, L
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 12/21/2006

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 23:11
Gender: Male
Observation Period Began: 23:21

Subject's Last Name: DUARTE-GALINDEZ

Date/Time of Test: 01/12/2007 23:49

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.078	23:44
Air Blank	0.000	23:44
Subject Sample #1	0.137	23:45
Breath Volume	1.785	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample #2	0.141	23:48
Breath Volume	2.617	23:48
Air Blank	0.000	23:48
Control Test	0.078	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 01/18/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO
Software Version: 8100.27
Breath Test Operator: COLDING, LINDA, D
Breath Test Operator Agency: Hillsborough Co SO
Arresting Officer: MORRILL, S
Arresting Officer Agency: Hillsborough Co SO
Last Agency Inspection: 12/21/2006

Date or Time Changed: No
Dry Gas Standard Lot #: 531505G
Dry Gas Expiry Date: 11/15/2007
Violation Code: DUI
Arrest Time: 01:50
Gender: Male
Observation Period Began: 02:17

Subject's Last Name: GARCIA

Date/Time of Test: 01/13/2007 02:48

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.077	02:41	
Air Blank	0.000	02:41	
Subject Sample #1	0.181	02:44	
Breath Volume	2.011	02:44	
Air Blank	0.000	02:44	
Air Blank	0.000	02:46	
Subject Sample #2	0.181	02:47	
Breath Volume	1.667	02:47	
Air Blank	0.000	02:47	
Control Test	0.076	02:47	RFI Detect
Air Blank		02:48	RFI Detect

Data Download Date/Time: 01/18/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: COLDING, LINDA, D

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: MORRILL, S

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 12/21/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 01:50

Gender: Male

Observation Period Began: 02:17

Subject's Last Name: GARCIA

Date/Time of Test: 01/13/2007 02:59

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.076	02:51
Air Blank	0.000	02:51
Subject Sample #1	0.168	02:54
Breath Volume	1.570	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.179	02:57
Breath Volume	2.039	02:57
Air Blank	0.000	02:58
Control Test	0.076	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 01/18/2007 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Tampa PD

Arresting Officer: ARENA, F

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 01:26

Gender: Male

Observation Period Began: 01:44

Subject's Last Name: APOSTOLSKIS

Date/Time of Test: 01/20/2007 02:11

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Subject Sample #1	0.080	02:07
Breath Volume	1.648	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.087	02:10
Breath Volume	2.968	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Tampa PD

Arresting Officer: BUCHER, P

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 02:56

Gender: Male

Observation Period Began: 03:09

Subject's Last Name: VELIZCALABUCHE

Date/Time of Test: 01/20/2007 03:44

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.079	03:33	
Air Blank	0.000	03:34	
Subject Sample #1	0.145	03:37	Volume Not Met
Breath Volume	0.757	03:37	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample #2	0.174	03:42	Volume Not Met
Breath Volume	0.843	03:42	
Air Blank	0.000	03:43	
Control Test	0.077	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Tampa PD

Arresting Officer: ARENA, F

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 03:34

Gender: Female

Observation Period Began: 03:53

Subject's Last Name: AYOUB

Date/Time of Test: 01/20/2007 04:22

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.076	04:16
Air Blank	0.000	04:17
Subject Sample #1	0.081	04:17
Breath Volume	2.410	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.078	04:21
Breath Volume	2.671	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: HYDER, D

Arresting Officer Agency: Hillsborough Co SO

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 22:51

Gender: Male

Observation Period Began: 23:38

Subject's Last Name: GOMEZ CABALLERO

Date/Time of Test: 01/21/2007 00:13

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.078	00:07
Air Blank	0.000	00:07
Subject Sample #1	0.163	00:08
Breath Volume	1.648	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample #2	0.164	00:12
Breath Volume	1.324	00:12
Air Blank	0.000	00:12
Control Test	0.077	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: GROVES, D

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 20:36

Gender: Male

Observation Period Began: 20:52

Subject's Last Name: ROBERTS

Date/Time of Test: 01/27/2007 21:23

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.075	21:18
Air Blank	0.000	21:18
Subject Sample #1	0.144	21:19
Breath Volume	3.371	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample #2	0.148	21:22
Breath Volume	3.292	21:22
Air Blank	0.000	21:22
Control Test	0.075	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: BUCHER, P

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 21:15

Gender: Male

Observation Period Began: 21:25

Subject's Last Name: ZARRALUGUI

Date/Time of Test: 01/27/2007 21:58

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.075	21:47	
Air Blank	0.000	21:47	
Subject Sample #1	0.177	21:50	Volume Not Met
Breath Volume	0.847	21:50	
Air Blank	0.000	21:51	
Air Blank	0.000	21:53	
Subject Sample #2	0.166	21:56	Volume Not Met
Breath Volume	0.406	21:56	
Air Blank	0.000	21:57	
Control Test	0.076	21:57	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: WILLIAMS, G

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 22:37

Gender: Male

Observation Period Began: 22:56

Subject's Last Name: WHEELLOCK

Date/Time of Test: 01/27/2007 23:27

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:23
Subject Sample #1	0.072	23:23
Breath Volume	2.050	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample #2	0.069	23:26
Breath Volume	1.886	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: WILSON, S

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 22:48

Gender: Female

Observation Period Began: 23:14

Subject's Last Name: PERKEY

Date/Time of Test: 01/27/2007 23:45

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.077	23:39
Air Blank	0.000	23:40
Subject Sample #1	0.143	23:40
Breath Volume	3.082	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample #2	0.141	23:43
Breath Volume	2.527	23:43
Air Blank	0.000	23:44
Control Test	0.077	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: GROVES, D

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 22:59

Gender: Male

Observation Period Began: 23:14

Subject's Last Name: MCADOO

Date/Time of Test: 01/27/2007 23:57

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.076	23:51
Air Blank	0.000	23:52
Subject Sample #1	0.187	23:52
Breath Volume	4.308	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample #2	0.190	23:55
Breath Volume	4.167	23:55
Air Blank	0.000	23:56
Control Test	0.077	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: WILSON, S

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 01:10

Gender: Female

Observation Period Began: 01:37

Subject's Last Name: ARMETTA

Date/Time of Test: 01/28/2007 02:08

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Subject Sample #1	0.117	02:04
Breath Volume	2.218	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #2	0.121	02:07
Breath Volume	2.421	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001731

Instrument Registered To: HILLSBOROUGH SO

Software Version: 8100.27

Breath Test Operator: GRAY, BEVERLY, A

Breath Test Operator Agency: Hillsborough Co SO

Arresting Officer: VALLEJO, J

Arresting Officer Agency: Tampa PD

Last Agency Inspection: 01/18/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 531505G

Dry Gas Expiry Date: 11/15/2007

Violation Code: DUI

Arrest Time: 02:00

Gender: Female

Observation Period Began: 02:31

Subject's Last Name: STACK

Date/Time of Test: 01/28/2007 03:00

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Subject Sample #1	0.054	02:56
Breath Volume	2.015	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.051	02:59
Breath Volume	1.511	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 02/21/2007 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001740

Instrument Registered To: CLEARWATER PD

Software Version: 8100.27

Breath Test Operator: YUEN, JONATHON, K

Breath Test Operator Agency: Clearwater PD

Arresting Officer: YUEN, JONATHON

Arresting Officer Agency: Clearwater PD

Last Agency Inspection: 12/15/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 6153011

Dry Gas Expiry Date: 06/05/2008

Violation Code: DUI

Arrest Time: 03:55

Gender: Female

Observation Period Began: 04:25

Subject's Last Name: CRANE

Date/Time of Test: 01/05/2007 04:53

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:47
Subject Sample #1	0.167	04:48
Breath Volume	2.531	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample #2	0.153	04:51
Breath Volume	1.433	04:51
Air Blank	0.000	04:52
Control Test	0.078	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 01/26/2007 07:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: DEVAUL, ERIC

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/08/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 610001G

Dry Gas Expiry Date: 04/12/2008

Violation Code: DUI

Arrest Time: 22:06

Gender: Male

Observation Period Began: 22:30

Subject's Last Name: POPISH

Date/Time of Test: 01/04/2007 23:18

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:12
Air Blank	0.000	23:12
Subject Sample #1	0.147	23:13
Breath Volume	1.769	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample #2	0.141	23:16
Breath Volume	1.308	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 01/12/2007 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: DEVAUL, ERIC

Arresting Officer Agency: FHP Troop L

Last Agency Inspection: 12/08/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 610001G

Dry Gas Expiry Date: 04/12/2008

Violation Code: DUI

Arrest Time: 21:11

Gender: Male

Observation Period Began: 21:53

Subject's Last Name: RAMIREZ

Date/Time of Test: 01/05/2007 22:27

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.077	22:21
Air Blank	0.000	22:22
Subject Sample #1	0.250	22:23
Breath Volume	1.863	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample #2	0.270	22:26
Breath Volume	1.867	22:26
Air Blank	0.000	22:27
Control Test	0.077	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 01/12/2007 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FLOWERS, GARY

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 12/08/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 610001G

Dry Gas Expiry Date: 04/12/2008

Violation Code: DUI

Arrest Time: 01:03

Gender: Female

Observation Period Began: 01:18

Subject's Last Name: SUMMERHILL

Date/Time of Test: 01/07/2007 02:11

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.077	02:04
Air Blank	0.000	02:05
Subject Sample #1	0.183	02:06
Breath Volume	2.355	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample #2	0.198	02:09
Breath Volume	3.273	02:09
Air Blank	0.000	02:10
Control Test	0.077	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/12/2007 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FLOWERS, GARY

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 01/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610001G

Dry Gas Expiry Date: 04/12/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: SANCHEZ

Date/Time of Test: 01/14/2007 02:25

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.078	02:11	
Air Blank	0.000	02:11	
Subject Sample #1	0.242	02:14	
Breath Volume	1.218	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:16	
Subject Sample #2	0.230	02:19	Volume Not Met
Breath Volume	0.019	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample #3	0.234	02:23	
Breath Volume	1.281	02:23	
Air Blank	0.000	02:24	
Control Test	0.074	02:25	Control Outside Tolerance
Air Blank	0.000	02:25	

Data Download Date/Time: 02/23/2007 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: FLOWERS, GARY

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 01/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 610001G

Dry Gas Expiry Date: 04/12/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 01:42

Subject's Last Name: SANCHEZ

Date/Time of Test: 01/14/2007 02:43

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.075	02:31	
Air Blank	0.000	02:32	
Subject Sample #1	0.226	02:33	
Breath Volume	1.132	02:33	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample #2	0.232	02:38	Volume Not Met
Breath Volume	0.542	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample #3	0.226	02:42	
Breath Volume	1.609	02:42	
Air Blank	0.000	02:42	
Control Test	0.074	02:43	Control Outside Tolerance
Air Blank	0.000	02:43	

Data Download Date/Time: 02/23/2007 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: SCHAFFER, NANCY, M

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: PARENT, G

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 01/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: OTHER

Arrest Time: 12:30

Gender: Male

Observation Period Began: 13:00

Subject's Last Name: DOE

Date/Time of Test: 01/18/2007 13:40

Diagnostics Check	OK	13:35
Air Blank	0.000	13:35
Control Test	0.078	13:35
Air Blank	0.000	13:36
Subject Sample #1	0.000	13:36
Breath Volume	2.015	13:36
Air Blank	0.000	13:37
Air Blank	0.000	13:39
Subject Sample #2	0.000	13:39
Breath Volume	2.531	13:39
Air Blank	0.000	13:39
Control Test	0.078	13:40
Air Blank	0.000	13:40
Diagnostics Check	OK	13:40

Data Download Date/Time: 02/23/2007 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: NEUMAN, CHRIS

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 01/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 23:25

Gender: Male

Observation Period Began: 23:30

Subject's Last Name: BEAMAN

Date/Time of Test: 01/20/2007 00:06

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.078	00:00
Air Blank	0.000	00:01
Subject Sample #1	0.077	00:01
Breath Volume	4.527	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample #2	0.077	00:04
Breath Volume	4.406	00:04
Air Blank	0.000	00:05
Control Test	0.077	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 02/23/2007 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001741

Instrument Registered To: PALM BEACH CO SO

Software Version: 8100.27

Breath Test Operator: PARENT, GARY, J

Breath Test Operator Agency: Palm Beach County SO

Arresting Officer: NOEL, KEN

Arresting Officer Agency: Palm Beach County SO

Last Agency Inspection: 01/12/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 632102I

Dry Gas Expiry Date: 11/20/2008

Violation Code: DUI

Arrest Time: 23:15

Gender: Male

Observation Period Began: 23:30

Subject's Last Name: SCHLETER

Date/Time of Test: 01/20/2007 01:02

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.077	00:56
Air Blank	0.000	00:56
Subject Sample #1	0.215	00:57
Breath Volume	1.335	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample #2	0.222	01:00
Breath Volume	2.320	01:00
Air Blank	0.000	01:01
Control Test	0.077	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 02/23/2007 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001866

Instrument Registered To: US AIR FORCE MACDILL

Software Version: 8100.27

Breath Test Operator: LUCERO, IVAN,

Breath Test Operator Agency: Macdill Air Force Ba

Arresting Officer: COGSWELL, ROBERT

Arresting Officer Agency: Macdill Air Force Ba

Last Agency Inspection: 12/28/2006

Date or Time Changed: Yes

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 20:00

Gender: Male

Observation Period Began: 22:20

Subject's Last Name: GREEN

Date/Time of Test: 01/13/2007 22:52

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.078	22:47
Air Blank	0.000	22:47
Subject Sample #1	0.051	22:48
Breath Volume	3.972	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample #2	0.049	22:51
Breath Volume	4.320	22:51
Air Blank	0.000	22:51
Control Test	0.078	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 02/12/2007 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001866

Instrument Registered To: US AIR FORCE MACDILL

Software Version: 8100.27

Breath Test Operator: SMITH, ELMER, H

Breath Test Operator Agency: Macdill Air Force Ba

Arresting Officer: COGSWELL, ROBERT

Arresting Officer Agency: Macdill Air Force Ba

Last Agency Inspection: 12/28/2006

Date or Time Changed: Yes

Dry Gas Standard Lot #: 535402I

Dry Gas Expiry Date: 12/22/2007

Violation Code: DUI

Arrest Time: 20:20

Gender: Male

Observation Period Began: 20:28

Subject's Last Name: STOTHERS

Date/Time of Test: 01/17/2007 21:13

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.078	21:08
Air Blank	0.000	21:08
Subject Sample #1	0.177	21:09
Breath Volume	1.785	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample #2	0.175	21:12
Breath Volume	2.734	21:12
Air Blank	0.000	21:13
Control Test	0.078	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 02/12/2007 16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BUTTERWORTH, DAVID, S

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: BUTTERWORTH, DAVID

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 01/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 22:57

Gender: Female

Observation Period Began: 22:57

Subject's Last Name: GALLAGHER

Date/Time of Test: 01/05/2007 23:54

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.077	23:48
Air Blank	0.000	23:49
Subject Sample #1	0.164	23:49
Breath Volume	2.460	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample #2	0.164	23:52
Breath Volume	2.339	23:52
Air Blank	0.000	23:53
Control Test	0.078	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 02/28/2007 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BUTTERWORTH, DAVID, S

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: FAZIO, JOHN

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 01/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 01:28

Gender: Female

Observation Period Began: 01:43

Subject's Last Name: BUCHHEIT

Date/Time of Test: 01/06/2007 02:12

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.077	02:06
Air Blank	0.000	02:07
Subject Sample #1	0.224	02:07
Breath Volume	1.687	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample #2	0.216	02:11
Breath Volume	1.695	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 02/28/2007 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD
Software Version: 8100.27
Breath Test Operator: BUTTERWORTH, DAVID, S
Breath Test Operator Agency: Port St. Lucie PD
Arresting Officer: BUTTERWORTH, DAVID
Arresting Officer Agency: Port St. Lucie PD
Last Agency Inspection: 01/03/2007

Date or Time Changed: No
Dry Gas Standard Lot #: 607401G
Dry Gas Expiry Date: 03/23/2008
Violation Code: DUI
Arrest Time: 02:07
Gender: Male
Observation Period Began: 02:30

Subject's Last Name: GRIFFIN

Date/Time of Test: 01/11/2007 03:16

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample #1	0.186	03:12
Breath Volume	2.761	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample #2	0.183	03:15
Breath Volume	2.957	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 02/28/2007 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BUTTERWORTH, DAVID, S

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: BUTTERWORTH, DAVID

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 01/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 23:57

Gender: Male

Observation Period Began: 00:00

Subject's Last Name: WHIPPLE

Date/Time of Test: 01/14/2007 00:48

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Subject Sample #1	0.179	00:43
Breath Volume	1.781	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample #2	0.188	00:46
Breath Volume	2.097	00:46
Air Blank	0.000	00:47
Control Test	0.078	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 02/28/2007 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: BUTTERWORTH, DAVID, S

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: NORMAN, DOUG

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 01/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 03:32

Gender: Female

Observation Period Began: 04:05

Subject's Last Name: PEJACK

Date/Time of Test: 01/16/2007 04:45

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:37
Subject Sample #1	0.232	04:37
Breath Volume	1.292	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample #2	0.208	04:40
Breath Volume	1.304	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample #3	0.215	04:44
Breath Volume	2.097	04:44
Air Blank	0.000	04:44
Control Test	0.077	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 02/28/2007 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-001962

Instrument Registered To: PORT ST LUCIE PD

Software Version: 8100.27

Breath Test Operator: HARRIS, OFC, I

Breath Test Operator Agency: Port St. Lucie PD

Arresting Officer: PIETRZAK, OFC

Arresting Officer Agency: Port St. Lucie PD

Last Agency Inspection: 01/03/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 607401G

Dry Gas Expiry Date: 03/23/2008

Violation Code: DUI

Arrest Time: 02:19

Gender: Male

Observation Period Began: 02:41

Subject's Last Name: JOHNSON

Date/Time of Test: 01/27/2007 03:11

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.078	03:06
Air Blank	0.000	03:06
Subject Sample #1	0.080	03:07
Breath Volume	1.738	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample #2	0.081	03:10
Breath Volume	1.750	03:10
Air Blank	0.000	03:10
Control Test	0.077	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 02/28/2007 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002097

Instrument Registered To: NAS JACKSONVILLE

Software Version: 8100.27

Breath Test Operator: KEITH, JOHN, M

Breath Test Operator Agency: NAS Jacksonville PD

Arresting Officer: KEITH, JOHN

Arresting Officer Agency: NAS Jacksonville PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 612501G

Dry Gas Expiry Date: 05/08/2008

Violation Code: OTHER

Arrest Time: 16:40

Gender: Male

Observation Period Began: 16:40

Subject's Last Name: DRURY

Date/Time of Test: 01/30/2007 17:26

Diagnostics Check	OK	17:20
Air Blank	0.000	17:20
Control Test	0.081	17:21
Air Blank	0.000	17:21
Subject Sample #1	0.029	17:22
Breath Volume	6.011	17:22
Air Blank	0.000	17:22
Air Blank	0.000	17:24
Subject Sample #2	0.028	17:24
Breath Volume	6.281	17:24
Air Blank	0.000	17:25
Control Test	0.081	17:25
Air Blank	0.000	17:26
Diagnostics Check	OK	17:26

Data Download Date/Time: 02/16/2007 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 23:06

Gender: Male

Observation Period Began: 23:10

Subject's Last Name: DAUGHTRY

Date/Time of Test: 01/05/2007 00:23

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.079	00:14	
Air Blank	0.000	00:15	
Subject Sample #1	0.250	00:15	
Breath Volume	4.503	00:15	
Air Blank	0.000	00:16	
Air Blank	0.000	00:18	
Subject Sample #2	0.215	00:18	
Breath Volume	2.519	00:18	
Air Blank	0.000	00:19	
Air Blank	0.000	00:21	
Subject Sample #3	0.228	00:21	
Breath Volume	4.828	00:21	
Air Blank	0.000	00:22	
Control Test	0.069	00:22	Control Outside Tolerance
Air Blank	0.000	00:23	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 23:06

Gender: Male

Observation Period Began: 23:10

Subject's Last Name: DAUGHTRY

Date/Time of Test: 01/05/2007 00:32

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Subject Sample #1	0.239	00:27
Breath Volume	4.445	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample #2	0.234	00:30
Breath Volume	4.292	00:30
Air Blank	0.000	00:31
Control Test	0.077	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: COSNER, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:58

Gender: Male

Observation Period Began: 04:08

Subject's Last Name: NARANJO

Date/Time of Test: 01/05/2007 04:35

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.079	04:32	
Air Blank	0.000	04:33	
Subject Sample #1		04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:34	
Control Test	0.079	04:34	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVEN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:20

Gender: Male

Observation Period Began: 04:21

Subject's Last Name: RAMIREZ-ALVAREZ

Date/Time of Test: 01/05/2007 05:09

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:03
Subject Sample #1	0.163	05:04
Breath Volume	1.667	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample #2	0.165	05:07
Breath Volume	1.500	05:07
Air Blank	0.000	05:08
Control Test	0.077	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: COSNER, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:28

Gender: Male

Observation Period Began: 05:35

Subject's Last Name: ROBERTS

Date/Time of Test: 01/05/2007 06:08

Diagnostics Check	OK	06:03	
Air Blank	0.000	06:04	
Control Test	0.079	06:04	
Air Blank	0.000	06:05	
Subject Sample #1		06:06	Subject Test Refused
Breath Volume	0.000	06:06	
Air Blank	0.000	06:07	
Control Test	0.079	06:07	
Air Blank	0.000	06:07	
Diagnostics Check	OK	06:08	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: WARNICK

Date/Time of Test: 01/06/2007 04:47

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.079	04:41	
Air Blank	0.000	04:42	
Subject Sample #1	0.152	04:43	
Breath Volume	1.328	04:43	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample #2	0.158	04:46	
Breath Volume	1.242	04:46	
Air Blank	0.000	04:46	
Control Test	0.072	04:47	Control Outside Tolerance
Air Blank	0.000	04:47	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: WARNICK

Date/Time of Test: 01/06/2007 04:56

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.080	04:50	
Air Blank	0.000	04:50	
Subject Sample #1	0.169	04:51	
Breath Volume	1.773	04:51	
Air Blank	0.000	04:52	
Air Blank	0.000	04:53	
Subject Sample #2	0.167	04:54	
Breath Volume	4.527	04:54	
Air Blank	0.000	04:55	
Control Test	0.073	04:55	Control Outside Tolerance
Air Blank	0.000	04:56	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: WARNICK

Date/Time of Test: 01/06/2007 05:03

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:58	
Control Test	0.079	04:58	
Air Blank	0.000	04:59	
Subject Sample #1	0.153	04:59	
Breath Volume	1.351	04:59	
Air Blank	0.000	05:00	
Air Blank	0.000	05:01	
Subject Sample #2	0.157	05:02	
Breath Volume	1.125	05:02	
Air Blank	0.000	05:03	
Control Test	0.074	05:03	Control Outside Tolerance
Air Blank	0.000	05:03	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: WARNICK

Date/Time of Test: 01/06/2007 05:16

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.076	05:07	
Air Blank	0.000	05:08	
Subject Sample #1	0.160	05:08	
Breath Volume	3.113	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample #2	0.138	05:11	
Breath Volume	1.949	05:11	
Air Blank	0.000	05:12	
Air Blank	0.000	05:14	
Subject Sample #3	0.146	05:14	
Breath Volume	4.066	05:14	
Air Blank	0.000	05:15	
Control Test	0.060	05:15	Control Outside Tolerance
Air Blank	0.000	05:16	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: WARNICK

Date/Time of Test: 01/06/2007 05:26

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.078	05:21
Air Blank	0.000	05:21
Subject Sample #1	0.152	05:22
Breath Volume	2.195	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample #2	0.153	05:25
Breath Volume	2.097	05:25
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:50

Gender: Male

Observation Period Began: 06:00

Subject's Last Name: SEPIASHVILI

Date/Time of Test: 01/06/2007 06:27

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Subject Sample #1	0.073	06:23
Breath Volume	1.289	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample #2	0.085	06:25
Breath Volume	1.507	06:25
Air Blank	0.000	06:26
Control Test	0.078	06:26
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 22:39

Gender: Male

Observation Period Began: 23:00

Subject's Last Name: GALDO

Date/Time of Test: 01/06/2007 23:28

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Subject Sample #1	0.072	23:23
Breath Volume	4.828	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample #2	0.060	23:26
Breath Volume	1.937	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: RIVERO, MARLEN, M

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MCVEY, GREG

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: Yes

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 01:55

Gender: Male

Observation Period Began: 02:05

Subject's Last Name: PEREZ

Date/Time of Test: 01/07/2007 02:42

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample #1	0.146	02:37
Breath Volume	2.089	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.146	02:41
Breath Volume	2.195	02:41
Air Blank	0.000	02:41
Control Test	0.078	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GUERRERO, SAMIR,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: GUERRERO, SAMIR

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 02:25

Gender: Male

Observation Period Began: 02:44

Subject's Last Name: PEREZ

Date/Time of Test: 01/07/2007 03:10

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Subject Sample #1	0.130	03:06
Breath Volume	2.941	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample #2	0.132	03:09
Breath Volume	2.953	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 02:43

Gender: Male

Observation Period Began: 03:00

Subject's Last Name: TOBEN

Date/Time of Test: 01/07/2007 03:29

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample #1	0.205	03:25
Breath Volume	4.289	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample #2	0.199	03:28
Breath Volume	3.902	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GONZALEZ, JOEL,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: GARCIA, EDUARDO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:55

Gender: Male

Observation Period Began: 04:10

Subject's Last Name: MUNOZ

Date/Time of Test: 01/07/2007 04:36

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.080	04:34	
Air Blank	0.000	04:35	
Subject Sample #1		04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.079	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:34

Gender: Male

Observation Period Began: 04:42

Subject's Last Name: JURADO

Date/Time of Test: 01/07/2007 05:17

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.080	05:11
Air Blank	0.000	05:11
Subject Sample #1	0.090	05:13
Breath Volume	1.816	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:15
Subject Sample #2	0.079	05:16
Breath Volume	1.351	05:16
Air Blank	0.000	05:16
Control Test	0.077	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVEN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:42

Gender: Male

Observation Period Began: 04:50

Subject's Last Name: MCCARTY

Date/Time of Test: 01/07/2007 05:26

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:24	
Control Test	0.079	05:24	
Air Blank	0.000	05:24	
Subject Sample #1		05:25	Subject Test Refused
Breath Volume	0.000	05:25	
Air Blank	0.000	05:25	
Control Test	0.079	05:25	
Air Blank	0.000	05:26	
Diagnostics Check	OK	05:26	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, GUILLERMO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 06:51

Gender: Male

Observation Period Began: 07:05

Subject's Last Name: GUIA

Date/Time of Test: 01/07/2007 07:30

Diagnostics Check	OK	07:26	
Air Blank	0.000	07:27	
Control Test	0.080	07:27	
Air Blank	0.000	07:28	
Subject Sample #1		07:28	Subject Test Refused
Breath Volume	0.000	07:28	
Air Blank	0.000	07:29	
Control Test	0.080	07:29	
Air Blank	0.000	07:30	
Diagnostics Check	OK	07:30	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:30

Gender: Male

Observation Period Began: 04:40

Subject's Last Name: HOLLAND

Date/Time of Test: 01/08/2007 05:13

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:04
Air Blank	0.000	05:04
Subject Sample #1	0.196	05:06
Breath Volume	1.207	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample #2	0.200	05:11
Breath Volume	1.890	05:11
Air Blank	0.000	05:12
Control Test	0.078	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WILDER, NEWELL

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 01:15

Gender: Male

Observation Period Began: 01:45

Subject's Last Name: PHILLIPS

Date/Time of Test: 01/10/2007 02:14

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:08	
Subject Sample #1	0.056	02:09	
Breath Volume	1.472	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample #2	0.063	02:13	
Breath Volume	1.226	02:13	
Air Blank	0.000	02:13	
Control Test	0.069	02:14	RFI Detect
Air Blank		02:14	RFI Detect

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GARCIA, EDUARDO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MARTINEZ, LLOYD

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:11

Gender: Male

Observation Period Began: 03:26

Subject's Last Name: GLADSTONE

Date/Time of Test: 01/10/2007 03:59

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Subject Sample #1	0.008	03:55
Breath Volume	3.562	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #2	0.007	03:58
Breath Volume	3.722	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:36

Gender: Female

Observation Period Began: 03:43

Subject's Last Name: LOPEZ

Date/Time of Test: 01/12/2007 04:06

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.080	04:04	
Air Blank	0.000	04:05	
Subject Sample #1		04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GARCIA, EDUARDO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVEN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:52

Gender: Male

Observation Period Began: 03:56

Subject's Last Name: GONZALEZ

Date/Time of Test: 01/12/2007 04:23

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample #1	0.176	04:19
Breath Volume	3.121	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample #2	0.171	04:22
Breath Volume	2.429	04:22
Air Blank	0.000	04:22
Control Test	0.077	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVEN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:20

Gender: Male

Observation Period Began: 05:35

Subject's Last Name: KELER

Date/Time of Test: 01/12/2007 06:03

Diagnostics Check	OK	05:56
Air Blank	0.000	05:57
Control Test	0.080	05:57
Air Blank	0.000	05:58
Subject Sample #1	0.230	05:58
Breath Volume	2.542	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample #2	0.228	06:01
Breath Volume	4.074	06:01
Air Blank	0.000	06:02
Control Test	0.078	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MESTRE, ENRIQUE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: OTERO, ALBERTO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 01:20

Gender: Male

Observation Period Began: 02:20

Subject's Last Name: MESTRES

Date/Time of Test: 01/13/2007 02:48

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample #1	0.000	02:43
Breath Volume	2.492	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample #2	0.000	02:46
Breath Volume	1.628	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:10

Gender: Male

Observation Period Began: 03:20

Subject's Last Name: AZOULAY

Date/Time of Test: 01/13/2007 03:48

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample #1	0.144	03:44
Breath Volume	3.339	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample #2	0.136	03:47
Breath Volume	2.222	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MACIAS, EDUARDO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:30

Gender: Male

Observation Period Began: 03:37

Subject's Last Name: SUERO KING

Date/Time of Test: 01/13/2007 04:34

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.079	04:32	
Air Blank	0.000	04:33	
Subject Sample #1		04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:34	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GARCIA, EDUARDO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: CAMPBELL, JAVON

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:19

Gender: Male

Observation Period Began: 04:27

Subject's Last Name: GEORGES

Date/Time of Test: 01/13/2007 04:55

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:49
Subject Sample #1	0.192	04:50
Breath Volume	2.062	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample #2	0.193	04:54
Breath Volume	2.507	04:54
Air Blank	0.000	04:54
Control Test	0.078	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:00

Gender: Male

Observation Period Began: 05:10

Subject's Last Name: CAMPUZANO

Date/Time of Test: 01/13/2007 05:41

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:35
Subject Sample #1	0.136	05:36
Breath Volume	1.121	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample #2	0.156	05:39
Breath Volume	1.574	05:39
Air Blank	0.000	05:40
Control Test	0.078	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: FLAHERTY, ARLEY,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVEN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 06:47

Gender: Female

Observation Period Began: 07:00

Subject's Last Name: DURAN

Date/Time of Test: 01/13/2007 07:33

Diagnostics Check	OK	07:24
Air Blank	0.000	07:25
Control Test	0.079	07:25
Air Blank	0.000	07:26
Subject Sample #1	0.139	07:29
Breath Volume	1.183	07:29
Air Blank	0.000	07:30
Air Blank	0.000	07:31
Subject Sample #2	0.130	07:32
Breath Volume	2.304	07:32
Air Blank	0.000	07:32
Control Test	0.075	07:33
Air Blank	0.000	07:33
Diagnostics Check	OK	07:33

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 23:30

Gender: Male

Observation Period Began: 23:50

Subject's Last Name: MURRAY

Date/Time of Test: 01/14/2007 00:14

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.080	00:11	
Air Blank	0.000	00:12	
Subject Sample #1		00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MACIAS, EDUARDO,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 02:32

Gender: Male

Observation Period Began: 02:44

Subject's Last Name: TICHY

Date/Time of Test: 01/14/2007 03:10

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Subject Sample #1		03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:48

Gender: Female

Observation Period Began: 03:57

Subject's Last Name: MARSHALL

Date/Time of Test: 01/14/2007 04:20

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:18	
Subject Sample #1		04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GUERRERO, SAMIR,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:13

Gender: Female

Observation Period Began: 04:30

Subject's Last Name: BENNETT

Date/Time of Test: 01/14/2007 05:09

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.075	05:04
Air Blank	0.000	05:04
Subject Sample #1	0.135	05:05
Breath Volume	1.476	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample #2	0.137	05:07
Breath Volume	1.707	05:07
Air Blank	0.000	05:08
Control Test	0.077	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:52

Gender: Male

Observation Period Began: 06:02

Subject's Last Name: PEREZ

Date/Time of Test: 01/14/2007 06:26

Diagnostics Check	OK	06:23	
Air Blank	0.000	06:23	
Control Test	0.080	06:24	
Air Blank	0.000	06:24	
Subject Sample #1		06:24	Subject Test Refused
Breath Volume	0.000	06:24	
Air Blank	0.000	06:25	
Control Test	0.080	06:25	
Air Blank	0.000	06:26	
Diagnostics Check	OK	06:26	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GARCIA, EDUARDO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: PEREZ, LUIS

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:48

Gender: Female

Observation Period Began: 06:10

Subject's Last Name: INCHAUSTI

Date/Time of Test: 01/14/2007 07:01

Diagnostics Check	OK	06:54
Air Blank	0.000	06:54
Control Test	0.079	06:55
Air Blank	0.000	06:55
Subject Sample #1	0.137	06:56
Breath Volume	2.875	06:56
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample #2	0.142	06:59
Breath Volume	2.273	06:59
Air Blank	0.000	07:00
Control Test	0.078	07:00
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MILLAN, KEVIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FERNANDEZ, GIRANNY

Arresting Officer Agency: Surfside PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 01:20

Gender: Male

Observation Period Began: 02:14

Subject's Last Name: MONTERO

Date/Time of Test: 01/15/2007 02:38

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Subject Sample #1		02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.081	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MILLAN, KEVIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:30

Gender: Male

Observation Period Began: 05:37

Subject's Last Name: MCLEAN

Date/Time of Test: 01/15/2007 06:11

Diagnostics Check	OK	06:04
Air Blank	0.000	06:04
Control Test	0.080	06:05
Air Blank	0.000	06:05
Subject Sample #1	0.147	06:06
Breath Volume	1.667	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:08
Subject Sample #2	0.139	06:09
Breath Volume	1.816	06:09
Air Blank	0.000	06:10
Control Test	0.080	06:10
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: PEREZ, LUIS, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ZAMUDIO, MARIA

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 00:41

Gender: Female

Observation Period Began: 00:55

Subject's Last Name: MARINUCCI

Date/Time of Test: 01/16/2007 01:34

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample #1		01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:02

Gender: Female

Observation Period Began: 04:16

Subject's Last Name: SMITH

Date/Time of Test: 01/17/2007 04:39

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:37	
Subject Sample #1		04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:38

Gender: Male

Observation Period Began: 05:04

Subject's Last Name: ANSPACH

Date/Time of Test: 01/17/2007 05:40

Diagnostics Check	OK	05:37	
Air Blank	0.000	05:38	
Control Test	0.080	05:38	
Air Blank	0.000	05:39	
Subject Sample #1		05:39	Subject Test Refused
Breath Volume	0.000	05:39	
Air Blank	0.000	05:39	
Control Test	0.080	05:40	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: PFROGNER, COLIN, L

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: PFROGNER, COLIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 10:39

Gender: Male

Observation Period Began: 11:12

Subject's Last Name: VRAHNOS

Date/Time of Test: 01/17/2007 11:48

Diagnostics Check	OK	11:40
Air Blank	0.000	11:41
Control Test	0.080	11:41
Air Blank	0.000	11:42
Subject Sample #1	0.128	11:43
Breath Volume	1.808	11:43
Air Blank	0.000	11:44
Air Blank	0.000	11:46
Subject Sample #2	0.131	11:47
Breath Volume	3.343	11:47
Air Blank	0.000	11:47
Control Test	0.078	11:48
Air Blank	0.000	11:48
Diagnostics Check	OK	11:48

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:37

Gender: Male

Observation Period Began: 04:00

Subject's Last Name: MILLER

Date/Time of Test: 01/18/2007 04:31

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.079	04:29	
Air Blank	0.000	04:30	
Subject Sample #1		04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:30	
Control Test	0.080	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:55

Gender: Male

Observation Period Began: 04:20

Subject's Last Name: BAKER

Date/Time of Test: 01/18/2007 04:59

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:55	
Control Test	0.079	04:55	
Air Blank	0.000	04:56	
Subject Sample #1		04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:58	
Control Test	0.080	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:10

Gender: Male

Observation Period Began: 05:15

Subject's Last Name: LINDSEY

Date/Time of Test: 01/18/2007 05:40

Diagnostics Check	OK	05:37	
Air Blank	0.000	05:37	
Control Test	0.080	05:38	
Air Blank	0.000	05:38	
Subject Sample #1		05:38	Subject Test Refused
Breath Volume	0.000	05:38	
Air Blank	0.000	05:39	
Control Test	0.081	05:39	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: COSNER, STEVEN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 06:00

Gender: Female

Observation Period Began: 06:12

Subject's Last Name: CIFERRI

Date/Time of Test: 01/18/2007 06:35

Diagnostics Check	OK	06:32	
Air Blank	0.000	06:33	
Control Test	0.080	06:33	
Air Blank	0.000	06:33	
Subject Sample #1		06:34	Subject Test Refused
Breath Volume	0.000	06:34	
Air Blank	0.000	06:34	
Control Test	0.080	06:34	
Air Blank	0.000	06:35	
Diagnostics Check	OK	06:35	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: COSNER, STEVE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:52

Gender: Female

Observation Period Began: 06:00

Subject's Last Name: SANCHEZ

Date/Time of Test: 01/18/2007 06:56

Diagnostics Check	OK	06:50
Air Blank	0.000	06:51
Control Test	0.081	06:51
Air Blank	0.000	06:52
Subject Sample #1	0.141	06:52
Breath Volume	2.464	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:54
Subject Sample #2	0.138	06:55
Breath Volume	1.632	06:55
Air Blank	0.000	06:56
Control Test	0.076	06:56
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MEDINA, FULGENCIO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 21:30

Gender: Female

Observation Period Began: 21:45

Subject's Last Name: HUMBLE

Date/Time of Test: 01/18/2007 22:16

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:09
Subject Sample #1	0.214	22:09
Breath Volume	1.144	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample #2	0.214	22:14
Breath Volume	1.957	22:14
Air Blank	0.000	22:15
Control Test	0.078	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MESTRE, ENIQUE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: COVLAT, ROBERT

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 01:12

Gender: Male

Observation Period Began: 01:18

Subject's Last Name: PEREZ GONZALEZ

Date/Time of Test: 01/19/2007 01:43

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.079	01:41	
Air Blank	0.000	01:42	
Subject Sample #1		01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.079	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:28

Gender: Female

Observation Period Began: 03:36

Subject's Last Name: MACCACHERO

Date/Time of Test: 01/19/2007 04:00

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Subject Sample #1		03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 12/20/2006

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:02

Gender: Male

Observation Period Began: 05:07

Subject's Last Name: COHEN

Date/Time of Test: 01/19/2007 05:31

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:29	
Control Test	0.080	05:29	
Air Blank	0.000	05:29	
Subject Sample #1		05:30	Subject Test Refused
Breath Volume	0.000	05:30	
Air Blank	0.000	05:30	
Control Test	0.081	05:31	
Air Blank	0.000	05:31	
Diagnostics Check	OK	05:31	

Data Download Date/Time: 01/19/2007 23:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: WIRTH, PHILLIP

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:15

Gender: Male

Observation Period Began: 03:35

Subject's Last Name: RUBIO DIAZ LEAL

Date/Time of Test: 01/20/2007 04:06

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample #1	0.150	04:00	Volume Not Met
Breath Volume	0.269	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample #2		04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.078	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BARRETO-VO, LIMURA

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:40

Gender: Male

Observation Period Began: 04:00

Subject's Last Name: FERNANDEZ

Date/Time of Test: 01/20/2007 04:30

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Subject Sample #1		04:28	Subject Test Refused
Breath Volume	0.164	04:28	
Air Blank	0.000	04:29	
Control Test	0.079	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: ROBINSON, SAMARIA, V

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:00

Gender: Male

Observation Period Began: 04:09

Subject's Last Name: MARSAL

Date/Time of Test: 01/20/2007 04:44

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.078	04:38
Air Blank	0.000	04:38
Subject Sample #1	0.081	04:39
Breath Volume	3.839	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample #2	0.076	04:42
Breath Volume	2.738	04:42
Air Blank	0.000	04:43
Control Test	0.078	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: GARCIA, EDUARDO, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: GARCIA, EDUARDO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:18

Gender: Female

Observation Period Began: 04:24

Subject's Last Name: WHITAKER

Date/Time of Test: 01/20/2007 05:14

Diagnostics Check	OK	05:01	
Air Blank	0.000	05:02	
Control Test	0.078	05:02	
Air Blank	0.000	05:03	
Subject Sample #1	0.167	05:06	Volume Not Met
Breath Volume	0.789	05:06	
Air Blank	0.000	05:07	
Air Blank	0.000	05:08	
Subject Sample #2	0.165	05:09	
Breath Volume	2.050	05:09	
Air Blank	0.000	05:10	
Air Blank	0.000	05:12	
Subject Sample #3	0.161	05:12	
Breath Volume	1.449	05:12	
Air Blank	0.000	05:13	
Control Test	0.076	05:13	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DELCASTILLO, GUILLERMO, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ANDERSON, DONALD

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:45

Gender: Male

Observation Period Began: 06:00

Subject's Last Name: NEVEU

Date/Time of Test: 01/20/2007 06:29

Diagnostics Check	OK	06:22
Air Blank	0.000	06:22
Control Test	0.079	06:22
Air Blank	0.000	06:23
Subject Sample #1	0.151	06:23
Breath Volume	2.406	06:23
Air Blank	0.000	06:24
Air Blank	0.000	06:26
Subject Sample #2	0.146	06:27
Breath Volume	2.132	06:27
Air Blank	0.000	06:28
Control Test	0.080	06:28
Air Blank	0.000	06:28
Diagnostics Check	OK	06:29

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: ROBINSON, SAMARIA, V

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: ROBINSON, SAMARIA

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:00

Gender: Male

Observation Period Began: 04:19

Subject's Last Name: HITE

Date/Time of Test: 01/21/2007 04:46

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample #1	0.042	04:42
Breath Volume	2.046	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample #2	0.042	04:45
Breath Volume	3.449	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MCVEY, GREGORY, B

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: BARRETO-VO, LIMURA

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:45

Gender: Male

Observation Period Began: 06:15

Subject's Last Name: SANTIAGO

Date/Time of Test: 01/21/2007 06:43

Diagnostics Check	OK	06:36
Air Blank	0.000	06:37
Control Test	0.079	06:37
Air Blank	0.000	06:37
Subject Sample #1	0.169	06:38
Breath Volume	4.156	06:38
Air Blank	0.000	06:39
Air Blank	0.000	06:41
Subject Sample #2	0.166	06:41
Breath Volume	2.808	06:41
Air Blank	0.000	06:42
Control Test	0.079	06:42
Air Blank	0.000	06:43
Diagnostics Check	OK	06:43

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: PEREZ, LUIS, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: COVALT, ROBERT

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 00:37

Gender: Female

Observation Period Began: 00:44

Subject's Last Name: MENDEZ

Date/Time of Test: 01/24/2007 01:20

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.081	01:14
Air Blank	0.000	01:14
Subject Sample #1	0.109	01:15
Breath Volume	1.445	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample #2	0.102	01:18
Breath Volume	2.140	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: CALLAN, ANTHONY

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:57

Gender: Male

Observation Period Began: 04:10

Subject's Last Name: REYNOLDS

Date/Time of Test: 01/24/2007 04:55

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.080	04:53	
Air Blank	0.000	04:53	
Subject Sample #1		04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	
Control Test	0.080	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: WILDER, NEWELL, G

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MILLAN, KEVIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 02:07

Gender: Male

Observation Period Began: 02:24

Subject's Last Name: DEGEN

Date/Time of Test: 01/25/2007 03:04

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Subject Sample #1		03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERIC

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 20:20

Gender: Female

Observation Period Began: 20:30

Subject's Last Name: FUTRELLE

Date/Time of Test: 01/25/2007 20:59

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:54
Air Blank	0.000	20:54
Subject Sample #1	0.142	20:54
Breath Volume	2.054	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample #2	0.138	20:57
Breath Volume	1.281	20:57
Air Blank	0.000	20:58
Control Test	0.079	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DIONNE, MARTIN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:54

Gender: Male

Observation Period Began: 05:03

Subject's Last Name: LACAYO

Date/Time of Test: 01/26/2007 05:30

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:27	
Control Test	0.079	05:27	
Air Blank	0.000	05:28	
Subject Sample #1		05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:29	
Control Test	0.081	05:29	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: VALDES, MANUEL,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DOMINGUEZ, ERICK

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:46

Gender: Male

Observation Period Began: 03:54

Subject's Last Name: GUILFOIL WRIGHT

Date/Time of Test: 01/27/2007 04:20

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:15
Air Blank	0.000	04:15
Subject Sample #1	0.188	04:16
Breath Volume	3.230	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample #2	0.188	04:19
Breath Volume	1.398	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: VALDES, MANUEL,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVE

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:42

Gender: Female

Observation Period Began: 03:46

Subject's Last Name: LEZCANO

Date/Time of Test: 01/27/2007 04:41

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Subject Sample #1	0.126	04:37
Breath Volume	1.582	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.124	04:40
Breath Volume	1.480	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DOMINGUEZ, ERIC,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FELDMAN, STEVEN

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:42

Gender: Male

Observation Period Began: 05:53

Subject's Last Name: CORREA

Date/Time of Test: 01/27/2007 06:22

Diagnostics Check	OK	06:14
Air Blank	0.000	06:15
Control Test	0.080	06:15
Air Blank	0.000	06:16
Subject Sample #1	0.165	06:17
Breath Volume	2.222	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample #2	0.149	06:20
Breath Volume	1.500	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: DIONNE, MARTIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: GUASTO, NICHOLAS

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 03:52

Gender: Male

Observation Period Began: 03:59

Subject's Last Name: FREIXAS

Date/Time of Test: 01/28/2007 04:25

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.078	04:22	
Air Blank	0.000	04:23	
Subject Sample #1		04:23	Subject Test Refused
Breath Volume	0.000	04:23	
Air Blank	0.000	04:24	
Control Test	0.080	04:24	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MACIAS, EDUARDO,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MACIAS, EDUARDO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:00

Gender: Male

Observation Period Began: 05:20

Subject's Last Name: SOMERVILLE

Date/Time of Test: 01/28/2007 05:52

Diagnostics Check	OK	05:49	
Air Blank	0.000	05:49	
Control Test	0.080	05:50	
Air Blank	0.000	05:50	
Subject Sample #1		05:51	Subject Test Refused
Breath Volume	0.000	05:51	
Air Blank	0.000	05:51	
Control Test	0.081	05:51	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MACIAS, EDUARDO,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: MACIAS, EDUARDO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:15

Gender: Male

Observation Period Began: 05:35

Subject's Last Name: GIORDANI

Date/Time of Test: 01/28/2007 06:14

Diagnostics Check	OK	06:07
Air Blank	0.000	06:08
Control Test	0.080	06:08
Air Blank	0.000	06:09
Subject Sample #1	0.158	06:09
Breath Volume	2.664	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample #2	0.161	06:12
Breath Volume	1.847	06:12
Air Blank	0.000	06:13
Control Test	0.079	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MESTRE, ENRIQUE,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: DELCASTILLO, FRANCISCO

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 02:25

Gender: Male

Observation Period Began: 03:12

Subject's Last Name: KOSINSKII

Date/Time of Test: 01/29/2007 03:43

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:33	
Subject Sample #1	0.108	03:37	Volume Not Met
Breath Volume	0.175	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:39	
Subject Sample #2		03:41	Subject Test Refused
Breath Volume	0.214	03:41	
Air Blank	0.000	03:42	
Control Test	0.076	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: HUMPHRY, KERRY, R

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: LESTER, R

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 05:35

Gender: Female

Observation Period Began: 06:09

Subject's Last Name: MEJIAS

Date/Time of Test: 01/29/2007 06:39

Diagnostics Check	OK	06:30
Air Blank	0.000	06:30
Control Test	0.080	06:31
Air Blank	0.000	06:31
Subject Sample #1	0.116	06:34
Breath Volume	1.570	06:34
Air Blank	0.000	06:35
Air Blank	0.000	06:36
Subject Sample #2	0.122	06:37
Breath Volume	1.601	06:37
Air Blank	0.000	06:38
Control Test	0.079	06:38
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: MILLAN, KEVIN,

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: VERNON, LASCELLES

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 00:50

Gender: Female

Observation Period Began: 00:55

Subject's Last Name: BAUMAN

Date/Time of Test: 01/30/2007 01:18

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Subject Sample #1		01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:17	
Control Test	0.078	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: PEREZ, LUIS, A

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: FLYNN, MELISSA

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 02:47

Gender: Male

Observation Period Began: 02:57

Subject's Last Name: BORGES

Date/Time of Test: 01/30/2007 03:36

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
Subject Sample #1		03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.081	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 02/27/2007 07:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number : 80-002232

Instrument Registered To: MIAMI BEACH PD

Software Version: 8100.27

Breath Test Operator: KUNCAS, ANDREW, J

Breath Test Operator Agency: Miami Beach PD

Arresting Officer: CALLAN, TONY

Arresting Officer Agency: Miami Beach PD

Last Agency Inspection: 01/19/2007

Date or Time Changed: No

Dry Gas Standard Lot #: 622802I

Dry Gas Expiry Date: 08/18/2008

Violation Code: DUI

Arrest Time: 04:40

Gender: Male

Observation Period Began: 05:55

Subject's Last Name: HEMERSON

Date/Time of Test: 01/31/2007 06:23

Diagnostics Check	OK	06:19	
Air Blank	0.000	06:20	
Control Test	0.080	06:20	
Air Blank	0.000	06:21	
Subject Sample #1		06:21	Subject Test Refused
Breath Volume	0.000	06:21	
Air Blank	0.000	06:22	
Control Test	0.080	06:22	
Air Blank	0.000	06:23	
Diagnostics Check	OK	06:23	

Data Download Date/Time: 02/27/2007 07:51