

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG, JESSE, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: YOUNG, JESSE  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/02/2011 05:50:06

---

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:46	
Control Test	0.078	05:47	
Air Blank	0.000	05:47	
Subject Sample #1		05:48	Subject Test Refused
Breath Volume	0.000	05:48	
Air Blank	0.000	05:49	
Control Test	0.078	05:49	
Air Blank	0.000	05:50	
Diagnostics Check	OK	05:50	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STOCKTON, JOSEPH, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: STOCKTON, JOSEPH  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 08/04/2011 01:50:17

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.077	01:44
Air Blank	0.000	01:45
Subject Sample #1	0.251	01:45
Breath Volume	1.433	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample #2	0.249	01:48
Breath Volume	1.343	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YOUNG, JESSE, A

BREATH TEST OPERATOR AGENCY: Key West PD

ARRESTING OFFICER: YOUNG, JESSE

ARRESTING OFFICER AGENCY: Key West PD

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 00:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: YASHASINOV

DATE/TIME OF TEST: 08/06/2011 03:28:28

---

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.078	03:26	
Air Blank	0.000	03:26	
Subject Sample #1		03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.078	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN-REYES, JUAN, C  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MARTIN-REYES, JUAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: KATSARELIS

DATE/TIME OF TEST: 08/06/2011 04:43:48

---

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.078	04:41	
Air Blank	0.000	04:42	
Subject Sample #1		04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.078	04:43	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG, JESSE, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: YOUNG, JESSE  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 08/07/2011 01:37:28

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Subject Sample #1	0.193	01:32
Breath Volume	4.156	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample #2	0.194	01:35
Breath Volume	3.632	01:35
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG, JESSE, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: YOUNG, JESSE  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 08/07/2011 04:54:03

---

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.077	04:43	
Air Blank	0.000	04:43	
Subject Sample #1		04:47	No Sample Provided
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample #2		04:52	No Sample Provided
Breath Volume	0.000	04:52	
Air Blank	0.000	04:53	
Control Test	0.077	04:53	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:54	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YOUNG, JESSE, A

BREATH TEST OPERATOR AGENCY: Key West PD

ARRESTING OFFICER: FRASER, CHRISTOPHER

ARRESTING OFFICER AGENCY: Monroe County SO

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 21:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: NUNNAMAKER

DATE/TIME OF TEST: 08/07/2011 23:59:31

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.077	23:53
Air Blank	0.000	23:54
Subject Sample #1	0.123	23:55
Breath Volume	3.750	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample #2	0.118	23:57
Breath Volume	2.437	23:57
Air Blank	0.000	23:58
Control Test	0.077	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG, JESSE, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: YOUNG, JESSE  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 08/08/2011 03:18:49

---

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.077	03:11	
Air Blank	0.000	03:11	
Subject Sample #1		03:14	No Sample Provided
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample #2		03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.077	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANCHEZ, LUIS, E  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: REILLY, THOMAS  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 08/08/2011 22:02:15

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.078	21:56
Air Blank	0.000	21:57
Subject Sample #1	0.293	21:57
Breath Volume	3.148	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample #2	0.292	22:00
Breath Volume	3.335	22:00
Air Blank	0.000	22:01
Control Test	0.076	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VALDES, JIMMY,

BREATH TEST OPERATOR AGENCY: Key West PD

ARRESTING OFFICER: FRASER, CHRIS

ARRESTING OFFICER AGENCY: Monroe County SO

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 22:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: FREY

DATE/TIME OF TEST: 08/10/2011 23:23:16

---

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Subject Sample #1	0.084	23:18
Breath Volume	4.867	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample #2	0.089	23:21
Breath Volume	4.019	23:21
Air Blank	0.000	23:22
Control Test	0.077	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG, JESSE, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: YOUNG, JESSE  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 08/11/2011 22:23:31

---

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.076	22:17
Air Blank	0.000	22:18
Subject Sample #1	0.277	22:18
Breath Volume	3.273	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample #2	0.277	22:21
Breath Volume	2.671	22:21
Air Blank	0.000	22:22
Control Test	0.077	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG, JESSE, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: YOUNG, JESSE  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 08/16/2011 01:58:59

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.077	01:51	
Air Blank	0.000	01:51	
Subject Sample #1		01:54	No Sample Provided
Breath Volume	0.000	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample #2		01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Control Test	0.077	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG, JESSE, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: FRASER, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: KEATON

DATE/TIME OF TEST: 08/16/2011 23:26:29

---

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.077	23:13	
Air Blank	0.000	23:13	
Subject Sample #1	0.256	23:15	
Breath Volume	3.777	23:15	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample #2		23:20	No Sample Provided
Breath Volume	0.000	23:20	
Air Blank	0.000	23:21	
Air Blank	0.000	23:23	
Subject Sample #3	0.251	23:24	
Breath Volume	3.671	23:24	
Air Blank	0.000	23:25	
Control Test	0.077	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YOUNG, JESSE, A

BREATH TEST OPERATOR AGENCY: Key West PD

ARRESTING OFFICER: YOUNG, JESSE

ARRESTING OFFICER AGENCY: Key West PD

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 21:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/19/2011 22:05:56

---

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.078	22:00
Air Blank	0.000	22:00
Subject Sample #1	0.187	22:01
Breath Volume	1.375	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample #2	0.191	22:04
Breath Volume	1.136	22:04
Air Blank	0.000	22:05
Control Test	0.078	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YOUNG, JESSE, A

BREATH TEST OPERATOR AGENCY: Key West PD

ARRESTING OFFICER: YOUNG, JESSE

ARRESTING OFFICER AGENCY: Key West PD

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 01:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GAITAN

DATE/TIME OF TEST: 08/20/2011 02:09:11

---

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.078	02:06	
Air Blank	0.000	02:07	
Subject Sample #1		02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.078	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG, JESSE, A  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: YOUNG, JESSE  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 08/20/2011 22:58:21

---

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.077	22:50	
Air Blank	0.000	22:51	
Subject Sample #1		22:54	No Sample Provided
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample #2		22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Control Test	0.078	22:57	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YOUNG, JESSE, A

BREATH TEST OPERATOR AGENCY: Key West PD

ARRESTING OFFICER: CELCER, GARY

ARRESTING OFFICER AGENCY: Key West PD

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 05:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: LONGTIN

DATE/TIME OF TEST: 08/21/2011 06:08:46

---

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.078	06:02
Air Blank	0.000	06:02
Subject Sample #1	0.202	06:03
Breath Volume	1.699	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample #2	0.192	06:07
Breath Volume	1.726	06:07
Air Blank	0.000	06:07
Control Test	0.076	06:08
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN-REYES, JUAN, C  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MARTIN-REYES, JUAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: ROUSSEAU

DATE/TIME OF TEST: 08/25/2011 05:36:00

---

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.078	05:30
Air Blank	0.000	05:30
Subject Sample #1	0.172	05:31
Breath Volume	3.792	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:33
Subject Sample #2	0.167	05:34
Breath Volume	4.265	05:34
Air Blank	0.000	05:35
Control Test	0.077	05:35
Air Blank	0.000	05:35
Diagnostics Check	OK	05:36

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN-REYES, JUAN, C  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FRASER, CHRIS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SCHIERING

DATE/TIME OF TEST: 08/26/2011 02:29:45

---

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.077	02:27	
Air Blank	0.000	02:28	
Subject Sample #1		02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.077	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN-REYES, JUAN, C  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FRASER, CHRIS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 08/27/2011 01:53:10

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.078	01:51	
Air Blank	0.000	01:51	
Subject Sample #1		01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.077	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GABAY, JOHN, K  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FRASER, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: MOORHEAD

DATE/TIME OF TEST: 08/30/2011 04:52:46

---

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:50	
Control Test	0.079	04:50	
Air Blank	0.000	04:51	
Subject Sample #1		04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:51	
Control Test	0.078	04:52	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAVARRIA, DEGLYS, E  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: CHAVARRIA, DEGLYS  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 08/31/2011 18:58:44

---

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:56	
Control Test	0.077	18:56	
Air Blank	0.000	18:57	
Subject Sample #1		18:57	Subject Test Refused
Breath Volume	0.000	18:57	
Air Blank	0.000	18:57	
Control Test	0.077	18:58	
Air Blank	0.000	18:58	
Diagnostics Check	OK	18:58	

---

Data Download Date/Time 08/31/2011 21:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DICKINSON, GORDON, L  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: ALEMAN, ANTONIO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/09/2011 20:15:44

---

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:12	
Control Test	0.081	20:12	
Air Blank	0.000	20:13	
Subject Sample #1		20:14	Subject Test Refused
Breath Volume	0.000	20:14	
Air Blank	0.000	20:14	
Control Test	0.080	20:15	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

---

Data Download Date/Time 08/28/2011 18:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, DANIEL, A  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: CABRAL, EDUARDO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 08:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/10/2011 10:03:11

---

Diagnostics Check	OK	09:56
Air Blank	0.000	09:56
Control Test	0.082	09:57
Air Blank	0.000	09:57
Subject Sample #1	0.149	09:58
Breath Volume	1.660	09:58
Air Blank	0.000	09:59
Air Blank	0.000	10:00
Subject Sample #2	0.141	10:01
Breath Volume	1.281	10:01
Air Blank	0.000	10:02
Control Test	0.081	10:02
Air Blank	0.000	10:03
Diagnostics Check	OK	10:03

---

Data Download Date/Time 08/28/2011 18:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEMAN, ANTONIO,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: STAMAND, STEVE  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CARRANZA

DATE/TIME OF TEST: 08/14/2011 22:01:31

---

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.081	21:56	
Air Blank	0.000	21:56	
Subject Sample #1	0.155	21:57	
Breath Volume	2.578	21:57	
Air Blank	0.000	21:57	
Air Blank	0.000	21:59	
Subject Sample #2		22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.080	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

---

Data Download Date/Time 08/28/2011 18:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA, MANUEL, A  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: MEDINA, MANUEL  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 08/23/2011 01:41:37

---

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.081	01:39	
Air Blank	0.000	01:40	
Subject Sample #1		01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:40	
Control Test	0.081	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

---

Data Download Date/Time 08/28/2011 18:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERO, MANNY,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: MATEU, O  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: TOSCA

DATE/TIME OF TEST: 08/23/2011 02:13:06

---

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:10	
Air Blank	0.000	02:10	
Subject Sample #1		02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.081	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time 08/28/2011 18:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARDIS, JOHN, D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: FARINAS, GEISY  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: TABORA

DATE/TIME OF TEST: 08/27/2011 03:30:29

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.082	03:24
Air Blank	0.000	03:25
Subject Sample #1	0.118	03:25
Breath Volume	2.148	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample #2	0.118	03:28
Breath Volume	1.792	03:28
Air Blank	0.000	03:29
Control Test	0.082	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time 08/28/2011 18:21

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALEXANDER, KETRA, J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ALEXANDER, K

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 23:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CAPETILLO

DATE/TIME OF TEST: 08/01/2011 00:18:58

---

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Subject Sample #1		00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Control Test	0.080	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGO, ROLANDO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LT. LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 01:52

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CONSTANDSE

DATE/TIME OF TEST: 08/06/2011 03:17:04

---

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:05	
Control Test	0.079	03:05	
Air Blank	0.000	03:06	
Subject Sample #1	0.048	03:09	Volume Not Met
Breath Volume	0.406	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:11	
Subject Sample #2	0.047	03:12	
Breath Volume	1.652	03:12	
Air Blank	0.000	03:13	
Air Blank	0.000	03:14	
Subject Sample #3	0.045	03:15	
Breath Volume	1.554	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:17	

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDRADE, CRISTIAN, A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ANDRADE, CRISTIAN

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 22:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 08/07/2011 23:22:47

---

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample #1	0.232	23:14
Breath Volume	2.398	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample #2	0.211	23:18
Breath Volume	1.476	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample #3	0.225	23:21
Breath Volume	2.218	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IRIAS, R,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: BLANCO, R  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG002801  
DRY GAS EXPIRY DATE: 01/28/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 08/11/2011 05:37:10

---

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:34	
Control Test	0.079	05:34	
Air Blank	0.000	05:35	
Subject Sample #1		05:35	Subject Test Refused
Breath Volume	0.000	05:35	
Air Blank	0.000	05:36	
Control Test	0.080	05:36	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CABRERA, ROGER, J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MANSOUR, AZEEZ

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 03:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 08/14/2011 03:42:31

---

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:40	
Subject Sample #1		03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CABRERA, ROGER, J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CABRERA, ROGER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 06:01

GENDER: Female

OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 08/14/2011 06:33:24

---

Diagnostics Check	OK	06:26
Air Blank	0.000	06:27
Control Test	0.079	06:27
Air Blank	0.000	06:28
Subject Sample #1	0.146	06:28
Breath Volume	2.019	06:28
Air Blank	0.000	06:29
Air Blank	0.000	06:31
Subject Sample #2	0.135	06:31
Breath Volume	1.921	06:31
Air Blank	0.000	06:32
Control Test	0.078	06:32
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARIN, JOHN, H  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: CHAI, KAMAL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG002801  
DRY GAS EXPIRY DATE: 01/28/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 08/15/2011 01:28:33

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample #1	0.190	01:23
Breath Volume	2.023	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample #2	0.195	01:26
Breath Volume	1.761	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDRADE, CRISTIAN, A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CALDERON, HERVEY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 01:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: RAMSAY

DATE/TIME OF TEST: 08/15/2011 01:56:55

---

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.080	01:54	
Air Blank	0.000	01:55	
Subject Sample #1		01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Control Test	0.080	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RACZEK, MATTHEW, J

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: YANEZ, RANDY

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: BUI

ARREST TIME: 18:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: HUET VALERA

DATE/TIME OF TEST: 08/17/2011 19:38:20

---

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.079	19:34	
Air Blank	0.000	19:35	
Subject Sample #1		19:36	Subject Test Refused
Breath Volume	0.000	19:36	
Air Blank	0.000	19:37	
Control Test	0.080	19:37	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGO, ROLANDO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LT. LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 08/20/2011 04:01:30

---

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:48	
Subject Sample #1	0.159	03:51	Volume Not Met
Breath Volume	0.171	03:51	
Air Blank	0.000	03:52	
Air Blank	0.000	03:53	
Subject Sample #2	0.175	03:56	
Breath Volume	1.660	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample #3	0.179	03:59	
Breath Volume	2.394	03:59	
Air Blank	0.000	04:00	
Control Test	0.080	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGO, ROLANDO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: DOMINGO, ROLANDO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 01:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BOLTON

DATE/TIME OF TEST: 08/21/2011 02:07:58

---

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Subject Sample #1		02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGO, ROLANDO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LT. LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 02:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 08/25/2011 03:06:20

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample #1	0.116	03:01
Breath Volume	1.367	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.116	03:04
Breath Volume	1.980	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time 08/25/2011 13:29

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NUNES, FABIANO, P

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 03:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/26/2011 04:22:06

---

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:13	
Subject Sample #1	0.065	04:17	Volume Not Met
Breath Volume	0.843	04:17	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample #2		04:20	Subject Test Refused
Breath Volume	0.132	04:20	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

---

Data Download Date/Time 09/27/2011 12:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FRANCO, GREGORY, R

BREATH TEST OPERATOR AGENCY: Miami-Dade PD

ARRESTING OFFICER: FRANCO, GREGORY

ARRESTING OFFICER AGENCY: Miami-Dade PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 01:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 08/27/2011 02:50:27

---

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Subject Sample #1	0.139	02:45
Breath Volume	1.710	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample #2	0.140	02:48
Breath Volume	2.019	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

---

Data Download Date/Time 09/27/2011 12:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGO, ROLANDO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LT. LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 02:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 08/27/2011 03:19:59

---

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample #1	0.093	03:15
Breath Volume	1.472	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample #2	0.089	03:18
Breath Volume	2.738	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time 09/27/2011 12:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VITAL, CHRISTOPHER, D

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 05:02

GENDER: Female

OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 08/27/2011 05:50:06

---

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:41
Subject Sample #1	0.105	05:43
Breath Volume	1.261	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample #2	0.095	05:48
Breath Volume	1.832	05:48
Air Blank	0.000	05:49
Control Test	0.078	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

---

Data Download Date/Time 09/27/2011 12:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOMINGO, ROLANDO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LT. LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 05:29

GENDER: Female

OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: ASDOURIAN

DATE/TIME OF TEST: 08/27/2011 06:19:13

---

Diagnostics Check	OK	06:16	
Air Blank	0.000	06:16	
Control Test	0.078	06:16	
Air Blank	0.000	06:17	
Subject Sample #1		06:17	Subject Test Refused
Breath Volume	0.000	06:17	
Air Blank	0.000	06:18	
Control Test	0.078	06:18	
Air Blank	0.000	06:19	
Diagnostics Check	OK	06:19	

---

Data Download Date/Time 09/27/2011 12:35

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDRADE, CRISTIAN, A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ALEXANDER, KETRA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: LOPEZ RUIZ

DATE/TIME OF TEST: 08/01/2011 04:40:19

---

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.082	04:34
Air Blank	0.000	04:35
Subject Sample #1	0.176	04:35
Breath Volume	3.050	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample #2	0.161	04:38
Breath Volume	1.937	04:38
Air Blank	0.000	04:39
Control Test	0.083	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVILA, EDDIE,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: TABORDA, LUIS  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LANDERO

DATE/TIME OF TEST: 08/01/2011 19:43:23

---

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:40	
Control Test	0.082	19:41	
Air Blank	0.000	19:41	
Subject Sample #1		19:42	Subject Test Refused
Breath Volume	0.000	19:42	
Air Blank	0.000	19:42	
Control Test	0.083	19:42	
Air Blank	0.000	19:43	
Diagnostics Check	OK	19:43	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENNEDY, JOSEPH, W

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: KENNEDY, JOSEPH

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 19:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LOUREIRO

DATE/TIME OF TEST: 08/02/2011 19:43:56

---

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:41	
Control Test	0.082	19:41	
Air Blank	0.000	19:42	
Subject Sample #1		19:42	Subject Test Refused
Breath Volume	0.000	19:42	
Air Blank	0.000	19:43	
Control Test	0.081	19:43	
Air Blank	0.000	19:43	
Diagnostics Check	OK	19:43	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENNEDY, JOSEPH, W

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: KENNEDY, JOSEPH

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 19:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LOUREIRO

DATE/TIME OF TEST: 08/02/2011 20:15:29

---

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.082	20:08
Air Blank	0.000	20:08
Subject Sample #1	0.250	20:10
Breath Volume	1.609	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:12
Subject Sample #2	0.246	20:13
Breath Volume	1.136	20:13
Air Blank	0.000	20:14
Control Test	0.082	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLANC, CAROLE,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: LOCKE, JEFFREY  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LAVOICE

DATE/TIME OF TEST: 08/03/2011 00:19:09

---

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.082	00:17	
Air Blank	0.000	00:17	
Subject Sample #1		00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Control Test	0.082	00:18	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR, JOSE, F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR, JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: CUADRA

DATE/TIME OF TEST: 08/03/2011 22:07:05

---

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:04	
Control Test	0.082	22:04	
Air Blank	0.000	22:05	
Subject Sample #1		22:05	Subject Test Refused
Breath Volume	0.000	22:05	
Air Blank	0.000	22:06	
Control Test	0.082	22:06	
Air Blank	0.000	22:07	
Diagnostics Check	OK	22:07	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR, JOSE, F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR, JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CUADRA

DATE/TIME OF TEST: 08/03/2011 23:01:13

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Subject Sample #1	0.136	22:56
Breath Volume	1.402	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample #2	0.135	22:59
Breath Volume	1.492	22:59
Air Blank	0.000	23:00
Control Test	0.082	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ, CRISTINO,

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: PEREZ, CRISTINO

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 18:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: MONTERREY

DATE/TIME OF TEST: 08/05/2011 19:46:00

---

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:42	
Control Test	0.082	19:42	
Air Blank	0.000	19:43	
Subject Sample #1	0.193	19:44	
Breath Volume	1.925	19:44	
Air Blank		19:45	Purge Fail
Air Blank	0.000	19:46	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ, CRISTINO,

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: PEREZ, CRISTINO

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 18:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: MONTERREY

DATE/TIME OF TEST: 08/05/2011 19:58:05

---

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.082	19:52
Air Blank	0.000	19:52
Subject Sample #1	0.195	19:53
Breath Volume	1.964	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample #2	0.177	19:56
Breath Volume	1.734	19:56
Air Blank	0.000	19:57
Control Test	0.081	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:58

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARDIS, JOHN, D

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: DEL MAZO, STEVEN

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 00:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: VARELA

DATE/TIME OF TEST: 08/06/2011 01:32:15

---

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.082	01:30	
Air Blank	0.000	01:30	
Subject Sample #1		01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.082	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDRADE, CRISTIAN, A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: BLANCO, R'

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 03:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 08/07/2011 04:18:27

---

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.082	04:16	
Air Blank	0.000	04:16	
Subject Sample #1		04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.082	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CABRERA, ROGER, J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GOMEZ, ALEX

ARRESTING OFFICER AGENCY: Miami-Dade School PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 00:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GARCIA GONZALEZ

DATE/TIME OF TEST: 08/08/2011 00:49:43

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.083	00:47	
Air Blank	0.000	00:48	
Subject Sample #1		00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.082	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ASKEW, JOHN, O  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: RODRIGUEZ, FABIAN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/11/2011 04:32:15

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:26
Air Blank	0.000	04:26
Subject Sample #1	0.102	04:27
Breath Volume	1.601	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample #2	0.106	04:30
Breath Volume	1.816	04:30
Air Blank	0.000	04:31
Control Test	0.084	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER, K, J  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: ALEXANDER, K  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SNIPES

DATE/TIME OF TEST: 08/13/2011 02:21:58

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Subject Sample #1	0.223	02:17
Breath Volume	1.433	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample #2	0.235	02:20
Breath Volume	2.636	02:20
Air Blank	0.000	02:21
Control Test	0.082	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ, CARLOS, E

BREATH TEST OPERATOR AGENCY: Miami-Dade PD

ARRESTING OFFICER: PEREZ, CARLOS

ARRESTING OFFICER AGENCY: Miami-Dade PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 03:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DOMINGO-VELAZQUEZ

DATE/TIME OF TEST: 08/13/2011 04:17:45

---

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.082	04:12
Air Blank	0.000	04:12
Subject Sample #1	0.175	04:13
Breath Volume	1.285	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample #2	0.177	04:16
Breath Volume	1.765	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NUNES, FABIANO, P

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: NUNES, FABIANO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: GALLOP

DATE/TIME OF TEST: 08/13/2011 05:43:18

---

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.082	05:41	
Air Blank	0.000	05:41	
Subject Sample #1		05:41	Subject Test Refused
Breath Volume	0.000	05:41	
Air Blank	0.000	05:42	
Control Test	0.082	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FRANCO, GREGORY,

BREATH TEST OPERATOR AGENCY: Miami-Dade PD

ARRESTING OFFICER: FRANCO, GREGORY

ARRESTING OFFICER AGENCY: Miami-Dade PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 02:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MUNGUIA-REYES

DATE/TIME OF TEST: 08/14/2011 03:04:10

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:59
Subject Sample #1	0.167	02:59
Breath Volume	2.234	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample #2	0.164	03:02
Breath Volume	2.722	03:02
Air Blank	0.000	03:03
Control Test	0.083	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NUNES, FABIANO, P

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: NUNES, FABIANO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 03:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HASSELBACH

DATE/TIME OF TEST: 08/14/2011 03:58:02

---

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.082	03:55	
Air Blank	0.000	03:56	
Subject Sample #1		03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.082	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:58	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARKS, TIMOTHY, C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MARKS, TIMOTHY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 08:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:46

SUBJECT'S LAST NAME: GRANADOS

DATE/TIME OF TEST: 08/14/2011 10:23:37

---

Diagnostics Check	OK	10:19	
Air Blank	0.000	10:20	
Control Test	0.082	10:20	
Air Blank	0.000	10:21	
Subject Sample #1		10:22	Subject Test Refused
Breath Volume	0.000	10:22	
Air Blank	0.000	10:22	
Control Test	0.082	10:23	
Air Blank	0.000	10:23	
Diagnostics Check	OK	10:23	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CABRERA, ROGER, J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CABRERA, ROGER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 08/15/2011 00:38:46

---

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Subject Sample #1	0.175	00:34
Breath Volume	1.648	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample #2	0.190	00:37
Breath Volume	3.058	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS, WALTER, A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RODRIGUEZ, F

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 01:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DAJO

DATE/TIME OF TEST: 08/16/2011 02:30:20

---

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.082	02:19	
Air Blank	0.000	02:20	
Subject Sample #1		02:23	No Sample Provided
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample #2		02:28	No Sample Provided
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.082	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS, WALTER, A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RODRIGUEZ, F

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 01:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DAJO

DATE/TIME OF TEST: 08/16/2011 02:36:26

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.082	02:34	
Air Blank	0.000	02:34	
Subject Sample #1		02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Control Test	0.082	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FRANCO, GREGORY, R

BREATH TEST OPERATOR AGENCY: Miami-Dade PD

ARRESTING OFFICER: FRANCO, GREGORY

ARRESTING OFFICER AGENCY: Miami-Dade PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 03:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: DE-PENA

DATE/TIME OF TEST: 08/18/2011 04:17:39

---

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.082	04:09	
Air Blank	0.000	04:10	
Subject Sample #1		04:13	No Sample Provided
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample #2		04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Control Test	0.082	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER, K,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: BLANCO, R  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: PEREDA

DATE/TIME OF TEST: 08/19/2011 05:49:30

---

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:47	
Control Test	0.082	05:47	
Air Blank	0.000	05:47	
Subject Sample #1		05:48	Subject Test Refused
Breath Volume	0.000	05:48	
Air Blank	0.000	05:48	
Control Test	0.082	05:48	
Air Blank	0.000	05:49	
Diagnostics Check	OK	05:49	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALEXANDER, K,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: PAULINO, A

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: PECAR HERNANDEZ

DATE/TIME OF TEST: 08/19/2011 06:09:24

---

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.081	06:06	
Air Blank	0.000	06:06	
Subject Sample #1	0.134	06:07	
Breath Volume	1.355	06:07	
Air Blank		06:08	Purge Fail
Air Blank	0.000	06:09	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALEXANDER, K,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: PAULINO, A

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: PECAR HERNANDEZ

DATE/TIME OF TEST: 08/19/2011 06:17:42

---

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.082	06:12
Air Blank	0.000	06:12
Subject Sample #1	0.141	06:13
Breath Volume	3.011	06:13
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample #2	0.138	06:16
Breath Volume	2.269	06:16
Air Blank	0.000	06:16
Control Test	0.082	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SERRA, UBALDO, F

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LOCKE, JEFF

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 00:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 08/20/2011 01:10:38

---

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:08	
Control Test	0.082	01:08	
Air Blank	0.000	01:09	
Subject Sample #1		01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.082	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MANSOUR, AZEEZ, F

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LEWIS, W

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 01:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MENDOZA MORALES

DATE/TIME OF TEST: 08/20/2011 02:23:34

---

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:22	
Subject Sample #1		02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.082	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARZOLA, PABLO,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: CABRERA, R  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/20/2011 04:12:43

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Subject Sample #1	0.124	04:08
Breath Volume	1.324	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.136	04:11
Breath Volume	2.484	04:11
Air Blank	0.000	04:11
Control Test	0.083	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASTILLO, WILSON, D

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MARKOWITZ, ANDREW

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 06:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/20/2011 07:31:20

---

Diagnostics Check	OK	07:27	
Air Blank	0.000	07:28	
Control Test	0.082	07:28	
Air Blank	0.000	07:28	
Subject Sample #1	0.169	07:29	
Breath Volume	2.542	07:29	
Air Blank		07:30	Purge Fail
Air Blank	0.000	07:31	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARKOWITZ, ANDREW, C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARTAGENA, Y

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 06:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/20/2011 07:40:13

---

Diagnostics Check	OK	07:33
Air Blank	0.000	07:34
Control Test	0.081	07:34
Air Blank	0.000	07:35
Subject Sample #1	0.173	07:35
Breath Volume	1.910	07:35
Air Blank	0.000	07:36
Air Blank	0.000	07:37
Subject Sample #2	0.160	07:38
Breath Volume	1.332	07:38
Air Blank	0.000	07:39
Control Test	0.081	07:39
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CABRERA, ROGER, J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CABRERA, ROGER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 22:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 08/20/2011 23:02:37

---

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.082	23:00	
Air Blank	0.000	23:00	
Subject Sample #1		23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	
Control Test	0.082	23:02	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SERRA, UBALDO, F

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ESPINOSA, FRANK

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/20/2011 23:55:45

---

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.082	23:50
Air Blank	0.000	23:50
Subject Sample #1	0.175	23:51
Breath Volume	1.320	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample #2	0.181	23:54
Breath Volume	1.304	23:54
Air Blank	0.000	23:54
Control Test	0.083	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MANSOUR, AZEEZ, F

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MENDEZ, H

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 01:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PALOMINO VIDAL

DATE/TIME OF TEST: 08/22/2011 01:33:14

---

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.081	01:31	
Air Blank	0.000	01:31	
Subject Sample #1		01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.082	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SERRA, UBALDO, F

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SERRA, UBALDO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 08/22/2011 05:39:14

---

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:27	
Control Test	0.082	05:28	
Air Blank	0.000	05:28	
Subject Sample #1	0.188	05:31	Volume Not Met
Breath Volume	0.183	05:31	
Air Blank	0.000	05:32	
Air Blank	0.000	05:34	
Subject Sample #2	0.188	05:37	Volume Not Met
Breath Volume	0.500	05:37	
Air Blank	0.000	05:38	
Control Test	0.081	05:38	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SERRA, UBALDO, F

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SERRA, UBALDO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 08/22/2011 07:08:03

---

Diagnostics Check	OK	07:05	
Air Blank	0.000	07:05	
Control Test	0.083	07:06	
Air Blank	0.000	07:06	
Subject Sample #1		07:06	Subject Test Refused
Breath Volume	0.000	07:06	
Air Blank	0.000	07:07	
Control Test	0.082	07:07	
Air Blank	0.000	07:07	
Diagnostics Check	OK	07:08	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLANC, CAROLE,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: RAMOS, VICTOR  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/24/2011 01:16:51

---

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.082	01:14	
Air Blank	0.000	01:15	
Subject Sample #1		01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.082	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREGOIRE, A,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GREGOIRE, A  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 08/25/2011 01:43:33

---

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.081	01:41	
Air Blank	0.000	01:41	
Subject Sample #1		01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

---

Data Download Date/Time 08/25/2011 13:09

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARZOLA, PABLO,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: ARZOLA, PABLO  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: NAVARRETE

DATE/TIME OF TEST: 08/25/2011 20:49:59

---

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.083	20:44
Air Blank	0.000	20:44
Subject Sample #1	0.033	20:45
Breath Volume	2.269	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:47
Subject Sample #2	0.033	20:48
Breath Volume	2.359	20:48
Air Blank	0.000	20:49
Control Test	0.082	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NUNES, FABIANO, P

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 03:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/26/2011 04:05:01

---

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.082	03:54	
Air Blank	0.000	03:54	
Subject Sample #1		03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Air Blank	0.000	04:00	
Subject Sample #2		04:03	No Sample Provided
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:05	

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REYES, BARBARA, T

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ACOSTA, R

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 09:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/26/2011 10:42:24

---

Diagnostics Check	OK	10:33	
Air Blank	0.000	10:34	
Control Test	0.082	10:34	
Air Blank	0.000	10:34	
Subject Sample #1		10:38	No Sample Provided
Breath Volume	0.000	10:38	
Air Blank	0.000	10:38	
Air Blank	0.000	10:40	
Subject Sample #2		10:41	Subject Test Refused
Breath Volume	0.000	10:41	
Air Blank	0.000	10:41	
Control Test	0.081	10:41	
Air Blank	0.000	10:42	
Diagnostics Check	OK	10:42	

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CABRERA, ROGER, J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CABRERA, ROGER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BARON DEL TORO

DATE/TIME OF TEST: 08/27/2011 00:02:06

---

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.081	00:00	
Air Blank	0.000	00:00	
Subject Sample #1		00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.081	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:02	

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VITAL, CHRISTOPHER, D

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: LOCKE, JEFFREY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 05:02

GENDER: Female

OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 08/27/2011 05:37:58

---

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:34	
Control Test	0.082	05:34	
Air Blank	0.000	05:34	
Subject Sample #1	0.108	05:36	
Breath Volume	1.300	05:36	
Air Blank		05:37	Purge Fail
Air Blank	0.000	05:37	

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARKOWITZ, ANDREW, C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MARKOWITZ, ANDREW

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 08:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/27/2011 09:40:10

---

Diagnostics Check	OK	09:37	
Air Blank	0.000	09:37	
Control Test	0.081	09:37	
Air Blank	0.000	09:38	
Subject Sample #1		09:38	Subject Test Refused
Breath Volume	0.000	09:38	
Air Blank	0.000	09:39	
Control Test	0.082	09:39	
Air Blank	0.000	09:40	
Diagnostics Check	OK	09:40	

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BENT, JONATHAN, L

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: BENT, JONATHAN

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 20:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: MELCHOR

DATE/TIME OF TEST: 08/27/2011 22:04:35

---

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.082	22:01	
Air Blank	0.000	22:01	
Subject Sample #1		22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:03	
Control Test	0.082	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ, CARLOS, E

BREATH TEST OPERATOR AGENCY: Miami-Dade PD

ARRESTING OFFICER: PEREZ, CARLOS

ARRESTING OFFICER AGENCY: Miami-Dade PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 00:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 08/28/2011 01:03:07

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.082	00:57
Air Blank	0.000	00:57
Subject Sample #1	0.160	00:58
Breath Volume	1.781	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:00
Subject Sample #2	0.160	01:01
Breath Volume	1.781	01:01
Air Blank	0.000	01:02
Control Test	0.082	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SERRA, UBALDO, F

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SERRA, UBALDO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 03:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MCDOUGALL

DATE/TIME OF TEST: 08/28/2011 03:51:35

---

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:49	
Control Test	0.082	03:49	
Air Blank	0.000	03:49	
Subject Sample #1		03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.082	03:51	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRERA, ROGER, J  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: YANEZ, RANDY  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: PUMO

DATE/TIME OF TEST: 08/28/2011 22:32:34

---

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.081	22:30	
Air Blank	0.000	22:31	
Subject Sample #1		22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.081	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

---

Data Download Date/Time 09/27/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN, SHANETTA, L

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GREEN, SHANETTA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 18:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 08/06/2011 19:40:20

---

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.079	19:34
Air Blank	0.000	19:34
Subject Sample #1	0.244	19:35
Breath Volume	1.703	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample #2	0.238	19:38
Breath Volume	2.128	19:38
Air Blank	0.000	19:39
Control Test	0.080	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

---

Data Download Date/Time 08/25/2011 13:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROJAS-LOPEZ, ALEXANDER,

BREATH TEST OPERATOR AGENCY: EI Portal PD

ARRESTING OFFICER: CASAIS, MANUEL

ARRESTING OFFICER AGENCY: EI Portal PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 19:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ROJAS-LOPEZ

DATE/TIME OF TEST: 08/06/2011 20:21:03

---

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.079	20:15
Air Blank	0.000	20:16
Subject Sample #1	0.158	20:16
Breath Volume	3.230	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample #2	0.163	20:19
Breath Volume	4.132	20:19
Air Blank	0.000	20:20
Control Test	0.079	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:21

---

Data Download Date/Time 08/25/2011 13:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRADLEY, ROSCOE, C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: JACKSON, ROGER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 08/11/2011 03:01:51

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:57	
Subject Sample #1	0.185	03:00	
Breath Volume	1.351	03:00	
Air Blank		03:01	Purge Fail
Air Blank		03:01	Purge Fail

---

Data Download Date/Time 08/25/2011 13:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRADLEY, ROSCOE, C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: JACKSON, ROGER

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 08/11/2011 03:17:14

---

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.077	03:11
Air Blank	0.000	03:11
Subject Sample #1	0.179	03:12
Breath Volume	1.113	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample #2	0.177	03:15
Breath Volume	1.183	03:15
Air Blank	0.000	03:16
Control Test	0.076	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time 08/25/2011 13:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS II, ALLIXEN, L

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HALL, S

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 19:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: BRAYBOY

DATE/TIME OF TEST: 08/11/2011 20:07:20

---

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:04	
Control Test	0.081	20:05	
Air Blank	0.000	20:05	
Subject Sample #1		20:05	Subject Test Refused
Breath Volume	0.000	20:05	
Air Blank	0.000	20:06	
Control Test	0.080	20:06	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

---

Data Download Date/Time 08/25/2011 13:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIVERA, RAFAEL, A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RIVERA, RAFAEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 13:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:51

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 08/16/2011 14:22:05

---

Diagnostics Check	OK	14:15
Air Blank	0.000	14:15
Control Test	0.080	14:15
Air Blank	0.000	14:16
Subject Sample #1	0.203	14:17
Breath Volume	1.968	14:17
Air Blank	0.000	14:17
Air Blank	0.000	14:19
Subject Sample #2	0.212	14:20
Breath Volume	1.917	14:20
Air Blank	0.000	14:21
Control Test	0.079	14:21
Air Blank	0.000	14:21
Diagnostics Check	OK	14:22

---

Data Download Date/Time 08/25/2011 13:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRADLEY, ROSCOE, C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: BRADLEY, ROSCOE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 01:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 08/20/2011 02:30:08

---

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:28	
Air Blank	0.000	02:28	
Subject Sample #1		02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

---

Data Download Date/Time 08/25/2011 13:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRADLEY, ROSCOE, C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: BRADLEY, ROSCOE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/14/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 22:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: POSAS

DATE/TIME OF TEST: 08/24/2011 23:52:53

---

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.079	23:50	
Air Blank	0.000	23:51	
Subject Sample #1		23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.080	23:52	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

---

Data Download Date/Time 08/25/2011 13:12

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRADLEY, ROSCOE, C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: PHILIPPE, JEAN

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 03:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 08/26/2011 04:22:28

---

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:16
Subject Sample #1	0.136	04:17
Breath Volume	1.578	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.133	04:20
Breath Volume	1.722	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time 09/27/2011 12:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEVENS II, ALLIXEN, L

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: STEVENS II, ALLIXEN

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG002801

DRY GAS EXPIRY DATE: 01/28/2012

VIOLATION CODE: DUI

ARREST TIME: 15:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/26/2011 16:09:05

---

Diagnostics Check	OK	16:05	
Air Blank	0.000	16:06	
Control Test	0.079	16:06	
Air Blank	0.000	16:06	
Subject Sample #1		16:07	Subject Test Refused
Breath Volume	0.000	16:07	
Air Blank	0.000	16:08	
Control Test	0.080	16:08	
Air Blank	0.000	16:09	
Diagnostics Check	OK	16:09	

---

Data Download Date/Time 09/27/2011 12:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHEETZ, JOHN, W  
BREATH TEST OPERATOR AGENCY: Calhoun County SO  
ARRESTING OFFICER: DALTON, WILLIAM  
ARRESTING OFFICER AGENCY: Calhoun County SO  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MEDLEY

DATE/TIME OF TEST: 08/12/2011 21:29:35

---

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.077	21:23
Air Blank	0.000	21:24
Subject Sample #1	0.106	21:24
Breath Volume	2.773	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample #2	0.105	21:28
Breath Volume	2.769	21:28
Air Blank	0.000	21:28
Control Test	0.077	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

---

Data Download Date/Time 08/31/2011 08:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORRIS, ANTHONY, S  
BREATH TEST OPERATOR AGENCY: Calhoun County SO  
ARRESTING OFFICER: NORRIS, ANTHONY  
ARRESTING OFFICER AGENCY: Calhoun County SO  
LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 685214  
DRY GAS EXPIRY DATE: 10/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CULPAN

DATE/TIME OF TEST: 08/26/2011 03:22:15

---

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.078	03:11	
Air Blank	0.000	03:11	
Subject Sample #1		03:15	No Sample Provided
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample #2	0.114	03:20	Volume Not Met
Breath Volume	0.453	03:20	
Air Blank	0.000	03:21	
Control Test	0.078	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

---

Data Download Date/Time 08/31/2011 08:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, L,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: ROMERO, F

ARRESTING OFFICER AGENCY: Medley PD

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 11:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 08/03/2011 17:41:49

---

Diagnostics Check	OK	17:35
Air Blank	0.000	17:36
Control Test	0.078	17:36
Air Blank	0.000	17:36
Subject Sample #1	0.000	17:37
Breath Volume	3.550	17:37
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample #2	0.000	17:40
Breath Volume	4.007	17:40
Air Blank	0.000	17:40
Control Test	0.077	17:41
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

---

Data Download Date/Time 08/25/2011 13:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARCIA, J,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: GARCIA, J

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 03:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CUELLAR VILLAFANA

DATE/TIME OF TEST: 08/12/2011 04:37:05

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.077	04:31
Air Blank	0.000	04:32
Subject Sample #1	0.255	04:32
Breath Volume	2.535	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample #2	0.251	04:35
Breath Volume	2.304	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:37

---

Data Download Date/Time 08/25/2011 13:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CONSALVO, ESTEBAN

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 05:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 08/18/2011 06:17:54

---

Diagnostics Check	OK	06:14	
Air Blank	0.000	06:15	
Control Test	0.079	06:15	
Air Blank	0.000	06:16	
Subject Sample #1		06:16	Subject Test Refused
Breath Volume	0.000	06:16	
Air Blank	0.000	06:16	
Control Test	0.078	06:17	
Air Blank	0.000	06:17	
Diagnostics Check	OK	06:17	

---

Data Download Date/Time 08/25/2011 13:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: RAMIREZ, SERGIO

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 01:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/20/2011 02:42:25

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.229	02:37
Breath Volume	2.164	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.236	02:40
Breath Volume	2.515	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time 08/25/2011 13:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: MONIER, JOSUE

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 704221

DRY GAS EXPIRY DATE: 12/11/2011

VIOLATION CODE: DUI

ARREST TIME: 03:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/27/2011 16:11:36

---

Diagnostics Check	OK	16:05
Air Blank	0.000	16:05
Control Test	0.079	16:06
Air Blank	0.000	16:06
Subject Sample #1	0.229	16:07
Breath Volume	1.500	16:07
Air Blank	0.000	16:07
Air Blank	0.000	16:09
Subject Sample #2	0.216	16:09
Breath Volume	1.410	16:09
Air Blank	0.000	16:10
Control Test	0.078	16:10
Air Blank	0.000	16:11
Diagnostics Check	OK	16:11

---

Data Download Date/Time 09/27/2011 12:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: OLVERA

DATE/TIME OF TEST: 08/02/2011 02:44:30

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample #1	0.162	02:40
Breath Volume	2.558	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.160	02:43
Breath Volume	2.445	02:43
Air Blank	0.000	02:43
Control Test	0.077	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: OLGUIN-BENITEZ

DATE/TIME OF TEST: 08/04/2011 01:45:12

---

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.078	01:43	
Air Blank	0.000	01:43	
Subject Sample #1		01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.078	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: KAVANAUGH, JAMES  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 08/04/2011 05:17:04

---

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:11
Subject Sample #1	0.249	05:12
Breath Volume	3.351	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample #2	0.248	05:15
Breath Volume	2.777	05:15
Air Blank	0.000	05:16
Control Test	0.077	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:17

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GARCIA-GARCIA

DATE/TIME OF TEST: 08/05/2011 01:51:08

---

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.078	01:49	
Air Blank	0.000	01:49	
Subject Sample #1		01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.078	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HURTADO, KENNETH,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: AGOSTO, VICTOR  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: FUNEUS

DATE/TIME OF TEST: 08/07/2011 23:24:26

---

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Subject Sample #1	0.211	23:19
Breath Volume	1.332	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample #2	0.226	23:22
Breath Volume	1.609	23:22
Air Blank	0.000	23:23
Control Test	0.077	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 08/08/2011 02:07:35

---

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.078	02:05	
Air Blank	0.000	02:06	
Subject Sample #1		02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.078	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURGUIDO, ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: MURGUIDO, ALEJANDRO  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/10/2011 21:52:18

---

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.079	21:47	
Air Blank	0.000	21:47	
Subject Sample #1	0.038	21:48	
Breath Volume	2.746	21:48	
Air Blank	0.000	21:48	
Air Blank	0.000	21:50	
Subject Sample #2		21:50	Subject Test Refused
Breath Volume	0.000	21:50	
Air Blank	0.000	21:51	
Control Test	0.077	21:51	
Air Blank	0.000	21:52	
Diagnostics Check	OK	21:52	

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LEMUS CUELLAR

DATE/TIME OF TEST: 08/12/2011 03:18:40

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:12
Air Blank	0.000	03:12
Subject Sample #1	0.098	03:14
Breath Volume	1.500	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample #2	0.094	03:17
Breath Volume	2.324	03:17
Air Blank	0.000	03:17
Control Test	0.077	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: ROJO, M  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 08/12/2011 03:46:59

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:41
Subject Sample #1	0.093	03:42
Breath Volume	3.695	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample #2	0.091	03:45
Breath Volume	3.625	03:45
Air Blank	0.000	03:46
Control Test	0.077	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLANCO, VERONICA, M  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: AGOSTO, VICTOR  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 08/13/2011 21:38:44

---

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Subject Sample #1	0.277	21:31
Breath Volume	1.503	21:31
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample #2	0.302	21:34
Breath Volume	2.847	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample #3	0.298	21:37
Breath Volume	2.042	21:37
Air Blank	0.000	21:37
Control Test	0.078	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLANCO, VERONICA, M  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: KENNEDY, BRIAN  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 17:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MATEO TORRES

DATE/TIME OF TEST: 08/14/2011 18:22:42

---

Diagnostics Check	OK	18:15
Air Blank	0.000	18:15
Control Test	0.077	18:16
Air Blank	0.000	18:16
Subject Sample #1	0.194	18:18
Breath Volume	1.203	18:18
Air Blank	0.000	18:18
Air Blank	0.000	18:20
Subject Sample #2	0.198	18:21
Breath Volume	1.511	18:21
Air Blank	0.000	18:21
Control Test	0.078	18:22
Air Blank	0.000	18:22
Diagnostics Check	OK	18:22

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GONZALEZ-PEREZ

DATE/TIME OF TEST: 08/15/2011 01:36:07

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Subject Sample #1	0.270	01:31
Breath Volume	2.398	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample #2	0.266	01:34
Breath Volume	1.941	01:34
Air Blank	0.000	01:35
Control Test	0.077	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 08/16/2011 01:35:58

---

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.078	01:33	
Air Blank	0.000	01:34	
Subject Sample #1		01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.079	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DALMAU

DATE/TIME OF TEST: 08/17/2011 02:13:51

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:12	
Subject Sample #1		02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 08/19/2011 02:32:57

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Subject Sample #1	0.125	02:28
Breath Volume	1.320	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample #2	0.122	02:31
Breath Volume	2.910	02:31
Air Blank	0.000	02:32
Control Test	0.076	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUBBARD, KENNETH, R  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: HUBBARD, KENNETH  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: TURCIOS

DATE/TIME OF TEST: 08/20/2011 01:45:18

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample #1	0.181	01:39
Breath Volume	1.730	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample #2	0.191	01:43
Breath Volume	1.230	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 08/22/2011 01:38:26

---

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.078	01:36	
Air Blank	0.000	01:36	
Subject Sample #1		01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GARCIA-PEREZ

DATE/TIME OF TEST: 08/24/2011 02:43:52

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample #1	0.126	02:39
Breath Volume	3.519	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample #2	0.126	02:42
Breath Volume	3.355	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: OWENS, RANDALL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: JIMENEZ-GOMEZ

DATE/TIME OF TEST: 08/24/2011 03:12:59

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.078	03:07
Air Blank	0.000	03:07
Subject Sample #1	0.162	03:08
Breath Volume	2.792	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample #2	0.164	03:11
Breath Volume	3.218	03:11
Air Blank	0.000	03:12
Control Test	0.075	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOSA, GABRIEL, J  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SOSA, GABRIEL  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: LOPEZ DIAZ

DATE/TIME OF TEST: 08/25/2011 02:19:26

---

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.078	02:17	
Air Blank	0.000	02:17	
Subject Sample #1		02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.078	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUBBARD, KENNETH, R  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: HUBBARD, KENNETH  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: ESPINEL

DATE/TIME OF TEST: 08/27/2011 19:31:32

---

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Subject Sample #1	0.230	19:26
Breath Volume	2.921	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample #2	0.219	19:29
Breath Volume	4.484	19:29
Air Blank	0.000	19:30
Control Test	0.078	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SURMAN, THOMAS, A  
BREATH TEST OPERATOR AGENCY: Homestead PD  
ARRESTING OFFICER: SURMAN, THOMAS  
ARRESTING OFFICER AGENCY: Homestead PD  
LAST AGENCY INSPECTION: 07/23/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770941  
DRY GAS EXPIRY DATE: 05/27/2012  
VIOLATION CODE: DUI  
ARREST TIME: 11:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/29/2011 11:46:48

---

Diagnostics Check	OK	11:40
Air Blank	0.000	11:40
Control Test	0.078	11:40
Air Blank	0.000	11:41
Subject Sample #1	0.271	11:42
Breath Volume	1.714	11:42
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample #2	0.264	11:45
Breath Volume	1.589	11:45
Air Blank	0.000	11:45
Control Test	0.078	11:46
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

---

Data Download Date/Time 08/30/2011 15:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ, JOEL,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DOMINGUEZ, E  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 08/13/2011 04:02:28

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:55
Subject Sample #1	0.168	03:56
Breath Volume	1.179	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample #2	0.159	04:00
Breath Volume	1.277	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time 08/30/2011 15:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAZONOV, VLADISLAV,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SAZONOV, VLADISLAV  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: AMIONE

DATE/TIME OF TEST: 08/14/2011 06:30:00

---

Diagnostics Check	OK	06:21	
Air Blank	0.000	06:22	
Control Test	0.079	06:22	
Air Blank	0.000	06:23	
Subject Sample #1	0.127	06:25	
Breath Volume	1.183	06:25	
Air Blank	0.000	06:26	
Air Blank	0.000	06:28	
Subject Sample #2		06:29	RFI Detect
Breath Volume	0.062	06:29	
Air Blank		06:29	RFI Detect
Air Blank		06:30	RFI Detect

---

Data Download Date/Time 08/30/2011 15:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAZONOV, VLADISLAV,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SAZONOV, VLADISLAV  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: AMIONE

DATE/TIME OF TEST: 08/14/2011 06:45:51

---

Diagnostics Check	OK	06:36	
Air Blank	0.000	06:36	
Control Test	0.078	06:36	
Air Blank	0.000	06:37	
Subject Sample #1	0.123	06:40	
Breath Volume	1.601	06:40	
Air Blank	0.000	06:40	
Air Blank	0.000	06:42	
Subject Sample #2	0.110	06:43	
Breath Volume	1.769	06:43	
Air Blank	0.000	06:44	
Control Test	0.078	06:44	
Air Blank	0.000	06:45	
Diagnostics Check	Fail	06:45	Voltage/Current Test
Air Blank	0.000	06:45	

---

Data Download Date/Time 08/30/2011 15:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAZONOV, VLADISLAV,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SAZONOV, VLADISLAV  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: AMIONE

DATE/TIME OF TEST: 08/14/2011 06:57:06

---

Diagnostics Check	OK	06:48	
Air Blank	0.000	06:48	
Control Test	0.078	06:49	
Air Blank	0.000	06:49	
Subject Sample #1	0.106	06:50	
Breath Volume	1.410	06:50	
Air Blank	0.000	06:51	
Air Blank	0.000	06:53	
Subject Sample #2	0.111	06:55	
Breath Volume	2.144	06:55	
Air Blank	0.000	06:55	
Control Test	0.077	06:56	
Air Blank	0.000	06:56	
Diagnostics Check	Fail	06:56	Voltage/Current Test
Air Blank	0.000	06:57	

---

Data Download Date/Time 08/30/2011 15:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ, ERIC,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: ROJO, MARIO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 08/24/2011 04:40:23

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:31
Subject Sample #1	0.122	04:34
Breath Volume	1.253	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample #2	0.111	04:38
Breath Volume	1.253	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

---

Data Download Date/Time 08/30/2011 15:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOMINGUEZ, ERIC,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SOCARRAS, PEDRO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CAMIS

DATE/TIME OF TEST: 08/25/2011 01:59:41

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.078	01:54
Air Blank	0.000	01:54
Subject Sample #1	0.232	01:55
Breath Volume	2.445	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample #2	0.233	01:58
Breath Volume	1.804	01:58
Air Blank	0.000	01:58
Control Test	0.078	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

---

Data Download Date/Time 08/30/2011 15:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WIRTH, PHILIP, A  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: WIRTH, P  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/25/2011 05:46:05

---

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.078	05:39
Air Blank	0.000	05:39
Subject Sample #1	0.206	05:40
Breath Volume	1.687	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:42
Subject Sample #2	0.192	05:44
Breath Volume	1.449	05:44
Air Blank	0.000	05:45
Control Test	0.078	05:45
Air Blank	0.000	05:45
Diagnostics Check	OK	05:46

---

Data Download Date/Time 08/30/2011 15:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MESTRE, ENRIQUE,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: VILLA, LORENCE  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 07:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: RENIEY

DATE/TIME OF TEST: 08/28/2011 07:59:16

---

Diagnostics Check	OK	07:56	
Air Blank	0.000	07:56	
Control Test	0.078	07:57	
Air Blank	0.000	07:57	
Subject Sample #1		07:57	Subject Test Refused
Breath Volume	0.000	07:57	
Air Blank	0.000	07:58	
Control Test	0.078	07:58	
Air Blank	0.000	07:59	
Diagnostics Check	OK	07:59	

---

Data Download Date/Time 08/30/2011 15:16

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUMPHRY, KERRY, R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: HUMPHRY, KERRY  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/30/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 750981  
DRY GAS EXPIRY DATE: 04/09/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/31/2011 02:58:46

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:57	
Subject Sample #1		02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.078	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

---

Data Download Date/Time 09/22/2011 15:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LARROSA, CLAUDIA, V  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: LOCKE, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RUB

DATE/TIME OF TEST: 08/03/2011 04:03:57

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:57
Subject Sample #1	0.211	03:59
Breath Volume	1.179	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample #2	0.222	04:02
Breath Volume	1.445	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time 08/17/2011 11:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCGRATH, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MCGRATH, ROBERT  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: RODRIGUEZ-DELRAY

DATE/TIME OF TEST: 08/05/2011 02:50:57

---

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	
Air Blank	0.000	02:38	
Subject Sample #1	0.144	02:41	Slope Not Level
Breath Volume	1.589	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:43	
Subject Sample #2	0.146	02:46	
Breath Volume	1.375	02:46	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample #3	0.158	02:49	
Breath Volume	1.589	02:49	
Air Blank	0.000	02:50	
Control Test	0.077	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

---

Data Download Date/Time 08/17/2011 11:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUNCAS, ANDREW, J  
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD  
ARRESTING OFFICER: KUNCAS, ANDREW  
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 08/12/2011 04:43:58

---

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.079	04:41	
Air Blank	0.000	04:42	
Subject Sample #1		04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:43	
Control Test	0.079	04:43	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

---

Data Download Date/Time 08/17/2011 11:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARIABELLO, MARCELINO,  
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD  
ARRESTING OFFICER: MARIABELLO, MARCELINO  
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 08/12/2011 21:48:24

---

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.078	21:46	
Air Blank	0.000	21:46	
Subject Sample #1		21:47	Subject Test Refused
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Control Test	0.078	21:47	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

---

Data Download Date/Time 08/17/2011 11:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWANDOWSKA, SABINA, M  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: LEWANDOWSKA, SABINA  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: S561010910620

DATE/TIME OF TEST: 08/14/2011 06:17:40

---

Diagnostics Check	OK	06:14	
Air Blank	0.000	06:15	
Control Test	0.078	06:15	
Air Blank	0.000	06:16	
Subject Sample #1		06:16	Subject Test Refused
Breath Volume	0.000	06:16	
Air Blank	0.000	06:16	
Control Test	0.079	06:17	
Air Blank	0.000	06:17	
Diagnostics Check	OK	06:17	

---

Data Download Date/Time 08/17/2011 11:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIETO, ANDRES, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PRIETO, ANDRES  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: STROZIER

DATE/TIME OF TEST: 08/14/2011 19:41:50

---

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.077	19:36
Air Blank	0.000	19:36
Subject Sample #1	0.275	19:37
Breath Volume	2.933	19:37
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample #2	0.256	19:40
Breath Volume	1.328	19:40
Air Blank	0.000	19:40
Control Test	0.077	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

---

Data Download Date/Time 08/17/2011 11:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ, JOHN-PAUL,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: WELLS, NICOLE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 08/16/2011 04:06:11

---

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample #1	0.173	04:01
Breath Volume	2.109	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample #2	0.175	04:04
Breath Volume	1.914	04:04
Air Blank	0.000	04:05
Control Test	0.077	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

---

Data Download Date/Time 08/17/2011 11:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON, ORLARRY, S  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: LIMA, PABLO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TUMBARELLO

DATE/TIME OF TEST: 08/17/2011 04:00:10

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample #1	0.207	03:55
Breath Volume	2.832	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:57
Subject Sample #2	0.192	03:58
Breath Volume	1.800	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time 08/17/2011 11:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOEL, LINDSLEY, J  
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD  
ARRESTING OFFICER: NOEL, LINDSLEY  
ARRESTING OFFICER AGENCY: Bay Harbor Island PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 08/20/2011 01:49:20

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample #1	0.180	01:42
Breath Volume	1.578	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample #2	0.181	01:47
Breath Volume	2.324	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KUNCAS, ANDREW, J

BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD

ARRESTING OFFICER: KUNCAS, ANDREW

ARRESTING OFFICER AGENCY: Bal Harbour Vil PD

LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 02:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BRANCOCCIO

DATE/TIME OF TEST: 08/20/2011 03:18:40

---

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:17	
Subject Sample #1		03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.078	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH, FELICIA, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH, FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: CAPELES

DATE/TIME OF TEST: 08/20/2011 04:56:42

---

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.078	04:50
Air Blank	0.000	04:51
Subject Sample #1	0.117	04:52
Breath Volume	1.750	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample #2	0.109	04:55
Breath Volume	1.390	04:55
Air Blank	0.000	04:55
Control Test	0.078	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELLS, NICOLE, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GOMEZ, PETE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/20/2011 05:10:52

---

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:08	
Control Test	0.078	05:08	
Air Blank	0.000	05:09	
Subject Sample #1		05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Control Test	0.079	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELLS, NICOLE, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MYRTIL, R  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: BRUNET

DATE/TIME OF TEST: 08/21/2011 04:27:32

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.078	04:19	
Air Blank	0.000	04:20	
Subject Sample #1	0.117	04:23	Volume Not Met
Breath Volume	0.777	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:25	
Subject Sample #2		04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.078	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITE, JASON, B  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: WAITE, JASON  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GALLOZA

DATE/TIME OF TEST: 08/23/2011 22:13:20

---

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.079	22:07
Air Blank	0.000	22:08
Subject Sample #1	0.205	22:08
Breath Volume	1.769	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample #2	0.212	22:11
Breath Volume	2.355	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASCLE, KEVIN, M  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: BASCLE, KEVIN  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: REY

DATE/TIME OF TEST: 08/26/2011 19:25:41

---

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.078	19:17
Air Blank	0.000	19:18
Subject Sample #1	0.191	19:21
Breath Volume	3.164	19:21
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample #2	0.200	19:24
Breath Volume	3.058	19:24
Air Blank	0.000	19:24
Control Test	0.078	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROIG, DAWN, M  
BREATH TEST OPERATOR AGENCY: Miami Shores PD  
ARRESTING OFFICER: ROIG, DAWN  
ARRESTING OFFICER AGENCY: Miami Shores PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: GARCON

DATE/TIME OF TEST: 08/27/2011 04:48:15

---

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.077	04:42
Air Blank	0.000	04:42
Subject Sample #1	0.170	04:43
Breath Volume	1.765	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample #2	0.172	04:46
Breath Volume	1.289	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROIG, DAWN, M  
BREATH TEST OPERATOR AGENCY: Miami Shores PD  
ARRESTING OFFICER: ROIG, DAWN  
ARRESTING OFFICER AGENCY: Miami Shores PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 08/28/2011 03:57:13

---

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.078	03:46	
Air Blank	0.000	03:46	
Subject Sample #1		03:50	No Sample Provided
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Air Blank	0.000	03:52	
Subject Sample #2	0.204	03:55	Volume Not Met
Breath Volume	0.082	03:55	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROIG, DAWN, M  
BREATH TEST OPERATOR AGENCY: Miami Shores PD  
ARRESTING OFFICER: ROIG, DAWN  
ARRESTING OFFICER AGENCY: Miami Shores PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 08/28/2011 04:14:38

---

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Subject Sample #1		04:03	No Sample Provided
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Air Blank	0.000	04:06	
Subject Sample #2	0.204	04:07	
Breath Volume	1.824	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:09	
Subject Sample #3	0.166	04:12	Volume Not Met
Breath Volume	0.691	04:12	
Air Blank	0.000	04:13	
Control Test	0.077	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROIG, DAWN, M  
BREATH TEST OPERATOR AGENCY: Miami Shores PD  
ARRESTING OFFICER: ROIG, DAWN  
ARRESTING OFFICER AGENCY: Miami Shores PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 08/28/2011 04:50:47

---

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.076	04:43
Air Blank	0.000	04:43
Subject Sample #1	0.193	04:44
Breath Volume	1.203	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample #2	0.187	04:49
Breath Volume	2.136	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COIMIN, JEAN, W  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ANGELINI, LIAM  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/28/2011 05:11:31

---

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.077	05:05
Air Blank	0.000	05:06
Subject Sample #1	0.079	05:06
Breath Volume	1.347	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample #2	0.076	05:09
Breath Volume	1.105	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL, REGINALD, J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ANGELINI, LIAM  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 08/28/2011 05:25:31

---

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.078	05:18
Air Blank	0.000	05:18
Subject Sample #1	0.105	05:21
Breath Volume	1.878	05:21
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample #2	0.103	05:24
Breath Volume	1.402	05:24
Air Blank	0.000	05:24
Control Test	0.077	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

---

Data Download Date/Time 09/06/2011 12:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS, STEVEN, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ADAMS, STEVEN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: LOOMER

DATE/TIME OF TEST: 08/12/2011 05:05:53

---

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.081	05:03	
Air Blank	0.000	05:04	
Subject Sample #1		05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

---

Data Download Date/Time 08/17/2011 12:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: BERNAL

DATE/TIME OF TEST: 08/13/2011 06:12:46

---

Diagnostics Check	OK	06:09	
Air Blank	0.000	06:10	
Control Test	0.081	06:10	
Air Blank	0.000	06:11	
Subject Sample #1		06:11	Subject Test Refused
Breath Volume	0.000	06:11	
Air Blank	0.000	06:11	
Control Test	0.081	06:12	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

---

Data Download Date/Time 08/17/2011 12:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITE, JASON, B  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: WAITE, JASON  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 07:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: MONTAS-NOLASCO

DATE/TIME OF TEST: 08/13/2011 07:37:19

---

Diagnostics Check	OK	07:35	
Air Blank	0.000	07:35	
Control Test	0.079	07:36	
Air Blank	0.000	07:36	
Subject Sample #1		07:37	RFI Detect
Breath Volume	0.000	07:37	
Air Blank		07:37	RFI Detect
Air Blank		07:37	RFI Detect

---

Data Download Date/Time 08/17/2011 12:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITE, JASON, B  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: WAITE, JASON  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 07:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: MONTAS NOLASCO

DATE/TIME OF TEST: 08/13/2011 07:41:36

---

Diagnostics Check	OK	07:38	
Air Blank	0.000	07:39	
Control Test	0.079	07:39	
Air Blank	0.000	07:40	
Subject Sample #1		07:40	Subject Test Refused
Breath Volume	0.000	07:40	
Air Blank	0.000	07:40	
Control Test	0.080	07:41	
Air Blank	0.000	07:41	
Diagnostics Check	OK	07:41	

---

Data Download Date/Time 08/17/2011 12:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SABOGAL

DATE/TIME OF TEST: 08/28/2011 02:59:51

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.082	02:53
Air Blank	0.000	02:54
Subject Sample #1	0.093	02:54
Breath Volume	2.437	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.092	02:58
Breath Volume	2.390	02:58
Air Blank	0.000	02:58
Control Test	0.083	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

---

Data Download Date/Time 09/06/2011 12:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS, STEVEN, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ADAMS, STEVEN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 08/31/2011 06:41:44

---

Diagnostics Check	OK	06:35
Air Blank	0.000	06:35
Control Test	0.080	06:35
Air Blank	0.000	06:36
Subject Sample #1	0.142	06:36
Breath Volume	2.953	06:36
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample #2	0.140	06:40
Breath Volume	3.675	06:40
Air Blank	0.000	06:40
Control Test	0.081	06:41
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

---

Data Download Date/Time 09/06/2011 12:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PRAVIA

DATE/TIME OF TEST: 08/04/2011 01:43:39

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample #1	0.000	01:39
Breath Volume	1.652	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample #2	0.000	01:42
Breath Volume	1.375	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time 08/17/2011 10:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REEVES, PETRA, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ESPINOSA, HARRY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SUNCEN

DATE/TIME OF TEST: 08/07/2011 04:33:13

---

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Subject Sample #1	0.216	04:28
Breath Volume	1.933	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample #2	0.231	04:31
Breath Volume	2.605	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time 08/17/2011 10:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 06:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 08/12/2011 07:03:23

---

Diagnostics Check	Fail	07:02	Temperature Regulation Test
Air Blank	0.000	07:03	

---

Data Download Date/Time 08/17/2011 10:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 06:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 08/12/2011 07:08:28

---

Diagnostics Check	OK	07:05	
Air Blank	0.000	07:05	
Control Test	0.079	07:06	
Air Blank	0.000	07:06	
Subject Sample #1		07:07	Subject Test Refused
Breath Volume	0.000	07:07	
Air Blank	0.000	07:07	
Control Test	0.079	07:07	
Air Blank	0.000	07:08	
Diagnostics Check	OK	07:08	

---

Data Download Date/Time 08/17/2011 10:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REYES, ERNESTO,  
BREATH TEST OPERATOR AGENCY: North Miami PD  
ARRESTING OFFICER: REYES, ERNESTO  
ARRESTING OFFICER AGENCY: North Miami PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: DARLING

DATE/TIME OF TEST: 08/12/2011 16:19:23

---

Diagnostics Check	OK	16:12
Air Blank	0.000	16:13
Control Test	0.079	16:13
Air Blank	0.000	16:13
Subject Sample #1	0.133	16:14
Breath Volume	2.976	16:14
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample #2	0.120	16:17
Breath Volume	1.949	16:17
Air Blank	0.000	16:18
Control Test	0.077	16:18
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

---

Data Download Date/Time 08/17/2011 10:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOREIRA, ERNESTO, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: WHITE, KHARY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: OTHER  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: AHMED

DATE/TIME OF TEST: 08/14/2011 04:57:15

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:52
Subject Sample #1	0.220	04:52
Breath Volume	3.640	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.224	04:55
Breath Volume	3.359	04:55
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time 08/17/2011 10:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 06:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: GILMA

DATE/TIME OF TEST: 08/26/2011 07:21:16

---

Diagnostics Check	OK	07:14
Air Blank	0.000	07:14
Control Test	0.079	07:15
Air Blank	0.000	07:15
Subject Sample #1	0.207	07:16
Breath Volume	1.757	07:16
Air Blank	0.000	07:17
Air Blank	0.000	07:19
Subject Sample #2	0.221	07:19
Breath Volume	1.386	07:19
Air Blank	0.000	07:20
Control Test	0.077	07:20
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

---

Data Download Date/Time 09/06/2011 12:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, PABLO,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: FETTERS, MAYLEE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: VALERIANO-ORELLANA

DATE/TIME OF TEST: 08/02/2011 04:02:17

---

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.082	04:00	
Air Blank	0.000	04:00	
Subject Sample #1		04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.083	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FETTERS, MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: GONZALEZ, ANGEL  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: DEPOO

DATE/TIME OF TEST: 08/03/2011 00:28:04

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.082	00:22
Air Blank	0.000	00:22
Subject Sample #1	0.304	00:23
Breath Volume	1.164	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:25
Subject Sample #2	0.302	00:26
Breath Volume	1.785	00:26
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:28

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 08/04/2011 02:46:57

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.083	02:34	
Air Blank	0.000	02:35	
Subject Sample #1	0.197	02:38	Volume Not Met
Breath Volume	0.070	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:40	
Subject Sample #2	0.195	02:41	
Breath Volume	1.566	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample #3	0.185	02:45	
Breath Volume	1.652	02:45	
Air Blank	0.000	02:46	
Control Test	0.083	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GALILOVA

DATE/TIME OF TEST: 08/05/2011 04:28:50

---

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.083	04:20
Air Blank	0.000	04:21
Subject Sample #1	0.117	04:23
Breath Volume	1.457	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample #2	0.103	04:27
Breath Volume	1.101	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, PABLO,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: FETTERS, MAYLEE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: ARISTEIGUIETA HERRER

DATE/TIME OF TEST: 08/06/2011 01:51:11

---

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.082	01:43	
Air Blank	0.000	01:44	
Subject Sample #1	0.148	01:47	Volume Not Met
Breath Volume	0.058	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample #2		01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.081	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, PABLO, M  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RODRIGUEZ, PABLO  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GUERRA CALLEJAS

DATE/TIME OF TEST: 08/06/2011 01:57:11

---

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.082	01:54	
Air Blank	0.000	01:55	
Subject Sample #1		01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.081	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FETTERS, MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RODRIGUEZ, PABLO  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ESCALONA

DATE/TIME OF TEST: 08/06/2011 05:48:22

---

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.082	05:42
Air Blank	0.000	05:42
Subject Sample #1	0.165	05:43
Breath Volume	1.320	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:45
Subject Sample #2	0.177	05:46
Breath Volume	1.636	05:46
Air Blank	0.000	05:47
Control Test	0.080	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, PABLO,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: FETTERS, MAYLEE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/07/2011 02:24:09

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.083	02:11	
Air Blank	0.000	02:11	
Subject Sample #1	0.227	02:15	Volume Not Met
Breath Volume	0.050	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:17	
Subject Sample #2	0.234	02:19	
Breath Volume	1.351	02:19	
Air Blank	0.000	02:19	
Air Blank	0.000	02:21	
Subject Sample #3	0.243	02:22	
Breath Volume	1.453	02:22	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLARDO, JORGE, L  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: DESIUS, DUCLAIRE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 08/09/2011 02:35:45

---

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.083	02:33	
Air Blank	0.000	02:34	
Subject Sample #1		02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.082	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

---

Data Download Date/Time 08/16/2011 09:42

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHANOR, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SHANOR, WILLIAM  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 16:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: CAMPUZANO CAMPO

DATE/TIME OF TEST: 08/09/2011 16:57:06

---

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:51	
Control Test	0.081	16:51	
Air Blank	0.000	16:52	
Subject Sample #1	0.201	16:55	Volume Not Met
Breath Volume	0.562	16:55	
Air Blank		16:56	Purge Fail
Air Blank	0.000	16:57	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHANOR, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SHANOR, WILLIAM  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 16:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: CAMPUZANO CAMPO

DATE/TIME OF TEST: 08/09/2011 17:31:13

---

Diagnostics Check	OK	17:18	
Air Blank	0.000	17:18	
Control Test	0.080	17:19	
Air Blank	0.000	17:19	
Subject Sample #1	0.216	17:22	Volume Not Met
Breath Volume	0.000	17:22	
Air Blank	0.000	17:24	
Air Blank	0.000	17:25	
Subject Sample #2	0.239	17:26	
Breath Volume	1.253	17:26	
Air Blank	0.000	17:27	
Air Blank	0.000	17:28	
Subject Sample #3	0.211	17:29	
Breath Volume	1.281	17:29	
Air Blank	0.000	17:30	
Control Test	0.073	17:30	Control Outside Tolerance
Air Blank	0.000	17:31	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHANOR, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SHANOR, WILLIAM  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 16:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: CAMPUZANO CAMPO

DATE/TIME OF TEST: 08/09/2011 17:40:45

---

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:34	
Control Test	0.076	17:34	
Air Blank	0.000	17:35	
Subject Sample #1	0.233	17:35	
Breath Volume	1.160	17:35	
Air Blank	0.000	17:36	
Air Blank	0.000	17:38	
Subject Sample #2	0.204	17:39	
Breath Volume	1.304	17:39	
Air Blank		17:40	Purge Fail
Air Blank	0.000	17:40	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ, CARLOS, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MELMER

DATE/TIME OF TEST: 08/11/2011 03:43:21

---

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.083	03:40	
Air Blank	0.000	03:41	
Subject Sample #1		03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.083	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FETTERS, MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: FETTERS, MAYLEE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 08/12/2011 01:48:49

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.083	01:43
Air Blank	0.000	01:43
Subject Sample #1	0.121	01:44
Breath Volume	2.375	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample #2	0.124	01:47
Breath Volume	1.906	01:47
Air Blank	0.000	01:47
Control Test	0.082	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: COLES

DATE/TIME OF TEST: 08/12/2011 02:34:51

---

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.083	02:32	
Air Blank	0.000	02:33	
Subject Sample #1		02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.083	02:34	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/12/2011 03:17:27

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.083	03:14	
Air Blank	0.000	03:15	
Subject Sample #1		03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.082	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 08/12/2011 03:29:19

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.083	03:23
Air Blank	0.000	03:24
Subject Sample #1	0.111	03:24
Breath Volume	1.308	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.122	03:27
Breath Volume	2.628	03:27
Air Blank	0.000	03:28
Control Test	0.083	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHANOR, WILLIAM, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SHANOR, WILLIAM  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: FERRIS

DATE/TIME OF TEST: 08/12/2011 21:44:45

---

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.082	21:39
Air Blank	0.000	21:40
Subject Sample #1	0.000	21:40
Breath Volume	3.925	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample #2	0.000	21:43
Breath Volume	3.890	21:43
Air Blank	0.000	21:43
Control Test	0.082	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCGRATH, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MCGRATH, ROBERT  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOMEZ GUITERREZ

DATE/TIME OF TEST: 08/13/2011 01:57:13

---

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.083	01:54	
Air Blank	0.000	01:55	
Subject Sample #1		01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.082	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: DE LA PAZ

DATE/TIME OF TEST: 08/13/2011 02:13:14

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.082	02:11	
Air Blank	0.000	02:11	
Subject Sample #1		02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.082	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 08/14/2011 00:37:30

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.083	00:32
Air Blank	0.000	00:32
Subject Sample #1	0.171	00:33
Breath Volume	3.917	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample #2	0.167	00:35
Breath Volume	3.449	00:35
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

---

Data Download Date/Time 08/16/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ASAD

DATE/TIME OF TEST: 08/17/2011 04:57:25

---

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.083	04:52
Air Blank	0.000	04:52
Subject Sample #1	0.124	04:53
Breath Volume	1.886	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.121	04:55
Breath Volume	1.886	04:55
Air Blank	0.000	04:56
Control Test	0.082	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO, GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO, GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: AKERS

DATE/TIME OF TEST: 08/19/2011 00:25:26

---

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.083	00:17
Air Blank	0.000	00:17
Subject Sample #1	0.144	00:18
Breath Volume	1.113	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample #2	0.143	00:23
Breath Volume	1.343	00:23
Air Blank	0.000	00:24
Control Test	0.082	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 08/19/2011 01:02:05

---

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:56
Subject Sample #1	0.272	00:57
Breath Volume	1.511	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample #2	0.269	01:00
Breath Volume	2.738	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:02

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK, JOSEPH, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK, JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: GADISON

DATE/TIME OF TEST: 08/19/2011 05:54:31

---

Diagnostics Check	OK	05:47
Air Blank	0.000	05:48
Control Test	0.083	05:48
Air Blank	0.000	05:49
Subject Sample #1	0.134	05:49
Breath Volume	3.425	05:49
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample #2	0.137	05:52
Breath Volume	3.550	05:52
Air Blank	0.000	05:53
Control Test	0.082	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FETTERS, MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RODRIGUEZ, PABLO  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VELIZ

DATE/TIME OF TEST: 08/20/2011 02:57:19

---

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.082	02:51
Air Blank	0.000	02:51
Subject Sample #1	0.275	02:52
Breath Volume	1.390	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample #2	0.260	02:55
Breath Volume	2.078	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FETTERS, MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RODRIGUEZ, PABLO  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TRAMONTANA

DATE/TIME OF TEST: 08/25/2011 02:34:00

---

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.083	02:31	
Air Blank	0.000	02:32	
Subject Sample #1		02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.084	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:34	

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARRERO, LEONARDO,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: SANCHEZ, ORLANDO  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: BARCENAS

DATE/TIME OF TEST: 08/25/2011 17:30:59

---

Diagnostics Check	OK	17:24
Air Blank	0.000	17:25
Control Test	0.082	17:25
Air Blank	0.000	17:25
Subject Sample #1	0.274	17:26
Breath Volume	3.652	17:26
Air Blank	0.000	17:27
Air Blank	0.000	17:28
Subject Sample #2	0.263	17:29
Breath Volume	2.785	17:29
Air Blank	0.000	17:30
Control Test	0.081	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, PABLO,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: FETTERS, MAYLE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/26/2011 00:49:48

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.082	00:47	
Air Blank	0.000	00:48	
Subject Sample #1		00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.083	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK, JOSEPH, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK, JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: PRADO MORA

DATE/TIME OF TEST: 08/27/2011 02:49:50

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.083	02:38	
Air Blank	0.000	02:39	
Subject Sample #1	0.168	02:42	Slope Not Level
Breath Volume	1.347	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:44	
Subject Sample #2	0.158	02:48	Volume Not Met
Breath Volume	0.757	02:48	
Air Blank	0.000	02:48	
Control Test	0.082	02:49	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK, JOSEPH, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK, JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: PRADO MORA

DATE/TIME OF TEST: 08/27/2011 03:00:16

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.157	02:54
Breath Volume	2.042	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.157	02:58
Breath Volume	1.250	02:58
Air Blank	0.000	02:59
Control Test	0.082	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: KREMER

DATE/TIME OF TEST: 08/27/2011 04:22:59

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.082	04:17
Air Blank	0.000	04:17
Subject Sample #1	0.067	04:18
Breath Volume	1.480	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.067	04:21
Breath Volume	1.894	04:21
Air Blank	0.000	04:22
Control Test	0.083	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESPINOZA, PABLO,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CELLA, DOMENICO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/28/2011 01:07:30

---

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.083	01:02
Air Blank	0.000	01:02
Subject Sample #1	0.234	01:03
Breath Volume	2.796	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample #2	0.242	01:05
Breath Volume	2.773	01:05
Air Blank	0.000	01:06
Control Test	0.083	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/28/2011 03:33:23

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.082	03:27
Air Blank	0.000	03:27
Subject Sample #1	0.243	03:28
Breath Volume	1.421	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample #2	0.242	03:31
Breath Volume	1.582	03:31
Air Blank	0.000	03:32
Control Test	0.083	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FETTERS, MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RODRIGUEZ, PABLO  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: HEREDIA

DATE/TIME OF TEST: 08/30/2011 01:26:17

---

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.083	01:20
Air Blank	0.000	01:20
Subject Sample #1	0.011	01:21
Breath Volume	2.085	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample #2	0.011	01:24
Breath Volume	2.464	01:24
Air Blank	0.000	01:25
Control Test	0.083	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FETTERS, MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: RODRIGUEZ, PABLO  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 779418  
DRY GAS EXPIRY DATE: 06/11/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 08/31/2011 05:15:04

---

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:12	
Control Test	0.083	05:12	
Air Blank	0.000	05:13	
Subject Sample #1		05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:14	
Control Test	0.083	05:14	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:15	

---

Data Download Date/Time 09/01/2011 13:24

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO, ALEX, B  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CASTILLO, ALEX  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 08/02/2011 02:41:37

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:36
Air Blank	0.000	02:36
Subject Sample #1	0.212	02:37
Breath Volume	1.910	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample #2	0.214	02:40
Breath Volume	3.062	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SLIMAK, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MELENDEZ MORAN

DATE/TIME OF TEST: 08/05/2011 02:03:37

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Subject Sample #1	0.128	01:59
Breath Volume	1.496	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample #2	0.119	02:02
Breath Volume	1.429	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SLIMAK, MARK  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: AMIL-MENDEZ

DATE/TIME OF TEST: 08/06/2011 06:52:53

---

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.080	06:46
Air Blank	0.000	06:47
Subject Sample #1	0.140	06:48
Breath Volume	1.460	06:48
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample #2	0.142	06:51
Breath Volume	2.023	06:51
Air Blank	0.000	06:51
Control Test	0.079	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: MONTES DE OCA

DATE/TIME OF TEST: 08/10/2011 06:45:14

---

Diagnostics Check	OK	06:42	
Air Blank	0.000	06:42	
Control Test	0.079	06:43	
Air Blank	0.000	06:43	
Subject Sample #1		06:43	Subject Test Refused
Breath Volume	0.000	06:43	
Air Blank	0.000	06:44	
Control Test	0.080	06:44	
Air Blank	0.000	06:45	
Diagnostics Check	OK	06:45	

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, R  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CORONA

DATE/TIME OF TEST: 08/12/2011 00:38:03

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Subject Sample #1	0.095	00:33
Breath Volume	1.140	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample #2	0.092	00:36
Breath Volume	1.363	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:38

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SLIMAK, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 08/12/2011 06:25:34

---

Diagnostics Check	OK	06:19
Air Blank	0.000	06:19
Control Test	0.079	06:20
Air Blank	0.000	06:20
Subject Sample #1	0.133	06:21
Breath Volume	4.117	06:21
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample #2	0.131	06:23
Breath Volume	3.972	06:23
Air Blank	0.000	06:24
Control Test	0.080	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SLIMAK, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 08/13/2011 02:06:54

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample #1	0.158	02:00
Breath Volume	1.718	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample #2	0.143	02:05
Breath Volume	1.671	02:05
Air Blank	0.000	02:05
Control Test	0.082	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ, CARLOS, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PEREZ, CARLOS  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 08/13/2011 21:52:31

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Subject Sample #1	0.185	21:47
Breath Volume	1.816	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample #2	0.174	21:50
Breath Volume	1.359	21:50
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ, CARLOS, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK, JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: LORENZO-ANTONIO

DATE/TIME OF TEST: 08/14/2011 00:26:21

---

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:20
Air Blank	0.000	00:20
Subject Sample #1	0.193	00:21
Breath Volume	2.347	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample #2	0.199	00:24
Breath Volume	2.878	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SLIMAK, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: OCASIO

DATE/TIME OF TEST: 08/14/2011 04:43:23

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample #1	0.162	04:38
Breath Volume	3.015	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample #2	0.166	04:41
Breath Volume	3.273	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ, CARLOS, A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SUAREZ, CARLOS  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 08/14/2011 05:50:20

---

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:43
Subject Sample #1	0.199	05:45
Breath Volume	2.242	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:47
Subject Sample #2	0.184	05:48
Breath Volume	2.128	05:48
Air Blank	0.000	05:49
Control Test	0.078	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

---

Data Download Date/Time 08/16/2011 09:45

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHANG, BRADLEY, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ROQUE, EUGENE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 17:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 08/19/2011 18:14:27

---

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.080	18:09
Air Blank	0.000	18:09
Subject Sample #1	0.000	18:10
Breath Volume	3.781	18:10
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample #2	0.000	18:13
Breath Volume	3.277	18:13
Air Blank	0.000	18:13
Control Test	0.081	18:13
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ, CARLOS,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SUAREZ, CARLOS  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/20/2011 03:30:44

---

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Subject Sample #1	0.141	03:25	Volume Not Met
Breath Volume	0.085	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample #2		03:29	Subject Test Refused
Breath Volume	0.054	03:29	
Air Blank	0.000	03:29	
Control Test	0.076	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ, CARLOS, A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SUAREZ, CARLOS  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: BASTO-LOPEZ

DATE/TIME OF TEST: 08/20/2011 07:03:41

---

Diagnostics Check	OK	06:58	
Air Blank	0.000	06:59	
Control Test	0.079	06:59	
Air Blank	0.000	07:00	
Subject Sample #1		07:02	Subject Test Refused
Breath Volume	0.000	07:02	
Air Blank	0.000	07:02	
Control Test	0.079	07:03	
Air Blank	0.000	07:03	
Diagnostics Check	OK	07:03	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SLIMAK, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 06:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: MULLER

DATE/TIME OF TEST: 08/20/2011 07:24:52

---

Diagnostics Check	OK	07:21	
Air Blank	0.000	07:22	
Control Test	0.078	07:22	
Air Blank	0.000	07:23	
Subject Sample #1		07:23	Subject Test Refused
Breath Volume	0.000	07:23	
Air Blank	0.000	07:23	
Control Test	0.078	07:24	
Air Blank	0.000	07:24	
Diagnostics Check	OK	07:24	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ, CARLOS, A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SUAREZ, CARLOS  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BARRERAS

DATE/TIME OF TEST: 08/21/2011 03:40:30

---

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.079	03:37	
Air Blank	0.000	03:38	
Subject Sample #1		03:38	Slope Not Met
Breath Volume	0.675	03:38	
Air Blank	0.000	03:39	
Control Test	0.080	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ, CARLOS, A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SUAREZ, CARLOS  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BARRERAS

DATE/TIME OF TEST: 08/21/2011 03:53:17

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Subject Sample #1	0.294	03:44
Breath Volume	1.363	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample #2	0.267	03:48
Breath Volume	1.257	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample #3	0.275	03:51
Breath Volume	1.300	03:51
Air Blank	0.000	03:52
Control Test	0.077	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIAMK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: BIANCHI, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: PERDOMO MARTINEZ

DATE/TIME OF TEST: 08/21/2011 05:29:44

---

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:27	
Control Test	0.078	05:27	
Air Blank	0.000	05:28	
Subject Sample #1		05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:28	
Control Test	0.079	05:29	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEARNEY, ANGELA, Y  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: RODRIGUEZ, JORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 08/25/2011 05:42:59

---

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.080	05:39	
Air Blank	0.000	05:40	
Subject Sample #1		05:41	Subject Test Refused
Breath Volume	0.054	05:41	
Air Blank	0.000	05:42	
Control Test	0.081	05:42	
Air Blank	0.000	05:42	
Diagnostics Check	OK	05:42	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: GOMEZ-BENAVIDEZ

DATE/TIME OF TEST: 08/26/2011 00:29:23

---

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.080	00:27	
Air Blank	0.000	00:27	
Subject Sample #1		00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: REGUERA

DATE/TIME OF TEST: 08/26/2011 02:14:49

---

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample #1	0.166	02:09
Breath Volume	1.937	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample #2	0.165	02:13
Breath Volume	1.730	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SLIMAK, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: RUANO

DATE/TIME OF TEST: 08/26/2011 03:20:29

---

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:18	
Air Blank	0.000	03:18	
Subject Sample #1		03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK, JOSEPH, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK, JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 08/27/2011 23:43:45

---

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.080	23:32	
Air Blank	0.000	23:32	
Subject Sample #1	0.161	23:35	Volume Not Met
Breath Volume	0.636	23:35	
Air Blank	0.000	23:36	
Air Blank	0.000	23:38	
Subject Sample #2	0.186	23:39	
Breath Volume	1.527	23:39	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample #3	0.199	23:42	
Breath Volume	2.023	23:42	
Air Blank	0.000	23:42	
Control Test	0.079	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ, CARLOS, A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SUAREZ, CARLOS  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 08/28/2011 05:14:45

---

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:09	
Control Test	0.080	05:09	
Air Blank	0.000	05:10	
Subject Sample #1		05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:13	
Control Test	0.080	05:14	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRUBBS, DARRYL, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRUBBS, DARRYL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/29/2011 00:32:20

---

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:29	
Subject Sample #1	0.191	00:30	
Breath Volume	1.523	00:30	
Air Blank		00:31	Purge Fail
Air Blank	0.000	00:32	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRUBBS, DARRYL, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRUBBS, DARRYL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/29/2011 00:38:14

---

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:35	
Subject Sample #1	0.184	00:36	
Breath Volume	1.136	00:36	
Air Blank		00:37	Purge Fail
Air Blank	0.000	00:38	

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRUBBS, DARRYL, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRUBBS, DARRYL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/29/2011 00:50:52

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Subject Sample #1	0.182	00:45
Breath Volume	1.808	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample #2	0.184	00:49
Breath Volume	1.960	00:49
Air Blank	0.000	00:49
Control Test	0.081	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

---

Data Download Date/Time 09/02/2011 13:43

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: DELATORRE

DATE/TIME OF TEST: 08/03/2011 04:36:51

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:31
Subject Sample #1	0.075	04:32
Breath Volume	1.675	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample #2	0.082	04:35
Breath Volume	2.605	04:35
Air Blank	0.000	04:35
Control Test	0.083	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS, ANTONIO, J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SOCARRAS, ANTONIO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 08/04/2011 01:16:03

---

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Subject Sample #1	0.163	01:11
Breath Volume	1.652	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample #2	0.168	01:14
Breath Volume	1.753	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:16

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK, JOSEPH, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK, JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MARTINEZ CAZORLA

DATE/TIME OF TEST: 08/04/2011 02:25:33

---

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:20
Air Blank	0.000	02:20
Subject Sample #1	0.196	02:21
Breath Volume	3.414	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample #2	0.191	02:24
Breath Volume	2.531	02:24
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLIMAK, M,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SLIMAK, M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SAN JUAN

DATE/TIME OF TEST: 08/04/2011 04:28:01

---

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.082	04:25	
Air Blank	0.000	04:26	
Subject Sample #1		04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Control Test	0.081	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:28	

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCGRATH, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MCGRATH, ROBERT  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SIRGANY

DATE/TIME OF TEST: 08/06/2011 03:44:51

---

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:43	
Subject Sample #1		03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.082	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCGRATH, ROBERT, E  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MCGRATH, ROBERT  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: JARAMILLO

DATE/TIME OF TEST: 08/11/2011 00:30:35

---

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.082	00:23
Air Blank	0.000	00:24
Subject Sample #1	0.148	00:25
Breath Volume	1.304	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample #2	0.145	00:29
Breath Volume	1.277	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/11/2011 05:01:32

---

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.083	04:56
Air Blank	0.000	04:56
Subject Sample #1	0.080	04:57
Breath Volume	2.027	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample #2	0.080	05:00
Breath Volume	2.210	05:00
Air Blank	0.000	05:00
Control Test	0.083	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOCARRAS, ANTONIO, J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SOCARRAS, ANTONIO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/12/2011 04:04:25

---

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.082	04:02	
Air Blank	0.000	04:02	
Subject Sample #1		04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.082	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOIVU, DAVID, J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: AZRAK, JOSEPH  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: OROZCO VASQUEZ

DATE/TIME OF TEST: 08/13/2011 03:05:07

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample #1	0.225	03:00
Breath Volume	1.410	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.227	03:03
Breath Volume	1.214	03:03
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO, GREGORY, R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO, GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 08/13/2011 03:36:49

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Subject Sample #1	0.162	03:30
Breath Volume	2.230	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample #2	0.158	03:35
Breath Volume	2.558	03:35
Air Blank	0.000	03:35
Control Test	0.083	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOIVU, DAVID,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MCGRATH, ROBERT  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 05:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: SIMBACO

DATE/TIME OF TEST: 08/13/2011 06:22:24

---

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.082	06:17
Air Blank	0.000	06:17
Subject Sample #1	0.126	06:18
Breath Volume	1.480	06:18
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample #2	0.126	06:20
Breath Volume	1.386	06:20
Air Blank	0.000	06:21
Control Test	0.082	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS, RICHARD, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS, RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ALONSO-TORRAS

DATE/TIME OF TEST: 08/14/2011 02:24:41

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.082	02:19
Air Blank	0.000	02:19
Subject Sample #1	0.035	02:20
Breath Volume	3.089	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample #2	0.034	02:23
Breath Volume	3.945	02:23
Air Blank	0.000	02:23
Control Test	0.083	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time 08/16/2011 09:46

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANCHEZ, LUCIANO,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SANCHEZ, LUCIANO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: ADMIN  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 08/17/2011 02:26:26

---

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.082	02:20
Air Blank	0.000	02:20
Subject Sample #1	0.132	02:21
Breath Volume	1.425	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample #2	0.131	02:24
Breath Volume	1.410	02:24
Air Blank	0.000	02:25
Control Test	0.082	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time 09/03/2011 22:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOIVU, DAVID, J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KOIVU, DAVID  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: QUEVEDO BOUZAMAYOR

DATE/TIME OF TEST: 08/21/2011 05:25:55

---

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.083	05:23	
Air Blank	0.000	05:24	
Subject Sample #1		05:24	Subject Test Refused
Breath Volume	0.000	05:24	
Air Blank	0.000	05:25	
Control Test	0.082	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

---

Data Download Date/Time 09/03/2011 22:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO, FRANCISCO,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MORENO, FRANCISCO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 08:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:26

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 08/24/2011 08:53:19

---

Diagnostics Check	OK	08:46
Air Blank	0.000	08:46
Control Test	0.082	08:47
Air Blank	0.000	08:47
Subject Sample #1	0.000	08:48
Breath Volume	1.460	08:48
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample #2	0.000	08:51
Breath Volume	1.617	08:51
Air Blank	0.000	08:52
Control Test	0.084	08:52
Air Blank	0.000	08:53
Diagnostics Check	OK	08:53

---

Data Download Date/Time 09/03/2011 22:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANCHEZ, LUCINO,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: SANCHEZ/SLIMAK, LUCIANO/M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MARTINEZ DIAZ

DATE/TIME OF TEST: 08/25/2011 03:21:41

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.082	03:15
Air Blank	0.000	03:15
Subject Sample #1	0.100	03:17
Breath Volume	2.398	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample #2	0.096	03:20
Breath Volume	1.886	03:20
Air Blank	0.000	03:20
Control Test	0.082	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time 09/03/2011 22:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COWART, RYAN, L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: COWART, RYAN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/26/2011 04:00:12

---

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.084	03:56	
Air Blank	0.000	03:56	
Subject Sample #1		03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.083	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

---

Data Download Date/Time 09/03/2011 22:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO, FRANCISCO,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MILLARES, BENINGO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 09:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:33

SUBJECT'S LAST NAME: VILLEDA

DATE/TIME OF TEST: 08/26/2011 10:00:46

---

Diagnostics Check	OK	09:53
Air Blank	0.000	09:54
Control Test	0.082	09:54
Air Blank	0.000	09:54
Subject Sample #1	0.189	09:55
Breath Volume	1.816	09:55
Air Blank	0.000	09:56
Air Blank	0.000	09:58
Subject Sample #2	0.185	09:59
Breath Volume	1.511	09:59
Air Blank	0.000	09:59
Control Test	0.083	10:00
Air Blank	0.000	10:00
Diagnostics Check	OK	10:00

---

Data Download Date/Time 09/03/2011 22:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOIVU, DAVID, J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KOIVU, DAVID  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/16/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/28/2011 05:07:19

---

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:05	
Air Blank	0.000	05:05	
Subject Sample #1		05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

---

Data Download Date/Time 09/03/2011 22:54

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FOGELGREN, D, S

BREATH TEST OPERATOR AGENCY: North Miami Beach PD

ARRESTING OFFICER: SOTERO, A

ARRESTING OFFICER AGENCY: North Miami Beach PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 731635

DRY GAS EXPIRY DATE: 03/04/2012

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MINDUBAEV

DATE/TIME OF TEST: 08/04/2011 02:17:10

---

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:15	
Air Blank	0.000	02:15	
Subject Sample #1		02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, KENNETH,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: BAKER, K  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: TABARES

DATE/TIME OF TEST: 08/04/2011 04:15:43

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:09
Subject Sample #1	0.130	04:11
Breath Volume	1.164	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample #2	0.131	04:14
Breath Volume	1.292	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SOTERO, ARMANDO, D

BREATH TEST OPERATOR AGENCY: North Miami Beach PD

ARRESTING OFFICER: SOTERO, ARMANDO

ARRESTING OFFICER AGENCY: North Miami Beach PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 731635

DRY GAS EXPIRY DATE: 03/04/2012

VIOLATION CODE: DUI

ARREST TIME: 01:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: JURGENSEN

DATE/TIME OF TEST: 08/05/2011 01:41:22

---

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.080	01:39	
Air Blank	0.000	01:39	
Subject Sample #1		01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTERO, ARMANDO, D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: AZEVEDO, MICHAEL  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: BOLANOS PEREZ

DATE/TIME OF TEST: 08/05/2011 05:48:22

---

Diagnostics Check	OK	05:42	
Air Blank	0.000	05:43	
Control Test	0.080	05:43	
Air Blank	0.000	05:44	
Subject Sample #1		05:46	Subject Test Refused
Breath Volume	0.125	05:46	
Air Blank	0.000	05:47	
Control Test	0.078	05:47	
Air Blank	0.000	05:48	
Diagnostics Check	OK	05:48	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FOGELGREN, D, S

BREATH TEST OPERATOR AGENCY: North Miami Beach PD

ARRESTING OFFICER: SOTERO, A

ARRESTING OFFICER AGENCY: North Miami Beach PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 731635

DRY GAS EXPIRY DATE: 03/04/2012

VIOLATION CODE: DUI

ARREST TIME: 01:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: NEUMANN-CISNEROS

DATE/TIME OF TEST: 08/09/2011 02:25:45

---

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:24	
Subject Sample #1		02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.080	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOGELGREN, D, S  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: SOTERO, A  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: JARAMILLO CARDONA

DATE/TIME OF TEST: 08/09/2011 03:58:50

---

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.081	03:56	
Air Blank	0.000	03:57	
Subject Sample #1		03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTERO, ARMANDO, D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: SOTERO, ARMANDO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MONSON

DATE/TIME OF TEST: 08/12/2011 02:29:10

---

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.081	02:27	
Air Blank	0.000	02:27	
Subject Sample #1		02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTERO, ARMANDO, D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, PEDRO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LASTARRIA

DATE/TIME OF TEST: 08/12/2011 02:53:22

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample #1	0.134	02:48
Breath Volume	1.617	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample #2	0.151	02:51
Breath Volume	1.640	02:51
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS, ZOILA, M  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: AZEVEDO, MICHAEL  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 06:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: TELLO

DATE/TIME OF TEST: 08/13/2011 06:39:45

---

Diagnostics Check	OK	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Subject Sample #1	0.097	06:34
Breath Volume	2.328	06:34
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample #2	0.093	06:38
Breath Volume	3.492	06:38
Air Blank	0.000	06:38
Control Test	0.078	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SOTERO, ARMANDO, D

BREATH TEST OPERATOR AGENCY: North Miami Beach PD

ARRESTING OFFICER: GOMEZ, PEDRO

ARRESTING OFFICER AGENCY: North Miami Beach PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 731635

DRY GAS EXPIRY DATE: 03/04/2012

VIOLATION CODE: DUI

ARREST TIME: 00:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LIM

DATE/TIME OF TEST: 08/16/2011 01:13:26

---

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.081	01:11	
Air Blank	0.000	01:11	
Subject Sample #1		01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOGELGREN, D, S  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: FOGELGREN, D  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BENSADOUN

DATE/TIME OF TEST: 08/17/2011 03:18:31

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample #1	0.145	03:12
Breath Volume	1.375	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.135	03:16
Breath Volume	1.781	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SOTERO, ARMANDO, D

BREATH TEST OPERATOR AGENCY: North Miami Beach PD

ARRESTING OFFICER: GOMEZ, PEDRO

ARRESTING OFFICER AGENCY: North Miami Beach PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 731635

DRY GAS EXPIRY DATE: 03/04/2012

VIOLATION CODE: DUI

ARREST TIME: 05:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: GRANIER PHELPS

DATE/TIME OF TEST: 08/17/2011 05:30:22

---

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.079	05:26	
Air Blank	0.000	05:26	
Subject Sample #1		05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:29	
Control Test	0.078	05:29	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOGELGREN, D, S  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: SOTERO, A  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ZAHARIEV

DATE/TIME OF TEST: 08/18/2011 03:09:27

---

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:07	
Subject Sample #1		03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTERO, ARMANDO, D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, PEDRO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/19/2011 04:17:58

---

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:15	
Air Blank	0.000	04:16	
Subject Sample #1		04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MAYA, JOSE,

BREATH TEST OPERATOR AGENCY: North Miami Beach PD

ARRESTING OFFICER: SIMON, SAMUEL

ARRESTING OFFICER AGENCY: North Miami Beach PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 731635

DRY GAS EXPIRY DATE: 03/04/2012

VIOLATION CODE: DUI

ARREST TIME: 00:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: RAAB

DATE/TIME OF TEST: 08/21/2011 00:49:43

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank	0.000	00:48	
Subject Sample #1		00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.080	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYA, JOSE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: MAYA, JOSE  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 08/21/2011 05:12:24

---

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Subject Sample #1	0.117	05:07
Breath Volume	1.269	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample #2	0.113	05:10
Breath Volume	1.207	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTERO, ARMANDO, D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, PEDRO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: FELLHAUER

DATE/TIME OF TEST: 08/24/2011 05:35:25

---

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Subject Sample #1	0.169	05:30
Breath Volume	1.886	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample #2	0.166	05:33
Breath Volume	1.238	05:33
Air Blank	0.000	05:34
Control Test	0.078	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTERO, ARMANDO, D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, PEDRO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ALEMAN

DATE/TIME OF TEST: 08/25/2011 02:27:08

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample #1	0.167	02:22
Breath Volume	2.558	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample #2	0.159	02:25
Breath Volume	2.699	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTERO, ARMANDO, D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, PEDRO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 08/26/2011 02:41:17

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.080	02:39	
Air Blank	0.000	02:39	
Subject Sample #1		02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.081	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, KENNETH,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: BAKER, K  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GAVERGUN

DATE/TIME OF TEST: 08/28/2011 02:36:36

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample #1	0.000	02:32
Breath Volume	2.000	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample #2	0.000	02:35
Breath Volume	1.890	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTERO, ARMANDO, D  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, PEDRO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: SCHWARTZMAN

DATE/TIME OF TEST: 08/30/2011 02:34:05

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample #1	0.184	02:28
Breath Volume	2.449	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample #2	0.189	02:32
Breath Volume	2.449	02:32
Air Blank	0.000	02:33
Control Test	0.075	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:34

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOGELGREN, DENNIS, S  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, P  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 05:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: SALLAM

DATE/TIME OF TEST: 08/30/2011 06:02:15

---

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.079	05:56
Air Blank	0.000	05:56
Subject Sample #1	0.152	05:57
Breath Volume	2.128	05:57
Air Blank	0.000	05:58
Air Blank	0.000	05:59
Subject Sample #2	0.142	06:00
Breath Volume	1.769	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYA, JOSE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: SOTERO, ARMANDO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 06:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: MONTAYNE

DATE/TIME OF TEST: 08/30/2011 06:56:17

---

Diagnostics Check	OK	06:44	
Air Blank	0.000	06:45	
Control Test	0.080	06:45	
Air Blank	0.000	06:46	
Subject Sample #1		06:49	No Sample Provided
Breath Volume	0.000	06:49	
Air Blank	0.000	06:49	
Air Blank	0.000	06:51	
Subject Sample #2		06:54	No Sample Provided
Breath Volume	0.000	06:54	
Air Blank	0.000	06:55	
Control Test	0.079	06:55	
Air Blank	0.000	06:56	
Diagnostics Check	OK	06:56	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FOGELGREN, D, S  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: SOTERO, A  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ROA

DATE/TIME OF TEST: 08/31/2011 00:45:35

---

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Subject Sample #1	0.196	00:41
Breath Volume	1.765	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample #2	0.187	00:44
Breath Volume	1.890	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPOTE, ALEXANDER, G  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: GOMEZ, PEDRO  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 731635  
DRY GAS EXPIRY DATE: 03/04/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 08/31/2011 22:01:36

---

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	22:00	
Subject Sample #1		22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

---

Data Download Date/Time 09/08/2011 20:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER, ROGER, G  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: TEST, TEST  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG110501  
DRY GAS EXPIRY DATE: 04/15/2013  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 15:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 08/03/2011 15:38:15

---

Diagnostics Check	OK	15:36	
Air Blank	0.000	15:37	
Control Test	0.000	15:37	Control Outside Tolerance
Air Blank	0.000	15:38	

---

Data Download Date/Time 08/03/2011 15:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER, ROGER, G  
BREATH TEST OPERATOR AGENCY: FDLE  
ARRESTING OFFICER: TEST, TEST  
ARRESTING OFFICER AGENCY: FDLE  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG110501  
DRY GAS EXPIRY DATE: 04/15/2013  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 15:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 08/03/2011 16:45:48

---

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:44	
Control Test	0.000	16:45	Control Outside Tolerance
Air Blank	0.000	16:45	

---

Data Download Date/Time 08/04/2011 07:50

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC DESTIN  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB, MATT,  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: KIRCHINGER, ALAN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG005701  
DRY GAS EXPIRY DATE: 02/26/2012  
VIOLATION CODE: BUI  
ARREST TIME: 16:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 08/07/2011 16:52:52

---

Diagnostics Check	OK	16:46
Air Blank	0.000	16:46
Control Test	0.078	16:47
Air Blank	0.000	16:47
Subject Sample #1	0.239	16:48
Breath Volume	3.296	16:48
Air Blank	0.000	16:48
Air Blank	0.000	16:50
Subject Sample #2	0.237	16:51
Breath Volume	2.386	16:51
Air Blank	0.000	16:51
Control Test	0.076	16:52
Air Blank	0.000	16:52
Diagnostics Check	OK	16:52

---

Data Download Date/Time 08/12/2011 13:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC DESTIN  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRCHINGER, ALAN, W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: KIRCHINGER, ALAN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 08/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG005701  
DRY GAS EXPIRY DATE: 02/26/2012  
VIOLATION CODE: BUI  
ARREST TIME: 18:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: KREBS

DATE/TIME OF TEST: 08/13/2011 19:17:28

---

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:13	
Control Test	0.078	19:13	
Air Blank	0.000	19:14	
Subject Sample #1		19:16	Subject Test Refused
Breath Volume	0.000	19:16	
Air Blank	0.000	19:16	
Control Test	0.079	19:16	
Air Blank	0.000	19:17	
Diagnostics Check	OK	19:17	

---

Data Download Date/Time 09/19/2011 12:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC DESTIN  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRCHINGER, ALAN, W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: KIRCHINGER, ALAN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 08/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG005701  
DRY GAS EXPIRY DATE: 02/26/2012  
VIOLATION CODE: BUI  
ARREST TIME: 17:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: LONDON

DATE/TIME OF TEST: 08/14/2011 17:35:41

---

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.079	17:29
Air Blank	0.000	17:29
Subject Sample #1	0.234	17:30
Breath Volume	2.496	17:30
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample #2	0.235	17:34
Breath Volume	1.355	17:34
Air Blank	0.000	17:34
Control Test	0.076	17:35
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

---

Data Download Date/Time 09/19/2011 12:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC DESTIN  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB, MATT,  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CULBERSON, DARYL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 08/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG005701  
DRY GAS EXPIRY DATE: 02/26/2012  
VIOLATION CODE: BUI  
ARREST TIME: 17:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 08/20/2011 18:39:11

---

Diagnostics Check	OK	18:24	
Air Blank	0.000	18:24	
Control Test	0.078	18:25	
Air Blank	0.000	18:25	
Subject Sample #1	0.139	18:28	No .020 agreement
Breath Volume	1.457	18:28	
Air Blank	0.000	18:28	
Air Blank	0.000	18:30	
Subject Sample #2	0.174	18:31	No .020 agreement
Breath Volume	1.218	18:31	
Air Blank	0.000	18:32	
Air Blank	0.000	18:34	
Subject Sample #3	0.157	18:37	Volume Not Met
Breath Volume	0.039	18:37	
Air Blank	0.000	18:38	
Control Test	0.076	18:38	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:39	

---

Data Download Date/Time 09/19/2011 12:49

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC PANAMA CITY  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSS, NEAL, G  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: GOSS, NEAL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 08/12/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG005701  
DRY GAS EXPIRY DATE: 02/26/2012  
VIOLATION CODE: BUI  
ARREST TIME: 18:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 08/27/2011 19:16:47

---

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:14	
Control Test	0.081	19:14	
Air Blank	0.000	19:15	
Subject Sample #1		19:15	Subject Test Refused
Breath Volume	0.000	19:15	
Air Blank	0.000	19:15	
Control Test	0.080	19:16	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

---

Data Download Date/Time 09/19/2011 12:38

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELLS, GEORGE, G  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: WISE, AARON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 08/04/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: BUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 08/20/2011 21:18:00

---

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:14	
Control Test	0.076	21:14	
Air Blank	0.000	21:14	
Subject Sample #1		21:16	Subject Test Refused
Breath Volume	0.000	21:16	
Air Blank	0.000	21:17	
Control Test	0.076	21:17	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:18	

---

Data Download Date/Time 09/21/2011 12:33

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

INSTRUMENT REGISTERED TO: FWC - OCALA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: YETTER, JAMES, R

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: YETTER, JAMES

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 08/04/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG005701

DRY GAS EXPIRY DATE: 02/26/2012

VIOLATION CODE: BUI

ARREST TIME: 23:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: NORDICK

DATE/TIME OF TEST: 08/21/2011 01:04:27

---

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.079	00:59	
Air Blank	0.000	01:00	
Subject Sample #1		01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

---

Data Download Date/Time 09/19/2011 13:41

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROLL, JOHN, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: PROLL, JOHN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 08/31/2011 22:27:56

---

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:23	
Control Test	0.078	22:24	
Air Blank	0.000	22:24	
Subject Sample #1		22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.078	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

---

Data Download Date/Time 09/28/2011 12:00

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, EDISON, B  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: CRUZ, EDISON  
ARRESTING OFFICER AGENCY: Village of Pinecrest  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11111080A1  
DRY GAS EXPIRY DATE: 05/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: CALZADA MAYOR

DATE/TIME OF TEST: 08/20/2011 01:25:25

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.076	01:22	
Air Blank	0.000	01:23	
Subject Sample #1		01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.076	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

---

Data Download Date/Time 09/07/2011 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, EDISON, B  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: CRUZ, EDISON  
ARRESTING OFFICER AGENCY: Village of Pinecrest  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11111080A1  
DRY GAS EXPIRY DATE: 05/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MAYOR

DATE/TIME OF TEST: 08/20/2011 01:43:01

---

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.078	01:32	
Air Blank	0.000	01:32	
Subject Sample #1		01:36	No Sample Provided
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample #2		01:41	No Sample Provided
Breath Volume	0.000	01:41	
Air Blank	0.000	01:42	
Control Test	0.077	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:43	

---

Data Download Date/Time 09/07/2011 11:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROLL, JOHN, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: PROLL, JOHN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 08/18/2011 00:35:05

---

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.078	00:26	
Air Blank	0.000	00:26	
Subject Sample #1		00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample #2		00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.078	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:35	

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARRISH, JAMES, S  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: PARRISH, JAMES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PARRADO

DATE/TIME OF TEST: 08/18/2011 01:33:58

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:28
Subject Sample #1	0.187	01:29
Breath Volume	1.964	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:31
Subject Sample #2	0.198	01:32
Breath Volume	2.199	01:32
Air Blank	0.000	01:33
Control Test	0.077	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPURLOCK, ROBERT, V

BREATH TEST OPERATOR AGENCY: Manatee County SO

ARRESTING OFFICER: SPURLOCK, ROBERT

ARRESTING OFFICER AGENCY: Manatee County SO

LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 23:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/22/2011 00:49:33

---

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.078	00:42
Air Blank	0.000	00:42
Subject Sample #1	0.213	00:43
Breath Volume	1.199	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample #2	0.208	00:47
Breath Volume	2.578	00:47
Air Blank	0.000	00:48
Control Test	0.076	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTHERLAND, MICHAEL, P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SUTHERLAND, MICHAEL  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 08/22/2011 02:50:23

---

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.076	02:49	RFI Detect
Air Blank	0.000	02:50	

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTHERLAND, MICHAEL, P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SUTHERLAND, MICHAEL  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 08/22/2011 02:59:21

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.077	02:53
Air Blank	0.000	02:53
Subject Sample #1	0.138	02:54
Breath Volume	4.480	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample #2	0.136	02:57
Breath Volume	1.691	02:57
Air Blank	0.000	02:58
Control Test	0.076	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PROLL, JOHN, A

BREATH TEST OPERATOR AGENCY: Manatee County SO

ARRESTING OFFICER: BABAUTA, JEFFREY

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 01:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ACEBO

DATE/TIME OF TEST: 08/24/2011 02:02:09

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.077	01:56
Air Blank	0.000	01:56
Subject Sample #1	0.208	01:57
Breath Volume	1.570	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample #2	0.196	02:00
Breath Volume	1.691	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPURLOCK, ROBERT, V

BREATH TEST OPERATOR AGENCY: Manatee County SO

ARRESTING OFFICER: SPURLOCK, ROBERT

ARRESTING OFFICER AGENCY: Manatee County SO

LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 01:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GIARRUSSO

DATE/TIME OF TEST: 08/26/2011 02:58:56

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:53
Subject Sample #1	0.185	02:54
Breath Volume	2.621	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample #2	0.169	02:57
Breath Volume	1.363	02:57
Air Blank	0.000	02:58
Control Test	0.077	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLINE, GLENN, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: CLINE, GLENN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: OLEARY

DATE/TIME OF TEST: 08/27/2011 22:40:00

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.078	22:33
Air Blank	0.000	22:34
Subject Sample #1	0.266	22:35
Breath Volume	1.949	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:37
Subject Sample #2	0.266	22:38
Breath Volume	1.949	22:38
Air Blank	0.000	22:39
Control Test	0.077	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:40

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARNETT, LISA, M

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: BARNETT, LISA

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 885290

DRY GAS EXPIRY DATE: 10/19/2012

VIOLATION CODE: DUI

ARREST TIME: 22:19

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/28/2011 23:23:47

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.078	23:18
Air Blank	0.000	23:18
Subject Sample #1	0.200	23:19
Breath Volume	2.433	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample #2	0.184	23:22
Breath Volume	1.757	23:22
Air Blank	0.000	23:22
Control Test	0.076	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000912

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTHERLAND, MICHAEL, P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SUTHERLAND, MICHAEL  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 885290  
DRY GAS EXPIRY DATE: 10/19/2012  
VIOLATION CODE: DUI  
ARREST TIME: 17:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 08/30/2011 19:14:13

---

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.078	19:09
Air Blank	0.000	19:09
Subject Sample #1	0.000	19:10
Breath Volume	2.609	19:10
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample #2	0.000	19:12
Breath Volume	2.675	19:12
Air Blank	0.000	19:13
Control Test	0.078	19:13
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

---

Data Download Date/Time 09/28/2011 11:52

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000913

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMDATH, ANTHONY,

BREATH TEST OPERATOR AGENCY: Bradenton PD

ARRESTING OFFICER: MEYER, MICHAEL

ARRESTING OFFICER AGENCY: Bradenton PD

LAST AGENCY INSPECTION: 07/25/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG003305

DRY GAS EXPIRY DATE: 02/02/2012

VIOLATION CODE: DUI

ARREST TIME: 01:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 08/12/2011 02:32:35

---

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.079	02:29	
Air Blank	0.000	02:29	
Subject Sample #1		02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.078	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

---

Data Download Date/Time 08/17/2011 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000913

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COLEMAN, WILLIAM  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: WISMER

DATE/TIME OF TEST: 08/19/2011 03:47:39

---

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Subject Sample #1	0.200	03:35	Volume Not Met
Breath Volume	0.761	03:35	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample #2	0.191	03:40	
Breath Volume	1.453	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:42	
Subject Sample #3	0.195	03:46	Volume Not Met
Breath Volume	0.105	03:46	
Air Blank	0.000	03:46	
Control Test	0.078	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

---

Data Download Date/Time 09/28/2011 11:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000913

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: COLEMAN, WILLIAM  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: WISMER

DATE/TIME OF TEST: 08/19/2011 04:03:32

---

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.078	03:52	
Air Blank	0.000	03:52	
Subject Sample #1	0.192	03:53	
Breath Volume	1.558	03:53	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample #2	0.199	03:58	Volume Not Met
Breath Volume	0.097	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample #3	0.195	04:02	
Breath Volume	1.582	04:02	
Air Blank	0.000	04:02	
Control Test	0.078	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

---

Data Download Date/Time 09/28/2011 11:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000913

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLY, JOSEPH, A  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: CRADDOCK, NASHEKA  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/20/2011 22:40:59

---

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.078	22:38	
Air Blank	0.000	22:39	
Subject Sample #1		22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:40	
Control Test	0.078	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

---

Data Download Date/Time 09/28/2011 11:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000913

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD, CARRIE, B  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: VAUGHN, RYAN  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LENKER

DATE/TIME OF TEST: 08/27/2011 04:28:59

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Subject Sample #1	0.196	04:22
Breath Volume	1.304	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.197	04:27
Breath Volume	1.296	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

---

Data Download Date/Time 09/28/2011 11:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000913

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, RUSSELL, W  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: CRUZ, RUSSELL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SEIBERT

DATE/TIME OF TEST: 08/27/2011 19:11:28

---

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:09	
Control Test	0.079	19:09	
Air Blank	0.000	19:09	
Subject Sample #1		19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:10	
Control Test	0.079	19:10	
Air Blank	0.000	19:11	
Diagnostics Check	OK	19:11	

---

Data Download Date/Time 09/28/2011 11:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000913

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROLL, JOHN, A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: PROLL, JOHN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 08/17/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG003305  
DRY GAS EXPIRY DATE: 02/02/2012  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 08/27/2011 20:58:22

---

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.078	20:52
Air Blank	0.000	20:52
Subject Sample #1	0.154	20:53
Breath Volume	1.613	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample #2	0.159	20:56
Breath Volume	1.812	20:56
Air Blank	0.000	20:57
Control Test	0.078	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

---

Data Download Date/Time 09/28/2011 11:56

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KONDAS, STEVE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HEFFNER

DATE/TIME OF TEST: 08/20/2011 01:39:44

---

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.078	01:33
Air Blank	0.000	01:33
Subject Sample #1	0.167	01:35
Breath Volume	2.136	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample #2	0.171	01:38
Breath Volume	2.367	01:38
Air Blank	0.000	01:38
Control Test	0.077	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time 09/28/2011 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: FILLAK, M  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: TORRESOLA

DATE/TIME OF TEST: 08/20/2011 01:58:58

---

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:52
Air Blank	0.000	01:52
Subject Sample #1	0.126	01:53
Breath Volume	1.761	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.131	01:57
Breath Volume	1.519	01:57
Air Blank	0.000	01:58
Control Test	0.076	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

---

Data Download Date/Time 09/28/2011 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ROWE, B  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 08/20/2011 02:36:27

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.078	02:30
Air Blank	0.000	02:30
Subject Sample #1	0.202	02:31
Breath Volume	1.558	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample #2	0.201	02:34
Breath Volume	1.777	02:34
Air Blank	0.000	02:35
Control Test	0.076	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time 09/28/2011 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD, LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD, LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SCAFFIDE

DATE/TIME OF TEST: 08/21/2011 22:34:52

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:28
Subject Sample #1	0.000	22:29
Breath Volume	1.257	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample #2	0.000	22:33
Breath Volume	1.152	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

---

Data Download Date/Time 09/28/2011 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SERRANO, R,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SERRANO, R  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 13:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:54

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/27/2011 15:51:23

---

Diagnostics Check	OK	15:42
Air Blank	0.000	15:42
Control Test	0.079	15:43
Air Blank	0.000	15:43
Subject Sample #1	0.059	15:45
Breath Volume	1.574	15:45
Air Blank	0.000	15:45
Air Blank	0.000	15:47
Subject Sample #2	0.061	15:49
Breath Volume	1.148	15:49
Air Blank	0.000	15:50
Control Test	0.080	15:50
Air Blank	0.000	15:51
Diagnostics Check	OK	15:51

---

Data Download Date/Time 09/28/2011 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SERRANO, R,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ANDUJAR, J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 16:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: MONTES-DELGADO

DATE/TIME OF TEST: 08/27/2011 17:51:25

---

Diagnostics Check	OK	17:43
Air Blank	0.000	17:43
Control Test	0.079	17:44
Air Blank	0.000	17:44
Subject Sample #1	0.199	17:46
Breath Volume	1.332	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample #2	0.203	17:49
Breath Volume	2.750	17:49
Air Blank	0.000	17:50
Control Test	0.078	17:50
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

---

Data Download Date/Time 09/28/2011 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD, LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD, LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 688217  
DRY GAS EXPIRY DATE: 11/09/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PRIHODA

DATE/TIME OF TEST: 08/30/2011 03:29:24

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample #1	0.152	03:24
Breath Volume	2.558	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.149	03:27
Breath Volume	2.546	03:27
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time 09/28/2011 13:37

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO, JUAN, M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: NAJARRO, CARLOS  
ARRESTING OFFICER AGENCY: Lee Port Authority  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 08/02/2011 00:03:35

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.077	23:57
Air Blank	0.000	23:58
Subject Sample #1	0.015	23:59
Breath Volume	1.285	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample #2	0.014	00:02
Breath Volume	1.242	00:02
Air Blank	0.000	00:02
Control Test	0.077	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TRASK, J  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CASANOVA

DATE/TIME OF TEST: 08/02/2011 23:58:26

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.077	23:52
Air Blank	0.000	23:52
Subject Sample #1	0.143	23:53
Breath Volume	3.246	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample #2	0.136	23:56
Breath Volume	1.515	23:56
Air Blank	0.000	23:57
Control Test	0.077	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ, LEIDY, Y  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BERDEN, CHELSEA  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SCHNITKER

DATE/TIME OF TEST: 08/04/2011 21:05:46

---

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.076	21:00
Air Blank	0.000	21:00
Subject Sample #1	0.027	21:01
Breath Volume	3.757	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample #2	0.027	21:04
Breath Volume	3.167	21:04
Air Blank	0.000	21:04
Control Test	0.078	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD, LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GOULD, LEONARD  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 08/05/2011 22:43:52

---

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.077	22:38
Air Blank	0.000	22:38
Subject Sample #1	0.183	22:39
Breath Volume	2.808	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample #2	0.181	22:42
Breath Volume	2.628	22:42
Air Blank	0.000	22:42
Control Test	0.077	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BREEN, WILLIAM, R

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: BREEN, WILLIAM

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35710080A3

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 03:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ARAGON

DATE/TIME OF TEST: 08/06/2011 04:30:23

---

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.078	04:25
Air Blank	0.000	04:25
Subject Sample #1	0.162	04:25
Breath Volume	1.773	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample #2	0.160	04:28
Breath Volume	1.621	04:28
Air Blank	0.000	04:29
Control Test	0.077	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FUNK, TIMOTHY, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: RONGISH, JEFF  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 11:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: DENHAM

DATE/TIME OF TEST: 08/06/2011 13:08:39

---

Diagnostics Check	OK	13:01
Air Blank	0.000	13:02
Control Test	0.077	13:02
Air Blank	0.000	13:02
Subject Sample #1	0.210	13:03
Breath Volume	1.515	13:03
Air Blank	0.000	13:04
Air Blank	0.000	13:05
Subject Sample #2	0.212	13:07
Breath Volume	1.617	13:07
Air Blank	0.000	13:07
Control Test	0.078	13:08
Air Blank	0.000	13:08
Diagnostics Check	OK	13:08

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURPHY, WILLIAM, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MURPHY, WILLIAM  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: HERNANDEZ GUEVARA

DATE/TIME OF TEST: 08/07/2011 20:34:23

---

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.076	20:27
Air Blank	0.000	20:27
Subject Sample #1	0.209	20:29
Breath Volume	2.304	20:29
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample #2	0.194	20:32
Breath Volume	1.718	20:32
Air Blank	0.000	20:33
Control Test	0.077	20:33
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO, JUAN, M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: FAULK, MATHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: BLANDON

DATE/TIME OF TEST: 08/07/2011 23:04:02

---

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.077	22:58
Air Blank	0.000	22:59
Subject Sample #1	0.206	22:59
Breath Volume	1.750	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample #2	0.210	23:02
Breath Volume	2.734	23:02
Air Blank	0.000	23:03
Control Test	0.078	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:04

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, SELENA, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEE, SELENA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 08/08/2011 03:12:54

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.077	03:07
Air Blank	0.000	03:07
Subject Sample #1	0.105	03:08
Breath Volume	1.832	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample #2	0.100	03:11
Breath Volume	2.093	03:11
Air Blank	0.000	03:12
Control Test	0.077	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, SELENA, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEE, SELENA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: MEESE

DATE/TIME OF TEST: 08/08/2011 19:48:25

---

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.076	19:43
Air Blank	0.000	19:43
Subject Sample #1	0.000	19:43
Breath Volume	1.812	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample #2	0.000	19:47
Breath Volume	1.437	19:47
Air Blank	0.000	19:47
Control Test	0.077	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELSAID, MARK, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: FETTER, JEFF  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MONASTERO

DATE/TIME OF TEST: 08/09/2011 23:56:46

---

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.077	23:51
Air Blank	0.000	23:51
Subject Sample #1	0.124	23:52
Breath Volume	2.050	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample #2	0.125	23:55
Breath Volume	1.269	23:55
Air Blank	0.000	23:55
Control Test	0.077	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMBRIA, ANDREW, R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS, GEORGE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PEACE

DATE/TIME OF TEST: 08/10/2011 03:17:26

---

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.076	03:11
Air Blank	0.000	03:11
Subject Sample #1	0.208	03:12
Breath Volume	2.875	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.206	03:15
Breath Volume	2.914	03:15
Air Blank	0.000	03:16
Control Test	0.077	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ROWE, BRANDON  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/11/2011 03:44:09

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.076	03:37
Air Blank	0.000	03:38
Subject Sample #1	0.151	03:39
Breath Volume	2.289	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample #2	0.147	03:42
Breath Volume	2.492	03:42
Air Blank	0.000	03:43
Control Test	0.076	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RONGISH, SHERRY, R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: RONGISH, SHERRY  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 14:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: NEWKIRK

DATE/TIME OF TEST: 08/12/2011 15:56:35

---

Diagnostics Check	OK	15:49
Air Blank	0.000	15:50
Control Test	0.078	15:50
Air Blank	0.000	15:51
Subject Sample #1	0.083	15:51
Breath Volume	2.648	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample #2	0.083	15:55
Breath Volume	1.957	15:55
Air Blank	0.000	15:55
Control Test	0.077	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELSAID, MARK, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DUNCAN, NICK  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/13/2011 23:13:49

---

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.077	23:09	
Air Blank	0.000	23:10	
Subject Sample #1	0.228	23:12	
Breath Volume	1.296	23:12	
Air Blank		23:13	Purge Fail
Air Blank	0.000	23:13	

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELSAID, MARK, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DUNCAN, NICK  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/13/2011 23:21:47

---

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.076	23:16
Air Blank	0.000	23:16
Subject Sample #1	0.203	23:17
Breath Volume	2.000	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample #2	0.213	23:20
Breath Volume	1.656	23:20
Air Blank	0.000	23:20
Control Test	0.077	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELSAID, MARK, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SHERMAN, KEN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SANCHEZ NARANJO

DATE/TIME OF TEST: 08/14/2011 01:25:33

---

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.078	01:20
Air Blank	0.000	01:20
Subject Sample #1	0.159	01:21
Breath Volume	2.773	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample #2	0.158	01:24
Breath Volume	2.246	01:24
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMBRIA, ANDREW, R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS, GEORGE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DETRICK

DATE/TIME OF TEST: 08/14/2011 04:13:04

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.077	04:07
Air Blank	0.000	04:07
Subject Sample #1	0.136	04:08
Breath Volume	1.398	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.132	04:11
Breath Volume	3.613	04:11
Air Blank	0.000	04:12
Control Test	0.077	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:13

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TINEO, NELSON, A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: TINEO, NELSON  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 04:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: TORRES CINTRON

DATE/TIME OF TEST: 08/14/2011 06:09:10

---

Diagnostics Check	OK	06:03
Air Blank	0.000	06:03
Control Test	0.077	06:03
Air Blank	0.000	06:04
Subject Sample #1	0.162	06:04
Breath Volume	2.335	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample #2	0.160	06:07
Breath Volume	1.199	06:07
Air Blank	0.000	06:08
Control Test	0.077	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLA, ADRIANA, M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BEATTY, CHARLES  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/15/2011 04:33:37

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.077	04:28
Air Blank	0.000	04:28
Subject Sample #1	0.134	04:29
Breath Volume	1.871	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample #2	0.133	04:32
Breath Volume	1.578	04:32
Air Blank	0.000	04:32
Control Test	0.077	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GORDON, J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MASSE

DATE/TIME OF TEST: 08/17/2011 04:07:30

---

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.077	04:05	
Air Blank	0.000	04:05	
Subject Sample #1		04:06	RFI Detect
Breath Volume	0.000	04:06	
Air Blank	0.000	04:06	
Control Test	0.077	04:06	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GORDON, J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MASSE

DATE/TIME OF TEST: 08/17/2011 04:23:26

---

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.077	04:09	
Air Blank	0.000	04:09	
Subject Sample #1	0.261	04:13	Volume Not Met
Breath Volume	0.082	04:13	
Air Blank	0.000	04:13	
Air Blank	0.000	04:15	
Subject Sample #2	0.264	04:16	No .020 agreement
Breath Volume	1.375	04:16	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample #3	0.240	04:21	No .020 agreement
Breath Volume	1.375	04:21	
Air Blank	0.000	04:22	
Control Test	0.079	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURPHY, WILLIAM, K  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MURPHY, WILLIAM  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: SIMONELLI

DATE/TIME OF TEST: 08/18/2011 00:21:30

---

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.077	00:09	
Air Blank	0.000	00:10	
Subject Sample #1	0.133	00:13	Volume Not Met
Breath Volume	0.648	00:13	
Air Blank	0.000	00:14	
Air Blank	0.000	00:15	
Subject Sample #2	0.157	00:16	
Breath Volume	1.648	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample #3	0.162	00:19	
Breath Volume	1.632	00:19	
Air Blank	0.000	00:20	
Control Test	0.076	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

---

Data Download Date/Time 08/18/2011 09:42

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAFLIN, DAVID, S  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MILLICAN, JOHN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01011080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 09:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:58

SUBJECT'S LAST NAME: COLDIRON

DATE/TIME OF TEST: 08/18/2011 11:27:57

---

Diagnostics Check	OK	11:21
Air Blank	0.000	11:21
Control Test	0.075	11:22
Air Blank	0.000	11:22
Subject Sample #1	0.150	11:23
Breath Volume	2.363	11:23
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample #2	0.154	11:26
Breath Volume	2.453	11:26
Air Blank	0.000	11:27
Control Test	0.075	11:27
Air Blank	0.000	11:27
Diagnostics Check	OK	11:27

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENEMAN, LESA,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BRENEMAN, LESA  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01011080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 11:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: HENDRON

DATE/TIME OF TEST: 08/18/2011 12:01:33

---

Diagnostics Check	OK	11:55
Air Blank	0.000	11:55
Control Test	0.075	11:55
Air Blank	0.000	11:56
Subject Sample #1	0.000	11:56
Breath Volume	1.554	11:56
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample #2	0.000	12:00
Breath Volume	1.769	12:00
Air Blank	0.000	12:00
Control Test	0.076	12:00
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, SELENA, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEE, SELENA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01011080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 08/18/2011 20:01:28

---

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.076	19:56
Air Blank	0.000	19:56
Subject Sample #1	0.000	19:57
Breath Volume	3.187	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample #2	0.000	20:00
Breath Volume	3.410	20:00
Air Blank	0.000	20:00
Control Test	0.077	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONIECZKI, DOMINIC, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KONIECZKI, DOMINIC  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01011080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/19/2011 03:05:24

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.076	02:59
Air Blank	0.000	02:59
Subject Sample #1	0.111	03:00
Breath Volume	1.605	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.115	03:03
Breath Volume	1.347	03:03
Air Blank	0.000	03:04
Control Test	0.076	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GUERREIRO, D  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 11:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 08/20/2011 01:28:56

---

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.078	01:19	
Air Blank	0.000	01:19	
Subject Sample #1	0.170	01:22	Volume Not Met
Breath Volume	0.054	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample #2		01:27	Subject Test Refused
Breath Volume	0.171	01:27	
Air Blank	0.000	01:28	
Control Test	0.076	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KONDAS, STEVE  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 08/20/2011 01:43:35

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.076	01:37
Air Blank	0.000	01:37
Subject Sample #1	0.085	01:38
Breath Volume	1.511	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample #2	0.086	01:42
Breath Volume	2.125	01:42
Air Blank	0.000	01:42
Control Test	0.076	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMBRIA, ANDREW, R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS, GEORGE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DWYER

DATE/TIME OF TEST: 08/20/2011 03:22:53

---

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.077	03:17
Air Blank	0.000	03:17
Subject Sample #1	0.173	03:18
Breath Volume	2.160	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample #2	0.170	03:21
Breath Volume	2.296	03:21
Air Blank	0.000	03:21
Control Test	0.077	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMBRIA, ANDREW, R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS, GEORGE  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 05:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 08/21/2011 06:07:39

---

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.077	06:01
Air Blank	0.000	06:01
Subject Sample #1	0.164	06:03
Breath Volume	1.746	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample #2	0.167	06:06
Breath Volume	2.035	06:06
Air Blank	0.000	06:06
Control Test	0.077	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENEMAN, LESA,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BRENEMAN, LESA  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: DUKAS

DATE/TIME OF TEST: 08/23/2011 04:24:09

---

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.077	04:18
Air Blank	0.000	04:18
Subject Sample #1	0.070	04:19
Breath Volume	2.296	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample #2	0.069	04:22
Breath Volume	1.695	04:22
Air Blank	0.000	04:23
Control Test	0.077	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIEROT, JEFFREY, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PIEROT, JEFFREY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 17:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: KNERAM

DATE/TIME OF TEST: 08/24/2011 19:03:38

---

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.077	18:57
Air Blank	0.000	18:58
Subject Sample #1	0.140	18:58
Breath Volume	1.164	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample #2	0.129	19:02
Breath Volume	2.097	19:02
Air Blank	0.000	19:02
Control Test	0.077	19:03
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HINTON, D  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ZEPERNICK

DATE/TIME OF TEST: 08/24/2011 23:32:43

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.077	23:27
Air Blank	0.000	23:27
Subject Sample #1	0.165	23:28
Breath Volume	3.980	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample #2	0.162	23:31
Breath Volume	3.843	23:31
Air Blank	0.000	23:31
Control Test	0.076	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIEROT, JEFFREY, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PIEROT, JEFFREY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 16:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: AN

DATE/TIME OF TEST: 08/25/2011 17:41:35

---

Diagnostics Check	OK	17:35
Air Blank	0.000	17:35
Control Test	0.077	17:35
Air Blank	0.000	17:36
Subject Sample #1	0.195	17:36
Breath Volume	2.859	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample #2	0.202	17:40
Breath Volume	3.042	17:40
Air Blank	0.000	17:40
Control Test	0.077	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEWMAN, LESLIE, D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GUERREIRO, D  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MEADORS

DATE/TIME OF TEST: 08/26/2011 01:22:49

---

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.077	01:16
Air Blank	0.000	01:17
Subject Sample #1	0.214	01:17
Breath Volume	1.925	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample #2	0.224	01:21
Breath Volume	2.644	01:21
Air Blank	0.000	01:21
Control Test	0.077	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEE, SELENA, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LEE, SELENA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/26/2011 02:00:42

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.077	01:54
Air Blank	0.000	01:55
Subject Sample #1	0.173	01:55
Breath Volume	2.917	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample #2	0.171	01:59
Breath Volume	2.273	01:59
Air Blank	0.000	01:59
Control Test	0.077	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RONGISH, SHERRY, R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: RONGISH, SHERRY  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KARCH

DATE/TIME OF TEST: 08/27/2011 01:21:40

---

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.077	01:16
Air Blank	0.000	01:16
Subject Sample #1	0.104	01:17
Breath Volume	3.714	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample #2	0.103	01:20
Breath Volume	2.890	01:20
Air Blank	0.000	01:20
Control Test	0.077	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAVOIE, KRYSTAL, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MITCHELL, L  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 17:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 08/27/2011 19:15:33

---

Diagnostics Check	OK	19:06
Air Blank	0.000	19:06
Control Test	0.076	19:07
Air Blank	0.000	19:07
Subject Sample #1	0.267	19:09
Breath Volume	1.175	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample #2	0.255	19:13
Breath Volume	1.355	19:13
Air Blank	0.000	19:14
Control Test	0.076	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULHERN, JAMES, A

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: SMYRNIOS, GEORGE

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35710080A3

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 02:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/28/2011 03:37:48

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.076	03:32
Air Blank	0.000	03:32
Subject Sample #1	0.143	03:33
Breath Volume	1.785	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample #2	0.140	03:36
Breath Volume	2.558	03:36
Air Blank	0.000	03:36
Control Test	0.076	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRYOR, M,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DEL PINO, R  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 14:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: VASQUEZ ORTIZ

DATE/TIME OF TEST: 08/28/2011 15:02:53

---

Diagnostics Check	OK	14:55	
Air Blank	0.000	14:55	
Control Test	0.076	14:56	
Air Blank	0.000	14:56	
Subject Sample #1	0.214	14:59	Volume Not Met
Breath Volume	0.160	14:59	
Air Blank	0.000	15:00	
Air Blank	0.000	15:02	
Subject Sample #2		15:02	Sequence Aborted
Breath Volume	0.000	15:02	
Air Blank	0.000	15:02	

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRYOR, M,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DEL PINO, R  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 14:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: VASQUEZ ORTIZ

DATE/TIME OF TEST: 08/28/2011 15:07:35

---

Diagnostics Check	OK	15:04	
Air Blank	0.000	15:05	
Control Test	0.077	15:05	
Air Blank	0.000	15:06	
Subject Sample #1		15:06	Subject Test Refused
Breath Volume	0.000	15:06	
Air Blank	0.000	15:06	
Control Test	0.076	15:06	
Air Blank	0.000	15:07	
Diagnostics Check	OK	15:07	

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAVOIE, KRYSTAL, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HILL, S  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: NOWLIN

DATE/TIME OF TEST: 08/29/2011 01:16:24

---

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.076	01:10
Air Blank	0.000	01:11
Subject Sample #1	0.150	01:11
Breath Volume	3.789	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample #2	0.151	01:14
Breath Volume	2.734	01:14
Air Blank	0.000	01:15
Control Test	0.075	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENEMAN, LESA,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BRENEMAN, LESA  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/18/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35710080A3  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 10:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:57

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 08/30/2011 11:24:07

---

Diagnostics Check	OK	11:17
Air Blank	0.000	11:18
Control Test	0.077	11:18
Air Blank	0.000	11:19
Subject Sample #1	0.000	11:19
Breath Volume	3.054	11:19
Air Blank	0.000	11:20
Air Blank	0.000	11:22
Subject Sample #2	0.000	11:22
Breath Volume	2.917	11:22
Air Blank	0.000	11:23
Control Test	0.078	11:23
Air Blank	0.000	11:24
Diagnostics Check	OK	11:24

---

Data Download Date/Time 09/28/2011 10:59

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAVN, SUSANNE, M

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: RAVN, SUSANNE

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 07/09/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG004701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 02:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LASHOMB

DATE/TIME OF TEST: 08/04/2011 03:50:45

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:45
Air Blank	0.000	03:45
Subject Sample #1	0.191	03:46
Breath Volume	2.828	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample #2	0.185	03:49
Breath Volume	4.050	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time 08/11/2011 14:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WAGNER, ROBERT, F

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: HARRI, SELVESTER

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 07/09/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG004701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 21:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: ROWSEY

DATE/TIME OF TEST: 08/04/2011 22:15:13

---

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.077	22:09
Air Blank	0.000	22:10
Subject Sample #1	0.207	22:10
Breath Volume	1.281	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample #2	0.208	22:13
Breath Volume	1.105	22:13
Air Blank	0.000	22:14
Control Test	0.076	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

---

Data Download Date/Time 08/11/2011 14:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WAGNER, ROBERT, F

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: RAVN, SUSANNE

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 07/09/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG004701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 23:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PASSAMONTE

DATE/TIME OF TEST: 08/11/2011 00:17:38

---

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:12
Subject Sample #1	0.147	00:13
Breath Volume	1.644	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample #2	0.157	00:16
Breath Volume	1.945	00:16
Air Blank	0.000	00:16
Control Test	0.077	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

---

Data Download Date/Time 08/11/2011 14:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WAGNER, ROBERT, F

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: WAGNER, ROBERT

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 07/09/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG004701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 02:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GUDMUNDSSON

DATE/TIME OF TEST: 08/11/2011 02:58:53

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.113	02:54
Breath Volume	1.960	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.111	02:57
Breath Volume	1.714	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time 08/11/2011 14:30

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WAGNER, ROBERT, F

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: WAGNER, ROBERT

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 08/11/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG004701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 22:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HOGLUND

DATE/TIME OF TEST: 08/11/2011 23:20:11

---

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.078	23:14
Air Blank	0.000	23:14
Subject Sample #1	0.170	23:15
Breath Volume	1.402	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample #2	0.182	23:18
Breath Volume	1.625	23:18
Air Blank	0.000	23:19
Control Test	0.077	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

---

Data Download Date/Time 09/16/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARRITY, THOMAS, A

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: GARRITY, THOMAS

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 08/11/2011

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: AG004701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 22:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 08/13/2011 23:03:50

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.077	22:57
Air Blank	0.000	22:57
Subject Sample #1	0.091	22:59
Breath Volume	1.753	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample #2	0.090	23:02
Breath Volume	1.812	23:02
Air Blank	0.000	23:02
Control Test	0.078	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

---

Data Download Date/Time 09/16/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOROS, LINDA, L

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: BROOKS, TRAVIS

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 08/11/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG004701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 22:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: YAMRICK

DATE/TIME OF TEST: 08/18/2011 23:35:13

---

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.076	23:29
Air Blank	0.000	23:29
Subject Sample #1	0.013	23:29
Breath Volume	1.500	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample #2	0.015	23:33
Breath Volume	1.875	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

---

Data Download Date/Time 09/16/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUSANTE, ANTHONY,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MUSANTE, ANTHONY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:21

SUBJECT'S LAST NAME: MAULLER

DATE/TIME OF TEST: 08/23/2011 13:50:31

---

Diagnostics Check	OK	13:49	
Air Blank	0.000	13:49	
Control Test	0.073	13:49	Control Outside Tolerance
Air Blank	0.000	13:50	

---

Data Download Date/Time 09/16/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUSANTE, ANTHONY,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MUSANTE, ANTHONY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:21

SUBJECT'S LAST NAME: MAULLER

DATE/TIME OF TEST: 08/23/2011 13:54:45

---

Diagnostics Check	OK	13:53	
Air Blank	0.000	13:53	
Control Test	0.074	13:54	Control Outside Tolerance
Air Blank	0.000	13:54	

---

Data Download Date/Time 09/16/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUSANTE, ANTHONY,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MUSANTE, ANTHONY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:21

SUBJECT'S LAST NAME: MAULLER

DATE/TIME OF TEST: 08/23/2011 14:08:02

---

Diagnostics Check	OK	14:01
Air Blank	0.000	14:01
Control Test	0.076	14:01
Air Blank	0.000	14:02
Subject Sample #1	0.272	14:03
Breath Volume	2.761	14:03
Air Blank	0.000	14:03
Air Blank	0.000	14:05
Subject Sample #2	0.272	14:06
Breath Volume	2.972	14:06
Air Blank	0.000	14:07
Control Test	0.077	14:07
Air Blank	0.000	14:07
Diagnostics Check	OK	14:08

---

Data Download Date/Time 09/16/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARRITY, THOMAS, A

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: LEISTER, LAURA

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 08/11/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG004701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: PROB/PAROL

ARREST TIME: 01:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 08/27/2011 01:27:04

---

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample #1	0.149	01:22
Breath Volume	2.054	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample #2	0.130	01:25
Breath Volume	1.148	01:25
Air Blank	0.000	01:26
Control Test	0.077	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:27

---

Data Download Date/Time 09/16/2011 10:14

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAVN, SUSANNE, M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: RAVN, SUSANNE  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG004701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 08/31/2011 03:20:01

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Subject Sample #1	0.151	03:15
Breath Volume	1.488	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample #2	0.153	03:18
Breath Volume	2.617	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:20

---

Data Download Date/Time 09/16/2011 10:14

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEABODY, MICHAEL, F  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: HINCHLIFFE, GREG  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BOURQUE

DATE/TIME OF TEST: 08/21/2011 00:20:51

---

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.148	00:20	Control Outside Tolerance
Air Blank	0.000	00:20	

---

Data Download Date/Time 09/01/2011 16:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEABODY, MICHAEL, F  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: HINCHLIFFE, GREGG  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: 89231

DATE/TIME OF TEST: 08/21/2011 00:30:53

---

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Subject Sample #1	0.159	00:26
Breath Volume	2.039	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample #2	0.161	00:29
Breath Volume	1.367	00:29
Air Blank	0.000	00:29
Control Test	0.078	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time 09/01/2011 16:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEABODY, MICHAEL, F  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: GERMAN, STEPHEN  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/21/2011 02:11:36

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:06
Air Blank	0.000	02:06
Subject Sample #1	0.066	02:06
Breath Volume	2.761	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.067	02:10
Breath Volume	2.460	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time 09/01/2011 16:03

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS, TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS, TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 08/02/2011 02:40:14

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Subject Sample #1	0.135	02:35
Breath Volume	2.175	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample #2	0.136	02:38
Breath Volume	2.300	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STUBBE, PATRICK, J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: HEBERBRAND, K  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 08/03/2011 23:34:45

---

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.078	23:32	
Air Blank	0.000	23:33	
Subject Sample #1		23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.078	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COMBS, SEAN, F  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: FOOTS, D.  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BEIERS

DATE/TIME OF TEST: 08/04/2011 23:09:08

---

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.078	23:03
Air Blank	0.000	23:04
Subject Sample #1	0.156	23:04
Breath Volume	1.468	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample #2	0.165	23:07
Breath Volume	1.257	23:07
Air Blank	0.000	23:08
Control Test	0.077	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS, TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS, TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BIGGART

DATE/TIME OF TEST: 08/06/2011 00:27:30

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.078	00:21
Air Blank	0.000	00:22
Subject Sample #1	0.081	00:22
Breath Volume	3.765	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample #2	0.075	00:26
Breath Volume	4.445	00:26
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PADILLA, MARIA,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KNOWLTON, J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: TOMAS FUNES

DATE/TIME OF TEST: 08/06/2011 01:39:37

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.077	01:31
Air Blank	0.000	01:31
Subject Sample #1	0.252	01:32
Breath Volume	3.082	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample #2	0.229	01:35
Breath Volume	2.242	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample #3	0.250	01:38
Breath Volume	2.652	01:38
Air Blank	0.000	01:38
Control Test	0.078	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEFFIN, DOUGLAS, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: LEFFIN, DOUGLAS  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 17:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: CRUZ CRUZ

DATE/TIME OF TEST: 08/06/2011 18:29:50

---

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.077	18:24
Air Blank	0.000	18:24
Subject Sample #1	0.226	18:24
Breath Volume	1.652	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample #2	0.243	18:28
Breath Volume	1.523	18:28
Air Blank	0.000	18:28
Control Test	0.078	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PULIZZOTTO, S  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 08/06/2011 21:12:48

---

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.077	21:06
Air Blank	0.000	21:07
Subject Sample #1	0.180	21:08
Breath Volume	1.714	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample #2	0.181	21:11
Breath Volume	1.699	21:11
Air Blank	0.000	21:11
Control Test	0.077	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS, TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS, TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: STROMBERG

DATE/TIME OF TEST: 08/07/2011 01:50:32

---

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.078	01:45
Air Blank	0.000	01:45
Subject Sample #1	0.182	01:46
Breath Volume	1.636	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample #2	0.178	01:49
Breath Volume	1.238	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CATANI, J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 08/07/2011 02:47:25

---

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Subject Sample #1	0.000	02:42
Breath Volume	3.804	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample #2	0.000	02:46
Breath Volume	3.917	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CATANI, J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 08/07/2011 23:14:20

---

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.077	23:07
Air Blank	0.000	23:08
Subject Sample #1	0.173	23:09
Breath Volume	2.820	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample #2	0.169	23:12
Breath Volume	2.039	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUARTE, ROGER,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CATANI, J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 08/09/2011 00:51:46

---

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:45
Subject Sample #1	0.054	00:46
Breath Volume	1.425	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample #2	0.055	00:50
Breath Volume	1.261	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ, RUBEN, J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: THOMAS, CHRIS  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 17:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: PRIZZI

DATE/TIME OF TEST: 08/10/2011 17:49:00

---

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.077	17:40
Air Blank	0.000	17:41
Subject Sample #1	0.097	17:42
Breath Volume	1.203	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample #2	0.095	17:47
Breath Volume	1.828	17:47
Air Blank	0.000	17:48
Control Test	0.079	17:48
Air Blank	0.000	17:48
Diagnostics Check	OK	17:49

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BENOIT, J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BASSETTE

DATE/TIME OF TEST: 08/11/2011 03:48:29

---

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:42
Subject Sample #1	0.045	03:44
Breath Volume	1.113	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample #2	0.049	03:46
Breath Volume	1.796	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MILLER, K  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RUIZ HERNANDEZ

DATE/TIME OF TEST: 08/12/2011 04:29:47

---

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.078	04:23
Air Blank	0.000	04:24
Subject Sample #1	0.188	04:25
Breath Volume	2.136	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample #2	0.186	04:28
Breath Volume	1.726	04:28
Air Blank	0.000	04:28
Control Test	0.078	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELANEY, PATRICK, T  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SHICKFUS, R  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: NINO

DATE/TIME OF TEST: 08/13/2011 03:31:21

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:25
Subject Sample #1	0.165	03:26
Breath Volume	2.281	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.173	03:29
Breath Volume	2.882	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYSON, GREGORY, D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: HEBEBRAND, K  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: TROCHESSETT

DATE/TIME OF TEST: 08/13/2011 04:21:39

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Subject Sample #1	0.126	04:16
Breath Volume	1.898	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample #2	0.124	04:20
Breath Volume	1.804	04:20
Air Blank	0.000	04:20
Control Test	0.078	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: RIEDY, J  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 08/16/2011 05:38:29

---

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.078	05:32
Air Blank	0.000	05:33
Subject Sample #1	0.000	05:34
Breath Volume	1.710	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample #2	0.000	05:37
Breath Volume	2.890	05:37
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOLIEK, P, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BOLIEK, P  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GISO

DATE/TIME OF TEST: 08/17/2011 04:12:29

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:06
Subject Sample #1	0.154	04:07
Breath Volume	1.496	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.154	04:10
Breath Volume	2.109	04:10
Air Blank	0.000	04:11
Control Test	0.077	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MULCAHEY, CHRISTOPHER, M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SHATTUCK, M  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 08/17/2011 23:06:39

---

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.079	22:53	
Air Blank	0.000	22:54	
Subject Sample #1	0.154	22:57	Volume Not Met
Breath Volume	0.925	22:57	
Air Blank	0.000	22:58	
Air Blank	0.000	23:00	
Subject Sample #2	0.164	23:01	
Breath Volume	1.222	23:01	
Air Blank	0.000	23:02	
Air Blank	0.000	23:04	
Subject Sample #3	0.159	23:05	
Breath Volume	1.121	23:05	
Air Blank	0.000	23:05	
Control Test	0.077	23:06	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COMBS, SEAN, F  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SHATTUCK, M.  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RUETER

DATE/TIME OF TEST: 08/18/2011 02:42:57

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.078	02:37
Air Blank	0.000	02:37
Subject Sample #1	0.132	02:38
Breath Volume	3.738	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.133	02:41
Breath Volume	3.910	02:41
Air Blank	0.000	02:42
Control Test	0.078	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: LINKENBERG, F  
ARRESTING OFFICER AGENCY: Marco Island PD  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 13:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: JAYNES

DATE/TIME OF TEST: 08/18/2011 15:16:21

---

Diagnostics Check	OK	15:04	
Air Blank	0.000	15:04	
Control Test	0.077	15:05	
Air Blank	0.000	15:05	
Subject Sample #1	0.184	15:08	Volume Not Met
Breath Volume	0.945	15:08	
Air Blank	0.000	15:09	
Air Blank	0.000	15:11	
Subject Sample #2	0.176	15:11	
Breath Volume	2.359	15:11	
Air Blank	0.000	15:12	
Air Blank	0.000	15:14	
Subject Sample #3	0.176	15:14	
Breath Volume	2.445	15:14	
Air Blank	0.000	15:15	
Control Test	0.077	15:15	
Air Blank	0.000	15:16	
Diagnostics Check	OK	15:16	

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BOOTH, J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: GAMEZ

DATE/TIME OF TEST: 08/21/2011 06:12:36

---

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.078	06:06
Air Blank	0.000	06:07
Subject Sample #1	0.189	06:08
Breath Volume	1.687	06:08
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample #2	0.175	06:11
Breath Volume	1.500	06:11
Air Blank	0.000	06:11
Control Test	0.078	06:12
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOLIEK, P, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BOLIEK, P  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/24/2011 23:19:06

---

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Subject Sample #1	0.228	23:14
Breath Volume	2.929	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample #2	0.221	23:17
Breath Volume	2.039	23:17
Air Blank	0.000	23:18
Control Test	0.077	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIBSON, J,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: GIBSON, J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 08/25/2011 01:57:05

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Subject Sample #1	0.033	01:52
Breath Volume	2.226	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample #2	0.033	01:55
Breath Volume	2.195	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:57

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY, ROSS, C  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BORCHERS, JOHN  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: FIGA

DATE/TIME OF TEST: 08/25/2011 02:44:41

---

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.242	02:39
Breath Volume	2.347	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.234	02:43
Breath Volume	3.921	02:43
Air Blank	0.000	02:43
Control Test	0.077	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CATANI, J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STRAUB

DATE/TIME OF TEST: 08/25/2011 03:28:26

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.078	03:22
Air Blank	0.000	03:23
Subject Sample #1	0.193	03:23
Breath Volume	1.832	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.200	03:26
Breath Volume	3.492	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLACKWOOD, GARFIELD, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PISANNO, F  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 08/26/2011 02:32:17

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Subject Sample #1	0.310	02:27
Breath Volume	3.089	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample #2	0.300	02:30
Breath Volume	1.933	02:30
Air Blank	0.000	02:31
Control Test	0.077	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPCKINS, J,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CASEY, J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: L'HEUREUX

DATE/TIME OF TEST: 08/27/2011 01:57:34

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:49
Subject Sample #1	0.243	01:49
Breath Volume	1.273	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample #2	0.271	01:52
Breath Volume	1.421	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample #3	0.273	01:56
Breath Volume	2.066	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MULCAHEY, CHRISTOPHER, M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SHATTUCK, M  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CLARO

DATE/TIME OF TEST: 08/28/2011 00:17:37

---

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.078	00:12
Air Blank	0.000	00:12
Subject Sample #1	0.000	00:13
Breath Volume	2.171	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample #2	0.000	00:16
Breath Volume	1.683	00:16
Air Blank	0.000	00:16
Control Test	0.078	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUARTE, ROGER,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SHAKURO, E  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CERON

DATE/TIME OF TEST: 08/28/2011 01:47:03

---

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Subject Sample #1	0.197	01:35	
Breath Volume	2.085	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:37	
Subject Sample #2	0.182	01:40	Volume Not Met
Breath Volume	0.875	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample #3	0.192	01:45	
Breath Volume	2.332	01:45	
Air Blank	0.000	01:46	
Control Test	0.078	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:47	

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYSON, GREGORY, D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CATANI, J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: GOULET

DATE/TIME OF TEST: 08/29/2011 00:25:41

---

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:20
Subject Sample #1	0.000	00:21
Breath Volume	2.183	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample #2	0.000	00:24
Breath Volume	2.515	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOLIEK, P, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BOLIEK, P  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: NAPOLES

DATE/TIME OF TEST: 08/30/2011 01:32:52

---

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.078	01:27
Air Blank	0.000	01:28
Subject Sample #1	0.000	01:28
Breath Volume	3.347	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample #2	0.000	01:31
Breath Volume	3.792	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOLIEK, P, A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CATANI, J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BROOKS-ROSALES

DATE/TIME OF TEST: 08/30/2011 02:19:55

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:14
Subject Sample #1	0.158	02:15
Breath Volume	1.675	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample #2	0.152	02:18
Breath Volume	1.859	02:18
Air Blank	0.000	02:19
Control Test	0.077	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS, TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS, TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 08/01/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770940  
DRY GAS EXPIRY DATE: 06/02/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: HERZOG

DATE/TIME OF TEST: 08/30/2011 04:22:26

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.078	04:20	
Air Blank	0.000	04:20	
Subject Sample #1		04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

---

Data Download Date/Time 09/04/2011 09:06

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, JOHN, H  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CIGNARALE, PAUL  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 18:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 08/06/2011 19:40:12

---

Diagnostics Check	OK	19:34
Air Blank	0.000	19:34
Control Test	0.078	19:35
Air Blank	0.000	19:35
Subject Sample #1	0.000	19:35
Breath Volume	2.988	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample #2	0.000	19:38
Breath Volume	2.945	19:38
Air Blank	0.000	19:39
Control Test	0.078	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

---

Data Download Date/Time 08/13/2011 14:39

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLEDZINSKI, ROBERT, A  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GAY, A  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: VOSBURGH

DATE/TIME OF TEST: 08/13/2011 02:58:14

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.077	02:52
Air Blank	0.000	02:52
Subject Sample #1	0.179	02:53
Breath Volume	1.304	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample #2	0.185	02:56
Breath Volume	1.914	02:56
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time 08/13/2011 14:39

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALTIMORE, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MUSGROVE, DAVID  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 16:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 08/13/2011 17:59:34

---

Diagnostics Check	OK	17:53
Air Blank	0.000	17:54
Control Test	0.078	17:54
Air Blank	0.000	17:55
Subject Sample #1	0.000	17:55
Breath Volume	2.011	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample #2	0.000	17:58
Breath Volume	1.902	17:58
Air Blank	0.000	17:58
Control Test	0.078	17:58
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHARTON, DANIEL, A  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WHARTON, DANIEL  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: OTHER  
ARREST TIME: 02:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 08/14/2011 02:49:12

---

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.077	02:44	
Air Blank	0.000	02:44	
Subject Sample #1	0.203	02:45	
Breath Volume	2.140	02:45	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample #2		02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:48	
Control Test	0.078	02:48	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALTIMORE, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BALTIMORE, KEVIN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 02:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 08/14/2011 10:54:13

---

Diagnostics Check	OK	10:48
Air Blank	0.000	10:48
Control Test	0.078	10:48
Air Blank	0.000	10:49
Subject Sample #1	0.080	10:49
Breath Volume	2.421	10:49
Air Blank	0.000	10:50
Air Blank	0.000	10:52
Subject Sample #2	0.078	10:52
Breath Volume	1.871	10:52
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:54
Diagnostics Check	OK	10:54

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALTIMORE, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BALTIMORE, KEVIN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 02:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 08/14/2011 14:12:22

---

Diagnostics Check	OK	14:06
Air Blank	0.000	14:06
Control Test	0.077	14:07
Air Blank	0.000	14:07
Subject Sample #1	0.024	14:08
Breath Volume	2.207	14:08
Air Blank	0.000	14:08
Air Blank	0.000	14:10
Subject Sample #2	0.022	14:10
Breath Volume	1.605	14:10
Air Blank	0.000	14:11
Control Test	0.078	14:11
Air Blank	0.000	14:12
Diagnostics Check	OK	14:12

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHENARD, JAMES, H  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SCHOEPFER, T  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 10:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: GAYLORD

DATE/TIME OF TEST: 08/15/2011 12:26:29

---

Diagnostics Check	OK	12:20
Air Blank	0.000	12:20
Control Test	0.078	12:21
Air Blank	0.000	12:21
Subject Sample #1	0.065	12:22
Breath Volume	2.273	12:22
Air Blank	0.000	12:22
Air Blank	0.000	12:24
Subject Sample #2	0.064	12:25
Breath Volume	2.015	12:25
Air Blank	0.000	12:25
Control Test	0.078	12:25
Air Blank	0.000	12:26
Diagnostics Check	OK	12:26

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON, CELESTE, M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WESTMORELAND, L  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: OTT

DATE/TIME OF TEST: 08/18/2011 02:36:24

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Subject Sample #1	0.093	02:31
Breath Volume	2.414	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample #2	0.083	02:34
Breath Volume	1.503	02:34
Air Blank	0.000	02:35
Control Test	0.077	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GOERL, B  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 08/18/2011 03:14:56

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.077	03:08
Air Blank	0.000	03:09
Subject Sample #1	0.114	03:09
Breath Volume	1.128	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.111	03:13
Breath Volume	2.101	03:13
Air Blank	0.000	03:14
Control Test	0.077	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GOERL, B  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 08/19/2011 00:45:22

---

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Subject Sample #1	0.164	00:40
Breath Volume	4.082	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample #2	0.154	00:43
Breath Volume	2.816	00:43
Air Blank	0.000	00:44
Control Test	0.077	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SCOTT, JOHN, H

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: COMBS, L

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 21:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ROUTES

DATE/TIME OF TEST: 08/19/2011 23:13:40

---

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.078	23:09	
Air Blank	0.000	23:10	
Subject Sample #1		23:12	Subject Test Refused
Breath Volume	0.070	23:12	
Air Blank	0.000	23:12	
Control Test	0.079	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SCOTT, JOHN, H

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: MENDLESON, M

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 08/20/2011 02:25:42

---

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Subject Sample #1	0.084	02:21
Breath Volume	2.480	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:23
Subject Sample #2	0.080	02:24
Breath Volume	1.976	02:24
Air Blank	0.000	02:24
Control Test	0.078	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, JOHN, H  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MENDELSON, MICHAEL  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CRAPANZANO

DATE/TIME OF TEST: 08/21/2011 03:27:02

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.077	03:21
Air Blank	0.000	03:22
Subject Sample #1	0.094	03:22
Breath Volume	2.941	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample #2	0.090	03:25
Breath Volume	2.339	03:25
Air Blank	0.000	03:26
Control Test	0.077	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STJEAN, SCOTT, M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: N/A, N/A  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: WAUGH

DATE/TIME OF TEST: 08/21/2011 17:26:51

---

Diagnostics Check	OK	17:19
Air Blank	0.000	17:20
Control Test	0.079	17:20
Air Blank	0.000	17:21
Subject Sample #1	0.268	17:22
Breath Volume	3.957	17:22
Air Blank	0.000	17:22
Air Blank	0.000	17:24
Subject Sample #2	0.269	17:25
Breath Volume	2.890	17:25
Air Blank	0.000	17:25
Control Test	0.078	17:26
Air Blank	0.000	17:26
Diagnostics Check	OK	17:26

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, JOSEPH, E  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BARLOW, JOSEPH  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: WAUGH

DATE/TIME OF TEST: 08/21/2011 22:27:49

---

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:22
Subject Sample #1	0.173	22:23
Breath Volume	1.593	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample #2	0.173	22:26
Breath Volume	1.632	22:26
Air Blank	0.000	22:26
Control Test	0.077	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, JOSEPH, E  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BARLOW, JOSEPH  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: WAUGH

DATE/TIME OF TEST: 08/22/2011 02:17:39

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:12
Subject Sample #1	0.137	02:13
Breath Volume	1.820	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.134	02:16
Breath Volume	1.667	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARLOW, JOSEPH, E  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BARLOW, JOSEPH  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: WAUGH

DATE/TIME OF TEST: 08/22/2011 05:24:07

---

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.078	05:18
Air Blank	0.000	05:19
Subject Sample #1	0.093	05:19
Breath Volume	1.226	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample #2	0.094	05:22
Breath Volume	1.773	05:22
Air Blank	0.000	05:23
Control Test	0.078	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALTIMORE, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BALTIMORE, KEVIN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 16:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: WAUGH

DATE/TIME OF TEST: 08/22/2011 11:37:27

---

Diagnostics Check	OK	11:32	
Air Blank	0.000	11:32	
Control Test	0.077	11:33	
Air Blank	0.000	11:33	
Subject Sample #1	0.026	11:34	
Breath Volume	2.578	11:34	
Air Blank	0.000	11:34	
Air Blank	0.000	11:36	
Subject Sample #2	0.025	11:36	
Breath Volume	2.554	11:36	
Air Blank		11:36	Improper Sample
Air Blank	0.000	11:37	

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALTIMORE, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BALTIMORE, KEVIN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: ADMIN  
ARREST TIME: 16:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: WAUGH

DATE/TIME OF TEST: 08/22/2011 12:00:55

---

Diagnostics Check	OK	11:54
Air Blank	0.000	11:55
Control Test	0.077	11:55
Air Blank	0.000	11:56
Subject Sample #1	0.021	11:56
Breath Volume	2.609	11:56
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample #2	0.022	11:59
Breath Volume	3.082	11:59
Air Blank	0.000	12:00
Control Test	0.078	12:00
Air Blank	0.000	12:00
Diagnostics Check	OK	12:00

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON, CELESTE, M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: VANARSDALE, JEFFREY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BACANER

DATE/TIME OF TEST: 08/22/2011 21:35:10

---

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Subject Sample #1	0.042	21:30
Breath Volume	2.636	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample #2	0.037	21:33
Breath Volume	1.910	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON, CELESTE, M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MORAGHAN, JOHN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LAVIGNE

DATE/TIME OF TEST: 08/23/2011 03:36:37

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.078	03:31	
Air Blank	0.000	03:31	
Subject Sample #1	0.166	03:32	
Breath Volume	4.089	03:32	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample #2	0.167	03:34	
Breath Volume	4.171	03:34	
Air Blank	0.000	03:35	
Control Test	0.077	03:35	
Air Blank		03:36	RFI Detect
Air Blank	0.000	03:36	

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, JOHN, H  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: COMTOIS, R  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 18:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: SOUTER

DATE/TIME OF TEST: 08/25/2011 19:44:48

---

Diagnostics Check	OK	19:38
Air Blank	0.000	19:39
Control Test	0.078	19:39
Air Blank	0.000	19:39
Subject Sample #1	0.300	19:40
Breath Volume	3.523	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:42
Subject Sample #2	0.291	19:43
Breath Volume	4.046	19:43
Air Blank	0.000	19:43
Control Test	0.076	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLEDZINSKI, ROBERT, A  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: RENZ, D  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/27/2011 00:00:39

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.078	23:55
Air Blank	0.000	23:55
Subject Sample #1	0.158	23:56
Breath Volume	1.867	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample #2	0.161	23:59
Breath Volume	3.015	23:59
Air Blank	0.000	23:59
Control Test	0.076	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON, CELESTE, M  
BREATH TEST OPERATOR AGENCY: Punta Gorda PD  
ARRESTING OFFICER: RENZ, DYLAN  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 08/27/2011 04:12:08

---

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Subject Sample #1	0.151	04:07
Breath Volume	3.835	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.148	04:10
Breath Volume	3.187	04:10
Air Blank	0.000	04:11
Control Test	0.076	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLEY, CDFC RICHARD, S  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: AYERS, HARVEY  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 10:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:14

SUBJECT'S LAST NAME: IZZO

DATE/TIME OF TEST: 08/27/2011 11:44:28

---

Diagnostics Check	OK	11:38
Air Blank	0.000	11:38
Control Test	0.078	11:39
Air Blank	0.000	11:39
Subject Sample #1	0.000	11:40
Breath Volume	2.675	11:40
Air Blank	0.000	11:40
Air Blank	0.000	11:42
Subject Sample #2	0.000	11:43
Breath Volume	2.460	11:43
Air Blank	0.000	11:43
Control Test	0.079	11:43
Air Blank	0.000	11:44
Diagnostics Check	OK	11:44

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN, ROBERT, J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: MARTIN, ROBERT  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06111080A1  
DRY GAS EXPIRY DATE: 03/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: TUTTLE

DATE/TIME OF TEST: 08/30/2011 17:24:59

---

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:21	
Control Test	0.078	17:21	
Air Blank	0.000	17:22	
Subject Sample #1		17:23	Subject Test Refused
Breath Volume	0.000	17:23	
Air Blank	0.000	17:24	
Control Test	0.078	17:24	
Air Blank	0.000	17:24	
Diagnostics Check	OK	17:24	

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAMERON, CELESTE, M

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: WHITE, RYAN

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 08/13/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06111080A1

DRY GAS EXPIRY DATE: 03/01/2013

VIOLATION CODE: DUI

ARREST TIME: 20:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: BRUCKER

DATE/TIME OF TEST: 08/31/2011 22:09:06

---

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.078	22:03
Air Blank	0.000	22:04
Subject Sample #1	0.000	22:04
Breath Volume	4.613	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample #2	0.000	22:07
Breath Volume	4.890	22:07
Air Blank	0.000	22:08
Control Test	0.079	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

---

Data Download Date/Time 09/20/2011 12:07

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BALTIMORE, KEVIN, L

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: WHITE, RYAN

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 10:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 08/03/2011 11:57:59

---

Diagnostics Check	OK	11:51
Air Blank	0.000	11:51
Control Test	0.079	11:52
Air Blank	0.000	11:52
Subject Sample #1	0.298	11:53
Breath Volume	2.238	11:53
Air Blank	0.000	11:54
Air Blank	0.000	11:56
Subject Sample #2	0.300	11:56
Breath Volume	1.941	11:56
Air Blank	0.000	11:57
Control Test	0.078	11:57
Air Blank	0.000	11:57
Diagnostics Check	OK	11:57

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WHARTON, DANIEL, A

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: MUSGROVE, D

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 23:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 08/04/2011 00:00:51

---

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:57	
Air Blank	0.000	23:58	
Subject Sample #1		23:59	Slope Not Met
Breath Volume	1.167	23:59	
Air Blank	0.000	23:59	
Control Test	0.080	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHARTON, DANIEL, A  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MUSGROVE, D  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 670754  
DRY GAS EXPIRY DATE: 09/22/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 08/04/2011 00:09:25

---

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Subject Sample #1	0.097	00:04
Breath Volume	2.304	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample #2	0.088	00:07
Breath Volume	1.179	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MARTINEZ, L  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 670754  
DRY GAS EXPIRY DATE: 09/22/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/04/2011 03:15:50

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample #1	0.168	03:11
Breath Volume	2.250	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:13
Subject Sample #2	0.176	03:14
Breath Volume	1.796	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PATTON, JOHN, G

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: PATTON, JOHN

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: ADMIN

ARREST TIME: 12:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 08/06/2011 13:42:17

---

Diagnostics Check	OK	13:35
Air Blank	0.000	13:35
Control Test	0.079	13:36
Air Blank	0.000	13:36
Subject Sample #1	0.341	13:37
Breath Volume	2.402	13:37
Air Blank	0.000	13:38
Air Blank	0.000	13:40
Subject Sample #2	0.344	13:40
Breath Volume	2.320	13:40
Air Blank	0.000	13:41
Control Test	0.080	13:41
Air Blank	0.000	13:42
Diagnostics Check	OK	13:42

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARLOW, JOSEPH, E

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: BUCHINSKY, W

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: ADMIN

ARREST TIME: 12:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 08/06/2011 19:10:35

---

Diagnostics Check	OK	19:03
Air Blank	0.000	19:03
Control Test	0.078	19:04
Air Blank	0.000	19:04
Subject Sample #1	0.174	19:05
Breath Volume	1.812	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample #2	0.164	19:09
Breath Volume	1.183	19:09
Air Blank	0.000	19:09
Control Test	0.079	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SAFFARAS, CELINA, L

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: JONES, J

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 14:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: HEIDLER

DATE/TIME OF TEST: 08/07/2011 15:23:30

---

Diagnostics Check	OK	15:17
Air Blank	0.000	15:17
Control Test	0.079	15:18
Air Blank	0.000	15:18
Subject Sample #1	0.172	15:19
Breath Volume	3.214	15:19
Air Blank	0.000	15:19
Air Blank	0.000	15:21
Subject Sample #2	0.171	15:22
Breath Volume	3.164	15:22
Air Blank	0.000	15:22
Control Test	0.078	15:22
Air Blank	0.000	15:23
Diagnostics Check	OK	15:23

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHENARD, JAMES, H  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DIAZ, O  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 670754  
DRY GAS EXPIRY DATE: 09/22/2011  
VIOLATION CODE: DUI  
ARREST TIME: 14:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: SLATON

DATE/TIME OF TEST: 08/11/2011 15:38:26

---

Diagnostics Check	OK	15:31
Air Blank	0.000	15:32
Control Test	0.080	15:32
Air Blank	0.000	15:33
Subject Sample #1	0.252	15:34
Breath Volume	2.457	15:34
Air Blank	0.000	15:34
Air Blank	0.000	15:36
Subject Sample #2	0.243	15:36
Breath Volume	2.085	15:36
Air Blank	0.000	15:37
Control Test	0.078	15:37
Air Blank	0.000	15:38
Diagnostics Check	OK	15:38

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARTIN, AMY, L

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: COLEMAN, D

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 670754

DRY GAS EXPIRY DATE: 09/22/2011

VIOLATION CODE: DUI

ARREST TIME: 16:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 08/12/2011 17:52:01

---

Diagnostics Check	OK	17:45
Air Blank	0.000	17:46
Control Test	0.078	17:46
Air Blank	0.000	17:47
Subject Sample #1	0.181	17:47
Breath Volume	2.839	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample #2	0.180	17:50
Breath Volume	2.414	17:50
Air Blank	0.000	17:51
Control Test	0.078	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:52

---

Data Download Date/Time 08/13/2011 15:55

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KEVIN, L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WESTMORELAND, L  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 670754  
DRY GAS EXPIRY DATE: 09/22/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ARMENIO

DATE/TIME OF TEST: 08/13/2011 02:46:47

---

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.077	02:41
Air Blank	0.000	02:41
Subject Sample #1	0.231	02:42
Breath Volume	1.843	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample #2	0.233	02:45
Breath Volume	2.910	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

---

Data Download Date/Time 08/13/2011 15:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAXTER, AUDREY, D  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: TAYLOR, ANDY  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 8177739  
DRY GAS EXPIRY DATE: 08/20/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 08/27/2011 20:22:59

---

Diagnostics Check	OK	20:09	
Air Blank	0.000	20:09	
Control Test	0.075	20:10	
Air Blank	0.000	20:10	
Subject Sample #1		20:13	No Sample Provided
Breath Volume	0.000	20:13	
Air Blank	0.000	20:14	
Air Blank	0.000	20:16	
Subject Sample #2	0.180	20:17	
Breath Volume	1.332	20:17	
Air Blank	0.000	20:18	
Air Blank	0.000	20:20	
Subject Sample #3	0.185	20:21	
Breath Volume	1.304	20:21	
Air Blank	0.000	20:22	
Control Test	0.075	20:22	
Air Blank		20:22	RFI Detect
Air Blank	0.000	20:22	

---

Data Download Date/Time 08/31/2011 15:57

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAUZZI, KRISTIN, G

BREATH TEST OPERATOR AGENCY: Glades County SO

ARRESTING OFFICER: CASSERLY, STEVE

ARRESTING OFFICER AGENCY: Glades County SO

LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 817739

DRY GAS EXPIRY DATE: 08/20/2012

VIOLATION CODE: DUI

ARREST TIME: 21:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WERNE

DATE/TIME OF TEST: 08/21/2011 21:55:54

---

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Subject Sample #1	0.000	21:51
Breath Volume	2.410	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample #2	0.000	21:54
Breath Volume	2.187	21:54
Air Blank	0.000	21:54
Control Test	0.078	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

---

Data Download Date/Time 08/31/2011 16:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAXTER, AUDREY, D  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: TAYLOR, ANDY  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 817739  
DRY GAS EXPIRY DATE: 08/20/2012  
VIOLATION CODE: DUI  
ARREST TIME: 18:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 08/27/2011 20:32:12

---

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.078	20:25
Air Blank	0.000	20:26
Subject Sample #1	0.154	20:27
Breath Volume	1.410	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:29
Subject Sample #2	0.171	20:30
Breath Volume	1.480	20:30
Air Blank	0.000	20:31
Control Test	0.078	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

---

Data Download Date/Time 08/31/2011 16:02

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/05/2011 00:00:49

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.081	23:55
Air Blank	0.000	23:55
Subject Sample #1	0.149	23:56
Breath Volume	4.707	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample #2	0.148	23:59
Breath Volume	4.738	23:59
Air Blank	0.000	23:59
Control Test	0.078	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

---

Data Download Date/Time 08/25/2011 09:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 08/06/2011 00:15:19

---

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.082	00:14	Interferent Detect
Air Blank	0.000	00:15	

---

Data Download Date/Time 08/25/2011 09:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 08/06/2011 00:23:06

---

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.078	00:17
Air Blank	0.000	00:17
Subject Sample #1	0.219	00:18
Breath Volume	4.246	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample #2	0.213	00:21
Breath Volume	4.945	00:21
Air Blank	0.000	00:22
Control Test	0.076	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

---

Data Download Date/Time 08/25/2011 09:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 08/18/2011 03:06:25

---

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.078	03:01	
Air Blank	0.000	03:01	
Subject Sample #1	0.126	03:02	
Breath Volume	3.421	03:02	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample #2	0.128	03:05	
Breath Volume	2.214	03:05	
Air Blank		03:05	Improper Sample
Air Blank	0.000	03:06	

---

Data Download Date/Time 08/25/2011 09:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 08/18/2011 03:10:50

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:08	
Subject Sample #1	0.122	03:09	
Breath Volume	3.332	03:09	
Air Blank		03:09	Improper Sample
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

---

Data Download Date/Time 08/25/2011 09:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 08/18/2011 03:19:25

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:13
Subject Sample #1	0.126	03:14
Breath Volume	3.531	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample #2	0.111	03:17
Breath Volume	1.898	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time 08/25/2011 09:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY, GARY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: HENDREN

DATE/TIME OF TEST: 08/19/2011 02:42:32

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.279	02:37
Breath Volume	2.699	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.261	02:41
Breath Volume	1.503	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time 08/25/2011 09:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11611080A2  
DRY GAS EXPIRY DATE: 06/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 08/23/2011 03:04:34

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:58
Subject Sample #1	0.320	02:59
Breath Volume	2.296	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample #2	0.317	03:03
Breath Volume	2.367	03:03
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time 08/25/2011 09:05

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUNT, ERNEST, E

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: HUNT, ERNEST

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 12:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/07/2011 02:07:22

---

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.077	02:05	
Air Blank	0.000	02:05	
Subject Sample #1		02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.077	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

---

Data Download Date/Time 09/29/2011 07:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REIS, REBECCA, A

BREATH TEST OPERATOR AGENCY: Havana PD

ARRESTING OFFICER: PYE, SHAUN

ARRESTING OFFICER AGENCY: Havana PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 00:56

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 08/12/2011 01:48:48

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.078	01:43
Air Blank	0.000	01:44
Subject Sample #1	0.000	01:44
Breath Volume	2.492	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample #2	0.000	01:47
Breath Volume	2.925	01:47
Air Blank	0.000	01:47
Control Test	0.077	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time 09/29/2011 07:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SAFFORD, STERLING,

BREATH TEST OPERATOR AGENCY: Havana PD

ARRESTING OFFICER: SAFFORD, STERLING

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 12:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 08/28/2011 01:59:26

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.076	01:54
Air Blank	0.000	01:54
Subject Sample #1	0.113	01:54
Breath Volume	1.929	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample #2	0.115	01:57
Breath Volume	1.902	01:57
Air Blank	0.000	01:58
Control Test	0.077	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

---

Data Download Date/Time 09/29/2011 07:13

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEELE, MARK, T

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: STEELE, MARK

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AGOO4701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 17:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: MORALEZ

DATE/TIME OF TEST: 08/11/2011 21:16:56

---

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:14	
Control Test	0.077	21:14	
Air Blank	0.000	21:15	
Subject Sample #1		21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:16	
Control Test	0.076	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

---

Data Download Date/Time 09/02/2011 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OLVERA, JESUS,

BREATH TEST OPERATOR AGENCY: Hendry County SO

ARRESTING OFFICER: BECKFORD, LEQUANDRA

ARRESTING OFFICER AGENCY: Hendry County SO

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AGOO4701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 08:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 08/13/2011 10:59:48

---

Diagnostics Check	OK	10:53
Air Blank	0.000	10:54
Control Test	0.077	10:54
Air Blank	0.000	10:55
Subject Sample #1	0.107	10:55
Breath Volume	2.265	10:55
Air Blank	0.000	10:56
Air Blank	0.000	10:57
Subject Sample #2	0.109	10:58
Breath Volume	2.218	10:58
Air Blank	0.000	10:58
Control Test	0.077	10:59
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

---

Data Download Date/Time 09/02/2011 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEELE, MARK, T

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: PARKS, T

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AGOO4701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 22:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MANZANO

DATE/TIME OF TEST: 08/17/2011 22:50:40

---

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:48	
Control Test	0.078	22:48	
Air Blank	0.000	22:49	
Subject Sample #1		22:49	RFI Detect
Breath Volume	0.000	22:49	
Air Blank	0.000	22:49	
Control Test	0.077	22:50	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

---

Data Download Date/Time 09/02/2011 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEELE, MARK, T

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: PARKS, TRAVIS

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 07/31/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: AGOO4701

DRY GAS EXPIRY DATE: 02/16/2012

VIOLATION CODE: DUI

ARREST TIME: 22:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MANZANO

DATE/TIME OF TEST: 08/17/2011 22:56:53

---

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.077	22:54	
Air Blank	0.000	22:55	
Subject Sample #1		22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:55	
Control Test	0.078	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

---

Data Download Date/Time 09/02/2011 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOODS, JOSHUA,  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: WOODS, JOSHUA  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AGOO4701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/27/2011 23:52:10

---

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:49	
Control Test	0.077	23:50	
Air Blank	0.000	23:50	
Subject Sample #1		23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Control Test	0.076	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

---

Data Download Date/Time 09/02/2011 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLVERA, JESUS,  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: REED, NICHOLAS  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AGOO4701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 06:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 08/28/2011 15:40:55

---

Diagnostics Check	OK	15:38	
Air Blank	0.000	15:38	
Control Test	0.076	15:38	
Air Blank	0.000	15:39	
Subject Sample #1		15:39	Subject Test Refused
Breath Volume	0.000	15:39	
Air Blank	0.000	15:40	
Control Test	0.076	15:40	
Air Blank	0.000	15:40	
Diagnostics Check	OK	15:40	

---

Data Download Date/Time 09/02/2011 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRK, NATHAN, E  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: KIRK, NATHAN  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 08/25/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AGOO4701  
DRY GAS EXPIRY DATE: 02/16/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 08/28/2011 22:36:40

---

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.078	22:31
Air Blank	0.000	22:31
Subject Sample #1	0.175	22:32
Breath Volume	2.511	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample #2	0.175	22:35
Breath Volume	2.371	22:35
Air Blank	0.000	22:35
Control Test	0.076	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

---

Data Download Date/Time 09/02/2011 16:55

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCANALLY, VONCILLE, S  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: RAKER, RUEL  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/04/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10310080A2  
DRY GAS EXPIRY DATE: 05/01/2012  
VIOLATION CODE: BUI  
ARREST TIME: 19:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BARFIELD

DATE/TIME OF TEST: 08/07/2011 20:38:20

---

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:32	
Control Test	0.077	20:32	
Air Blank	0.000	20:33	
Subject Sample #1	0.113	20:33	
Breath Volume	2.039	20:33	
Air Blank	0.000	20:34	
Air Blank	0.000	20:36	
Subject Sample #2	0.112	20:36	
Breath Volume	2.453	20:36	
Air Blank	0.000	20:37	
Control Test	0.074	20:37	Control Outside Tolerance
Air Blank	0.000	20:38	

---

Data Download Date/Time 09/27/2011 11:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN, JOHN, W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: RAKER, RUEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 08/04/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10310080A2  
DRY GAS EXPIRY DATE: 05/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/13/2011 22:04:00

---

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.075	21:56	
Air Blank	0.000	21:57	
Subject Sample #1	0.085	21:58	
Breath Volume	2.898	21:58	
Air Blank	0.000	21:59	
Air Blank	0.000	22:00	
Subject Sample #2	0.079	22:02	
Breath Volume	3.742	22:02	
Air Blank	0.000	22:03	
Control Test	0.074	22:03	Control Outside Tolerance
Air Blank	0.000	22:04	

---

Data Download Date/Time 09/27/2011 11:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN, JOHN, W  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: RAKER, RUEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 08/04/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10310080A2  
DRY GAS EXPIRY DATE: 05/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/13/2011 22:16:03

---

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:10	
Control Test	0.075	22:10	
Air Blank	0.000	22:10	
Subject Sample #1	0.071	22:11	
Breath Volume	1.468	22:11	
Air Blank	0.000	22:12	
Air Blank	0.000	22:13	
Subject Sample #2	0.075	22:14	
Breath Volume	1.773	22:14	
Air Blank	0.000	22:15	
Control Test	0.074	22:15	Control Outside Tolerance
Air Blank	0.000	22:16	

---

Data Download Date/Time 09/27/2011 11:27

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS, MICAH, L  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: BLACKMAN, JOSEPH  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: CYL018  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: EURICH

DATE/TIME OF TEST: 08/01/2011 01:27:21

---

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:22
Subject Sample #1	0.115	01:23
Breath Volume	1.144	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample #2	0.110	01:25
Breath Volume	1.859	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

---

Data Download Date/Time 08/15/2011 13:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILON, JIMMY, R  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: PILON, JIMMY  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 07/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: CYL018  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/07/2011 03:03:06

---

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.078	03:00	
Air Blank	0.000	03:01	
Subject Sample #1		03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.079	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

---

Data Download Date/Time 08/15/2011 13:40

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PILON, JIMMY, R  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: SCRUGGS, DEMETRIUS  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 08/15/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: CYL018  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: MORMAN

DATE/TIME OF TEST: 08/16/2011 20:54:16

---

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49
Subject Sample #1	0.200	20:49
Breath Volume	1.398	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample #2	0.206	20:52
Breath Volume	2.324	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

---

Data Download Date/Time 09/30/2011 23:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: THOMAS, MICAH, L

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: MALLEY, PATRICK

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 08/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: CYL018

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 22:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 08/17/2011 23:31:27

---

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.080	23:29	
Air Blank	0.000	23:29	
Subject Sample #1		23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Control Test	0.079	23:30	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

---

Data Download Date/Time 09/30/2011 23:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: THOMAS, MICAH, L

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: MALLEY, PATRICK

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 08/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: CYL018

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 20:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/28/2011 21:38:06

---

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:36	
Air Blank	0.000	21:36	
Subject Sample #1		21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:37	
Control Test	0.079	21:37	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

---

Data Download Date/Time 09/30/2011 23:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PILON, JIMMY, R

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: PILON, JIMMY

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 08/15/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: CYL018

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 22:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 08/29/2011 22:45:56

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Subject Sample #1	0.280	22:41
Breath Volume	3.589	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:43
Subject Sample #2	0.284	22:44
Breath Volume	3.914	22:44
Air Blank	0.000	22:45
Control Test	0.077	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time 09/30/2011 23:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWINDLE, CHARLES, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SWINDLE, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/02/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: SWEARENGIN

DATE/TIME OF TEST: 08/26/2011 22:21:34

---

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.079	22:16
Air Blank	0.000	22:16
Subject Sample #1	0.238	22:17
Breath Volume	3.582	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample #2	0.236	22:20
Breath Volume	4.062	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

---

Data Download Date/Time 09/30/2011 12:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DAWSON, GARY, S

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: DAWSON, GARY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/02/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 05911080A2

DRY GAS EXPIRY DATE: 02/01/2013

VIOLATION CODE: DUI

ARREST TIME: 03:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/27/2011 03:48:56

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:43
Subject Sample #1	0.091	03:44
Breath Volume	2.945	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample #2	0.089	03:47
Breath Volume	2.292	03:47
Air Blank	0.000	03:48
Control Test	0.077	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time 09/30/2011 12:04

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAINES III, ROBERT, L  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: LOPEZ, RICARDO  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 08/01/2011 01:15:48

---

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.078	01:13	
Air Blank	0.000	01:14	
Subject Sample #1		01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.078	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

---

Data Download Date/Time 08/03/2011 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY, G  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: HEINEMANN

DATE/TIME OF TEST: 08/03/2011 03:54:56

---

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.077	03:49
Air Blank	0.000	03:49
Subject Sample #1	0.072	03:50
Breath Volume	2.652	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample #2	0.071	03:53
Breath Volume	2.832	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time 08/03/2011 09:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREGO, M, P  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: KAPLAN, R  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: MCGAVIN

DATE/TIME OF TEST: 08/12/2011 02:08:09

---

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.077	02:02
Air Blank	0.000	02:03
Subject Sample #1	0.231	02:03
Breath Volume	3.035	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #2	0.225	02:06
Breath Volume	3.414	02:06
Air Blank	0.000	02:07
Control Test	0.076	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORDICA, FELICIA, N  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY, GARY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/12/2011 02:29:34

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.077	02:23
Air Blank	0.000	02:23
Subject Sample #1	0.165	02:24
Breath Volume	2.003	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample #2	0.163	02:28
Breath Volume	1.386	02:28
Air Blank	0.000	02:28
Control Test	0.075	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPEIGNER, BRIAN, W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SPEIGNER, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 08/12/2011 20:39:30

---

Diagnostics Check	OK	20:32
Air Blank	0.000	20:33
Control Test	0.076	20:33
Air Blank	0.000	20:34
Subject Sample #1	0.149	20:34
Breath Volume	2.691	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample #2	0.149	20:38
Breath Volume	3.480	20:38
Air Blank	0.000	20:38
Control Test	0.076	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEANEY, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MEGNA, JEROME  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 08/13/2011 03:05:20

---

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.077	03:03	
Air Blank	0.000	03:03	
Subject Sample #1		03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.077	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCOTT, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: VANSANT JR

DATE/TIME OF TEST: 08/13/2011 03:43:28

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.077	03:37
Air Blank	0.000	03:37
Subject Sample #1	0.160	03:38
Breath Volume	3.441	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.161	03:41
Breath Volume	2.027	03:41
Air Blank	0.000	03:42
Control Test	0.076	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCOTT, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: UPENIEKS

DATE/TIME OF TEST: 08/14/2011 03:07:50

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.077	03:02
Air Blank	0.000	03:02
Subject Sample #1	0.156	03:03
Breath Volume	1.847	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.155	03:06
Breath Volume	1.972	03:06
Air Blank	0.000	03:06
Control Test	0.077	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SMIDT, M  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/14/2011 04:11:58

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.076	04:06
Air Blank	0.000	04:06
Subject Sample #1	0.191	04:07
Breath Volume	1.617	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample #2	0.207	04:10
Breath Volume	2.132	04:10
Air Blank	0.000	04:11
Control Test	0.077	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPEIGNER, BRIAN, W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SPEIGNER, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/16/2011 04:00:21

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.077	03:53
Air Blank	0.000	03:53
Subject Sample #1	0.197	03:55
Breath Volume	2.742	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample #2	0.202	03:58
Breath Volume	2.015	03:58
Air Blank	0.000	03:59
Control Test	0.077	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULZE, JON,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCHULZE, JON  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RUDD

DATE/TIME OF TEST: 08/17/2011 00:51:21

---

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Subject Sample #1	0.030	00:46
Breath Volume	4.589	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample #2	0.028	00:49
Breath Volume	3.757	00:49
Air Blank	0.000	00:50
Control Test	0.076	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCOTT, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: PHIPPS

DATE/TIME OF TEST: 08/20/2011 01:21:35

---

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.076	01:10	
Air Blank	0.000	01:11	
Subject Sample #1		01:14	No Sample Provided
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample #2		01:20	No Sample Provided
Breath Volume	0.000	01:20	
Air Blank	0.000	01:20	
Control Test	0.078	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SMIDT, M  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DOLAR

DATE/TIME OF TEST: 08/20/2011 02:35:40

---

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.077	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.093	02:31
Breath Volume	1.781	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample #2	0.099	02:34
Breath Volume	1.792	02:34
Air Blank	0.000	02:34
Control Test	0.076	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER, MARK, J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: PAUL, MATT  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 08/21/2011 21:35:16

---

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.076	21:33	
Air Blank	0.000	21:33	
Subject Sample #1		21:33	Subject Test Refused
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Control Test	0.076	21:34	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, JAMES, J  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HARRIS, JAMES  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DEMERCADO

DATE/TIME OF TEST: 08/22/2011 03:11:04

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:01
Air Blank	0.000	03:02
Subject Sample #1	0.189	03:03
Breath Volume	1.320	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.211	03:06
Breath Volume	1.847	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample #3	0.214	03:09
Breath Volume	1.207	03:09
Air Blank	0.000	03:10
Control Test	0.076	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:11

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCHWAB, M  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: KIDNEY

DATE/TIME OF TEST: 08/23/2011 00:07:31

---

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.077	00:00
Air Blank	0.000	00:00
Subject Sample #1	0.051	00:02
Breath Volume	1.343	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample #2	0.049	00:06
Breath Volume	1.597	00:06
Air Blank	0.000	00:06
Control Test	0.076	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREGO, M, P  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: CREGO, M  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RACICOT

DATE/TIME OF TEST: 08/26/2011 02:27:14

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Subject Sample #1	0.204	02:22
Breath Volume	2.644	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample #2	0.205	02:25
Breath Volume	3.183	02:25
Air Blank	0.000	02:26
Control Test	0.076	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY, G  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: KAUFFMAN

DATE/TIME OF TEST: 08/26/2011 04:05:16

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.077	03:57
Air Blank	0.000	03:57
Subject Sample #1	0.243	03:59
Breath Volume	1.253	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample #2	0.239	04:03
Breath Volume	1.691	04:03
Air Blank	0.000	04:04
Control Test	0.075	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCOTT, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 08/27/2011 01:43:50

---

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.076	01:37	
Air Blank	0.000	01:37	
Subject Sample #1	0.155	01:39	
Breath Volume	1.359	01:39	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample #2	0.153	01:42	
Breath Volume	1.234	01:42	
Air Blank	0.000	01:42	
Control Test	0.074	01:43	Control Outside Tolerance
Air Blank	0.000	01:43	

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCOTT, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 08/27/2011 01:52:09

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.075	01:46
Air Blank	0.000	01:47
Subject Sample #1	0.154	01:47
Breath Volume	1.332	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #2	0.149	01:50
Breath Volume	1.242	01:50
Air Blank	0.000	01:51
Control Test	0.076	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEANEY, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MEANEY, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: PALLAS

DATE/TIME OF TEST: 08/27/2011 03:54:47

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.077	03:48
Air Blank	0.000	03:48
Subject Sample #1	0.107	03:50
Breath Volume	1.890	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample #2	0.114	03:53
Breath Volume	1.507	03:53
Air Blank	0.000	03:53
Control Test	0.076	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOHIN, ERIK, J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BOHIN, ERIK  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: BLOCKER

DATE/TIME OF TEST: 08/27/2011 04:25:31

---

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.076	04:22	
Air Blank	0.000	04:23	
Subject Sample #1		04:23	Subject Test Refused
Breath Volume	0.000	04:23	
Air Blank	0.000	04:24	
Control Test	0.077	04:24	
Air Blank	0.000	04:24	
Diagnostics Check	Fail	04:25	Voltage/Current Test
Air Blank	0.000	04:25	

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWERS, CHRISTOPHER, J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: POWERS, CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 08/28/2011 23:20:25

---

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.077	23:13
Air Blank	0.000	23:14
Subject Sample #1	0.193	23:15
Breath Volume	1.734	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample #2	0.182	23:18
Breath Volume	3.625	23:18
Air Blank	0.000	23:19
Control Test	0.076	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULZE, JON,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCHULZE, JON  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 08/30/2011 01:07:53

---

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.078	01:00
Air Blank	0.000	01:01
Subject Sample #1	0.079	01:01
Breath Volume	1.148	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample #2	0.077	01:06
Breath Volume	1.523	01:06
Air Blank	0.000	01:06
Control Test	0.077	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MEGNA, JEROME  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 08/31/2011 03:15:12

---

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.077	03:09
Air Blank	0.000	03:10
Subject Sample #1	0.073	03:10
Breath Volume	2.304	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.077	03:13
Breath Volume	4.027	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time 09/22/2011 10:28

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCOTT, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 08/03/2011 03:34:28

---

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.077	03:22	
Air Blank	0.000	03:23	
Subject Sample #1	0.173	03:26	Volume Not Met
Breath Volume	0.300	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:28	
Subject Sample #2	0.208	03:29	
Breath Volume	3.417	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample #3	0.200	03:32	
Breath Volume	2.230	03:32	
Air Blank	0.000	03:33	
Control Test	0.074	03:33	Control Outside Tolerance
Air Blank	0.000	03:34	

---

Data Download Date/Time 08/03/2011 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCOTT, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 08/03/2011 03:44:32

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.076	03:37
Air Blank	0.000	03:37
Subject Sample #1	0.205	03:39
Breath Volume	2.128	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.208	03:42
Breath Volume	1.414	03:42
Air Blank	0.000	03:43
Control Test	0.075	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time 08/03/2011 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, JAMES, J  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HARRIS, JAMES  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/11/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05911080A2  
DRY GAS EXPIRY DATE: 02/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: STEUART

DATE/TIME OF TEST: 08/03/2011 03:53:41

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.076	03:48
Air Blank	0.000	03:48
Subject Sample #1	0.185	03:49
Breath Volume	1.445	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample #2	0.165	03:52
Breath Volume	1.144	03:52
Air Blank	0.000	03:52
Control Test	0.075	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time 08/03/2011 09:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPEIGNER, BRIAN, W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SPEIGNER, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MCCLOUD

DATE/TIME OF TEST: 08/05/2011 01:01:38

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.078	00:55
Air Blank	0.000	00:56
Subject Sample #1	0.013	00:56
Breath Volume	1.718	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample #2	0.012	01:00
Breath Volume	2.601	01:00
Air Blank	0.000	01:00
Control Test	0.078	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREGO, M, P  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: GLOVER, J  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 08/06/2011 02:04:49

---

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.077	01:59
Air Blank	0.000	01:59
Subject Sample #1	0.079	02:00
Breath Volume	2.761	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample #2	0.080	02:03
Breath Volume	2.609	02:03
Air Blank	0.000	02:03
Control Test	0.077	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPEIGNER, BRIAN, W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SPEIGNER, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 08/11/2011 03:30:00

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.078	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.190	03:25
Breath Volume	1.406	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample #2	0.188	03:28
Breath Volume	1.527	03:28
Air Blank	0.000	03:29
Control Test	0.077	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:30

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPEIGNER, BRIAN, W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SPEIGNER, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ALDEN

DATE/TIME OF TEST: 08/13/2011 03:56:09

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:50
Subject Sample #1	0.198	03:51
Breath Volume	1.207	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample #2	0.218	03:54
Breath Volume	1.222	03:54
Air Blank	0.000	03:55
Control Test	0.077	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWERS, CHRISTOPHER, J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: POWERS, CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: BECKWITH

DATE/TIME OF TEST: 08/17/2011 23:47:27

---

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.077	23:45	
Air Blank	0.000	23:45	
Subject Sample #1		23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.077	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, JAMES, J  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BIEDERMANN, DALE  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 08/19/2011 01:56:41

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:51
Subject Sample #1	0.129	01:52
Breath Volume	1.765	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample #2	0.124	01:55
Breath Volume	2.535	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPEIGNER, BRIAN, W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SPEIGNER, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: NAJDOWSKI

DATE/TIME OF TEST: 08/19/2011 02:26:51

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample #1	0.172	02:22
Breath Volume	1.761	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample #2	0.179	02:25
Breath Volume	2.222	02:25
Air Blank	0.000	02:25
Control Test	0.078	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BALLARD, J  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HANDSCOMB

DATE/TIME OF TEST: 08/20/2011 02:06:38

---

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Subject Sample #1	0.221	02:01
Breath Volume	2.546	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample #2	0.220	02:05
Breath Volume	2.925	02:05
Air Blank	0.000	02:05
Control Test	0.077	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS, JAMES, J  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HARRIS, JAMES  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/21/2011 02:58:14

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:53
Subject Sample #1	0.128	02:53
Breath Volume	2.292	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample #2	0.126	02:56
Breath Volume	3.035	02:56
Air Blank	0.000	02:57
Control Test	0.076	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAUFORD, JIM, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BEAUFORD, JIM  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GREENWOOD

DATE/TIME OF TEST: 08/27/2011 01:52:59

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.076	01:51	
Air Blank	0.000	01:52	
		01:52	Improper Sample
Air Blank	0.000	01:52	

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEAUFORD, JIM, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BEAUFORD, JIM  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 12:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GREENWOOD

DATE/TIME OF TEST: 08/27/2011 02:01:09

---

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.076	01:55
Air Blank	0.000	01:55
Subject Sample #1	0.190	01:56
Breath Volume	1.105	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample #2	0.196	01:59
Breath Volume	1.753	01:59
Air Blank	0.000	02:00
Control Test	0.077	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IZQUIERDO, JEREMY, M  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: IZQUIERDO, JEREMY  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BOLTON

DATE/TIME OF TEST: 08/27/2011 03:53:22

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.077	03:51	
Air Blank	0.000	03:51	
Subject Sample #1		03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.077	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IZQUIERDO, JEREMY, M  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: PRUETT, LYNN  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 08/31/2011 02:59:38

---

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.078	02:57	
Air Blank	0.000	02:58	
Subject Sample #1		02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.077	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERESSE, A,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: MERESSE, A  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 675059  
DRY GAS EXPIRY DATE: 09/25/2011  
VIOLATION CODE: DUI  
ARREST TIME: 04:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 08/31/2011 05:38:33

---

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.078	05:32
Air Blank	0.000	05:33
Subject Sample #1	0.198	05:34
Breath Volume	2.726	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample #2	0.189	05:37
Breath Volume	1.628	05:37
Air Blank	0.000	05:37
Control Test	0.077	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

---

Data Download Date/Time 09/22/2011 10:23

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT, ERNEST, E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HUNT, ERNEST  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: ERVIN

DATE/TIME OF TEST: 08/13/2011 20:40:57

---

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:34	
Control Test	0.075	20:34	
Air Blank	0.000	20:34	
Subject Sample #1	0.195	20:36	
Breath Volume	1.652	20:36	
Air Blank	0.000	20:36	
Air Blank	0.000	20:38	
Subject Sample #2	0.206	20:39	
Breath Volume	1.910	20:39	
Air Blank	0.000	20:40	
Control Test	0.073	20:40	Control Outside Tolerance
Air Blank	0.000	20:40	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WADE, CHARLES, W

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: WADE, CHARLES

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/18/2011 00:14:50

---

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.075	00:05	
Air Blank	0.000	00:05	
Subject Sample #1	0.197	00:07	
Breath Volume	2.046	00:07	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample #2	0.201	00:12	Volume Not Met
Breath Volume	0.542	00:12	
Air Blank		00:12	Improper Sample
Air Blank	0.000	00:13	
Control Test	0.076	00:14	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADE, CHARLES, W  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: WADE, CHARLES  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/18/2011 00:23:10

---

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.073	00:22	Control Outside Tolerance
Air Blank	0.000	00:23	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADE, CHARLES, W  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: WADE, CHARLES  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/18/2011 00:26:47

---

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.074	00:26	Control Outside Tolerance
Air Blank	0.000	00:26	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADE, CHARLES, W  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: WADE, CHARLES  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/18/2011 01:28:24

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.075	01:22	
Air Blank	0.000	01:22	
Subject Sample #1	0.177	01:23	
Breath Volume	2.207	01:23	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample #2	0.164	01:26	
Breath Volume	2.003	01:26	
Air Blank	0.000	01:27	
Control Test	0.074	01:27	Control Outside Tolerance
Air Blank	0.000	01:28	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT, ERNEST, E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HUNT, ERNEST  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 08/27/2011 02:01:52

---

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.076	01:54	
Air Blank	0.000	01:54	
Subject Sample #1	0.139	01:56	
Breath Volume	2.265	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:58	
Subject Sample #2	0.125	02:00	
Breath Volume	1.542	02:00	
Air Blank	0.000	02:00	
Control Test	0.074	02:01	Control Outside Tolerance
Air Blank	0.000	02:01	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT, ERNEST, E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HUNT, ERNEST  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: LOPEZ MARROQUIN

DATE/TIME OF TEST: 08/27/2011 05:38:06

---

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.075	05:27	
Air Blank	0.000	05:27	
Subject Sample #1	0.128	05:30	Volume Not Met
Breath Volume	0.328	05:30	
Air Blank	0.000	05:31	
Air Blank	0.000	05:33	
Subject Sample #2	0.107	05:36	Volume Not Met
Breath Volume	0.609	05:36	
Air Blank	0.000	05:37	
Control Test	0.073	05:37	Control Outside Tolerance
Air Blank	0.000	05:38	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT, ERNEST, E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HUNT, ERNEST  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: LOPEZ MARROQUIN

DATE/TIME OF TEST: 08/27/2011 05:52:00

---

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:40	
Control Test	0.075	05:41	
Air Blank	0.000	05:41	
Subject Sample #1	0.104	05:44	Slope Not Level
Breath Volume	1.179	05:44	
Air Blank	0.000	05:45	
Air Blank	0.000	05:47	
Subject Sample #2	0.101	05:50	Volume Not Met
Breath Volume	1.082	05:50	
Air Blank	0.000	05:51	
Control Test	0.073	05:51	Control Outside Tolerance
Air Blank	0.000	05:52	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADE, CHARLES, W  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: WADE, CHARLES  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/28/2011 00:56:20

---

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.074	00:55	Control Outside Tolerance
Air Blank	0.000	00:56	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADE, CHARLES, W  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: KING, JASON  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 08/28/2011 01:39:51

---

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.075	01:37	
Air Blank	0.000	01:37	
Subject Sample #1		01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Control Test	0.074	01:39	Control Outside Tolerance
Air Blank	0.000	01:39	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNT, ERNEST, E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HUNT, ERNEST  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/30/2011 02:28:21

---

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.075	02:21	
Air Blank	0.000	02:22	
Subject Sample #1	0.145	02:23	
Breath Volume	1.433	02:23	
Air Blank	0.000	02:24	
Air Blank	0.000	02:26	
Subject Sample #2	0.152	02:26	
Breath Volume	1.726	02:26	
Air Blank	0.000	02:27	
Control Test	0.072	02:27	Control Outside Tolerance
Air Blank	0.000	02:28	

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAFFORD, STERLING,  
BREATH TEST OPERATOR AGENCY: Gadsden County SO  
ARRESTING OFFICER: SAFFORD, STERLING  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 08/30/2011 15:32:50

---

Diagnostics Check	OK	15:26
Air Blank	0.000	15:27
Control Test	0.075	15:27
Air Blank	0.000	15:28
Subject Sample #1	0.000	15:28
Breath Volume	2.718	15:28
Air Blank	0.000	15:29
Air Blank	0.000	15:31
Subject Sample #2	0.000	15:31
Breath Volume	2.207	15:31
Air Blank	0.000	15:31
Control Test	0.075	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAFFORD, STERLING,  
BREATH TEST OPERATOR AGENCY: Gadsden County SO  
ARRESTING OFFICER: SAFFORD, STERLING  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/09/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 13:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 08/30/2011 15:46:36

---

Diagnostics Check	OK	15:40
Air Blank	0.000	15:40
Control Test	0.075	15:41
Air Blank	0.000	15:41
Subject Sample #1	0.063	15:42
Breath Volume	2.296	15:42
Air Blank	0.000	15:42
Air Blank	0.000	15:44
Subject Sample #2	0.060	15:45
Breath Volume	2.050	15:45
Air Blank	0.000	15:45
Control Test	0.075	15:46
Air Blank	0.000	15:46
Diagnostics Check	OK	15:46

---

Data Download Date/Time 09/29/2011 08:48

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAINES, CRAIG, A  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: GAINES, CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/04/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 710643  
DRY GAS EXPIRY DATE: 12/28/2011  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RUDD

DATE/TIME OF TEST: 08/24/2011 01:45:14

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Subject Sample #1	0.091	01:40
Breath Volume	1.902	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample #2	0.081	01:43
Breath Volume	2.128	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time 09/29/2011 10:32

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREENE, JOSHUA, M

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: ZARATE, JOHN

ARRESTING OFFICER AGENCY: Wakulla County SO

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 18:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 08/08/2011 19:56:08

---

Diagnostics Check	OK	19:49
Air Blank	0.000	19:50
Control Test	0.078	19:50
Air Blank	0.000	19:50
Subject Sample #1	0.043	19:51
Breath Volume	1.734	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample #2	0.041	19:54
Breath Volume	1.632	19:54
Air Blank	0.000	19:55
Control Test	0.076	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

---

Data Download Date/Time 09/27/2011 11:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEANEY, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: PHILLIPS, RANDALL  
ARRESTING OFFICER AGENCY: Wakulla County SO  
LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: QUICK

DATE/TIME OF TEST: 08/18/2011 21:02:52

---

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.076	20:54
Air Blank	0.000	20:54
Subject Sample #1	0.341	20:55
Breath Volume	3.011	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample #2	0.318	20:57
Breath Volume	2.300	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample #3	0.337	21:01
Breath Volume	2.660	21:01
Air Blank	0.000	21:01
Control Test	0.075	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

---

Data Download Date/Time 09/27/2011 11:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPEIGNER, BRIAN, W

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: SPEIGNER, BRIAN

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13811080A1

DRY GAS EXPIRY DATE: 07/01/2013

VIOLATION CODE: DUI

ARREST TIME: 22:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: BOZEMAN

DATE/TIME OF TEST: 08/26/2011 23:11:04

---

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.077	23:05
Air Blank	0.000	23:05
Subject Sample #1	0.000	23:06
Breath Volume	2.617	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample #2	0.000	23:09
Breath Volume	2.769	23:09
Air Blank	0.000	23:10
Control Test	0.077	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:11

---

Data Download Date/Time 09/27/2011 11:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPEIGNER, BRIAN, W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SPEIGNER, BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13811080A1  
DRY GAS EXPIRY DATE: 07/01/2013  
VIOLATION CODE: DUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/27/2011 21:18:20

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.077	21:11
Air Blank	0.000	21:12
Subject Sample #1	0.129	21:13
Breath Volume	3.257	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:15
Subject Sample #2	0.127	21:16
Breath Volume	3.359	21:16
Air Blank	0.000	21:17
Control Test	0.077	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

---

Data Download Date/Time 09/27/2011 11:17

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: JAMES, JOHN

ARRESTING OFFICER AGENCY: Orlando PD

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 08/03/2011 03:17:31

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.078	03:12
Air Blank	0.000	03:12
Subject Sample #1	0.235	03:13
Breath Volume	3.269	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.239	03:16
Breath Volume	3.937	03:16
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time 08/04/2011 15:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRITY, SAMUEL, S  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS, DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/04/2011 03:05:32

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.078	03:00
Air Blank	0.000	03:00
Subject Sample #1	0.110	03:01
Breath Volume	3.089	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.104	03:04
Breath Volume	2.988	03:04
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time 08/04/2011 15:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: HARWOOD, ARMANDO

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LAMBETH

DATE/TIME OF TEST: 08/04/2011 04:00:01

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:55
Subject Sample #1	0.108	03:55
Breath Volume	4.277	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample #2	0.110	03:58
Breath Volume	4.605	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	04:00

---

Data Download Date/Time 08/04/2011 15:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: WHALEN, DANIEL

ARRESTING OFFICER AGENCY: Orlando PD

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: DICOSTANZO

DATE/TIME OF TEST: 08/12/2011 04:09:05

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.077	04:03
Air Blank	0.000	04:04
Subject Sample #1	0.174	04:04
Breath Volume	2.738	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.173	04:07
Breath Volume	3.089	04:07
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

---

Data Download Date/Time 08/18/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GERRITY, SAMUEL, S

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: BLAKE, TRISTA

ARRESTING OFFICER AGENCY: Ocoee PD

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GALVEZDIAZ

DATE/TIME OF TEST: 08/13/2011 03:42:44

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Subject Sample #1	0.151	03:38
Breath Volume	2.503	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample #2	0.164	03:41
Breath Volume	4.093	03:41
Air Blank	0.000	03:41
Control Test	0.078	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

---

Data Download Date/Time 08/18/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DANJOU, SCOTT  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 08/14/2011 03:03:03

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Subject Sample #1	0.008	02:58
Breath Volume	1.480	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample #2	0.000	03:01
Breath Volume	1.691	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:03

---

Data Download Date/Time 08/18/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: RANDO, STEPHEN

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ANGER

DATE/TIME OF TEST: 08/14/2011 04:07:51

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:02
Subject Sample #1	0.165	04:03
Breath Volume	1.312	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.157	04:06
Breath Volume	1.792	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

---

Data Download Date/Time 08/18/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: YOUNG, MICHAEL

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 03:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 08/14/2011 04:25:01

---

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:19
Air Blank	0.000	04:19
Subject Sample #1	0.039	04:20
Breath Volume	1.699	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.037	04:23
Breath Volume	1.300	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:25

---

Data Download Date/Time 08/18/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN, JAMES, E

BREATH TEST OPERATOR AGENCY: Orange County SO

ARRESTING OFFICER: WALWYN, CELVIN

ARRESTING OFFICER AGENCY: Orange County SO

LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 904241

DRY GAS EXPIRY DATE: 11/12/2012

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 08/15/2011 03:49:40

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Subject Sample #1	0.226	03:44
Breath Volume	2.589	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample #2	0.211	03:48
Breath Volume	2.554	03:48
Air Blank	0.000	03:48
Control Test	0.077	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

---

Data Download Date/Time 08/18/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN, JAMES, E  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LORENZO, STEVE  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 07/21/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 904241  
DRY GAS EXPIRY DATE: 11/12/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 08/17/2011 03:43:37

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample #1	0.147	03:39
Breath Volume	1.789	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.152	03:42
Breath Volume	2.039	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time 08/18/2011 12:31

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WADE, CHARLES, W

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: WADE, CHARLES

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/18/2011 00:53:07

---

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.075	00:46	
Air Blank	0.000	00:47	
Subject Sample #1	0.196	00:47	
Breath Volume	1.882	00:47	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample #2	0.183	00:51	
Breath Volume	2.078	00:51	
Air Blank	0.000	00:52	
Control Test	0.073	00:52	Control Outside Tolerance
Air Blank	0.000	00:53	

---

Data Download Date/Time 09/29/2011 08:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WADE, CHARLES, W

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: WADE, CHARLES

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/18/2011 01:05:18

---

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.075	00:59	
Air Blank	0.000	01:00	
Subject Sample #1	0.178	01:00	
Breath Volume	2.039	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample #2	0.169	01:03	
Breath Volume	1.468	01:03	
Air Blank	0.000	01:04	
Control Test	0.074	01:04	Control Outside Tolerance
Air Blank	0.000	01:05	

---

Data Download Date/Time 09/29/2011 08:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WADE, CHARLES, W

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: WADE, CHARLES

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 08/03/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 00:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/28/2011 00:49:31

---

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.074	00:48	Control Outside Tolerance
Air Blank	0.000	00:49	

---

Data Download Date/Time 09/29/2011 08:51

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VISSER, ANDREW, C

BREATH TEST OPERATOR AGENCY: St. Cloud PD

ARRESTING OFFICER: VISSER, ANDREW

ARRESTING OFFICER AGENCY: St. Cloud PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 23:52

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 08/03/2011 00:26:33

---

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:24	
Control Test	0.081	00:24	
Air Blank	0.000	00:24	
Subject Sample #1		00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:25	
Control Test	0.082	00:25	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAY, THEODIS, M  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: DAVIDSON, JAMES  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 08/05/2011 03:45:32

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Subject Sample #1	0.136	03:40
Breath Volume	2.597	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample #2	0.127	03:43
Breath Volume	1.414	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNTON, ALVIN, F  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: COLLAZO, LOUIS  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: TUBBS

DATE/TIME OF TEST: 08/05/2011 07:55:05

---

Diagnostics Check	OK	07:50	
Air Blank	0.000	07:51	
Control Test	0.081	07:51	
Air Blank	0.000	07:51	
Subject Sample #1		07:53	Subject Test Refused
Breath Volume	0.000	07:53	
Air Blank	0.000	07:54	
Control Test	0.081	07:54	
Air Blank	0.000	07:55	
Diagnostics Check	OK	07:55	

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, HECTOR, E  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RODRIGUEZ, HECTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:07

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 08/05/2011 11:29:45

---

Diagnostics Check	OK	11:29	
Air Blank		11:29	Improper Sample
Air Blank		11:29	Improper Sample

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, HECTOR, E  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RODRIGUEZ, HECTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 08/05/2011 11:32:51

---

Diagnostics Check	OK	11:31	
Air Blank		11:32	Ambient Fail
Air Blank	0.000	11:32	

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, HECTOR, E  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RODRIGUEZ, HECTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 08/05/2011 11:36:14

---

Diagnostics Check	OK	11:35	
Air Blank		11:36	Improper Sample
Air Blank		11:36	Improper Sample

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, HECTOR, E  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RODRIGUEZ, HECTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 09:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 08/05/2011 11:43:11

---

Diagnostics Check	OK	11:42	
Air Blank		11:42	Ambient Fail
Air Blank	0.000	11:43	

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, HECTOR, E

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: RODRIGUEZ, HECTOR

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 09:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 08/05/2011 11:57:16

---

Diagnostics Check	OK	11:51
Air Blank	0.000	11:51
Control Test	0.081	11:52
Air Blank	0.000	11:52
Subject Sample #1	0.065	11:52
Breath Volume	1.421	11:52
Air Blank	0.000	11:53
Air Blank	0.000	11:55
Subject Sample #2	0.062	11:55
Breath Volume	1.402	11:55
Air Blank	0.000	11:56
Control Test	0.080	11:56
Air Blank	0.000	11:57
Diagnostics Check	OK	11:57

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VISSER, ANDREW, C

BREATH TEST OPERATOR AGENCY: St. Cloud PD

ARRESTING OFFICER: VISSER, A

ARRESTING OFFICER AGENCY: St. Cloud PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 02:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LUGO-VAZQUEZ

DATE/TIME OF TEST: 08/06/2011 03:55:54

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample #1	0.040	03:51
Breath Volume	1.296	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample #2	0.040	03:54
Breath Volume	1.550	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIZARRO, JESUS, M  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRUZ, EDDIE  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: OTHER  
ARREST TIME: 12:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:38

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/06/2011 14:47:49

---

Diagnostics Check	OK	14:42	
Air Blank	0.000	14:42	
Control Test	0.080	14:42	
Air Blank	0.000	14:43	
Subject Sample #1	0.168	14:43	
Breath Volume	2.789	14:43	
Air Blank	0.000	14:44	
Air Blank	0.000	14:46	
Subject Sample #2		14:46	Subject Test Refused
Breath Volume	0.000	14:46	
Air Blank	0.000	14:46	
Control Test	0.080	14:47	
Air Blank	0.000	14:47	
Diagnostics Check	OK	14:47	

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNANDEZ, CANDELARIO,

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: HARTMANN, PATRICK

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 11:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: BASLEE

DATE/TIME OF TEST: 08/07/2011 13:28:41

---

Diagnostics Check	OK	13:22
Air Blank	0.000	13:23
Control Test	0.080	13:23
Air Blank	0.000	13:23
Subject Sample #1	0.128	13:24
Breath Volume	2.265	13:24
Air Blank	0.000	13:24
Air Blank	0.000	13:26
Subject Sample #2	0.120	13:27
Breath Volume	2.109	13:27
Air Blank	0.000	13:27
Control Test	0.080	13:28
Air Blank	0.000	13:28
Diagnostics Check	OK	13:28

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEDULLA, VINCENT, A

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: PEDULLA, VINCENT

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 17:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: AADDI

DATE/TIME OF TEST: 08/08/2011 18:50:44

---

Diagnostics Check	OK	18:42	
Air Blank	0.000	18:42	
Control Test	0.080	18:42	
Air Blank	0.000	18:43	
Subject Sample #1	0.233	18:45	
Breath Volume	2.667	18:45	
Air Blank	0.000	18:45	
Air Blank	0.000	18:47	
Subject Sample #2	0.234	18:49	
Breath Volume	1.531	18:49	
Air Blank	0.000	18:49	
Control Test	0.078	18:50	RFI Detect
Air Blank	0.000	18:50	

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEDULLA, VINCNET, A

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: PEDULLA, VINCENT

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 17:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: AADDI

DATE/TIME OF TEST: 08/08/2011 19:02:59

---

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.079	18:57
Air Blank	0.000	18:57
Subject Sample #1	0.241	18:58
Breath Volume	1.664	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:00
Subject Sample #2	0.247	19:01
Breath Volume	2.324	19:01
Air Blank	0.000	19:02
Control Test	0.078	19:02
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COGGINS, MICHELLE, M

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: PEOLLE, JAMES

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 19:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: AADDI

DATE/TIME OF TEST: 08/09/2011 01:13:25

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:08
Subject Sample #1	0.082	01:08
Breath Volume	2.660	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample #2	0.081	01:11
Breath Volume	1.660	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COGGINS, MICHELLE, M  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PEDOLLE, JAMES  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: AADDI

DATE/TIME OF TEST: 08/09/2011 03:03:42

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample #1	0.029	02:59
Breath Volume	1.792	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample #2	0.027	03:02
Breath Volume	1.148	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VISSER, ANDREW, C

BREATH TEST OPERATOR AGENCY: St. Cloud PD

ARRESTING OFFICER: VISSER, ANDREW

ARRESTING OFFICER AGENCY: St. Cloud PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 04:03

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: BARRETO

DATE/TIME OF TEST: 08/12/2011 05:13:18

---

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:58	
Control Test	0.081	04:58	
Air Blank	0.000	04:58	
Subject Sample #1	0.140	05:02	Volume Not Met
Breath Volume	0.910	05:02	
Air Blank	0.000	05:03	
Air Blank	0.000	05:04	
Subject Sample #2	0.157	05:06	
Breath Volume	1.617	05:06	
Air Blank	0.000	05:07	
Air Blank	0.000	05:08	
Subject Sample #3	0.148	05:11	
Breath Volume	1.671	05:11	
Air Blank	0.000	05:12	
Control Test	0.078	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COGGINS, MICHELLE, M

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: HERRERA, AGNEL

ARRESTING OFFICER AGENCY: St. Cloud PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 00:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CHALMERS

DATE/TIME OF TEST: 08/28/2011 01:34:09

---

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Subject Sample #1	0.080	01:21	
Breath Volume	1.878	01:21	
Air Blank	0.000	01:22	
Air Blank	0.000	01:24	
Subject Sample #2	0.070	01:27	Volume Not Met
Breath Volume	0.238	01:27	
Air Blank	0.000	01:27	
Air Blank	0.000	01:29	
Subject Sample #3	0.071	01:32	
Breath Volume	1.257	01:32	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

---

Data Download Date/Time 08/29/2011 09:10

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLLEY, DANIEL, A

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: COLLEY, DANIEL

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 07/02/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 17:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 08/05/2011 18:24:24

---

Diagnostics Check	OK	18:17
Air Blank	0.000	18:17
Control Test	0.080	18:18
Air Blank	0.000	18:18
Subject Sample #1	0.042	18:19
Breath Volume	1.433	18:19
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample #2	0.038	18:22
Breath Volume	1.843	18:22
Air Blank	0.000	18:23
Control Test	0.078	18:23
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

---

Data Download Date/Time 08/06/2011 16:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SULLIVAN, ASHLEY, A

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: SULLIVAN, ASHLEY

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 17:40

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SWARTZ

DATE/TIME OF TEST: 08/17/2011 18:29:59

---

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.079	18:24
Air Blank	0.000	18:24
Subject Sample #1	0.149	18:25
Breath Volume	1.878	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample #2	0.148	18:28
Breath Volume	2.125	18:28
Air Blank	0.000	18:29
Control Test	0.079	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

---

Data Download Date/Time 09/03/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VELEZ, HOMMY,

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: PETULA, KYLE

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 21:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: ARBELAEZ

DATE/TIME OF TEST: 08/19/2011 21:47:30

---

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.078	21:38	
Air Blank	0.000	21:39	
Subject Sample #1	0.194	21:41	
Breath Volume	2.078	21:41	
Air Blank	0.000	21:42	
Air Blank	0.000	21:43	
Subject Sample #2	0.179	21:46	
Breath Volume	1.445	21:46	
Air Blank		21:46	Purge Fail
Air Blank	0.000	21:47	

---

Data Download Date/Time 09/03/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VELEZ, HOMMY,

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: PETULLA, KYLE

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 21:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: ARBELAEZ

DATE/TIME OF TEST: 08/19/2011 22:20:49

---

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.077	22:14
Air Blank	0.000	22:14
Subject Sample #1	0.178	22:15
Breath Volume	1.656	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample #2	0.174	22:19
Breath Volume	1.996	22:19
Air Blank	0.000	22:19
Control Test	0.077	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

---

Data Download Date/Time 09/03/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FILLMON, JIMMY, O

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: LOVELACE, AMBER

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 23:11

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 08/21/2011 23:51:26

---

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample #1	0.234	23:47	Volume Not Met
Breath Volume	0.386	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:49	
Subject Sample #2		23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:50	
Control Test	0.076	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

---

Data Download Date/Time 09/03/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CAMPBELL, ROBERT, J

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: ELLER, ASHLEY

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 01:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 08/27/2011 02:14:46

---

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:09
Subject Sample #1	0.185	02:10
Breath Volume	2.074	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample #2	0.176	02:13
Breath Volume	2.558	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

---

Data Download Date/Time 09/03/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DICKEY, JARED, D

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: ELLER, OFC

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 04:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: INIGUEZ

DATE/TIME OF TEST: 08/27/2011 05:30:12

---

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample #1	0.109	05:23
Breath Volume	1.359	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample #2	0.094	05:28
Breath Volume	1.449	05:28
Air Blank	0.000	05:29
Control Test	0.078	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

---

Data Download Date/Time 09/03/2011 13:22

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FILLMON, JIMMY, O

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: ELLER, ASHLEY

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 03:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/28/2011 04:56:38

---

Diagnostics Check	OK	04:55	
Air Blank		04:56	RFI Detect
Air Blank	0.000	04:56	

---

Data Download Date/Time	09/03/2011	13:22
-------------------------	------------	-------

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FILLMON, JIMMY, O

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: ELLER, ASHLEY

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 03:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/28/2011 05:06:29

---

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	05:00
Air Blank	0.000	05:00
Subject Sample #1	0.195	05:01
Breath Volume	1.414	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:03
Subject Sample #2	0.188	05:04
Breath Volume	1.378	05:04
Air Blank	0.000	05:05
Control Test	0.077	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

---

Data Download Date/Time 09/03/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FILLMON, JIMMY, O

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: ELLER, ASHLEY

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 08/08/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 759279

DRY GAS EXPIRY DATE: 04/20/2012

VIOLATION CODE: DUI

ARREST TIME: 04:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BOULTBEE

DATE/TIME OF TEST: 08/30/2011 05:28:23

---

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:26	
Control Test	0.080	05:26	
Air Blank	0.000	05:26	
Subject Sample #1		05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:27	
Control Test	0.080	05:27	
Air Blank	0.000	05:28	
Diagnostics Check	OK	05:28	

---

Data Download Date/Time 09/03/2011 13:22

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ECHEVARRIA, FELIX,

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: ECHEVARRIA, FELIX

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 16:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: CABALLERO

DATE/TIME OF TEST: 08/06/2011 17:02:02

---

Diagnostics Check	OK	16:55
Air Blank	0.000	16:55
Control Test	0.079	16:55
Air Blank	0.000	16:56
Subject Sample #1	0.183	16:57
Breath Volume	1.992	16:57
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample #2	0.197	17:00
Breath Volume	2.843	17:00
Air Blank	0.000	17:01
Control Test	0.077	17:01
Air Blank	0.000	17:01
Diagnostics Check	OK	17:02

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FROST, DIANA, L

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: MONACO, A

ARRESTING OFFICER AGENCY: St. Cloud PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 02:34

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/17/2011 03:19:06

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample #1	0.000	03:14
Breath Volume	2.820	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample #2	0.000	03:17
Breath Volume	2.785	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIEHL, EDWARD,

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: DIEHL, EDWARD

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 18:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: OROURKE

DATE/TIME OF TEST: 08/19/2011 19:43:36

---

Diagnostics Check	OK	19:36
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:38
Subject Sample #1	0.229	19:38
Breath Volume	2.609	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample #2	0.225	19:42
Breath Volume	1.632	19:42
Air Blank	0.000	19:42
Control Test	0.079	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEDULLA, VINCNET, A

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: LAMB, JAMES

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 01:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/20/2011 02:43:53

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Subject Sample #1	0.153	02:39
Breath Volume	2.625	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample #2	0.156	02:42
Breath Volume	2.550	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIEHL, EDWARD,  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: DIEHL, EDWARD  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 646826  
DRY GAS EXPIRY DATE: 07/27/2011  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JALAFF

DATE/TIME OF TEST: 08/20/2011 03:23:06

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample #1	0.203	03:18
Breath Volume	1.171	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample #2	0.205	03:21
Breath Volume	1.929	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIEHL, EDWARD,

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: DIEHL, EDWARD

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 02:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CONTRERAS-MARTIN

DATE/TIME OF TEST: 08/21/2011 03:23:02

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample #1	0.210	03:18
Breath Volume	1.257	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample #2	0.210	03:21
Breath Volume	1.421	03:21
Air Blank	0.000	03:22
Control Test	0.078	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:23

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEDULLA, VINCENT, A

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: LAMB, JAMES

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 18:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/24/2011 19:02:17

---

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Subject Sample #1	0.000	18:57
Breath Volume	3.398	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample #2	0.000	19:00
Breath Volume	3.281	19:00
Air Blank	0.000	19:01
Control Test	0.079	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DELANY, STEPHEN, P

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: FRANCO, CARLOS

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 02:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: DELANY

DATE/TIME OF TEST: 08/25/2011 03:37:26

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.159	03:32
Breath Volume	1.445	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample #2	0.166	03:35
Breath Volume	1.707	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALVAREZ, ROSA, A

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: WACKER, NICHOLAS

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 06:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 08/28/2011 07:14:38

---

Diagnostics Check	OK	07:07
Air Blank	0.000	07:08
Control Test	0.079	07:08
Air Blank	0.000	07:08
Subject Sample #1	0.161	07:09
Breath Volume	2.515	07:09
Air Blank	0.000	07:10
Air Blank	0.000	07:11
Subject Sample #2	0.159	07:12
Breath Volume	2.785	07:12
Air Blank	0.000	07:13
Control Test	0.079	07:14
Air Blank	0.000	07:14
Diagnostics Check	OK	07:14

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONSALVES, KEISHA, A

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: LUGO, N

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 07/27/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 03:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CERVONI

DATE/TIME OF TEST: 08/29/2011 04:37:30

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample #1	0.191	04:32
Breath Volume	1.917	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample #2	0.181	04:35
Breath Volume	1.230	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

---

Data Download Date/Time 08/29/2011 09:18

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DIAZ, ANGEL, M

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: LAMB, JAMES

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 08/29/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 721301

DRY GAS EXPIRY DATE: 02/01/2012

VIOLATION CODE: DUI

ARREST TIME: 22:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: GOMEZ BURGOS

DATE/TIME OF TEST: 08/29/2011 23:55:29

---

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:50
Air Blank	0.000	23:50
Subject Sample #1	0.000	23:51
Breath Volume	2.527	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample #2	0.000	23:54
Breath Volume	2.218	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

---

Data Download Date/Time 09/29/2011 07:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUIDEN, CAROLYN, F  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: LAMB, J  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/29/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 721301  
DRY GAS EXPIRY DATE: 02/01/2012  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: HUTCHINS

DATE/TIME OF TEST: 08/31/2011 00:43:37

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.077	00:37
Air Blank	0.000	00:38
Subject Sample #1	0.000	00:39
Breath Volume	1.386	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample #2	0.000	00:42
Breath Volume	1.324	00:42
Air Blank	0.000	00:42
Control Test	0.076	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time 09/29/2011 07:53

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

INSTRUMENT REGISTERED TO: WINTER PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEADOWS, SHAWN, M  
BREATH TEST OPERATOR AGENCY: Winter Park PD  
ARRESTING OFFICER: MEADOWS, SHAWN  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 07/26/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: BIER

DATE/TIME OF TEST: 08/02/2011 03:33:18

---

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:23	
Subject Sample #1		03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample #2		03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

---

Data Download Date/Time 08/29/2011 11:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

INSTRUMENT REGISTERED TO: WINTER PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAIRBANKS, ANTHONY, M  
BREATH TEST OPERATOR AGENCY: Winter Park PD  
ARRESTING OFFICER: FAIRBANKS, ANTHONY  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 07/26/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 15:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/02/2011 16:47:59

---

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:44	
Control Test	0.080	16:44	
Air Blank	0.000	16:44	
Subject Sample #1		16:46	Subject Test Refused
Breath Volume	0.000	16:46	
Air Blank	0.000	16:47	
Control Test	0.078	16:47	
Air Blank	0.000	16:47	
Diagnostics Check	OK	16:47	

---

Data Download Date/Time 08/29/2011 11:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

INSTRUMENT REGISTERED TO: WINTER PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEADOWS, SHAWN, M  
BREATH TEST OPERATOR AGENCY: Winter Park PD  
ARRESTING OFFICER: MEADOWS, SHAWN  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 07/26/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BURKHALTER

DATE/TIME OF TEST: 08/03/2011 03:43:54

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample #1	0.155	03:39
Breath Volume	1.753	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample #2	0.151	03:42
Breath Volume	1.167	03:42
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time 08/29/2011 11:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

INSTRUMENT REGISTERED TO: WINTER PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COMBAS, JOHN, J  
BREATH TEST OPERATOR AGENCY: Winter Park PD  
ARRESTING OFFICER: WAGGANER, SHARRON  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 07/26/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 08/05/2011 23:26:52

---

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:19	
Subject Sample #1	0.251	23:22	Volume Not Met
Breath Volume	0.753	23:22	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample #2		23:25	Subject Test Refused
Breath Volume	0.000	23:25	
Air Blank	0.000	23:25	
Control Test	0.078	23:26	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

---

Data Download Date/Time 08/29/2011 11:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

INSTRUMENT REGISTERED TO: WINTER PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE, C,  
BREATH TEST OPERATOR AGENCY: Winter Park PD  
ARRESTING OFFICER: VIDRO, K  
ARRESTING OFFICER AGENCY: Winter Park PD  
LAST AGENCY INSPECTION: 07/26/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 781787  
DRY GAS EXPIRY DATE: 06/10/2012  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: DOW

DATE/TIME OF TEST: 08/09/2011 01:52:48

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample #1	0.174	01:48
Breath Volume	5.484	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #2	0.192	01:51
Breath Volume	5.648	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time 08/29/2011 11:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

INSTRUMENT REGISTERED TO: WINTER PARK P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SHORT, WILLIAM, S

BREATH TEST OPERATOR AGENCY: Winter Park PD

ARRESTING OFFICER: SASLO, JOLE

ARRESTING OFFICER AGENCY: Winter Park PD

LAST AGENCY INSPECTION: 07/26/2011

Date or Time Changed: No

DRY GAS STANDARD LOT #: 781787

DRY GAS EXPIRY DATE: 06/10/2012

VIOLATION CODE: DUI

ARREST TIME: 19:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 08/10/2011 20:54:00

---

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:45
Air Blank	0.000	20:46
Subject Sample #1	0.000	20:48
Breath Volume	2.324	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample #2	0.000	20:52
Breath Volume	1.785	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:54

---

Data Download Date/Time 08/29/2011 11:58

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000997

INSTRUMENT REGISTERED TO: ALTAMONTE SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILA, GEORGE, N  
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD  
ARRESTING OFFICER: VILA, GEORGE  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 07/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 08:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: CHERNEGA

DATE/TIME OF TEST: 08/01/2011 09:16:06

---

Diagnostics Check	OK	09:10
Air Blank	0.000	09:10
Control Test	0.080	09:10
Air Blank	0.000	09:11
Subject Sample #1	0.000	09:11
Breath Volume	1.660	09:11
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample #2	0.000	09:14
Breath Volume	2.085	09:14
Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

---

Data Download Date/Time 08/30/2011 04:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000997

INSTRUMENT REGISTERED TO: ALTAMONTE SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DESTEFANO, MICHAEL, N  
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD  
ARRESTING OFFICER: TOZZI, SHAUN  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 07/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 12:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 08/01/2011 14:10:29

---

Diagnostics Check	OK	14:04
Air Blank	0.000	14:04
Control Test	0.079	14:05
Air Blank	0.000	14:05
Subject Sample #1	0.000	14:06
Breath Volume	3.531	14:06
Air Blank	0.000	14:06
Air Blank	0.000	14:08
Subject Sample #2	0.000	14:09
Breath Volume	2.265	14:09
Air Blank	0.000	14:09
Control Test	0.079	14:09
Air Blank	0.000	14:10
Diagnostics Check	OK	14:10

---

Data Download Date/Time 08/30/2011 04:47

# Florida Department of Law Enforcement Alcohol Testing Program

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000997

INSTRUMENT REGISTERED TO: ALTAMONTE SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELOACH, STEVEN, F  
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD  
ARRESTING OFFICER: DELOACH, STEVEN  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 07/24/2011

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 761073  
DRY GAS EXPIRY DATE: 05/06/2012  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CANTERBURY

DATE/TIME OF TEST: 08/09/2011 20:31:15

---

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:25	
Control Test	0.080	20:26	
Air Blank	0.000	20:26	
Subject Sample #1	0.031	20:27	
Breath Volume	2.890	20:27	
Air Blank	0.000	20:27	
Air Blank	0.000	20:29	
Subject Sample #2		20:29	Subject Test Refused
Breath Volume	0.000	20:29	
Air Blank	0.000	20:30	
Control Test	0.080	20:30	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

---

Data Download Date/Time 08/30/2011 04:47