

Suereth, Don

80-000943

LDB

From: John Neiderhiser - 1287 [John.Neiderhiser@colliersheriff.org]
Sent: Monday, February 22, 2010 10:08 AM
To: Suereth, Don
Subject: Electronic Data Review

Don

As to the expiration date for lot 581913 it is 02-23-2011, I must have put in the year for the day. I did use the right date on the inspection done on 01-03-2010 and the instrument was not place in-service until 01-20-2010 when I reentered the lot and date.

I have contact Deputy Lucente as to the login on 80-000943, the additional names were the subject he was attempting to retrieve. I have given him additional training on recalling test and cautioned him as to entering information into the instrument.

Cpl John Neiderhiser
Collier County Sheriff's Office

This communication is confidential and may contain privileged information intended solely for the named addressee(s). It may not be used or disclosed except for the purpose for which it has been sent. If you are not the intended recipient, you must not copy, distribute or take any action induced by or in reliance on information contained in this message. Unless expressly stated, opinions in this message are those of the individual sender and not of the Collier County Sheriff's Office. If you have received this communication in error, please notify the Collier County Sheriff's Office by emailing helpdesk@colliersheriff.org quoting the sender and delete the message and any attached documents. the Collier County Sheriff's Office accepts no liability or responsibility for any onward transmission or use of emails and attachments having left the colliersheriff domain.

RECEIVED**MAR 04 2010****FDLE
Alcohol Testing Program**