

Florida Department of Law Enforcement Alcohol Testing Program
 July 2009 Subject Tests Statistical Data Review
 Messages Report

*Updated
 7/22/10
 UAB*

Serial Number	Agency	# of Tests	Messages (Total Number of Occurrences)																			
			VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	CII Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr	DSP Test	
80-000200	FDLE																					
80-000202	FDLE Training																					
80-000206	FDLE	10	4					1														
80-000219	FDLE																					
80-000220	FDLE Training																					
80-000221	FDLE Training	52					1															
80-000222	FDLE Training																					
80-000223	FDLE																					
80-000224	FDLE																					
80-000225	FDLE Training																					
80-000227	FDLE																					
80-000228	FDLE																					
80-000229	FDLE	28	1			1	1	1	12													
80-000230	FDLE Training																					
80-000737	Alachua County S																					
80-000738	Fort Walton Beach	17				2	1	5														
80-000739	Okaloosa County S	21	2			2	2	6	1													
80-000742	Valparaiso PD																					
80-000744	Escambia County	8	2					4														
80-000749	Escambia County																					
80-000751	Escambia County	59	2			3	4	4	12													
80-000752	Marion County SO																					
80-000759	Jackson County S																					
80-000760	Alachua County S	13	2			2		2														
80-000763	Okaloosa County S	50	4			1		19														
80-000765	Santa Rosa County																					
80-000768	Suwannee County																					
80-000770	Hamilton County S	3	3					2														
80-000771	Perry PD																					
80-000772	Santa Rosa County																					
80-000773	Lafayette County S																					
80-000775	Columbia County S																					
80-000776	Lafayette County S																					

Florida Department of Law Enforcement Alcohol Testing Program
 July 2009 Subject Tests Statistical Data Review
 Messages Report

Messages (Total Number of Occurrences)																						
Serial Number	Agency	# of Tests	VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	Ch Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr	DSP Test	
80-000777	Columbia County S	8	3				1						1									
80-000778	NAS - Jacksonville	2																				
80-000779	Madison County S																					
80-000780	Suwannee County	11	1				3															
80-000781	Washington Count	3																				
80-000782	Washington Count	5																				
80-000784	Holmes County SO	5	1				1	1						2								
80-000785	Holmes County SO																					
80-000786	Broward County S																					
80-000787	Gulf Breeze PD	8				1	3		1													
80-000788	St Lucie County S	3																				
80-000790	Vero Beach PD	3							1													
80-000791	Fort Pierce PD																					
80-000792	St Lucie County S																					
80-000794	St Lucie County S	49	7					3	1													
80-000795	Gulf County SO	6																				
80-000796	Longboat Key PD	2	2																			
80-000797	Hamilton County S																					
80-000808	Hillsborough Count																					
80-000809	Dixie County SO																					
80-000810	Dixie County SO	3	1			1															1	
80-000811	Perry PD	4	2																		1	
80-000812	Hernando County																					
80-000813	Marion County SO	56				2	10								2							
80-000814	Hernando County	15		4			3															
80-000815	Sumter County SO																					
80-000816	Sumter County SO	6					1	4														
80-000817	Citrus County SO																					
80-000819	Citrus County SO	20	4					8														
80-000822	Lake County SO	18					1	5														
80-000823	Lake County SO																					
80-000824	Leesburg PD	9					1	1					1									
80-000825	Mount Dora PD	6					3															

Florida Department of Law Enforcement Alcohol Testing Program

July 2009 Subject Tests Statistical Data Review

Messages Report

Serial Number	Agency	# of Tests	Messages (Total Number of Occurrences)																			
			VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	Cll Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr	DSP Test	
80-000827	Marion County SO	6	1		1																	
80-000829	Hillsborough Count																					
80-000830	Hillsborough Count																					
80-000831	Marin County SO																					
80-000832	Marin County SO	2																				
80-000833	Hillsborough Count																					
80-000836	Umatilla PD																					
80-000838	Panama City Beac																					
80-000840	Lynn Haven PD	1						1														
80-000859	Cape Coral PD	21					2															
80-000861	Sarasota County S																					
80-000862	Blountstown PD	4						2														
80-000867	Monroe County SO																					
80-000868	Monroe County SO	29	3					6												1		
80-000869	Monroe County SO	16						8														
80-000870	Monroe County SO	8			3			5														
80-000871	Coral Gables PD	14	1		2			4			1											
80-000872	Miami PD	43	4		3		1	16												4		
80-000873	Miami PD	9	2		1		1	6											3			
80-000874	Miami PD	16	6		1			6											1			
80-000875	Calhoun County S	1																				
80-000876	Hialeah PD	6			1			4														
80-000878	Homestead PD	6						2														
80-000879	Miami Beach PD	51	4				2	21											3			
80-000880	Miami Dade PD	34	1		1			19														
80-000881	Miami Dade PD	7						3														
80-000882	Miami Dade PD	10	1				2	5														
80-000883	Miami Dade PD	54		10		2		21			1									1		
80-000884	Miami Dade PD	13	1					3												1		
80-000885	Miami Dade PD	26	2				2	6												1		
80-000886	North Miami Beach																					
80-000887	Chattahoochee PD																					
80-000888	Pinellas County SO																					

Florida Department of Law Enforcement Alcohol Testing Program

July 2009 Subject Tests Statistical Data Review

Messages Report

Serial Number	Agency	# of Tests	Messages (Total Number of Occurrences)																			
			VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	Cll Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr	DSP Test	
80-000951	Hendry County SO	6																				
80-000952	Franklin County S	5						1														
80-000953	Franklin County S																					
80-000954	Jefferson County S																					
80-000955	Clewiston PD	2					1															
80-000956	Jefferson County S	4																				
80-000957	Leon County SO	33				1		4														
80-000958	Leon County SO	9						1	2			1										
80-000959	Gadsden County S	17						1	3											1		
80-000960	Liberty County SO	2				1																2
80-000961	Wakulla County S																					
80-000962	Wakulla County S	6							3													
80-000963	Orange County SO	50				6		1	2	5												
80-000964	Orange County SO	54				9		2														
80-000965	Quincy PD																					
80-000966	Osceola County Co	29					1		1	3						1						
80-000967	Apopka PD	3				1				2												
80-000968	Osceola County Co	19				1																
80-000994	Winter Park PD	11						3		4												
80-000995	Cocoa Beach PD	6				1																
80-000996	Titusville PD	12				1				2												
80-000997	Allamonte Springs	4				1																
80-000998	Seminole County S	19				6				1										1		
80-000999	Seminole County S																					
80-001000	Cocoa PD																					
80-001001	Satellite Beach PD	6																				
80-001002	Hardee County SO	3						2		1												
80-001003	Pinellas County SO																					
80-001004	Brevard County SO	7				1																3
80-001005	Pinellas County SO	12								2						1						
80-001006	Indianatic PD	3																				
80-001007	Haines City PD	2																				
80-001008	West Melbourne P	7				2				3												1

Florida Department of Law Enforcement Alcohol Testing Program
 July 2009 Subject Tests Statistical Data Review
 Messages Report

Messages (Total Number of Occurrences)																						
Serial Number	Agency	# of Tests	VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	CH Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr	DSP Test	
80-001230	Tyndall AFB	2																				
80-001231	Auburndale PD	2																				
80-001232	NAS - Pensacola																					
80-001233	Palm Beach Count	60	10		3		3							2								
80-001234	Palm Beach Count	6	1				1										1					
80-001235	West Palm Beach	6	1				1															
80-001236	Palm Beach Count	82	11		3		4	2														
80-001237	Polk County SO																					
80-001238	Groveland PD																					
80-001242	Daytona Beach Sh	5																				
80-001243	Edgewater PD	2																				
80-001244	Putnam County SO	5	1										1									
80-001245	Port Orange PD	3															3					
80-001246	South Daytona PD	22	1				2															1
80-001247	Broward County S	91	4	4	3	3	1															
80-001248	Broward County S																					
80-001249	Broward County S																					
80-001250	Broward County S	9	1		2		3															
80-001253	Brevard County SO	5																				
80-001254	Monroe County SO																					
80-001255	Orange County SO	45	6	2	7		2															
80-001256	Orange County SO	46	11		1		1	2	1													
80-001257	Orange County SO																					
80-001258	Orange County SO																					
80-001259	Pinellas County SO																					
80-001260	Cocoa PD	7	2				1															
80-001261	Rockledge PD	2																				
80-001262	Melbourne PD	32	3	1			3	1							1							
80-001263	Indian Harbour PD	2																				
80-001264	Brevard County SO	4																				
80-001265	Palm Bay PD																					
80-001266	Palm Bay PD	1																				
80-001267	Palm Beach Count	20	6		1		1															

Florida Department of Law Enforcement Alcohol Testing Program

July 2009 Subject Tests Statistical Data Review

Messages Report

Messages (Total Number of Occurrences)																						
Serial Number	Agency	# of Tests	VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	Cll Out Tol	Term Reg	Analy Stabil	EE PROM	Volt Curr	DSP Test	
80-001268																						
80-001269	Palm Beach Court																					
80-001270	Palm Beach Court																					
80-001271	Seminole County S																					
80-001272	Seminole County S	22	1																			
80-001273	Pinellas County SO																					
80-001274	Pinellas County SO																					
80-001275	Hiialeah Gardens P																					
80-001276	Miami Dade PD	9					2															
80-001277	Miami Dade PD	38			1		12															
80-001278	Jacksonville SO	61	1	1			18															
80-001279	Jacksonville SO	85	1			4	34															
80-001280	Jacksonville SO	43	1				14				1											
80-001281	Nassau County SO	11					3															
80-001282	University of Florid	8			4		2															
80-001283	High Springs PD	3			2					1												
80-001284	Gilchrist County S	1																				
80-001285	Gilchrist County S																					
80-001286	Baker County SO																					
80-001287	Baker County SO																					
80-001288	Union County SO	1																				
80-001289	Union County SO																					
80-001290	Stark PD																					
80-001291	Bradford County S	1																				
80-001292	Levy County SO																					
80-001293	Williston PD																					
80-001302	Santa Rosa County	3				1	1															
80-001303	Santa Rosa County	40	1		4		12															
80-001305	Marlana PD																					
80-001306	Mexico Beach PD																					
80-001307	Madison County S	1																				
80-001308	Bay County SO																					
80-001309	Bay County SO	18		2	1		4															

Florida Department of Law Enforcement Alcohol Testing Program

July 2009 Subject Tests Statistical Data Review

Messages Report

Messages (Total Number of Occurrences)																						
Serial Number	Agency	# of Tests	VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	CII Out Tol	Tem Reg	Analy Stabil	EE PROM	Voit Curr	DSP Test	
80-002097	NAS Jacksonville																					
80-002170	NAS Key West																					
80-002171	FSU PD																					
80-002231	New Smyrna Beac	18	1		2		7															
80-002232	Miami Beach PD																					
80-002462	Hiialeah PD																					
80-002463	Hiialeah PD	30	4		1		14									3						
80-002464	Hiialeah PD	3	1		1		2															
80-002566	Surfside PD	10					3															
80-002987	Miami Gardens PD	1																				
80-002988	Miami Gardens PD	4	1			1		1	3													
80-002989	Tallahassee PD	3				1	1									1						
80-003195	FFWCC																					
80-003196	FFWCC																					
80-003197	FWCC Carrabelle																					
80-003224	Miami Beach PD	71	7		2	2	22															
80-003347	Pasco County Sher	2																				
80-003348	Sweetwater PD	8					3															
80-003387	Hillsborough Count	303	19		1	6	40		1							1						
80-003388	Hillsborough Count	67	5				2	1						2		2						
80-003406	Florida Highway Pa																					
80-003407	Florida Highway Pa																					
80-003408	Florida Highway Pa																					
80-003410	Florida Highway Pa																					
80-003411	Florida Highway Pa	42					13															
80-003412	Florida Highway Pa	6					1		1													
80-003413	Florida Highway Pa	1					1															
80-003935	Osceola County Sh	6	1																			
80-003937	Osceola County Sh	12	3													4						
80-003945	Lakeland PD																					
80-004587	Osceola County Sh																					
80-004770	Hurlburt Field Air F																					
80-004771	Hurlburt Field Air F																					

Florida Department of Law Enforcement Alcohol Testing Program
 July 2009 Subject Tests Statistical Data Review
 Messages Report

		Messages (Total Number of Occurrences)																				
Serial Number	Agency	# of Tests	VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	CHI Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr	DSP Test	
80-005057	Okaloosa County S																					
80-005058	Crestview Police D																					
80-005076	Sarasota County S																					
80-005077	North Miami Police																					

		Messages (Total Number of Occurrences)																					
Serial Number	Agency	# of Tests	VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	CHI Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr	DSP Test		
Totals		4058	292	43	139	76	802	9	33	7	16	0	9	22	32	70	1	0	0	4	0		
%			7.20	1.06	3.43	1.87	19.76	0.22	0.81	0.17	0.39	0.00	0.22	0.54	0.79	1.72	0.02	0.00	0.00	0.10	0.00		

VNM = Volume Not Met
 SNM = Slope Not Met
 NSP = No Sample Provided
 RFI = RFI Detect
 REF = Subject Test Refused
 SNL = Slope Not Level
 ABT = Sequence Aborted
 INT = Interferent Detect
 IPS = Improper Sample

RNG = Range Exceeded
 AMB = Ambient Fail
 No .020 = No .020 Agreement
 PUR = Purge Fail
 CHI Out Tol = Control Outside Tolerance
 Temp Reg = Diagnostic Fail - Temperature Regulation Test
 Analy Stabil = Diagnostic Fail - Analytical Stability Test
 EE PROM = Diagnostic Fail - EEPROM Checksum Test
 Volt/Curr = Diagnostic Fail - Voltage/Current Test
 DSP Test = Diagnostic Fail - Digital Signal Processor