









Florida Department of Law Enforcement Alcohol Testing Program

December 2007 Subject Tests Statistical Data Review

Messages Report

Serial Number	Agency	# of Tests	Messages (Total Number of Occurrences)																Vlt Curt	
			VNM	SNM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	Ctl Out Tol	Tem Reg	Analy Stabil		EE PROM
80-000988	Osceola County Co	44	7				4	5												
80-000994	Winter Park PD	11			3		1	3												
80-000985	Cocoa Beach PD	8	1				3													
80-000996	Titusville PD	9	2					1				2								
80-000997	Altamonte Springs																			
80-000998	Seminole County S																			
80-000999	Seminole County S																			
80-001000	Cocoa PD	6																		
80-001001	Satellite Beach PD	9	2																	
80-001002	Hardee County SO	1																		
80-001003	Pinellas County SO	28						6							1	1				
80-001004	Brevard County SO																			
80-001005	Pinellas County SO	93	4		1			30			2									
80-001006	Indianatic PD	1																		
80-001007	Haines City PD	7																		
80-001008	West Melbourne P	2									1									
80-001009	Palm Bay PD	8	2		1			1												
80-001010	Brevard County SO	8	2					1												
80-001018	Jacksonville SO	85	1		1		1	26							3					
80-001042	Frostproof PD																			
80-001043	Highlands County	5																		
80-001044	Coconut Creek PD	11						4												
80-001045	Jacksonville SO	74	2					21												
80-001046	Clay County SO	48	5		12			15							1					
80-001047	Broward County S	90	7		3			6			1									
80-001048	Lake Alfred PD																			
80-001049	Pasco County SO	23	2		5						2									
80-001050	Coral Springs PD	6	1																	
80-001051	St. Pete PD	43	1				3	10			1									
80-001052	Jacksonville SO	81	4				1	23												
80-001053	Nassau County SO	9	1		1		2													
80-001054	Highlands County	25	2					4			1				1					
80-001055	Polk County SO																			
80-001056	Davie PD	7	4																	
80-001057	Lakeland PD	49			4		4	5												
80-001058	Lake Wales PD																			













Florida Department of Law Enforcement Alcohol Testing Program

December 2007 Subject Tests Statistical Data Review

Messages Report

Messages (Total Number of Occurrences)																					
Serial Number	Agency	# of Tests	VNM	SNIM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	Ctl Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr	
80-002463	Hialeah PD	28	1		2		13				2					1					
80-002464	Hialeah PD	13	1			2	3									1					
80-002566	Surfside PD	6					1														
80-002989	Tallahassee PD																				

Messages (Total Number of Occurrences)																				
Serial Number	Agency	# of Tests	VNM	SNIM	NSP	RFI	REF	SNL	ABT	INT	IPS	RNG	AMB	No 0.020	PUR	Ctl Out Tol	Tem Reg	Analy Stabil	EE PROM	Volt Curr
Totals		4747	417	9	153	158	854	11	42	15	14	0	9	21	85	58	9	2	0	1
%			8.78	0.19	3.22	3.33	17.99	0.23	0.88	0.32	0.29	0.00	0.19	0.44	1.79	1.22	0.19	0.04	0.00	0.02

VNM = Volume Not Met  
 SNM = Slope Not Met  
 NSP = No Sample Provided  
 RFI = RFI Detect  
 REF = Subject Test Refused  
 SNL = Slope Not Level  
 ABT = Sequence Aborted  
 INT = Interferent Detect  
 IPS = Improper Sample

RNG = Range Exceeded  
 AMB = Ambient Fall  
 No .020 = No .020 Agreement  
 PUR = Purge Fail  
 Ctl Out Tol = Control Outside Tolerance  
 Temp Reg = Diagnostic Fail - Temperature Regulation Test  
 Analy Stabil = Analytical Stability Test  
 EE PROM = Diagnostic Fail - EEPROM Checksum Test  
 Volt/Curr = Diagnostic Fail - Voltage/Current Test