

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY, ROBERT, P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: KENNEY, SCOTT

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/01/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 01:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: DOBBS

DATE/TIME OF TEST: 06/05/2010 02:22:14

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Subject Sample #1	0.166	02:17
Breath Volume	1.933	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample #2	0.169	02:20
Breath Volume	2.515	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time 06/12/2010 17:57

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONIN, SEAN, J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CRONIN, SEAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/01/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: VAN-HARREN

DATE/TIME OF TEST: 06/06/2010 03:01:35

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Subject Sample #1	0.120	02:56
Breath Volume	2.121	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample #2	0.119	03:00
Breath Volume	2.218	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time 06/12/2010 17:57

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONIN, SEAN, J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CRONIN, SEAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 05/01/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BREAUX

DATE/TIME OF TEST: 06/07/2010 03:51:21

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Subject Sample #1	0.137	03:46
Breath Volume	1.433	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample #2	0.139	03:49
Breath Volume	2.734	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time 06/12/2010 17:57

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MORGAN, DARIN, P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MORGAN, DARIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/01/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 08:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MOYER

DATE/TIME OF TEST: 06/08/2010 11:08:02

Diagnostics Check	OK	11:02
Air Blank	0.000	11:02
Control Test	0.081	11:02
Air Blank	0.000	11:03
Subject Sample #1	0.000	11:03
Breath Volume	1.367	11:03
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample #2	0.000	11:06
Breath Volume	1.304	11:06
Air Blank	0.000	11:07
Control Test	0.080	11:07
Air Blank	0.000	11:07
Diagnostics Check	OK	11:08

Data Download Date/Time 06/12/2010 17:57

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ZIMMERMAN, SCOTT, M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: WERNER, MATT

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/12/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 675059

DRY GAS EXPIRY DATE: 09/25/2011

VIOLATION CODE: DUI

ARREST TIME: 23:34

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PONTREMOLI

DATE/TIME OF TEST: 06/15/2010 00:20:58

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:15
Subject Sample #1	0.187	00:16
Breath Volume	1.695	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample #2	0.193	00:19
Breath Volume	2.625	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time 07/24/2010 09:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONIN, SEAN, J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CRONIN, SEAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 06/18/2010 03:38:45

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample #1	0.140	03:32
Breath Volume	1.175	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample #2	0.141	03:37
Breath Volume	1.398	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time 07/24/2010 09:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONIN, SEAN, J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CRONIN, SEAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/12/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 675059
DRY GAS EXPIRY DATE: 09/25/2011
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BENJAMIN-STANFORD

DATE/TIME OF TEST: 06/19/2010 01:40:01

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Subject Sample #1	0.168	01:35
Breath Volume	1.265	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample #2	0.165	01:38
Breath Volume	1.933	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:40

Data Download Date/Time 07/24/2010 09:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVOIE, KRYSTAL, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ROSARIO, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: FELLNOR

DATE/TIME OF TEST: 06/01/2010 00:51:21

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample #1	0.053	00:46
Breath Volume	1.792	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample #2	0.048	00:49
Breath Volume	2.085	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN, KURTIS, J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ESTERLE, HUGH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: HUSS

DATE/TIME OF TEST: 06/02/2010 05:42:49

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:37
Subject Sample #1	0.251	05:37
Breath Volume	3.847	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample #2	0.249	05:41
Breath Volume	3.332	05:41
Air Blank	0.000	05:41
Control Test	0.077	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/05/2010 01:16:55

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:11
Subject Sample #1	0.119	01:12
Breath Volume	3.398	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:14
Subject Sample #2	0.118	01:15
Breath Volume	3.484	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW, ROBERT, T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KAMALFF, DONAVON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SUESS

DATE/TIME OF TEST: 06/05/2010 03:39:24

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Subject Sample #1	0.239	03:37	Volume Not Met
Breath Volume	0.285	03:37	
Air Blank		03:38	Purge Fail
Air Blank	0.000	03:39	

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/05/2010 04:08:17

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.076	04:01
Air Blank	0.000	04:02
Subject Sample #1	0.250	04:02
Breath Volume	1.925	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.251	04:06
Breath Volume	2.441	04:06
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISINGER, ANDREAS,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PATTON, ROB
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MCGIRT

DATE/TIME OF TEST: 06/05/2010 21:16:47

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.080	21:10
Air Blank	0.000	21:10
Subject Sample #1	0.165	21:12
Breath Volume	1.750	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample #2	0.164	21:15
Breath Volume	1.382	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/06/2010 02:57:19

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:54	
Air Blank	0.000	02:54	
Subject Sample #1		02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/06/2010 03:02:35

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	02:59	
Air Blank	0.000	03:00	
Subject Sample #1		03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIDIO, JAMES, F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DIDIO, JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 06/06/2010 04:41:39

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Subject Sample #1	0.168	04:37
Breath Volume	1.402	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.161	04:39
Breath Volume	1.652	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TABARRACCI, KRISTIN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WALSH, PAUL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 09:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: WOODWARD

DATE/TIME OF TEST: 06/06/2010 10:46:51

Diagnostics Check	OK	10:40
Air Blank	0.000	10:41
Control Test	0.080	10:41
Air Blank	0.000	10:42
Subject Sample #1	0.000	10:42
Breath Volume	1.667	10:42
Air Blank	0.000	10:43
Air Blank	0.000	10:45
Subject Sample #2	0.000	10:45
Breath Volume	1.417	10:45
Air Blank	0.000	10:45
Control Test	0.079	10:46
Air Blank	0.000	10:46
Diagnostics Check	OK	10:46

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERBIEST, RICHARD, M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VERBIEST, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: HULGAS

DATE/TIME OF TEST: 06/06/2010 19:38:56

Diagnostics Check	OK	19:32
Air Blank	0.000	19:32
Control Test	0.079	19:33
Air Blank	0.000	19:33
Subject Sample #1	0.244	19:34
Breath Volume	1.101	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:36
Subject Sample #2	0.240	19:37
Breath Volume	2.242	19:37
Air Blank	0.000	19:37
Control Test	0.078	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBOUR, DANIEL, B
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BROWN, A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/07/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: KORTIS

DATE/TIME OF TEST: 06/07/2010 22:07:28

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample #1	0.145	22:02
Breath Volume	2.425	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample #2	0.148	22:05
Breath Volume	3.242	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time 06/11/2010 11:01

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLAK, ALLAN, S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ROSS, ERIK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: ROUSH

DATE/TIME OF TEST: 06/11/2010 12:21:55

Diagnostics Check	OK	12:15
Air Blank	0.000	12:15
Control Test	0.081	12:16
Air Blank	0.000	12:16
Subject Sample #1	0.000	12:17
Breath Volume	1.273	12:17
Air Blank	0.000	12:17
Air Blank	0.000	12:19
Subject Sample #2	0.000	12:20
Breath Volume	1.234	12:20
Air Blank	0.000	12:20
Control Test	0.080	12:21
Air Blank	0.000	12:21
Diagnostics Check	OK	12:21

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LOPEZ-SOLORZANO

DATE/TIME OF TEST: 06/12/2010 00:26:41

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Subject Sample #1	0.131	00:21
Breath Volume	2.164	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample #2	0.131	00:24
Breath Volume	1.964	00:24
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: LAPERA

DATE/TIME OF TEST: 06/12/2010 02:32:22

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Subject Sample #1	0.034	02:27
Breath Volume	3.492	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample #2	0.034	02:30
Breath Volume	3.613	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYSE, JAIME, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCCLURG, J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: ROUSH

DATE/TIME OF TEST: 06/12/2010 15:31:30

Diagnostics Check	OK	15:25
Air Blank	0.000	15:25
Control Test	0.080	15:26
Air Blank	0.000	15:26
Subject Sample #1	0.000	15:27
Breath Volume	1.652	15:27
Air Blank	0.000	15:27
Air Blank	0.000	15:29
Subject Sample #2	0.000	15:30
Breath Volume	1.910	15:30
Air Blank	0.000	15:30
Control Test	0.079	15:30
Air Blank	0.000	15:31
Diagnostics Check	OK	15:31

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYSE, JAIME, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VIDAL, J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MUNNS

DATE/TIME OF TEST: 06/12/2010 16:22:13

Diagnostics Check	OK	16:16
Air Blank	0.000	16:16
Control Test	0.081	16:16
Air Blank	0.000	16:17
Subject Sample #1	0.000	16:17
Breath Volume	3.351	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample #2	0.000	16:20
Breath Volume	2.382	16:20
Air Blank	0.000	16:21
Control Test	0.077	16:21
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/12/2010 23:42:48

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:37
Air Blank	0.000	23:37
Subject Sample #1	0.136	23:38
Breath Volume	2.960	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample #2	0.136	23:41
Breath Volume	1.882	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE, MICHAEL,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MOORE, MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 06/13/2010 03:15:28

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:09
Subject Sample #1	0.331	03:10
Breath Volume	1.789	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.325	03:13
Breath Volume	1.402	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE, MICHAEL,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MOORE, MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CARIGNAN

DATE/TIME OF TEST: 06/17/2010 02:17:44

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:12
Air Blank	0.000	02:12
Subject Sample #1	0.088	02:13
Breath Volume	1.285	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.085	02:16
Breath Volume	1.898	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONDAS, STEVE, W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KONDAS, STEVE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/17/2010 23:19:06

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.080	23:07	
Air Blank	0.000	23:07	
Subject Sample #1	0.142	23:08	
Breath Volume	1.753	23:08	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample #2	0.139	23:13	Volume Not Met
Breath Volume	0.363	23:13	
Air Blank	0.000	23:14	
Air Blank	0.000	23:16	
Subject Sample #3	0.136	23:17	
Breath Volume	1.464	23:17	
Air Blank	0.000	23:18	
Control Test	0.075	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 06/19/2010 01:44:18

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample #1	0.140	01:38
Breath Volume	1.437	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample #2	0.147	01:42
Breath Volume	2.867	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYLAND, JOSEPH, T
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WAYLAND, JOSEPH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 06/19/2010 23:37:06

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:31
Subject Sample #1	0.223	23:32
Breath Volume	2.000	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample #2	0.215	23:35
Breath Volume	1.519	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYLAND, JOSEPH, T
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CIESLENSKI, JOSEPH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 06/20/2010 01:35:47

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.081	01:30
Air Blank	0.000	01:30
Subject Sample #1	0.000	01:31
Breath Volume	1.816	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample #2	0.000	01:34
Breath Volume	1.285	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI, JOSEPH, P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: OLMER, TODD
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: PAWLOWSKI

DATE/TIME OF TEST: 06/20/2010 02:45:28

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample #1	0.107	02:40
Breath Volume	2.558	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample #2	0.102	02:43
Breath Volume	2.664	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DIAS JR

DATE/TIME OF TEST: 06/20/2010 03:43:45

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample #1	0.161	03:38
Breath Volume	2.332	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.155	03:41
Breath Volume	1.292	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVOIE, KRYSTAL, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE, CARL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 06/20/2010 20:07:24

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:02
Subject Sample #1	0.278	20:02
Breath Volume	2.148	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample #2	0.283	20:05
Breath Volume	2.398	20:05
Air Blank	0.000	20:06
Control Test	0.078	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIDIO, JAMES, F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DIDIO, JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/21/2010 03:08:51

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample #1	0.222	03:00
Breath Volume	1.183	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.200	03:04
Breath Volume	1.175	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample #3	0.199	03:07
Breath Volume	1.539	03:07
Air Blank	0.000	03:07
Control Test	0.076	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS, ROBERT, L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELLIS, ROBERT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BARTOK

DATE/TIME OF TEST: 06/25/2010 22:36:45

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.081	22:29
Air Blank	0.000	22:29
Subject Sample #1	0.188	22:31
Breath Volume	1.437	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:33
Subject Sample #2	0.171	22:34
Breath Volume	1.246	22:34
Air Blank	0.000	22:35
Control Test	0.077	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time 07/09/2010 11:09

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIDIO, JAMES, F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DIDIO, JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FOLGER

DATE/TIME OF TEST: 06/30/2010 02:09:53

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Subject Sample #1	0.010	02:05
Breath Volume	1.175	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample #2	0.000	02:08
Breath Volume	1.164	02:08
Air Blank	0.000	02:08
Control Test	0.077	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVOIE, KRYSTAL, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: OLMER, TODD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 747256
DRY GAS EXPIRY DATE: 03/25/2012
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CUTTING

DATE/TIME OF TEST: 06/30/2010 03:52:24

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.135	03:47
Breath Volume	1.921	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample #2	0.129	03:50
Breath Volume	1.566	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time 07/09/2010 11:09

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUMER, WERNER, R
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: RODRIGUEZ, CARLOS
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822501I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: SOLER

DATE/TIME OF TEST: 06/18/2010 07:59:51

Diagnostics Check	OK	07:56	
Air Blank	0.000	07:56	
Control Test	0.078	07:56	
Air Blank	0.000	07:57	
Subject Sample #1		07:58	Subject Test Refused
Breath Volume	0.000	07:58	
Air Blank	0.000	07:58	
Control Test	0.078	07:59	
Air Blank	0.000	07:59	
Diagnostics Check	OK	07:59	

Data Download Date/Time 06/30/2010 23:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ, JOSE, A
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: DIAZ, J
ARRESTING OFFICER AGENCY: West Miami PD
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822501I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PERDIGON

DATE/TIME OF TEST: 06/19/2010 02:54:55

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.078	02:52	
Air Blank	0.000	02:53	
Subject Sample #1		02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time 06/30/2010 23:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ, JOSE, A
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: VIJIL, J
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822501I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BARANDIARAN

DATE/TIME OF TEST: 06/23/2010 22:22:40

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.077	22:20	
Air Blank	0.000	22:21	
Subject Sample #1		22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.077	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time 06/30/2010 23:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLEY, KEVIN, C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: JOHNSON, CHRISTOPHER
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 822501I
DRY GAS EXPIRY DATE: 08/21/2010
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: IBARRA MARTINEZ

DATE/TIME OF TEST: 06/28/2010 00:04:38

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.077	23:58
Air Blank	0.000	23:58
Subject Sample #1	0.201	23:59
Breath Volume	1.617	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample #2	0.206	00:02
Breath Volume	2.226	00:02
Air Blank	0.000	00:03
Control Test	0.076	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time 06/30/2010 23:04

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESLER, OFC.K., G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RESLER, OFC.K.
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 06/02/2010 03:28:46

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.083	03:23
Air Blank	0.000	03:23
Subject Sample #1	0.174	03:24
Breath Volume	4.105	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample #2	0.172	03:27
Breath Volume	3.718	03:27
Air Blank	0.000	03:27
Control Test	0.082	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time 06/18/2010 07:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU, DAVID, N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU, DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: NEIBERGER

DATE/TIME OF TEST: 06/13/2010 22:07:36

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.082	22:01
Air Blank	0.000	22:01
Subject Sample #1	0.255	22:02
Breath Volume	3.984	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample #2	0.255	22:05
Breath Volume	4.210	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time 06/18/2010 07:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RESLER, OFC.K., G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: RESLER, OFC.K.

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 05/20/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 00:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: ROMO-LOPEZ

DATE/TIME OF TEST: 06/14/2010 00:44:57

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.082	00:39
Air Blank	0.000	00:39
Subject Sample #1	0.207	00:40
Breath Volume	3.648	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample #2	0.206	00:43
Breath Volume	3.578	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time 06/18/2010 07:26

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BODLE, RANDY, S

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: COMINI, PAUL

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/21/2011

VIOLATION CODE: DUI

ARREST TIME: 19:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: EKLUND

DATE/TIME OF TEST: 06/18/2010 19:59:32

Diagnostics Check	OK	19:56	
Air Blank	0.000	19:57	
Control Test	0.083	19:57	
Air Blank	0.000	19:57	
Subject Sample #1		19:58	Subject Test Refused
Breath Volume	0.000	19:58	
Air Blank	0.000	19:58	
Control Test	0.082	19:58	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

Data Download Date/Time 07/17/2010 07:18

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAULK, BRETT, A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: FAULK, BRETT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HAIGHT

DATE/TIME OF TEST: 06/23/2010 02:53:49

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:48
Air Blank	0.000	02:48
Subject Sample #1	0.234	02:49
Breath Volume	3.460	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample #2	0.231	02:52
Breath Volume	3.203	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time 07/17/2010 07:18

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 01:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MOHAMMED

DATE/TIME OF TEST: 06/01/2010 01:58:05

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:53
Subject Sample #1	0.127	01:53
Breath Volume	2.804	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.123	01:56
Breath Volume	2.085	01:56
Air Blank	0.000	01:57
Control Test	0.082	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:58

Data Download Date/Time 06/13/2010 07:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 03:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/01/2010 03:25:23

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Subject Sample #1		03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.082	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time 06/13/2010 07:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOSSER, D, M

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MOSSER, D

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: MCQUILLEN

DATE/TIME OF TEST: 06/01/2010 13:19:36

Diagnostics Check	OK	13:12
Air Blank	0.000	13:12
Control Test	0.082	13:13
Air Blank	0.000	13:13
Subject Sample #1	0.288	13:15
Breath Volume	2.183	13:15
Air Blank	0.000	13:15
Air Blank	0.000	13:17
Subject Sample #2	0.300	13:18
Breath Volume	1.339	13:18
Air Blank	0.000	13:18
Control Test	0.080	13:19
Air Blank	0.000	13:19
Diagnostics Check	OK	13:19

Data Download Date/Time 06/13/2010 07:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BRASWELL, MICHAEL, L

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: KISHTA, RAMEZ

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 05/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 23:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/05/2010 00:25:43

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:20
Subject Sample #1	0.000	00:21
Breath Volume	2.824	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample #2	0.000	00:24
Breath Volume	2.882	00:24
Air Blank	0.000	00:24
Control Test	0.081	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time 06/13/2010 07:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 01:42

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 06/10/2010 02:25:14

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:17
Air Blank	0.000	02:17
Subject Sample #1	0.157	02:19
Breath Volume	1.738	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample #2	0.156	02:23
Breath Volume	1.902	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time 06/13/2010 07:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MCNAMEE

DATE/TIME OF TEST: 06/11/2010 03:18:21

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample #1	0.164	03:12
Breath Volume	2.089	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample #2	0.156	03:16
Breath Volume	1.640	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time 06/13/2010 07:38

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/13/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 00:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 06/16/2010 01:37:12

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Subject Sample #1	0.076	01:32
Breath Volume	1.589	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample #2	0.079	01:35
Breath Volume	1.636	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time 07/11/2010 08:16

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON, TYLER,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON, TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/13/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 01:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DEBATS

DATE/TIME OF TEST: 06/19/2010 02:04:55

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.082	01:59
Air Blank	0.000	01:59
Subject Sample #1	0.212	02:00
Breath Volume	2.250	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample #2	0.216	02:03
Breath Volume	1.894	02:03
Air Blank	0.000	02:04
Control Test	0.082	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time 07/11/2010 08:16

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORTES, C,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: CORTES, C

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/13/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 22:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: BINNS JR

DATE/TIME OF TEST: 06/28/2010 23:13:13

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.081	23:04	
Air Blank	0.000	23:04	
Subject Sample #1		23:07	No Sample Provided
Breath Volume	0.000	23:07	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample #2		23:11	Subject Test Refused
Breath Volume	0.000	23:11	
Air Blank	0.000	23:12	
Control Test	0.081	23:12	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time 07/11/2010 08:16

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINNEY, S, E

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: FINNEY, S

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/13/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 02:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VERRILL JR

DATE/TIME OF TEST: 06/29/2010 03:09:02

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample #1	0.175	03:04
Breath Volume	2.589	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample #2	0.174	03:07
Breath Volume	2.656	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:09

Data Download Date/Time 07/11/2010 08:16

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORTES, C,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: CORTES, C

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/13/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 23:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LETCHWORTH

DATE/TIME OF TEST: 06/30/2010 00:42:07

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.083	00:36
Air Blank	0.000	00:36
Subject Sample #1	0.206	00:37
Breath Volume	1.492	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.198	00:40
Breath Volume	1.238	00:40
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time 07/11/2010 08:16

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADILLA, JESSE, A
BREATH TEST OPERATOR AGENCY: NAS Key West
ARRESTING OFFICER: PADILLA, JESSE
ARRESTING OFFICER AGENCY: NAS Key West
LAST AGENCY INSPECTION: 05/21/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 685214
DRY GAS EXPIRY DATE: 10/28/2011
VIOLATION CODE: OTHER
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 06/08/2010 01:46:07

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Subject Sample #1	0.228	01:41	
Breath Volume	1.101	01:41	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample #2		01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.078	01:45	
Air Blank		01:45	RFI Detect
Air Blank		01:46	RFI Detect

Data Download Date/Time 06/24/2010 11:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUITT, JERRAMI, L
BREATH TEST OPERATOR AGENCY: NAS Key West
ARRESTING OFFICER: PADILLA, JESSE
ARRESTING OFFICER AGENCY: NAS Key West
LAST AGENCY INSPECTION: 05/21/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 685214
DRY GAS EXPIRY DATE: 10/28/2011
VIOLATION CODE: ADMIN
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 06/08/2010 08:54:10

Diagnostics Check	OK	08:47
Air Blank	0.000	08:47
Control Test	0.079	08:48
Air Blank	0.000	08:48
Subject Sample #1	0.116	08:49
Breath Volume	1.433	08:49
Air Blank	0.000	08:50
Air Blank	0.000	08:52
Subject Sample #2	0.122	08:52
Breath Volume	1.921	08:52
Air Blank	0.000	08:53
Control Test	0.079	08:53
Air Blank	0.000	08:54
Diagnostics Check	OK	08:54

Data Download Date/Time 06/24/2010 11:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUITT, JERRAMI, L
BREATH TEST OPERATOR AGENCY: NAS Key West
ARRESTING OFFICER: PADILLA, JESSE
ARRESTING OFFICER AGENCY: NAS Key West
LAST AGENCY INSPECTION: 05/21/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 685214
DRY GAS EXPIRY DATE: 10/28/2011
VIOLATION CODE: ADMIN
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 06/08/2010 09:35:40

Diagnostics Check	OK	09:29
Air Blank	0.000	09:29
Control Test	0.079	09:29
Air Blank	0.000	09:30
Subject Sample #1	0.114	09:31
Breath Volume	2.480	09:31
Air Blank	0.000	09:31
Air Blank	0.000	09:33
Subject Sample #2	0.112	09:34
Breath Volume	2.433	09:34
Air Blank	0.000	09:34
Control Test	0.078	09:35
Air Blank	0.000	09:35
Diagnostics Check	OK	09:35

Data Download Date/Time 06/24/2010 11:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DATKA, JOSHUA, L
BREATH TEST OPERATOR AGENCY: NAS Key West
ARRESTING OFFICER: DATKA, JOSHUA
ARRESTING OFFICER AGENCY: NAS Key West
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 685214
DRY GAS EXPIRY DATE: 10/28/2011
VIOLATION CODE: OTHER
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: BLAGG

DATE/TIME OF TEST: 06/25/2010 12:58:08

Diagnostics Check	OK	12:51	
Air Blank	0.000	12:51	
Control Test	0.079	12:52	
Air Blank	0.000	12:52	
Subject Sample #1	0.000	12:53	
Breath Volume	3.417	12:53	
Air Blank	0.000	12:53	
Air Blank	0.000	12:55	
Subject Sample #2		12:56	Subject Test Refused
Breath Volume	0.269	12:56	
Air Blank	0.000	12:57	
Control Test	0.080	12:57	
Air Blank	0.000	12:58	
Diagnostics Check	OK	12:58	

Data Download Date/Time 07/09/2010 11:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUITT, JERRAMI, L
BREATH TEST OPERATOR AGENCY: NAS Key West
ARRESTING OFFICER: SANDERS, MICHAEL
ARRESTING OFFICER AGENCY: NAS Key West
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 685214
DRY GAS EXPIRY DATE: 10/28/2011
VIOLATION CODE: ADMIN
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: TUOHY

DATE/TIME OF TEST: 06/29/2010 06:12:22

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:58	
Control Test	0.079	05:58	
Air Blank	0.000	05:58	
Subject Sample #1	0.218	06:02	Volume Not Met
Breath Volume	0.218	06:02	
Air Blank	0.000	06:02	
Air Blank	0.000	06:04	
Subject Sample #2	0.214	06:06	
Breath Volume	2.683	06:06	
Air Blank	0.000	06:06	
Air Blank	0.000	06:08	
Subject Sample #3	0.213	06:10	
Breath Volume	1.957	06:10	
Air Blank	0.000	06:11	
Control Test	0.077	06:11	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

Data Download Date/Time 07/09/2010 11:54

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, REGINALD, B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GUZMAN, ANTHONY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/11/2010 04:20:21

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample #1	0.136	04:15
Breath Volume	1.429	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.135	04:18
Breath Volume	1.550	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIRTH, PHILIP, A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: WIRTH, PHILIP
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: LEAR

DATE/TIME OF TEST: 06/11/2010 05:57:19

Diagnostics Check	OK	05:50
Air Blank	0.000	05:51
Control Test	0.080	05:51
Air Blank	0.000	05:52
Subject Sample #1	0.188	05:52
Breath Volume	2.046	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample #2	0.179	05:55
Breath Volume	1.574	05:55
Air Blank	0.000	05:56
Control Test	0.080	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ, ROGER, E
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: LOBELLO, LISA
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: PADILLA ACOSTA

DATE/TIME OF TEST: 06/12/2010 02:15:34

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.082	02:10
Air Blank	0.000	02:10
Subject Sample #1	0.132	02:11
Breath Volume	2.515	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample #2	0.136	02:14
Breath Volume	2.828	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ, JOEL,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MC VEY, G
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: HYMAN

DATE/TIME OF TEST: 06/12/2010 04:14:59

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:01	
Subject Sample #1		04:04	No Sample Provided
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample #2	0.180	04:09	
Breath Volume	1.222	04:09	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample #3	0.184	04:13	
Breath Volume	1.113	04:13	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIRTH, PHILIP, A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: WIRTH, PHILIP
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/12/2010 05:07:01

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample #1	0.159	05:02
Breath Volume	1.871	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample #2	0.164	05:05
Breath Volume	2.171	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:07

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, IAN, M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MARTINEZ, ROGER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: SOLANO

DATE/TIME OF TEST: 06/12/2010 19:28:00

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:25	
Control Test	0.081	19:25	
Air Blank	0.000	19:26	
Subject Sample #1		19:26	Subject Test Refused
Breath Volume	0.000	19:26	
Air Blank	0.000	19:27	
Control Test	0.080	19:27	
Air Blank	0.000	19:27	
Diagnostics Check	OK	19:28	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ, ROGER, E
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GRANT, REID
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 06/12/2010 23:49:41

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.081	23:47	
Air Blank	0.000	23:48	
Subject Sample #1		23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:48	
Control Test	0.080	23:49	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSNER, STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ESTRAVIZ, ALBERTO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SEAS

DATE/TIME OF TEST: 06/16/2010 03:13:57

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:07
Air Blank	0.000	03:07
Subject Sample #1	0.224	03:09
Breath Volume	2.062	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample #2	0.210	03:12
Breath Volume	1.375	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSNER, STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER, STEVE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: RIVERA-FUNEZ

DATE/TIME OF TEST: 06/16/2010 05:36:09

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Subject Sample #1	0.091	05:31
Breath Volume	1.523	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample #2	0.092	05:34
Breath Volume	1.367	05:34
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCVEY, GREGORY, B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOUZA, A
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SIBILIA

DATE/TIME OF TEST: 06/17/2010 04:08:01

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.082	03:55	
Air Blank	0.000	03:56	
Subject Sample #1	0.122	03:59	Slope Not Level
Breath Volume	1.332	03:59	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample #2	0.120	04:03	
Breath Volume	1.222	04:03	
Air Blank	0.000	04:04	
Air Blank	0.000	04:05	
Subject Sample #3	0.119	04:06	
Breath Volume	1.421	04:06	
Air Blank	0.000	04:07	
Control Test	0.077	04:07	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:08	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSNER, STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER, STEVE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/19/2010 03:05:48

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample #1	0.213	03:01
Breath Volume	3.269	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.209	03:04
Breath Volume	2.238	03:04
Air Blank	0.000	03:04
Control Test	0.078	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, REGINALD, B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: POLANCO, RICHARD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 06/19/2010 04:09:48

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample #1	0.102	04:05
Breath Volume	1.691	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample #2	0.094	04:08
Breath Volume	1.394	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, REGINALD, B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FLEISCHHAUER, THOMAS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ZAMBRANO

DATE/TIME OF TEST: 06/19/2010 05:29:03

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:22
Subject Sample #1	0.133	05:22
Breath Volume	2.140	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:24
Subject Sample #2	0.129	05:27
Breath Volume	1.820	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:29

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYEGH, MICHELLE, M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SAYEGH, MICHELLE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 06/19/2010 22:15:40

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.082	22:09
Air Blank	0.000	22:10
Subject Sample #1	0.127	22:11
Breath Volume	1.414	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample #2	0.131	22:14
Breath Volume	1.769	22:14
Air Blank	0.000	22:14
Control Test	0.079	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORISME, SMITH,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: BLACIO, GUSTAVO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SALCEDO

DATE/TIME OF TEST: 06/20/2010 06:17:29

Diagnostics Check	OK	06:14	
Air Blank	0.000	06:14	
Control Test	0.081	06:15	
Air Blank	0.000	06:15	
Subject Sample #1		06:16	Subject Test Refused
Breath Volume	0.000	06:16	
Air Blank	0.000	06:16	
Control Test	0.081	06:16	
Air Blank	0.000	06:17	
Diagnostics Check	OK	06:17	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYEGH, MICHELLE, M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ROJAS, JAIME
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: VIEYTO

DATE/TIME OF TEST: 06/20/2010 17:21:21

Diagnostics Check	OK	17:14
Air Blank	0.000	17:15
Control Test	0.080	17:15
Air Blank	0.000	17:16
Subject Sample #1	0.202	17:16
Breath Volume	2.152	17:16
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample #2	0.205	17:19
Breath Volume	1.925	17:19
Air Blank	0.000	17:20
Control Test	0.079	17:20
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYEGH, MICHELLE, M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CHIN, ALAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: GROBACK

DATE/TIME OF TEST: 06/20/2010 22:41:01

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample #1	0.116	22:35
Breath Volume	2.628	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample #2	0.114	22:39
Breath Volume	2.332	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:41

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCVEY, GREGORY, B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VILLA, L
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: OKUR

DATE/TIME OF TEST: 06/21/2010 03:23:35

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample #1		03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.080	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYEGH, MICHELLE, M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FLEISCHAUER, THOMAS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: FERRARI

DATE/TIME OF TEST: 06/22/2010 04:55:01

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.081	04:52	
Air Blank	0.000	04:53	
Subject Sample #1		04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	
Control Test	0.081	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:55	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYEGH, MICHELLE, M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ESTRAVIZ, ALBERTO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: PERNAS

DATE/TIME OF TEST: 06/22/2010 05:45:50

Diagnostics Check	OK	05:42	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Subject Sample #1		05:44	Subject Test Refused
Breath Volume	0.000	05:44	
Air Blank	0.000	05:44	
Control Test	0.080	05:45	
Air Blank	0.000	05:45	
Diagnostics Check	OK	05:45	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUNCAS, ANDREW, J
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: BOLDUC, KENNETH
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: PEREZ ICABALZEA

DATE/TIME OF TEST: 06/23/2010 04:50:04

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.081	04:42
Air Blank	0.000	04:42
Subject Sample #1	0.184	04:45
Breath Volume	1.582	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample #2	0.172	04:48
Breath Volume	1.375	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:50

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSNER, STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER, STEVE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 06/24/2010 04:45:12

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:39
Subject Sample #1	0.244	04:40
Breath Volume	2.125	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample #2	0.239	04:43
Breath Volume	1.894	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ, ROGER, E
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ROBINSON, IAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PALACIOS-NODA

DATE/TIME OF TEST: 06/24/2010 22:46:52

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample #1	0.191	22:42
Breath Volume	3.230	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample #2	0.199	22:45
Breath Volume	2.011	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON, IAN, M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MARTINEZ, ROGER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/26/2010 19:13:08

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:10	
Control Test	0.081	19:11	
Air Blank	0.000	19:11	
Subject Sample #1		19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:12	
Control Test	0.080	19:12	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCVEY, GREGORY, B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOUZA, A
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ERRO

DATE/TIME OF TEST: 06/27/2010 05:41:40

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.080	05:39	
Air Blank	0.000	05:40	
Subject Sample #1		05:40	Subject Test Refused
Breath Volume	0.000	05:40	
Air Blank	0.000	05:40	
Control Test	0.080	05:41	
Air Blank	0.000	05:41	
Diagnostics Check	OK	05:41	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESORE, STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FLEISCHHAUR, THOMAS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 06/29/2010 04:38:13

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:31
Subject Sample #1	0.134	04:33
Breath Volume	4.437	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample #2	0.137	04:36
Breath Volume	4.300	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUNCAS, ANDREW, J
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ESTRAVIZ, ALBERTO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: DE ARMAS

DATE/TIME OF TEST: 06/30/2010 05:35:14

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:34	
Control Test	0.152	05:34	Control Outside Tolerance
Air Blank	0.000	05:35	

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUNCAS, ANDREW, J
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ESTRAVIZ, ALBERTO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/10/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: DE ARMAS

DATE/TIME OF TEST: 06/30/2010 05:43:31

Diagnostics Check	OK	05:37
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Subject Sample #1	0.153	05:39
Breath Volume	2.566	05:39
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample #2	0.147	05:42
Breath Volume	2.398	05:42
Air Blank	0.000	05:42
Control Test	0.079	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time 07/06/2010 11:35

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CONSALVO, ESTEBAN

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 04:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: PAEZ

DATE/TIME OF TEST: 06/11/2010 05:11:46

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:09	
Control Test	0.079	05:09	
Air Blank	0.000	05:10	
Subject Sample #1		05:10	Subject Test Refused
Breath Volume	0.000	05:10	
Air Blank	0.000	05:10	
Control Test	0.079	05:11	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

Data Download Date/Time 06/18/2010 12:19

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CRUZ, EVI,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: CRUZ, EVI

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 00:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/16/2010 01:41:16

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Subject Sample #1	0.137	01:35	
Breath Volume	1.199	01:35	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample #2	0.141	01:39	
Breath Volume	1.160	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank		01:40	Improper Sample
Air Blank	0.000	01:41	

Data Download Date/Time 06/18/2010 12:19

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: FABIEN, J

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 03:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/20/2010 04:12:49

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:06
Subject Sample #1	0.258	04:07
Breath Volume	1.496	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.249	04:11
Breath Volume	1.480	04:11
Air Blank	0.000	04:11
Control Test	0.077	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time 07/21/2010 11:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SANCHEZ, RAUL,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: SANCHEZ, RAUL

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 07:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 06/20/2010 08:40:09

Diagnostics Check	OK	08:37	
Air Blank	0.000	08:37	
Control Test	0.079	08:38	
Air Blank	0.000	08:38	
Subject Sample #1		08:38	Subject Test Refused
Breath Volume	0.000	08:38	
Air Blank	0.000	08:39	
Control Test	0.080	08:39	
Air Blank	0.000	08:40	
Diagnostics Check	OK	08:40	

Data Download Date/Time 07/21/2010 11:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CONSALVO, ESTEBAN, A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: ESPINOSA, VICTOR

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CHAVARRIA

DATE/TIME OF TEST: 06/25/2010 02:45:04

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample #1	0.175	02:37
Breath Volume	2.335	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample #2	0.152	02:40
Breath Volume	1.148	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample #3	0.168	02:43
Breath Volume	2.593	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:45

Data Download Date/Time 07/21/2010 11:17

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, M,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: GOITIA, R

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 18:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 06/03/2010 18:57:29

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:51	
Control Test	0.079	18:51	
Air Blank	0.000	18:52	
Subject Sample #1	0.262	18:52	
Breath Volume	1.269	18:52	
Air Blank	0.000	18:53	
Air Blank	0.000	18:55	
Subject Sample #2	0.271	18:55	
Breath Volume	2.320	18:55	
Air Blank	0.000	18:56	
Control Test	0.074	18:56	Control Outside Tolerance
Air Blank	0.000	18:57	

Data Download Date/Time 06/18/2010 12:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MEDINA, ANTHONY,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: LOPEZ, NICK

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 20:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: AMARO

DATE/TIME OF TEST: 06/13/2010 21:15:09

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.079	21:08
Air Blank	0.000	21:08
Subject Sample #1	0.267	21:10
Breath Volume	1.871	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:12
Subject Sample #2	0.252	21:13
Breath Volume	1.867	21:13
Air Blank	0.000	21:14
Control Test	0.077	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time 06/18/2010 12:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ, C,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DELIMA, G

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 21:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: REDONDO

DATE/TIME OF TEST: 06/16/2010 22:24:40

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.077	22:19
Air Blank	0.000	22:19
Subject Sample #1	0.250	22:19
Breath Volume	2.316	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample #2	0.250	22:22
Breath Volume	2.949	22:22
Air Blank	0.000	22:23
Control Test	0.076	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time 06/18/2010 12:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DELIMA, G,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DIEUDONNE, K

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 21:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 06/18/2010 21:51:51

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:44	
Subject Sample #1	0.179	21:44	
Breath Volume	1.507	21:44	
Air Blank	0.000	21:45	
Air Blank	0.000	21:47	
Subject Sample #2		21:49	RFI Detect
Breath Volume	0.878	21:49	
Air Blank	0.000	21:50	
Control Test	0.077	21:51	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time 07/21/2010 11:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DELIMA, G,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DIEUDONNE, K

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 21:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 06/18/2010 22:03:38

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:56	
Control Test	0.076	21:56	
Air Blank	0.000	21:57	
Subject Sample #1	0.170	21:57	
Breath Volume	1.187	21:57	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample #2	0.167	22:01	
Breath Volume	1.789	22:01	
Air Blank	0.000	22:02	
Control Test	0.074	22:03	Control Outside Tolerance
Air Blank	0.000	22:03	

Data Download Date/Time 07/21/2010 11:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DELIMA, G,

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DIEUDONNE, K

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 06/18/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 822101I

DRY GAS EXPIRY DATE: 08/11/2010

VIOLATION CODE: DUI

ARREST TIME: 21:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 06/18/2010 22:16:24

Diagnostics Check	OK	22:07
Air Blank	0.000	22:08
Control Test	0.076	22:08
Air Blank	0.000	22:09
Subject Sample #1	0.171	22:10
Breath Volume	2.222	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:12
Subject Sample #2	0.169	22:14
Breath Volume	2.175	22:14
Air Blank	0.000	22:15
Control Test	0.076	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time 07/21/2010 11:13

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, JORGE, L
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: RODRIGUEZ, JORGE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SUMMERALL

DATE/TIME OF TEST: 06/05/2010 22:14:52

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
Control Test	0.079	22:07	
Air Blank	0.000	22:07	
Subject Sample #1		22:10	No Sample Provided
Breath Volume	0.000	22:10	
Air Blank	0.000	22:11	
Air Blank	0.000	22:13	
Subject Sample #2		22:14	Sequence Aborted
Breath Volume	0.000	22:14	
Air Blank	0.000	22:14	

Data Download Date/Time 07/01/2010 14:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, JORGE, L
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: RODRIGUEZ, JORGE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: PANHOO

DATE/TIME OF TEST: 06/06/2010 19:08:09

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.079	19:02
Air Blank	0.000	19:03
Subject Sample #1	0.220	19:03
Breath Volume	2.843	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample #2	0.208	19:06
Breath Volume	1.671	19:06
Air Blank	0.000	19:07
Control Test	0.077	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time 07/01/2010 14:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIOS, JAVIER,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARRIOS, JAVIER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 06/12/2010 18:05:34

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:02	
Control Test	0.079	18:02	
Air Blank	0.000	18:02	
Subject Sample #1		18:04	Subject Test Refused
Breath Volume	0.000	18:04	
Air Blank	0.000	18:04	
Control Test	0.079	18:04	
Air Blank	0.000	18:05	
Diagnostics Check	OK	18:05	

Data Download Date/Time 07/01/2010 14:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARRIOS, JAVIER,

BREATH TEST OPERATOR AGENCY: OTHER Local

ARRESTING OFFICER: BARRIOS, JAVIER

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 688217

DRY GAS EXPIRY DATE: 11/09/2011

VIOLATION CODE: DUI

ARREST TIME: 21:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/17/2010 22:00:13

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.080	21:56	
Air Blank	0.000	21:57	
Subject Sample #1		21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.079	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time 07/01/2010 14:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, JORGE, L
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: RODRIGUEZ, JORGE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/18/2010 21:34:20

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:33	
Control Test	0.079	21:33	
Air Blank		21:33	Improper Sample
Air Blank	0.000	21:34	

Data Download Date/Time 07/01/2010 14:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, JORGE, L
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: RODRIGUEZ, JORGE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/18/2010 21:42:42

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:37
Subject Sample #1	0.227	21:38
Breath Volume	2.191	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:40
Subject Sample #2	0.223	21:41
Breath Volume	2.863	21:41
Air Blank	0.000	21:41
Control Test	0.076	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time 07/01/2010 14:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, J,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: WILLIAMS, J
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 06/20/2010 14:36:31

Diagnostics Check	OK	14:33	
Air Blank	0.000	14:33	
Control Test	0.078	14:34	
Air Blank	0.000	14:34	
Subject Sample #1		14:35	Subject Test Refused
Breath Volume	0.000	14:35	
Air Blank	0.000	14:35	
Control Test	0.078	14:35	
Air Blank	0.000	14:36	
Diagnostics Check	OK	14:36	

Data Download Date/Time 07/01/2010 14:42

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILONE, JEREMY, M
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TRAPAGA, STEVE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/23/2010 02:31:48

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:25
Subject Sample #1	0.160	02:27
Breath Volume	2.285	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample #2	0.168	02:30
Breath Volume	3.820	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time 07/01/2010 14:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILONE, JEREMY, M
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: RIVAS, ENRIQUE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 688217
DRY GAS EXPIRY DATE: 11/09/2011
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GARCIA FLORES

DATE/TIME OF TEST: 06/28/2010 01:55:39

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.078	01:52	
Air Blank	0.000	01:53	
Subject Sample #1		01:54	Slope Not Met
Breath Volume	1.429	01:54	
Air Blank	0.000	01:54	
Control Test	0.078	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time 07/01/2010 14:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PILONE, JEREMY, M

BREATH TEST OPERATOR AGENCY: OTHER Local

ARRESTING OFFICER: RIVAS, ENRIQUE

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 688217

DRY GAS EXPIRY DATE: 11/09/2011

VIOLATION CODE: DUI

ARREST TIME: 01:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GARCIA FLORES

DATE/TIME OF TEST: 06/28/2010 02:00:32

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.077	01:58	
Air Blank	0.000	01:58	
Subject Sample #1		01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.078	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time 07/01/2010 14:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE, A,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MERESSE, A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 11/21/2011
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BRASHER

DATE/TIME OF TEST: 06/03/2010 23:27:03

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Subject Sample #1	0.213	23:22
Breath Volume	1.949	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample #2	0.216	23:25
Breath Volume	1.375	23:25
Air Blank	0.000	23:26
Control Test	0.076	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:27

Data Download Date/Time 07/01/2010 12:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE, A,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MERESSE, A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 11/21/2011
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: SPINCIC

DATE/TIME OF TEST: 06/09/2010 22:56:24

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:55	
Control Test	0.078	22:55	
Air Blank		22:55	RFI Detect
Air Blank		22:56	RFI Detect

Data Download Date/Time 07/01/2010 12:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE, A,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MERESSE, A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 11/21/2011
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/11/2010 00:26:23

Diagnostics Check	OK	00:25	
Air Blank		00:25	RFI Detect
Air Blank	0.000	00:26	

Data Download Date/Time	07/01/2010	12:23
-------------------------	------------	-------

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE, A,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MERESSE, A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/20/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 11/21/2011
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/11/2010 00:37:33

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:28
Subject Sample #1	0.190	00:31
Breath Volume	1.171	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample #2	0.180	00:35
Breath Volume	1.332	00:35
Air Blank	0.000	00:36
Control Test	0.076	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time 07/01/2010 12:23

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETO-VO, L, O
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: N/A, N/A
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: COURT ORD
ARREST TIME: 09:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:22

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 06/01/2010 09:48:52

Diagnostics Check	OK	09:43
Air Blank	0.000	09:43
Control Test	0.078	09:43
Air Blank	0.000	09:44
Subject Sample #1	0.000	09:44
Breath Volume	1.996	09:44
Air Blank	0.000	09:45
Air Blank	0.000	09:47
Subject Sample #2	0.000	09:47
Breath Volume	2.148	09:47
Air Blank	0.000	09:47
Control Test	0.078	09:48
Air Blank	0.000	09:48
Diagnostics Check	OK	09:48

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVAGNI, R, W
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SILVAGNI, R
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ABRIL

DATE/TIME OF TEST: 06/02/2010 03:00:42

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:58	
Air Blank	0.000	02:58	
Subject Sample #1		02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time 06/10/2010 11:27

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMPHRY, KERRY, R
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER, S
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: LO YACONO

DATE/TIME OF TEST: 06/02/2010 05:01:24

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Subject Sample #1	0.031	04:56
Breath Volume	2.199	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample #2	0.030	04:59
Breath Volume	2.589	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, REGINALD, B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SHIMKO, RONALD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 06/03/2010 01:42:35

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.075	01:36
Air Blank	0.000	01:36
Subject Sample #1	0.155	01:37
Breath Volume	1.589	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample #2	0.147	01:40
Breath Volume	1.386	01:40
Air Blank	0.000	01:41
Control Test	0.077	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVAGNI, R, W
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SILVAGNI, R
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: TRABANINO

DATE/TIME OF TEST: 06/04/2010 01:59:15

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Subject Sample #1	0.077	01:54
Breath Volume	3.402	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample #2	0.070	01:57
Breath Volume	1.730	01:57
Air Blank	0.000	01:58
Control Test	0.077	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVAGNI, R, W
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SILVAGNI, R
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: TENEKEDES

DATE/TIME OF TEST: 06/05/2010 04:10:46

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Subject Sample #1	0.147	04:05
Breath Volume	1.480	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample #2	0.155	04:09
Breath Volume	2.277	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIRTH, PHILIP, A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ESTRAVIZ, A
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 06/05/2010 05:04:56

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:02	
Control Test	0.079	05:02	
Air Blank	0.000	05:03	
Subject Sample #1		05:03	Subject Test Refused
Breath Volume	0.000	05:03	
Air Blank	0.000	05:04	
Control Test	0.078	05:04	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:04	

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ, JOEL,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SAZONAF, V
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: VINAS

DATE/TIME OF TEST: 06/05/2010 05:50:22

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.078	05:43
Air Blank	0.000	05:44
Subject Sample #1	0.138	05:45
Breath Volume	1.773	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample #2	0.137	05:48
Breath Volume	2.351	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ, FREDERICK, A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PEREZ, LUIS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: MANZANARES

DATE/TIME OF TEST: 06/05/2010 19:30:17

Diagnostics Check	OK	19:23
Air Blank	0.000	19:24
Control Test	0.077	19:24
Air Blank	0.000	19:24
Subject Sample #1	0.289	19:25
Breath Volume	2.296	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:27
Subject Sample #2	0.286	19:28
Breath Volume	2.050	19:28
Air Blank	0.000	19:29
Control Test	0.077	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON, REGINALD, B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ROBINSON, IAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/05/2010 23:07:23

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.075	23:05	
Air Blank	0.000	23:05	
Subject Sample #1		23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.077	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETO-VO, L, O
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: NONE, NONE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: COURT ORD
ARREST TIME: 07:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 06/08/2010 08:23:34

Diagnostics Check	OK	08:17
Air Blank	0.000	08:18
Control Test	0.078	08:18
Air Blank	0.000	08:19
Subject Sample #1	0.000	08:19
Breath Volume	1.843	08:19
Air Blank	0.000	08:19
Air Blank	0.000	08:21
Subject Sample #2	0.000	08:22
Breath Volume	1.945	08:22
Air Blank	0.000	08:22
Control Test	0.078	08:22
Air Blank	0.000	08:23
Diagnostics Check	OK	08:23

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES, MANUEL,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CHIN, ALAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/26/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 750981
DRY GAS EXPIRY DATE: 04/09/2012
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 06/08/2010 16:25:03

Diagnostics Check	OK	16:17
Air Blank	0.000	16:18
Control Test	0.077	16:18
Air Blank	0.000	16:19
Subject Sample #1	0.320	16:20
Breath Volume	1.371	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:22
Subject Sample #2	0.318	16:23
Breath Volume	1.191	16:23
Air Blank	0.000	16:24
Control Test	0.076	16:24
Air Blank	0.000	16:24
Diagnostics Check	OK	16:25

Data Download Date/Time 06/10/2010 11:27

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MONSEGUR, ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: FAIN

DATE/TIME OF TEST: 06/24/2010 00:48:20

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Subject Sample #1	0.000	00:44
Breath Volume	3.230	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample #2	0.000	00:46
Breath Volume	2.792	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time 07/14/2010 11:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, R, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EVANS, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/25/2010 00:44:21

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Subject Sample #1	0.097	00:39
Breath Volume	3.101	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample #2	0.096	00:42
Breath Volume	3.093	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time 07/14/2010 11:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: NIMONS

DATE/TIME OF TEST: 06/25/2010 01:51:30

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.078	01:39
Air Blank	0.000	01:39
Subject Sample #1	0.245	01:41
Breath Volume	1.410	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample #2	0.214	01:46
Breath Volume	1.835	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample #3	0.210	01:50
Breath Volume	1.152	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time 07/14/2010 11:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUCK, SCOTT, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SCHUCK, SCOTT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: PEREZ-JIMENES

DATE/TIME OF TEST: 06/26/2010 04:01:50

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:56
Subject Sample #1	0.242	03:57
Breath Volume	2.226	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample #2	0.248	04:00
Breath Volume	2.492	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time 07/14/2010 11:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR, ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MONSEGUR, ROBERT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/26/2010 05:02:07

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.079	04:51	
Air Blank	0.000	04:51	
Subject Sample #1	0.221	04:54	Volume Not Met
Breath Volume	0.101	04:54	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample #2	0.206	05:00	Volume Not Met
Breath Volume	0.195	05:00	
Air Blank	0.000	05:01	
Control Test	0.077	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time 07/14/2010 11:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, R, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EVANS, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/23/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BETZ

DATE/TIME OF TEST: 06/30/2010 22:14:41

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.078	22:11	
Air Blank	0.000	22:12	
Subject Sample #1		22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:13	
Control Test	0.077	22:14	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time 07/14/2010 11:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLLADO, E, R

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: ORTEGA, R

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 05/31/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/28/2011

VIOLATION CODE: DUI

ARREST TIME: 01:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ONEGA

DATE/TIME OF TEST: 06/06/2010 03:06:52

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Subject Sample #1	0.249	03:01
Breath Volume	2.035	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.237	03:05
Breath Volume	1.769	03:05
Air Blank	0.000	03:05
Control Test	0.078	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time 06/11/2010 05:51

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLM, CARL, A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HOLM, CARL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/28/2011
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CHOSED

DATE/TIME OF TEST: 06/13/2010 23:52:56

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.079	23:41	
Air Blank	0.000	23:42	
Subject Sample #1	0.186	23:45	Volume Not Met
Breath Volume	0.894	23:45	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample #2	0.165	23:51	Volume Not Met
Breath Volume	0.101	23:51	
Air Blank	0.000	23:52	
Control Test	0.074	23:52	Control Outside Tolerance
Air Blank	0.000	23:52	

Data Download Date/Time 07/29/2010 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FETTERS, MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: FETTERS, MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/28/2011
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 06/25/2010 23:37:13

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.077	23:29
Air Blank	0.000	23:30
Subject Sample #1	0.131	23:32
Breath Volume	1.582	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample #2	0.127	23:35
Breath Volume	1.472	23:35
Air Blank	0.000	23:36
Control Test	0.075	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time 07/29/2010 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANIO, PIERRE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: UBIETA, GUIERMO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/28/2011
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 06/28/2010 00:34:39

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.078	00:31	
Air Blank	0.000	00:32	
Subject Sample #1		00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.077	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time 07/29/2010 19:44

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARLISLE, J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: AMAYA

DATE/TIME OF TEST: 06/01/2010 02:09:17

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Subject Sample #1	0.187	02:04
Breath Volume	1.996	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample #2	0.189	02:07
Breath Volume	2.062	02:07
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: CORONA-MEDINA

DATE/TIME OF TEST: 06/01/2010 05:10:32

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.082	04:56	
Air Blank	0.000	04:56	
Subject Sample #1	0.147	04:59	Volume Not Met
Breath Volume	0.460	04:59	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample #2	0.150	05:03	
Breath Volume	1.335	05:03	
Air Blank	0.000	05:03	
Air Blank	0.000	05:05	
Subject Sample #3	0.148	05:08	Volume Not Met
Breath Volume	0.542	05:08	
Air Blank	0.000	05:09	
Control Test	0.079	05:09	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: SOLORIO NAVARRO

DATE/TIME OF TEST: 06/02/2010 02:04:51

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.081	01:52	
Air Blank	0.000	01:52	
Subject Sample #1	0.205	01:54	
Breath Volume	1.210	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample #2	0.197	02:00	Volume Not Met
Breath Volume	0.085	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:02	
Subject Sample #3	0.195	02:03	
Breath Volume	1.320	02:03	
Air Blank	0.000	02:03	
Control Test	0.078	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 06/02/2010 04:19:17

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Subject Sample #1	0.125	04:14
Breath Volume	1.632	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample #2	0.126	04:17
Breath Volume	1.988	04:17
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: KELLY-PERRY

DATE/TIME OF TEST: 06/02/2010 04:33:42

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:23	
Subject Sample #1	0.165	04:26	Volume Not Met
Breath Volume	0.093	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:28	
Subject Sample #2	0.152	04:32	Volume Not Met
Breath Volume	0.144	04:32	
Air Blank	0.000	04:32	
Control Test	0.079	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: KELLY-PERRY

DATE/TIME OF TEST: 06/02/2010 04:49:44

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.079	04:36	
Air Blank	0.000	04:36	
Subject Sample #1	0.166	04:39	Volume Not Met
Breath Volume	0.308	04:39	
Air Blank	0.000	04:40	
Air Blank	0.000	04:42	
Subject Sample #2	0.169	04:43	
Breath Volume	1.164	04:43	
Air Blank	0.000	04:44	
Air Blank	0.000	04:46	
Subject Sample #3	0.164	04:48	
Breath Volume	1.722	04:48	
Air Blank	0.000	04:48	
Control Test	0.079	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/02/2010 05:18:53

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Subject Sample #1	0.150	05:14
Breath Volume	2.601	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:16
Subject Sample #2	0.146	05:17
Breath Volume	2.109	05:17
Air Blank	0.000	05:17
Control Test	0.079	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SPANGLER, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 06/02/2010 15:56:16

Diagnostics Check	OK	15:50
Air Blank	0.000	15:50
Control Test	0.081	15:50
Air Blank	0.000	15:51
Subject Sample #1	0.145	15:51
Breath Volume	2.488	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample #2	0.149	15:54
Breath Volume	2.593	15:54
Air Blank	0.000	15:55
Control Test	0.080	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THOMPSON, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: WYATTE

DATE/TIME OF TEST: 06/02/2010 17:39:41

Diagnostics Check	OK	17:32
Air Blank	0.000	17:32
Control Test	0.081	17:33
Air Blank	0.000	17:33
Subject Sample #1	0.217	17:34
Breath Volume	1.929	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample #2	0.211	17:38
Breath Volume	2.406	17:38
Air Blank	0.000	17:38
Control Test	0.077	17:39
Air Blank	0.000	17:39
Diagnostics Check	OK	17:39

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCLAIN, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 06/02/2010 21:06:14

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.081	20:50	
Air Blank	0.000	20:51	
Subject Sample #1	0.252	20:53	
Breath Volume	1.257	20:53	
Air Blank	0.000	20:53	
Air Blank	0.000	20:55	
Subject Sample #2	0.224	20:58	Volume Not Met
Breath Volume	0.886	20:58	
Air Blank	0.000	20:59	
Air Blank	0.000	21:01	
Subject Sample #3	0.217	21:04	Volume Not Met
Breath Volume	0.148	21:04	
Air Blank	0.000	21:05	
Control Test	0.077	21:05	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCLAIN, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 06/02/2010 21:15:18

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.077	21:08
Air Blank	0.000	21:09
Subject Sample #1	0.248	21:10
Breath Volume	1.550	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample #2	0.252	21:13
Breath Volume	1.480	21:13
Air Blank	0.000	21:14
Control Test	0.078	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOSTER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: LINDE

DATE/TIME OF TEST: 06/02/2010 23:19:28

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample #1	0.089	23:15
Breath Volume	2.457	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample #2	0.085	23:17
Breath Volume	2.671	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PELLETIER

DATE/TIME OF TEST: 06/03/2010 00:10:14

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Subject Sample #1	0.195	00:05
Breath Volume	1.273	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample #2	0.190	00:08
Breath Volume	1.152	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 06/03/2010 01:58:21

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Subject Sample #1	0.133	01:53
Breath Volume	3.156	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample #2	0.134	01:56
Breath Volume	3.171	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PARISH

DATE/TIME OF TEST: 06/03/2010 04:12:34

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Subject Sample #1	0.174	04:08
Breath Volume	1.164	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample #2	0.173	04:11
Breath Volume	1.238	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUTHERS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 06/03/2010 04:49:13

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Subject Sample #1	0.153	04:44
Breath Volume	1.828	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample #2	0.158	04:47
Breath Volume	1.656	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FAUSKEE, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/03/2010 05:15:32

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Subject Sample #1	0.243	05:07
Breath Volume	1.160	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample #2	0.216	05:10
Breath Volume	1.246	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample #3	0.252	05:13
Breath Volume	1.238	05:13
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORRELL, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/03/2010 05:34:44

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.080	05:29
Air Blank	0.000	05:29
Subject Sample #1	0.182	05:30
Breath Volume	1.597	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample #2	0.184	05:33
Breath Volume	1.703	05:33
Air Blank	0.000	05:33
Control Test	0.080	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BODAMER, C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: RINGOLD

DATE/TIME OF TEST: 06/04/2010 02:18:01

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.082	02:09
Air Blank	0.000	02:10
Subject Sample #1	0.177	02:12
Breath Volume	2.101	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample #2	0.175	02:16
Breath Volume	1.820	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:18

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORET, F
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: FERREIRA FERREIRA

DATE/TIME OF TEST: 06/04/2010 05:06:48

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Subject Sample #1	0.258	05:01
Breath Volume	1.898	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample #2	0.250	05:05
Breath Volume	1.886	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: PRADO

DATE/TIME OF TEST: 06/04/2010 21:07:38

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.081	21:05	
Air Blank	0.000	21:05	
Subject Sample #1		21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.081	21:07	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORREY, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MILLERE

DATE/TIME OF TEST: 06/04/2010 22:42:11

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.082	22:37
Air Blank	0.000	22:37
Subject Sample #1	0.000	22:37
Breath Volume	2.265	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample #2	0.000	22:40
Breath Volume	2.894	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BENENATI

DATE/TIME OF TEST: 06/05/2010 00:11:33

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:06
Air Blank	0.000	00:06
Subject Sample #1	0.013	00:07
Breath Volume	2.964	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample #2	0.011	00:10
Breath Volume	2.792	00:10
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/05/2010 01:11:11

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:05
Air Blank	0.000	01:05
Subject Sample #1	0.187	01:06
Breath Volume	2.531	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample #2	0.195	01:09
Breath Volume	2.839	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCMURPHY, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DIMAS

DATE/TIME OF TEST: 06/05/2010 02:26:59

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample #1	0.240	02:22
Breath Volume	2.796	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample #2	0.251	02:25
Breath Volume	1.417	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORRELL, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ARMENTEROS

DATE/TIME OF TEST: 06/05/2010 03:06:46

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Subject Sample #1	0.109	03:02
Breath Volume	2.140	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample #2	0.117	03:05
Breath Volume	2.789	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOSTER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SARMIENTO

DATE/TIME OF TEST: 06/05/2010 03:43:35

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Subject Sample #1	0.163	03:38
Breath Volume	2.765	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample #2	0.166	03:42
Breath Volume	2.187	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAMER, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PALMISANO

DATE/TIME OF TEST: 06/05/2010 04:25:49

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:20
Subject Sample #1	0.198	04:20
Breath Volume	1.292	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample #2	0.191	04:24
Breath Volume	1.660	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: TUBBS

DATE/TIME OF TEST: 06/05/2010 20:13:17

Diagnostics Check	OK	20:06
Air Blank	0.000	20:07
Control Test	0.082	20:07
Air Blank	0.000	20:08
Subject Sample #1	0.000	20:08
Breath Volume	2.035	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample #2	0.000	20:11
Breath Volume	2.578	20:11
Air Blank	0.000	20:12
Control Test	0.082	20:12
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOSTER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 06/05/2010 22:35:19

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:29	
Control Test	0.081	22:29	
Air Blank	0.000	22:30	
Subject Sample #1	0.145	22:30	
Breath Volume	1.937	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample #2	0.140	22:33	
Breath Volume	1.722	22:33	
Air Blank	0.000	22:34	
Control Test	0.081	22:34	Interferent Detect
Air Blank	0.000	22:35	

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEAD, K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 06/06/2010 00:06:46

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample #1	0.000	00:02
Breath Volume	1.601	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample #2	0.000	00:05
Breath Volume	1.242	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LAROSE, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 06/06/2010 00:54:07

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Subject Sample #1	0.168	00:49
Breath Volume	2.367	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample #2	0.170	00:52
Breath Volume	1.414	00:52
Air Blank	0.000	00:53
Control Test	0.081	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCCALLA, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 06/06/2010 02:04:31

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.082	01:57
Air Blank	0.000	01:57
Subject Sample #1	0.188	01:59
Breath Volume	1.843	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample #2	0.193	02:02
Breath Volume	1.515	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: CANDELARIA

DATE/TIME OF TEST: 06/06/2010 03:21:53

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Subject Sample #1	0.143	03:16
Breath Volume	2.042	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample #2	0.149	03:20
Breath Volume	1.765	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAMER, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HESS

DATE/TIME OF TEST: 06/06/2010 03:31:05

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Subject Sample #1	0.064	03:26
Breath Volume	2.777	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample #2	0.063	03:29
Breath Volume	3.406	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORREY, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/06/2010 03:54:21

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample #1	0.017	03:49
Breath Volume	1.570	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample #2	0.017	03:52
Breath Volume	1.640	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: ZELKOFF

DATE/TIME OF TEST: 06/06/2010 22:49:35

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.082	22:43
Air Blank	0.000	22:44
Subject Sample #1	0.101	22:44
Breath Volume	2.613	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample #2	0.102	22:48
Breath Volume	2.906	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WHITEHEAD, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 06/07/2010 00:10:00

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.082	00:02
Air Blank	0.000	00:02
Subject Sample #1	0.174	00:04
Breath Volume	1.125	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample #2	0.170	00:08
Breath Volume	1.226	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:10

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLAN-CINTRON, E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 06/07/2010 00:44:59

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Subject Sample #1	0.296	00:40
Breath Volume	1.871	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample #2	0.308	00:43
Breath Volume	1.597	00:43
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 06/07/2010 04:18:50

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:12
Subject Sample #1	0.174	04:14
Breath Volume	1.562	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample #2	0.185	04:17
Breath Volume	1.707	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLAN-CINTRON, E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: IVARRA-SILES

DATE/TIME OF TEST: 06/08/2010 02:14:35

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Subject Sample #1	0.121	02:09
Breath Volume	3.125	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample #2	0.117	02:13
Breath Volume	3.074	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time 06/08/2010 13:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURAN, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: WEIR

DATE/TIME OF TEST: 06/08/2010 21:41:01

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.081	21:34
Air Blank	0.000	21:35
Subject Sample #1	0.032	21:35
Breath Volume	2.277	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample #2	0.029	21:39
Breath Volume	2.148	21:39
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:41

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUTHERS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/08/2010 23:13:53

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample #1	0.252	23:08
Breath Volume	1.449	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:10
Subject Sample #2	0.257	23:12
Breath Volume	1.257	23:12
Air Blank	0.000	23:12
Control Test	0.079	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ISOM

DATE/TIME OF TEST: 06/09/2010 01:38:30

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:33
Air Blank	0.000	01:33
Subject Sample #1	0.122	01:34
Breath Volume	3.984	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample #2	0.124	01:36
Breath Volume	3.699	01:36
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BRITT-STATZ

DATE/TIME OF TEST: 06/09/2010 03:20:53

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.081	03:15
Air Blank	0.000	03:15
Subject Sample #1	0.098	03:16
Breath Volume	1.472	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample #2	0.091	03:19
Breath Volume	1.187	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BRICAUD

DATE/TIME OF TEST: 06/10/2010 00:12:29

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:06
Subject Sample #1	0.089	00:07
Breath Volume	1.398	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample #2	0.083	00:10
Breath Volume	1.578	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBOSA, F
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/10/2010 02:29:28

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.082	02:23
Air Blank	0.000	02:23
Subject Sample #1	0.195	02:24
Breath Volume	1.503	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample #2	0.184	02:27
Breath Volume	1.207	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GILSON

DATE/TIME OF TEST: 06/10/2010 02:51:28

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Subject Sample #1	0.078	02:46
Breath Volume	1.816	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample #2	0.083	02:49
Breath Volume	2.316	02:49
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRCIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOSTER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 06/10/2010 06:05:20

Diagnostics Check	OK	05:59
Air Blank	0.000	05:59
Control Test	0.081	06:00
Air Blank	0.000	06:00
Subject Sample #1	0.000	06:01
Breath Volume	1.113	06:01
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample #2	0.000	06:03
Breath Volume	2.351	06:03
Air Blank	0.000	06:04
Control Test	0.081	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE, RICHARD, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIAMPORCARO, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: ADMIN
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: GARY

DATE/TIME OF TEST: 06/10/2010 10:03:12

Diagnostics Check	OK	09:56	
Air Blank	0.000	09:57	
Control Test	0.081	09:57	
Air Blank	0.000	09:57	
Subject Sample #1	0.382	09:58	
Breath Volume	2.550	09:58	
Air Blank	0.000	09:59	
Air Blank	0.000	10:01	
Subject Sample #2		10:01	Slope Not Met
Breath Volume	1.324	10:01	
Air Blank	0.000	10:02	
Control Test	0.080	10:02	
Air Blank	0.000	10:03	
Diagnostics Check	OK	10:03	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE, RICHARD, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GEE, W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/10/2010 17:45:26

Diagnostics Check	OK	17:36
Air Blank	0.000	17:36
Control Test	0.081	17:36
Air Blank	0.000	17:37
Subject Sample #1	0.105	17:40
Breath Volume	1.789	17:40
Air Blank	0.000	17:41
Air Blank	0.000	17:42
Subject Sample #2	0.105	17:43
Breath Volume	1.984	17:43
Air Blank	0.000	17:44
Control Test	0.080	17:44
Air Blank	0.000	17:45
Diagnostics Check	OK	17:45

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORREY, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: SUMNER

DATE/TIME OF TEST: 06/10/2010 18:57:23

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.081	18:52
Air Blank	0.000	18:52
Subject Sample #1	0.000	18:53
Breath Volume	2.375	18:53
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample #2	0.000	18:56
Breath Volume	2.347	18:56
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORRELL, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: VARNEDOE

DATE/TIME OF TEST: 06/11/2010 03:05:32

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.082	02:59
Air Blank	0.000	03:00
Subject Sample #1	0.150	03:00
Breath Volume	2.597	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample #2	0.152	03:03
Breath Volume	2.203	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 06/11/2010 03:45:15

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Subject Sample #1	0.142	03:40
Breath Volume	1.890	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample #2	0.140	03:43
Breath Volume	1.378	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BAKALYAR

DATE/TIME OF TEST: 06/11/2010 04:08:22

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:01
Subject Sample #1	0.183	04:03
Breath Volume	2.718	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.183	04:06
Breath Volume	2.160	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUSTEK, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/11/2010 04:21:00

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample #1	0.205	04:15
Breath Volume	1.957	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample #2	0.199	04:19
Breath Volume	2.296	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:21

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOSTER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: LAINA

DATE/TIME OF TEST: 06/11/2010 05:29:53

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.082	05:24
Air Blank	0.000	05:24
Subject Sample #1	0.088	05:25
Breath Volume	3.636	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:27
Subject Sample #2	0.088	05:28
Breath Volume	3.226	05:28
Air Blank	0.000	05:28
Control Test	0.082	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 06/11/2010 05:59:50

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:47	
Control Test	0.082	05:47	
Air Blank	0.000	05:48	
Subject Sample #1	0.261	05:51	Volume Not Met
Breath Volume	0.117	05:51	
Air Blank	0.000	05:52	
Air Blank	0.000	05:53	
Subject Sample #2	0.263	05:55	
Breath Volume	1.222	05:55	
Air Blank	0.000	05:56	
Air Blank	0.000	05:57	
Subject Sample #3	0.270	05:58	
Breath Volume	1.597	05:58	
Air Blank	0.000	05:58	
Control Test	0.079	05:59	
Air Blank	0.000	05:59	
Diagnostics Check	OK	05:59	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: INSIGNARES, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/11/2010 21:17:05

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Subject Sample #1	0.207	21:12
Breath Volume	1.324	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample #2	0.204	21:15
Breath Volume	1.375	21:15
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEAD, K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: HARROD

DATE/TIME OF TEST: 06/11/2010 22:33:42

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:18	
Control Test	0.081	22:18	
Air Blank	0.000	22:19	
Subject Sample #1	0.322	22:22	
Breath Volume	1.191	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:24	
Subject Sample #2	0.280	22:28	Volume Not Met
Breath Volume	0.265	22:28	
Air Blank	0.000	22:28	
Air Blank	0.000	22:30	
Subject Sample #3	0.323	22:31	
Breath Volume	1.253	22:31	
Air Blank	0.000	22:32	
Control Test	0.078	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DONAWAY

DATE/TIME OF TEST: 06/11/2010 23:16:13

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample #1	0.188	23:11
Breath Volume	1.960	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample #2	0.183	23:14
Breath Volume	1.281	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERNANDEZ, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: BROSHEARS

DATE/TIME OF TEST: 06/12/2010 00:27:17

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Subject Sample #1	0.153	00:22
Breath Volume	1.750	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample #2	0.146	00:25
Breath Volume	1.402	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: INDERWIESEN

DATE/TIME OF TEST: 06/12/2010 00:42:50

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample #1	0.196	00:38
Breath Volume	2.152	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.190	00:41
Breath Volume	1.785	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOBSON, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CAIRNIE

DATE/TIME OF TEST: 06/12/2010 01:29:38

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample #1	0.153	01:25
Breath Volume	2.210	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample #2	0.144	01:28
Breath Volume	1.601	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 06/12/2010 02:15:06

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:09
Subject Sample #1	0.147	02:10
Breath Volume	1.156	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample #2	0.145	02:13
Breath Volume	1.648	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/12/2010 02:35:55

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:30
Subject Sample #1	0.140	02:31
Breath Volume	3.500	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample #2	0.138	02:34
Breath Volume	4.078	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FAUSKEE, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GODFREY

DATE/TIME OF TEST: 06/12/2010 02:48:54

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:43
Subject Sample #1	0.189	02:44
Breath Volume	1.316	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample #2	0.190	02:47
Breath Volume	1.492	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOBSON, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ZERANSKI

DATE/TIME OF TEST: 06/12/2010 04:07:38

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:00
Subject Sample #1	0.153	04:02
Breath Volume	1.382	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample #2	0.155	04:06
Breath Volume	1.519	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: INSIGNARES, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 06/12/2010 04:24:35

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Subject Sample #1	0.187	04:19
Breath Volume	1.410	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample #2	0.186	04:22
Breath Volume	1.742	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SIRMANS

DATE/TIME OF TEST: 06/12/2010 04:42:06

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Subject Sample #1	0.097	04:37
Breath Volume	1.164	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:39
Subject Sample #2	0.106	04:40
Breath Volume	1.234	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: COLLIE

DATE/TIME OF TEST: 06/12/2010 04:54:18

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:45
Subject Sample #1	0.108	04:48
Breath Volume	1.382	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample #2	0.110	04:52
Breath Volume	1.875	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURAN, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 06/12/2010 05:48:59

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.081	05:42
Air Blank	0.000	05:43
Subject Sample #1	0.088	05:44
Breath Volume	1.968	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample #2	0.082	05:47
Breath Volume	1.687	05:47
Air Blank	0.000	05:48
Control Test	0.080	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 06/12/2010 21:31:45

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.082	21:26
Air Blank	0.000	21:26
Subject Sample #1	0.000	21:27
Breath Volume	2.105	21:27
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample #2	0.000	21:30
Breath Volume	1.750	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 06/12/2010 23:05:11

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:02	
Control Test	0.081	23:02	
Air Blank	0.000	23:03	
Subject Sample #1		23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.082	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 06/12/2010 23:15:13

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:10
Subject Sample #1	0.101	23:10
Breath Volume	1.785	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample #2	0.105	23:13
Breath Volume	2.183	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHANAN, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: LARREA

DATE/TIME OF TEST: 06/13/2010 00:06:19

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.082	00:00
Air Blank	0.000	00:00
Subject Sample #1	0.127	00:01
Breath Volume	1.796	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample #2	0.128	00:04
Breath Volume	1.421	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEAD, K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: RESTALL

DATE/TIME OF TEST: 06/13/2010 01:08:18

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Subject Sample #1		01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 06/13/2010 03:00:48

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Subject Sample #1	0.212	02:55
Breath Volume	2.468	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample #2	0.208	02:59
Breath Volume	2.683	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MELO, JR
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: OLEARY

DATE/TIME OF TEST: 06/13/2010 03:30:37

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Subject Sample #1	0.110	03:23
Breath Volume	1.363	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample #2	0.133	03:25
Breath Volume	3.296	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample #3	0.120	03:29
Breath Volume	2.070	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEAD, K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CAPPACETTI-COSME

DATE/TIME OF TEST: 06/13/2010 04:21:30

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.081	04:16
Air Blank	0.000	04:16
Subject Sample #1	0.108	04:17
Breath Volume	2.335	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample #2	0.103	04:19
Breath Volume	1.945	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAUSS, MARTIN, J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CASTRO, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 06/13/2010 08:42:51

Diagnostics Check	OK	08:36
Air Blank	0.000	08:36
Control Test	0.082	08:37
Air Blank	0.000	08:37
Subject Sample #1	0.000	08:38
Breath Volume	1.574	08:38
Air Blank	0.000	08:38
Air Blank	0.000	08:40
Subject Sample #2	0.000	08:41
Breath Volume	1.816	08:41
Air Blank	0.000	08:41
Control Test	0.082	08:42
Air Blank	0.000	08:42
Diagnostics Check	OK	08:42

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAUSS, MARTIN, J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SPANGLER, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:28

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/13/2010 13:54:10

Diagnostics Check	OK	13:48	
Air Blank	0.000	13:48	
Control Test	0.081	13:49	
Air Blank	0.000	13:49	
Subject Sample #1		13:52	Subject Test Refused
Breath Volume	0.152	13:52	
Air Blank	0.000	13:53	
Control Test	0.081	13:53	
Air Blank	0.000	13:54	
Diagnostics Check	OK	13:54	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAUSS, MARTIN, J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SPANGLER, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:28

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/13/2010 14:15:12

Diagnostics Check	OK	14:00	
Air Blank	0.000	14:01	
Control Test	0.081	14:01	
Air Blank	0.000	14:02	
Subject Sample #1	0.278	14:05	Volume Not Met
Breath Volume	0.375	14:05	
Air Blank	0.000	14:06	
Air Blank	0.000	14:07	
Subject Sample #2	0.239	14:09	No .020 agreement
Breath Volume	1.597	14:09	
Air Blank	0.000	14:10	
Air Blank	0.000	14:12	
Subject Sample #3	0.264	14:13	No .020 agreement
Breath Volume	1.648	14:13	
Air Blank	0.000	14:14	
Control Test	0.078	14:14	
Air Blank	0.000	14:15	
Diagnostics Check	OK	14:15	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAUSS, MARTIN, J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SPANGLER, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:28

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/13/2010 14:27:34

Diagnostics Check	OK	14:20
Air Blank	0.000	14:20
Control Test	0.079	14:20
Air Blank	0.000	14:21
Subject Sample #1	0.279	14:23
Breath Volume	1.820	14:23
Air Blank	0.000	14:23
Air Blank	0.000	14:25
Subject Sample #2	0.272	14:25
Breath Volume	2.605	14:25
Air Blank	0.000	14:26
Control Test	0.078	14:26
Air Blank	0.000	14:27
Diagnostics Check	OK	14:27

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 06/13/2010 21:21:13

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.081	21:15
Air Blank	0.000	21:16
Subject Sample #1	0.291	21:16
Breath Volume	1.542	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample #2	0.287	21:19
Breath Volume	1.996	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 06/14/2010 02:11:25

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Subject Sample #1	0.134	02:06
Breath Volume	2.582	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample #2	0.138	02:09
Breath Volume	3.207	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEAD, K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: MARROQUIN

DATE/TIME OF TEST: 06/14/2010 04:51:35

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.081	04:45
Air Blank	0.000	04:45
Subject Sample #1	0.120	04:47
Breath Volume	2.960	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample #2	0.120	04:50
Breath Volume	2.558	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 06/14/2010 19:16:29

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.082	19:10
Air Blank	0.000	19:11
Subject Sample #1	0.218	19:12
Breath Volume	2.609	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample #2	0.221	19:14
Breath Volume	2.441	19:14
Air Blank	0.000	19:15
Control Test	0.081	19:15
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: PFRIEM

DATE/TIME OF TEST: 06/14/2010 22:28:51

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.082	22:23
Air Blank	0.000	22:23
Subject Sample #1	0.350	22:24
Breath Volume	1.828	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample #2	0.347	22:27
Breath Volume	2.570	22:27
Air Blank	0.000	22:27
Control Test	0.081	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ALLENDE

DATE/TIME OF TEST: 06/14/2010 23:14:04

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.082	23:07
Air Blank	0.000	23:08
Subject Sample #1	0.000	23:09
Breath Volume	2.722	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample #2	0.000	23:12
Breath Volume	2.453	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:14

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COOPER, W
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BOLAND

DATE/TIME OF TEST: 06/15/2010 02:35:31

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.082	02:30
Air Blank	0.000	02:30
Subject Sample #1	0.000	02:31
Breath Volume	1.917	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample #2	0.000	02:34
Breath Volume	1.535	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVERAS, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DELASALA

DATE/TIME OF TEST: 06/15/2010 23:41:06

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.082	23:34
Air Blank	0.000	23:34
Subject Sample #1	0.187	23:36
Breath Volume	1.757	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample #2	0.195	23:39
Breath Volume	1.566	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: VISNESKY

DATE/TIME OF TEST: 06/17/2010 00:34:16

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Subject Sample #1	0.143	00:29
Breath Volume	1.152	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample #2	0.156	00:32
Breath Volume	2.062	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: LAMBE

DATE/TIME OF TEST: 06/17/2010 01:52:06

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.000	01:47
Breath Volume	2.363	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample #2	0.000	01:50
Breath Volume	2.699	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 06/17/2010 03:07:25

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Subject Sample #1	0.193	03:03
Breath Volume	2.113	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample #2	0.193	03:05
Breath Volume	1.625	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BOUTWELL

DATE/TIME OF TEST: 06/17/2010 04:04:19

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample #1	0.136	03:59
Breath Volume	2.652	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample #2	0.134	04:02
Breath Volume	2.332	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUSTEK, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 06/17/2010 04:14:40

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:09
Subject Sample #1	0.097	04:10
Breath Volume	2.363	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample #2	0.099	04:13
Breath Volume	2.269	04:13
Air Blank	0.000	04:13
Control Test	0.082	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAUSS, MARTIN, J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEVOE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: QUEEN

DATE/TIME OF TEST: 06/17/2010 17:03:26

Diagnostics Check	OK	16:57
Air Blank	0.000	16:57
Control Test	0.081	16:57
Air Blank	0.000	16:58
Subject Sample #1	0.070	16:58
Breath Volume	4.785	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample #2	0.071	17:01
Breath Volume	5.019	17:01
Air Blank	0.000	17:02
Control Test	0.081	17:02
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WISLON, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/18/2010 00:33:07

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample #1	0.264	00:28
Breath Volume	1.917	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample #2	0.259	00:31
Breath Volume	2.046	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCLAIN, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: HANKS

DATE/TIME OF TEST: 06/18/2010 01:07:27

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:02
Air Blank	0.000	01:02
Subject Sample #1	0.089	01:03
Breath Volume	2.511	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample #2	0.095	01:05
Breath Volume	2.167	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVERAS, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FARROW

DATE/TIME OF TEST: 06/18/2010 01:53:01

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample #1	0.150	01:48
Breath Volume	1.687	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample #2	0.153	01:51
Breath Volume	1.679	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:53

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORET, F
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BLANCO LIZARDO

DATE/TIME OF TEST: 06/18/2010 02:21:56

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.082	02:15
Air Blank	0.000	02:16
Subject Sample #1	0.114	02:17
Breath Volume	3.382	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample #2	0.111	02:20
Breath Volume	2.675	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: KINSER

DATE/TIME OF TEST: 06/18/2010 02:35:02

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:33	
Subject Sample #1		02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:35	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HOLLETT

DATE/TIME OF TEST: 06/18/2010 02:53:04

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.082	02:47
Air Blank	0.000	02:48
Subject Sample #1	0.211	02:48
Breath Volume	1.789	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample #2	0.199	02:51
Breath Volume	1.800	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:53

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: KINSER

DATE/TIME OF TEST: 06/18/2010 03:02:17

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample #1	0.184	02:57
Breath Volume	2.324	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample #2	0.177	03:00
Breath Volume	2.039	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GANNASCOLI

DATE/TIME OF TEST: 06/18/2010 04:00:41

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.082	03:54
Air Blank	0.000	03:55
Subject Sample #1	0.111	03:56
Breath Volume	2.343	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample #2	0.111	03:59
Breath Volume	2.121	03:59
Air Blank	0.000	03:59
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 06/18/2010 04:29:05

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Subject Sample #1	0.110	04:24
Breath Volume	1.871	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample #2	0.115	04:27
Breath Volume	2.386	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER, B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: REARDON

DATE/TIME OF TEST: 06/18/2010 20:06:10

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.081	20:03	
Air Blank	0.000	20:04	
Subject Sample #1	0.211	20:04	
Breath Volume	1.812	20:04	
Air Blank		20:04	RFI Detect
Air Blank	0.000	20:05	
Control Test	0.081	20:05	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:06	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER, B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: REARDON

DATE/TIME OF TEST: 06/18/2010 20:15:38

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.081	20:08
Air Blank	0.000	20:09
Subject Sample #1	0.210	20:10
Breath Volume	1.992	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:12
Subject Sample #2	0.212	20:13
Breath Volume	2.054	20:13
Air Blank	0.000	20:14
Control Test	0.080	20:15
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ROEDING

DATE/TIME OF TEST: 06/19/2010 01:13:46

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	00:59	
Subject Sample #1	0.165	01:03	Volume Not Met
Breath Volume	1.097	01:03	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample #2	0.163	01:06	
Breath Volume	1.746	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:08	
Subject Sample #3	0.153	01:12	Volume Not Met
Breath Volume	0.632	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ROEDING

DATE/TIME OF TEST: 06/19/2010 01:23:02

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample #1	0.174	01:16
Breath Volume	2.058	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample #2	0.161	01:21
Breath Volume	1.750	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:23

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SCHULMAN

DATE/TIME OF TEST: 06/19/2010 02:05:13

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.081	01:59
Air Blank	0.000	01:59
Subject Sample #1	0.137	02:00
Breath Volume	1.296	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample #2	0.133	02:03
Breath Volume	1.550	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/19/2010 02:15:58

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:10
Subject Sample #1	0.225	02:11
Breath Volume	2.488	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample #2	0.213	02:14
Breath Volume	1.312	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCCALLA, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/19/2010 02:22:14

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:18	
Subject Sample #1		02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORET, F
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ELLERBEE

DATE/TIME OF TEST: 06/19/2010 02:33:23

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample #1	0.164	02:28
Breath Volume	1.796	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample #2	0.164	02:31
Breath Volume	2.230	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUSTEK, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: DAUSE

DATE/TIME OF TEST: 06/19/2010 03:08:00

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Subject Sample #1	0.141	03:03
Breath Volume	2.289	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample #2	0.133	03:06
Breath Volume	1.718	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:08

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WHITEHEAD, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/19/2010 03:15:43

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Subject Sample #1	0.236	03:10	
Breath Volume	3.640	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample #2	0.229	03:13	
Breath Volume	3.398	03:13	
Air Blank	0.000	03:14	
Control Test	0.081	03:14	Interferent Detect
Air Blank		03:15	Purge Fail

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WHITEHEAD, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/19/2010 03:23:27

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Subject Sample #1	0.220	03:18
Breath Volume	3.468	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample #2	0.222	03:21
Breath Volume	3.632	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOSTER, E
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: TORCASIO

DATE/TIME OF TEST: 06/19/2010 04:01:53

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Subject Sample #1		04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Control Test	0.081	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: ANGLIN

DATE/TIME OF TEST: 06/19/2010 04:32:53

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.082	04:27
Air Blank	0.000	04:27
Subject Sample #1	0.142	04:28
Breath Volume	2.082	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample #2	0.149	04:31
Breath Volume	1.910	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 06/19/2010 04:41:31

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample #1	0.047	04:37
Breath Volume	1.996	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample #2	0.047	04:39
Breath Volume	2.949	04:39
Air Blank	0.000	04:40
Control Test	0.082	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER, MELONIE, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: KEANE

DATE/TIME OF TEST: 06/19/2010 04:56:09

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Subject Sample #1	0.153	04:51
Breath Volume	1.488	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample #2	0.153	04:54
Breath Volume	1.210	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS, ANGELA, V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 06/19/2010 07:47:40

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.081	07:42
Air Blank	0.000	07:42
Subject Sample #1	0.194	07:43
Breath Volume	1.914	07:43
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample #2	0.184	07:46
Breath Volume	1.925	07:46
Air Blank	0.000	07:46
Control Test	0.080	07:47
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: SALAS-ALCOCER

DATE/TIME OF TEST: 06/19/2010 21:21:38

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:08	
Control Test	0.081	21:09	
Air Blank	0.000	21:09	
Subject Sample #1	0.255	21:10	
Breath Volume	1.238	21:10	
Air Blank	0.000	21:11	
Air Blank	0.000	21:13	
Subject Sample #2	0.231	21:16	Volume Not Met
Breath Volume	0.875	21:16	
Air Blank	0.000	21:17	
Air Blank	0.000	21:18	
Subject Sample #3	0.250	21:19	
Breath Volume	1.121	21:19	
Air Blank	0.000	21:20	
Control Test	0.079	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/19/2010 23:11:00

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Subject Sample #1	0.141	23:06
Breath Volume	2.972	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample #2	0.146	23:09
Breath Volume	3.250	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:11

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEAD, K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ROSADO RAMIREZ

DATE/TIME OF TEST: 06/20/2010 00:42:36

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:35
Subject Sample #1	0.174	00:38
Breath Volume	1.269	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample #2	0.170	00:41
Breath Volume	1.265	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MELO, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/20/2010 03:02:49

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample #1	0.098	02:58
Breath Volume	2.035	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample #2	0.098	03:01
Breath Volume	1.957	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARSON, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: TOTH

DATE/TIME OF TEST: 06/20/2010 03:15:40

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Subject Sample #1	0.000	03:11
Breath Volume	3.488	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample #2	0.000	03:14
Breath Volume	3.261	03:14
Air Blank	0.000	03:14
Control Test	0.081	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BODAMER, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/20/2010 04:03:47

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample #1	0.143	03:59
Breath Volume	1.585	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample #2	0.145	04:02
Breath Volume	1.546	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: MARCH

DATE/TIME OF TEST: 06/20/2010 04:34:09

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:29
Air Blank	0.000	04:29
Subject Sample #1	0.088	04:29
Breath Volume	2.363	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample #2	0.090	04:32
Breath Volume	2.324	04:32
Air Blank	0.000	04:33
Control Test	0.082	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/20/2010 06:12:19

Diagnostics Check	OK	06:04
Air Blank	0.000	06:04
Control Test	0.081	06:05
Air Blank	0.000	06:05
Subject Sample #1	0.164	06:07
Breath Volume	1.835	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:09
Subject Sample #2	0.168	06:10
Breath Volume	1.523	06:10
Air Blank	0.000	06:11
Control Test	0.080	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS, ANGELA, V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCMURTRY, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: WASDEN

DATE/TIME OF TEST: 06/20/2010 17:13:15

Diagnostics Check	OK	17:07
Air Blank	0.000	17:07
Control Test	0.081	17:07
Air Blank	0.000	17:08
Subject Sample #1	0.000	17:08
Breath Volume	1.789	17:08
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample #2	0.000	17:11
Breath Volume	2.000	17:11
Air Blank	0.000	17:12
Control Test	0.081	17:12
Air Blank	0.000	17:13
Diagnostics Check	OK	17:13

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER, B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 06/20/2010 21:52:19

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.082	21:46
Air Blank	0.000	21:47
Subject Sample #1	0.353	21:47
Breath Volume	1.792	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample #2	0.356	21:50
Breath Volume	1.503	21:50
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARDWELL, L
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GIVENS

DATE/TIME OF TEST: 06/20/2010 23:18:18

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:13
Subject Sample #1	0.125	23:13
Breath Volume	2.171	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample #2	0.128	23:16
Breath Volume	1.738	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: CRILLY

DATE/TIME OF TEST: 06/20/2010 23:31:21

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample #1	0.000	23:26
Breath Volume	2.914	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample #2	0.000	23:29
Breath Volume	3.085	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BODAMER, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: MARDER

DATE/TIME OF TEST: 06/21/2010 01:57:10

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample #1	0.141	01:52
Breath Volume	2.656	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample #2	0.132	01:55
Breath Volume	1.738	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: LEONARDS

DATE/TIME OF TEST: 06/21/2010 06:18:31

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.081	06:11
Air Blank	0.000	06:12
Subject Sample #1	0.174	06:13
Breath Volume	2.691	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample #2	0.169	06:16
Breath Volume	3.066	06:16
Air Blank	0.000	06:17
Control Test	0.080	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE, RICHARD, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BOUFFARD, C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: GALINDEZ

DATE/TIME OF TEST: 06/21/2010 07:23:55

Diagnostics Check	OK	07:17
Air Blank	0.000	07:17
Control Test	0.081	07:18
Air Blank	0.000	07:18
Subject Sample #1	0.202	07:19
Breath Volume	2.121	07:19
Air Blank	0.000	07:20
Air Blank	0.000	07:21
Subject Sample #2	0.199	07:22
Breath Volume	1.972	07:22
Air Blank	0.000	07:23
Control Test	0.081	07:23
Air Blank	0.000	07:23
Diagnostics Check	OK	07:23

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE, RICHARD, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ELSEDOUDI, A
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: SILVA-FLORES

DATE/TIME OF TEST: 06/21/2010 15:18:58

Diagnostics Check	OK	15:11
Air Blank	0.000	15:11
Control Test	0.082	15:12
Air Blank	0.000	15:12
Subject Sample #1	0.275	15:13
Breath Volume	1.570	15:13
Air Blank	0.000	15:14
Air Blank	0.000	15:15
Subject Sample #2	0.269	15:17
Breath Volume	2.507	15:17
Air Blank	0.000	15:18
Control Test	0.080	15:18
Air Blank	0.000	15:18
Diagnostics Check	OK	15:18

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JONES, O
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: GRANGER

DATE/TIME OF TEST: 06/21/2010 23:00:28

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample #1	0.088	22:55
Breath Volume	1.808	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample #2	0.089	22:58
Breath Volume	1.554	22:58
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHANAN, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 06/22/2010 02:33:50

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample #1	0.165	02:29
Breath Volume	1.878	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample #2	0.162	02:32
Breath Volume	2.632	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MOREHOUSE

DATE/TIME OF TEST: 06/22/2010 02:54:21

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:48
Subject Sample #1	0.070	02:49
Breath Volume	2.910	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample #2	0.072	02:52
Breath Volume	2.921	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: GUGGINO

DATE/TIME OF TEST: 06/22/2010 04:11:30

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.082	04:05
Air Blank	0.000	04:06
Subject Sample #1	0.207	04:06
Breath Volume	1.449	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample #2	0.217	04:09
Breath Volume	2.472	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MELO, JR
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/22/2010 04:50:06

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Subject Sample #1	0.153	04:45
Breath Volume	2.003	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample #2	0.154	04:48
Breath Volume	1.425	04:48
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MELO, JR
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ZAGAMI

DATE/TIME OF TEST: 06/22/2010 23:06:29

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.081	23:00
Air Blank	0.000	23:00
Subject Sample #1	0.000	23:02
Breath Volume	1.332	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample #2	0.000	23:05
Breath Volume	1.488	23:05
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VALLEJO, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: TORRERO

DATE/TIME OF TEST: 06/23/2010 02:31:12

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample #1	0.172	02:26
Breath Volume	1.972	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample #2	0.166	02:29
Breath Volume	1.398	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER, MELONIE, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BODAMER, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KOONTZ

DATE/TIME OF TEST: 06/23/2010 04:42:08

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.080	04:40	
Air Blank	0.000	04:41	
Subject Sample #1		04:41	Sequence Aborted
Breath Volume	0.000	04:41	
Air Blank	0.000	04:42	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBOSA, F
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 06/23/2010 04:57:37

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:51
Subject Sample #1	0.224	04:52
Breath Volume	1.625	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample #2	0.224	04:55
Breath Volume	1.675	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HYDER, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: GIGANTE

DATE/TIME OF TEST: 06/23/2010 05:51:28

Diagnostics Check	OK	05:45
Air Blank	0.000	05:45
Control Test	0.081	05:45
Air Blank	0.000	05:46
Subject Sample #1	0.190	05:46
Breath Volume	1.140	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample #2	0.202	05:49
Breath Volume	2.105	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LAROSE, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: WALKE

DATE/TIME OF TEST: 06/24/2010 00:36:37

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Subject Sample #1	0.137	00:31
Breath Volume	1.617	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample #2	0.134	00:35
Breath Volume	1.171	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HARGROVE

DATE/TIME OF TEST: 06/24/2010 04:15:57

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Subject Sample #1	0.034	04:11
Breath Volume	3.125	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample #2	0.032	04:14
Breath Volume	3.167	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVARES, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 06/24/2010 04:34:30

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Subject Sample #1	0.105	04:30
Breath Volume	1.960	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample #2	0.106	04:32
Breath Volume	1.347	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCCALLA, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: HUBBELL

DATE/TIME OF TEST: 06/24/2010 05:06:07

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.080	05:03	
Air Blank	0.000	05:04	
Subject Sample #1		05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:05	
Control Test	0.081	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time 06/24/2010 13:34

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: WYNN

DATE/TIME OF TEST: 06/25/2010 00:48:35

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:43
Air Blank	0.000	00:43
Subject Sample #1	0.172	00:44
Breath Volume	1.277	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample #2	0.170	00:47
Breath Volume	1.593	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GLOVER, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: WARGO

DATE/TIME OF TEST: 06/25/2010 01:23:21

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.082	01:17
Air Blank	0.000	01:17
Subject Sample #1	0.246	01:18
Breath Volume	2.074	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample #2	0.244	01:21
Breath Volume	1.371	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CHILS

DATE/TIME OF TEST: 06/25/2010 01:57:07

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Subject Sample #1	0.164	01:52
Breath Volume	1.421	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample #2	0.168	01:55
Breath Volume	1.296	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARSON, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: VERONEAU

DATE/TIME OF TEST: 06/25/2010 02:17:03

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Subject Sample #1	0.000	02:12
Breath Volume	3.921	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample #2	0.000	02:15
Breath Volume	2.339	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:17

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FAUSKEE, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 06/25/2010 02:25:04

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:19
Subject Sample #1	0.181	02:20
Breath Volume	2.312	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample #2	0.178	02:23
Breath Volume	1.800	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:25

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCMURPHY, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: RODRIGUEZ-MARTINEZ

DATE/TIME OF TEST: 06/25/2010 04:04:58

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample #1	0.169	04:00
Breath Volume	1.570	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample #2	0.163	04:03
Breath Volume	1.835	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 06/25/2010 04:13:43

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.081	04:08	
Air Blank	0.000	04:09	
Subject Sample #1		04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.081	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCCALLA, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/25/2010 04:27:56

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample #1	0.158	04:22
Breath Volume	1.335	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample #2	0.171	04:26
Breath Volume	1.589	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: KIRIISA-ALIBAKI

DATE/TIME OF TEST: 06/25/2010 04:41:06

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample #1	0.115	04:36
Breath Volume	1.656	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample #2	0.118	04:39
Breath Volume	1.667	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FAUSKEE, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 06/25/2010 04:49:58

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Subject Sample #1	0.221	04:45
Breath Volume	1.429	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample #2	0.224	04:48
Breath Volume	2.843	04:48
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARBOSA, F
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/25/2010 05:25:23

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:17
Subject Sample #1	0.274	05:19
Breath Volume	1.496	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample #2	0.274	05:23
Breath Volume	1.746	05:23
Air Blank	0.000	05:24
Control Test	0.079	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: BATCHELOR

DATE/TIME OF TEST: 06/25/2010 20:57:46

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.081	20:51
Air Blank	0.000	20:51
Subject Sample #1	0.211	20:52
Breath Volume	1.335	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample #2	0.218	20:56
Breath Volume	1.699	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GARCIA-ORTIZ

DATE/TIME OF TEST: 06/25/2010 22:52:14

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:44
Subject Sample #1	0.094	22:47
Breath Volume	1.585	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample #2	0.088	22:50
Breath Volume	1.734	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 06/26/2010 00:03:16

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.081	23:57
Air Blank	0.000	23:58
Subject Sample #1	0.132	23:58
Breath Volume	2.121	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample #2	0.129	00:01
Breath Volume	1.980	00:01
Air Blank	0.000	00:02
Control Test	0.081	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 06/26/2010 01:30:32

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.082	01:23
Air Blank	0.000	01:23
Subject Sample #1	0.181	01:24
Breath Volume	1.519	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample #2	0.182	01:28
Breath Volume	1.738	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BEDORE

DATE/TIME OF TEST: 06/26/2010 01:39:34

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Subject Sample #1	0.209	01:35
Breath Volume	3.203	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample #2	0.207	01:37
Breath Volume	2.312	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUTHERS, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/26/2010 03:40:40

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Subject Sample #1	0.135	03:36
Breath Volume	2.031	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample #2	0.139	03:39
Breath Volume	1.355	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: VALALAS

DATE/TIME OF TEST: 06/26/2010 04:22:32

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.081	04:16
Air Blank	0.000	04:16
Subject Sample #1	0.166	04:17
Breath Volume	1.488	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample #2	0.177	04:20
Breath Volume	2.585	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINTON, JENNA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCCALLA, C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 06/26/2010 05:22:56

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:15
Air Blank	0.000	05:15
Subject Sample #1	0.092	05:18
Breath Volume	2.492	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample #2	0.094	05:21
Breath Volume	1.937	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALES, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/26/2010 19:34:20

Diagnostics Check	OK	19:22	
Air Blank	0.000	19:22	
Control Test	0.082	19:23	
Air Blank	0.000	19:23	
Subject Sample #1	0.212	19:26	Volume Not Met
Breath Volume	0.902	19:26	
Air Blank	0.000	19:27	
Air Blank	0.000	19:29	
Subject Sample #2	0.222	19:29	
Breath Volume	3.089	19:29	
Air Blank	0.000	19:30	
Air Blank	0.000	19:32	
Subject Sample #3	0.220	19:32	
Breath Volume	3.343	19:32	
Air Blank	0.000	19:33	
Control Test	0.078	19:33	
Air Blank	0.000	19:34	
Diagnostics Check	OK	19:34	

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MCGOWAN, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RIVERON

DATE/TIME OF TEST: 06/26/2010 22:31:47

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.081	22:26
Air Blank	0.000	22:26
Subject Sample #1	0.028	22:27
Breath Volume	1.304	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample #2	0.028	22:30
Breath Volume	2.144	22:30
Air Blank	0.000	22:30
Control Test	0.080	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AMAR, G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: PENYA

DATE/TIME OF TEST: 06/26/2010 23:07:52

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:02
Subject Sample #1	0.180	23:03
Breath Volume	1.773	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample #2	0.185	23:06
Breath Volume	1.113	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LAROSE, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: CARPINTIER

DATE/TIME OF TEST: 06/27/2010 00:17:36

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Subject Sample #1	0.081	00:12
Breath Volume	1.230	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample #2	0.081	00:16
Breath Volume	1.238	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: SEPULVEDA

DATE/TIME OF TEST: 06/27/2010 02:44:52

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:34	
Subject Sample #1	0.165	02:37	Volume Not Met
Breath Volume	0.332	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:39	
Subject Sample #2	0.200	02:40	
Breath Volume	2.027	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample #3	0.194	02:43	
Breath Volume	2.394	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 06/27/2010 04:07:41

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample #1	0.066	04:03
Breath Volume	2.199	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample #2	0.068	04:06
Breath Volume	2.585	04:06
Air Blank	0.000	04:06
Control Test	0.081	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVERES, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: COMACHO-FERNANDEZ

DATE/TIME OF TEST: 06/27/2010 04:31:29

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:25
Subject Sample #1	0.217	04:26
Breath Volume	2.554	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample #2	0.214	04:29
Breath Volume	1.992	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MATAS, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/27/2010 05:12:31

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample #1	0.087	05:06
Breath Volume	1.742	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample #2	0.084	05:10
Breath Volume	1.644	05:10
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/27/2010 23:09:40

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:55	
Control Test	0.081	22:55	
Air Blank	0.000	22:56	
Subject Sample #1	0.087	22:59	Volume Not Met
Breath Volume	0.675	22:59	
Air Blank	0.000	23:00	
Air Blank	0.000	23:02	
Subject Sample #2	0.098	23:03	
Breath Volume	1.136	23:03	
Air Blank	0.000	23:04	
Air Blank	0.000	23:06	
Subject Sample #3	0.098	23:08	
Breath Volume	1.535	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:09	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHANAN, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 06/28/2010 02:18:30

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Subject Sample #1	0.177	02:14
Breath Volume	2.117	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample #2	0.180	02:16
Breath Volume	2.062	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHANAN, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/29/2010 00:22:30

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:11	
Control Test	0.081	00:11	
Air Blank	0.000	00:12	
Subject Sample #1	0.086	00:15	Volume Not Met
Breath Volume	0.023	00:15	
Air Blank	0.000	00:15	
Air Blank	0.000	00:17	
Subject Sample #2	0.092	00:20	Volume Not Met
Breath Volume	0.210	00:20	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHANAN, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/29/2010 00:32:53

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:24	
Control Test	0.080	00:24	
Air Blank	0.000	00:25	
Subject Sample #1	0.077	00:28	Volume Not Met
Breath Volume	0.085	00:28	
Air Blank	0.000	00:28	
Air Blank	0.000	00:30	
Subject Sample #2		00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BODAMER, E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/29/2010 02:55:00

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample #1	0.075	02:50
Breath Volume	3.000	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample #2	0.074	02:53
Breath Volume	3.859	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE, RICHARD, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MULDROW, D
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 06/29/2010 17:36:56

Diagnostics Check	OK	17:29
Air Blank	0.000	17:30
Control Test	0.081	17:30
Air Blank	0.000	17:31
Subject Sample #1	0.042	17:32
Breath Volume	1.535	17:32
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample #2	0.041	17:35
Breath Volume	1.777	17:35
Air Blank	0.000	17:36
Control Test	0.080	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, CHARNELLE, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KACZMAREK, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 06/30/2010 00:55:46

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Subject Sample #1	0.145	00:51
Breath Volume	1.957	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:53
Subject Sample #2	0.149	00:54
Breath Volume	1.425	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, TALIA, R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUCHER, P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: OUNG

DATE/TIME OF TEST: 06/30/2010 02:37:27

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:32
Subject Sample #1	0.101	02:32
Breath Volume	1.828	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample #2	0.088	02:35
Breath Volume	1.359	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POGORZALA, KRISTEN, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BORDEN, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/24/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646831
DRY GAS EXPIRY DATE: 07/30/2011
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SAILS

DATE/TIME OF TEST: 06/30/2010 04:02:10

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.082	03:55
Air Blank	0.000	03:55
Subject Sample #1	0.131	03:56
Breath Volume	1.750	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample #2	0.140	04:00
Breath Volume	1.574	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time 07/02/2010 11:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, STEWART, G

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: HINTON, DAN

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 05/06/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: \643837

DRY GAS EXPIRY DATE: 07/17/2011

VIOLATION CODE: DUI

ARREST TIME: 20:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: REPPER

DATE/TIME OF TEST: 06/04/2010 22:15:18

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.078	22:06
Air Blank	0.000	22:07
Subject Sample #1	0.251	22:10
Breath Volume	2.757	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample #2	0.245	22:13
Breath Volume	1.894	22:13
Air Blank	0.000	22:14
Control Test	0.075	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time 06/09/2010 11:14

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE, DAVID, L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ROE, DAVID
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011802
DRY GAS EXPIRY DATE: 04/28/2012
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 06/04/2010 23:29:28

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.077	23:23
Air Blank	0.000	23:24
Subject Sample #1	0.206	23:24
Breath Volume	3.289	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample #2	0.202	23:27
Breath Volume	3.105	23:27
Air Blank	0.000	23:28
Control Test	0.076	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time 06/23/2010 20:50

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NADAL, JOSE, V

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: CONSALVO, ESTEBAN

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 668180

DRY GAS EXPIRY DATE: 09/28/2011

VIOLATION CODE: DUI

ARREST TIME: 22:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 06/11/2010 23:26:07

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.077	23:21
Air Blank	0.000	23:21
Subject Sample #1	0.204	23:21
Breath Volume	1.714	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample #2	0.204	23:24
Breath Volume	2.195	23:24
Air Blank	0.000	23:25
Control Test	0.076	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time 07/29/2010 19:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON, RAYMOND, E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LAKS, ERIC
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/28/2011
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/12/2010 00:36:30

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.077	00:34	
Air Blank	0.000	00:34	
Subject Sample #1		00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.078	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time 07/29/2010 19:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON, RAYMOND, E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MUNSCHY, PATRICK
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 06/11/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/28/2011
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/12/2010 02:08:22

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Subject Sample #1	0.128	02:00
Breath Volume	3.332	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample #2	0.106	02:03
Breath Volume	1.500	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample #3	0.117	02:06
Breath Volume	2.335	02:06
Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time 07/29/2010 19:40

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FHP TROOP B

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALKER, SHELIA,

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: WALKER, SHELIA

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 06/14/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 825601I

DRY GAS EXPIRY DATE: 10/28/2010

VIOLATION CODE: DUI

ARREST TIME: 19:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: MADDUX

DATE/TIME OF TEST: 06/26/2010 19:44:25

Diagnostics Check	OK	19:37	
Air Blank	0.000	19:38	
Control Test	0.077	19:38	
Air Blank	0.000	19:38	
Subject Sample #1	0.075	19:39	
Breath Volume	3.132	19:39	
Air Blank	0.000	19:40	
Air Blank	0.000	19:41	
Subject Sample #2	0.066	19:42	
Breath Volume	2.082	19:42	
Air Blank	0.000	19:42	
Control Test	0.077	19:43	
Air Blank	0.000	19:43	
Diagnostics Check	Fail	19:43	Internal Printer Test
Air Blank	0.000	19:44	

Data Download Date/Time 07/19/2010 16:21

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY CORR
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEEMS, JASON, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TEEMS, JASON
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/19/2010 02:53:21

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.076	02:44	
Air Blank	0.000	02:45	
Subject Sample #1	0.193	02:48	Volume Not Met
Breath Volume	0.644	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:50	
Subject Sample #2		02:51	Subject Test Refused
Breath Volume	0.105	02:51	
Air Blank	0.000	02:52	
Control Test	0.074	02:52	Control Outside Tolerance
Air Blank	0.000	02:53	

Data Download Date/Time 07/29/2010 10:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY CORR
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEEMS, JASON, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TEEMS, JASON
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 06/19/2010 22:33:20

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.077	22:27
Air Blank	0.000	22:28
Subject Sample #1	0.051	22:29
Breath Volume	1.335	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample #2	0.048	22:31
Breath Volume	1.394	22:31
Air Blank	0.000	22:32
Control Test	0.076	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time 07/29/2010 10:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY CORR
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEEMS, JASON, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TEEMS, JASON
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 06/20/2010 01:47:50

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:42
Subject Sample #1	0.127	01:43
Breath Volume	1.949	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample #2	0.124	01:46
Breath Volume	1.664	01:46
Air Blank	0.000	01:46
Control Test	0.077	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time 07/29/2010 10:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY CORR
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEEMS, JASON, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TEEMS, JASON
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: BLACKMAN

DATE/TIME OF TEST: 06/24/2010 20:27:02

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.077	20:19
Air Blank	0.000	20:20
Subject Sample #1	0.230	20:21
Breath Volume	1.195	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample #2	0.226	20:25
Breath Volume	1.289	20:25
Air Blank	0.000	20:26
Control Test	0.077	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:27

Data Download Date/Time 07/29/2010 10:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY CORR
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEEMS, JASON, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TEEMS, JASON
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HOLLIER

DATE/TIME OF TEST: 06/29/2010 03:14:33

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.076	03:08
Air Blank	0.000	03:09
Subject Sample #1	0.208	03:10
Breath Volume	1.371	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample #2	0.204	03:13
Breath Volume	1.929	03:13
Air Blank	0.000	03:13
Control Test	0.076	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time 07/29/2010 10:39

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOFOCO, JOSEPH, B
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOFOCO, JOSEPH
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/25/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 622245
DRY GAS EXPIRY DATE: 05/28/2011
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 06/03/2010 23:22:03

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.079	23:16
Air Blank	0.000	23:16
Subject Sample #1	0.196	23:17
Breath Volume	1.769	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample #2	0.188	23:20
Breath Volume	1.851	23:20
Air Blank	0.000	23:21
Control Test	0.076	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:22

Data Download Date/Time 06/17/2010 10:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOFOCO, JOSEPH, B

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: LOFOCO, JOSEPH

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 622245

DRY GAS EXPIRY DATE: 05/28/2011

VIOLATION CODE: DUI

ARREST TIME: 01:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/26/2010 01:42:20

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:36
Subject Sample #1	0.261	01:37
Breath Volume	1.132	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample #2	0.274	01:40
Breath Volume	1.375	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time 07/29/2010 10:45

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORTES, C,

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: CORTES, C

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 21:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ENCARNACION

DATE/TIME OF TEST: 06/03/2010 22:28:20

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:23
Subject Sample #1	0.224	22:23
Breath Volume	1.441	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample #2	0.241	22:26
Breath Volume	2.554	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time 06/13/2010 07:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WEAVER, SCOTT, A

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: WEAVER, SCOTT

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/02/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 832201I

DRY GAS EXPIRY DATE: 11/22/2010

VIOLATION CODE: DUI

ARREST TIME: 22:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/07/2010 23:20:21

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Subject Sample #1	0.000	23:15
Breath Volume	1.316	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample #2	0.000	23:18
Breath Volume	1.218	23:18
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time 06/13/2010 07:32

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HILLEY, MICHAEL, E

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: HILLEY, MICHAEL

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 02:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MORALES ALVARADO

DATE/TIME OF TEST: 06/26/2010 02:58:00

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.077	02:51
Air Blank	0.000	02:51
Subject Sample #1	0.210	02:53
Breath Volume	2.066	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample #2	0.207	02:56
Breath Volume	1.324	02:56
Air Blank	0.000	02:57
Control Test	0.075	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:58

Data Download Date/Time 07/29/2010 10:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLEY, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HILLEY, MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646826
DRY GAS EXPIRY DATE: 07/27/2011
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 06/26/2010 23:59:05

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.078	23:53
Air Blank	0.000	23:54
Subject Sample #1	0.125	23:54
Breath Volume	3.050	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample #2	0.124	23:57
Breath Volume	2.859	23:57
Air Blank	0.000	23:58
Control Test	0.077	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time 07/29/2010 10:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HILLEY, MICHAEL, E

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: HILLEY, MICHAEL

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 03:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SANCHEZ GUTIERREZ

DATE/TIME OF TEST: 06/27/2010 03:25:10

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample #1	0.223	03:20
Breath Volume	1.500	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample #2	0.229	03:23
Breath Volume	2.203	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time 07/29/2010 10:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HILLEY, MICHAEL, E

BREATH TEST OPERATOR AGENCY: Osceola County SO

ARRESTING OFFICER: HILLEY, MICHAEL

ARRESTING OFFICER AGENCY: Osceola County SO

LAST AGENCY INSPECTION: 06/17/2010

Date or Time Changed: No

DRY GAS STANDARD LOT #: 646826

DRY GAS EXPIRY DATE: 07/27/2011

VIOLATION CODE: DUI

ARREST TIME: 20:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MCCLINTOCK

DATE/TIME OF TEST: 06/27/2010 20:51:20

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.078	20:46
Air Blank	0.000	20:46
Subject Sample #1	0.151	20:46
Breath Volume	2.382	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample #2	0.148	20:49
Breath Volume	2.722	20:49
Air Blank	0.000	20:50
Control Test	0.076	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time 07/29/2010 10:36

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: HURLBURT FIELD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERSTLER, ESTHER, M
BREATH TEST OPERATOR AGENCY: Hurlburt Field PD
ARRESTING OFFICER: GAMBRELL, PAUL
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 05/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 08/21/2011
VIOLATION CODE: OTHER
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: AVOTRI

DATE/TIME OF TEST: 06/29/2010 01:11:52

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample #1	0.065	01:07
Breath Volume	2.445	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample #2	0.059	01:10
Breath Volume	2.164	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time 06/29/2010 16:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: HURLBURT FIELD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERSTLER, ESTHER, M
BREATH TEST OPERATOR AGENCY: Hurlburt Field PD
ARRESTING OFFICER: GAMBRELL, PAUL
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 05/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: OTHER
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 06/29/2010 02:19:20

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Subject Sample #1	0.102	02:14	
Breath Volume	1.828	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:17	
Subject Sample #2		02:17	Interferent Detect
Breath Volume	1.792	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time 06/29/2010 16:46

Florida Department of Law Enforcement Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: HURLBURT FIELD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERSTLER, ESTHER, M
BREATH TEST OPERATOR AGENCY: Hurlburt Field PD
ARRESTING OFFICER: GAMBRELL, PAUL
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 05/28/2010

Date or Time Changed: No
DRY GAS STANDARD LOT #: 668180
DRY GAS EXPIRY DATE: 09/21/2011
VIOLATION CODE: OTHER
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 06/29/2010 02:29:38

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample #1	0.100	02:25
Breath Volume	3.785	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample #2	0.092	02:28
Breath Volume	3.507	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time 06/29/2010 16:46