Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD,LANNY,A
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: HASTERT,SAMUEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: ADMIN
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: KLEAU

DATE/TIME OF TEST: 11/13/2025 15:42:42

Diagnostics Check		OK	15:36
Air Blank		0.000	15:37
Control Test		0.079	15:37
Air Blank		0.000	15:37
Subject Sample	# 1	0.000	15:38
Breath Volume		1.273	15:38
Air Blank		0.000	15:38
Air Blank		0.000	15:40
Subject Sample	# 2	0.000	15:41
Breath Volume		1.570	15:41
Air Blank		0.000	15:41
Control Test		0.080	15:42
Air Blank		0.000	15:42
Diagnostics Check		OK	15:42

Data Download Date/Time: 12/05/2025 16:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: ROJAS FARGAS

DATE/TIME OF TEST: 11/01/2025 06:07:35

Diagnostics Check		OK	06:01
Air Blank		0.000	06:01
Control Test		0.080	06:01
Air Blank		0.000	06:02
Subject Sample	# 1	0.205	06:02
Breath Volume		2.179	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.212	06:06
Breath Volume		1.859	06:06
Air Blank		0.000	06:06
Control Test		0.080	06:06
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 11/05/2025 21:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ST VIL,R
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 11/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/08/2025 23:31:31

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.078	23:25
Air Blank		0.000	23:26
Subject Sample	# 1	0.047	23:26
Breath Volume		1.242	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.043	23:30
Breath Volume		1.199	23:30
Air Blank		0.000	23:30
Control Test		0.079	23:30
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 12/04/2025 03:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MONTESINO

DATE/TIME OF TEST: 11/27/2025 03:14:29

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1	0.149	03:10	
Breath Volume		2.242	03:10	
Air Blank		0.000	03:11	
Air Blank		0.000	03:13	
Subject Sample	# 2	0.146	03:13	
Breath Volume		1.769	03:13	
Air Blank			03:14	RFI Detect
Air Blank		0.000	03:14	

Data Download Date/Time: 12/04/2025 03:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MONTESINO

DATE/TIME OF TEST: 11/27/2025 03:33:10

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.139	03:28
Breath Volume		2.113	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.141	03:31
Breath Volume		1.750	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 12/04/2025 03:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,MICHELLE,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LIVINGSTON,RON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 11/07/2025 21:26:27

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.076	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.148	21:20
Breath Volume		2.035	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.161	21:24
Breath Volume		1.777	21:24
Air Blank		0.000	21:25
Control Test		0.077	21:25
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 11/25/2025 16:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 11/10/2025 09:06:32

Diagnostics Check		OK	09:03	
Air Blank		0.000	09:03	
Control Test		0.077	09:03	
Air Blank		0.000	09:04	
Subject Sample	# 1		09:05	Subject Test Refused
Breath Volume		0.000	09:05	
Air Blank		0.000	09:05	
Control Test		0.077	09:05	
Air Blank		0.000	09:06	
Diagnostics Check		OK	09:06	

Data Download Date/Time: 11/25/2025 16:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MALCOLM,JARRED
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: LEMON

DATE/TIME OF TEST: 11/10/2025 18:12:11

Diagnostics Check		OK	18:05
Air Blank		0.000	18:06
Control Test		0.078	18:06
Air Blank		0.000	18:06
Subject Sample	# 1	0.000	18:07
Breath Volume		2.023	18:07
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 2	0.000	18:10
Breath Volume		2.855	18:10
Air Blank		0.000	18:11
Control Test		0.076	18:11
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 11/25/2025 16:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORTENBERRY,ZACHARY,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORTENBERRY,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: LOPEZ EMETERIO

DATE/TIME OF TEST: 11/11/2025 19:13:17

Diagnostics Check		OK	19:03
Air Blank		0.000	19:03
Control Test		0.078	19:03
Air Blank		0.000	19:04
Subject Sample	# 1	0.234	19:05
Breath Volume		3.164	19:05
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.207	19:08
Breath Volume		1.988	19:08
Air Blank		0.000	19:09
Air Blank		0.000	19:10
Subject Sample	# 3	0.236	19:11
Breath Volume		3.355	19:11
Air Blank		0.000	19:12
Control Test		0.077	19:12
Air Blank		0.000	19:13
Diagnostics Check		OK	19:13

Data Download Date/Time: 11/25/2025 16:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI,MARK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JABLONSKI,MARK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 11/15/2025 04:57:23

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:53	
Control Test		0.077	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 1	0.193	04:56	
Breath Volume		2.554	04:56	
Air Blank			04:56	Purge Fail
Air Blank		0.000	04:57	

Data Download Date/Time: 11/25/2025 16:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI,MARK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JABLONSKI,MARK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 11/15/2025 05:05:50

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.075	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.182	05:01
Breath Volume		1.730	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:03
Subject Sample	# 2	0.180	05:04
Breath Volume		1.921	05:04
Air Blank		0.000	05:04
Control Test		0.075	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 11/25/2025 16:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALIEL,AMANDA,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: AVERY,?
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 11/23/2025 14:26:08

Diagnostics Check		OK	14:23	
Air Blank		0.000	14:23	
Control Test		0.077	14:23	
Air Blank		0.000	14:24	
Subject Sample	# 1		14:24	Subject Test Refused
Breath Volume		0.000	14:24	
Air Blank		0.000	14:25	
Control Test		0.077	14:25	
Air Blank		0.000	14:26	
Diagnostics Check		OK	14:26	

Data Download Date/Time: 11/25/2025 16:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,MARC,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TRUJILLO,MARC
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 11/01/2025 04:10:52

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:04	
Control Test		0.078	04:05	
Air Blank		0.000	04:05	
Subject Sample	# 1	0.167	04:06	
Breath Volume		1.250	04:06	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 2		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.078	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 11/19/2025 21:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,MARC,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TRUJILLO,MARC
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 11/01/2025 04:27:21

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.077	04:14	
Air Blank		0.000	04:15	
Subject Sample	# 1	0.129	04:18	Volume Not Met
Breath Volume		0.769	04:18	
Air Blank		0.000	04:19	
Air Blank		0.000	04:20	
Subject Sample	# 2	0.170	04:21	
Breath Volume		1.308	04:21	
Air Blank		0.000	04:22	
Air Blank		0.000	04:23	
Subject Sample	# 3	0.151	04:25	
Breath Volume		1.851	04:25	
Air Blank		0.000	04:26	
Control Test		0.076	04:26	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 11/19/2025 21:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,MARC,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TRUJILLO,MARC
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: TULLY

DATE/TIME OF TEST: 11/01/2025 19:55:06

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:42	
Control Test		0.080	19:43	
Air Blank		0.000	19:43	
Subject Sample	# 1	0.167	19:44	
Breath Volume		1.242	19:44	
Air Blank		0.000	19:45	
Air Blank		0.000	19:47	
Subject Sample	# 2		19:50	No Sample Provided
Breath Volume		0.000	19:50	
Air Blank		0.000	19:50	
Air Blank		0.000	19:52	
Subject Sample	# 3	0.167	19:53	
Breath Volume		1.460	19:53	
Air Blank		0.000	19:54	
Control Test		0.079	19:54	
Air Blank		0.000	19:55	
Diagnostics Check		OK	19:55	

Data Download Date/Time: 11/19/2025 21:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TEETS,SHAWN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 11/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: SCHMOLTZE

DATE/TIME OF TEST: 11/22/2025 00:15:40

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.101	00:10
Breath Volume		4.730	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.100	00:14
Breath Volume		4.683	00:14
Air Blank		0.000	00:14
Control Test		0.080	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 12/06/2025 23:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERZIG,NATHAN,G
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TENNANT,DAVID
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 11/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BAUCOM

DATE/TIME OF TEST: 11/25/2025 01:16:14

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.080	01:14	
Air Blank		0.000	01:14	
Subject Sample	# 1		01:15	Sequence Aborted
Breath Volume		0.000	01:15	
Air Blank		0.000	01:16	

Data Download Date/Time: 12/06/2025 23:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERZIG,NATHAN,G
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TENNANT,DAVID
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 11/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BAUCOM

DATE/TIME OF TEST: 11/25/2025 01:31:31

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.079	01:20	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:24	No Sample Provided
Breath Volume		0.000	01:24	
Air Blank		0.000	01:25	
Air Blank		0.000	01:27	
Subject Sample	# 2		01:30	No Sample Provided
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Control Test		0.080	01:30	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 12/06/2025 23:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GODFREY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 11/02/2025 01:23:42

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:20	
Control Test		0.081	01:20	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.080	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 11/24/2025 12:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER,MONICA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,J
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 11/08/2025 00:24:59

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.224	00:20
Breath Volume		1.652	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.222	00:23
Breath Volume		2.066	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 11/24/2025 12:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEEL,TOMMY,J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: NEEL,TOMMY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 11/08/2025 07:01:39

Diagnostics Check		OK	06:50	
Air Blank		0.000	06:50	
Control Test		0.081	06:51	
Air Blank		0.000	06:51	
Subject Sample	# 1		06:54	No Sample Provided
Breath Volume		0.000	06:54	
Air Blank		0.000	06:55	
Air Blank		0.000	06:57	
Subject Sample	# 2		07:00	No Sample Provided
Breath Volume		0.000	07:00	
Air Blank		0.000	07:00	
Control Test		0.081	07:01	
Air Blank		0.000	07:01	
Diagnostics Check		OK	07:01	

Data Download Date/Time: 11/24/2025 12:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEEL,TOMMY,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: STANTON,MICHAEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GONZALEZ-ESTRADA

DATE/TIME OF TEST: 11/09/2025 01:01:06

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.242	00:59
Breath Volume		1.609	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEEL,TOMMY,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: STANTON,MICHAEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GONZALEZ-ESTRADA

DATE/TIME OF TEST: 11/09/2025 01:25:11

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.081	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.229	01:20
Breath Volume		1.828	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:22
Subject Sample	# 2	0.227	01:23
Breath Volume		3.015	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,RICHARD,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: DAIGLE,TYLER
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 11/09/2025 21:37:01

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.082	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.233	21:32
Breath Volume		1.105	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.223	21:35
Breath Volume		1.281	21:35
Air Blank		0.000	21:36
Control Test		0.076	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:37

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 11/11/2025 22:18:40

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.082	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.000	22:14
Breath Volume		2.035	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.000	22:17
Breath Volume		2.441	22:17
Air Blank		0.000	22:17
Control Test		0.082	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER,MONICA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALBRITTON,ROY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HAMRICK

DATE/TIME OF TEST: 11/15/2025 02:50:31

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.097	02:37	
Breath Volume		1.175	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2	0.087	02:43	Volume Not Met
Breath Volume		0.148	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.083	02:49	Volume Not Met
Breath Volume		0.035	02:49	
Air Blank		0.000	02:49	
Control Test		0.082	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: YOUNG,TIMYRON
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 11/15/2025 04:36:17

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.081	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.000	04:31
Breath Volume		3.121	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.000	04:34
Breath Volume		3.339	04:34
Air Blank		0.000	04:35
Control Test		0.081	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER,MONICA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER,MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: ARCHIE

DATE/TIME OF TEST: 11/15/2025 05:22:13

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.081	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.158	05:17
Breath Volume		1.390	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:19
Subject Sample	# 2	0.157	05:20
Breath Volume		1.765	05:20
Air Blank		0.000	05:21
Control Test		0.080	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRE,ADAM,J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HERNANDEZ,MELCHOR
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 11/15/2025 18:41:52

Diagnostics Check		OK	18:35
Air Blank		0.000	18:35
Control Test		0.081	18:36
Air Blank		0.000	18:36
Subject Sample	# 1	0.211	18:37
Breath Volume		1.316	18:37
Air Blank		0.000	18:38
Air Blank		0.000	18:39
Subject Sample	# 2	0.229	18:40
Breath Volume		1.343	18:40
Air Blank		0.000	18:40
Control Test		0.079	18:41
Air Blank		0.000	18:41
Diagnostics Check		OK	18:41

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: CARRIGAN

DATE/TIME OF TEST: 11/16/2025 00:39:18

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.081	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.105	00:34
Breath Volume		2.101	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.105	00:37
Breath Volume		2.507	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODFREY,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GODFREY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 11/16/2025 01:27:08

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.259	01:22
Breath Volume		2.699	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.254	01:25
Breath Volume		3.167	01:25
Air Blank		0.000	01:26
Control Test		0.082	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEEL,TOMMY,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: NEMERGUT,JOHN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 11/16/2025 05:00:15

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.081	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.125	04:55
Breath Volume		2.312	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.121	04:58
Breath Volume		1.785	04:58
Air Blank		0.000	04:59
Control Test		0.081	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALBRITTON,ROY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 11/16/2025 23:43:24

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.081	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.000	23:39
Breath Volume		2.449	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.000	23:42
Breath Volume		1.789	23:42
Air Blank		0.000	23:42
Control Test		0.081	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,IAN,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WINFREE,SAIGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 11/20/2025 14:38:25

Diagnostics Check		OK	14:31
Air Blank		0.000	14:31
Control Test		0.080	14:32
Air Blank		0.000	14:32
Subject Sample	# 1	0.265	14:33
Breath Volume		5.144	14:33
Air Blank		0.000	14:34
Air Blank		0.000	14:35
Subject Sample	# 2	0.271	14:36
Breath Volume		4.746	14:36
Air Blank		0.000	14:37
Control Test		0.079	14:37
Air Blank		0.000	14:38
Diagnostics Check		OK	14:38

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FREEMAN,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 11/21/2025 23:21:01

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.076	23:16
Breath Volume		1.683	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.086	23:19
Breath Volume		2.468	23:19
Air Blank		0.000	23:20
Control Test		0.081	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:21

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: YOUNG,TIMYRON
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SHELBY

DATE/TIME OF TEST: 11/22/2025 02:09:45

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.081	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.154	02:04
Breath Volume		1.324	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.163	02:08
Breath Volume		1.378	02:08
Air Blank		0.000	02:08
Control Test		0.081	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FREEMAN,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/22/2025 05:14:04

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.081	05:07
Air Blank		0.000	05:07
Subject Sample	# 1	0.153	05:08
Breath Volume		1.554	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:10
Subject Sample	# 2	0.159	05:12
Breath Volume		2.113	05:12
Air Blank		0.000	05:13
Control Test		0.081	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:14

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GLASS,JONATHON
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RAINWATER

DATE/TIME OF TEST: 11/24/2025 00:03:10

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:51	
Control Test		0.081	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 2	0.333	23:52	
Breath Volume		1.679	23:52	
Air Blank		0.000	23:53	
Air Blank		0.000	23:55	
Subject Sample	# 3		23:58	No Sample Provided
Breath Volume		0.000	23:58	
Air Blank		0.000	23:58	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.325	00:01	
Breath Volume		1.468	00:01	
Air Blank		0.000	00:02	
Control Test		0.082	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 11/24/2025 12:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JOSHUA,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BIGGS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 11/01/2025 22:06:58

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:02
Subject Sample	# 1	0.198	22:02
Breath Volume		2.542	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.192	22:05
Breath Volume		2.648	22:05
Air Blank		0.000	22:06
Control Test		0.079	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 11/25/2025 22:54

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JOSHUA,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BIGGS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 11/12/2025 00:01:57

Diagnostics Check		OK	23:59	
Air Blank		0.000	23:59	
Control Test		0.080	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:00	RFI Detect
Breath Volume		0.000	00:00	
Air Blank		0.000	00:01	
Control Test		0.079	00:01	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 11/25/2025 22:54

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JOSHUA,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BIGGS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 11/12/2025 00:04:59

Diagnostics Check	OK	00:04	
Air Blank		00:04	RFI Detect
Air Blank		00:04	RFI Detect

Data Download Date/Time: 11/25/2025 22:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JOSHUA,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BIGGS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 11/12/2025 00:16:14

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.081	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.223	00:11
Breath Volume		2.113	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:13
Subject Sample	# 2	0.239	00:14
Breath Volume		1.890	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 11/25/2025 22:54

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIFT,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BREND

DATE/TIME OF TEST: 11/18/2025 19:43:53

Diagnostics Check		OK	19:37
Air Blank		0.000	19:38
Control Test		0.080	19:38
Air Blank		0.000	19:38
Subject Sample	# 1	0.000	19:39
Breath Volume		3.027	19:39
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.000	19:42
Breath Volume		3.558	19:42
Air Blank		0.000	19:42
Control Test		0.079	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 11/25/2025 22:54

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BULLOCK,GINO
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CASS

DATE/TIME OF TEST: 11/23/2025 01:04:07

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.078	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:03	
Control Test		0.078	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 11/25/2025 22:54

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 11/23/2025 02:29:01

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.308	02:23
Breath Volume		1.730	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.306	02:27
Breath Volume		2.316	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:29

Data Download Date/Time: 11/25/2025 22:54

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,MICHELLE,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REEDER,GRACEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/26/2025 23:59:56

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.237	23:55
Breath Volume		1.789	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.232	23:58
Breath Volume		2.101	23:58
Air Blank		0.000	23:59
Control Test		0.078	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 12/31/2025 14:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,NATHAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SUMNER,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: MARTINEZ HERNANDEZ

DATE/TIME OF TEST: 11/14/2025 06:56:14

Diagnostics Check		OK	06:50
Air Blank		0.000	06:50
Control Test		0.081	06:50
Air Blank		0.000	06:51
Subject Sample	# 1	0.147	06:51
Breath Volume		2.042	06:51
Air Blank		0.000	06:52
Air Blank		0.000	06:54
Subject Sample	# 2	0.150	06:54
Breath Volume		1.886	06:54
Air Blank		0.000	06:55
Control Test		0.080	06:55
Air Blank		0.000	06:56
Diagnostics Check		OK	06:56

Data Download Date/Time: 11/24/2025 02:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNEY,NOAH,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: O'CONNEY,NOAH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: CRUMITY

DATE/TIME OF TEST: 11/26/2025 11:50:59

Diagnostics Check		OK	11:43
Air Blank		0.000	11:43
Control Test		0.080	11:44
Air Blank		0.000	11:44
Subject Sample	# 1	0.000	11:46
Breath Volume		1.394	11:46
Air Blank		0.000	11:46
Air Blank		0.000	11:48
Subject Sample	# 2	0.000	11:49
Breath Volume		1.179	11:49
Air Blank		0.000	11:50
Control Test		0.079	11:50
Air Blank		0.000	11:50
Diagnostics Check		OK	11:50

Data Download Date/Time: 12/30/2025 01:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WELFORD,DERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CRIBBS

DATE/TIME OF TEST: 11/30/2025 04:32:10

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.082	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.169	04:27
Breath Volume		2.867	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.163	04:30
Breath Volume		2.097	04:30
Air Blank		0.000	04:31
Control Test		0.081	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 12/05/2025 20:33

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEDJESKE,JAMES,R
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CADY,CHARLES
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 11/02/2025 00:58:15

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.120	00:53
Breath Volume		2.093	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.119	00:56
Breath Volume		2.316	00:56
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 11/27/2025 20:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 11/04/2025 22:58:53

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.078	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.126	22:52
Breath Volume		1.476	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:54
Subject Sample	# 2	0.123	22:57
Breath Volume		1.746	22:57
Air Blank		0.000	22:57
Control Test		0.079	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 11/27/2025 20:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 11/07/2025 23:47:47

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.078	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.128	23:43
Breath Volume		3.175	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.127	23:46
Breath Volume		3.324	23:46
Air Blank		0.000	23:46
Control Test		0.076	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 11/27/2025 20:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 11/17/2025 21:14:05

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.080	21:07
Air Blank		0.000	21:08
Subject Sample	# 1	0.184	21:09
Breath Volume		2.675	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.185	21:12
Breath Volume		2.656	21:12
Air Blank		0.000	21:13
Control Test		0.076	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 11/27/2025 20:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEMMONS,L,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: HARDEN,J
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: REVIS

DATE/TIME OF TEST: 11/21/2025 22:21:26

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.079	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.263	22:16
Breath Volume		2.734	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.267	22:19
Breath Volume		1.113	22:19
Air Blank		0.000	22:20
Control Test		0.076	22:20
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 11/27/2025 20:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RITCH

DATE/TIME OF TEST: 11/22/2025 02:40:04

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.086	02:35
Breath Volume		3.500	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.084	02:38
Breath Volume		3.472	02:38
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:40

Data Download Date/Time: 11/27/2025 20:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FISHER,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: OCHOAVEGA

DATE/TIME OF TEST: 11/23/2025 01:13:05

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.146	01:08
Breath Volume		2.039	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.140	01:11
Breath Volume		1.597	01:11
Air Blank		0.000	01:12
Control Test		0.077	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 11/27/2025 20:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: RATHEL

DATE/TIME OF TEST: 11/23/2025 17:23:46

Diagnostics Check		OK	17:17
Air Blank		0.000	17:18
Control Test		0.079	17:18
Air Blank		0.000	17:18
Subject Sample	# 1	0.134	17:19
Breath Volume		1.347	17:19
Air Blank		0.000	17:19
Air Blank		0.000	17:21
Subject Sample	# 2	0.130	17:22
Breath Volume		2.515	17:22
Air Blank		0.000	17:22
Control Test		0.077	17:23
Air Blank		0.000	17:23
Diagnostics Check		OK	17:23

Data Download Date/Time: 11/27/2025 20:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FISHER,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MAPP

DATE/TIME OF TEST: 11/28/2025 01:55:04

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.081	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.076	01:49
Breath Volume		1.144	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.076	01:53
Breath Volume		1.453	01:53
Air Blank		0.000	01:54
Control Test		0.077	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:55

Data Download Date/Time: 01/04/2026 21:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTSHALL,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GUTSHALL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SIPOS

DATE/TIME OF TEST: 11/29/2025 02:46:00

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.080	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1	0.226	02:42	
Breath Volume		2.484	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 2	0.227	02:45	
Breath Volume		2.585	02:45	
Air Blank			02:45	RFI Detect
Air Blank		0.000	02:46	

Data Download Date/Time: 01/04/2026 21:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEDJESKE,JAMES,R
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: COOK,SAMUEL
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 11/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: HILGERSON

DATE/TIME OF TEST: 11/29/2025 21:50:42

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.079	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.174	21:46
Breath Volume		2.589	21:46
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.169	21:49
Breath Volume		2.988	21:49
Air Blank		0.000	21:49
Control Test		0.078	21:50
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 01/04/2026 21:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,STACEY,C
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: THIGPEN,D
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 11/10/2025 01:32:25

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.081	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.215	01:27
Breath Volume		2.203	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.221	01:30
Breath Volume		2.425	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 11/18/2025 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,F,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALLEN,F
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: MATHERS

DATE/TIME OF TEST: 11/01/2025 20:59:55

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.081	20:52
Air Blank		0.000	20:52
Subject Sample	# 1	0.291	20:53
Breath Volume		1.988	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:55
Subject Sample	# 2	0.277	20:58
Breath Volume		1.320	20:58
Air Blank		0.000	20:58
Control Test		0.080	20:59
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 11/07/2025 07:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSBORNE,EVAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 11/29/2025 23:38:10

Diagnostics Check		OK	23:31
Air Blank		0.000	23:32
Control Test		0.081	23:32
Air Blank		0.000	23:32
Subject Sample	# 1	0.151	23:33
Breath Volume		1.265	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.150	23:36
Breath Volume		1.457	23:36
Air Blank		0.000	23:37
Control Test		0.081	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 12/05/2025 21:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSBORNE,EVAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: CASTILLO MEDINA

DATE/TIME OF TEST: 11/30/2025 22:34:16

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.160	22:29
Breath Volume		2.152	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.162	22:32
Breath Volume		3.179	22:32
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 12/05/2025 21:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: TATE,AYDEN
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 11/09/2025 15:44:27

Diagnostics Check		OK	15:31	
Air Blank		0.000	15:32	
Control Test		0.081	15:32	
Air Blank		0.000	15:33	
Subject Sample	# 1		15:36	No Sample Provided
Breath Volume		0.000	15:36	
Air Blank		0.000	15:36	
Air Blank		0.000	15:38	
Subject Sample	# 2	0.111	15:39	
Breath Volume		1.558	15:39	
Air Blank		0.000	15:40	
Air Blank		0.000	15:41	
Subject Sample	# 3	0.105	15:42	
Breath Volume		2.457	15:42	
Air Blank		0.000	15:43	
Control Test		0.079	15:43	
Air Blank		0.000	15:44	
Diagnostics Check		OK	15:44	

Data Download Date/Time: 11/28/2025 19:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ELMORE,TT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: MARBUTT

DATE/TIME OF TEST: 11/20/2025 16:01:01

Diagnostics Check		OK	15:58	
Air Blank		0.000	15:58	
Control Test		0.082	15:58	
Air Blank		0.000	15:59	
Subject Sample	# 1		15:59	RFI Detect
Breath Volume		0.000	15:59	
Air Blank		0.000	16:00	
Control Test		0.082	16:00	
Air Blank		0.000	16:00	
Diagnostics Check		OK	16:01	

Data Download Date/Time: 11/28/2025 19:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ELMORE,TT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: MARBUTT

DATE/TIME OF TEST: 11/20/2025 16:05:43

Diagnostics Check		OK	16:02	
Air Blank		0.000	16:03	
Control Test		0.082	16:03	
Air Blank		0.000	16:04	
Subject Sample	# 1		16:04	RFI Detect
Breath Volume		0.000	16:04	
Air Blank		0.000	16:04	
Control Test		0.082	16:05	
Air Blank		0.000	16:05	
Diagnostics Check		OK	16:05	

Data Download Date/Time: 11/28/2025 19:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ELMORE,TT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: MARBUTT

DATE/TIME OF TEST: 11/20/2025 16:16:40

Diagnostics Check		OK	16:08
Air Blank		0.000	16:08
Control Test		0.081	16:09
Air Blank		0.000	16:09
Subject Sample	# 1	0.180	16:11
Breath Volume		2.425	16:11
Air Blank		0.000	16:12
Air Blank		0.000	16:13
Subject Sample	# 2	0.185	16:14
Breath Volume		3.382	16:14
Air Blank		0.000	16:15
Control Test		0.079	16:16
Air Blank		0.000	16:16
Diagnostics Check		OK	16:16

Data Download Date/Time: 11/28/2025 19:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DURKA BOWHALL

DATE/TIME OF TEST: 11/28/2025 02:56:53

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.204	02:52
Breath Volume		2.250	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.199	02:55
Breath Volume		2.914	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 12/22/2025 08:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KICKLIGHTER,KEVIN,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MARTINEK,KEITH
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 11/02/2025 02:14:23

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.067	02:09
Breath Volume		1.781	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.067	02:12
Breath Volume		1.667	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 11/29/2025 17:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BLACK,EDWARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: AURELIEN

DATE/TIME OF TEST: 11/05/2025 00:26:51

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.081	00:21	
Air Blank		0.000	00:21	
Subject Sample	# 1	0.292	00:22	
Breath Volume		2.468	00:22	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2		00:25	Slope Not Met
Breath Volume		2.734	00:25	
Air Blank		0.000	00:25	
Control Test		0.080	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 11/29/2025 17:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BLACK,EDWARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: AURELIEN

DATE/TIME OF TEST: 11/05/2025 00:36:39

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.081	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.308	00:32
Breath Volume		3.234	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.308	00:35
Breath Volume		3.183	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 11/29/2025 17:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITES,DEJAS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MORENO,JAIRO
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: OLIVA REYES

DATE/TIME OF TEST: 11/07/2025 04:50:18

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.080	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.199	04:45
Breath Volume		1.808	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.208	04:48
Breath Volume		3.757	04:48
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 11/29/2025 17:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 11/07/2025 22:19:45

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.081	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.258	22:14
Breath Volume		1.957	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.256	22:18
Breath Volume		2.839	22:18
Air Blank		0.000	22:18
Control Test		0.079	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 11/29/2025 17:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: HAREN

DATE/TIME OF TEST: 11/08/2025 04:15:11

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.200	04:10
Breath Volume		3.941	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.198	04:13
Breath Volume		4.027	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 11/29/2025 17:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/08/2025 23:55:48

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.080	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.203	23:51
Breath Volume		3.871	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.184	23:54
Breath Volume		1.890	23:54
Air Blank		0.000	23:54
Control Test		0.080	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 11/29/2025 17:28

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GALEAS ZUNIGA

DATE/TIME OF TEST: 11/09/2025 02:06:14

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.081	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.129	02:01
Breath Volume		1.289	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.139	02:04
Breath Volume		3.242	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 11/29/2025 17:28

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: YARISH,DYLAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/09/2025 19:14:16

Diagnostics Check		OK	19:07
Air Blank		0.000	19:07
Control Test		0.080	19:08
Air Blank		0.000	19:08
Subject Sample	# 1	0.160	19:09
Breath Volume		3.292	19:09
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.147	19:12
Breath Volume		3.128	19:12
Air Blank		0.000	19:13
Control Test		0.079	19:13
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 11/29/2025 17:28

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PERKINS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 11/09/2025 21:55:42

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.080	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.167	21:51
Breath Volume		1.378	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:53
Subject Sample	# 2	0.171	21:54
Breath Volume		1.484	21:54
Air Blank		0.000	21:54
Control Test		0.078	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 11/29/2025 17:28

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ETIENNE,EMMANUEL,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ETIENNE,EMMANUEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 11/10/2025 01:47:20

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.081	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.015	01:46	
Air Blank		0.000	01:47	
			01:47	RFI Detect
Air Blank			01:47	RFI Detect

Data Download Date/Time: 11/29/2025 17:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DECHERT,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DORSAINRRE

DATE/TIME OF TEST: 11/10/2025 04:07:59

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.148	04:03
Breath Volume		1.480	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.143	04:06
Breath Volume		1.171	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 11/29/2025 17:28

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 11/12/2025 19:47:34

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.081	19:42
Air Blank		0.000	19:42
Subject Sample	# 1	0.088	19:43
Breath Volume		2.902	19:43
Air Blank		0.000	19:43
Air Blank		0.000	19:45
Subject Sample	# 2	0.082	19:46
Breath Volume		2.765	19:46
Air Blank		0.000	19:46
Control Test		0.081	19:46
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BAILES

DATE/TIME OF TEST: 11/13/2025 02:55:14

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.166	02:50
Breath Volume		3.640	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.162	02:53
Breath Volume		2.035	02:53
Air Blank		0.000	02:54
Control Test		0.081	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 11/13/2025 20:33:17

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.082	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.178	20:28
Breath Volume		2.789	20:28
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.178	20:31
Breath Volume		3.128	20:31
Air Blank		0.000	20:32
Control Test		0.080	20:32
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: N,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SHEFTALL

DATE/TIME OF TEST: 11/14/2025 03:10:23

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.081	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1	0.144	03:03	Volume Not Met
Breath Volume		0.113	03:03	
Air Blank		0.000	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 2		03:08	Subject Test Refused
Breath Volume		0.082	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SHEFTALL

DATE/TIME OF TEST: 11/14/2025 03:21:38

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:13	
Control Test		0.079	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 1	0.145	03:17	Volume Not Met
Breath Volume		0.097	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.079	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 11/18/2025 02:18:11

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.081	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.127	02:13
Breath Volume		2.664	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.122	02:16
Breath Volume		1.984	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/19/2025 02:13:41

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.081	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.081	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/19/2025 02:18:43

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.081	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.081	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DECHERT,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: BALDIE

DATE/TIME OF TEST: 11/22/2025 04:59:41

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.081	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.049	04:55
Breath Volume		2.281	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.046	04:58
Breath Volume		3.542	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: AHRENS,DAVID
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 11/22/2025 23:16:14

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.081	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.134	23:11
Breath Volume		1.464	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.132	23:14
Breath Volume		2.207	23:14
Air Blank		0.000	23:15
Control Test		0.080	23:15
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 11/29/2025 17:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,ANGEL,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ADAMS,THOMAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 11/22/2025 23:30:31

Diagnostics Check		OK	23:24
Air Blank		0.000	23:24
Control Test		0.080	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.101	23:26
Breath Volume		5.882	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.097	23:28
Breath Volume		4.628	23:28
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 11/29/2025 17:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: PIRRO

DATE/TIME OF TEST: 11/22/2025 23:44:23

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.080	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.161	23:39
Breath Volume		2.683	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.162	23:42
Breath Volume		2.570	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 11/29/2025 17:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA,ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SWINGLE

DATE/TIME OF TEST: 11/24/2025 00:48:26

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	RFI Detect
Air Blank	0.000	00:48	

Data Download Date/Time: 11/29/2025 17:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA,ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SWINGLE

DATE/TIME OF TEST: 11/24/2025 00:58:08

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.116	00:53
Breath Volume		1.675	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.117	00:56
Breath Volume		1.257	00:56
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 11/29/2025 17:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SWIERKOWSKI,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DIEZ

DATE/TIME OF TEST: 11/24/2025 01:32:59

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.000	01:28
Breath Volume		1.882	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.000	01:31
Breath Volume		1.125	01:31
Air Blank		0.000	01:32
Control Test		0.081	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 11/29/2025 17:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTENS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BRADY,Z
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 11/25/2025 04:32:12

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.081	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.135	04:27
Breath Volume		3.660	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.124	04:30
Breath Volume		2.207	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 11/29/2025 17:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: BENAVIDEZ

DATE/TIME OF TEST: 11/26/2025 09:01:08

Diagnostics Check		OK	08:55
Air Blank		0.000	08:55
Control Test		0.082	08:56
Air Blank		0.000	08:56
Subject Sample	# 1	0.192	08:56
Breath Volume		1.562	08:56
Air Blank		0.000	08:57
Air Blank		0.000	08:59
Subject Sample	# 2	0.181	08:59
Breath Volume		1.808	08:59
Air Blank		0.000	09:00
Control Test		0.081	09:00
Air Blank		0.000	09:01
Diagnostics Check		OK	09:01

Data Download Date/Time: 11/29/2025 17:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 11/07/2025 20:12:03

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.080	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.143	20:07
Breath Volume		1.460	20:07
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.148	20:10
Breath Volume		1.718	20:10
Air Blank		0.000	20:11
Control Test		0.079	20:11
Air Blank		0.000	20:11
Diagnostics Check		OK	20:12

Data Download Date/Time: 11/08/2025 16:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THEIS,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: THEIS,ZACHARY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: BROOKINS

DATE/TIME OF TEST: 11/08/2025 18:30:09

Diagnostics Check		OK	18:23
Air Blank		0.000	18:24
Control Test		0.080	18:24
Air Blank		0.000	18:24
Subject Sample	# 1	0.255	18:25
Breath Volume		2.355	18:25
Air Blank		0.000	18:26
Air Blank		0.000	18:27
Subject Sample	# 2	0.245	18:28
Breath Volume		2.386	18:28
Air Blank		0.000	18:29
Control Test		0.079	18:29
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 11/19/2025 15:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUILLARD,ALEXANDER,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GUILLARD,ALEXANDER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: HEROUX

DATE/TIME OF TEST: 11/08/2025 20:17:43

Diagnostics Check		OK	20:11
Air Blank		0.000	20:11
Control Test		0.080	20:12
Air Blank		0.000	20:12
Subject Sample	# 1	0.290	20:12
Breath Volume		1.210	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.296	20:16
Breath Volume		1.453	20:16
Air Blank		0.000	20:16
Control Test		0.079	20:17
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 11/19/2025 15:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: BIEBER

DATE/TIME OF TEST: 11/14/2025 19:29:41

Diagnostics Check		OK	19:23
Air Blank		0.000	19:23
Control Test		0.081	19:24
Air Blank		0.000	19:24
Subject Sample	# 1	0.151	19:24
Breath Volume		2.628	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.162	19:28
Breath Volume		1.734	19:28
Air Blank		0.000	19:28
Control Test		0.080	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 11/19/2025 15:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERICO,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PERICO,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PINTARD

DATE/TIME OF TEST: 11/15/2025 02:40:23

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:33	
Control Test		0.081	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1	0.182	02:37	Volume Not Met
Breath Volume		0.136	02:37	
Air Blank		0.000	02:38	
Air Blank			02:39	Ambient Fail
Air Blank		0.000	02:40	

Data Download Date/Time: 11/19/2025 15:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: BENNINGHOFF

DATE/TIME OF TEST: 11/15/2025 05:32:33

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.215	05:27
Breath Volume		3.960	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.207	05:30
Breath Volume		2.902	05:30
Air Blank		0.000	05:31
Control Test		0.079	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 11/19/2025 15:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINNERT,AJAY,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MEINNERT,AJAY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ENGLAND

DATE/TIME OF TEST: 11/15/2025 07:27:03

Diagnostics Check		OK	07:20
Air Blank		0.000	07:21
Control Test		0.081	07:21
Air Blank		0.000	07:22
Subject Sample	# 1	0.007	07:22
Breath Volume		5.527	07:22
Air Blank		0.000	07:23
Air Blank		0.000	07:25
Subject Sample	# 2	0.000	07:25
Breath Volume		6.136	07:25
Air Blank		0.000	07:26
Control Test		0.079	07:26
Air Blank		0.000	07:26
Diagnostics Check		OK	07:27

Data Download Date/Time: 11/19/2025 15:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/16/2025 01:55:56

Diagnostics Check	OK	01:55	
Air Blank		01:55	RFI Detect
Air Blank	0.000	01:55	

Data Download Date/Time: 11/19/2025 15:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/16/2025 01:59:46

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:59	
Air Blank		01:59	RFI Detect
Air Blank		01:59	RFI Detect

Data Download Date/Time: 11/19/2025 15:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/16/2025 02:09:51

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.081	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.217	02:03
Breath Volume		1.574	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.202	02:08
Breath Volume		1.761	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 11/19/2025 15:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA',JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/18/2025 22:20:27

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.080	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.238	22:15
Breath Volume		2.542	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.228	22:18
Breath Volume		1.742	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 11/19/2025 15:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JULIO,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SCOTT,C.
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 12:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 11/19/2025 13:15:31

Diagnostics Check		OK	13:09
Air Blank		0.000	13:09
Control Test		0.080	13:09
Air Blank		0.000	13:10
Subject Sample	# 1	0.249	13:10
Breath Volume		1.992	13:10
Air Blank		0.000	13:11
Air Blank		0.000	13:13
Subject Sample	# 2	0.252	13:13
Breath Volume		2.792	13:13
Air Blank		0.000	13:14
Control Test		0.079	13:14
Air Blank		0.000	13:15
Diagnostics Check		OK	13:15

Data Download Date/Time: 11/19/2025 15:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JAYCOB,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: AMSLER,K
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RICO CANTADOR

DATE/TIME OF TEST: 11/02/2025 00:23:44

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:21	
Control Test		0.079	00:21	
Air Blank		0.000	00:22	
Subject Sample	# 1		00:22	Subject Test Refused
Breath Volume		0.000	00:22	
Air Blank		0.000	00:22	
Control Test		0.079	00:23	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 11/29/2025 07:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ROUSE

DATE/TIME OF TEST: 11/07/2025 22:49:02

Diagnostics Check		OK	22:42
Air Blank		0.000	22:43
Control Test		0.080	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.253	22:44
Breath Volume		1.925	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:46
Subject Sample	# 2	0.254	22:47
Breath Volume		2.007	22:47
Air Blank		0.000	22:48
Control Test		0.079	22:48
Air Blank		0.000	22:48
Diagnostics Check		OK	22:49

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSELL,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOGGS,ZACHARY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: RUFFIN

DATE/TIME OF TEST: 11/08/2025 00:03:05

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.079	00:00	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.079	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSELL,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: LATDURETTE

DATE/TIME OF TEST: 11/08/2025 22:20:00

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.080	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.166	22:13
Breath Volume		1.554	22:13
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.173	22:18
Breath Volume		2.148	22:18
Air Blank		0.000	22:19
Control Test		0.079	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:20

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JAYCOB,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WADDELL,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 11/15/2025 04:12:05

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.158	04:07
Breath Volume		2.480	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.161	04:10
Breath Volume		2.273	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MARTINIS,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PRUDENTE

DATE/TIME OF TEST: 11/16/2025 03:38:36

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.080	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.199	03:33
Breath Volume		3.089	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.192	03:36
Breath Volume		3.167	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JAYCOB,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MARTINIS,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MC DONOUGH

DATE/TIME OF TEST: 11/16/2025 22:03:06

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.079	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.180	21:57
Breath Volume		1.738	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.171	22:01
Breath Volume		1.660	22:01
Air Blank		0.000	22:02
Control Test		0.080	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JAYCOB,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: JEFFERSON,K
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 11/16/2025 23:02:43

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.177	22:58
Breath Volume		1.414	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.186	23:01
Breath Volume		2.253	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DAVENPORT,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/17/2025 03:11:55

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.131	03:03
Breath Volume		1.125	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.152	03:06
Breath Volume		2.902	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 3	0.151	03:10
Breath Volume		2.949	03:10
Air Blank		0.000	03:10
Control Test		0.078	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSELL,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SURIEL,ENMANUEL
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: JULSON DELP

DATE/TIME OF TEST: 11/18/2025 18:57:08

Diagnostics Check		OK	18:54	
Air Blank		0.000	18:54	
Control Test		0.079	18:54	
Air Blank		0.000	18:55	
Subject Sample	# 1		18:55	Subject Test Refused
Breath Volume		0.000	18:55	
Air Blank		0.000	18:56	
Control Test		0.080	18:56	
Air Blank		0.000	18:57	
Diagnostics Check		OK	18:57	

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JAYCOB,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HRADECKY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: GONYO

DATE/TIME OF TEST: 11/20/2025 19:06:13

Diagnostics Check		OK	19:03	
Air Blank		0.000	19:03	
Control Test		0.080	19:04	
Air Blank		0.000	19:04	
Subject Sample	# 1		19:04	Subject Test Refused
Breath Volume		0.000	19:04	
Air Blank		0.000	19:05	
Control Test		0.080	19:05	
Air Blank		0.000	19:06	
Diagnostics Check		OK	19:06	

Data Download Date/Time: 11/29/2025 07:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: FALLICA

DATE/TIME OF TEST: 11/01/2025 22:45:35

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.078	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.000	22:41
Breath Volume		3.527	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.000	22:44
Breath Volume		3.160	22:44
Air Blank		0.000	22:44
Control Test		0.078	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 11/30/2025 05:49

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PEREDA

DATE/TIME OF TEST: 11/05/2025 22:56:23

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.078	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.157	22:51
Breath Volume		2.207	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.150	22:54
Breath Volume		2.269	22:54
Air Blank		0.000	22:55
Control Test		0.078	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 11/30/2025 05:49

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,FRANCISCO,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: FERRER,FRANCISCO
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 11/08/2025 15:56:30

Diagnostics Check	OK	15:55	
Air Blank	0.000	15:56	
		15:56	RFI Detect
Air Blank		15:56	RFI Detect

Data Download Date/Time: 11/30/2025 05:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,FRANCISCO,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: FERRER,FRANCISCO
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: V420813656420

DATE/TIME OF TEST: 11/08/2025 16:06:12

Diagnostics Check		OK	15:59
Air Blank		0.000	15:59
Control Test		0.078	16:00
Air Blank		0.000	16:00
Subject Sample	# 1	0.309	16:01
Breath Volume		1.683	16:01
Air Blank		0.000	16:02
Air Blank		0.000	16:04
Subject Sample	# 2	0.301	16:04
Breath Volume		1.890	16:04
Air Blank		0.000	16:05
Control Test		0.077	16:05
Air Blank		0.000	16:06
Diagnostics Check		OK	16:06

Data Download Date/Time: 11/30/2025 05:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: PETT

DATE/TIME OF TEST: 11/08/2025 20:25:47

Diagnostics Check		OK	20:19
Air Blank		0.000	20:19
Control Test		0.079	20:20
Air Blank		0.000	20:20
Subject Sample	# 1	0.129	20:21
Breath Volume		1.152	20:21
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.126	20:24
Breath Volume		1.281	20:24
Air Blank		0.000	20:24
Control Test		0.078	20:25
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 11/30/2025 05:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLMAN JR,DAVID,E
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 11/09/2025 02:13:37

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:10	
Control Test		0.078	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.079	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 11/30/2025 05:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LEVINSON

DATE/TIME OF TEST: 11/15/2025 04:13:56

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.078	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.061	04:09
Breath Volume		1.648	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.059	04:12
Breath Volume		1.320	04:12
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 11/30/2025 05:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,R
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: TROUT

DATE/TIME OF TEST: 11/16/2025 23:38:59

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.079	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.158	23:34
Breath Volume		1.605	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.156	23:37
Breath Volume		1.308	23:37
Air Blank		0.000	23:38
Control Test		0.077	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 11/30/2025 05:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,R
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 11/22/2025 23:43:24

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.079	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.152	23:37	
Breath Volume		1.320	23:37	
Air Blank		0.000	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 2		23:42	Subject Test Refused
Breath Volume		0.812	23:42	
Air Blank		0.000	23:42	
Control Test		0.008	23:42	RFI Detect
Air Blank		0.000	23:43	

Data Download Date/Time: 11/30/2025 05:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,R
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250713
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/26/2025 23:55:39

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.078	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.134	23:51
Breath Volume		1.179	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.133	23:54
Breath Volume		1.566	23:54
Air Blank		0.000	23:54
Control Test		0.078	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 11/30/2025 05:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSHONG,N,R
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: REYNOLDS,G
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 11/01/2025 04:37:20

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.082	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.076	04:33
Breath Volume		1.878	04:33
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.076	04:35
Breath Volume		1.585	04:35
Air Blank		0.000	04:36
Control Test		0.082	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ALMODOVAR

DATE/TIME OF TEST: 11/02/2025 03:37:38

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.082	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.181	03:33
Breath Volume		1.261	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.189	03:36
Breath Volume		1.542	03:36
Air Blank		0.000	03:36
Control Test		0.081	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEMBY,N,
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: ALEXANDER,A
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: TAMBURRO

DATE/TIME OF TEST: 11/02/2025 13:35:30

Diagnostics Check		OK	13:31	
Air Blank		0.000	13:31	
Control Test		0.082	13:31	
Air Blank		0.000	13:32	
Subject Sample	# 1		13:34	Sequence Aborted
Breath Volume		0.148	13:34	
Air Blank		0.000	13:35	

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEMBY,N,
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: ALEXANDER,A
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: TAMBURRO

DATE/TIME OF TEST: 11/02/2025 14:03:35

Diagnostics Check		OK	13:56
Air Blank		0.000	13:57
Control Test		0.081	13:57
Air Blank		0.000	13:58
Subject Sample	# 1	0.213	13:58
Breath Volume		2.683	13:58
Air Blank		0.000	13:59
Air Blank		0.000	14:01
Subject Sample	# 2	0.221	14:01
Breath Volume		2.523	14:01
Air Blank		0.000	14:02
Control Test		0.081	14:02
Air Blank		0.000	14:03
Diagnostics Check		OK	14:03

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,KEVIN,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DOUGLAS,ASHLEY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: TIBBITTS

DATE/TIME OF TEST: 11/07/2025 03:04:46

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:02	
Control Test		0.081	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:03	Subject Test Refused
Breath Volume		0.000	03:03	
Air Blank		0.000	03:03	
Control Test		0.081	03:04	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARGE,MIRANDA,H
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 11/07/2025 21:39:11

Diagnostics Check		OK	21:36	
Air Blank		0.000	21:36	
Control Test		0.082	21:37	
Air Blank		0.000	21:37	
Subject Sample	# 1		21:37	Subject Test Refused
Breath Volume		0.000	21:37	
Air Blank		0.000	21:38	
Control Test		0.082	21:38	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MAZNARITZ,KAYLA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 11/08/2025 03:58:00

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.082	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.181	03:53
Breath Volume		2.144	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.180	03:56
Breath Volume		1.562	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:58

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CABEZO,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ZOUGUI

DATE/TIME OF TEST: 11/08/2025 04:33:26

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.082	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.154	04:29
Breath Volume		2.398	04:29
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.153	04:31
Breath Volume		1.636	04:31
Air Blank		0.000	04:32
Control Test		0.081	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: BURNS,JEFFREY
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CHAVEZ ZAMBRANO

DATE/TIME OF TEST: 11/08/2025 22:37:46

Diagnostics Check		OK	22:31
Air Blank		0.000	22:31
Control Test		0.082	22:32
Air Blank		0.000	22:32
Subject Sample	# 1	0.200	22:33
Breath Volume		1.324	22:33
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.201	22:36
Breath Volume		2.046	22:36
Air Blank		0.000	22:36
Control Test		0.080	22:37
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LOWE,MAYCEE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/09/2025 01:44:23

Diagnostics Check	OK	01:43	
Air Blank		01:43	RFI Detect
Air Blank	0.000	01:44	

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTRIDGE,JAMES,T
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BARKER,D
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: STADELBAUER

DATE/TIME OF TEST: 11/09/2025 02:48:55

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:38	
Control Test		0.082	02:39	
Air Blank		0.000	02:39	
Subject Sample	# 1	0.248	02:40	No .020 agreement
Breath Volume		1.269	02:40	
Air Blank		0.000	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.227	02:43	No .020 agreement
Breath Volume		1.210	02:43	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.195	02:47	No .020 agreement
Breath Volume		2.015	02:47	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOERING,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HONORE,BRIAN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 11/10/2025 20:00:23

Diagnostics Check		OK	19:55	
Air Blank		0.000	19:55	
Control Test		0.080	19:56	
Air Blank		0.000	19:56	
Subject Sample	# 1	0.313	19:56	
Breath Volume		2.582	19:56	
Air Blank		0.000	19:57	
Air Blank		0.000	19:59	
Subject Sample	# 2	0.295	19:59	
Breath Volume		2.457	19:59	
Air Blank			19:59	RFI Detect
Air Blank		0.000	20:00	

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOERING,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HONORE,BRIAN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 11/10/2025 20:10:52

Diagnostics Check		OK	20:04
Air Blank		0.000	20:05
Control Test		0.081	20:05
Air Blank		0.000	20:06
Subject Sample	# 1	0.309	20:06
Breath Volume		1.183	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:08
Subject Sample	# 2	0.306	20:09
Breath Volume		2.367	20:09
Air Blank		0.000	20:09
Control Test		0.080	20:10
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: BURNS,JEFFREY
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/13/2025 00:37:02

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.083	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.245	00:32
Breath Volume		2.535	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.236	00:35
Breath Volume		1.101	00:35
Air Blank		0.000	00:36
Control Test		0.082	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:37

Data Download Date/Time: 11/24/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MARRERO,EMMANUEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DOS SANTOS

DATE/TIME OF TEST: 11/15/2025 00:20:04

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.082	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.174	00:15
Breath Volume		2.597	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:17
Subject Sample	# 2	0.175	00:18
Breath Volume		2.167	00:18
Air Blank		0.000	00:19
Control Test		0.082	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:20

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERDT,ANDRU,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: ECKERDT,ANDRU
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: SALINAS RUBIO

DATE/TIME OF TEST: 11/15/2025 23:54:13

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.082	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.091	23:49
Breath Volume		1.140	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:51
Subject Sample	# 2	0.090	23:52
Breath Volume		1.410	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SMITH,NICHOLAS
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/17/2025 02:40:01

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.082	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.229	02:35
Breath Volume		1.203	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.227	02:38
Breath Volume		1.128	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:40

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: PITMAN,WESLEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 11/18/2025 00:03:23

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.082	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:02	
Control Test		0.081	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: MURRAY,C
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: RIEGLER

DATE/TIME OF TEST: 11/18/2025 01:08:48

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.000	01:04
Breath Volume		1.207	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.000	01:07
Breath Volume		1.894	01:07
Air Blank		0.000	01:07
Control Test		0.081	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MARINACCI,JOSEPH
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CONNOR

DATE/TIME OF TEST: 11/21/2025 04:33:29

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:18	
Control Test		0.082	04:18	
Air Blank		0.000	04:18	
Subject Sample	# 1		04:22	No Sample Provided
Breath Volume		0.000	04:22	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 2	0.078	04:27	
Breath Volume		1.718	04:27	
Air Blank		0.000	04:28	
Air Blank		0.000	04:30	
Subject Sample	# 3	0.079	04:31	
Breath Volume		1.609	04:31	
Air Blank		0.000	04:32	
Control Test		0.082	04:32	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REBANDO,BRIAN,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: RODRIGUEZ,YEROMME
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: DOZIER

DATE/TIME OF TEST: 11/22/2025 04:59:44

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.082	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.060	04:55
Breath Volume		1.660	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.058	04:58
Breath Volume		1.656	04:58
Air Blank		0.000	04:58
Control Test		0.081	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 11/23/2025 02:18:33

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.081	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.082	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILLIS,JOSEPH
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LATORRE LAO

DATE/TIME OF TEST: 11/23/2025 03:01:54

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.082	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.000	02:57
Breath Volume		3.761	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.000	03:00
Breath Volume		3.703	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REBANDO,BRIAN,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KAY,MICHAEL
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CARLOS

DATE/TIME OF TEST: 11/23/2025 03:23:46

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.191	03:19
Breath Volume		1.718	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.195	03:22
Breath Volume		1.175	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 11/24/2025 11:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,KEVIN,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: EDWARDS,KEVIN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 11/25/2025 01:07:30

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.082	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.199	01:02
Breath Volume		1.425	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.213	01:05
Breath Volume		2.105	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 12/23/2025 10:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES,MONIQUE,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BARNES,MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RAYO MALDONADO

DATE/TIME OF TEST: 11/27/2025 03:26:07

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.081	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1	0.139	03:18	Volume Not Met
Breath Volume		0.968	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 2	0.165	03:24	Volume Not Met
Breath Volume		0.109	03:24	
Air Blank		0.000	03:25	
Control Test		0.078	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 12/23/2025 10:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES,MONIQUE,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BARNES,MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RAYO MALDONADO

DATE/TIME OF TEST: 11/27/2025 03:41:59

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.161	03:37
Breath Volume		1.980	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:39
Subject Sample	# 2	0.163	03:40
Breath Volume		1.144	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 12/23/2025 10:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,PHILIP,J
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: BUEHLER,ANTHONY
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MCCORD

DATE/TIME OF TEST: 11/29/2025 01:35:55

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.083	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.041	01:31
Breath Volume		2.421	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:33
Subject Sample	# 2	0.035	01:34
Breath Volume		2.335	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 12/23/2025 10:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: TOLEDO ALEGRIA

DATE/TIME OF TEST: 11/01/2025 01:19:04

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.130	01:14
Breath Volume		1.292	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.120	01:17
Breath Volume		1.503	01:17
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:19

Data Download Date/Time: 11/24/2025 11:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: REBANDO,J
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WILLARD

DATE/TIME OF TEST: 11/01/2025 06:08:08

Diagnostics Check		OK	06:01
Air Blank		0.000	06:01
Control Test		0.081	06:01
Air Blank		0.000	06:02
Subject Sample	# 1	0.053	06:03
Breath Volume		4.449	06:03
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.054	06:06
Breath Volume		4.269	06:06
Air Blank		0.000	06:07
Control Test		0.080	06:07
Air Blank		0.000	06:08
Diagnostics Check		OK	06:08

Data Download Date/Time: 11/24/2025 11:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LOWE,MAYCEE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/09/2025 01:50:46

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.295	01:45
Breath Volume		4.128	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.294	01:49
Breath Volume		4.117	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 11/24/2025 11:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MAZNARITZ,KAYLA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 10/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SEMAN

DATE/TIME OF TEST: 11/23/2025 02:31:07

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:26	
Control Test		0.080	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Control Test		0.080	02:30	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 11/24/2025 11:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 11/30/2025 18:46:52

Diagnostics Check		OK	18:40	
Air Blank		0.000	18:40	
Control Test		0.081	18:40	
Air Blank		0.000	18:41	
Subject Sample	# 1	0.193	18:42	
Breath Volume		1.570	18:42	
Air Blank		0.000	18:42	
Air Blank		0.000	18:44	
Subject Sample	# 2		18:45	RFI Detect
Breath Volume		1.546	18:45	
Air Blank		0.000	18:45	
Control Test		0.080	18:46	
Air Blank		0.000	18:46	
Diagnostics Check		OK	18:46	

Data Download Date/Time: 12/19/2025 23:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 11/30/2025 18:54:26

Diagnostics Check		OK	18:48	
Air Blank		0.000	18:48	
Control Test		0.081	18:49	
Air Blank		0.000	18:49	
Subject Sample	# 1	0.205	18:50	
Breath Volume		2.265	18:50	
Air Blank		0.000	18:51	
Air Blank		0.000	18:53	
Subject Sample	# 2		18:53	RFI Detect
Breath Volume		0.000	18:53	
Air Blank			18:53	RFI Detect
Air Blank		0.000	18:54	

Data Download Date/Time: 12/19/2025 23:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 11/06/2025 21:24:03

Diagnostics Check		OK	21:17
Air Blank		0.000	21:18
Control Test		0.079	21:18
Air Blank		0.000	21:18
Subject Sample	# 1	0.155	21:19
Breath Volume		1.953	21:19
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.158	21:22
Breath Volume		1.285	21:22
Air Blank		0.000	21:23
Control Test		0.078	21:23
Air Blank		0.000	21:23
Diagnostics Check		OK	21:24

Data Download Date/Time: 11/12/2025 17:33

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ZACE

DATE/TIME OF TEST: 11/16/2025 02:19:02

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.132	02:14
Breath Volume		2.992	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.125	02:17
Breath Volume		2.429	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:19

Data Download Date/Time: 12/10/2025 22:18

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSME,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MILLER,ZACHARY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: CATLETT

DATE/TIME OF TEST: 11/01/2025 02:52:34

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.081	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.000	02:48
Breath Volume		1.757	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.000	02:51
Breath Volume		2.519	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 11/26/2025 23:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSME,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: NELSON,JADEN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: HESLOP

DATE/TIME OF TEST: 11/02/2025 03:04:52

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.188	03:00
Breath Volume		2.035	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:02
Subject Sample	# 2	0.190	03:03
Breath Volume		2.699	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSME,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CUMBIE,JAMES
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MCNEAL

DATE/TIME OF TEST: 11/03/2025 03:07:38

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.150	03:03
Breath Volume		1.605	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.150	03:06
Breath Volume		1.234	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 11/05/2025 20:16:36

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.081	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.000	20:12
Breath Volume		2.722	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.000	20:15
Breath Volume		2.277	20:15
Air Blank		0.000	20:15
Control Test		0.080	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/07/2025 00:49:07

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.055	00:44
Breath Volume		1.847	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.054	00:47
Breath Volume		1.410	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ASHFORD

DATE/TIME OF TEST: 11/12/2025 20:11:47

Diagnostics Check		OK	20:05
Air Blank		0.000	20:06
Control Test		0.080	20:06
Air Blank		0.000	20:07
Subject Sample	# 1	0.219	20:07
Breath Volume		1.386	20:07
Air Blank		0.000	20:08
Air Blank		0.000	20:09
Subject Sample	# 2	0.215	20:10
Breath Volume		1.605	20:10
Air Blank		0.000	20:10
Control Test		0.080	20:11
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAK,STEVEN,M
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: VIDAK,STEVEN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 11/12/2025 22:00:31

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:58	
Control Test		0.079	21:58	
Air Blank		0.000	21:58	
Subject Sample	# 1		21:59	Subject Test Refused
Breath Volume		0.000	21:59	
Air Blank		0.000	21:59	
Control Test		0.080	21:59	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIPF,RICHARD,K
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ROLAND

DATE/TIME OF TEST: 11/16/2025 01:13:50

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1	0.174	01:03	
Breath Volume		1.207	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2	0.169	01:09	Volume Not Met
Breath Volume		1.058	01:09	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 3	0.175	01:12	
Breath Volume		1.953	01:12	
Air Blank		0.000	01:12	
Control Test		0.078	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY,ELISE,N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CADY,ELISE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SAPP

DATE/TIME OF TEST: 11/16/2025 04:32:49

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.157	04:28
Breath Volume		1.718	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.156	04:31
Breath Volume		1.992	04:31
Air Blank		0.000	04:31
Control Test		0.079	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOROS,CHUWAN,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BOROS,CHUWAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 11/17/2025 17:10:20

Diagnostics Check		OK	17:03
Air Blank		0.000	17:03
Control Test		0.079	17:04
Air Blank		0.000	17:04
Subject Sample	# 1	0.185	17:05
Breath Volume		1.171	17:05
Air Blank		0.000	17:06
Air Blank		0.000	17:07
Subject Sample	# 2	0.193	17:08
Breath Volume		2.066	17:08
Air Blank		0.000	17:09
Control Test		0.077	17:09
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOROS,CHUWAN,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BOROS,CHUWAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: DRIBAN

DATE/TIME OF TEST: 11/20/2025 23:04:30

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:53	
Control Test		0.080	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 1	0.160	22:57	Volume Not Met
Breath Volume		0.136	22:57	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 2		23:03	No Sample Provided
Breath Volume		0.000	23:03	
Air Blank		0.000	23:03	
Control Test		0.077	23:03	
Air Blank		0.000	23:04	
Diagnostics Check		OK	23:04	

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOROS,CHUWAN,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BOROS,CHUWAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DRIBAN

DATE/TIME OF TEST: 11/20/2025 23:23:43

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.079	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.079	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 11/26/2025 23:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KYLE,F
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOLMES,KYLE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 11/21/2025 00:13:32

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.080	00:07
Air Blank		0.000	00:08
Subject Sample	# 1	0.091	00:09
Breath Volume		1.839	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.092	00:12
Breath Volume		2.238	00:12
Air Blank		0.000	00:12
Control Test		0.080	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 11/26/2025 23:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIPF,RICHARD,K
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 11/25/2025 02:43:59

Diagnostics Check		OK	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.205	02:38
Breath Volume		1.941	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.205	02:42
Breath Volume		2.230	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 11/26/2025 23:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY,ELISE,N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BROWN,CHRISTIAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: AVENDANO GUILLEN

DATE/TIME OF TEST: 11/26/2025 23:49:38

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.169	23:45
Breath Volume		1.527	23:45
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.160	23:48
Breath Volume		1.984	23:48
Air Blank		0.000	23:48
Control Test		0.081	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 12/22/2025 21:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIPF,RICHARD,K
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 11/27/2025 04:06:57

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.149	04:02
Breath Volume		2.843	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.146	04:05
Breath Volume		2.937	04:05
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 12/22/2025 21:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIPF,RICHARD,K
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 11/27/2025 22:30:48

Diagnostics Check		OK	22:19	
Air Blank		0.000	22:19	
Control Test		0.080	22:20	
Air Blank		0.000	22:20	
Subject Sample	# 1		22:23	No Sample Provided
Breath Volume		0.000	22:23	
Air Blank		0.000	22:24	
Air Blank		0.000	22:26	
Subject Sample	# 2	0.119	22:29	Volume Not Met
Breath Volume		0.660	22:29	
Air Blank		0.000	22:29	
Control Test		0.080	22:30	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 12/22/2025 21:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIPF,RICHARD,K
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 11/27/2025 22:57:06

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.116	22:52
Breath Volume		2.328	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.115	22:55
Breath Volume		1.464	22:55
Air Blank		0.000	22:56
Control Test		0.081	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 12/22/2025 21:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOROS,CHUWAN,
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: GLASS,B
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 11/29/2025 01:02:39

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.000	00:58
Breath Volume		2.253	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.000	01:01
Breath Volume		3.089	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 12/22/2025 21:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMUEL,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BUZZO,HEATHER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CASTILLO-MARTINEZ

DATE/TIME OF TEST: 11/30/2025 04:33:06

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.174	04:28
Breath Volume		2.273	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.173	04:31
Breath Volume		2.820	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 12/22/2025 21:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403278599
DRY GAS EXPIRY DATE: 02/24/2028
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/30/2025 16:28:28

Diagnostics Check		OK	16:22
Air Blank		0.000	16:22
Control Test		0.080	16:23
Air Blank		0.000	16:23
Subject Sample	# 1	0.095	16:24
Breath Volume		1.761	16:24
Air Blank		0.000	16:24
Air Blank		0.000	16:26
Subject Sample	# 2	0.093	16:27
Breath Volume		1.460	16:27
Air Blank		0.000	16:27
Control Test		0.080	16:27
Air Blank		0.000	16:28
Diagnostics Check		OK	16:28

Data Download Date/Time: 12/22/2025 21:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOULIHAN,RILEY,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MICKEL,CONNOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BOCIO

DATE/TIME OF TEST: 11/01/2025 02:15:46

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.082	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.214	02:11
Breath Volume		2.921	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.206	02:14
Breath Volume		2.605	02:14
Air Blank		0.000	02:14
Control Test		0.082	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGATTI,BRADLEY,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KNOWLTON,CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MENJIVAR

DATE/TIME OF TEST: 11/01/2025 03:55:36

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.185	03:51
Breath Volume		1.113	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.187	03:54
Breath Volume		1.425	03:54
Air Blank		0.000	03:54
Control Test		0.081	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDEMANN,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CAMPOS,ABRIL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: REVOLLEDO MACHADO

DATE/TIME OF TEST: 11/02/2025 00:28:02

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:12	
Control Test		0.082	00:13	
Air Blank		0.000	00:13	
Subject Sample	# 1	0.136	00:15	
Breath Volume		1.109	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 2	0.133	00:21	Volume Not Met
Breath Volume		0.015	00:21	
Air Blank		0.000	00:21	
Air Blank		0.000	00:23	
Subject Sample	# 3	0.147	00:26	
Breath Volume		1.136	00:26	
Air Blank		0.000	00:27	
Control Test		0.081	00:27	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JACKSON,JACOB
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SAO GONZALEZ

DATE/TIME OF TEST: 11/08/2025 02:09:18

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.082	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.150	02:04
Breath Volume		1.214	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.156	02:07
Breath Volume		1.328	02:07
Air Blank		0.000	02:08
Control Test		0.082	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARTBAUER,DAKOTA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BERNAL TORRES

DATE/TIME OF TEST: 11/08/2025 03:15:24

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.170	03:10
Breath Volume		2.230	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.177	03:13
Breath Volume		2.589	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JACKSON,JACOB
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/09/2025 03:48:44

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.196	03:43
Breath Volume		1.332	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.180	03:47
Breath Volume		1.386	03:47
Air Blank		0.000	03:47
Control Test		0.082	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HILL,TIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/09/2025 04:43:29

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.081	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.164	04:39
Breath Volume		3.496	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.161	04:42
Breath Volume		3.488	04:42
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HANLOTXOMPHOU

DATE/TIME OF TEST: 11/16/2025 01:10:12

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.082	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.022	01:05
Breath Volume		1.453	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.021	01:08
Breath Volume		1.750	01:08
Air Blank		0.000	01:09
Control Test		0.082	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGATTI,BRADLEY,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON,WILLIE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 11/16/2025 04:19:28

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.081	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:12	No Sample Provided
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2		04:18	No Sample Provided
Breath Volume		0.000	04:18	
Air Blank		0.000	04:18	
Control Test		0.082	04:18	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: GUILLEN GONZALEZ

DATE/TIME OF TEST: 11/17/2025 19:41:30

Diagnostics Check		OK	19:34
Air Blank		0.000	19:34
Control Test		0.082	19:35
Air Blank		0.000	19:35
Subject Sample	# 1	0.095	19:36
Breath Volume		2.429	19:36
Air Blank		0.000	19:37
Air Blank		0.000	19:39
Subject Sample	# 2	0.095	19:40
Breath Volume		2.402	19:40
Air Blank		0.000	19:40
Control Test		0.081	19:40
Air Blank		0.000	19:41
Diagnostics Check		OK	19:41

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LANGEVIN

DATE/TIME OF TEST: 11/18/2025 19:11:20

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.081	19:05
Air Blank		0.000	19:05
Subject Sample	# 1	0.320	19:06
Breath Volume		1.554	19:06
Air Blank		0.000	19:07
Air Blank		0.000	19:09
Subject Sample	# 2	0.323	19:09
Breath Volume		1.574	19:09
Air Blank		0.000	19:10
Control Test		0.081	19:10
Air Blank		0.000	19:11
Diagnostics Check		OK	19:11

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP,RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP,RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: ALANIA ESTRELLA

DATE/TIME OF TEST: 11/20/2025 21:32:14

Diagnostics Check		OK	21:24
Air Blank		0.000	21:25
Control Test		0.082	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.155	21:27
Breath Volume		1.449	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.146	21:30
Breath Volume		1.503	21:30
Air Blank		0.000	21:31
Control Test		0.081	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: STINNETT

DATE/TIME OF TEST: 11/22/2025 02:25:07

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.082	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.133	02:20
Breath Volume		2.753	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.133	02:23
Breath Volume		2.570	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 11/22/2025 02:37:21

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.082	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.177	02:32
Breath Volume		2.500	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.167	02:35
Breath Volume		1.847	02:35
Air Blank		0.000	02:36
Control Test		0.081	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,D,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: ORTIZ VAZQUEZ

DATE/TIME OF TEST: 11/23/2025 20:26:50

Diagnostics Check		OK	20:20
Air Blank		0.000	20:20
Control Test		0.082	20:21
Air Blank		0.000	20:21
Subject Sample	# 1	0.263	20:22
Breath Volume		2.085	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:24
Subject Sample	# 2	0.272	20:25
Breath Volume		2.914	20:25
Air Blank		0.000	20:25
Control Test		0.080	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 11/24/2025 19:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/27/2025 23:35:28

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.082	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.000	23:31
Breath Volume		3.718	23:31
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.000	23:34
Breath Volume		4.058	23:34
Air Blank		0.000	23:34
Control Test		0.082	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 12/12/2025 20:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: GIAMBALVO

DATE/TIME OF TEST: 11/29/2025 20:16:55

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:13	
Control Test		0.081	20:13	
Air Blank		0.000	20:13	
Subject Sample	# 1		20:15	Subject Test Refused
Breath Volume		0.000	20:15	
Air Blank		0.000	20:16	
Control Test		0.083	20:16	
Air Blank		0.000	20:16	
Diagnostics Check		OK	20:16	

Data Download Date/Time: 12/12/2025 20:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: RAGSDALE

DATE/TIME OF TEST: 11/30/2025 01:35:06

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.082	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.155	01:30
Breath Volume		2.738	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.141	01:33
Breath Volume		1.570	01:33
Air Blank		0.000	01:34
Control Test		0.082	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 12/12/2025 20:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 11/06/2025 23:20:52

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:09	
Control Test		0.079	23:10	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:13	No Sample Provided
Breath Volume		0.000	23:13	
Air Blank		0.000	23:14	
Air Blank		0.000	23:16	
Subject Sample	# 2		23:19	No Sample Provided
Breath Volume		0.000	23:19	
Air Blank		0.000	23:19	
Control Test		0.080	23:20	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:20	

Data Download Date/Time: 12/23/2025 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BOWMAN III

DATE/TIME OF TEST: 11/11/2025 23:37:28

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:26	
Control Test		0.081	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:30	No Sample Provided
Breath Volume		0.000	23:30	
Air Blank		0.000	23:30	
Air Blank		0.000	23:32	
Subject Sample	# 2		23:36	No Sample Provided
Breath Volume		0.000	23:36	
Air Blank		0.000	23:36	
Control Test		0.080	23:36	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 12/23/2025 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LOVGREN,KAVEN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: WREN

DATE/TIME OF TEST: 11/01/2025 03:22:33

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:15	
Control Test		0.080	03:16	
Air Blank		0.000	03:16	
Subject Sample	# 1	0.121	03:17	
Breath Volume		2.101	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 2	0.114	03:20	
Breath Volume		1.824	03:20	
Air Blank		0.000	03:21	
Control Test		0.079	03:21	
Air Blank		0.000	03:22	
Diagnostics Check		Fail	03:22	Voltage/Current Test
Air Blank		0.000	03:22	

Data Download Date/Time: 11/14/2025 10:52

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MATSAGAS

DATE/TIME OF TEST: 11/01/2025 21:56:48

Diagnostics Check		OK	21:50
Air Blank		0.000	21:50
Control Test		0.080	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.133	21:52
Breath Volume		3.199	21:52
Air Blank		0.000	21:52
Air Blank		0.000	21:54
Subject Sample	# 2	0.134	21:55
Breath Volume		3.476	21:55
Air Blank		0.000	21:55
Control Test		0.079	21:56
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 11/14/2025 10:52

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: MEDICH

DATE/TIME OF TEST: 11/02/2025 19:02:34

Diagnostics Check		OK	18:55
Air Blank		0.000	18:56
Control Test		0.079	18:56
Air Blank		0.000	18:57
Subject Sample	# 1	0.108	18:57
Breath Volume		1.308	18:57
Air Blank		0.000	18:58
Air Blank		0.000	19:00
Subject Sample	# 2	0.113	19:01
Breath Volume		1.617	19:01
Air Blank		0.000	19:01
Control Test		0.079	19:01
Air Blank		0.000	19:02
Diagnostics Check		OK	19:02

Data Download Date/Time: 11/14/2025 10:52

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: GAUDETTE

DATE/TIME OF TEST: 11/06/2025 01:31:58

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.160	01:27
Breath Volume		2.226	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.159	01:30
Breath Volume		2.664	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 11/14/2025 10:52

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: LYON

DATE/TIME OF TEST: 11/06/2025 22:25:53

Diagnostics Check	Fail	22:25	Voltage/Current Test
Air Blank	0.000	22:25	

Data Download Date/Time: 11/14/2025 10:52

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: LYON

DATE/TIME OF TEST: 11/06/2025 22:35:33

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.079	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.187	22:30
Breath Volume		2.906	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:32
Subject Sample	# 2	0.171	22:33
Breath Volume		1.812	22:33
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 11/14/2025 10:52

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETROWSKI,DAVID
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HEIDEN

DATE/TIME OF TEST: 11/12/2025 01:23:42

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.171	01:19
Breath Volume		1.789	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.180	01:22
Breath Volume		2.292	01:22
Air Blank		0.000	01:22
Control Test		0.080	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 11/14/2025 10:52

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LOVGREN,KAVEN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 11/16/2025 20:07:24

Diagnostics Check		OK	20:01
Air Blank		0.000	20:01
Control Test		0.079	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.175	20:02
Breath Volume		5.203	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.171	20:05
Breath Volume		4.449	20:05
Air Blank		0.000	20:06
Control Test		0.079	20:06
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 12/12/2025 08:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO-PEREZ,RODOLFO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CASTILLO-PEREZ,RODOLFO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: GUILLAUME

DATE/TIME OF TEST: 11/01/2025 21:37:39

Diagnostics Check		OK	21:28
Air Blank		0.000	21:29
Control Test		0.079	21:29
Air Blank		0.000	21:29
Subject Sample	# 1	0.342	21:30
Breath Volume		1.335	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.331	21:35
Breath Volume		1.488	21:35
Air Blank		0.000	21:36
Control Test		0.078	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:13

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 11/02/2025 13:36:19

Diagnostics Check		OK	13:33	
Air Blank		0.000	13:33	
Control Test		0.079	13:34	
Air Blank		0.000	13:34	
Subject Sample	# 1		13:34	RFI Detect
Breath Volume		0.000	13:34	
Air Blank		0.000	13:35	
Control Test		0.078	13:35	
Air Blank		0.000	13:36	
Diagnostics Check		OK	13:36	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:13

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 11/02/2025 13:46:38

Diagnostics Check		OK	13:40
Air Blank		0.000	13:41
Control Test		0.080	13:41
Air Blank		0.000	13:41
Subject Sample	# 1	0.000	13:42
Breath Volume		2.320	13:42
Air Blank		0.000	13:42
Air Blank		0.000	13:44
Subject Sample	# 2	0.000	13:45
Breath Volume		1.390	13:45
Air Blank		0.000	13:45
Control Test		0.079	13:46
Air Blank		0.000	13:46
Diagnostics Check		OK	13:46

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 11/03/2025 01:04:06

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.096	00:59
Breath Volume		2.480	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.095	01:02
Breath Volume		1.484	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: FLORYANCE,CODY
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/04/2025 01:18:12

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.079	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.078	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: RAMOS RENTERIA

DATE/TIME OF TEST: 11/07/2025 04:48:34

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:45	
Control Test		0.079	04:46	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:47	Subject Test Refused
Breath Volume		0.000	04:47	
Air Blank		0.000	04:47	
Control Test		0.079	04:47	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: RODAS RODRIGUEZ

DATE/TIME OF TEST: 11/08/2025 02:04:33

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.183	02:00
Breath Volume		2.664	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.170	02:03
Breath Volume		1.117	02:03
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ASHER

DATE/TIME OF TEST: 11/09/2025 03:39:42

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.079	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.223	03:35
Breath Volume		3.039	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.212	03:38
Breath Volume		2.218	03:38
Air Blank		0.000	03:38
Control Test		0.078	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,TY,D
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/11/2025 21:55:48

Diagnostics Check		OK	21:50	
Air Blank		0.000	21:50	
Control Test		0.079	21:51	
Air Blank		0.000	21:51	
Subject Sample	# 1		21:54	Subject Test Refused
Breath Volume		0.000	21:54	
Air Blank		0.000	21:54	
Control Test		0.079	21:55	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: ANDRADE,ESTEBAN
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ROJAS RAMIREZ

DATE/TIME OF TEST: 11/12/2025 23:14:15

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:10	
Control Test		0.079	23:10	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:12	Subject Test Refused
Breath Volume		0.000	23:12	
Air Blank		0.000	23:13	
Control Test		0.079	23:13	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:14	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,AYDEN,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: FLETCHER,AYDEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FARLAND

DATE/TIME OF TEST: 11/13/2025 00:23:00

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.115	00:18
Breath Volume		2.726	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:20
Subject Sample	# 2	0.104	00:21
Breath Volume		1.699	00:21
Air Blank		0.000	00:22
Control Test		0.079	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:23

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 11/13/2025 13:43:34

Diagnostics Check		OK	13:40	
Air Blank		0.000	13:40	
Control Test		0.078	13:41	
Air Blank		0.000	13:41	
Subject Sample	# 1		13:42	Subject Test Refused
Breath Volume		0.000	13:42	
Air Blank		0.000	13:42	
Control Test		0.079	13:42	
Air Blank		0.000	13:43	
Diagnostics Check		OK	13:43	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: ADORNO,SAMUEL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/13/2025 21:23:46

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:21	
Control Test		0.079	21:21	
Air Blank		0.000	21:21	
Subject Sample	# 1		21:22	Subject Test Refused
Breath Volume		0.000	21:22	
Air Blank		0.000	21:22	
Control Test		0.079	21:23	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: ADORNO,SAMUEL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: OSBORN

DATE/TIME OF TEST: 11/14/2025 04:13:52

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.173	04:08
Breath Volume		1.199	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.166	04:12
Breath Volume		1.945	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO PEREZ,RODOLFO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CASTILLO PEREZ,RODOLFO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: VALVERDE LEZAMA

DATE/TIME OF TEST: 11/14/2025 22:43:35

Diagnostics Check	Fail	22:43	Voltage/Current Test
Air Blank	0.000	22:43	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO PEREZ,RODOLFO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CASTILLO PEREZ,RODOLFO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: VALVERDE LEZAMA

DATE/TIME OF TEST: 11/14/2025 22:51:23

Diagnostics Check		OK	22:48	
Air Blank		0.000	22:49	
Control Test		0.079	22:49	
Air Blank		0.000	22:49	
Subject Sample	# 1		22:50	Subject Test Refused
Breath Volume		0.000	22:50	
Air Blank		0.000	22:50	
Control Test		0.079	22:50	
Air Blank		0.000	22:51	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO PEREZ,RODOLFO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CASTILLO PEREZ,RODOLFO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 11/15/2025 03:03:14

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.080	03:01	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:02	
Control Test		0.079	03:02	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CAMPO

DATE/TIME OF TEST: 11/15/2025 03:13:52

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.079	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1	0.169	03:09	
Breath Volume		4.105	03:09	
Air Blank		0.000	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 2	0.170	03:12	
Breath Volume		4.242	03:12	
Air Blank		0.000	03:12	
Control Test		0.079	03:13	
Air Blank			03:13	RFI Detect
Air Blank		0.000	03:13	

Data Download Date/Time: 11/18/2025 13:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: GOLOVANOV,IVAN
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: VAN DE MEER

DATE/TIME OF TEST: 11/19/2025 01:22:32

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.079	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.190	01:19	
Breath Volume		1.320	01:19	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2	0.191	01:21	
Breath Volume		1.640	01:21	
Air Blank			01:22	Improper Sample
Air Blank		0.000	01:22	

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: FLORES,MADILYN
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 11/20/2025 02:40:37

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.144	02:36
Breath Volume		2.898	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.142	02:39
Breath Volume		2.707	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LIZ

DATE/TIME OF TEST: 11/22/2025 22:41:50

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:39	
Control Test		0.079	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.079	22:41	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ROSS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HUGHES,ROSS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MORRAN

DATE/TIME OF TEST: 11/23/2025 02:43:22

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.078	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: CAHILL,CHRISTOPHER
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: THEISING

DATE/TIME OF TEST: 11/24/2025 02:20:08

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.080	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:19	
Control Test		0.079	02:19	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRERA,VICTOR,R
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JEAN BAPTISTE,STANLEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BORA

DATE/TIME OF TEST: 11/26/2025 04:16:35

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.092	04:12
Breath Volume		2.414	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.093	04:15
Breath Volume		1.402	04:15
Air Blank		0.000	04:15
Control Test		0.078	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ROSS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HUGHES,ROSS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 11/27/2025 04:13:06

Diagnostics Check		OK	03:59	
Air Blank		0.000	04:00	
Control Test		0.080	04:00	
Air Blank		0.000	04:00	
Subject Sample	# 1	0.171	04:04	Volume Not Met
Breath Volume		0.683	04:04	
Air Blank		0.000	04:04	
Air Blank		0.000	04:06	
Subject Sample	# 2	0.180	04:07	
Breath Volume		1.648	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 3	0.168	04:11	
Breath Volume		1.816	04:11	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: STOREY,T
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 11/27/2025 16:22:37

Diagnostics Check		OK	16:19	
Air Blank		0.000	16:20	
Control Test		0.080	16:20	
Air Blank		0.000	16:20	
Subject Sample	# 1		16:21	Subject Test Refused
Breath Volume		0.000	16:21	
Air Blank		0.000	16:21	
Control Test		0.080	16:22	
Air Blank		0.000	16:22	
Diagnostics Check		OK	16:22	

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: ANDRADE,ESTEBAN
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: HENGELBROK

DATE/TIME OF TEST: 11/28/2025 05:24:02

Diagnostics Check	OK	05:23	
Air Blank		05:23	RFI Detect
Air Blank		05:24	RFI Detect

Data Download Date/Time: 12/11/2025 11:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: VOLK

DATE/TIME OF TEST: 11/29/2025 01:16:19

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.108	01:11
Breath Volume		2.402	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.109	01:14
Breath Volume		1.511	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 11/30/2025 00:31:16

Diagnostics Check	OK	00:30	
Air Blank		00:31	RFI Detect
Air Blank		00:31	RFI Detect

Data Download Date/Time: 12/11/2025 11:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 11/30/2025 00:39:35

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.079	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:35
Breath Volume		2.000	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.000	00:38
Breath Volume		3.460	00:38
Air Blank		0.000	00:38
Control Test		0.078	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SHARIPOV

DATE/TIME OF TEST: 11/30/2025 04:41:53

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.195	04:37
Breath Volume		1.472	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:39
Subject Sample	# 2	0.186	04:40
Breath Volume		2.550	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 12/11/2025 11:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: KEY WEST PD
ARRESTING OFFICER: WRIGHT,DANIEL
ARRESTING OFFICER AGENCY: KEY WEST PD
LAST AGENCY INSPECTION: 11/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: DARBOUZE

DATE/TIME OF TEST: 11/30/2025 05:23:43

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:21	
Control Test		0.079	05:21	
Air Blank		0.000	05:22	
Subject Sample	# 1		05:22	Subject Test Refused
Breath Volume		0.000	05:22	
Air Blank		0.000	05:22	
Control Test		0.080	05:23	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 12/11/2025 12:00

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.