Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ZAMORA,EVERLYN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 11/01/2025 08:20:16

Diagnostics Check		OK	08:06	
Air Blank		0.000	08:06	
Control Test		0.081	08:07	
Air Blank		0.000	08:07	
Subject Sample	# 1	0.207	08:10	Volume Not Met
Breath Volume		0.238	08:10	
Air Blank		0.000	08:11	
Air Blank		0.000	08:13	
Subject Sample	# 2	0.249	08:15	
Breath Volume		1.910	08:15	
Air Blank		0.000	08:16	
Air Blank		0.000	08:17	
Subject Sample	# 3	0.250	08:18	
Breath Volume		1.882	08:18	
Air Blank		0.000	08:19	
Control Test		0.077	08:19	
Air Blank		0.000	08:20	
Diagnostics Check		OK	08:20	

Data Download Date/Time: 11/22/2025 06:46

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,OSCAR,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROQUE,OSCAR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: RODRIGUEZ RAMIREZ

DATE/TIME OF TEST: 11/07/2025 23:43:56

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:36	
Control Test		0.081	23:37	
Air Blank		0.000	23:37	
Subject Sample	# 1	0.387	23:40	Volume Not Met
Breath Volume		0.082	23:40	
Air Blank		0.000	23:41	
Air Blank		0.000	23:43	
Subject Sample	# 2		23:43	Sequence Aborted
Breath Volume		0.000	23:43	
Air Blank		0.000	23:43	

Data Download Date/Time: 11/22/2025 06:46

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ZAMORA,E
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GARCIA DOMINGO

DATE/TIME OF TEST: 11/22/2025 04:08:08

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.169	04:03
Breath Volume		2.480	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.170	04:06
Breath Volume		2.152	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 11/22/2025 06:46

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CORIO CHAVEZ

DATE/TIME OF TEST: 11/27/2025 02:21:04

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.082	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.211	02:16
Breath Volume		1.386	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:18
Subject Sample	# 2	0.207	02:19
Breath Volume		1.679	02:19
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:21

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MONZON ALVAREZ

DATE/TIME OF TEST: 11/28/2025 22:16:28

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:08	
Control Test		0.081	22:08	
Air Blank		0.000	22:09	
Subject Sample	# 1		22:12	No Sample Provided
Breath Volume		0.000	22:12	
Air Blank		0.000	22:12	
Air Blank		0.000	22:14	
Subject Sample	# 2		22:15	Subject Test Refused
Breath Volume		0.000	22:15	
Air Blank		0.000	22:15	
Control Test		0.081	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBERNARDINO,BRANDON,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FARRAJ,SANAD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/01/2025 02:46:44

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.084	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.111	02:41
Breath Volume		1.738	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.098	02:44
Breath Volume		1.832	02:44
Air Blank		0.000	02:45
Control Test		0.083	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 11/14/2025 10:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIENTOS,KENNETH,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GONZALEZ,OSCAR
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: CASTILLO-LEAL

DATE/TIME OF TEST: 11/15/2025 04:56:17

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:52	
Control Test		0.083	04:52	
Air Blank		0.000	04:52	
Subject Sample	# 1		04:54	Subject Test Refused
Breath Volume		0.000	04:54	
Air Blank		0.000	04:55	
Control Test		0.084	04:55	
Air Blank		0.000	04:56	
Diagnostics Check		OK	04:56	

Data Download Date/Time: 12/12/2025 10:00

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LARRY,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HARTWELL,DEVON
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 11/16/2025 19:45:46

Diagnostics Check		OK	19:37
Air Blank		0.000	19:37
Control Test		0.083	19:37
Air Blank		0.000	19:38
Subject Sample	# 1	0.138	19:39
Breath Volume		1.101	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.126	19:43
Breath Volume		1.363	19:43
Air Blank		0.000	19:44
Control Test		0.082	19:45
Air Blank		0.000	19:45
Diagnostics Check		OK	19:45

Data Download Date/Time: 12/12/2025 10:00

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEITH,JOHN,W
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SOSA,GABRIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: HERNANDEZ MENDOZA

DATE/TIME OF TEST: 11/19/2025 20:17:27

Diagnostics Check		OK	20:05	
Air Blank		0.000	20:05	
Control Test		0.083	20:06	
Air Blank		0.000	20:06	
Subject Sample	# 1		20:09	No Sample Provided
Breath Volume		0.000	20:09	
Air Blank		0.000	20:10	
Air Blank		0.000	20:12	
Subject Sample	# 2	0.223	20:15	Volume Not Met
Breath Volume		0.656	20:15	
Air Blank		0.000	20:16	
Control Test		0.083	20:16	
Air Blank		0.000	20:17	
Diagnostics Check		OK	20:17	

Data Download Date/Time: 12/12/2025 10:00

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEITH,JOHN,W
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SOSA,GABRIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: HERNANDEZ MENDOZA

DATE/TIME OF TEST: 11/19/2025 20:29:15

Diagnostics Check		OK	20:26	
Air Blank		0.000	20:26	
Control Test		0.080	20:27	
Air Blank		0.000	20:27	
Subject Sample	# 1		20:27	Subject Test Refused
Breath Volume		0.000	20:27	
Air Blank		0.000	20:28	
Control Test		0.082	20:28	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 12/12/2025 10:00

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KRETCHMAN

DATE/TIME OF TEST: 11/01/2025 21:32:13

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:26	
Control Test		0.082	21:26	
Air Blank		0.000	21:27	
Subject Sample	# 1	0.332	21:30	
Breath Volume		1.187	21:30	
Air Blank			21:31	Purge Fail
Air Blank		0.000	21:32	

Data Download Date/Time: 11/10/2025 14:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KRETCHMAN

DATE/TIME OF TEST: 11/01/2025 21:41:13

Diagnostics Check		OK	21:33
Air Blank		0.000	21:34
Control Test		0.077	21:34
Air Blank		0.000	21:35
Subject Sample	# 1	0.335	21:36
Breath Volume		2.222	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.337	21:39
Breath Volume		1.128	21:39
Air Blank		0.000	21:40
Control Test		0.079	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 11/10/2025 14:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: ZAMORA GUILLERMO

DATE/TIME OF TEST: 11/02/2025 18:11:52

Diagnostics Check		OK	18:05
Air Blank		0.000	18:05
Control Test		0.081	18:06
Air Blank		0.000	18:06
Subject Sample	# 1	0.147	18:07
Breath Volume		3.023	18:07
Air Blank		0.000	18:08
Air Blank		0.000	18:09
Subject Sample	# 2	0.137	18:10
Breath Volume		2.433	18:10
Air Blank		0.000	18:10
Control Test		0.080	18:11
Air Blank		0.000	18:11
Diagnostics Check		OK	18:11

Data Download Date/Time: 11/10/2025 14:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ALONZO

DATE/TIME OF TEST: 11/04/2025 02:10:02

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:59	
Control Test		0.082	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		02:03	No Sample Provided
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2		02:08	No Sample Provided
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.081	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 11/10/2025 14:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: AROSEN

DATE/TIME OF TEST: 11/04/2025 22:15:59

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:13	
Control Test		0.081	22:13	
Air Blank		0.000	22:13	
Subject Sample	# 1	0.285	22:14	
Breath Volume		1.371	22:14	
Air Blank			22:15	Purge Fail
Air Blank		0.000	22:15	

Data Download Date/Time: 11/10/2025 14:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: AROSEN

DATE/TIME OF TEST: 11/04/2025 22:24:57

Diagnostics Check		OK	22:17
Air Blank		0.000	22:18
Control Test		0.078	22:18
Air Blank		0.000	22:19
Subject Sample	# 1	0.298	22:19
Breath Volume		2.273	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.296	22:23
Breath Volume		2.171	22:23
Air Blank		0.000	22:23
Control Test		0.078	22:24
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 11/10/2025 14:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: BONDS

DATE/TIME OF TEST: 11/09/2025 19:31:08

Diagnostics Check		OK	19:24
Air Blank		0.000	19:25
Control Test		0.081	19:25
Air Blank		0.000	19:26
Subject Sample	# 1	0.104	19:26
Breath Volume		1.671	19:26
Air Blank		0.000	19:27
Air Blank		0.000	19:29
Subject Sample	# 2	0.100	19:29
Breath Volume		1.992	19:29
Air Blank		0.000	19:30
Control Test		0.079	19:30
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 11/10/2025 14:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SARAH,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BROWN,SARAH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 11/09/2025 21:00:21

Diagnostics Check		OK	20:53
Air Blank		0.000	20:53
Control Test		0.080	20:54
Air Blank		0.000	20:54
Subject Sample	# 1	0.184	20:55
Breath Volume		2.175	20:55
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.179	20:58
Breath Volume		1.796	20:58
Air Blank		0.000	20:59
Control Test		0.079	20:59
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 11/10/2025 14:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FUENTES,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WERNET

DATE/TIME OF TEST: 11/01/2025 00:47:31

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.132	00:42
Breath Volume		1.132	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.122	00:45
Breath Volume		1.734	00:45
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/01/2025 00:55:44

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.080	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.129	00:51
Breath Volume		4.640	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.127	00:54
Breath Volume		4.523	00:54
Air Blank		0.000	00:54
Control Test		0.080	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WALDROP,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: CHAUNCEY

DATE/TIME OF TEST: 11/01/2025 19:30:30

Diagnostics Check		OK	19:23
Air Blank		0.000	19:24
Control Test		0.081	19:24
Air Blank		0.000	19:25
Subject Sample	# 1	0.000	19:25
Breath Volume		1.550	19:25
Air Blank		0.000	19:26
Air Blank		0.000	19:28
Subject Sample	# 2	0.000	19:29
Breath Volume		2.859	19:29
Air Blank		0.000	19:29
Control Test		0.081	19:29
Air Blank		0.000	19:30
Diagnostics Check		OK	19:30

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: RUNN

DATE/TIME OF TEST: 11/02/2025 01:43:58

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.134	01:39
Breath Volume		2.011	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.136	01:42
Breath Volume		1.300	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAERTZ,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: HAVIRD

DATE/TIME OF TEST: 11/05/2025 05:43:36

Diagnostics Check		OK	05:34	
Air Blank		0.000	05:35	
Control Test		0.081	05:35	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:39	No Sample Provided
Breath Volume		0.000	05:39	
Air Blank		0.000	05:39	
Air Blank		0.000	05:41	
Subject Sample	# 2		05:42	Subject Test Refused
Breath Volume		0.000	05:42	
Air Blank		0.000	05:42	
Control Test		0.080	05:42	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:43	

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BLEVINS

DATE/TIME OF TEST: 11/07/2025 23:00:08

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:50	
Control Test		0.080	22:51	
Air Blank		0.000	22:51	
Subject Sample	# 1		22:54	No Sample Provided
Breath Volume		0.000	22:54	
Air Blank		0.000	22:55	
Air Blank		0.000	22:57	
Subject Sample	# 2		22:58	Subject Test Refused
Breath Volume		0.000	22:58	
Air Blank		0.000	22:59	
Control Test		0.080	22:59	
Air Blank		0.000	23:00	
Diagnostics Check		OK	23:00	

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAERTZ,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/08/2025 02:10:17

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.081	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:05	No Sample Provided
Breath Volume		0.000	02:05	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 2		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.081	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAHLKE,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HOODENPYLE

DATE/TIME OF TEST: 11/09/2025 03:46:23

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.244	03:41
Breath Volume		2.441	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.237	03:44
Breath Volume		2.773	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HERNANDEZ,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/13/2025 00:21:49

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.082	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.000	00:17
Breath Volume		3.734	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.000	00:20
Breath Volume		4.066	00:20
Air Blank		0.000	00:20
Control Test		0.082	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HERNANDEZ',C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/13/2025 21:54:18

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:46	
Control Test		0.081	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 1		21:50	No Sample Provided
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Air Blank		0.000	21:52	
Subject Sample	# 2		21:52	Subject Test Refused
Breath Volume		0.000	21:52	
Air Blank		0.000	21:53	
Control Test		0.082	21:53	
Air Blank		0.000	21:54	
Diagnostics Check		OK	21:54	

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,HAILEY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAERTZ,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/14/2025 03:59:30

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.081	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:55	No Sample Provided
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 2		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Control Test		0.081	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/15/2025 05:15:34

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.081	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.126	05:10
Breath Volume		1.464	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.133	05:13
Breath Volume		1.722	05:13
Air Blank		0.000	05:14
Control Test		0.081	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 11/19/2025 23:25:14

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.081	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.000	23:20
Breath Volume		1.449	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.000	23:23
Breath Volume		1.687	23:23
Air Blank		0.000	23:24
Control Test		0.081	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BOLENA,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: TAKARA

DATE/TIME OF TEST: 11/19/2025 23:53:03

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.081	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1	0.278	23:45	Volume Not Met
Breath Volume		1.078	23:45	
Air Blank		0.000	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 2	0.296	23:48	
Breath Volume		1.785	23:48	
Air Blank		0.000	23:49	
Air Blank		0.000	23:50	
Subject Sample	# 3	0.294	23:51	
Breath Volume		2.132	23:51	
Air Blank		0.000	23:52	
Control Test		0.080	23:52	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 11/30/2025 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FERRANO,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: LEPLEY

DATE/TIME OF TEST: 11/20/2025 20:50:31

Diagnostics Check		OK	20:41	
Air Blank		0.000	20:42	
Control Test		0.081	20:42	
Air Blank		0.000	20:42	
Subject Sample	# 1		20:46	No Sample Provided
Breath Volume		0.000	20:46	
Air Blank		0.000	20:46	
Air Blank		0.000	20:48	
Subject Sample	# 2		20:49	Subject Test Refused
Breath Volume		0.000	20:49	
Air Blank		0.000	20:49	
Control Test		0.081	20:49	
Air Blank		0.000	20:50	
Diagnostics Check		OK	20:50	

Data Download Date/Time: 11/30/2025 11:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SMITH,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: NEESE

DATE/TIME OF TEST: 11/22/2025 02:11:45

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.203	02:06
Breath Volume		1.714	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.184	02:10
Breath Volume		1.988	02:10
Air Blank		0.000	02:10
Control Test		0.081	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 11/30/2025 11:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GARRISON,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 11/22/2025 03:16:16

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.081	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.153	03:11
Breath Volume		1.269	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.159	03:14
Breath Volume		2.089	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 11/30/2025 11:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,T,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAHLKE,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: WHEALEN

DATE/TIME OF TEST: 11/23/2025 03:38:22

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.121	03:33
Breath Volume		1.132	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.118	03:36
Breath Volume		1.378	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 11/30/2025 11:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HERNANDEZ,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 11/23/2025 05:09:55

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.081	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.139	05:04
Breath Volume		1.109	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.141	05:08
Breath Volume		1.753	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 11/30/2025 11:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/24/2025 05:45:57

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:37	
Control Test		0.081	05:38	
Air Blank		0.000	05:38	
Subject Sample	# 1		05:41	No Sample Provided
Breath Volume		0.000	05:41	
Air Blank		0.000	05:42	
Air Blank		0.000	05:44	
Subject Sample	# 2		05:44	Subject Test Refused
Breath Volume		0.000	05:44	
Air Blank		0.000	05:45	
Control Test		0.081	05:45	
Air Blank		0.000	05:45	
Diagnostics Check		OK	05:45	

Data Download Date/Time: 11/30/2025 11:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 11/01/2025 01:39:10

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.230	01:34
Breath Volume		1.324	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.237	01:37
Breath Volume		1.296	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 11/14/2025 14:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: JIMENEZ GARCIA

DATE/TIME OF TEST: 11/01/2025 13:31:52

Diagnostics Check		OK	13:25
Air Blank		0.000	13:25
Control Test		0.081	13:25
Air Blank		0.000	13:26
Subject Sample	# 1	0.227	13:27
Breath Volume		1.347	13:27
Air Blank		0.000	13:28
Air Blank		0.000	13:29
Subject Sample	# 2	0.232	13:30
Breath Volume		1.902	13:30
Air Blank		0.000	13:30
Control Test		0.080	13:31
Air Blank		0.000	13:31
Diagnostics Check		OK	13:31

Data Download Date/Time: 11/14/2025 14:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BESTERMAN

DATE/TIME OF TEST: 11/02/2025 01:28:46

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.082	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.162	01:23
Breath Volume		1.609	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:25
Subject Sample	# 2	0.164	01:27
Breath Volume		1.335	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 11/14/2025 14:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HECTOR,BRADY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: VALCEUS

DATE/TIME OF TEST: 11/02/2025 03:50:50

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.061	03:46
Breath Volume		1.375	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.060	03:49
Breath Volume		1.218	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 11/14/2025 14:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SALMERON

DATE/TIME OF TEST: 11/02/2025 04:47:02

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.081	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.103	04:38
Breath Volume		1.171	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.132	04:42
Breath Volume		2.746	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 3	0.134	04:45
Breath Volume		3.882	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:47

Data Download Date/Time: 11/14/2025 14:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RICCI,MATTHEW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 11/03/2025 01:07:21

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.143	01:01
Breath Volume		1.464	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:03
Subject Sample	# 2	0.147	01:05
Breath Volume		2.203	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 11/14/2025 14:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: MARICLE

DATE/TIME OF TEST: 11/05/2025 19:24:22

Diagnostics Check		OK	19:17
Air Blank		0.000	19:18
Control Test		0.081	19:18
Air Blank		0.000	19:19
Subject Sample	# 1	0.107	19:19
Breath Volume		1.683	19:19
Air Blank		0.000	19:20
Air Blank		0.000	19:22
Subject Sample	# 2	0.106	19:22
Breath Volume		1.375	19:22
Air Blank		0.000	19:23
Control Test		0.081	19:23
Air Blank		0.000	19:24
Diagnostics Check		OK	19:24

Data Download Date/Time: 11/14/2025 14:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 11/05/2025 20:47:20

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.081	20:41
Air Blank		0.000	20:42
Subject Sample	# 1	0.183	20:42
Breath Volume		2.515	20:42
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.186	20:45
Breath Volume		1.953	20:45
Air Blank		0.000	20:46
Control Test		0.080	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: KLINGENSTEIN

DATE/TIME OF TEST: 11/06/2025 03:06:24

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.146	03:01
Breath Volume		1.457	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.146	03:04
Breath Volume		2.542	03:04
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DITOMMASO,CHRISTIAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BOGUE

DATE/TIME OF TEST: 11/07/2025 03:02:15

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.200	02:57
Breath Volume		1.613	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.218	03:00
Breath Volume		1.683	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: DASS

DATE/TIME OF TEST: 11/08/2025 00:54:51

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.081	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.167	00:50
Breath Volume		2.195	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.162	00:53
Breath Volume		2.003	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: CHUNG

DATE/TIME OF TEST: 11/08/2025 19:50:27

Diagnostics Check		OK	19:43
Air Blank		0.000	19:44
Control Test		0.080	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.097	19:45
Breath Volume		3.230	19:45
Air Blank		0.000	19:46
Air Blank		0.000	19:48
Subject Sample	# 2	0.092	19:48
Breath Volume		2.335	19:48
Air Blank		0.000	19:49
Control Test		0.080	19:49
Air Blank		0.000	19:50
Diagnostics Check		OK	19:50

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMIGELSKI,TREVOR
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: MCCAULEY

DATE/TIME OF TEST: 11/09/2025 06:43:16

Diagnostics Check		OK	06:36
Air Blank		0.000	06:37
Control Test		0.080	06:37
Air Blank		0.000	06:37
Subject Sample	# 1	0.279	06:38
Breath Volume		2.269	06:38
Air Blank		0.000	06:39
Air Blank		0.000	06:41
Subject Sample	# 2	0.262	06:41
Breath Volume		2.082	06:41
Air Blank		0.000	06:42
Control Test		0.080	06:42
Air Blank		0.000	06:43
Diagnostics Check		OK	06:43

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MAMIN

DATE/TIME OF TEST: 11/10/2025 02:06:45

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.081	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.045	01:58	Volume Not Met
Breath Volume		0.812	01:58	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2	0.055	02:01	
Breath Volume		1.515	02:01	
Air Blank		0.000	02:02	
Air Blank		0.000	02:04	
Subject Sample	# 3	0.051	02:05	
Breath Volume		1.277	02:05	
Air Blank		0.000	02:05	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BUESO ROSALES

DATE/TIME OF TEST: 11/10/2025 04:47:15

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.166	04:42
Breath Volume		1.628	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.166	04:45
Breath Volume		1.808	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: DERILUS

DATE/TIME OF TEST: 11/11/2025 05:02:55

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.082	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.167	04:58
Breath Volume		2.132	04:58
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.164	05:01
Breath Volume		1.878	05:01
Air Blank		0.000	05:02
Control Test		0.080	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MASON,CHAD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 11/11/2025 05:39:25

Diagnostics Check		OK	05:26	
Air Blank		0.000	05:27	
Control Test		0.082	05:27	
Air Blank		0.000	05:27	
Subject Sample	# 1	0.073	05:31	Volume Not Met
Breath Volume		0.718	05:31	
Air Blank		0.000	05:31	
Air Blank		0.000	05:33	
Subject Sample	# 2	0.090	05:34	
Breath Volume		1.343	05:34	
Air Blank		0.000	05:34	
Air Blank		0.000	05:36	
Subject Sample	# 3	0.100	05:37	
Breath Volume		2.070	05:37	
Air Blank		0.000	05:38	
Control Test		0.081	05:38	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 11/14/2025 03:01:55

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.082	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.000	02:57
Breath Volume		2.781	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.000	03:00
Breath Volume		2.566	03:00
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 11/14/2025 14:11

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMIGELSKI,TREVOR
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: PELICO VICENTE

DATE/TIME OF TEST: 11/14/2025 23:15:11

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.081	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.133	23:09
Breath Volume		2.062	23:09
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.119	23:13
Breath Volume		1.671	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMIGELSKI,TREVOR
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: LIND

DATE/TIME OF TEST: 11/15/2025 05:25:21

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.081	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.211	05:20
Breath Volume		1.773	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.212	05:23
Breath Volume		1.953	05:23
Air Blank		0.000	05:24
Control Test		0.080	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/15/2025 19:14:53

Diagnostics Check		OK	19:08
Air Blank		0.000	19:08
Control Test		0.081	19:09
Air Blank		0.000	19:09
Subject Sample	# 1	0.140	19:10
Breath Volume		1.398	19:10
Air Blank		0.000	19:11
Air Blank		0.000	19:12
Subject Sample	# 2	0.137	19:13
Breath Volume		1.406	19:13
Air Blank		0.000	19:13
Control Test		0.080	19:14
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MORALES VASQUEZ

DATE/TIME OF TEST: 11/16/2025 02:37:26

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.220	02:32
Breath Volume		2.515	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:34
Subject Sample	# 2	0.215	02:35
Breath Volume		1.976	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THESSE,JACQUES
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: JIMENEZ PEREZ

DATE/TIME OF TEST: 11/16/2025 20:10:21

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.081	20:04
Air Blank		0.000	20:05
Subject Sample	# 1	0.226	20:05
Breath Volume		1.296	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.226	20:08
Breath Volume		1.257	20:08
Air Blank		0.000	20:09
Control Test		0.081	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 11/17/2025 05:32:44

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.080	05:25
Air Blank		0.000	05:25
Subject Sample	# 1	0.203	05:27
Breath Volume		3.394	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.189	05:31
Breath Volume		1.218	05:31
Air Blank		0.000	05:31
Control Test		0.079	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/17/2025 23:52:11

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.163	23:47
Breath Volume		2.714	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.164	23:50
Breath Volume		2.191	23:50
Air Blank		0.000	23:51
Control Test		0.081	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CARTER III

DATE/TIME OF TEST: 11/18/2025 02:36:35

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.210	02:31
Breath Volume		2.500	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:33
Subject Sample	# 2	0.216	02:34
Breath Volume		1.136	02:34
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: STANFORD

DATE/TIME OF TEST: 11/20/2025 19:48:39

Diagnostics Check		OK	19:42
Air Blank		0.000	19:42
Control Test		0.081	19:43
Air Blank		0.000	19:43
Subject Sample	# 1	0.221	19:44
Breath Volume		1.457	19:44
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.208	19:47
Breath Volume		1.347	19:47
Air Blank		0.000	19:47
Control Test		0.080	19:48
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 11/20/2025 22:44:00

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.081	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.145	22:39
Breath Volume		1.515	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:41
Subject Sample	# 2	0.155	22:42
Breath Volume		1.917	22:42
Air Blank		0.000	22:43
Control Test		0.080	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:44

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LNDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FRANCISCO ESTEBAN

DATE/TIME OF TEST: 11/20/2025 23:36:51

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.081	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.127	23:32
Breath Volume		1.410	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.130	23:35
Breath Volume		1.851	23:35
Air Blank		0.000	23:35
Control Test		0.081	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HUTCHINSON HUNT

DATE/TIME OF TEST: 11/21/2025 00:32:21

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.162	00:27
Breath Volume		1.746	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.167	00:30
Breath Volume		1.929	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/21/2025 02:04:16

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.086	01:59
Breath Volume		3.054	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.087	02:02
Breath Volume		1.718	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YEPEZ,GHERMAIME
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: VIOLA

DATE/TIME OF TEST: 11/21/2025 04:22:12

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.081	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.158	04:17
Breath Volume		1.683	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.155	04:20
Breath Volume		1.476	04:20
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: METZGER

DATE/TIME OF TEST: 11/21/2025 22:48:32

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.076	22:43
Breath Volume		1.296	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.076	22:47
Breath Volume		1.269	22:47
Air Blank		0.000	22:47
Control Test		0.081	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DALRYMPLE,KIMONNE
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 11/22/2025 00:31:40

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.081	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.266	00:27
Breath Volume		2.300	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.255	00:30
Breath Volume		2.605	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: OMBRES

DATE/TIME OF TEST: 11/22/2025 00:59:03

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.000	00:54
Breath Volume		1.921	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.000	00:57
Breath Volume		1.851	00:57
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:59

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AUGUST,ASHLEIGH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: HERNANDEZ HERRERA

DATE/TIME OF TEST: 11/22/2025 01:36:23

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.080	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.152	01:31
Breath Volume		1.976	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.134	01:34
Breath Volume		1.179	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBINSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: VELASQUEZ VERDE

DATE/TIME OF TEST: 11/22/2025 02:02:11

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.136	01:57
Breath Volume		2.824	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.127	02:00
Breath Volume		1.742	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JAMES,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: MENDEZ LOPEZ

DATE/TIME OF TEST: 11/22/2025 17:46:28

Diagnostics Check		OK	17:40
Air Blank		0.000	17:40
Control Test		0.081	17:40
Air Blank		0.000	17:41
Subject Sample	# 1	0.092	17:42
Breath Volume		1.906	17:42
Air Blank		0.000	17:42
Air Blank		0.000	17:44
Subject Sample	# 2	0.089	17:44
Breath Volume		3.648	17:44
Air Blank		0.000	17:45
Control Test		0.081	17:45
Air Blank		0.000	17:46
Diagnostics Check		OK	17:46

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: BITTNER

DATE/TIME OF TEST: 11/22/2025 23:17:46

Diagnostics Check		OK	23:11
Air Blank		0.000	23:12
Control Test		0.081	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.123	23:13
Breath Volume		2.675	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:15
Subject Sample	# 2	0.116	23:16
Breath Volume		3.078	23:16
Air Blank		0.000	23:16
Control Test		0.081	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FITTIPALDI,VINCENT
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ROBLERO GOMEZ

DATE/TIME OF TEST: 11/23/2025 01:18:41

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.247	01:13
Breath Volume		2.070	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.251	01:17
Breath Volume		2.957	01:17
Air Blank		0.000	01:17
Control Test		0.080	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ORTEGA,JUAM
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: AGUILAR SANCHEZ

DATE/TIME OF TEST: 11/23/2025 20:53:38

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.081	20:48
Air Blank		0.000	20:48
Subject Sample	# 1	0.194	20:49
Breath Volume		1.300	20:49
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.193	20:52
Breath Volume		1.441	20:52
Air Blank		0.000	20:52
Control Test		0.080	20:53
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: THEOPHILE

DATE/TIME OF TEST: 11/24/2025 05:26:26

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.238	05:21
Breath Volume		2.796	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.245	05:24
Breath Volume		2.652	05:24
Air Blank		0.000	05:25
Control Test		0.080	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: PEREZ CASTRO

DATE/TIME OF TEST: 11/24/2025 23:22:58

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.081	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.206	23:18
Breath Volume		2.335	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:20
Subject Sample	# 2	0.207	23:21
Breath Volume		2.941	23:21
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEFILIPPIS,MICHAEL
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GOODLETT

DATE/TIME OF TEST: 11/25/2025 20:44:32

Diagnostics Check		OK	20:37
Air Blank		0.000	20:37
Control Test		0.081	20:38
Air Blank		0.000	20:38
Subject Sample	# 1	0.041	20:39
Breath Volume		1.738	20:39
Air Blank		0.000	20:40
Air Blank		0.000	20:42
Subject Sample	# 2	0.040	20:43
Breath Volume		1.875	20:43
Air Blank		0.000	20:43
Control Test		0.081	20:43
Air Blank		0.000	20:44
Diagnostics Check		OK	20:44

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 11/26/2025 01:54:30

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.197	01:50
Breath Volume		2.027	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.194	01:52
Breath Volume		1.882	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ROBB

DATE/TIME OF TEST: 11/26/2025 02:21:38

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.127	02:16
Breath Volume		1.488	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.120	02:20
Breath Volume		1.113	02:20
Air Blank		0.000	02:20
Control Test		0.081	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ALONZO GERVACIO

DATE/TIME OF TEST: 11/27/2025 01:48:14

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.081	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:42	No Sample Provided
Breath Volume		0.000	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2		01:46	Subject Test Refused
Breath Volume		0.175	01:46	
Air Blank		0.000	01:47	
Control Test		0.082	01:47	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 11/27/2025 04:06:23

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.000	04:02
Breath Volume		3.750	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:05
Breath Volume		3.375	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: VIZUETH ROJO

DATE/TIME OF TEST: 11/27/2025 22:39:26

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.081	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.141	22:35
Breath Volume		1.753	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.144	22:37
Breath Volume		1.542	22:37
Air Blank		0.000	22:38
Control Test		0.080	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: NAZAIRE

DATE/TIME OF TEST: 11/28/2025 04:03:48

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.174	03:58
Breath Volume		1.191	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.171	04:02
Breath Volume		1.695	04:02
Air Blank		0.000	04:02
Control Test		0.082	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER,MASON
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 11/28/2025 22:30:05

Diagnostics Check		OK	22:23
Air Blank		0.000	22:24
Control Test		0.082	22:24
Air Blank		0.000	22:25
Subject Sample	# 1	0.250	22:25
Breath Volume		1.777	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.244	22:28
Breath Volume		2.191	22:28
Air Blank		0.000	22:29
Control Test		0.081	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACK,CHEYENNE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: LURIUS

DATE/TIME OF TEST: 11/29/2025 02:45:44

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.132	02:40
Breath Volume		1.476	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:42
Subject Sample	# 2	0.121	02:44
Breath Volume		1.476	02:44
Air Blank		0.000	02:44
Control Test		0.081	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 11/29/2025 05:46:32

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.081	05:41
Air Blank		0.000	05:41
Subject Sample	# 1	0.155	05:42
Breath Volume		1.675	05:42
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.160	05:44
Breath Volume		2.476	05:44
Air Blank		0.000	05:45
Control Test		0.082	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 11/29/2025 23:28:47

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.082	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.226	23:24
Breath Volume		1.167	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.218	23:27
Breath Volume		1.117	23:27
Air Blank		0.000	23:27
Control Test		0.080	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GONZALES LOPEZ

DATE/TIME OF TEST: 11/30/2025 00:43:06

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.081	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.249	00:38
Breath Volume		1.363	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:40
Subject Sample	# 2	0.249	00:41
Breath Volume		1.656	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TUANG,HGIN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LAFAYE

DATE/TIME OF TEST: 11/30/2025 05:32:23

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:21	
Control Test		0.081	05:21	
Air Blank		0.000	05:22	
Subject Sample	# 1	0.000	05:25	Volume Not Met
Breath Volume		0.039	05:25	
Air Blank		0.000	05:25	
Air Blank		0.000	05:27	
Subject Sample	# 2	0.131	05:30	Volume Not Met
Breath Volume		0.019	05:30	
Air Blank		0.000	05:31	
Control Test		0.080	05:31	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 11/07/2025 22:00:56

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.078	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.165	21:56
Breath Volume		2.550	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:58
Subject Sample	# 2	0.165	21:59
Breath Volume		3.199	21:59
Air Blank		0.000	21:59
Control Test		0.078	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 11/09/2025 00:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CARATACHEA

DATE/TIME OF TEST: 11/08/2025 00:57:20

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.097	00:52
Breath Volume		2.453	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.095	00:55
Breath Volume		2.277	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 11/09/2025 00:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: GREWAL

DATE/TIME OF TEST: 11/15/2025 03:52:49

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.121	03:48
Breath Volume		2.218	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.125	03:51
Breath Volume		1.734	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 12/20/2025 02:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: SAVINSKY

DATE/TIME OF TEST: 11/16/2025 02:35:10

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.072	02:30
Breath Volume		1.867	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.082	02:33
Breath Volume		1.355	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 12/20/2025 02:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/23/2025 01:19:00

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.000	01:14
Breath Volume		3.066	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.000	01:17
Breath Volume		2.457	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:19

Data Download Date/Time: 12/20/2025 02:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/27/2025 04:48:28

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.135	04:43
Breath Volume		3.414	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.134	04:46
Breath Volume		3.238	04:46
Air Blank		0.000	04:47
Control Test		0.078	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 12/20/2025 02:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 11/01/2025 01:55:47

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.105	01:50
Breath Volume		1.230	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.107	01:54
Breath Volume		1.859	01:54
Air Blank		0.000	01:54
Control Test		0.076	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 11/20/2025 18:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAUENHOFER,TIMOTHY,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FRAUENHOFER,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 11/01/2025 18:07:04

Diagnostics Check	OK	18:06	
Air Blank		18:06	Ambient Fail
Air Blank	0.000	18:07	

Data Download Date/Time: 11/20/2025 18:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAUENHOFER,TIMOTHY,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FRAUENHOFER,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 11/01/2025 18:15:59

Diagnostics Check		OK	18:09
Air Blank		0.000	18:09
Control Test		0.075	18:09
Air Blank		0.000	18:10
Subject Sample	# 1	0.198	18:11
Breath Volume		2.425	18:11
Air Blank		0.000	18:12
Air Blank		0.000	18:13
Subject Sample	# 2	0.206	18:14
Breath Volume		2.675	18:14
Air Blank		0.000	18:15
Control Test		0.075	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 11/20/2025 18:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/14/2025 23:42:29

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:36	
Control Test		0.075	23:37	
Air Blank		0.000	23:37	
Subject Sample	# 1	0.195	23:38	
Breath Volume		2.644	23:38	
Air Blank		0.000	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 2		23:41	RFI Detect
Breath Volume		0.000	23:41	
Air Blank		0.000	23:41	
Control Test		0.078	23:41	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 11/20/2025 18:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/14/2025 23:50:44

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.075	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.197	23:46
Breath Volume		2.769	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.192	23:49
Breath Volume		2.851	23:49
Air Blank		0.000	23:49
Control Test		0.075	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 11/20/2025 18:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: CERVANTES LOPEZ

DATE/TIME OF TEST: 11/29/2025 18:44:40

Diagnostics Check		OK	18:37
Air Blank		0.000	18:38
Control Test		0.076	18:38
Air Blank		0.000	18:39
Subject Sample	# 1	0.294	18:40
Breath Volume		1.652	18:40
Air Blank		0.000	18:41
Air Blank		0.000	18:42
Subject Sample	# 2	0.284	18:43
Breath Volume		1.656	18:43
Air Blank		0.000	18:43
Control Test		0.075	18:44
Air Blank		0.000	18:44
Diagnostics Check		OK	18:44

Data Download Date/Time: 12/27/2025 18:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BARBAROSSO

DATE/TIME OF TEST: 11/29/2025 23:35:02

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.075	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.237	23:30
Breath Volume		3.609	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:32
Subject Sample	# 2	0.238	23:33
Breath Volume		3.179	23:33
Air Blank		0.000	23:34
Control Test		0.078	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:35

Data Download Date/Time: 12/27/2025 18:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: AMELINCKX LABRADOR

DATE/TIME OF TEST: 11/30/2025 21:59:56

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:53	
Control Test		0.075	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.253	21:55	
Breath Volume		1.742	21:55	
Air Blank		0.000	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 2	0.258	21:58	
Breath Volume		2.421	21:58	
Air Blank		0.000	21:59	
Control Test		0.074	21:59	Control Outside Tolerance
Air Blank		0.000	21:59	

Data Download Date/Time: 12/27/2025 18:08

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: AMELINCKX LABRADOR

DATE/TIME OF TEST: 11/30/2025 22:09:49

Diagnostics Check		OK	22:02	
Air Blank		0.000	22:02	
Control Test		0.075	22:03	
Air Blank		0.000	22:03	
Subject Sample	# 1	0.224	22:04	
Breath Volume		2.113	22:04	
Air Blank		0.000	22:05	
Air Blank		0.000	22:06	
Subject Sample	# 2	0.232	22:08	
Breath Volume		2.949	22:08	
Air Blank		0.000	22:08	
Control Test		0.074	22:09	Control Outside Tolerance
Air Blank		0.000	22:09	

Data Download Date/Time: 12/27/2025 18:08

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:58

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 11/25/2025 08:25:12

Diagnostics Check		OK	08:19
Air Blank		0.000	08:19
Control Test		0.082	08:19
Air Blank		0.000	08:20
Subject Sample	# 1	0.000	08:21
Breath Volume		1.671	08:21
Air Blank		0.000	08:21
Air Blank		0.000	08:23
Subject Sample	# 2	0.000	08:23
Breath Volume		1.398	08:23
Air Blank		0.000	08:24
Control Test		0.082	08:24
Air Blank		0.000	08:25
Diagnostics Check		OK	08:25

Data Download Date/Time: 12/22/2025 11:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/02/2025 01:15:17

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.163	01:09
Breath Volume		1.707	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.150	01:13
Breath Volume		1.507	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 11/24/2025 11:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER,MONICA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER,MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: GARCIA QUINTANILLA

DATE/TIME OF TEST: 11/16/2025 05:22:55

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.079	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.076	05:18
Breath Volume		1.812	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:20
Subject Sample	# 2	0.073	05:21
Breath Volume		1.582	05:21
Air Blank		0.000	05:22
Control Test		0.078	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 11/24/2025 11:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER,MONICA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER,MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 11/19/2025 01:18:22

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.079	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.182	01:05	
Breath Volume		1.378	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 2	0.158	01:10	Volume Not Met
Breath Volume		0.777	01:10	
Air Blank		0.000	01:11	
Air Blank		0.000	01:13	
Subject Sample	# 3	0.178	01:16	
Breath Volume		1.437	01:16	
Air Blank		0.000	01:17	
Control Test		0.076	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 11/24/2025 11:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDD,RACHEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUDD,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 11/29/2025 01:55:19

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.204	01:50
Breath Volume		2.175	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.214	01:53
Breath Volume		1.105	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 11/29/2025 04:38:12

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.000	04:33
Breath Volume		2.093	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.000	04:36
Breath Volume		2.761	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 11/30/2025 00:11:39

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.030	00:07
Breath Volume		4.472	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.030	00:10
Breath Volume		4.398	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 11/30/2025 02:39:36

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.164	02:34
Breath Volume		3.000	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.172	02:38
Breath Volume		3.589	02:38
Air Blank		0.000	02:38
Control Test		0.078	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 11/30/2025 05:31:14

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.079	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.090	05:25
Breath Volume		1.246	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.092	05:29
Breath Volume		2.238	05:29
Air Blank		0.000	05:30
Control Test		0.078	05:30
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTER,DYLAN,F
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: FLAMING

DATE/TIME OF TEST: 11/30/2025 19:33:41

Diagnostics Check		OK	19:27
Air Blank		0.000	19:27
Control Test		0.079	19:27
Air Blank		0.000	19:28
Subject Sample	# 1	0.146	19:29
Breath Volume		2.648	19:29
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.149	19:32
Breath Volume		2.574	19:32
Air Blank		0.000	19:32
Control Test		0.077	19:33
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAVO,GEORGE,A
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: BRAVO,GEORGE
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: MANCHENO TORO

DATE/TIME OF TEST: 11/02/2025 06:42:10

Diagnostics Check		OK	06:35
Air Blank		0.000	06:36
Control Test		0.081	06:36
Air Blank		0.000	06:37
Subject Sample	# 1	0.173	06:37
Breath Volume		2.757	06:37
Air Blank		0.000	06:38
Air Blank		0.000	06:40
Subject Sample	# 2	0.170	06:40
Breath Volume		2.945	06:40
Air Blank		0.000	06:41
Control Test		0.082	06:41
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 12/08/2025 20:39

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUENTES,BIANCA,M
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: REDWAN,NADER
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 10/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TORREZ

DATE/TIME OF TEST: 11/16/2025 00:30:28

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.082	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.211	00:25
Breath Volume		2.527	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.221	00:28
Breath Volume		2.535	00:28
Air Blank		0.000	00:29
Control Test		0.081	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 12/08/2025 20:39

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: KUNILOVSKAIA

DATE/TIME OF TEST: 11/02/2025 05:26:10

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.077	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.167	05:20
Breath Volume		1.292	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.156	05:24
Breath Volume		1.519	05:24
Air Blank		0.000	05:25
Control Test		0.077	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 11/06/2025 04:04

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.