

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PRUNEDA,ANDREA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: RAYMUNDO MATOM

DATE/TIME OF TEST: 02/05/2026 18:28:49

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.082	18:22
Air Blank	0.000	18:22
Subject Sample # 1	0.248	18:23
Breath Volume	1.625	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.251	18:27
Breath Volume	1.828	18:27
Air Blank	0.000	18:27
Control Test	0.080	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 02/16/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOQUE, OSCAR,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: SUAREZ, LOGAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MATOM MARCOS

DATE/TIME OF TEST: 02/15/2026 21:37:16

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.202	21:32
Breath Volume	2.328	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.197	21:35
Breath Volume	2.218	21:35
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 02/16/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 02/16/2026 04:02:21

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.083	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.243	03:57
Breath Volume		1.437	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.237	04:00
Breath Volume		2.304	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 02/16/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRBY,DOMINIC,R
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: SUAREZ,LOGAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MENDEZ PEREZ

DATE/TIME OF TEST: 02/21/2026 01:20:25

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.080	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:12	No Sample Provided
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Air Blank		0.000	01:14	
Subject Sample	# 2	0.231	01:15	
Breath Volume		1.882	01:15	
Air Blank		0.000	01:16	
Air Blank		0.000	01:18	
Subject Sample	# 3	0.228	01:18	
Breath Volume		1.273	01:18	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 03/23/2026 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRBY,DOMINIC,R
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: IRBY,DOMINIC
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: PIERRE PAUL

DATE/TIME OF TEST: 02/23/2026 01:38:50

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.082	01:27	
Air Blank		0.000	01:28	
Subject Sample	# 1	0.248	01:31	Volume Not Met
Breath Volume		0.085	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 2	0.234	01:37	Volume Not Met
Breath Volume		0.152	01:37	
Air Blank		0.000	01:37	
Control Test		0.078	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 03/23/2026 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRBY,DOMINIC,R
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: IRBY,DOMINIC
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: PIERRE PAUL

DATE/TIME OF TEST: 02/23/2026 01:43:33

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.078	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.078	01:42	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 03/23/2026 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE, OSCAR,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROQUE, OSCAR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: MEJIA SANTAY

DATE/TIME OF TEST: 02/23/2026 02:05:51

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.147	02:01
Breath Volume	3.546	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.147	02:04
Breath Volume	3.382	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 03/23/2026 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBERNARDINO,BRANDON,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIRA,BRYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 02/18/2026 22:47:45

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.083	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.000	22:42
Breath Volume	1.386	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.000	22:46
Breath Volume	1.574	22:46
Air Blank	0.000	22:46
Control Test	0.082	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 03/06/2026 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FARRAJ,S
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: LEITZEL

DATE/TIME OF TEST: 02/25/2026 21:00:33

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.085	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.164	20:55
Breath Volume	3.824	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.163	20:58
Breath Volume	3.753	20:58
Air Blank	0.000	20:59
Control Test	0.083	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 03/06/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BOLENA, B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SEYBOLT

DATE/TIME OF TEST: 02/06/2026 23:49:12

Diagnostics Check		OK	23:42
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.254	23:44
Breath Volume		4.089	23:44
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.244	23:47
Breath Volume		4.007	23:47
Air Blank		0.000	23:48
Control Test		0.080	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 04/30/2026 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: ORTIZ ROLON

DATE/TIME OF TEST: 02/01/2026 01:08:29

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:59	
Control Test		0.080	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1	0.139	01:03	Volume Not Met
Breath Volume		0.160	01:03	
Air Blank		0.000	01:03	
Air Blank		0.000	01:05	
Subject Sample	# 2		01:06	Subject Test Refused
Breath Volume		0.183	01:06	
Air Blank		0.000	01:07	
Control Test		0.078	01:07	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMIGELSKI, TREVOR
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: KIMANI

DATE/TIME OF TEST: 02/03/2026 22:47:43

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:32	
Control Test		0.079	22:32	
Air Blank		0.000	22:33	
Subject Sample	# 1	0.290	22:36	Volume Not Met
Breath Volume		0.964	22:36	
Air Blank		0.000	22:37	
Air Blank		0.000	22:39	
Subject Sample	# 2	0.279	22:40	
Breath Volume		1.417	22:40	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 3	0.285	22:45	
Breath Volume		2.320	22:45	
Air Blank		0.000	22:46	
Control Test		0.075	22:47	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COELHO,GABRIEL
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:06

SUBJECT'S LAST NAME: DEMESYEUX

DATE/TIME OF TEST: 02/04/2026 11:41:04

Diagnostics Check	OK	11:34
Air Blank	0.000	11:34
Control Test	0.078	11:35
Air Blank	0.000	11:35
Subject Sample # 1	0.000	11:36
Breath Volume	2.214	11:36
Air Blank	0.000	11:37
Air Blank	0.000	11:39
Subject Sample # 2	0.000	11:39
Breath Volume	2.011	11:39
Air Blank	0.000	11:40
Control Test	0.079	11:40
Air Blank	0.000	11:40
Diagnostics Check	OK	11:41

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMPSON,CHASE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: HILDRETH

DATE/TIME OF TEST: 02/04/2026 22:48:47

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.079	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.000	22:44
Breath Volume	1.527	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.000	22:47
Breath Volume	1.285	22:47
Air Blank	0.000	22:47
Control Test	0.078	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BRONNER

DATE/TIME OF TEST: 02/05/2026 02:24:23

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.078	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.041	02:19
Breath Volume	2.015	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.039	02:22
Breath Volume	2.496	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/05/2026 21:00:58

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.199	20:56
Breath Volume	1.574	20:56
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.192	20:59
Breath Volume	1.820	20:59
Air Blank	0.000	21:00
Control Test	0.077	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS, VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PELLETIER

DATE/TIME OF TEST: 02/06/2026 01:09:22

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.061	01:04
Breath Volume		2.312	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.058	01:07
Breath Volume		2.683	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 02/06/2026 23:48:08

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.282	23:43
Breath Volume	1.746	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.283	23:46
Breath Volume	1.632	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/07/2026 02:22:48

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.161	02:18
Breath Volume	1.339	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.166	02:21
Breath Volume	1.558	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 02/07/2026 07:35:32

Diagnostics Check		OK	07:29
Air Blank		0.000	07:29
Control Test		0.079	07:29
Air Blank		0.000	07:30
Subject Sample	# 1	0.174	07:31
Breath Volume		2.277	07:31
Air Blank		0.000	07:31
Air Blank		0.000	07:33
Subject Sample	# 2	0.173	07:33
Breath Volume		2.382	07:33
Air Blank		0.000	07:34
Control Test		0.078	07:34
Air Blank		0.000	07:35
Diagnostics Check		OK	07:35

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: ESTIVERNE

DATE/TIME OF TEST: 02/08/2026 02:23:28

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.126	02:19
Breath Volume	1.453	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.120	02:21
Breath Volume	1.507	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 02/08/2026 02:54:15

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.094	02:48
Breath Volume	1.199	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.092	02:52
Breath Volume	1.406	02:52
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BATES,IVAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CHIVALAN HERNANDEZ

DATE/TIME OF TEST: 02/08/2026 03:15:26

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.078	03:09	
Air Blank	0.000	03:09	
Subject Sample # 1	0.197	03:10	
Breath Volume	2.949	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample # 2	0.188	03:13	
Breath Volume	2.328	03:13	
Air Blank	0.000	03:14	
Control Test	0.078	03:14	Interferent Detect
Air Blank	0.000	03:15	

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TIA, L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/09/2026 02:43:54

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.000	02:39
Breath Volume	1.386	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:42
Breath Volume	1.246	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/10/2026 22:40:52

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.078	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.227	22:35
Breath Volume	1.796	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.223	22:39
Breath Volume	1.507	22:39
Air Blank	0.000	22:39
Control Test	0.075	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 02/13/2026 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASAZZA, KIETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: MERA

DATE/TIME OF TEST: 02/11/2026 05:07:45

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.078	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:03	No Sample Provided
Breath Volume	0.000	05:03	
Air Blank	0.000	05:03	
Air Blank	0.000	05:05	
Subject Sample	# 2	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.078	05:07	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 02/13/2026 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 02/11/2026 22:09:57

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.078	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.148	22:05
Breath Volume	1.351	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.139	22:08
Breath Volume	1.707	22:08
Air Blank	0.000	22:09
Control Test	0.078	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 02/13/2026 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/14/2026 05:58:19

Diagnostics Check	OK	05:50
Air Blank	0.000	05:51
Control Test	0.079	05:51
Air Blank	0.000	05:52
Subject Sample # 1	0.171	05:52
Breath Volume	1.210	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.174	05:56
Breath Volume	1.437	05:56
Air Blank	0.000	05:57
Control Test	0.079	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,ARIEL
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MAWN

DATE/TIME OF TEST: 02/15/2026 00:39:57

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.066	00:35
Breath Volume	2.367	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.061	00:38
Breath Volume	2.574	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SOLORZANO BASTIAN

DATE/TIME OF TEST: 02/15/2026 02:33:46

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.155	02:29
Breath Volume	1.105	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.160	02:32
Breath Volume	1.339	02:32
Air Blank	0.000	02:32
Control Test	0.078	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,PAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: LOZANO PERDOMO

DATE/TIME OF TEST: 02/15/2026 07:47:30

Diagnostics Check	OK	07:40
Air Blank	0.000	07:40
Control Test	0.078	07:40
Air Blank	0.000	07:41
Subject Sample # 1	0.163	07:42
Breath Volume	1.394	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample # 2	0.175	07:45
Breath Volume	2.050	07:45
Air Blank	0.000	07:46
Control Test	0.077	07:46
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,ARIEL
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SWANSKI

DATE/TIME OF TEST: 02/19/2026 03:40:14

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.137	03:35
Breath Volume	3.292	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.128	03:38
Breath Volume	1.835	03:38
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS, KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BARRIOS VELASQUEZ

DATE/TIME OF TEST: 02/19/2026 04:36:31

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.248	04:31
Breath Volume	1.457	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.254	04:34
Breath Volume	1.394	04:34
Air Blank	0.000	04:35
Control Test	0.078	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RIN,LAURENCE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 02/19/2026 17:51:29

Diagnostics Check	OK	17:44
Air Blank	0.000	17:44
Control Test	0.079	17:45
Air Blank	0.000	17:45
Subject Sample # 1	0.108	17:46
Breath Volume	1.296	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:48
Subject Sample # 2	0.105	17:49
Breath Volume	1.562	17:49
Air Blank	0.000	17:50
Control Test	0.077	17:50
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: PEREZ VALASQUEZ

DATE/TIME OF TEST: 02/21/2026 03:18:40

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.150	03:13
Breath Volume	1.183	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.147	03:17
Breath Volume	1.621	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENYMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 02/21/2026 20:53:34

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.079	20:48
Air Blank		0.000	20:48
Subject Sample	# 1	0.226	20:49
Breath Volume		2.421	20:49
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.217	20:52
Breath Volume		2.761	20:52
Air Blank		0.000	20:52
Control Test		0.078	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: BACSO

DATE/TIME OF TEST: 02/22/2026 01:40:02

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.154	01:35
Breath Volume		2.445	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.155	01:38
Breath Volume		1.726	01:38
Air Blank		0.000	01:39
Control Test		0.078	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:40

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BLEWETT

DATE/TIME OF TEST: 02/22/2026 03:22:38

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.012	03:18
Breath Volume		1.906	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.012	03:21
Breath Volume		1.648	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: FELDER

DATE/TIME OF TEST: 02/24/2026 06:18:31

Diagnostics Check	OK	06:09
Air Blank	0.000	06:09
Control Test	0.079	06:10
Air Blank	0.000	06:10
Subject Sample # 1	0.148	06:11
Breath Volume	1.562	06:11
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.140	06:16
Breath Volume	1.386	06:16
Air Blank	0.000	06:17
Control Test	0.079	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN, JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 02/24/2026 22:50:16

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.207	22:45
Breath Volume	2.988	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.201	22:48
Breath Volume	3.058	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: OBOY

DATE/TIME OF TEST: 02/27/2026 04:46:22

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.078	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:44	Interferent Detect
Breath Volume	1.644	04:44	
Air Blank	0.000	04:45	
Control Test	0.078	04:45	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: OBOY

DATE/TIME OF TEST: 02/27/2026 05:14:56

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.077	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:13	Interferent Detect
Breath Volume	2.066	05:13	
Air Blank	0.000	05:14	
Control Test	0.078	05:14	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

Data Download Date/Time: 03/13/2026 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEMEN,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 02/07/2026 03:13:34

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.150	03:08
Breath Volume		2.421	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.150	03:11
Breath Volume		1.382	03:11
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 02/19/2026 01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BEJARANO

DATE/TIME OF TEST: 02/08/2026 01:08:26

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.116	01:03
Breath Volume		2.683	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.112	01:06
Breath Volume		3.421	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/19/2026 01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BURNETTE

DATE/TIME OF TEST: 02/08/2026 02:43:39

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.124	02:39
Breath Volume	1.917	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.115	02:42
Breath Volume	2.128	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 02/19/2026 01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CHOC CHOC

DATE/TIME OF TEST: 02/08/2026 03:21:17

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.081	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.124	03:16
Breath Volume	1.558	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 2	0.134	03:19
Breath Volume	1.574	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 02/19/2026 01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,COLBY,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SEARS,COLBY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/21/2026 03:33:06

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.075	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1 0.145	03:27	
Breath Volume	1.824	03:27	
Air Blank	0.000	03:28	
Air Blank	0.000	03:30	
Subject Sample	# 2 0.139	03:31	
Breath Volume	1.492	03:31	
Air Blank	0.000	03:32	
Control Test	0.072	03:32	Control Outside Tolerance
Air Blank	0.000	03:33	

Data Download Date/Time: 03/06/2026 18:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,COLBY,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SEARS,COLBY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/21/2026 03:35:35

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.074	03:34	Control Outside Tolerance
Air Blank	0.000	03:35	

Data Download Date/Time: 03/06/2026 18:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS, COLBY, D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SEARS, COLBY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/21/2026 03:37:49

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.074	03:37	Control Outside Tolerance
Air Blank	0.000	03:37	

Data Download Date/Time: 03/06/2026 18:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/04/2026 02:10:47

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.083	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.131	02:05
Breath Volume	2.160	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.130	02:09
Breath Volume	2.292	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 02/27/2026 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 02/06/2026 23:28:07

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.082	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.255	23:23
Breath Volume	1.457	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.238	23:26
Breath Volume	2.628	23:26
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 02/27/2026 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RAINGE,MARIAH
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MCDERMOTT

DATE/TIME OF TEST: 02/07/2026 03:54:02

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.082	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.231	03:49
Breath Volume		1.765	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.228	03:52
Breath Volume		1.531	03:52
Air Blank		0.000	03:53
Control Test		0.081	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:54

Data Download Date/Time: 02/27/2026 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LOVELESS,PAYTON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: CHAPIN

DATE/TIME OF TEST: 02/27/2026 05:54:02

Diagnostics Check		OK	05:47
Air Blank		0.000	05:47
Control Test		0.081	05:48
Air Blank		0.000	05:48
Subject Sample	# 1	0.193	05:49
Breath Volume		1.820	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.193	05:52
Breath Volume		1.175	05:52
Air Blank		0.000	05:53
Control Test		0.080	05:53
Air Blank		0.000	05:53
Diagnostics Check		OK	05:54

Data Download Date/Time: 02/27/2026 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLVERA, JESUS,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HARTWELL, DEVON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: GORE

DATE/TIME OF TEST: 02/12/2026 12:20:39

Diagnostics Check	OK	12:14
Air Blank	0.000	12:15
Control Test	0.078	12:15
Air Blank	0.000	12:15
Subject Sample # 1	0.000	12:16
Breath Volume	2.835	12:16
Air Blank	0.000	12:16
Air Blank	0.000	12:18
Subject Sample # 2	0.000	12:19
Breath Volume	2.140	12:19
Air Blank	0.000	12:19
Control Test	0.076	12:19
Air Blank	0.000	12:20
Diagnostics Check	OK	12:20

Data Download Date/Time: 02/13/2026 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PINKNEY

DATE/TIME OF TEST: 02/14/2026 23:35:25

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	RFI Detect
Air Blank		23:35	RFI Detect

Data Download Date/Time: 03/17/2026 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PINKNEY

DATE/TIME OF TEST: 02/14/2026 23:43:12

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.085	23:38
Breath Volume	1.808	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.085	23:41
Breath Volume	1.273	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 03/17/2026 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEE,BRENNON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: LEONARDO RAMIREZ

DATE/TIME OF TEST: 02/15/2026 16:34:23

Diagnostics Check	OK	16:29	
Air Blank	0.000	16:29	
Control Test	0.081	16:29	
Air Blank	0.000	16:30	
Subject Sample	# 1	16:32	Subject Test Refused
Breath Volume	0.253	16:32	
Air Blank	0.000	16:33	
Control Test	0.081	16:33	
Air Blank	0.000	16:34	
Diagnostics Check	OK	16:34	

Data Download Date/Time: 03/17/2026 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORCE,TAYLOR,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORCE,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: KUHN

DATE/TIME OF TEST: 02/21/2026 17:39:54

Diagnostics Check	OK	17:39	
Air Blank		17:39	RFI Detect
Air Blank		17:39	RFI Detect

Data Download Date/Time: 03/17/2026 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORCE,TAYLOR,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORCE,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: KUHN

DATE/TIME OF TEST: 02/21/2026 17:50:37

Diagnostics Check	OK	17:43
Air Blank	0.000	17:44
Control Test	0.080	17:44
Air Blank	0.000	17:45
Subject Sample # 1	0.109	17:45
Breath Volume	1.789	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:48
Subject Sample # 2	0.105	17:49
Breath Volume	1.609	17:49
Air Blank	0.000	17:49
Control Test	0.080	17:50
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 03/17/2026 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FREEMAN,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 02/28/2026 05:46:21

Diagnostics Check	OK	05:39
Air Blank	0.000	05:39
Control Test	0.080	05:40
Air Blank	0.000	05:40
Subject Sample # 1	0.137	05:41
Breath Volume	1.261	05:41
Air Blank	0.000	05:42
Air Blank	0.000	05:43
Subject Sample # 2	0.133	05:44
Breath Volume	3.320	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 03/17/2026 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUQUE,DANIELA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,MELANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80006456
DRY GAS EXPIRY DATE: 12/18/2026
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 02/21/2026 06:20:30

Diagnostics Check	OK	06:12	
Air Blank	0.000	06:13	
Control Test	0.078	06:13	
Air Blank	0.000	06:14	
Subject Sample	# 1 0.150	06:14	
Breath Volume	2.429	06:14	
Air Blank	0.000	06:15	
Air Blank	0.000	06:17	
Subject Sample	# 2 0.147	06:18	
Breath Volume	2.347	06:18	
Air Blank	0.000	06:19	
Control Test	0.072	06:19	Control Outside Tolerance
Air Blank	0.000	06:20	

Data Download Date/Time: 03/23/2026 18:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, WILLIAM,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ, DAMIAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: TORO MARQUEZ

DATE/TIME OF TEST: 02/28/2026 05:56:47

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.078	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.154	05:49
Breath Volume	2.257	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.129	05:51
Breath Volume	1.347	05:51
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 3	0.149	05:55
Breath Volume	2.800	05:55
Air Blank	0.000	05:55
Control Test	0.076	05:56
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 03/23/2026 18:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,MATTHEW
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BALDONADO

DATE/TIME OF TEST: 02/23/2026 03:54:11

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 03/12/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEREIRA,SALISHA,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FEREIRA,SALISHA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 02/23/2026 20:09:56

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.080	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.174	20:05
Breath Volume	3.078	20:05
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.159	20:08
Breath Volume	3.296	20:08
Air Blank	0.000	20:09
Control Test	0.079	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 03/12/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERREIRA,SALISHA,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TAPANES,LAURA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GONZALEZ-LOPEZ

DATE/TIME OF TEST: 02/24/2026 02:50:03

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.244	02:44
Breath Volume		1.980	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:46
Subject Sample	# 2	0.235	02:48
Breath Volume		2.175	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:50

Data Download Date/Time: 03/12/2026 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/26/2026 00:21:24

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.000	00:17
Breath Volume	2.882	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.000	00:20
Breath Volume	2.664	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 03/12/2026 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 02/28/2026 03:03:03

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.079	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.231	02:58
Breath Volume		1.687	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.233	03:01
Breath Volume		2.132	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:03

Data Download Date/Time: 03/12/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: BOLEY

DATE/TIME OF TEST: 02/28/2026 18:43:29

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:31	
Control Test	0.079	18:31	
Air Blank	0.000	18:32	
Subject Sample	# 1	18:35	No Sample Provided
Breath Volume	0.000	18:35	
Air Blank	0.000	18:35	
Air Blank	0.000	18:37	
Subject Sample	# 2	0.111	18:38
Breath Volume	3.808	18:38	
Air Blank	0.000	18:39	
Air Blank	0.000	18:41	
Subject Sample	# 3	0.117	18:41
Breath Volume	4.113	18:41	
Air Blank	0.000	18:42	
Control Test	0.079	18:42	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 03/12/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND-SHIRAH,MIKAELA,D
BREATH TEST OPERATOR AGENCY: Bonifay PD
ARRESTING OFFICER: STRICKLAND-SHIRAH,MIKAELA
ARRESTING OFFICER AGENCY: Bonifay PD
LAST AGENCY INSPECTION: 02/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: HAGANS

DATE/TIME OF TEST: 02/27/2026 16:02:49

Diagnostics Check	OK	15:59	
Air Blank	0.000	15:59	
Control Test	0.079	15:59	
Air Blank	0.000	16:00	
Subject Sample	# 1	16:02	Sequence Aborted
Breath Volume	0.000	16:02	
Air Blank	0.000	16:02	

Data Download Date/Time: 03/12/2026 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDWAN,NADER,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: REDWAN,NADER
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 01/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 02/01/2026 05:14:17

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.082	05:12	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:12	Subject Test Refused
Breath Volume	0.000	05:12	
Air Blank	0.000	05:13	
Control Test	0.082	05:13	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

Data Download Date/Time: 02/13/2026 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIGA,JUAN,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: MCEACHIN,SCOTT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: BELAUSTEGUI

DATE/TIME OF TEST: 02/01/2026 16:41:57

Diagnostics Check	OK	16:39	
Air Blank	0.000	16:39	
Control Test	0.085	16:39	
Air Blank	0.000	16:40	
Subject Sample	# 1	16:40	Subject Test Refused
Breath Volume	0.000	16:40	
Air Blank	0.000	16:41	
Control Test	0.083	16:41	
Air Blank	0.000	16:41	
Diagnostics Check	OK	16:41	

Data Download Date/Time: 02/13/2026 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIL,ALEJANDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GIL,ALEJANDRO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: COLMENARES

DATE/TIME OF TEST: 02/04/2026 19:06:59

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:04	
Control Test	0.081	19:04	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:05	Subject Test Refused
Breath Volume	0.000	19:05	
Air Blank	0.000	19:06	
Control Test	0.082	19:06	
Air Blank	0.000	19:06	
Diagnostics Check	OK	19:06	

Data Download Date/Time: 02/13/2026 15:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 02/22/2026 20:46:27

Diagnostics Check	OK	20:42	
Air Blank	0.000	20:42	
Control Test	0.081	20:42	
Air Blank	0.000	20:43	
Subject Sample	# 1	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Control Test	0.082	20:45	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 03/06/2026 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIGA,JUAN,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: POSSE,ARMANDO
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 02/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/25/2026 15:02:57

Diagnostics Check	OK	14:54
Air Blank	0.000	14:55
Control Test	0.082	14:55
Air Blank	0.000	14:55
Subject Sample # 1	0.096	14:56
Breath Volume	1.945	14:56
Air Blank	0.000	14:57
Air Blank	0.000	14:59
Subject Sample # 2	0.086	15:01
Breath Volume	1.597	15:01
Air Blank	0.000	15:02
Control Test	0.082	15:02
Air Blank	0.000	15:02
Diagnostics Check	OK	15:02

Data Download Date/Time: 03/06/2026 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GRAZIOSE,MADISON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 02/08/2026 20:37:52

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.078	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.157	20:33
Breath Volume	2.101	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.152	20:36
Breath Volume	1.589	20:36
Air Blank	0.000	20:36
Control Test	0.078	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 02/09/2026 02:15:48

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.078	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.224	02:11
Breath Volume		1.269	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.229	02:14
Breath Volume		1.539	02:14
Air Blank		0.000	02:14
Control Test		0.078	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RANGEL,DANNER
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: STORR

DATE/TIME OF TEST: 02/14/2026 03:41:59

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.078	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.223	03:37
Breath Volume		1.269	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.216	03:40
Breath Volume		1.324	03:40
Air Blank		0.000	03:41
Control Test		0.075	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: BOBBER

DATE/TIME OF TEST: 02/14/2026 05:01:51

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.077	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.147	04:57
Breath Volume	1.542	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.150	05:00
Breath Volume	2.531	05:00
Air Blank	0.000	05:00
Control Test	0.077	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: FRAGOSO VARGAS

DATE/TIME OF TEST: 02/14/2026 05:21:42

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.077	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.117	05:16
Breath Volume	2.058	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.109	05:20
Breath Volume	1.531	05:20
Air Blank	0.000	05:20
Control Test	0.077	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: CADEAU

DATE/TIME OF TEST: 02/16/2026 01:55:49

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.078	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.224	01:51
Breath Volume		1.304	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.227	01:54
Breath Volume		1.746	01:54
Air Blank		0.000	01:54
Control Test		0.078	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: KLOCKENBRINK

DATE/TIME OF TEST: 02/21/2026 01:58:11

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.077	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.257	01:53
Breath Volume		2.722	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.258	01:56
Breath Volume		2.941	01:56
Air Blank		0.000	01:57
Control Test		0.077	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 02/22/2026 04:24:28

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.077	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.133	04:20
Breath Volume	1.687	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.138	04:22
Breath Volume	1.640	04:22
Air Blank	0.000	04:23
Control Test	0.077	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 02/22/2026 04:54:28

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:43	
Control Test		0.077	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.094	04:47	Volume Not Met
Breath Volume		0.503	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 2	0.089	04:52	Volume Not Met
Breath Volume		0.667	04:52	
Air Blank		0.000	04:53	
Control Test		0.075	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 02/22/2026 05:03:14

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.075	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.114	04:58
Breath Volume	1.859	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:00
Subject Sample # 2	0.104	05:01
Breath Volume	1.414	05:01
Air Blank	0.000	05:02
Control Test	0.076	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GIL,RUBIN
ARRESTING OFFICER AGENCY: Lauderdale PD
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: PANGAN

DATE/TIME OF TEST: 02/26/2026 12:50:07

Diagnostics Check	OK	12:43
Air Blank	0.000	12:44
Control Test	0.078	12:44
Air Blank	0.000	12:45
Subject Sample # 1	0.000	12:45
Breath Volume	1.480	12:45
Air Blank	0.000	12:46
Air Blank	0.000	12:48
Subject Sample # 2	0.000	12:48
Breath Volume	1.601	12:48
Air Blank	0.000	12:49
Control Test	0.078	12:49
Air Blank	0.000	12:50
Diagnostics Check	OK	12:50

Data Download Date/Time: 03/04/2026 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MASEILLE,BRYANT
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: MACALUSO

DATE/TIME OF TEST: 02/27/2026 21:44:45

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.161	21:38
Breath Volume	1.121	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.155	21:43
Breath Volume	1.363	21:43
Air Blank	0.000	21:43
Control Test	0.078	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 03/04/2026 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TOWNSEND,KIRK-PATRICK
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: GORDYUSHKINA

DATE/TIME OF TEST: 02/28/2026 05:37:27

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.078	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.230	05:31
Breath Volume	1.882	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:33
Subject Sample # 2	0.230	05:35
Breath Volume	2.183	05:35
Air Blank	0.000	05:36
Control Test	0.075	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 03/04/2026 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 02/01/2026 01:45:30

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1 0.269	01:40	
Breath Volume	1.226	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank		01:44	Ambient Fail
Air Blank	0.000	01:45	

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: GONZALEZ PERALTA

DATE/TIME OF TEST: 02/01/2026 23:50:05

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.160	23:45
Breath Volume	3.875	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.159	23:48
Breath Volume	3.773	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: ROMERO MUNGUIA

DATE/TIME OF TEST: 02/02/2026 23:43:31

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.167	23:38
Breath Volume	2.488	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.165	23:41
Breath Volume	2.964	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GARVEY

DATE/TIME OF TEST: 02/03/2026 00:32:07

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.207	00:27
Breath Volume		2.132	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.203	00:30
Breath Volume		2.605	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 02/03/2026 19:21:51

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.079	19:16
Air Blank	0.000	19:16
Subject Sample # 1	0.167	19:17
Breath Volume	1.937	19:17
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.168	19:20
Breath Volume	2.089	19:20
Air Blank	0.000	19:20
Control Test	0.080	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CABALLERO

DATE/TIME OF TEST: 02/03/2026 20:52:16

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.262	20:42
Breath Volume	1.488	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.231	20:46
Breath Volume	1.433	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:48
Subject Sample # 3	0.252	20:50
Breath Volume	1.179	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: LOMRANTZ

DATE/TIME OF TEST: 02/03/2026 22:31:30

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.237	22:25
Breath Volume	1.722	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.245	22:29
Breath Volume	2.421	22:29
Air Blank	0.000	22:30
Control Test	0.078	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SIMON,JAIDEM
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KOUNCK

DATE/TIME OF TEST: 02/06/2026 22:15:01

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.079	22:06	
Air Blank	0.000	22:06	
Subject Sample	# 1 0.000	22:10	Volume Not Met
Breath Volume	0.253	22:10	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:14	
Control Test	0.079	22:14	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: AGUILAR HERNANDEZ

DATE/TIME OF TEST: 02/06/2026 23:21:56

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.129	23:17
Breath Volume	2.050	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.128	23:20
Breath Volume	2.183	23:20
Air Blank	0.000	23:20
Control Test	0.078	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LOPEZ LPOEZ

DATE/TIME OF TEST: 02/06/2026 23:55:32

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.224	23:50
Breath Volume		1.667	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.212	23:53
Breath Volume		1.457	23:53
Air Blank		0.000	23:54
Control Test		0.078	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS, TIA, L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MULHAM, DONALD
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 02/07/2026 01:43:39

Diagnostics Check		OK	01:26	
Air Blank		0.000	01:27	
Control Test		0.079	01:27	
Air Blank		0.000	01:28	
Subject Sample	# 1	0.129	01:31	Volume Not Met
Breath Volume		0.195	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 2	0.137	01:36	
Breath Volume		1.312	01:36	
Air Blank		0.000	01:37	
Air Blank		0.000	01:38	
Subject Sample	# 3	0.130	01:41	
Breath Volume		1.218	01:41	
Air Blank		0.000	01:42	
Control Test		0.077	01:43	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: AGILAR ESCOBAR

DATE/TIME OF TEST: 02/07/2026 21:21:38

Diagnostics Check		OK	21:06	
Air Blank		0.000	21:06	
Control Test		0.078	21:06	
Air Blank		0.000	21:07	
Subject Sample	# 1	0.172	21:10	Volume Not Met
Breath Volume		0.363	21:10	
Air Blank		0.000	21:11	
Air Blank		0.000	21:13	
Subject Sample	# 2	0.219	21:15	
Breath Volume		2.777	21:15	
Air Blank		0.000	21:16	
Air Blank		0.000	21:17	
Subject Sample	# 3	0.214	21:19	
Breath Volume		1.789	21:19	
Air Blank		0.000	21:20	
Control Test		0.077	21:20	
Air Blank		0.000	21:21	
Diagnostics Check		OK	21:21	

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 02/07/2026 23:11:13

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.087	23:06
Breath Volume	3.066	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.088	23:09
Breath Volume	3.500	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: D'ANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: ARTAU

DATE/TIME OF TEST: 02/08/2026 16:32:51

Diagnostics Check	OK	16:26
Air Blank	0.000	16:26
Control Test	0.079	16:27
Air Blank	0.000	16:27
Subject Sample # 1	0.000	16:28
Breath Volume	1.929	16:28
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 2	0.000	16:31
Breath Volume	1.585	16:31
Air Blank	0.000	16:31
Control Test	0.078	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR,MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: CERRATO

DATE/TIME OF TEST: 02/08/2026 21:27:42

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.171	21:22
Breath Volume	3.679	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.174	21:26
Breath Volume	4.308	21:26
Air Blank	0.000	21:26
Control Test	0.078	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD, ELIZABATH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SAINT MARTIN

DATE/TIME OF TEST: 02/08/2026 21:47:44

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.079	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.244	21:41
Breath Volume	1.531	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.241	21:46
Breath Volume	1.402	21:46
Air Blank	0.000	21:46
Control Test	0.078	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLASCO,LUIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: PETIT

DATE/TIME OF TEST: 02/09/2026 00:21:02

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.078	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.240	00:16
Breath Volume		1.765	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.234	00:19
Breath Volume		1.832	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:21

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU, YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 02/10/2026 03:35:28

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.078	03:30
Breath Volume	1.890	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.086	03:33
Breath Volume	2.441	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HERNANDEZ ROSARIO

DATE/TIME OF TEST: 02/11/2026 01:14:07

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.182	01:09
Breath Volume	2.113	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.188	01:12
Breath Volume	2.144	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 02/11/2026 03:00:57

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.074	02:56
Breath Volume		1.441	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.075	02:59
Breath Volume		1.140	02:59
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/11/2026 23:00:49

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.116	22:56
Breath Volume		2.027	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.122	22:59
Breath Volume		1.765	22:59
Air Blank		0.000	22:59
Control Test		0.077	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SURIN

DATE/TIME OF TEST: 02/12/2026 02:29:09

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.066	02:24
Breath Volume	1.890	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:26
Subject Sample # 2	0.066	02:27
Breath Volume	1.804	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ARENCIBIA

DATE/TIME OF TEST: 02/12/2026 03:07:42

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.080	03:03
Breath Volume		2.066	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.080	03:06
Breath Volume		1.109	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: GALE

DATE/TIME OF TEST: 02/12/2026 21:39:50

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.089	21:34
Breath Volume		1.367	21:34
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.093	21:38
Breath Volume		1.457	21:38
Air Blank		0.000	21:38
Control Test		0.079	21:39
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 02/13/2026 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/13/2026 05:05:41

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.078	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.166	04:59
Breath Volume	1.609	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.173	05:04
Breath Volume	1.679	05:04
Air Blank	0.000	05:04
Control Test	0.079	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 02/13/2026 15:23