

Bell, Danielle

From: Wilson, Barbie <BWilson@polksheriff.org>
Sent: Wednesday, September 06, 2017 2:46 PM
To: Bell, Danielle
Subject: Afternoon

Afternoon,

I was hoping for a good day today but unfortunately Instrument 80-001237 did not cooperate. I conducted 3 different test and it did not pass the Interferent Detect test on either inspection. I checked the to ensure that it was at the correct temperature, it appeared to be at 33 degrees. I waited another 10-15 minutes for it to heat up even more and this did not solve the problem. On the third attempt, I switched simulators and again it did not pass. I have printed and uploaded the test results and have removed it from use.

Do you want me to send it to you or to Enforcement Electronics? Also, could you please send me a copy of FDLE Form 40 for Instrument 80-001127?

Thank you for your assistance and stay safe during the storm.

Detention Deputy Barbie Wilson, 6493
Inmate Intake Trusty Supervisor
Agency Inspector

Office: 863-534-6248 (ext. 7919)
Cell: (863)287-9128

"One measure of your success will be the degree to which you build up others who work with you. While building up others, you will build up yourself."

— James E. Casey

From: Bell, Danielle [DanielleBell@fdle.state.fl.us]
Sent: Friday, September 1, 2017 12:34 PM
To: Wilson, Barbie
Subject: Re: Form-40-Agency-Insp-Rpt-Intox-80001

Perfect. If you'll just sign it. That's great.

Thanks,
Danielle

Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: "Wilson, Barbie" <BWilson@polksheriff.org>
Date: 9/1/17 12:31 PM (GMT-05:00)
To: "Bell, Danielle" <DanielleBell@fdle.state.fl.us>, "Wilson, Barbie" <BWilson@polksheriff.org>

Bell, Danielle

From: Bell, Danielle
Sent: Tuesday, May 09, 2017 5:03 PM
To: 'Wilson, Barbie'
Subject: RE: Scan Job from Copier

80-001237 @B

I'm sorry for the 2nd email. I just finished reviewing your inspections and on 80-001237 it looks like you entered the incorrect lot number and expiration date for the 0.20 alcohol reference solution. You used the lot number for the 0.05. Could you single line strike through the incorrect information, write in the correct information, initial and date the change then send me a copy?

Thanks,

Danielle

Danielle M. Bell
Department Inspector
Alcohol Testing Program
Florida Department of Law Enforcement
(850) 617-1286 (office)
(850) 274-2757 (cell)
(850) 921-3787 (fax)

-----Original Message-----

From: Wilson, Barbie [<mailto:BWilson@polksheriff.org>]
Sent: Tuesday, May 09, 2017 4:07 PM
To: Bell, Danielle
Subject: FW: Scan Job from Copier

I truly had an off day today conducting my inspections. I have indicated why I received the errors on the test, here is a copy of all three test. Have you received BA instrument from Enforcement Electronics belonging to the Polk County Sheriff's Office? I will be out of the office until tomorrow morning at 7:30 if I need to explain further and have court tomorrow afternoon.

Wilson

From: copier@polksheriff.org [copier@polksheriff.org]
Sent: Tuesday, May 9, 2017 3:17 PM
To: Wilson, Barbie
Subject: Scan Job from Copier

Bell, Danielle

From: Wilson, Barbie <BWilson@polksheriff.org>
Sent: Tuesday, May 09, 2017 4:07 PM
To: Bell, Danielle
Subject: FW: Scan Job from Copier
Attachments: scan.pdf

80-001127

80-001237

80-001080

I truly had an off day today conducting my inspections. I have indicated why I received the errors on the test, here is a copy of all three test. Have you received BA instrument from Enforcement Electronics belonging to the Polk County Sheriff's Office? I will be out of the office until tomorrow morning at 7:30 if I need to explain further and have court tomorrow afternoon.

Wilson

From: copier@polksheriff.org [copier@polksheriff.org]
Sent: Tuesday, May 9, 2017 3:17 PM
To: Wilson, Barbie
Subject: Scan Job from Copier