Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HASMI,CHRISTIAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MORALESGARCIADEGARCI

DATE/TIME OF TEST: 12/05/2025 03:33:22

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.085	03:28
Breath Volume		2.187	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.082	03:31
Breath Volume		1.824	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: PAU

DATE/TIME OF TEST: 12/09/2025 19:18:09

Diagnostics Check		OK	19:11
Air Blank		0.000	19:11
Control Test		0.080	19:11
Air Blank		0.000	19:12
Subject Sample	# 1	0.368	19:13
Breath Volume		1.632	19:13
Air Blank		0.000	19:14
Air Blank		0.000	19:15
Subject Sample	# 2	0.352	19:16
Breath Volume		1.750	19:16
Air Blank		0.000	19:17
Control Test		0.079	19:17
Air Blank		0.000	19:18
Diagnostics Check		OK	19:18

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HILSON,RANDY
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ALMEIDA PEREZ

DATE/TIME OF TEST: 12/15/2025 22:55:54

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.082	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1	0.219	22:44	
Breath Volume		1.945	22:44	
Air Blank		0.000	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 2	0.189	22:49	Volume Not Met
Breath Volume		0.527	22:49	
Air Blank		0.000	22:50	
Air Blank		0.000	22:52	
Subject Sample	# 3		22:54	Subject Test Refused
Breath Volume		1.078	22:54	
Air Blank		0.000	22:54	
Control Test		0.077	22:55	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:55	

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: MANUELDASILVAVIEIRA

DATE/TIME OF TEST: 12/17/2025 07:26:26

Diagnostics Check		OK	07:12	
Air Blank		0.000	07:13	
Control Test		0.080	07:13	
Air Blank		0.000	07:14	
Subject Sample	# 1		07:17	No Sample Provided
Breath Volume		0.000	07:17	
Air Blank		0.000	07:17	
Air Blank		0.000	07:19	
Subject Sample	# 2	0.142	07:21	
Breath Volume		1.937	07:21	
Air Blank		0.000	07:22	
Air Blank		0.000	07:23	
Subject Sample	# 3	0.136	07:24	
Breath Volume		1.933	07:24	
Air Blank		0.000	07:25	
Control Test		0.079	07:25	
Air Blank		0.000	07:26	
Diagnostics Check		OK	07:26	

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERE,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: IRBY,DOMINIC
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: MATOM BOP

DATE/TIME OF TEST: 12/21/2025 01:03:38

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.171	00:58
Breath Volume		1.429	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.172	01:02
Breath Volume		1.886	01:02
Air Blank		0.000	01:02
Control Test		0.081	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,OSCAR,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HILSON,RANDY
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: BASQUES PEREZ

DATE/TIME OF TEST: 12/21/2025 08:21:06

Diagnostics Check		OK	08:14
Air Blank		0.000	08:14
Control Test		0.081	08:15
Air Blank		0.000	08:15
Subject Sample	# 1	0.121	08:16
Breath Volume		2.203	08:16
Air Blank		0.000	08:17
Air Blank		0.000	08:19
Subject Sample	# 2	0.121	08:19
Breath Volume		3.164	08:19
Air Blank		0.000	08:20
Control Test		0.081	08:20
Air Blank		0.000	08:21
Diagnostics Check		OK	08:21

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: PEREZ MATEO

DATE/TIME OF TEST: 12/24/2025 19:21:44

Diagnostics Check		OK	19:15
Air Blank		0.000	19:16
Control Test		0.082	19:16
Air Blank		0.000	19:16
Subject Sample	# 1	0.243	19:17
Breath Volume		1.621	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:19
Subject Sample	# 2	0.247	19:20
Breath Volume		1.539	19:20
Air Blank		0.000	19:20
Control Test		0.081	19:21
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 12/30/2025 19:21

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBERNARDINO,BRANDON,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DIBERNARDINO,BRANDON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HERNANDEZ RAMIREZ

DATE/TIME OF TEST: 12/05/2025 23:26:32

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.084	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.202	23:21
Breath Volume		3.179	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.196	23:24
Breath Volume		2.257	23:24
Air Blank		0.000	23:25
Control Test		0.082	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 12/12/2025 10:00

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HULL,GARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 11:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 12/11/2025 12:37:51

Diagnostics Check		OK	12:29	
Air Blank		0.000	12:29	
Control Test		0.083	12:30	
Air Blank		0.000	12:30	
Subject Sample	# 1	0.263	12:33	Volume Not Met
Breath Volume		0.343	12:33	
Air Blank		0.000	12:34	
Air Blank		0.000	12:36	
Subject Sample	# 2		12:36	Subject Test Refused
Breath Volume		0.000	12:36	
Air Blank		0.000	12:36	
Control Test		0.082	12:37	
Air Blank		0.000	12:37	
Diagnostics Check		OK	12:37	

Data Download Date/Time: 12/12/2025 10:00

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FARRAJ,S
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PEREZ-DIAZ

DATE/TIME OF TEST: 12/28/2025 04:25:20

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.084	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.192	04:20
Breath Volume		1.421	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.184	04:23
Breath Volume		1.687	04:23
Air Blank		0.000	04:24
Control Test		0.082	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 01/16/2026 10:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: PEAY

DATE/TIME OF TEST: 12/24/2025 23:15:53

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.081	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.156	23:11
Breath Volume		2.109	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:13
Subject Sample	# 2	0.156	23:14
Breath Volume		2.289	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 01/12/2026 08:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: MOULD

DATE/TIME OF TEST: 12/31/2025 21:09:53

Diagnostics Check		OK	21:03
Air Blank		0.000	21:04
Control Test		0.082	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.007	21:05
Breath Volume		3.476	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.000	21:08
Breath Volume		2.363	21:08
Air Blank		0.000	21:08
Control Test		0.079	21:09
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 01/12/2026 08:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JORDAN,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 12/02/2025 22:52:37

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.099	22:47
Breath Volume		2.140	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.092	22:51
Breath Volume		2.230	22:51
Air Blank		0.000	22:51
Control Test		0.080	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 12/30/2025 22:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: HERRON

DATE/TIME OF TEST: 12/05/2025 23:43:46

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.081	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:39	No Sample Provided
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 2		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:42	
Control Test		0.080	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 12/30/2025 22:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ALTAMIRANO-VENEGAS

DATE/TIME OF TEST: 12/06/2025 02:57:26

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.081	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1	0.239	02:46	No .020 agreement
Breath Volume		1.726	02:46	
Air Blank		0.000	02:46	
Air Blank		0.000	02:48	
Subject Sample	# 2	0.208	02:49	No .020 agreement
Breath Volume		1.273	02:49	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 3	0.217	02:55	Volume Not Met
Breath Volume		0.375	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 12/30/2025 22:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SNOW,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: BURMEISTER

DATE/TIME OF TEST: 12/06/2025 17:06:31

Diagnostics Check		OK	16:59
Air Blank		0.000	16:59
Control Test		0.081	17:00
Air Blank		0.000	17:00
Subject Sample	# 1	0.157	17:01
Breath Volume		2.035	17:01
Air Blank		0.000	17:02
Air Blank		0.000	17:04
Subject Sample	# 2	0.161	17:04
Breath Volume		2.351	17:04
Air Blank		0.000	17:05
Control Test		0.080	17:05
Air Blank		0.000	17:06
Diagnostics Check		OK	17:06

Data Download Date/Time: 12/30/2025 22:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: SIZEMORE

DATE/TIME OF TEST: 12/07/2025 00:06:54

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.080	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 3		23:57	No Sample Provided
Breath Volume		0.000	23:57	
Air Blank		0.000	23:57	
Air Blank		0.000	23:59	
Subject Sample	# 1	0.155	00:00	
Breath Volume		1.949	00:00	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2	0.151	00:05	
Breath Volume		2.246	00:05	
Air Blank		0.000	00:05	
Control Test		0.080	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 12/30/2025 22:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ROMIG,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 12/07/2025 01:35:51

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:27	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:31	No Sample Provided
Breath Volume		0.000	01:31	
Air Blank		0.000	01:31	
Air Blank		0.000	01:33	
Subject Sample	# 2		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.081	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 12/30/2025 22:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 12/11/2025 01:23:59

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.082	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.153	01:18
Breath Volume		1.828	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.154	01:22
Breath Volume		2.058	01:22
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 12/30/2025 22:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SNOW,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/15/2025 00:12:20

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.082	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.263	00:07
Breath Volume		1.644	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:09
Subject Sample	# 2	0.266	00:10
Breath Volume		1.238	00:10
Air Blank		0.000	00:11
Control Test		0.082	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HANNA,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: DOUGALL

DATE/TIME OF TEST: 12/19/2025 00:05:52

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.267	00:00
Breath Volume		3.671	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.268	00:04
Breath Volume		4.062	00:04
Air Blank		0.000	00:04
Control Test		0.080	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 12/19/2025 22:57:54

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:49	
Control Test		0.082	22:49	
Air Blank		0.000	22:50	
Subject Sample	# 1		22:53	No Sample Provided
Breath Volume		0.000	22:53	
Air Blank		0.000	22:53	
Air Blank		0.000	22:55	
Subject Sample	# 2		22:56	Subject Test Refused
Breath Volume		0.000	22:56	
Air Blank		0.000	22:56	
Control Test		0.081	22:57	
Air Blank		0.000	22:57	
Diagnostics Check		OK	22:57	

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAHLKE,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: NORWOOD

DATE/TIME OF TEST: 12/20/2025 03:50:22

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.081	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.222	03:45
Breath Volume		3.089	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.229	03:48
Breath Volume		1.824	03:48
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCQUEEN,G
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: MORONTA MARTE

DATE/TIME OF TEST: 12/23/2025 00:20:29

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:09	
Control Test		0.080	00:09	
Air Blank		0.000	00:09	
Subject Sample	# 1	0.222	00:13	Volume Not Met
Breath Volume		0.261	00:13	
Air Blank		0.000	00:13	
Air Blank		0.000	00:15	
Subject Sample	# 2	0.230	00:15	
Breath Volume		1.265	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 3	0.241	00:18	
Breath Volume		1.875	00:18	
Air Blank		0.000	00:19	
Control Test		0.080	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILLIAMS,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MARBIN

DATE/TIME OF TEST: 12/23/2025 00:57:16

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.147	00:52
Breath Volume		1.785	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.138	00:55
Breath Volume		1.238	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/23/2025 03:28:35

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:20	
Control Test		0.081	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:24	No Sample Provided
Breath Volume		0.000	03:24	
Air Blank		0.000	03:24	
Air Blank		0.000	03:26	
Subject Sample	# 2		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.081	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JORDAN,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 12/25/2025 01:03:36

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.082	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.178	00:58
Breath Volume		2.648	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.173	01:01
Breath Volume		3.738	01:01
Air Blank		0.000	01:02
Control Test		0.081	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SHOEMAKER

DATE/TIME OF TEST: 12/26/2025 04:34:33

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.081	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.103	04:29
Breath Volume		2.238	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.099	04:32
Breath Volume		2.722	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 12/30/2025 22:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: GRACEY

DATE/TIME OF TEST: 12/01/2025 23:41:57

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.081	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.000	23:37
Breath Volume		1.984	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.000	23:40
Breath Volume		2.093	23:40
Air Blank		0.000	23:41
Control Test		0.081	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 12/03/2025 23:15:50

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:04	
Control Test		0.082	23:05	
Air Blank		0.000	23:05	
Subject Sample	# 1	0.138	23:08	Slope Not Level
Breath Volume		1.449	23:08	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2	0.132	23:14	Slope Not Level
Breath Volume		1.367	23:14	
Air Blank		0.000	23:14	
Control Test		0.079	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALONY,T
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CHILEL LOPEZ

DATE/TIME OF TEST: 12/06/2025 00:52:13

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.081	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.247	00:47
Breath Volume		2.640	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:49
Subject Sample	# 2	0.254	00:50
Breath Volume		1.570	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: BAEZ JR

DATE/TIME OF TEST: 12/06/2025 01:31:02

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.132	01:26
Breath Volume		1.316	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:28
Subject Sample	# 2	0.128	01:29
Breath Volume		1.746	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:31

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBINSON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MAZZURCO

DATE/TIME OF TEST: 12/06/2025 02:08:44

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.250	02:03
Breath Volume		1.707	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.231	02:07
Breath Volume		1.246	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/07/2025 01:45:28

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.081	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.139	01:40
Breath Volume		1.191	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.138	01:43
Breath Volume		3.675	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBESON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BURSEK

DATE/TIME OF TEST: 12/07/2025 03:09:50

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.081	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.103	03:05
Breath Volume		2.152	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.103	03:08
Breath Volume		2.144	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARDARELLI,CARA
ARRESTING OFFICER AGENCY: Manalapan Pub Safety
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/07/2025 05:07:24

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.081	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.147	05:02
Breath Volume		1.187	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.151	05:05
Breath Volume		1.332	05:05
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: DAMES

DATE/TIME OF TEST: 12/07/2025 05:24:53

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.080	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.186	05:20
Breath Volume		1.398	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.190	05:23
Breath Volume		2.691	05:23
Air Blank		0.000	05:23
Control Test		0.080	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: CHENG

DATE/TIME OF TEST: 12/07/2025 20:27:02

Diagnostics Check		OK	20:20
Air Blank		0.000	20:20
Control Test		0.080	20:21
Air Blank		0.000	20:21
Subject Sample	# 1	0.224	20:22
Breath Volume		1.976	20:22
Air Blank		0.000	20:22
Air Blank		0.000	20:24
Subject Sample	# 2	0.223	20:25
Breath Volume		2.136	20:25
Air Blank		0.000	20:26
Control Test		0.080	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:27

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ETIENNE

DATE/TIME OF TEST: 12/07/2025 22:06:35

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:55	
Control Test		0.080	21:55	
Air Blank		0.000	21:56	
Subject Sample	# 1	0.234	21:59	Volume Not Met
Breath Volume		0.066	21:59	
Air Blank		0.000	22:00	
Air Blank		0.000	22:01	
Subject Sample	# 2	0.227	22:04	Volume Not Met
Breath Volume		0.105	22:04	
Air Blank		0.000	22:05	
Control Test		0.079	22:05	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GOMEZ ORDONEZ

DATE/TIME OF TEST: 12/08/2025 01:06:44

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.080	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.185	01:02
Breath Volume		1.917	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.187	01:05
Breath Volume		2.023	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HU,HAROLD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 12/08/2025 01:59:50

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.080	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.167	01:55
Breath Volume		1.214	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.171	01:58
Breath Volume		1.378	01:58
Air Blank		0.000	01:58
Control Test		0.081	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: LOZADA

DATE/TIME OF TEST: 12/08/2025 05:43:33

Diagnostics Check		OK	05:36
Air Blank		0.000	05:37
Control Test		0.081	05:37
Air Blank		0.000	05:38
Subject Sample	# 1	0.178	05:38
Breath Volume		1.343	05:38
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.180	05:41
Breath Volume		1.929	05:41
Air Blank		0.000	05:42
Control Test		0.080	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: NZIMANDE

DATE/TIME OF TEST: 12/09/2025 04:55:00

Diagnostics Check		OK	04:48
Air Blank		0.000	04:49
Control Test		0.081	04:49
Air Blank		0.000	04:49
Subject Sample	# 1	0.070	04:50
Breath Volume		1.347	04:50
Air Blank		0.000	04:51
Air Blank		0.000	04:52
Subject Sample	# 2	0.070	04:53
Breath Volume		1.660	04:53
Air Blank		0.000	04:54
Control Test		0.081	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:55

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/10/2025 03:58:28

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.082	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.224	03:53
Breath Volume		1.800	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.220	03:56
Breath Volume		2.257	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 12/12/2025 16:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RIOS OSPINA

DATE/TIME OF TEST: 12/11/2025 02:37:29

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.253	02:32
Breath Volume		1.441	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.257	02:35
Breath Volume		2.164	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 12/12/2025 16:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,A
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: MOON

DATE/TIME OF TEST: 12/11/2025 22:25:59

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.081	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.161	22:21
Breath Volume		1.449	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.162	22:24
Breath Volume		2.105	22:24
Air Blank		0.000	22:25
Control Test		0.081	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 12/12/2025 16:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GRIGLIO

DATE/TIME OF TEST: 12/12/2025 01:58:23

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:52	
Control Test		0.075	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.190	01:53	
Breath Volume		2.210	01:53	
Air Blank		0.000	01:54	
Air Blank		0.000	01:56	
Subject Sample	# 2	0.189	01:56	
Breath Volume		1.968	01:56	
Air Blank		0.000	01:57	
Control Test		0.081	01:57	Interferent Detect
Air Blank		0.000	01:58	

Data Download Date/Time: 12/27/2025 18:08

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GRIGLIO

DATE/TIME OF TEST: 12/12/2025 02:06:56

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.076	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.182	02:02
Breath Volume		2.359	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:04
Subject Sample	# 2	0.179	02:05
Breath Volume		2.003	02:05
Air Blank		0.000	02:06
Control Test		0.076	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 12/27/2025 18:08

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HAYES-WHALEY

DATE/TIME OF TEST: 12/27/2025 02:19:13

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.074	02:18	Control Outside Tolerance
Air Blank	0.000	02:19	

Data Download Date/Time: 12/27/2025 18:08

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HAYES-WHALEY

DATE/TIME OF TEST: 12/27/2025 02:27:03

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.231	02:22
Breath Volume		2.710	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:24
Subject Sample	# 2	0.226	02:25
Breath Volume		2.140	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:27

Data Download Date/Time: 12/27/2025 18:08

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HATGIDEK

DATE/TIME OF TEST: 12/28/2025 02:41:19

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.075	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1	0.121	02:37	
Breath Volume		4.695	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 2		02:40	RFI Detect
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.077	02:40	
Air Blank			02:41	RFI Detect
Air Blank			02:41	RFI Detect

Data Download Date/Time: 01/09/2026 18:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HATGIDEK

DATE/TIME OF TEST: 12/28/2025 02:47:02

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:44	
Control Test		0.075	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	RFI Detect
Breath Volume		0.000	02:45	
Air Blank		0.000	02:46	
Control Test		0.076	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 01/09/2026 18:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HATGIDEK

DATE/TIME OF TEST: 12/28/2025 02:55:27

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.075	02:49	
Air Blank		0.000	02:49	
Subject Sample	# 1	0.121	02:50	
Breath Volume		5.761	02:50	
Air Blank		0.000	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.110	02:53	
Breath Volume		4.300	02:53	
Air Blank		0.000	02:54	
Control Test		0.077	02:54	
Air Blank		0.000	02:54	
Diagnostics Check		Fail	02:54	Analytical Stability Test
Air Blank		0.000	02:55	

Data Download Date/Time: 01/09/2026 18:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: YOUNG,TIMYRON
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HUNOLD

DATE/TIME OF TEST: 12/01/2025 02:29:34

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.000	02:24
Breath Volume		2.324	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.000	02:28
Breath Volume		1.890	02:28
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 12/03/2025 02:29:07

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.168	02:24
Breath Volume		2.167	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.171	02:27
Breath Volume		2.667	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FREEMAN,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MCCOVERY

DATE/TIME OF TEST: 12/05/2025 04:05:23

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.129	04:00
Breath Volume		1.863	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.129	04:03
Breath Volume		1.707	04:03
Air Blank		0.000	04:04
Control Test		0.077	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PALMERO

DATE/TIME OF TEST: 12/06/2025 04:16:43

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.162	04:12
Breath Volume		2.910	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.162	04:15
Breath Volume		2.746	04:15
Air Blank		0.000	04:15
Control Test		0.077	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MCGRUDER

DATE/TIME OF TEST: 12/07/2025 03:50:13

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.077	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1	0.109	03:39	
Breath Volume		1.675	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.096	03:45	Volume Not Met
Breath Volume		0.722	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 3	0.099	03:48	
Breath Volume		1.625	03:48	
Air Blank		0.000	03:49	
Control Test		0.075	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: TIFFANY

DATE/TIME OF TEST: 12/10/2025 00:21:25

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.215	00:16
Breath Volume		2.746	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.206	00:19
Breath Volume		2.078	00:19
Air Blank		0.000	00:20
Control Test		0.077	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 12/12/2025 02:22:55

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.078	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.148	02:18
Breath Volume		1.875	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.135	02:21
Breath Volume		1.214	02:21
Air Blank		0.000	02:21
Control Test		0.078	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 12/14/2025 00:23:49

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.078	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.000	00:19
Breath Volume		1.644	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.000	00:22
Breath Volume		2.183	00:22
Air Blank		0.000	00:22
Control Test		0.078	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 12/15/2025 00:28:22

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.000	00:23
Breath Volume		2.660	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:26
Breath Volume		2.140	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 12/23/2025 13:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FREEMAN,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 12/15/2025 02:07:52

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.141	02:03
Breath Volume		2.773	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.139	02:06
Breath Volume		2.796	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HANDELAND

DATE/TIME OF TEST: 12/17/2025 01:04:23

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.132	00:56
Breath Volume		1.105	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:58
Subject Sample	# 2	0.154	00:59
Breath Volume		2.785	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 3	0.129	01:02
Breath Volume		1.453	01:02
Air Blank		0.000	01:03
Control Test		0.078	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN,DELRICKUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN,DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: SOUBLET JR

DATE/TIME OF TEST: 12/18/2025 05:24:48

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.079	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.135	05:20
Breath Volume		3.144	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:22
Subject Sample	# 2	0.129	05:23
Breath Volume		3.132	05:23
Air Blank		0.000	05:23
Control Test		0.078	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:59

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/20/2025 14:25:40

Diagnostics Check		OK	14:19
Air Blank		0.000	14:19
Control Test		0.079	14:20
Air Blank		0.000	14:20
Subject Sample	# 1	0.000	14:21
Breath Volume		2.449	14:21
Air Blank		0.000	14:21
Air Blank		0.000	14:23
Subject Sample	# 2	0.000	14:24
Breath Volume		3.085	14:24
Air Blank		0.000	14:24
Control Test		0.079	14:25
Air Blank		0.000	14:25
Diagnostics Check		OK	14:25

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 11:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WINGFIELD

DATE/TIME OF TEST: 12/21/2025 00:16:37

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.187	00:12
Breath Volume		1.515	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.189	00:15
Breath Volume		2.007	00:15
Air Blank		0.000	00:15
Control Test		0.077	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 12/23/2025 13:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAVO,GEORGE,A
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: BRAVO,GEORGE
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 12/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 12/18/2025 22:11:50

Diagnostics Check		OK	22:07	
Air Blank		0.000	22:08	
Control Test		0.082	22:08	
Air Blank		0.000	22:08	
Subject Sample	# 1		22:10	Subject Test Refused
Breath Volume		0.000	22:10	
Air Blank		0.000	22:10	
Control Test		0.080	22:11	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:11	

Data Download Date/Time: 01/15/2026 09:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JAVIER,A
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: GONZALEZ,F
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 12/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 12/20/2025 19:56:17

Diagnostics Check		OK	19:49
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Subject Sample	# 1	0.000	19:52
Breath Volume		2.062	19:52
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.000	19:54
Breath Volume		2.273	19:54
Air Blank		0.000	19:55
Control Test		0.078	19:55
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 01/15/2026 09:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JAVIER,A
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: ACOSTA,S
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: SANTIAGO PEREZ

DATE/TIME OF TEST: 12/20/2025 20:42:09

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.081	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.226	20:37
Breath Volume		2.847	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.208	20:40
Breath Volume		1.707	20:40
Air Blank		0.000	20:41
Control Test		0.081	20:41
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 01/15/2026 09:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,CARLOS,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LABRADOR,A
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: BLYSKAL

DATE/TIME OF TEST: 12/26/2025 13:20:17

Diagnostics Check		OK	13:17	
Air Blank		0.000	13:17	
Control Test		0.082	13:18	
Air Blank		0.000	13:18	
Subject Sample	# 1		13:18	Subject Test Refused
Breath Volume		0.000	13:18	
Air Blank		0.000	13:19	
Control Test		0.082	13:19	
Air Blank		0.000	13:20	
Diagnostics Check		OK	13:20	

Data Download Date/Time: 01/15/2026 09:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 12/12/2025 23:52:21

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.078	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.000	23:47
Breath Volume		1.335	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.000	23:50
Breath Volume		1.714	23:50
Air Blank		0.000	23:51
Control Test		0.079	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 12/14/2025 00:25:55

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.078	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.161	00:20
Breath Volume		2.503	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.157	00:24
Breath Volume		2.484	00:24
Air Blank		0.000	00:24
Control Test		0.078	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCOY,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 12/14/2025 01:51:18

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.077	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.195	01:46
Breath Volume		1.500	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.197	01:49
Breath Volume		1.253	01:49
Air Blank		0.000	01:50
Control Test		0.077	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: VAZQUEZ LOPEZ

DATE/TIME OF TEST: 12/15/2025 23:19:18

Diagnostics Check		OK	23:12
Air Blank		0.000	23:13
Control Test		0.078	23:13
Air Blank		0.000	23:13
Subject Sample	# 1	0.207	23:14
Breath Volume		3.558	23:14
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.205	23:17
Breath Volume		3.210	23:17
Air Blank		0.000	23:18
Control Test		0.078	23:18
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: JUAREZSOLOJ

DATE/TIME OF TEST: 12/17/2025 00:51:46

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.078	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.243	00:46
Breath Volume		3.347	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.235	00:50
Breath Volume		3.207	00:50
Air Blank		0.000	00:50
Control Test		0.077	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMIGELSKI,TREVOR
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 12/17/2025 02:38:42

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.078	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.181	02:33
Breath Volume		1.308	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.178	02:36
Breath Volume		1.382	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GUTIERREZ,DAVID
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: SMALDONE

DATE/TIME OF TEST: 12/18/2025 01:02:13

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.078	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.119	00:57
Breath Volume		1.648	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.120	01:00
Breath Volume		3.117	01:00
Air Blank		0.000	01:01
Control Test		0.078	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO,JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 12/18/2025 05:17:20

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.079	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.130	05:12
Breath Volume		1.796	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.134	05:15
Breath Volume		2.457	05:15
Air Blank		0.000	05:16
Control Test		0.078	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAPARCHE,CAMDEN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 12/19/2025 02:42:33

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.078	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.231	02:35	Volume Not Met
Breath Volume		0.304	02:35	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.216	02:40	Volume Not Met
Breath Volume		0.035	02:40	
Air Blank		0.000	02:41	
Control Test		0.077	02:41	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: ORTIZ MALDONALDO

DATE/TIME OF TEST: 12/19/2025 19:55:34

Diagnostics Check		OK	19:47
Air Blank		0.000	19:48
Control Test		0.077	19:48
Air Blank		0.000	19:49
Subject Sample	# 1	0.287	19:50
Breath Volume		2.367	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:52
Subject Sample	# 2	0.288	19:53
Breath Volume		2.566	19:53
Air Blank		0.000	19:54
Control Test		0.076	19:54
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/19/2025 22:35:02

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.077	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.000	22:30
Breath Volume		3.304	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.000	22:33
Breath Volume		3.003	22:33
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ESCOBAR VILLATARO

DATE/TIME OF TEST: 12/20/2025 04:11:33

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.078	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.203	04:06
Breath Volume		2.128	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.189	04:09
Breath Volume		1.410	04:09
Air Blank		0.000	04:10
Control Test		0.077	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: LOPEZ RIVAS

DATE/TIME OF TEST: 12/21/2025 06:04:55

Diagnostics Check		OK	05:57
Air Blank		0.000	05:57
Control Test		0.078	05:58
Air Blank		0.000	05:58
Subject Sample	# 1	0.217	06:00
Breath Volume		1.242	06:00
Air Blank		0.000	06:00
Air Blank		0.000	06:02
Subject Sample	# 2	0.218	06:03
Breath Volume		2.003	06:03
Air Blank		0.000	06:03
Control Test		0.077	06:04
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORRISON,ERIC
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: MAYORGA MARTINEZ

DATE/TIME OF TEST: 12/21/2025 11:32:52

Diagnostics Check		OK	11:26
Air Blank		0.000	11:26
Control Test		0.078	11:26
Air Blank		0.000	11:27
Subject Sample	# 1	0.166	11:28
Breath Volume		2.566	11:28
Air Blank		0.000	11:29
Air Blank		0.000	11:30
Subject Sample	# 2	0.168	11:31
Breath Volume		2.074	11:31
Air Blank		0.000	11:31
Control Test		0.077	11:32
Air Blank		0.000	11:32
Diagnostics Check		OK	11:32

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 12/21/2025 22:40:44

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.078	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.060	22:36
Breath Volume		1.855	22:36
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.060	22:39
Breath Volume		1.160	22:39
Air Blank		0.000	22:39
Control Test		0.078	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CAJUSTE,AIRD
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 12/22/2025 10:10:56

Diagnostics Check		OK	10:04
Air Blank		0.000	10:04
Control Test		0.079	10:04
Air Blank		0.000	10:05
Subject Sample	# 1	0.375	10:06
Breath Volume		1.585	10:06
Air Blank		0.000	10:07
Air Blank		0.000	10:08
Subject Sample	# 2	0.369	10:09
Breath Volume		1.722	10:09
Air Blank		0.000	10:09
Control Test		0.078	10:10
Air Blank		0.000	10:10
Diagnostics Check		OK	10:10

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: HERMAN

DATE/TIME OF TEST: 12/24/2025 23:45:42

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.078	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.138	23:40
Breath Volume		4.175	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.142	23:44
Breath Volume		3.972	23:44
Air Blank		0.000	23:44
Control Test		0.077	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GABRIEL,JACKYME
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: MACEDA NINO

DATE/TIME OF TEST: 12/25/2025 21:15:31

Diagnostics Check		OK	21:08
Air Blank		0.000	21:09
Control Test		0.078	21:09
Air Blank		0.000	21:09
Subject Sample	# 1	0.204	21:10
Breath Volume		1.457	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.207	21:13
Breath Volume		1.417	21:13
Air Blank		0.000	21:14
Control Test		0.077	21:14
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROWE,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TAMARIZ AVENDANO

DATE/TIME OF TEST: 12/26/2025 19:10:55

Diagnostics Check		OK	19:03
Air Blank		0.000	19:04
Control Test		0.078	19:04
Air Blank		0.000	19:05
Subject Sample	# 1	0.173	19:05
Breath Volume		1.542	19:05
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.166	19:09
Breath Volume		1.937	19:09
Air Blank		0.000	19:09
Control Test		0.077	19:10
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FISHBAUGH,CHARLES
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: PRITCHARD

DATE/TIME OF TEST: 12/27/2025 04:58:50

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.078	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.198	04:54
Breath Volume		1.167	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.199	04:57
Breath Volume		1.996	04:57
Air Blank		0.000	04:57
Control Test		0.078	04:58
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARCIA,ISRAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: LORES MONTALVO

DATE/TIME OF TEST: 12/27/2025 19:23:50

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.078	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.189	19:18
Breath Volume		1.566	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.180	19:22
Breath Volume		1.167	19:22
Air Blank		0.000	19:22
Control Test		0.076	19:23
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD,RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MONTANEZ

DATE/TIME OF TEST: 12/28/2025 00:23:04

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.077	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.197	00:17
Breath Volume		1.304	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.201	00:21
Breath Volume		1.253	00:21
Air Blank		0.000	00:22
Control Test		0.077	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:23

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 12/29/2025 21:31:49

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:28	
Control Test		0.078	21:29	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:30	Subject Test Refused
Breath Volume		0.000	21:30	
Air Blank		0.000	21:30	
Control Test		0.079	21:31	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/29/2025 22:17:38

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.078	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.197	22:12
Breath Volume		1.757	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.201	22:15
Breath Volume		1.234	22:15
Air Blank		0.000	22:16
Control Test		0.078	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A2
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: FLEURY

DATE/TIME OF TEST: 12/31/2025 01:41:28

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.214	01:36
Breath Volume		1.449	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.215	01:39
Breath Volume		2.117	01:39
Air Blank		0.000	01:40
Control Test		0.078	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 01/16/2026 13:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.