Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LEWIS,S
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/24/2025 03:30:25

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.081	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.152	03:24
Breath Volume		1.906	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.148	03:28
Breath Volume		1.832	03:28
Air Blank		0.000	03:29
Control Test		0.082	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 01/22/2026 05:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARTINI

DATE/TIME OF TEST: 12/24/2025 04:10:34

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.081	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.175	04:05
Breath Volume		1.886	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.162	04:09
Breath Volume		1.464	04:09
Air Blank		0.000	04:09
Control Test		0.082	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 01/22/2026 05:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MUNOZ,W
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/25/2025 03:38:18

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.080	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.209	03:33
Breath Volume		1.464	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.201	03:36
Breath Volume		1.527	03:36
Air Blank		0.000	03:37
Control Test		0.081	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 01/22/2026 05:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANDOVAL,R
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 12/25/2025 04:22:13

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.204	04:17
Breath Volume		1.371	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.211	04:20
Breath Volume		1.906	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 01/22/2026 05:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,JAMES,A
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: CHITTY,LAURA
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:08

SUBJECT'S LAST NAME: GAMBOA PINTO

DATE/TIME OF TEST: 12/25/2025 11:46:01

Diagnostics Check		OK	11:38	
Air Blank		0.000	11:38	
Control Test		0.081	11:39	
Air Blank		0.000	11:39	
Subject Sample	# 1	0.152	11:40	
Breath Volume		1.796	11:40	
Air Blank		0.000	11:41	
Air Blank		0.000	11:43	
Subject Sample	# 2	0.153	11:44	
Breath Volume		1.500	11:44	
Air Blank		0.000	11:44	
Control Test		0.080	11:44	
Air Blank		0.000	11:45	
Diagnostics Check		Fail	11:45	Temperature Regulation Test
Air Blank		0.000	11:46	

Data Download Date/Time: 01/22/2026 05:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: YOUNGBLOOD,B
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/26/2025 01:22:02

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.078	01:17
Breath Volume		1.757	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.078	01:20
Breath Volume		1.312	01:20
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:22

Data Download Date/Time: 01/22/2026 05:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LAYNE,N
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/30/2025 22:16:14

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.081	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.119	22:11
Breath Volume		1.507	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.112	22:14
Breath Volume		1.601	22:14
Air Blank		0.000	22:15
Control Test		0.081	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 01/22/2026 05:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DELROSSO,D
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/30/2025 23:17:53

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:05	
Control Test		0.082	23:06	
Air Blank		0.000	23:06	
Subject Sample	# 1	0.044	23:09	Volume Not Met
Breath Volume		0.414	23:09	
Air Blank		0.000	23:10	
Air Blank		0.000	23:12	
Subject Sample	# 2	0.205	23:13	
Breath Volume		1.148	23:13	
Air Blank		0.000	23:13	
Air Blank		0.000	23:15	
Subject Sample	# 3	0.209	23:16	
Breath Volume		1.785	23:16	
Air Blank		0.000	23:16	
Control Test		0.081	23:17	
Air Blank		0.000	23:17	
Diagnostics Check		OK	23:17	

Data Download Date/Time: 01/22/2026 05:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MCMAHAN,J
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: ALBANESE

DATE/TIME OF TEST: 12/31/2025 22:15:29

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.082	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.000	22:11
Breath Volume		1.789	22:11
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.000	22:14
Breath Volume		1.460	22:14
Air Blank		0.000	22:14
Control Test		0.082	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 01/22/2026 05:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: SCOZZARI,KEN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: VALENCIA PINEDA

DATE/TIME OF TEST: 12/04/2025 08:36:15

Diagnostics Check		OK	08:19	
Air Blank		0.000	08:20	
Control Test		0.080	08:20	
Air Blank		0.000	08:20	
Subject Sample	# 1	0.096	08:23	Volume Not Met
Breath Volume		0.605	08:23	
Air Blank		0.000	08:24	
Air Blank		0.000	08:26	
Subject Sample	# 2	0.088	08:28	
Breath Volume		2.113	08:28	
Air Blank		0.000	08:29	
Air Blank		0.000	08:31	
Subject Sample	# 3	0.080	08:34	Volume Not Met
Breath Volume		0.351	08:34	
Air Blank		0.000	08:35	
Control Test		0.077	08:35	
Air Blank		0.000	08:36	
Diagnostics Check		OK	08:36	

Data Download Date/Time: 12/13/2025 17:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: SCOZZARI,KEN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: VALENCIA PINEDA

DATE/TIME OF TEST: 12/04/2025 08:47:02

Diagnostics Check		OK	08:40
Air Blank		0.000	08:40
Control Test		0.078	08:40
Air Blank		0.000	08:41
Subject Sample	# 1	0.088	08:42
Breath Volume		1.425	08:42
Air Blank		0.000	08:42
Air Blank		0.000	08:44
Subject Sample	# 2	0.083	08:45
Breath Volume		2.636	08:45
Air Blank		0.000	08:46
Control Test		0.077	08:46
Air Blank		0.000	08:46
Diagnostics Check		OK	08:47

Data Download Date/Time: 12/13/2025 17:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WOCK,MICHAEL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: TRIPP

DATE/TIME OF TEST: 12/12/2025 22:42:59

Diagnostics Check	OK	22:42	
Air Blank		22:42	RFI Detect
Air Blank	0.000	22:42	

Data Download Date/Time: 12/13/2025 17:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WOCK,MICHAEL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: TRIPP

DATE/TIME OF TEST: 12/12/2025 22:47:49

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:45	
Control Test		0.079	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1		22:46	Subject Test Refused
Breath Volume		0.000	22:46	
Air Blank		0.000	22:46	
Control Test		0.079	22:47	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 12/13/2025 17:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 12/22/2025 03:56:25

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.113	03:52
Breath Volume		1.179	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.123	03:54
Breath Volume		1.570	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 01/24/2026 14:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BJURHOLM,JACOB,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KILIAN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 12/01/2025 03:18:03

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:15	
Control Test		0.081	03:16	
Air Blank		0.000	03:16	
Subject Sample	# 1		03:16	Subject Test Refused
Breath Volume		0.000	03:16	
Air Blank		0.000	03:17	
Control Test		0.082	03:17	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DENHARD,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: GOLD

DATE/TIME OF TEST: 12/02/2025 03:49:50

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.082	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.000	03:45
Breath Volume		3.003	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.000	03:48
Breath Volume		2.238	03:48
Air Blank		0.000	03:48
Control Test		0.082	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: AREAS

DATE/TIME OF TEST: 12/03/2025 04:33:40

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:31	
Control Test		0.082	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1		04:32	Subject Test Refused
Breath Volume		0.000	04:32	
Air Blank		0.000	04:32	
Control Test		0.082	04:33	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: AREAS

DATE/TIME OF TEST: 12/03/2025 04:45:13

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.081	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:44	
Control Test		0.081	04:44	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LONGANO,TRACY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MARLOW

DATE/TIME OF TEST: 12/04/2025 02:06:18

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.082	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.222	02:01
Breath Volume		1.730	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.220	02:04
Breath Volume		1.527	02:04
Air Blank		0.000	02:05
Control Test		0.081	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BJURHOLM,JACOB,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KILIAN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 12/05/2025 02:29:55

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.081	02:27	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:29	
Control Test		0.082	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 06:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 12/05/2025 07:27:25

Diagnostics Check		OK	07:21
Air Blank		0.000	07:21
Control Test		0.081	07:21
Air Blank		0.000	07:22
Subject Sample	# 1	0.051	07:22
Breath Volume		1.855	07:22
Air Blank		0.000	07:23
Air Blank		0.000	07:25
Subject Sample	# 2	0.050	07:25
Breath Volume		1.851	07:25
Air Blank		0.000	07:26
Control Test		0.081	07:26
Air Blank		0.000	07:27
Diagnostics Check		OK	07:27

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,CONNOR,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MACK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/07/2025 00:47:47

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.081	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:46	Subject Test Refused
Breath Volume		0.000	00:46	
Air Blank		0.000	00:46	
Control Test		0.082	00:47	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,CONNOR,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KING,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GODHO FLORES

DATE/TIME OF TEST: 12/07/2025 01:15:51

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.081	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.182	01:11
Breath Volume		2.527	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.169	01:14
Breath Volume		2.585	01:14
Air Blank		0.000	01:14
Control Test		0.081	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 12/07/2025 05:34:02

Diagnostics Check		OK	05:27
Air Blank		0.000	05:27
Control Test		0.082	05:28
Air Blank		0.000	05:28
Subject Sample	# 1	0.212	05:29
Breath Volume		2.648	05:29
Air Blank		0.000	05:30
Air Blank		0.000	05:31
Subject Sample	# 2	0.214	05:32
Breath Volume		1.476	05:32
Air Blank		0.000	05:33
Control Test		0.081	05:33
Air Blank		0.000	05:33
Diagnostics Check		OK	05:34

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSS-JOHNSON,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SEEDERS

DATE/TIME OF TEST: 12/09/2025 03:12:10

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.082	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.154	03:07
Breath Volume		2.300	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.151	03:10
Breath Volume		2.085	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BJURHOLM,JACOB,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUZUKI,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: PALENIK

DATE/TIME OF TEST: 12/09/2025 04:56:24

Diagnostics Check		OK	04:53	
Air Blank		0.000	04:53	
Control Test		0.082	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 1		04:55	Subject Test Refused
Breath Volume		0.000	04:55	
Air Blank		0.000	04:55	
Control Test		0.081	04:55	
Air Blank		0.000	04:56	
Diagnostics Check		OK	04:56	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROWBACK,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: MILIAN

DATE/TIME OF TEST: 12/10/2025 17:36:06

Diagnostics Check		OK	17:32	
Air Blank		0.000	17:33	
Control Test		0.081	17:33	
Air Blank		0.000	17:34	
Subject Sample	# 1		17:34	Subject Test Refused
Breath Volume		0.000	17:34	
Air Blank		0.000	17:35	
Control Test		0.081	17:35	
Air Blank		0.000	17:36	
Diagnostics Check		OK	17:36	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES GONZALEZ,PEDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REYES GONZALEZ,PEDRO
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 12/10/2025 22:35:03

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.082	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.000	22:30
Breath Volume		1.535	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.000	22:33
Breath Volume		2.761	22:33
Air Blank		0.000	22:34
Control Test		0.081	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:35

Data Download Date/Time: 12/17/2025 09:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 12/11/2025 22:10:05

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.081	22:04
Air Blank		0.000	22:04
Subject Sample	# 1	0.073	22:05
Breath Volume		2.605	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:07
Subject Sample	# 2	0.071	22:08
Breath Volume		2.671	22:08
Air Blank		0.000	22:09
Control Test		0.080	22:09
Air Blank		0.000	22:09
Diagnostics Check		OK	22:10

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YANICK,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: DOLCH

DATE/TIME OF TEST: 12/11/2025 22:44:18

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.081	22:38
Air Blank		0.000	22:39
Subject Sample	# 1	0.000	22:39
Breath Volume		2.484	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.000	22:42
Breath Volume		3.222	22:42
Air Blank		0.000	22:43
Control Test		0.081	22:43
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ANDERSON,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: VALKA

DATE/TIME OF TEST: 12/12/2025 02:13:54

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.356	02:06
Breath Volume		1.875	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.329	02:09
Breath Volume		1.730	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:11
Subject Sample	# 3	0.329	02:12
Breath Volume		1.472	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: POOLE,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: DEMERS

DATE/TIME OF TEST: 12/12/2025 03:48:45

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.081	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.103	03:44
Breath Volume		1.605	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.108	03:47
Breath Volume		1.285	03:47
Air Blank		0.000	03:47
Control Test		0.082	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NUSSBAUM,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/12/2025 03:57:11

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.081	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.081	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANIEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COLE,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: WHATLEY

DATE/TIME OF TEST: 12/12/2025 07:49:25

Diagnostics Check		OK	07:46	
Air Blank		0.000	07:47	
Control Test		0.081	07:47	
Air Blank		0.000	07:47	
Subject Sample	# 1		07:48	Subject Test Refused
Breath Volume		0.000	07:48	
Air Blank		0.000	07:48	
Control Test		0.081	07:48	
Air Blank		0.000	07:49	
Diagnostics Check		OK	07:49	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BONANNO,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: LIACURIS

DATE/TIME OF TEST: 12/12/2025 20:16:47

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.082	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.000	20:12
Breath Volume		1.929	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.000	20:15
Breath Volume		2.152	20:15
Air Blank		0.000	20:15
Control Test		0.082	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BELHOSTE

DATE/TIME OF TEST: 12/12/2025 21:50:12

Diagnostics Check		OK	21:43
Air Blank		0.000	21:44
Control Test		0.082	21:44
Air Blank		0.000	21:45
Subject Sample	# 1	0.000	21:45
Breath Volume		2.785	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.000	21:48
Breath Volume		2.527	21:48
Air Blank		0.000	21:49
Control Test		0.082	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHRAMM,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SMYRSKI

DATE/TIME OF TEST: 12/12/2025 23:23:14

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:20	
Control Test		0.082	23:21	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:21	Subject Test Refused
Breath Volume		0.000	23:21	
Air Blank		0.000	23:22	
Control Test		0.082	23:22	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EXTINE,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: DANIELLE

DATE/TIME OF TEST: 12/13/2025 00:03:11

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:54	
Control Test		0.082	23:54	
Air Blank		0.000	23:55	
Subject Sample	# 2		23:58	No Sample Provided
Breath Volume		0.000	23:58	
Air Blank		0.000	23:58	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.082	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LONGANO,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BORDIER

DATE/TIME OF TEST: 12/13/2025 00:20:16

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:08	
Control Test		0.082	00:08	
Air Blank		0.000	00:09	
Subject Sample	# 1	0.234	00:10	No .020 agreement
Breath Volume		1.136	00:10	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.233	00:15	Volume Not Met
Breath Volume		0.050	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 3	0.261	00:18	No .020 agreement
Breath Volume		1.199	00:18	
Air Blank		0.000	00:19	
Control Test		0.080	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LONGANO,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BORDIER

DATE/TIME OF TEST: 12/13/2025 00:42:15

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.081	00:40	
Air Blank		0.000	00:40	
Subject Sample	# 1		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:41	
Control Test		0.081	00:41	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEWS,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GARZON-GONZALEZ

DATE/TIME OF TEST: 12/14/2025 01:35:48

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.082	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.081	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NUSSBAUM,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCLENDON

DATE/TIME OF TEST: 12/15/2025 23:48:00

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.082	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.267	23:43
Breath Volume		4.074	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.269	23:46
Breath Volume		3.253	23:46
Air Blank		0.000	23:47
Control Test		0.081	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:48

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRUNETTI,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: TITTLE

DATE/TIME OF TEST: 12/16/2025 20:51:21

Diagnostics Check		OK	20:44
Air Blank		0.000	20:44
Control Test		0.082	20:45
Air Blank		0.000	20:45
Subject Sample	# 1	0.235	20:46
Breath Volume		1.125	20:46
Air Blank		0.000	20:47
Air Blank		0.000	20:48
Subject Sample	# 2	0.232	20:49
Breath Volume		1.144	20:49
Air Blank		0.000	20:50
Control Test		0.081	20:50
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CANTRELL

DATE/TIME OF TEST: 12/17/2025 02:19:57

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.082	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:19	
Control Test		0.082	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 12/17/2025 09:09

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHASE,ADDISON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ALMAZAN

DATE/TIME OF TEST: 12/18/2025 00:17:33

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.177	00:12
Breath Volume		3.382	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.183	00:16
Breath Volume		2.085	00:16
Air Blank		0.000	00:16
Control Test		0.081	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/18/2025 23:20:57

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.138	23:16
Breath Volume		1.382	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.137	23:19
Breath Volume		1.605	23:19
Air Blank		0.000	23:20
Control Test		0.081	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUZUKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BOYCE-HEAVLIN

DATE/TIME OF TEST: 12/18/2025 23:48:23

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.080	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:47	
Control Test		0.081	23:47	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BICKERTON

DATE/TIME OF TEST: 12/19/2025 04:05:24

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.082	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:04	
Control Test		0.081	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE,M,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIMBERLY,TAYLOR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: SIPE

DATE/TIME OF TEST: 12/19/2025 17:34:08

Diagnostics Check		OK	17:27
Air Blank		0.000	17:28
Control Test		0.082	17:28
Air Blank		0.000	17:29
Subject Sample	# 1	0.000	17:29
Breath Volume		1.550	17:29
Air Blank		0.000	17:30
Air Blank		0.000	17:32
Subject Sample	# 2	0.000	17:32
Breath Volume		1.945	17:32
Air Blank		0.000	17:33
Control Test		0.082	17:33
Air Blank		0.000	17:34
Diagnostics Check		OK	17:34

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAMMOCK,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: TINDALL

DATE/TIME OF TEST: 12/19/2025 19:40:19

Diagnostics Check		OK	19:32
Air Blank		0.000	19:32
Control Test		0.081	19:32
Air Blank		0.000	19:33
Subject Sample	# 1	0.289	19:35
Breath Volume		2.046	19:35
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.297	19:38
Breath Volume		1.332	19:38
Air Blank		0.000	19:39
Control Test		0.080	19:39
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: ARFUSO

DATE/TIME OF TEST: 12/19/2025 20:28:23

Diagnostics Check		OK	20:22
Air Blank		0.000	20:22
Control Test		0.081	20:22
Air Blank		0.000	20:23
Subject Sample	# 1	0.071	20:23
Breath Volume		1.726	20:23
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.069	20:26
Breath Volume		1.429	20:26
Air Blank		0.000	20:27
Control Test		0.081	20:27
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLLOWELL,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: DROSIS

DATE/TIME OF TEST: 12/19/2025 22:41:40

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.081	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.000	22:36
Breath Volume		1.566	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.000	22:40
Breath Volume		1.859	22:40
Air Blank		0.000	22:40
Control Test		0.082	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: POOLE,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LANGE

DATE/TIME OF TEST: 12/19/2025 22:54:55

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:52	
Control Test		0.081	22:52	
Air Blank		0.000	22:53	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.000	22:53	
Air Blank		0.000	22:54	
Control Test		0.082	22:54	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CASTILLO-URATAAL

DATE/TIME OF TEST: 12/21/2025 00:44:46

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.082	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.215	00:40
Breath Volume		2.324	00:40
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.200	00:43
Breath Volume		1.273	00:43
Air Blank		0.000	00:43
Control Test		0.082	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GLYNN

DATE/TIME OF TEST: 12/21/2025 03:09:45

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.082	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.211	03:04
Breath Volume		1.562	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.211	03:08
Breath Volume		2.156	03:08
Air Blank		0.000	03:08
Control Test		0.081	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: TOMASSETTI

DATE/TIME OF TEST: 12/21/2025 04:38:43

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.082	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:37	
Control Test		0.081	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIRKENDALL,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: GIORGI

DATE/TIME OF TEST: 12/22/2025 02:11:56

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.000	02:07
Breath Volume		2.675	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.000	02:10
Breath Volume		2.429	02:10
Air Blank		0.000	02:11
Control Test		0.082	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JUNE,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 12/22/2025 02:28:27

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.082	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.212	02:23
Breath Volume		1.398	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.219	02:26
Breath Volume		1.816	02:26
Air Blank		0.000	02:27
Control Test		0.082	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILSON,DESMOND
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: MC GINN

DATE/TIME OF TEST: 12/22/2025 06:13:33

Diagnostics Check		OK	06:10	
Air Blank		0.000	06:11	
Control Test		0.082	06:11	
Air Blank		0.000	06:12	
Subject Sample	# 1		06:12	Subject Test Refused
Breath Volume		0.000	06:12	
Air Blank		0.000	06:12	
Control Test		0.081	06:12	
Air Blank		0.000	06:13	
Diagnostics Check		OK	06:13	

Data Download Date/Time: 01/07/2026 09:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOATENG,DERRICK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: OTWELL

DATE/TIME OF TEST: 12/23/2025 19:30:34

Diagnostics Check		OK	19:23
Air Blank		0.000	19:24
Control Test		0.081	19:24
Air Blank		0.000	19:25
Subject Sample	# 1	0.106	19:26
Breath Volume		1.929	19:26
Air Blank		0.000	19:26
Air Blank		0.000	19:28
Subject Sample	# 2	0.105	19:29
Breath Volume		2.175	19:29
Air Blank		0.000	19:29
Control Test		0.081	19:29
Air Blank		0.000	19:30
Diagnostics Check		OK	19:30

Data Download Date/Time: 01/07/2026 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,CPL,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUZUKI,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: GAZPAR-ZACARIAS

DATE/TIME OF TEST: 12/24/2025 02:12:01

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.082	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.139	02:07
Breath Volume		1.804	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.150	02:10
Breath Volume		2.046	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 01/07/2026 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JUNE,MAX
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 12/24/2025 22:43:05

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:40	
Control Test		0.081	22:41	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:41	Subject Test Refused
Breath Volume		0.000	22:41	
Air Blank		0.000	22:42	
Control Test		0.081	22:42	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 01/07/2026 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,CONNOR,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRENNAN,CONNOR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HOCKEY

DATE/TIME OF TEST: 12/25/2025 01:40:53

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.082	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.203	01:36
Breath Volume		2.375	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.205	01:39
Breath Volume		1.582	01:39
Air Blank		0.000	01:39
Control Test		0.081	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 01/07/2026 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,CONNOR,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEDLIE,CHELSEA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KWAPIL

DATE/TIME OF TEST: 12/26/2025 03:13:57

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:11	
Control Test		0.082	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:13	
Control Test		0.082	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 01/07/2026 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,CONNOR,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RENAKER,DAVID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 12/26/2025 04:12:47

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.081	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.134	04:08
Breath Volume		2.628	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.136	04:11
Breath Volume		1.886	04:11
Air Blank		0.000	04:11
Control Test		0.082	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 01/07/2026 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: KNICKERBOCKER

DATE/TIME OF TEST: 12/27/2025 08:49:13

Diagnostics Check		OK	08:43
Air Blank		0.000	08:43
Control Test		0.081	08:43
Air Blank		0.000	08:44
Subject Sample	# 1	0.150	08:44
Breath Volume		1.429	08:44
Air Blank		0.000	08:45
Air Blank		0.000	08:47
Subject Sample	# 2	0.155	08:47
Breath Volume		1.707	08:47
Air Blank		0.000	08:48
Control Test		0.081	08:48
Air Blank		0.000	08:49
Diagnostics Check		OK	08:49

Data Download Date/Time: 01/07/2026 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KOWALSKI,BRANDON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 12/28/2025 15:23:25

Diagnostics Check		OK	15:20	
Air Blank		0.000	15:21	
Control Test		0.081	15:21	
Air Blank		0.000	15:21	
Subject Sample	# 1		15:22	Subject Test Refused
Breath Volume		0.000	15:22	
Air Blank		0.000	15:22	
Control Test		0.081	15:22	
Air Blank		0.000	15:23	
Diagnostics Check		OK	15:23	

Data Download Date/Time: 01/07/2026 09:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: HABIMANA

DATE/TIME OF TEST: 12/04/2025 00:02:51

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.080	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.227	23:58
Breath Volume		2.003	23:58
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.215	00:01
Breath Volume		1.714	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WRIGHT,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VIOTA ARCHIBALD

DATE/TIME OF TEST: 12/07/2025 01:45:34

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.305	01:41
Breath Volume		1.265	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.309	01:44
Breath Volume		2.121	01:44
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BARTMAN,BRANDT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: KISH

DATE/TIME OF TEST: 12/07/2025 20:38:10

Diagnostics Check		OK	20:35	
Air Blank		0.000	20:35	
Control Test		0.080	20:36	
Air Blank		0.000	20:36	
Subject Sample	# 1		20:36	Subject Test Refused
Breath Volume		0.000	20:36	
Air Blank		0.000	20:37	
Control Test		0.080	20:37	
Air Blank		0.000	20:38	
Diagnostics Check		OK	20:38	

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARRITT,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: KOPCZYNSKI

DATE/TIME OF TEST: 12/07/2025 22:19:39

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.079	22:12	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.204	22:15	Volume Not Met
Breath Volume		0.296	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:18	
Subject Sample	# 2		22:18	Subject Test Refused
Breath Volume		0.000	22:18	
Air Blank		0.000	22:18	
Control Test		0.078	22:19	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PHILLIPS,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: WOMRATH

DATE/TIME OF TEST: 12/12/2025 04:24:02

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.079	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 1	0.167	04:14	Volume Not Met
Breath Volume		0.085	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 2	0.196	04:17	
Breath Volume		1.371	04:17	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 3		04:22	Subject Test Refused
Breath Volume		0.121	04:22	
Air Blank		0.000	04:23	
Control Test		0.079	04:23	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CHAVARRO GONZALEZ

DATE/TIME OF TEST: 12/13/2025 03:47:04

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.170	03:42
Breath Volume		2.261	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.169	03:45
Breath Volume		1.765	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROSENBERG,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: BUI

DATE/TIME OF TEST: 12/13/2025 17:35:26

Diagnostics Check		OK	17:29
Air Blank		0.000	17:29
Control Test		0.079	17:29
Air Blank		0.000	17:30
Subject Sample	# 1	0.009	17:31
Breath Volume		2.375	17:31
Air Blank		0.000	17:31
Air Blank		0.000	17:33
Subject Sample	# 2	0.007	17:34
Breath Volume		2.289	17:34
Air Blank		0.000	17:34
Control Test		0.079	17:34
Air Blank		0.000	17:35
Diagnostics Check		OK	17:35

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORGAN,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 07:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 12/14/2025 08:53:41

Diagnostics Check		OK	08:47
Air Blank		0.000	08:47
Control Test		0.080	08:48
Air Blank		0.000	08:48
Subject Sample	# 1	0.000	08:49
Breath Volume		2.425	08:49
Air Blank		0.000	08:49
Air Blank		0.000	08:51
Subject Sample	# 2	0.000	08:52
Breath Volume		2.859	08:52
Air Blank		0.000	08:52
Control Test		0.080	08:53
Air Blank		0.000	08:53
Diagnostics Check		OK	08:53

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: FEACHER

DATE/TIME OF TEST: 12/15/2025 04:12:56

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.126	04:08
Breath Volume		1.613	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.120	04:11
Breath Volume		1.144	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 12/17/2025 09:44

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WRIGHT,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LEMAY JR

DATE/TIME OF TEST: 12/18/2025 22:25:17

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:21	
Control Test		0.080	22:21	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:24	
Control Test		0.080	22:24	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 01/07/2026 09:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 12/20/2025 04:41:23

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:25	
Control Test		0.080	04:26	
Air Blank		0.000	04:26	
Subject Sample	# 1	0.134	04:29	Volume Not Met
Breath Volume		0.429	04:29	
Air Blank		0.000	04:30	
Air Blank		0.000	04:32	
Subject Sample	# 2	0.143	04:34	
Breath Volume		2.554	04:34	
Air Blank		0.000	04:35	
Air Blank		0.000	04:36	
Subject Sample	# 3		04:39	Subject Test Refused
Breath Volume		0.035	04:39	
Air Blank		0.000	04:40	
Control Test		0.079	04:40	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: ESPARZA JR

DATE/TIME OF TEST: 12/20/2025 07:22:39

Diagnostics Check		OK	07:16
Air Blank		0.000	07:16
Control Test		0.080	07:16
Air Blank		0.000	07:17
Subject Sample	# 1	0.117	07:17
Breath Volume		1.671	07:17
Air Blank		0.000	07:18
Air Blank		0.000	07:20
Subject Sample	# 2	0.120	07:21
Breath Volume		1.589	07:21
Air Blank		0.000	07:21
Control Test		0.080	07:22
Air Blank		0.000	07:22
Diagnostics Check		OK	07:22

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: CHANCE

DATE/TIME OF TEST: 12/21/2025 06:48:06

Diagnostics Check		OK	06:40
Air Blank		0.000	06:41
Control Test		0.079	06:41
Air Blank		0.000	06:42
Subject Sample	# 1	0.111	06:42
Breath Volume		1.785	06:42
Air Blank		0.000	06:43
Air Blank		0.000	06:45
Subject Sample	# 2	0.112	06:46
Breath Volume		1.476	06:46
Air Blank		0.000	06:47
Control Test		0.080	06:47
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERGUSON,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: PAULE

DATE/TIME OF TEST: 12/21/2025 16:02:33

Diagnostics Check		OK	15:56
Air Blank		0.000	15:56
Control Test		0.078	15:57
Air Blank		0.000	15:57
Subject Sample	# 1	0.219	15:58
Breath Volume		2.042	15:58
Air Blank		0.000	15:58
Air Blank		0.000	16:00
Subject Sample	# 2	0.217	16:01
Breath Volume		1.542	16:01
Air Blank		0.000	16:01
Control Test		0.079	16:01
Air Blank		0.000	16:02
Diagnostics Check		OK	16:02

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 12/22/2025 02:00:59

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.080	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.203	01:55
Breath Volume		2.718	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.206	01:59
Breath Volume		2.378	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ALVAREZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: ERVIN

DATE/TIME OF TEST: 12/23/2025 21:57:03

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.158	21:52
Breath Volume		1.367	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.161	21:55
Breath Volume		1.234	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:56
Diagnostics Check		OK	21:57

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 12/24/2025 23:07:58

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.080	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.137	23:03
Breath Volume		1.773	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.129	23:06
Breath Volume		1.613	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SEVOSTJANOV,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 12/25/2025 00:24:30

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.011	00:20
Breath Volume		3.437	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.010	00:23
Breath Volume		3.445	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIRKENDALL,JESSIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: AVILA PERDOMO

DATE/TIME OF TEST: 12/26/2025 02:43:25

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:35	No Sample Provided
Breath Volume		0.000	02:35	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.152	02:38	
Breath Volume		1.378	02:38	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 3	0.157	02:41	
Breath Volume		2.628	02:41	
Air Blank		0.000	02:42	
Control Test		0.080	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEWS,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: TIFFT

DATE/TIME OF TEST: 12/27/2025 02:50:23

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.141	02:45
Breath Volume		3.070	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.142	02:48
Breath Volume		3.136	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,ALBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: OLVERA ALCANTARA

DATE/TIME OF TEST: 12/28/2025 02:50:19

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.080	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.226	02:45
Breath Volume		1.457	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.235	02:48
Breath Volume		2.816	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 01/07/2026 09:53

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: LEAL HERNANDEZ

DATE/TIME OF TEST: 12/04/2025 09:46:19

Diagnostics Check		OK	09:40	
Air Blank		0.000	09:41	
Control Test		0.081	09:41	
Air Blank		0.000	09:42	
Subject Sample	# 1	0.000	09:42	
Breath Volume		1.285	09:42	
Air Blank		0.000	09:43	
Air Blank			09:44	RFI Detect
Air Blank		0.000	09:45	
Control Test		0.081	09:45	
Air Blank		0.000	09:46	
Diagnostics Check		OK	09:46	

Data Download Date/Time: 12/05/2025 16:55

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: LEAL HERNANDEZ

DATE/TIME OF TEST: 12/04/2025 09:54:35

Diagnostics Check		OK	09:48
Air Blank		0.000	09:49
Control Test		0.080	09:49
Air Blank		0.000	09:49
Subject Sample	# 1	0.000	09:50
Breath Volume		1.343	09:50
Air Blank		0.000	09:50
Air Blank		0.000	09:52
Subject Sample	# 2	0.000	09:53
Breath Volume		1.417	09:53
Air Blank		0.000	09:53
Control Test		0.079	09:53
Air Blank		0.000	09:54
Diagnostics Check		OK	09:54

Data Download Date/Time: 12/05/2025 16:55

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: ALFRED

DATE/TIME OF TEST: 12/04/2025 10:12:12

Diagnostics Check		OK	10:06
Air Blank		0.000	10:06
Control Test		0.080	10:06
Air Blank		0.000	10:07
Subject Sample	# 1	0.000	10:07
Breath Volume		1.156	10:07
Air Blank		0.000	10:08
Air Blank		0.000	10:10
Subject Sample	# 2	0.000	10:10
Breath Volume		1.757	10:10
Air Blank		0.000	10:11
Control Test		0.080	10:11
Air Blank		0.000	10:12
Diagnostics Check		OK	10:12

Data Download Date/Time: 12/05/2025 16:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: SAUCEDO TELLO

DATE/TIME OF TEST: 12/04/2025 15:32:41

Diagnostics Check		OK	15:26
Air Blank		0.000	15:27
Control Test		0.080	15:27
Air Blank		0.000	15:28
Subject Sample	# 1	0.000	15:28
Breath Volume		1.390	15:28
Air Blank		0.000	15:28
Air Blank		0.000	15:30
Subject Sample	# 2	0.000	15:31
Breath Volume		1.664	15:31
Air Blank		0.000	15:31
Control Test		0.080	15:32
Air Blank		0.000	15:32
Diagnostics Check		OK	15:32

Data Download Date/Time: 12/05/2025 16:55

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,C
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 15:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 12/04/2025 16:32:35

Diagnostics Check		OK	16:24
Air Blank		0.000	16:24
Control Test		0.081	16:25
Air Blank		0.000	16:25
Subject Sample	# 1	0.000	16:28
Breath Volume		1.203	16:28
Air Blank		0.000	16:28
Air Blank		0.000	16:30
Subject Sample	# 2	0.000	16:31
Breath Volume		2.160	16:31
Air Blank		0.000	16:31
Control Test		0.078	16:31
Air Blank		0.000	16:32
Diagnostics Check		OK	16:32

Data Download Date/Time: 12/05/2025 16:55

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: CRUZ RODRIGUEZ

DATE/TIME OF TEST: 12/05/2025 18:03:47

Diagnostics Check		OK	18:00	
Air Blank		0.000	18:01	
Control Test		0.080	18:01	
Air Blank		0.000	18:02	
Subject Sample	# 1		18:02	Subject Test Refused
Breath Volume		0.000	18:02	
Air Blank		0.000	18:02	
Control Test		0.080	18:03	
Air Blank		0.000	18:03	
Diagnostics Check		OK	18:03	

Data Download Date/Time: 01/07/2026 18:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: VEGA VEGA

DATE/TIME OF TEST: 12/05/2025 21:44:29

Diagnostics Check		OK	21:37
Air Blank		0.000	21:37
Control Test		0.080	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.209	21:40
Breath Volume		1.351	21:40
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.199	21:42
Breath Volume		1.195	21:42
Air Blank		0.000	21:43
Control Test		0.078	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 01/07/2026 18:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/06/2025 03:47:08

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.139	03:42
Breath Volume		1.968	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.147	03:45
Breath Volume		1.914	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 01/07/2026 18:12

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: IRVIN

DATE/TIME OF TEST: 12/06/2025 04:31:34

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.074	04:26
Breath Volume		1.273	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.069	04:30
Breath Volume		1.121	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: FLAR

DATE/TIME OF TEST: 12/07/2025 22:59:06

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.080	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.018	22:54
Breath Volume		1.828	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.016	22:57
Breath Volume		1.800	22:57
Air Blank		0.000	22:58
Control Test		0.079	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JOHNSON,GREGORY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: REYNOSO

DATE/TIME OF TEST: 12/07/2025 23:07:56

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.200	23:03
Breath Volume		1.367	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.199	23:06
Breath Volume		1.730	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/08/2025 08:45:27

Diagnostics Check		OK	08:39
Air Blank		0.000	08:40
Control Test		0.080	08:40
Air Blank		0.000	08:40
Subject Sample	# 1	0.000	08:41
Breath Volume		2.273	08:41
Air Blank		0.000	08:41
Air Blank		0.000	08:43
Subject Sample	# 2	0.000	08:44
Breath Volume		1.980	08:44
Air Blank		0.000	08:44
Control Test		0.079	08:44
Air Blank		0.000	08:45
Diagnostics Check		OK	08:45

Data Download Date/Time: 01/07/2026 18:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 12/11/2025 14:13:08

Diagnostics Check		OK	14:07
Air Blank		0.000	14:07
Control Test		0.080	14:07
Air Blank		0.000	14:08
Subject Sample	# 1	0.000	14:08
Breath Volume		1.585	14:08
Air Blank		0.000	14:09
Air Blank		0.000	14:11
Subject Sample	# 2	0.000	14:11
Breath Volume		1.308	14:11
Air Blank		0.000	14:12
Control Test		0.079	14:12
Air Blank		0.000	14:13
Diagnostics Check		OK	14:13

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: COIL

DATE/TIME OF TEST: 12/11/2025 14:57:06

Diagnostics Check		OK	14:51
Air Blank		0.000	14:51
Control Test		0.079	14:51
Air Blank		0.000	14:52
Subject Sample	# 1	0.000	14:52
Breath Volume		1.777	14:52
Air Blank		0.000	14:53
Air Blank		0.000	14:55
Subject Sample	# 2	0.000	14:55
Breath Volume		1.800	14:55
Air Blank		0.000	14:56
Control Test		0.079	14:56
Air Blank		0.000	14:57
Diagnostics Check		OK	14:57

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 15:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: RIVERA PENA

DATE/TIME OF TEST: 12/11/2025 15:46:40

Diagnostics Check		OK	15:40
Air Blank		0.000	15:40
Control Test		0.080	15:41
Air Blank		0.000	15:41
Subject Sample	# 1	0.000	15:42
Breath Volume		1.636	15:42
Air Blank		0.000	15:42
Air Blank		0.000	15:44
Subject Sample	# 2	0.000	15:45
Breath Volume		2.636	15:45
Air Blank		0.000	15:45
Control Test		0.079	15:46
Air Blank		0.000	15:46
Diagnostics Check		OK	15:46

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 15:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: RIVERA PENA

DATE/TIME OF TEST: 12/11/2025 15:56:26

Diagnostics Check		OK	15:50
Air Blank		0.000	15:50
Control Test		0.080	15:51
Air Blank		0.000	15:51
Subject Sample	# 1	0.000	15:52
Breath Volume		2.429	15:52
Air Blank		0.000	15:52
Air Blank		0.000	15:54
Subject Sample	# 2	0.000	15:55
Breath Volume		2.820	15:55
Air Blank		0.000	15:55
Control Test		0.079	15:55
Air Blank		0.000	15:56
Diagnostics Check		OK	15:56

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: PROB/PAROL
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 12/11/2025 16:41:53

Diagnostics Check		OK	16:35
Air Blank		0.000	16:36
Control Test		0.080	16:36
Air Blank		0.000	16:37
Subject Sample	# 1	0.000	16:37
Breath Volume		1.660	16:37
Air Blank		0.000	16:38
Air Blank		0.000	16:40
Subject Sample	# 2	0.000	16:40
Breath Volume		1.734	16:40
Air Blank		0.000	16:40
Control Test		0.079	16:41
Air Blank		0.000	16:41
Diagnostics Check		OK	16:41

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RORDIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RORIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:57

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 12/12/2025 15:25:09

Diagnostics Check		OK	15:19
Air Blank		0.000	15:19
Control Test		0.080	15:20
Air Blank		0.000	15:20
Subject Sample	# 1	0.000	15:20
Breath Volume		1.519	15:20
Air Blank		0.000	15:21
Air Blank		0.000	15:23
Subject Sample	# 2	0.000	15:23
Breath Volume		1.867	15:23
Air Blank		0.000	15:24
Control Test		0.078	15:24
Air Blank		0.000	15:25
Diagnostics Check		OK	15:25

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 12/17/2025 12:53:56

Diagnostics Check		OK	12:48
Air Blank		0.000	12:48
Control Test		0.080	12:48
Air Blank		0.000	12:49
Subject Sample	# 1	0.000	12:49
Breath Volume		1.449	12:49
Air Blank		0.000	12:50
Air Blank		0.000	12:52
Subject Sample	# 2	0.000	12:52
Breath Volume		1.355	12:52
Air Blank		0.000	12:53
Control Test		0.080	12:53
Air Blank		0.000	12:53
Diagnostics Check		OK	12:53

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 12/19/2025 21:54:12

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:51	
Control Test		0.080	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 1		21:52	Subject Test Refused
Breath Volume		0.000	21:52	
Air Blank		0.000	21:53	
Control Test		0.080	21:53	
Air Blank		0.000	21:54	
Diagnostics Check		OK	21:54	

Data Download Date/Time: 01/07/2026 18:13

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 12/21/2025 00:17:16

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.081	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.061	00:12
Breath Volume		1.292	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.064	00:15
Breath Volume		1.667	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 01/07/2026 18:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: LOBATO

DATE/TIME OF TEST: 12/21/2025 04:37:19

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.168	04:32
Breath Volume		1.972	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.171	04:35
Breath Volume		1.339	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 01/07/2026 18:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SCALF

DATE/TIME OF TEST: 12/21/2025 23:37:03

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.079	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.130	23:32
Breath Volume		1.324	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.133	23:35
Breath Volume		1.472	23:35
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:37

Data Download Date/Time: 01/07/2026 18:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 12/25/2025 02:16:52

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:02	
Control Test		0.081	02:03	
Air Blank		0.000	02:03	
Subject Sample	# 1	0.094	02:06	Volume Not Met
Breath Volume		0.480	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:09	
Subject Sample	# 2	0.096	02:11	
Breath Volume		1.468	02:11	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 3	0.098	02:15	
Breath Volume		1.363	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 01/07/2026 18:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: VEGA-CHAVEZ

DATE/TIME OF TEST: 12/25/2025 03:42:18

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.080	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.082	03:37
Breath Volume		1.460	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.080	03:40
Breath Volume		2.101	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 01/07/2026 18:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: GONZALEZ MAISONET

DATE/TIME OF TEST: 12/25/2025 04:09:18

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:06	
Control Test		0.080	04:07	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 01/07/2026 18:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: MUTON RAMIREZ

DATE/TIME OF TEST: 12/25/2025 05:51:26

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.081	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.147	05:47
Breath Volume		1.871	05:47
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.152	05:49
Breath Volume		1.292	05:49
Air Blank		0.000	05:50
Control Test		0.079	05:50
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 01/07/2026 18:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: GUERRERO DAZA

DATE/TIME OF TEST: 12/25/2025 21:47:16

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.079	21:41
Air Blank		0.000	21:42
Subject Sample	# 1	0.196	21:42
Breath Volume		1.855	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.184	21:45
Breath Volume		1.789	21:45
Air Blank		0.000	21:46
Control Test		0.079	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 01/07/2026 18:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ESPINOSA LOPEZ

DATE/TIME OF TEST: 12/26/2025 03:00:52

Diagnostics Check	OK	02:59	
Air Blank		03:00	RFI Detect
Air Blank	0.000	03:00	

Data Download Date/Time: 01/07/2026 18:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: OBRIEN,ROBERT
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LAUREANO NAVARRO

DATE/TIME OF TEST: 12/28/2025 01:31:05

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:19	
Control Test		0.080	01:20	
Air Blank		0.000	01:20	
Subject Sample	# 1	0.140	01:23	Volume Not Met
Breath Volume		0.488	01:23	
Air Blank		0.000	01:24	
Air Blank		0.000	01:26	
Subject Sample	# 2	0.115	01:29	Volume Not Met
Breath Volume		0.429	01:29	
Air Blank		0.000	01:30	
Control Test		0.079	01:30	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 01/07/2026 18:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RIVEROS,REY
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/05/2025 00:45:22

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.102	00:40
Breath Volume		1.324	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.102	00:43
Breath Volume		1.937	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 12/27/2025 07:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CRISSEY,SCOTT
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BARAN

DATE/TIME OF TEST: 12/06/2025 01:21:46

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.216	01:17
Breath Volume		2.066	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:19
Subject Sample	# 2	0.227	01:20
Breath Volume		2.332	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 12/27/2025 07:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CRISSEY,SCOTT
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BARAN

DATE/TIME OF TEST: 12/06/2025 02:00:06

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.217	01:55
Breath Volume		2.136	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.211	01:58
Breath Volume		1.859	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 12/27/2025 07:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDOUX,JACOB,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEDOUX,JACOB
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/15/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BARBOSA

DATE/TIME OF TEST: 12/17/2025 02:13:43

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.082	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.105	02:09
Breath Volume		2.832	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.111	02:12
Breath Volume		1.855	02:12
Air Blank		0.000	02:12
Control Test		0.080	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 12/27/2025 07:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,JASON,B
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KING,STEPHEN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: PESKAR

DATE/TIME OF TEST: 12/19/2025 13:07:18

Diagnostics Check		OK	13:01
Air Blank		0.000	13:01
Control Test		0.080	13:02
Air Blank		0.000	13:02
Subject Sample	# 1	0.000	13:02
Breath Volume		2.644	13:02
Air Blank		0.000	13:03
Air Blank		0.000	13:05
Subject Sample	# 2	0.000	13:05
Breath Volume		3.914	13:05
Air Blank		0.000	13:06
Control Test		0.080	13:06
Air Blank		0.000	13:07
Diagnostics Check		OK	13:07

Data Download Date/Time: 12/27/2025 07:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: TELCY

DATE/TIME OF TEST: 12/22/2025 23:29:42

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.081	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.140	23:24
Breath Volume		2.164	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.134	23:28
Breath Volume		1.937	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 12/27/2025 07:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 12/01/2025 06:20:23

Diagnostics Check		OK	06:17	
Air Blank		0.000	06:17	
Control Test		0.079	06:18	
Air Blank		0.000	06:18	
Subject Sample	# 1		06:18	Subject Test Refused
Breath Volume		0.000	06:18	
Air Blank		0.000	06:19	
Control Test		0.079	06:19	
Air Blank		0.000	06:20	
Diagnostics Check		OK	06:20	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,T,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: QUINONEZ

DATE/TIME OF TEST: 12/02/2025 00:55:04

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.079	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:54	
Control Test		0.079	00:54	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 12/03/2025 22:00:19

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:57	
Control Test		0.079	21:58	
Air Blank		0.000	21:58	
Subject Sample	# 1		21:58	Subject Test Refused
Breath Volume		0.000	21:58	
Air Blank		0.000	21:59	
Control Test		0.079	21:59	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/05/2025 04:22:13

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.254	04:17
Breath Volume		2.199	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:19
Subject Sample	# 2	0.245	04:20
Breath Volume		2.695	04:20
Air Blank		0.000	04:21
Control Test		0.077	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CINOTTI,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 12/06/2025 00:05:44

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:03	
Control Test		0.079	00:03	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:04	Subject Test Refused
Breath Volume		0.000	00:04	
Air Blank		0.000	00:04	
Control Test		0.079	00:05	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON JR.,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 12/06/2025 02:46:29

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.079	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:42	No Sample Provided
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Air Blank		0.000	02:44	
Subject Sample	# 2		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.078	02:45	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 12/28/2025 17:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CINOTTI,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: L605546213000

DATE/TIME OF TEST: 12/06/2025 04:48:10

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.079	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.116	04:42
Breath Volume		1.417	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.107	04:46
Breath Volume		1.761	04:46
Air Blank		0.000	04:47
Control Test		0.077	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JORDAN,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/06/2025 05:49:54

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.077	05:43
Air Blank		0.000	05:44
Subject Sample	# 1	0.189	05:45
Breath Volume		2.824	05:45
Air Blank		0.000	05:45
Air Blank		0.000	05:47
Subject Sample	# 2	0.193	05:48
Breath Volume		3.707	05:48
Air Blank		0.000	05:48
Control Test		0.077	05:49
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 12/28/2025 17:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 12/06/2025 16:54:20

Diagnostics Check		OK	16:48
Air Blank		0.000	16:48
Control Test		0.079	16:48
Air Blank		0.000	16:49
Subject Sample	# 1	0.342	16:49
Breath Volume		1.992	16:49
Air Blank		0.000	16:50
Air Blank		0.000	16:52
Subject Sample	# 2	0.324	16:52
Breath Volume		2.121	16:52
Air Blank		0.000	16:53
Control Test		0.078	16:53
Air Blank		0.000	16:54
Diagnostics Check		OK	16:54

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: HIGGINBOTHAM

DATE/TIME OF TEST: 12/06/2025 19:49:18

Diagnostics Check		OK	19:42
Air Blank		0.000	19:43
Control Test		0.078	19:43
Air Blank		0.000	19:44
Subject Sample	# 1	0.190	19:44
Breath Volume		1.675	19:44
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.188	19:47
Breath Volume		1.601	19:47
Air Blank		0.000	19:48
Control Test		0.078	19:48
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 12/06/2025 23:49:08

Diagnostics Check		OK	23:42
Air Blank		0.000	23:42
Control Test		0.078	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.121	23:44
Breath Volume		1.621	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:46
Subject Sample	# 2	0.120	23:47
Breath Volume		1.214	23:47
Air Blank		0.000	23:48
Control Test		0.077	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PEPPER

DATE/TIME OF TEST: 12/07/2025 02:18:24

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.078	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.078	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOPKINS,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 12/07/2025 04:44:49

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.078	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.243	04:39
Breath Volume		1.437	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:41
Subject Sample	# 2	0.231	04:43
Breath Volume		1.492	04:43
Air Blank		0.000	04:43
Control Test		0.077	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: GHAZNAWI

DATE/TIME OF TEST: 12/07/2025 21:08:30

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.078	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.155	21:03
Breath Volume		1.355	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.162	21:06
Breath Volume		1.421	21:06
Air Blank		0.000	21:07
Control Test		0.077	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: POLKE

DATE/TIME OF TEST: 12/09/2025 04:26:27

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.178	04:21
Breath Volume		1.488	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.169	04:24
Breath Volume		1.332	04:24
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 12/09/2025 20:41:24

Diagnostics Check		OK	20:38	
Air Blank		0.000	20:38	
Control Test		0.079	20:39	
Air Blank		0.000	20:39	
Subject Sample	# 1		20:39	Subject Test Refused
Breath Volume		0.000	20:39	
Air Blank		0.000	20:40	
Control Test		0.079	20:40	
Air Blank		0.000	20:41	
Diagnostics Check		OK	20:41	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,H,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: ZEHNER

DATE/TIME OF TEST: 12/09/2025 22:38:11

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.080	22:36	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:37	
Control Test		0.080	22:37	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/10/2025 00:06:05

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:55	
Control Test		0.079	23:55	
Air Blank		0.000	23:55	
Subject Sample	# 2		23:59	No Sample Provided
Breath Volume		0.000	23:59	
Air Blank		0.000	23:59	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:04	No Sample Provided
Breath Volume		0.000	00:04	
Air Blank		0.000	00:05	
Control Test		0.079	00:05	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/10/2025 00:19:59

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.079	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1		00:18	Subject Test Refused
Breath Volume		0.000	00:18	
Air Blank		0.000	00:19	
Control Test		0.080	00:19	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 12/10/2025 01:05:07

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:03	Subject Test Refused
Breath Volume		0.000	01:03	
Air Blank		0.000	01:04	
Control Test		0.079	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MCCLENDON

DATE/TIME OF TEST: 12/10/2025 03:29:11

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.079	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.142	03:23
Breath Volume		1.511	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.144	03:27
Breath Volume		1.500	03:27
Air Blank		0.000	03:28
Control Test		0.077	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 12/10/2025 18:32:27

Diagnostics Check		OK	18:26
Air Blank		0.000	18:26
Control Test		0.079	18:26
Air Blank		0.000	18:27
Subject Sample	# 1	0.114	18:27
Breath Volume		1.652	18:27
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.113	18:30
Breath Volume		1.488	18:30
Air Blank		0.000	18:31
Control Test		0.079	18:31
Air Blank		0.000	18:32
Diagnostics Check		OK	18:32

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M.F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/11/2025 03:58:30

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.191	03:53
Breath Volume		2.300	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.198	03:56
Breath Volume		2.281	03:56
Air Blank		0.000	03:57
Control Test		0.076	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 12/28/2025 17:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M.F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: VANCSEK

DATE/TIME OF TEST: 12/12/2025 01:37:13

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.081	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:35	Subject Test Refused
Breath Volume		0.000	01:35	
Air Blank		0.000	01:36	
Control Test		0.079	01:36	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/12/2025 21:16:31

Diagnostics Check		OK	21:13	
Air Blank		0.000	21:13	
Control Test		0.079	21:14	
Air Blank		0.000	21:14	
Subject Sample	# 1		21:15	Subject Test Refused
Breath Volume		0.000	21:15	
Air Blank		0.000	21:15	
Control Test		0.080	21:15	
Air Blank		0.000	21:16	
Diagnostics Check		OK	21:16	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BOYNTON

DATE/TIME OF TEST: 12/13/2025 01:22:39

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:20	
Control Test		0.079	01:20	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:21	Subject Test Refused
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:22	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ELLISON,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: HERRIN

DATE/TIME OF TEST: 12/13/2025 03:19:29

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.093	03:14
Breath Volume		1.171	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.090	03:17
Breath Volume		1.382	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/13/2025 04:32:41

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.214	04:27
Breath Volume		2.777	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.199	04:30
Breath Volume		1.199	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORCHARDT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 12/13/2025 09:47:47

Diagnostics Check		OK	09:41
Air Blank		0.000	09:41
Control Test		0.080	09:41
Air Blank		0.000	09:42
Subject Sample	# 1	0.164	09:43
Breath Volume		1.175	09:43
Air Blank		0.000	09:43
Air Blank		0.000	09:45
Subject Sample	# 2	0.165	09:46
Breath Volume		1.648	09:46
Air Blank		0.000	09:46
Control Test		0.078	09:47
Air Blank		0.000	09:47
Diagnostics Check		OK	09:47

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMSON,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: WORTHY

DATE/TIME OF TEST: 12/13/2025 14:06:19

Diagnostics Check		OK	14:03	
Air Blank		0.000	14:03	
Control Test		0.078	14:04	
Air Blank		0.000	14:04	
Subject Sample	# 1		14:04	Subject Test Refused
Breath Volume		0.000	14:04	
Air Blank		0.000	14:05	
Control Test		0.079	14:05	
Air Blank		0.000	14:06	
Diagnostics Check		OK	14:06	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 12/13/2025 18:36:01

Diagnostics Check		OK	18:28
Air Blank		0.000	18:29
Control Test		0.079	18:29
Air Blank		0.000	18:30
Subject Sample	# 1	0.168	18:30
Breath Volume		1.343	18:30
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.172	18:34
Breath Volume		1.476	18:34
Air Blank		0.000	18:35
Control Test		0.078	18:35
Air Blank		0.000	18:35
Diagnostics Check		OK	18:36

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUHAS,SAMANTHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,D
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: DUGAN

DATE/TIME OF TEST: 12/13/2025 23:07:31

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:04	
Control Test		0.079	23:05	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:06	Subject Test Refused
Breath Volume		0.000	23:06	
Air Blank		0.000	23:06	
Control Test		0.079	23:06	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: STOCKTON

DATE/TIME OF TEST: 12/14/2025 01:27:19

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.079	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Control Test		0.078	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: NEVES GUIMARAES JORG

DATE/TIME OF TEST: 12/14/2025 03:45:30

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.079	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.079	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: RAGLAND

DATE/TIME OF TEST: 12/14/2025 04:18:06

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.131	04:13
Breath Volume		1.125	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.127	04:16
Breath Volume		1.406	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RICHARDSON,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/14/2025 16:27:54

Diagnostics Check		OK	16:17	
Air Blank		0.000	16:18	
Control Test		0.079	16:18	
Air Blank		0.000	16:19	
Subject Sample	# 1	0.377	16:19	No .020 agreement
Breath Volume		1.851	16:19	
Air Blank		0.000	16:20	
Air Blank		0.000	16:22	
Subject Sample	# 2	0.346	16:22	No .020 agreement
Breath Volume		2.117	16:22	
Air Blank		0.000	16:23	
Air Blank		0.000	16:24	
Subject Sample	# 3	0.322	16:26	No .020 agreement
Breath Volume		1.304	16:26	
Air Blank		0.000	16:26	
Control Test		0.078	16:27	
Air Blank		0.000	16:27	
Diagnostics Check		OK	16:27	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,V,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HOLLING

DATE/TIME OF TEST: 12/14/2025 23:57:25

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.079	23:54	
Air Blank		0.000	23:54	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:56	
Control Test		0.080	23:56	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,V,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A.W.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 12/15/2025 00:50:44

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:47	
Control Test		0.080	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.080	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:38

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 12/15/2025 12:05:10

Diagnostics Check		OK	11:58
Air Blank		0.000	11:59
Control Test		0.080	11:59
Air Blank		0.000	12:00
Subject Sample	# 1	0.204	12:00
Breath Volume		2.050	12:00
Air Blank		0.000	12:01
Air Blank		0.000	12:03
Subject Sample	# 2	0.190	12:03
Breath Volume		1.140	12:03
Air Blank		0.000	12:04
Control Test		0.079	12:04
Air Blank		0.000	12:05
Diagnostics Check		OK	12:05

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: BOSTICK

DATE/TIME OF TEST: 12/17/2025 15:36:23

Diagnostics Check		OK	15:33	
Air Blank		0.000	15:33	
Control Test		0.079	15:34	
Air Blank		0.000	15:34	
Subject Sample	# 1		15:34	Subject Test Refused
Breath Volume		0.000	15:34	
Air Blank		0.000	15:35	
Control Test		0.078	15:35	
Air Blank		0.000	15:36	
Diagnostics Check		OK	15:36	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHAW,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 12/17/2025 21:44:17

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.080	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.000	21:40
Breath Volume		1.210	21:40
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.000	21:42
Breath Volume		1.117	21:42
Air Blank		0.000	21:43
Control Test		0.079	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 12/18/2025 22:16:52

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.078	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.107	22:12
Breath Volume		1.492	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.109	22:15
Breath Volume		3.695	22:15
Air Blank		0.000	22:15
Control Test		0.078	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: WOODSON

DATE/TIME OF TEST: 12/19/2025 19:35:54

Diagnostics Check		OK	19:32	
Air Blank		0.000	19:33	
Control Test		0.078	19:33	
Air Blank		0.000	19:34	
Subject Sample	# 1		19:34	Subject Test Refused
Breath Volume		0.000	19:34	
Air Blank		0.000	19:34	
Control Test		0.079	19:35	
Air Blank		0.000	19:35	
Diagnostics Check		OK	19:35	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: KHOUGALI

DATE/TIME OF TEST: 12/19/2025 22:24:25

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.079	22:17
Air Blank		0.000	22:18
Subject Sample	# 1	0.129	22:19
Breath Volume		1.468	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.128	22:22
Breath Volume		1.566	22:22
Air Blank		0.000	22:23
Control Test		0.078	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/20/2025 00:10:07

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:07	
Control Test		0.080	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 1		00:08	Subject Test Refused
Breath Volume		0.000	00:08	
Air Blank		0.000	00:09	
Control Test		0.079	00:09	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/20/2025 01:53:44

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.136	01:48
Breath Volume		1.312	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.127	01:52
Breath Volume		1.351	01:52
Air Blank		0.000	01:52
Control Test		0.079	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HODGE MOORE

DATE/TIME OF TEST: 12/20/2025 04:19:09

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.159	04:14
Breath Volume		2.171	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:16
Subject Sample	# 2	0.160	04:17
Breath Volume		1.582	04:17
Air Blank		0.000	04:18
Control Test		0.078	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/20/2025 22:27:04

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.079	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.000	22:22
Breath Volume		1.511	22:22
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.000	22:25
Breath Volume		2.171	22:25
Air Blank		0.000	22:26
Control Test		0.079	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:27

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: BORGAN

DATE/TIME OF TEST: 12/21/2025 02:24:52

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.079	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.019	02:20
Breath Volume		1.414	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.019	02:23
Breath Volume		1.941	02:23
Air Blank		0.000	02:23
Control Test		0.079	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 12/28/2025 17:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BOLDEN

DATE/TIME OF TEST: 12/21/2025 04:16:39

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Control Test		0.079	04:16	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 12/28/2025 17:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSBORN,SHANE,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ORTAGUS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: HIGHLAND

DATE/TIME OF TEST: 12/21/2025 17:40:25

Diagnostics Check		OK	17:37	
Air Blank		0.000	17:37	
Control Test		0.079	17:38	
Air Blank		0.000	17:38	
Subject Sample	# 1		17:39	Subject Test Refused
Breath Volume		0.000	17:39	
Air Blank		0.000	17:39	
Control Test		0.079	17:39	
Air Blank		0.000	17:40	
Diagnostics Check		OK	17:40	

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: CRUSAN

DATE/TIME OF TEST: 12/21/2025 19:46:28

Diagnostics Check		OK	19:40
Air Blank		0.000	19:40
Control Test		0.080	19:40
Air Blank		0.000	19:41
Subject Sample	# 1	0.096	19:41
Breath Volume		1.609	19:41
Air Blank		0.000	19:42
Air Blank		0.000	19:44
Subject Sample	# 2	0.093	19:44
Breath Volume		1.234	19:44
Air Blank		0.000	19:45
Control Test		0.079	19:45
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/22/2025 05:52:08

Diagnostics Check		OK	05:49	
Air Blank		0.000	05:49	
Control Test		0.079	05:49	
Air Blank		0.000	05:50	
Subject Sample	# 1		05:50	Subject Test Refused
Breath Volume		0.000	05:50	
Air Blank		0.000	05:51	
Control Test		0.080	05:51	
Air Blank		0.000	05:52	
Diagnostics Check		OK	05:52	

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCCRARY,D
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: DENHAM

DATE/TIME OF TEST: 12/22/2025 22:00:17

Diagnostics Check		OK	21:54
Air Blank		0.000	21:54
Control Test		0.079	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.041	21:55
Breath Volume		1.167	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.039	21:58
Breath Volume		1.238	21:58
Air Blank		0.000	21:59
Control Test		0.079	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 12/23/2025 05:11:29

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.203	05:06
Breath Volume		1.464	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:08
Subject Sample	# 2	0.201	05:09
Breath Volume		1.468	05:09
Air Blank		0.000	05:10
Control Test		0.077	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PALLAIS

DATE/TIME OF TEST: 12/24/2025 02:33:31

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.224	02:28
Breath Volume		2.601	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.205	02:31
Breath Volume		2.781	02:31
Air Blank		0.000	02:32
Control Test		0.078	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: LOPEZ AGUILAR

DATE/TIME OF TEST: 12/24/2025 21:16:43

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.078	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.200	21:12
Breath Volume		1.757	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.188	21:15
Breath Volume		1.511	21:15
Air Blank		0.000	21:15
Control Test		0.077	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHAW,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 12/24/2025 22:46:46

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.049	22:41
Breath Volume		1.210	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.049	22:45
Breath Volume		1.246	22:45
Air Blank		0.000	22:45
Control Test		0.078	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUHAS,SAMANTHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/25/2025 02:47:44

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:45	
Control Test		0.078	02:45	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:46	Subject Test Refused
Breath Volume		0.000	02:46	
Air Blank		0.000	02:46	
Control Test		0.078	02:47	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WU,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: VICTORIANO

DATE/TIME OF TEST: 12/25/2025 05:03:55

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.064	04:59
Breath Volume		1.398	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.067	05:02
Breath Volume		1.851	05:02
Air Blank		0.000	05:02
Control Test		0.078	05:03
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILEY,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: JASTINCARRIO

DATE/TIME OF TEST: 12/25/2025 22:34:44

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.079	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.160	22:30
Breath Volume		1.492	22:30
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.161	22:33
Breath Volume		1.625	22:33
Air Blank		0.000	22:33
Control Test		0.078	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUHAS,SAMANTHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: JETTE

DATE/TIME OF TEST: 12/25/2025 23:41:06

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.078	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1	0.109	23:36	Volume Not Met
Breath Volume		0.035	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Control Test		0.076	23:40	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRAY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/26/2025 09:52:34

Diagnostics Check		OK	09:45
Air Blank		0.000	09:46
Control Test		0.079	09:46
Air Blank		0.000	09:47
Subject Sample	# 1	0.110	09:48
Breath Volume		1.460	09:48
Air Blank		0.000	09:48
Air Blank		0.000	09:50
Subject Sample	# 2	0.116	09:50
Breath Volume		1.304	09:50
Air Blank		0.000	09:51
Control Test		0.078	09:51
Air Blank		0.000	09:52
Diagnostics Check		OK	09:52

Data Download Date/Time: 12/28/2025 17:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUHAS,SAMANTHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOOD,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 12/27/2025 06:05:49

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.078	06:00
Air Blank		0.000	06:00
Subject Sample	# 1	0.196	06:01
Breath Volume		1.277	06:01
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.189	06:04
Breath Volume		1.726	06:04
Air Blank		0.000	06:04
Control Test		0.077	06:05
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/27/2025 21:32:28

Diagnostics Check		OK	21:23
Air Blank		0.000	21:23
Control Test		0.078	21:24
Air Blank		0.000	21:24
Subject Sample	# 1	0.151	21:26
Breath Volume		1.726	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.147	21:30
Breath Volume		1.160	21:30
Air Blank		0.000	21:31
Control Test		0.077	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: SIMMS

DATE/TIME OF TEST: 12/28/2025 00:13:40

Diagnostics Check		OK	00:06
Air Blank		0.000	00:07
Control Test		0.078	00:07
Air Blank		0.000	00:08
Subject Sample	# 1	0.092	00:08
Breath Volume		1.500	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.082	00:12
Breath Volume		1.292	00:12
Air Blank		0.000	00:12
Control Test		0.078	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 12/28/2025 17:26

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 12/28/2025 02:40:36

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.149	02:35
Breath Volume		2.085	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.145	02:38
Breath Volume		2.296	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 12/28/2025 17:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MARCOTTE

DATE/TIME OF TEST: 12/29/2025 23:35:12

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.000	23:30
Breath Volume		1.550	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.000	23:33
Breath Volume		3.777	23:33
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 01/08/2026 10:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VONESCHE,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/30/2025 04:15:07

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.140	04:10
Breath Volume		1.480	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:12
Subject Sample	# 2	0.148	04:13
Breath Volume		1.675	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 01/08/2026 10:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VONESCHE,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/31/2025 03:47:29

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:45	
Control Test		0.080	03:45	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:46	Subject Test Refused
Breath Volume		0.000	03:46	
Air Blank		0.000	03:46	
Control Test		0.080	03:46	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 01/08/2026 10:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: MARIANYI GALVEZ

DATE/TIME OF TEST: 12/31/2025 21:54:19

Diagnostics Check		OK	21:47
Air Blank		0.000	21:48
Control Test		0.080	21:48
Air Blank		0.000	21:49
Subject Sample	# 1	0.051	21:49
Breath Volume		1.738	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.048	21:52
Breath Volume		1.863	21:52
Air Blank		0.000	21:53
Control Test		0.080	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 01/08/2026 10:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: CUNARD

DATE/TIME OF TEST: 12/03/2025 19:29:41

Diagnostics Check		OK	19:23
Air Blank		0.000	19:23
Control Test		0.078	19:24
Air Blank		0.000	19:24
Subject Sample	# 1	0.166	19:25
Breath Volume		2.914	19:25
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.160	19:28
Breath Volume		2.800	19:28
Air Blank		0.000	19:28
Control Test		0.076	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 12/04/2025 00:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER,SEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: NUFFER,KURT
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/05/2025 00:42:30

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.077	00:39	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:41	Subject Test Refused
Breath Volume		0.000	00:41	
Air Blank		0.000	00:41	
Control Test		0.077	00:41	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEEHAN,MATTHEW,T
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEEHAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: PROCO

DATE/TIME OF TEST: 12/06/2025 20:32:59

Diagnostics Check		OK	20:26
Air Blank		0.000	20:26
Control Test		0.077	20:26
Air Blank		0.000	20:27
Subject Sample	# 1	0.168	20:28
Breath Volume		1.843	20:28
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.171	20:31
Breath Volume		1.863	20:31
Air Blank		0.000	20:32
Control Test		0.076	20:32
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: KEENAM

DATE/TIME OF TEST: 12/09/2025 03:07:30

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.000	03:02
Breath Volume		1.335	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.000	03:06
Breath Volume		1.546	03:06
Air Blank		0.000	03:06
Control Test		0.077	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEEHAN,MATTHEW,T
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEEHAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 12/10/2025 21:16:37

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.078	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.000	21:12
Breath Volume		4.011	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.000	21:15
Breath Volume		3.882	21:15
Air Blank		0.000	21:15
Control Test		0.078	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMABILE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: AMABILE,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 12/13/2025 03:18:24

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.212	03:13
Breath Volume		1.886	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.195	03:16
Breath Volume		1.375	03:16
Air Blank		0.000	03:17
Control Test		0.076	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: PHAM

DATE/TIME OF TEST: 12/13/2025 23:50:13

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.077	23:48	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:48	Subject Test Refused
Breath Volume		0.000	23:48	
Air Blank		0.000	23:49	
Control Test		0.076	23:49	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/15/2025 02:42:29

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:40	
Control Test		0.078	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.078	02:41	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CHAVEZ TELLEZ,ALBERTO
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: CSABON

DATE/TIME OF TEST: 12/19/2025 19:17:25

Diagnostics Check		OK	19:10
Air Blank		0.000	19:10
Control Test		0.078	19:10
Air Blank		0.000	19:11
Subject Sample	# 1	0.125	19:11
Breath Volume		1.503	19:11
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.130	19:15
Breath Volume		2.214	19:15
Air Blank		0.000	19:16
Control Test		0.078	19:16
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CHAVEZ TELLEZ,ALBERTO
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: MORJANI

DATE/TIME OF TEST: 12/20/2025 03:18:51

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.077	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.000	03:14
Breath Volume		1.621	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.000	03:17
Breath Volume		1.468	03:17
Air Blank		0.000	03:17
Control Test		0.078	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEEHAN,MATTHEW,T
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEEHAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 12/21/2025 04:07:05

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.078	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.049	04:02
Breath Volume		3.812	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.049	04:05
Breath Volume		3.652	04:05
Air Blank		0.000	04:06
Control Test		0.077	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:07

Data Download Date/Time: 02/01/2026 05:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: TEJEDA

DATE/TIME OF TEST: 12/22/2025 04:10:29

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.077	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.163	04:06
Breath Volume		1.578	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.163	04:09
Breath Volume		1.308	04:09
Air Blank		0.000	04:09
Control Test		0.077	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: VETH

DATE/TIME OF TEST: 12/23/2025 00:12:27

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.077	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.155	00:07
Breath Volume		2.046	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.142	00:11
Breath Volume		1.691	00:11
Air Blank		0.000	00:11
Control Test		0.077	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,GARY,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CALDWELL,GARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:56

SUBJECT'S LAST NAME: COREY

DATE/TIME OF TEST: 12/23/2025 13:24:28

Diagnostics Check		OK	13:17
Air Blank		0.000	13:18
Control Test		0.077	13:18
Air Blank		0.000	13:18
Subject Sample	# 1	0.000	13:19
Breath Volume		1.558	13:19
Air Blank		0.000	13:19
Air Blank		0.000	13:21
Subject Sample	# 2	0.000	13:23
Breath Volume		1.714	13:23
Air Blank		0.000	13:23
Control Test		0.077	13:23
Air Blank		0.000	13:24
Diagnostics Check		OK	13:24

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: HIERSCHE

DATE/TIME OF TEST: 12/24/2025 05:17:06

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.077	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.000	05:12
Breath Volume		1.718	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.000	05:15
Breath Volume		1.605	05:15
Air Blank		0.000	05:16
Control Test		0.077	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCORDINO,DOUGLAS,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CAMPADORE,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: BECRAFT

DATE/TIME OF TEST: 12/25/2025 15:18:09

Diagnostics Check		OK	15:15	
Air Blank		0.000	15:15	
Control Test		0.078	15:16	
Air Blank		0.000	15:16	
Subject Sample	# 1		15:16	Subject Test Refused
Breath Volume		0.000	15:16	
Air Blank		0.000	15:17	
Control Test		0.077	15:17	
Air Blank		0.000	15:18	
Diagnostics Check		OK	15:18	

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEEHAN,GEORGE,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEEHAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WEGELIN

DATE/TIME OF TEST: 12/25/2025 21:18:42

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:16	
Control Test		0.078	21:16	
Air Blank		0.000	21:17	
Subject Sample	# 1		21:17	Subject Test Refused
Breath Volume		0.000	21:17	
Air Blank		0.000	21:17	
Control Test		0.076	21:18	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CONTRERAS GARCIA

DATE/TIME OF TEST: 12/27/2025 02:01:52

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.078	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.088	01:57
Breath Volume		1.937	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.086	02:00
Breath Volume		2.472	02:00
Air Blank		0.000	02:00
Control Test		0.076	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CRUZ PEREZ

DATE/TIME OF TEST: 12/28/2025 02:04:47

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.077	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.091	02:00
Breath Volume		3.664	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:02
Subject Sample	# 2	0.086	02:03
Breath Volume		1.914	02:03
Air Blank		0.000	02:03
Control Test		0.077	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAYSON,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SMITH,JAYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: LIAN

DATE/TIME OF TEST: 12/28/2025 20:36:26

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.077	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.296	20:32
Breath Volume		1.777	20:32
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.284	20:34
Breath Volume		2.179	20:34
Air Blank		0.000	20:35
Control Test		0.077	20:35
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: BARTHOLOMEW

DATE/TIME OF TEST: 12/31/2025 21:28:39

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.079	21:23
Air Blank		0.000	21:23
Subject Sample	# 1	0.000	21:24
Breath Volume		1.746	21:24
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.000	21:27
Breath Volume		1.441	21:27
Air Blank		0.000	21:27
Control Test		0.079	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 02/01/2026 05:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKIANOV,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: RODRIGUEZ,ETHAN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: POOVEY

DATE/TIME OF TEST: 12/07/2025 03:39:00

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.064	03:34
Breath Volume		1.265	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.063	03:37
Breath Volume		1.250	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:39

Data Download Date/Time: 12/29/2025 08:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: HYATT

DATE/TIME OF TEST: 12/18/2025 01:46:37

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.162	01:40
Breath Volume		1.503	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.164	01:45
Breath Volume		1.601	01:45
Air Blank		0.000	01:45
Control Test		0.079	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 12/29/2025 08:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DONALD,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HEDGEMON,KAI
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: COLE-MILLER

DATE/TIME OF TEST: 12/20/2025 23:27:55

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.079	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.157	23:23
Breath Volume		1.457	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.149	23:26
Breath Volume		1.578	23:26
Air Blank		0.000	23:27
Control Test		0.078	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 12/29/2025 08:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKIANOV,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXANDER
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 12/25/2025 02:21:03

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.172	02:16
Breath Volume		1.925	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:18
Subject Sample	# 2	0.171	02:19
Breath Volume		2.808	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:21

Data Download Date/Time: 12/29/2025 08:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFEY,KYLE,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: DELACAMPA,BRYANT
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: DOSSANTOSQUARESMA

DATE/TIME OF TEST: 12/30/2025 23:55:27

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.079	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.000	23:49	Volume Not Met
Breath Volume		0.000	23:49	
Air Blank		0.000	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 2		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:54	
Control Test		0.080	23:54	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 01/27/2026 09:49

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 12/08/2025 02:03:24

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.079	01:58
Breath Volume		1.402	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.081	02:01
Breath Volume		1.570	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 12/30/2025 23:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ERNEST,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HALPIN,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 12/12/2025 16:58:25

Diagnostics Check		OK	16:54	
Air Blank		0.000	16:55	
Control Test		0.079	16:55	
Air Blank		0.000	16:55	
Subject Sample	# 1		16:57	Subject Test Refused
Breath Volume		0.000	16:57	
Air Blank		0.000	16:57	
Control Test		0.079	16:57	
Air Blank		0.000	16:58	
Diagnostics Check		OK	16:58	

Data Download Date/Time: 12/30/2025 23:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ERNEST,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CLYDE,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 12/17/2025 17:48:35

Diagnostics Check		OK	17:45	
Air Blank		0.000	17:46	
Control Test		0.080	17:46	
Air Blank		0.000	17:46	
Subject Sample	# 1		17:47	Subject Test Refused
Breath Volume		0.000	17:47	
Air Blank		0.000	17:47	
Control Test		0.079	17:48	
Air Blank		0.000	17:48	
Diagnostics Check		OK	17:48	

Data Download Date/Time: 12/30/2025 23:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BLAKE,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: BOTELLO

DATE/TIME OF TEST: 12/18/2025 19:23:08

Diagnostics Check		OK	19:16
Air Blank		0.000	19:16
Control Test		0.079	19:17
Air Blank		0.000	19:17
Subject Sample	# 1	0.352	19:18
Breath Volume		2.582	19:18
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.355	19:21
Breath Volume		2.917	19:21
Air Blank		0.000	19:22
Control Test		0.078	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 12/30/2025 23:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: SOUTHERLAND

DATE/TIME OF TEST: 12/19/2025 19:16:07

Diagnostics Check		OK	19:09
Air Blank		0.000	19:09
Control Test		0.079	19:10
Air Blank		0.000	19:10
Subject Sample	# 1	0.120	19:11
Breath Volume		1.726	19:11
Air Blank		0.000	19:12
Air Blank		0.000	19:13
Subject Sample	# 2	0.115	19:14
Breath Volume		1.414	19:14
Air Blank		0.000	19:15
Control Test		0.079	19:15
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 12/30/2025 23:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,K,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: GOULD

DATE/TIME OF TEST: 12/29/2025 14:51:26

Diagnostics Check		OK	14:40	
Air Blank		0.000	14:40	
Control Test		0.079	14:40	
Air Blank		0.000	14:41	
Subject Sample	# 1		14:44	No Sample Provided
Breath Volume		0.000	14:44	
Air Blank		0.000	14:44	
Air Blank		0.000	14:46	
Subject Sample	# 2		14:50	No Sample Provided
Breath Volume		0.000	14:50	
Air Blank		0.000	14:50	
Control Test		0.079	14:50	
Air Blank		0.000	14:51	
Diagnostics Check		OK	14:51	

Data Download Date/Time: 12/30/2025 23:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MEGOWAN,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/31/2025 02:21:42

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.079	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1	0.172	02:10	
Breath Volume		4.671	02:10	
Air Blank		0.000	02:11	
Air Blank		0.000	02:13	
Subject Sample	# 2		02:16	No Sample Provided
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Air Blank		0.000	02:18	
Subject Sample	# 3	0.167	02:20	
Breath Volume		1.820	02:20	
Air Blank		0.000	02:20	
Control Test		0.080	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 02/03/2026 18:17

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ROBERT,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 12/02/2025 02:17:24

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.260	02:12
Breath Volume		1.699	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.269	02:15
Breath Volume		2.183	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ADRIANNE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WATSON,S
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: PETTIT

DATE/TIME OF TEST: 12/03/2025 01:50:37

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.220	01:45
Breath Volume		3.527	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.205	01:49
Breath Volume		1.238	01:49
Air Blank		0.000	01:49
Control Test		0.078	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,GRANT,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY,GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: MIRANDA PEDRAZA

DATE/TIME OF TEST: 12/03/2025 18:21:17

Diagnostics Check		OK	18:14
Air Blank		0.000	18:15
Control Test		0.079	18:15
Air Blank		0.000	18:16
Subject Sample	# 1	0.270	18:16
Breath Volume		2.300	18:16
Air Blank		0.000	18:17
Air Blank		0.000	18:19
Subject Sample	# 2	0.269	18:19
Breath Volume		2.144	18:19
Air Blank		0.000	18:20
Control Test		0.078	18:20
Air Blank		0.000	18:21
Diagnostics Check		OK	18:21

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORREGO,MAXWELL,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: BEJANO

DATE/TIME OF TEST: 12/04/2025 20:23:40

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.078	20:18
Air Blank		0.000	20:18
Subject Sample	# 1	0.131	20:19
Breath Volume		2.164	20:19
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.125	20:22
Breath Volume		2.070	20:22
Air Blank		0.000	20:22
Control Test		0.079	20:23
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADOCCHIO,CASEY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CHURUKIAN,BRANDON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: HILLARD

DATE/TIME OF TEST: 12/11/2025 03:11:05

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.131	03:06
Breath Volume		4.691	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.134	03:09
Breath Volume		3.695	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:11

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,GRANT,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY,GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:27

SUBJECT'S LAST NAME: ALDERSON

DATE/TIME OF TEST: 12/14/2025 11:00:20

Diagnostics Check		OK	10:53
Air Blank		0.000	10:54
Control Test		0.079	10:54
Air Blank		0.000	10:55
Subject Sample	# 1	0.198	10:55
Breath Volume		1.765	10:55
Air Blank		0.000	10:56
Air Blank		0.000	10:58
Subject Sample	# 2	0.198	10:58
Breath Volume		2.003	10:58
Air Blank		0.000	10:59
Control Test		0.078	10:59
Air Blank		0.000	11:00
Diagnostics Check		OK	11:00

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERARDI,JOHN,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LAMANDIA,VINCENT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SABALEWSKI

DATE/TIME OF TEST: 12/24/2025 01:41:02

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.147	01:36
Breath Volume		1.683	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.148	01:39
Breath Volume		1.812	01:39
Air Blank		0.000	01:40
Control Test		0.077	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:41

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: CORNELIUS

DATE/TIME OF TEST: 12/27/2025 02:25:56

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.132	02:21
Breath Volume		1.734	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.133	02:24
Breath Volume		2.246	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTO,JOHN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FLIETSRA JR

DATE/TIME OF TEST: 12/29/2025 05:03:14

Diagnostics Check		OK	04:59	
Air Blank		0.000	05:00	
Control Test		0.079	05:00	
Air Blank		0.000	05:01	
Subject Sample	# 1	0.121	05:01	
Breath Volume		2.722	05:01	
Air Blank			05:01	Improper Sample
Air Blank		0.000	05:02	
Control Test		0.079	05:02	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTO,JOHN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FLIETSTRA JR

DATE/TIME OF TEST: 12/29/2025 05:11:42

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.137	05:07
Breath Volume		2.421	05:07
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.131	05:10
Breath Volume		1.656	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASPER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON,SAMUEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: CHEEVER

DATE/TIME OF TEST: 12/30/2025 21:17:55

Diagnostics Check		OK	21:11
Air Blank		0.000	21:12
Control Test		0.079	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.000	21:13
Breath Volume		3.250	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.000	21:16
Breath Volume		3.121	21:16
Air Blank		0.000	21:17
Control Test		0.080	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 12/31/2025 15:07

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LANIER,Z
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: FRITZ

DATE/TIME OF TEST: 12/01/2025 00:01:20

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.083	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.143	23:56
Breath Volume		1.355	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.136	23:59
Breath Volume		1.968	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 12/27/2025 12:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MAGALSKI,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 12/01/2025 20:20:19

Diagnostics Check		OK	20:15	
Air Blank		0.000	20:15	
Control Test		0.081	20:15	
Air Blank		0.000	20:16	
Subject Sample	# 1		20:18	Subject Test Refused
Breath Volume		0.000	20:18	
Air Blank		0.000	20:19	
Control Test		0.081	20:19	
Air Blank		0.000	20:20	
Diagnostics Check		OK	20:20	

Data Download Date/Time: 12/27/2025 12:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CLIFTON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: VENA

DATE/TIME OF TEST: 12/01/2025 23:55:28

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:43	
Control Test		0.082	23:43	
Air Blank		0.000	23:44	
Subject Sample	# 1	0.170	23:47	Volume Not Met
Breath Volume		0.117	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 2	0.169	23:50	
Breath Volume		1.593	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 3	0.173	23:53	
Breath Volume		1.187	23:53	
Air Blank		0.000	23:54	
Control Test		0.079	23:54	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 12/27/2025 12:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,A,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MAGALSKI,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: WILLAIMS

DATE/TIME OF TEST: 12/03/2025 03:36:29

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.082	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.134	03:31
Breath Volume		1.109	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.125	03:34
Breath Volume		1.332	03:34
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/27/2025 12:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: REDMON

DATE/TIME OF TEST: 12/04/2025 20:47:13

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.081	20:41
Air Blank		0.000	20:42
Subject Sample	# 1	0.117	20:42
Breath Volume		2.558	20:42
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.117	20:45
Breath Volume		1.980	20:45
Air Blank		0.000	20:46
Control Test		0.081	20:46
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 12/27/2025 12:38

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SAFAR,B
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: CADDELL

DATE/TIME OF TEST: 12/06/2025 03:13:22

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.082	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.192	03:08
Breath Volume		1.210	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.200	03:11
Breath Volume		1.398	03:11
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 12/27/2025 12:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GAINER

DATE/TIME OF TEST: 12/06/2025 03:49:36

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:40	
Control Test		0.082	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 1	0.265	03:42	
Breath Volume		1.632	03:42	
Air Blank		0.000	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 2	0.275	03:48	Volume Not Met
Breath Volume		0.710	03:48	
Air Blank			03:48	Purge Fail
Air Blank		0.000	03:49	

Data Download Date/Time: 12/27/2025 12:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GAINER

DATE/TIME OF TEST: 12/06/2025 04:06:13

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:57	
Control Test		0.079	03:58	
Air Blank		0.000	03:58	
Subject Sample	# 1	0.277	03:59	
Breath Volume		1.269	03:59	
Air Blank		0.000	03:59	
Air Blank		0.000	04:01	
Subject Sample	# 2		04:04	Subject Test Refused
Breath Volume		0.253	04:04	
Air Blank		0.000	04:05	
Control Test		0.078	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 12/27/2025 12:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SAFAR,B
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GERSHAW

DATE/TIME OF TEST: 12/07/2025 01:44:48

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:43	
Control Test		0.081	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 12/27/2025 12:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ELIZABETH,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HERRINGTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: ROZIER

DATE/TIME OF TEST: 12/08/2025 22:13:00

Diagnostics Check		OK	22:04	
Air Blank		0.000	22:05	
Control Test		0.079	22:05	
Air Blank		0.000	22:05	
Subject Sample	# 1	0.147	22:09	Volume Not Met
Breath Volume		0.410	22:09	
Air Blank		0.000	22:09	
Air Blank		0.000	22:11	
Subject Sample	# 2		22:11	Subject Test Refused
Breath Volume		0.000	22:11	
Air Blank		0.000	22:12	
Control Test		0.078	22:12	
Air Blank		0.000	22:12	
Diagnostics Check		OK	22:13	

Data Download Date/Time: 12/27/2025 12:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MILLAN,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 12/08/2025 22:21:24

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:19	
Control Test		0.079	22:19	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:20	Subject Test Refused
Breath Volume		0.000	22:20	
Air Blank		0.000	22:20	
Control Test		0.079	22:20	
Air Blank		0.000	22:21	
Diagnostics Check		OK	22:21	

Data Download Date/Time: 12/27/2025 12:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ELIZABETH,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SELMANOVIC,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 12/09/2025 22:29:33

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:27	
Control Test		0.081	22:27	
Air Blank		0.000	22:28	
Subject Sample	# 1		22:28	Subject Test Refused
Breath Volume		0.000	22:28	
Air Blank		0.000	22:28	
Control Test		0.080	22:28	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 12/27/2025 12:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 12/10/2025 00:10:31

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:07	
Control Test		0.081	00:08	
Air Blank		0.000	00:08	
Subject Sample	# 1		00:09	Subject Test Refused
Breath Volume		0.000	00:09	
Air Blank		0.000	00:09	
Control Test		0.081	00:09	
Air Blank		0.000	00:10	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 12/27/2025 12:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ELIZABETH,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 12/12/2025 23:10:18

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.081	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.000	23:05
Breath Volume		2.074	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.000	23:08
Breath Volume		2.085	23:08
Air Blank		0.000	23:09
Control Test		0.080	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 12/27/2025 12:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CLAY,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: DEGUIERA

DATE/TIME OF TEST: 12/13/2025 23:26:20

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.180	23:21
Breath Volume		2.035	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.186	23:24
Breath Volume		2.695	23:24
Air Blank		0.000	23:25
Control Test		0.079	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/14/2025 01:14:00

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.063	01:09
Breath Volume		1.941	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:11
Subject Sample	# 2	0.065	01:12
Breath Volume		1.707	01:12
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:14

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HERRINGTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/14/2025 02:07:43

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.191	02:02
Breath Volume		2.136	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.181	02:06
Breath Volume		1.378	02:06
Air Blank		0.000	02:06
Control Test		0.078	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDEN,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 12/15/2025 02:49:02

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:41	
Control Test		0.081	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1	0.085	02:44	
Breath Volume		1.996	02:44	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 2	0.078	02:48	
Breath Volume		1.621	02:48	
Air Blank			02:48	RFI Detect
Air Blank		0.000	02:49	

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: FONTANA,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: KORTEKAAS

DATE/TIME OF TEST: 12/17/2025 01:00:37

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.081	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.235	00:56
Breath Volume		2.152	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.225	00:58
Breath Volume		1.609	00:58
Air Blank		0.000	00:59
Control Test		0.078	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THORNBURG,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GUINN

DATE/TIME OF TEST: 12/17/2025 01:32:48

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.000	01:28
Breath Volume		2.425	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.000	01:31
Breath Volume		1.144	01:31
Air Blank		0.000	01:31
Control Test		0.081	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 12/27/2025 12:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ELIZABETH,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/18/2025 20:46:34

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.081	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.000	20:42
Breath Volume		4.675	20:42
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.000	20:45
Breath Volume		4.726	20:45
Air Blank		0.000	20:45
Control Test		0.080	20:45
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,A,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BLAKENSHIP

DATE/TIME OF TEST: 12/21/2025 02:29:29

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.096	02:24
Breath Volume		1.277	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.102	02:27
Breath Volume		1.683	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWMAN,TYLER,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PERALTA,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: NEESE

DATE/TIME OF TEST: 12/22/2025 16:14:32

Diagnostics Check		OK	16:08
Air Blank		0.000	16:08
Control Test		0.079	16:09
Air Blank		0.000	16:09
Subject Sample	# 1	0.075	16:10
Breath Volume		1.507	16:10
Air Blank		0.000	16:10
Air Blank		0.000	16:12
Subject Sample	# 2	0.075	16:12
Breath Volume		2.113	16:12
Air Blank		0.000	16:13
Control Test		0.079	16:13
Air Blank		0.000	16:14
Diagnostics Check		OK	16:14

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/23/2025 04:04:44

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.081	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.000	04:00
Breath Volume		3.746	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.000	04:03
Breath Volume		3.796	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ELIZABETH,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MIXON

DATE/TIME OF TEST: 12/23/2025 23:48:12

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.081	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.000	23:43
Breath Volume		1.644	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.000	23:46
Breath Volume		2.390	23:46
Air Blank		0.000	23:47
Control Test		0.081	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MAGALSKI,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: KESSEL

DATE/TIME OF TEST: 12/25/2025 02:21:37

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.081	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.098	02:17
Breath Volume		1.542	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.093	02:20
Breath Volume		1.429	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,T,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MAGALSKI,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ODONNELL

DATE/TIME OF TEST: 12/26/2025 02:08:52

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.081	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:07	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 12/26/2025 21:44:58

Diagnostics Check		OK	21:38
Air Blank		0.000	21:39
Control Test		0.081	21:39
Air Blank		0.000	21:40
Subject Sample	# 1	0.000	21:40
Breath Volume		1.609	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.000	21:43
Breath Volume		1.808	21:43
Air Blank		0.000	21:44
Control Test		0.081	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GRIFFTH

DATE/TIME OF TEST: 12/27/2025 02:38:18

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.081	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.148	02:33
Breath Volume		1.101	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.166	02:36
Breath Volume		1.121	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 12/27/2025 12:42

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SAFAR,B
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 12/28/2025 01:13:02

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.078	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:12	
Control Test		0.078	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 02/28/2026 21:46

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,ROBERT,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: RODEFFER

DATE/TIME OF TEST: 12/28/2025 03:23:27

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.081	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 02/28/2026 21:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: CLOWERS

DATE/TIME OF TEST: 12/28/2025 20:09:17

Diagnostics Check		OK	20:03
Air Blank		0.000	20:03
Control Test		0.081	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.000	20:05
Breath Volume		1.738	20:05
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.000	20:07
Breath Volume		1.265	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 02/28/2026 21:46

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 12/29/2025 04:56:50

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.081	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.000	04:52
Breath Volume		1.554	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.000	04:55
Breath Volume		1.609	04:55
Air Blank		0.000	04:55
Control Test		0.081	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 02/28/2026 21:46

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,B,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232062
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: CASTLE

DATE/TIME OF TEST: 12/31/2025 23:00:53

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.081	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.054	22:56
Breath Volume		1.289	22:56
Air Blank		0.000	22:57
Air Blank		0.000	22:58
Subject Sample	# 2	0.054	22:59
Breath Volume		1.593	22:59
Air Blank		0.000	22:59
Control Test		0.078	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 02/28/2026 21:46

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SANCHEZ,ALEJANDRO
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GOMEZ MENDOZA

DATE/TIME OF TEST: 12/01/2025 01:18:56

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.079	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1		01:11	No Sample Provided
Breath Volume		0.000	01:11	
Air Blank		0.000	01:12	
Air Blank		0.000	01:14	
Subject Sample	# 2		01:17	No Sample Provided
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.079	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 12/23/2025 07:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 12/01/2025 01:31:24

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.064	01:26
Breath Volume		4.230	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.064	01:29
Breath Volume		4.277	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 12/23/2025 07:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SALSMAN

DATE/TIME OF TEST: 12/03/2025 01:02:01

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.079	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:55	No Sample Provided
Breath Volume		0.000	00:55	
Air Blank		0.000	00:55	
Air Blank		0.000	00:57	
Subject Sample	# 2		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Control Test		0.079	01:01	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SALSMAN

DATE/TIME OF TEST: 12/03/2025 01:19:09

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.079	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1	0.205	01:06	
Breath Volume		1.203	01:06	
Air Blank		0.000	01:07	
Air Blank		0.000	01:08	
Subject Sample	# 2		01:12	No Sample Provided
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Air Blank		0.000	01:14	
Subject Sample	# 3		01:17	No Sample Provided
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.079	01:18	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 12/06/2025 04:11:35

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.144	04:06
Breath Volume		2.238	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.144	04:10
Breath Volume		1.902	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BALL,GARY
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FIGUEROA-NAZARIO

DATE/TIME OF TEST: 12/07/2025 00:23:59

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.066	00:19
Breath Volume		2.183	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.069	00:22
Breath Volume		1.812	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MARTIN,SAMUEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/07/2025 02:01:37

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.156	01:56
Breath Volume		2.554	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.176	02:00
Breath Volume		1.140	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:09

SUBJECT'S LAST NAME: NUTT

DATE/TIME OF TEST: 12/08/2025 10:36:47

Diagnostics Check		OK	10:29
Air Blank		0.000	10:30
Control Test		0.080	10:30
Air Blank		0.000	10:31
Subject Sample	# 1	0.302	10:32
Breath Volume		1.765	10:32
Air Blank		0.000	10:32
Air Blank		0.000	10:34
Subject Sample	# 2	0.320	10:35
Breath Volume		1.625	10:35
Air Blank		0.000	10:35
Control Test		0.078	10:36
Air Blank		0.000	10:36
Diagnostics Check		OK	10:36

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:18

SUBJECT'S LAST NAME: JUDAH

DATE/TIME OF TEST: 12/08/2025 10:47:32

Diagnostics Check		OK	10:43	
Air Blank		0.000	10:44	
Control Test		0.079	10:44	
Air Blank		0.000	10:45	
Subject Sample	# 1		10:46	Interferent Detect
Breath Volume		2.652	10:46	
Air Blank		0.000	10:46	
Control Test		0.079	10:46	
Air Blank		0.000	10:47	
Diagnostics Check		OK	10:47	

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:18

SUBJECT'S LAST NAME: JUDAH

DATE/TIME OF TEST: 12/08/2025 10:59:20

Diagnostics Check		OK	10:49	
Air Blank		0.000	10:50	
Control Test		0.079	10:50	
Air Blank		0.000	10:51	
Subject Sample	# 1		10:54	No Sample Provided
Breath Volume		0.000	10:54	
Air Blank		0.000	10:54	
Air Blank		0.000	10:56	
Subject Sample	# 2		10:57	Interferent Detect
Breath Volume		3.167	10:57	
Air Blank		0.000	10:58	
Control Test		0.079	10:58	
Air Blank		0.000	10:59	
Diagnostics Check		OK	10:59	

Data Download Date/Time: 12/23/2025 07:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:18

SUBJECT'S LAST NAME: JUDAH

DATE/TIME OF TEST: 12/08/2025 11:05:14

Diagnostics Check		OK	11:01	
Air Blank		0.000	11:02	
Control Test		0.079	11:02	
Air Blank		0.000	11:03	
Subject Sample	# 1		11:03	Interferent Detect
Breath Volume		1.324	11:03	
Air Blank		0.000	11:04	
Control Test		0.079	11:04	
Air Blank		0.000	11:05	
Diagnostics Check		OK	11:05	

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: JUDAH

DATE/TIME OF TEST: 12/08/2025 11:49:52

Diagnostics Check		OK	11:46	
Air Blank		0.000	11:46	
Control Test		0.080	11:47	
Air Blank		0.000	11:47	
Subject Sample	# 1		11:48	Interferent Detect
Breath Volume		4.468	11:48	
Air Blank		0.000	11:48	
Control Test		0.080	11:49	
Air Blank		0.000	11:49	
Diagnostics Check		OK	11:49	

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DAWSON,MICHAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: SUBJECT

DATE/TIME OF TEST: 12/08/2025 12:10:48

Diagnostics Check		OK	12:04
Air Blank		0.000	12:05
Control Test		0.079	12:05
Air Blank		0.000	12:05
Subject Sample	# 1	0.000	12:06
Breath Volume		2.644	12:06
Air Blank		0.000	12:06
Air Blank		0.000	12:08
Subject Sample	# 2	0.000	12:09
Breath Volume		4.019	12:09
Air Blank		0.000	12:09
Control Test		0.078	12:10
Air Blank		0.000	12:10
Diagnostics Check		OK	12:10

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: JUDAH

DATE/TIME OF TEST: 12/08/2025 13:41:11

Diagnostics Check		OK	13:37	
Air Blank		0.000	13:38	
Control Test		0.079	13:38	
Air Blank		0.000	13:38	
Subject Sample	# 1		13:39	Interferent Detect
Breath Volume		3.402	13:39	
Air Blank		0.000	13:40	
Control Test		0.080	13:40	
Air Blank		0.000	13:41	
Diagnostics Check		OK	13:41	

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,AARON,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 12/09/2025 14:02:51

Diagnostics Check		OK	13:56
Air Blank		0.000	13:57
Control Test		0.080	13:57
Air Blank		0.000	13:58
Subject Sample	# 1	0.000	13:58
Breath Volume		1.500	13:58
Air Blank		0.000	13:59
Air Blank		0.000	14:01
Subject Sample	# 2	0.000	14:01
Breath Volume		1.855	14:01
Air Blank		0.000	14:01
Control Test		0.079	14:02
Air Blank		0.000	14:02
Diagnostics Check		OK	14:02

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 12/10/2025 23:14:47

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.080	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.272	23:10
Breath Volume		2.109	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:12
Subject Sample	# 2	0.276	23:13
Breath Volume		2.062	23:13
Air Blank		0.000	23:13
Control Test		0.079	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: GREGORIO FLORES

DATE/TIME OF TEST: 12/11/2025 02:33:49

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.176	02:28
Breath Volume		1.527	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:30
Subject Sample	# 2	0.167	02:32
Breath Volume		1.414	02:32
Air Blank		0.000	02:32
Control Test		0.077	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 12/23/2025 07:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: STIDHAM

DATE/TIME OF TEST: 12/11/2025 03:18:16

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.000	03:12
Breath Volume		1.968	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.000	03:16
Breath Volume		2.605	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,PHILIP,P
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: KING,PHILIP
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: KULIKOWSKI

DATE/TIME OF TEST: 12/11/2025 03:42:01

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.079	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:35	No Sample Provided
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2		03:40	No Sample Provided
Breath Volume		0.000	03:40	
Air Blank		0.000	03:41	
Control Test		0.079	03:41	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULES,EMMANUELLO,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: JULES,EMMANUELLO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: ARROYO GARRIDO

DATE/TIME OF TEST: 12/12/2025 19:38:32

Diagnostics Check		OK	19:32
Air Blank		0.000	19:33
Control Test		0.079	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.190	19:34
Breath Volume		4.113	19:34
Air Blank		0.000	19:34
Air Blank		0.000	19:36
Subject Sample	# 2	0.180	19:36
Breath Volume		1.804	19:36
Air Blank		0.000	19:37
Control Test		0.078	19:37
Air Blank		0.000	19:38
Diagnostics Check		OK	19:38

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ,JACOB,M
BREATH TEST OPERATOR AGENCY: Davenport PD
ARRESTING OFFICER: JIMENEZ,JACOB
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: MALDONADO MERCADO

DATE/TIME OF TEST: 12/13/2025 08:22:44

Diagnostics Check		OK	08:16
Air Blank		0.000	08:16
Control Test		0.079	08:17
Air Blank		0.000	08:17
Subject Sample	# 1	0.065	08:18
Breath Volume		1.492	08:18
Air Blank		0.000	08:18
Air Blank		0.000	08:20
Subject Sample	# 2	0.060	08:21
Breath Volume		2.605	08:21
Air Blank		0.000	08:21
Control Test		0.079	08:22
Air Blank		0.000	08:22
Diagnostics Check		OK	08:22

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,AARON,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: ARRIAGA-GARCIA

DATE/TIME OF TEST: 12/13/2025 18:00:49

Diagnostics Check		OK	17:49	
Air Blank		0.000	17:49	
Control Test		0.080	17:50	
Air Blank		0.000	17:50	
Subject Sample	# 1		17:53	No Sample Provided
Breath Volume		0.000	17:53	
Air Blank		0.000	17:54	
Air Blank		0.000	17:56	
Subject Sample	# 2		17:59	No Sample Provided
Breath Volume		0.000	17:59	
Air Blank		0.000	17:59	
Control Test		0.080	18:00	
Air Blank		0.000	18:00	
Diagnostics Check		OK	18:00	

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,AARON,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: ARRIAGA-GARCIA

DATE/TIME OF TEST: 12/13/2025 18:14:15

Diagnostics Check		OK	18:07
Air Blank		0.000	18:07
Control Test		0.079	18:07
Air Blank		0.000	18:08
Subject Sample	# 1	0.129	18:09
Breath Volume		1.957	18:09
Air Blank		0.000	18:10
Air Blank		0.000	18:12
Subject Sample	# 2	0.119	18:12
Breath Volume		1.355	18:12
Air Blank		0.000	18:13
Control Test		0.079	18:13
Air Blank		0.000	18:14
Diagnostics Check		OK	18:14

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GARCIA,LUIS
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/14/2025 05:10:26

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.080	05:02
Air Blank		0.000	05:02
Subject Sample	# 1	0.000	05:05
Breath Volume		2.816	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.000	05:09
Breath Volume		1.976	05:09
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE,BRANDON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/14/2025 06:16:23

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.080	06:10
Air Blank		0.000	06:11
Subject Sample	# 1	0.135	06:11
Breath Volume		2.257	06:11
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.134	06:14
Breath Volume		1.761	06:14
Air Blank		0.000	06:15
Control Test		0.080	06:15
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: LARA REYES

DATE/TIME OF TEST: 12/14/2025 09:19:32

Diagnostics Check		OK	09:07	
Air Blank		0.000	09:08	
Control Test		0.079	09:08	
Air Blank		0.000	09:09	
Subject Sample	# 1	0.124	09:12	Volume Not Met
Breath Volume		0.082	09:12	
Air Blank		0.000	09:13	
Air Blank		0.000	09:14	
Subject Sample	# 2	0.100	09:18	Volume Not Met
Breath Volume		0.105	09:18	
Air Blank		0.000	09:18	
Control Test		0.078	09:18	
Air Blank		0.000	09:19	
Diagnostics Check		OK	09:19	

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: LARA REYES

DATE/TIME OF TEST: 12/14/2025 09:46:11

Diagnostics Check		OK	09:34	
Air Blank		0.000	09:35	
Control Test		0.078	09:35	
Air Blank		0.000	09:35	
Subject Sample	# 1	0.122	09:39	Volume Not Met
Breath Volume		0.082	09:39	
Air Blank		0.000	09:39	
Air Blank		0.000	09:41	
Subject Sample	# 2	0.114	09:44	Volume Not Met
Breath Volume		0.066	09:44	
Air Blank		0.000	09:45	
Control Test		0.078	09:45	
Air Blank		0.000	09:46	
Diagnostics Check		OK	09:46	

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ZAMOT

DATE/TIME OF TEST: 12/15/2025 03:15:44

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.080	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:08	No Sample Provided
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2		03:14	No Sample Provided
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.081	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 12/16/2025 23:38:02

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:27	
Control Test		0.079	23:27	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:31	No Sample Provided
Breath Volume		0.000	23:31	
Air Blank		0.000	23:31	
Air Blank		0.000	23:33	
Subject Sample	# 2		23:36	No Sample Provided
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Control Test		0.080	23:37	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 12/23/2025 07:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 12/17/2025 02:52:16

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.198	02:46
Breath Volume		2.324	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.200	02:50
Breath Volume		1.921	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SUSANA JR

DATE/TIME OF TEST: 12/17/2025 06:14:49

Diagnostics Check		OK	06:01	
Air Blank		0.000	06:01	
Control Test		0.080	06:02	
Air Blank		0.000	06:02	
Subject Sample	# 1		06:05	No Sample Provided
Breath Volume		0.000	06:05	
Air Blank		0.000	06:06	
Air Blank		0.000	06:08	
Subject Sample	# 2	0.166	06:10	
Breath Volume		2.394	06:10	
Air Blank		0.000	06:11	
Air Blank		0.000	06:12	
Subject Sample	# 3	0.170	06:13	
Breath Volume		2.144	06:13	
Air Blank		0.000	06:13	
Control Test		0.079	06:14	
Air Blank		0.000	06:14	
Diagnostics Check		OK	06:14	

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 12/19/2025 00:30:36

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.080	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.000	00:26
Breath Volume		1.285	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.000	00:29
Breath Volume		1.132	00:29
Air Blank		0.000	00:29
Control Test		0.078	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,PHILIP,P
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: KING,PHILIP
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: TOMAS

DATE/TIME OF TEST: 12/20/2025 02:56:15

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.044	02:51
Breath Volume		1.890	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.044	02:54
Breath Volume		2.640	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HARDIN,JEREMIAH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: AUGUSTINOWICZ

DATE/TIME OF TEST: 12/21/2025 01:23:58

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.206	01:19
Breath Volume		1.312	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.198	01:22
Breath Volume		1.238	01:22
Air Blank		0.000	01:23
Control Test		0.078	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BALL,GARY
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/21/2025 02:06:24

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.277	02:01
Breath Volume		2.156	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.277	02:04
Breath Volume		1.949	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON JR,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PURSE

DATE/TIME OF TEST: 12/21/2025 23:09:22

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.196	23:04
Breath Volume		2.625	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:06
Subject Sample	# 2	0.205	23:07
Breath Volume		1.562	23:07
Air Blank		0.000	23:08
Control Test		0.077	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: LIVELY

DATE/TIME OF TEST: 12/22/2025 01:03:24

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.190	00:58
Breath Volume		1.949	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.172	01:01
Breath Volume		1.187	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/22/2025 07:39:39

Diagnostics Check		OK	07:32
Air Blank		0.000	07:33
Control Test		0.080	07:33
Air Blank		0.000	07:33
Subject Sample	# 1	0.172	07:34
Breath Volume		1.500	07:34
Air Blank		0.000	07:35
Air Blank		0.000	07:37
Subject Sample	# 2	0.169	07:38
Breath Volume		1.871	07:38
Air Blank		0.000	07:38
Control Test		0.078	07:39
Air Blank		0.000	07:39
Diagnostics Check		OK	07:39

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBER,RYAN,C
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ELBER,RYAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HIGDON

DATE/TIME OF TEST: 12/22/2025 21:06:52

Diagnostics Check		OK	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:01
Subject Sample	# 1	0.000	21:02
Breath Volume		2.406	21:02
Air Blank		0.000	21:02
Air Blank		0.000	21:05
Subject Sample	# 2	0.000	21:05
Breath Volume		2.984	21:05
Air Blank		0.000	21:05
Control Test		0.080	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 12/23/2025 01:25:21

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.080	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:18	No Sample Provided
Breath Volume		0.000	01:18	
Air Blank		0.000	01:18	
Air Blank		0.000	01:20	
Subject Sample	# 2		01:23	No Sample Provided
Breath Volume		0.000	01:23	
Air Blank		0.000	01:24	
Control Test		0.080	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 12/23/2025 07:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALLEYNE,ALIYAH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GUERTIN

DATE/TIME OF TEST: 12/23/2025 03:22:56

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:12	
Control Test		0.078	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:15	No Sample Provided
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Air Blank		0.000	03:18	
Subject Sample	# 2		03:21	No Sample Provided
Breath Volume		0.000	03:21	
Air Blank		0.000	03:22	
Control Test		0.079	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 12/23/2025 07:25

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FIREBAUGH,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MENDZ RAMOS

DATE/TIME OF TEST: 12/23/2025 17:50:30

Diagnostics Check		OK	17:42
Air Blank		0.000	17:43
Control Test		0.080	17:43
Air Blank		0.000	17:44
Subject Sample	# 1	0.165	17:44
Breath Volume		1.937	17:44
Air Blank		0.000	17:45
Air Blank		0.000	17:47
Subject Sample	# 2	0.168	17:48
Breath Volume		1.277	17:48
Air Blank		0.000	17:49
Control Test		0.079	17:49
Air Blank		0.000	17:50
Diagnostics Check		OK	17:50

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: VERDUGO RODRIGUEZ

DATE/TIME OF TEST: 12/24/2025 22:29:29

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.109	22:24
Breath Volume		1.250	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.105	22:28
Breath Volume		2.027	22:28
Air Blank		0.000	22:28
Control Test		0.079	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MENDOZA PEREZ

DATE/TIME OF TEST: 12/24/2025 23:15:55

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.079	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.158	23:11
Breath Volume		1.703	23:11
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.153	23:14
Breath Volume		1.273	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OUTLAW,ADAM
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BROXEY-THOMAS

DATE/TIME OF TEST: 12/25/2025 01:03:25

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.011	00:59
Breath Volume		2.242	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.010	01:02
Breath Volume		1.535	01:02
Air Blank		0.000	01:02
Control Test		0.079	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: HASLEM

DATE/TIME OF TEST: 12/25/2025 01:27:01

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.080	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1	0.208	01:19	Volume Not Met
Breath Volume		0.601	01:19	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2	0.169	01:25	Volume Not Met
Breath Volume		0.316	01:25	
Air Blank		0.000	01:26	
Control Test		0.075	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITCHIE,SHELBY,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RITCHIE,SHELBY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/25/2025 04:12:14

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.129	04:07
Breath Volume		2.199	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.127	04:10
Breath Volume		1.726	04:10
Air Blank		0.000	04:11
Control Test		0.077	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/25/2025 13:34:22

Diagnostics Check		OK	13:28
Air Blank		0.000	13:28
Control Test		0.080	13:29
Air Blank		0.000	13:29
Subject Sample	# 1	0.150	13:29
Breath Volume		1.132	13:29
Air Blank		0.000	13:30
Air Blank		0.000	13:32
Subject Sample	# 2	0.151	13:32
Breath Volume		2.507	13:32
Air Blank		0.000	13:33
Control Test		0.078	13:33
Air Blank		0.000	13:34
Diagnostics Check		OK	13:34

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: QUIROZ

DATE/TIME OF TEST: 12/26/2025 22:06:55

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:55	
Control Test		0.080	21:55	
Air Blank		0.000	21:55	
Subject Sample	# 1	0.087	21:59	Volume Not Met
Breath Volume		0.449	21:59	
Air Blank		0.000	21:59	
Air Blank		0.000	22:01	
Subject Sample	# 2	0.115	22:02	
Breath Volume		2.972	22:02	
Air Blank		0.000	22:02	
Air Blank		0.000	22:04	
Subject Sample	# 3	0.116	22:05	
Breath Volume		2.503	22:05	
Air Blank		0.000	22:06	
Control Test		0.080	22:06	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 12/27/2025 00:18:03

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.109	00:12
Breath Volume		3.484	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.104	00:16
Breath Volume		2.910	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:18

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BRAHAM

DATE/TIME OF TEST: 12/27/2025 03:45:06

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.080	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:38	No Sample Provided
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2		03:43	No Sample Provided
Breath Volume		0.000	03:43	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 12/28/2025 00:37:41

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.184	00:32
Breath Volume		1.917	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.188	00:36
Breath Volume		2.031	00:36
Air Blank		0.000	00:36
Control Test		0.079	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: URIBE

DATE/TIME OF TEST: 12/28/2025 00:58:14

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.211	00:53
Breath Volume		1.566	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.214	00:56
Breath Volume		1.121	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 01/20/2026 06:35

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GARCIA,LUIS
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 12/28/2025 04:45:34

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:32	
Control Test		0.080	04:33	
Air Blank		0.000	04:33	
Subject Sample	# 1	0.156	04:36	Volume Not Met
Breath Volume		0.890	04:36	
Air Blank		0.000	04:37	
Air Blank		0.000	04:39	
Subject Sample	# 2	0.138	04:40	
Breath Volume		1.648	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 3	0.150	04:43	
Breath Volume		1.433	04:43	
Air Blank		0.000	04:44	
Control Test		0.079	04:44	
Air Blank			04:45	RFI Detect
Air Blank		0.000	04:45	

Data Download Date/Time: 01/20/2026 06:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GARCIA,LUIS
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 12/28/2025 05:24:31

Diagnostics Check		OK	05:13	
Air Blank		0.000	05:13	
Control Test		0.079	05:13	
Air Blank		0.000	05:14	
Subject Sample	# 1		05:17	No Sample Provided
Breath Volume		0.000	05:17	
Air Blank		0.000	05:18	
Air Blank		0.000	05:20	
Subject Sample	# 2		05:23	No Sample Provided
Breath Volume		0.000	05:23	
Air Blank		0.000	05:23	
Control Test		0.079	05:23	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 01/20/2026 06:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CORDERO-TRESTON,KASSANDRA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: STEELMAN

DATE/TIME OF TEST: 12/28/2025 09:23:23

Diagnostics Check		OK	09:16
Air Blank		0.000	09:17
Control Test		0.079	09:17
Air Blank		0.000	09:18
Subject Sample	# 1	0.113	09:18
Breath Volume		2.289	09:18
Air Blank		0.000	09:19
Air Blank		0.000	09:21
Subject Sample	# 2	0.115	09:21
Breath Volume		2.800	09:21
Air Blank		0.000	09:22
Control Test		0.079	09:22
Air Blank		0.000	09:23
Diagnostics Check		OK	09:23

Data Download Date/Time: 01/20/2026 06:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: SUMNER

DATE/TIME OF TEST: 12/28/2025 19:20:52

Diagnostics Check		OK	19:09	
Air Blank		0.000	19:09	
Control Test		0.080	19:10	
Air Blank		0.000	19:10	
Subject Sample	# 1		19:13	No Sample Provided
Breath Volume		0.000	19:13	
Air Blank		0.000	19:14	
Air Blank		0.000	19:16	
Subject Sample	# 2		19:19	No Sample Provided
Breath Volume		0.000	19:19	
Air Blank		0.000	19:19	
Control Test		0.080	19:20	
Air Blank		0.000	19:20	
Diagnostics Check		OK	19:20	

Data Download Date/Time: 01/20/2026 06:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULES,EMMANUELLO,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: JULES,EMMANUELLO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 12/28/2025 21:52:40

Diagnostics Check		OK	21:46
Air Blank		0.000	21:47
Control Test		0.080	21:47
Air Blank		0.000	21:47
Subject Sample	# 1	0.116	21:48
Breath Volume		1.820	21:48
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.120	21:51
Breath Volume		3.097	21:51
Air Blank		0.000	21:51
Control Test		0.079	21:52
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 01/20/2026 06:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MAISONET,JIMMY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: DRIGGERS

DATE/TIME OF TEST: 12/29/2025 00:41:55

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:33	
Control Test		0.080	00:34	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:37	No Sample Provided
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Air Blank		0.000	00:40	
Subject Sample	# 2		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:41	
Control Test		0.080	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 01/20/2026 06:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: CAPE

DATE/TIME OF TEST: 12/30/2025 22:47:51

Diagnostics Check		OK	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.175	22:43
Breath Volume		1.296	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.172	22:46
Breath Volume		1.171	22:46
Air Blank		0.000	22:46
Control Test		0.079	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 01/20/2026 06:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGULO,DALLAS,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: DELTUFO

DATE/TIME OF TEST: 12/31/2025 18:22:04

Diagnostics Check		OK	18:11
Air Blank		0.000	18:11
Control Test		0.079	18:11
Air Blank		0.000	18:12
Subject Sample	# 1	0.189	18:13
Breath Volume		3.195	18:13
Air Blank		0.000	18:14
Air Blank		0.000	18:16
Subject Sample	# 2	0.159	18:17
Breath Volume		1.281	18:17
Air Blank		0.000	18:18
Air Blank		0.000	18:19
Subject Sample	# 3	0.183	18:20
Breath Volume		3.425	18:20
Air Blank		0.000	18:21
Control Test		0.079	18:21
Air Blank		0.000	18:21
Diagnostics Check		OK	18:22

Data Download Date/Time: 01/20/2026 06:36

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PARIS

DATE/TIME OF TEST: 12/23/2025 04:48:30

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.078	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.069	04:44
Breath Volume		2.113	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.068	04:47
Breath Volume		2.312	04:47
Air Blank		0.000	04:47
Control Test		0.079	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 01/28/2026 13:27

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: CALLARI,MICHAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 12/17/2025 23:56:29

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:45	
Control Test		0.080	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.116	23:46	
Breath Volume		1.757	23:46	
Air Blank		0.000	23:47	
Air Blank		0.000	23:49	
Subject Sample	# 2		23:52	No Sample Provided
Breath Volume		0.000	23:52	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 3		23:55	Subject Test Refused
Breath Volume		0.000	23:55	
Air Blank		0.000	23:55	
Control Test		0.080	23:55	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 01/14/2026 11:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORENSTEIN,RANDY,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WISLON,ALLEN
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 11/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: LOPEZ MORENO

DATE/TIME OF TEST: 12/07/2025 22:49:52

Diagnostics Check		OK	22:43
Air Blank		0.000	22:44
Control Test		0.080	22:44
Air Blank		0.000	22:45
Subject Sample	# 1	0.196	22:45
Breath Volume		1.769	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:47
Subject Sample	# 2	0.187	22:48
Breath Volume		1.984	22:48
Air Blank		0.000	22:48
Control Test		0.079	22:49
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 12/26/2025 08:15

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 11/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CARMACK

DATE/TIME OF TEST: 12/20/2025 03:18:22

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.116	03:14
Breath Volume		1.539	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.111	03:16
Breath Volume		2.015	03:16
Air Blank		0.000	03:17
Control Test		0.081	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 12/26/2025 08:16

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPANA,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CAMPANA,NICHOLAS
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 11/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/23/2025 14:42:23

Diagnostics Check		OK	14:36
Air Blank		0.000	14:36
Control Test		0.081	14:36
Air Blank		0.000	14:37
Subject Sample	# 1	0.000	14:37
Breath Volume		2.093	14:37
Air Blank		0.000	14:38
Air Blank		0.000	14:40
Subject Sample	# 2	0.000	14:41
Breath Volume		1.679	14:41
Air Blank		0.000	14:41
Control Test		0.080	14:41
Air Blank		0.000	14:42
Diagnostics Check		OK	14:42

Data Download Date/Time: 12/26/2025 08:16

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,K,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CURBELO,A
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 12/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 12/31/2025 22:53:05

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.082	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.098	22:48
Breath Volume		2.000	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.097	22:51
Breath Volume		1.417	22:51
Air Blank		0.000	22:52
Control Test		0.080	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:53

Data Download Date/Time: 01/28/2026 10:38

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ELIZABETH,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BROWN,ELIZABETH
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 11/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: OTHER
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 12/08/2025 19:01:44

Diagnostics Check		OK	18:55
Air Blank		0.000	18:56
Control Test		0.078	18:56
Air Blank		0.000	18:57
Subject Sample	# 1	0.000	18:57
Breath Volume		2.937	18:57
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.000	19:00
Breath Volume		3.480	19:00
Air Blank		0.000	19:00
Control Test		0.078	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:01

Data Download Date/Time: 12/26/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: OTHER
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 12/08/2025 19:10:23

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.078	19:05
Air Blank		0.000	19:05
Subject Sample	# 1	0.000	19:06
Breath Volume		1.824	19:06
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.000	19:09
Breath Volume		1.820	19:09
Air Blank		0.000	19:09
Control Test		0.077	19:09
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 12/26/2025 11:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-KEMP,DANIELLE,T
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: SPEAKMAN,DUSTIN
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 12/26/2025 15:52:06

Diagnostics Check		OK	15:45
Air Blank		0.000	15:46
Control Test		0.077	15:46
Air Blank		0.000	15:47
Subject Sample	# 1	0.118	15:47
Breath Volume		2.296	15:47
Air Blank		0.000	15:48
Air Blank		0.000	15:50
Subject Sample	# 2	0.111	15:50
Breath Volume		1.710	15:50
Air Blank		0.000	15:51
Control Test		0.077	15:51
Air Blank		0.000	15:52
Diagnostics Check		OK	15:52

Data Download Date/Time: 01/23/2026 14:41

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 07:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 12/06/2025 08:19:43

Diagnostics Check		OK	08:16	
Air Blank		0.000	08:16	
Control Test		0.080	08:17	
Air Blank		0.000	08:17	
Subject Sample	# 1		08:17	Subject Test Refused
Breath Volume		0.000	08:17	
Air Blank		0.000	08:18	
Control Test		0.080	08:18	
Air Blank		0.000	08:19	
Diagnostics Check		Fail	08:19	Voltage/Current Test
Air Blank		0.000	08:19	

Data Download Date/Time: 12/17/2025 10:16

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 07:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 12/06/2025 08:26:03

Diagnostics Check		OK	08:23	
Air Blank		0.000	08:23	
Control Test		0.080	08:24	
Air Blank		0.000	08:24	
Subject Sample	# 1		08:24	Subject Test Refused
Breath Volume		0.000	08:24	
Air Blank		0.000	08:25	
Control Test		0.080	08:25	
Air Blank		0.000	08:25	
Diagnostics Check		OK	08:26	

Data Download Date/Time: 12/17/2025 10:16

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 10:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:39

SUBJECT'S LAST NAME: LOUTTIT

DATE/TIME OF TEST: 12/06/2025 11:06:10

Diagnostics Check		OK	10:59
Air Blank		0.000	11:00
Control Test		0.080	11:00
Air Blank		0.000	11:01
Subject Sample	# 1	0.000	11:01
Breath Volume		1.425	11:01
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.000	11:04
Breath Volume		2.273	11:04
Air Blank		0.000	11:05
Control Test		0.079	11:05
Air Blank		0.000	11:06
Diagnostics Check		OK	11:06

Data Download Date/Time: 12/17/2025 10:16

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,HOWARD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 12/07/2025 13:59:01

Diagnostics Check		OK	13:52
Air Blank		0.000	13:53
Control Test		0.080	13:53
Air Blank		0.000	13:54
Subject Sample	# 1	0.068	13:54
Breath Volume		2.105	13:54
Air Blank		0.000	13:55
Air Blank		0.000	13:57
Subject Sample	# 2	0.068	13:57
Breath Volume		1.472	13:57
Air Blank		0.000	13:58
Control Test		0.081	13:58
Air Blank		0.000	13:58
Diagnostics Check		OK	13:59

Data Download Date/Time: 12/17/2025 10:16

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: RAMROOP

DATE/TIME OF TEST: 12/21/2025 01:45:46

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.163	01:40
Breath Volume		1.285	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.165	01:44
Breath Volume		1.257	01:44
Air Blank		0.000	01:44
Control Test		0.081	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 01/07/2026 14:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RHASIATRY

DATE/TIME OF TEST: 12/21/2025 02:40:03

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.154	02:35
Breath Volume		1.777	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.165	02:38
Breath Volume		1.636	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:40

Data Download Date/Time: 01/07/2026 14:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DERISCA

DATE/TIME OF TEST: 12/21/2025 03:00:44

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.143	02:55
Breath Volume		1.574	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.138	02:59
Breath Volume		1.675	02:59
Air Blank		0.000	02:59
Control Test		0.079	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 01/07/2026 14:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NURY,KELSIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ALHASANI

DATE/TIME OF TEST: 12/21/2025 03:34:44

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.081	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.000	03:30
Breath Volume		1.175	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.000	03:33
Breath Volume		1.339	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 01/07/2026 14:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUQUE,S,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: GARCIA,E
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/1018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 12/24/2025 02:39:41

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:23	
Control Test		0.080	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1	0.147	02:27	
Breath Volume		1.363	02:27	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2	0.136	02:32	Volume Not Met
Breath Volume		0.117	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 3		02:38	No Sample Provided
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Control Test		0.080	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 01/14/2026 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUQUE,S,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: WILLIAMS,A
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/1018
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: GRIFFITHS

DATE/TIME OF TEST: 12/25/2025 07:35:59

Diagnostics Check		OK	07:29
Air Blank		0.000	07:30
Control Test		0.080	07:30
Air Blank		0.000	07:30
Subject Sample	# 1	0.104	07:31
Breath Volume		1.476	07:31
Air Blank		0.000	07:32
Air Blank		0.000	07:34
Subject Sample	# 2	0.106	07:34
Breath Volume		1.910	07:34
Air Blank		0.000	07:35
Control Test		0.080	07:35
Air Blank		0.000	07:35
Diagnostics Check		OK	07:35

Data Download Date/Time: 01/14/2026 11:22

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:57

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 12/05/2025 12:23:36

Diagnostics Check		OK	12:17
Air Blank		0.000	12:18
Control Test		0.081	12:18
Air Blank		0.000	12:19
Subject Sample	# 1	0.000	12:19
Breath Volume		4.433	12:19
Air Blank		0.000	12:19
Air Blank		0.000	12:21
Subject Sample	# 2	0.000	12:22
Breath Volume		3.679	12:22
Air Blank		0.000	12:22
Control Test		0.080	12:23
Air Blank		0.000	12:23
Diagnostics Check		OK	12:23

Data Download Date/Time: 12/17/2025 14:49

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 12/05/2025 17:10:35

Diagnostics Check		OK	16:59	
Air Blank		0.000	16:59	
Control Test		0.081	16:59	
Air Blank		0.000	17:00	
Subject Sample	# 1		17:03	No Sample Provided
Breath Volume		0.000	17:03	
Air Blank		0.000	17:04	
Air Blank		0.000	17:06	
Subject Sample	# 2		17:09	No Sample Provided
Breath Volume		0.000	17:09	
Air Blank		0.000	17:09	
Control Test		0.081	17:09	
Air Blank		0.000	17:10	
Diagnostics Check		OK	17:10	

Data Download Date/Time: 12/17/2025 14:49

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RAY,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 12/05/2025 19:22:01

Diagnostics Check		OK	19:15
Air Blank		0.000	19:16
Control Test		0.081	19:16
Air Blank		0.000	19:16
Subject Sample	# 1	0.063	19:17
Breath Volume		1.722	19:17
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.058	19:20
Breath Volume		1.484	19:20
Air Blank		0.000	19:21
Control Test		0.078	19:21
Air Blank		0.000	19:21
Diagnostics Check		OK	19:22

Data Download Date/Time: 12/17/2025 14:49

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GARNICA

DATE/TIME OF TEST: 12/07/2025 00:03:08

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.081	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.163	23:58
Breath Volume		3.359	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.165	00:01
Breath Volume		3.843	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 12/17/2025 14:49

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: BOYER ARROYO

DATE/TIME OF TEST: 12/07/2025 00:45:40

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.081	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.081	00:41
Breath Volume		2.945	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.080	00:44
Breath Volume		2.839	00:44
Air Blank		0.000	00:44
Control Test		0.081	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 12/17/2025 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANDLER,RYAN,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CHANDLER,RYAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/08/2025 21:55:56

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.081	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.000	21:51
Breath Volume		1.937	21:51
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.000	21:54
Breath Volume		1.566	21:54
Air Blank		0.000	21:55
Control Test		0.081	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 12/17/2025 14:49

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BLUNT

DATE/TIME OF TEST: 12/09/2025 01:51:19

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.081	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.143	01:46
Breath Volume		1.628	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.139	01:49
Breath Volume		1.988	01:49
Air Blank		0.000	01:50
Control Test		0.081	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 12/17/2025 14:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: PESTKE

DATE/TIME OF TEST: 12/09/2025 20:29:46

Diagnostics Check		OK	20:18	
Air Blank		0.000	20:18	
Control Test		0.082	20:19	
Air Blank		0.000	20:19	
Subject Sample	# 1		20:22	No Sample Provided
Breath Volume		0.000	20:22	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 2		20:28	No Sample Provided
Breath Volume		0.000	20:28	
Air Blank		0.000	20:28	
Control Test		0.081	20:29	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 12/17/2025 14:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 12/10/2025 00:38:23

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.081	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.264	00:30
Breath Volume		1.355	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.242	00:33
Breath Volume		1.281	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 3	0.252	00:36
Breath Volume		2.292	00:36
Air Blank		0.000	00:37
Control Test		0.081	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 12/17/2025 14:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DEUSEN,COLE,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VAN DEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: VANLACKEN

DATE/TIME OF TEST: 12/10/2025 22:09:32

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:54	
Control Test		0.081	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.214	21:56	
Breath Volume		1.691	21:56	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 2	0.214	22:02	Volume Not Met
Breath Volume		0.039	22:02	
Air Blank		0.000	22:03	
Air Blank		0.000	22:04	
Subject Sample	# 3	0.209	22:07	Volume Not Met
Breath Volume		0.281	22:07	
Air Blank		0.000	22:08	
Control Test		0.076	22:08	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 12/17/2025 14:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 12/13/2025 01:48:32

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.082	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.000	01:44
Breath Volume		3.382	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.000	01:47
Breath Volume		2.523	01:47
Air Blank		0.000	01:47
Control Test		0.082	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 12/17/2025 14:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/13/2025 17:54:29

Diagnostics Check		OK	17:47
Air Blank		0.000	17:48
Control Test		0.081	17:48
Air Blank		0.000	17:48
Subject Sample	# 1	0.109	17:50
Breath Volume		1.750	17:50
Air Blank		0.000	17:50
Air Blank		0.000	17:52
Subject Sample	# 2	0.108	17:52
Breath Volume		1.246	17:52
Air Blank		0.000	17:53
Control Test		0.081	17:53
Air Blank		0.000	17:54
Diagnostics Check		OK	17:54

Data Download Date/Time: 12/17/2025 14:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIDD,DAVID,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KIDD,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: COWART

DATE/TIME OF TEST: 12/16/2025 18:45:58

Diagnostics Check		OK	18:39
Air Blank		0.000	18:40
Control Test		0.082	18:40
Air Blank		0.000	18:41
Subject Sample	# 1	0.096	18:41
Breath Volume		2.757	18:41
Air Blank		0.000	18:42
Air Blank		0.000	18:43
Subject Sample	# 2	0.094	18:44
Breath Volume		3.511	18:44
Air Blank		0.000	18:45
Control Test		0.082	18:45
Air Blank		0.000	18:45
Diagnostics Check		OK	18:45

Data Download Date/Time: 12/17/2025 14:50

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SCOTT,YEATES
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: PEREZ JIMENEZ

DATE/TIME OF TEST: 12/17/2025 17:09:29

Diagnostics Check		OK	17:02
Air Blank		0.000	17:02
Control Test		0.081	17:02
Air Blank		0.000	17:03
Subject Sample	# 1	0.150	17:04
Breath Volume		1.523	17:04
Air Blank		0.000	17:05
Air Blank		0.000	17:06
Subject Sample	# 2	0.137	17:07
Breath Volume		1.359	17:07
Air Blank		0.000	17:08
Control Test		0.082	17:08
Air Blank		0.000	17:09
Diagnostics Check		OK	17:09

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIDD,DAVID,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KIDD,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 12/17/2025 21:02:22

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.081	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.160	20:57
Breath Volume		2.937	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.164	21:00
Breath Volume		3.199	21:00
Air Blank		0.000	21:01
Control Test		0.081	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 12/19/2025 23:00:24

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:52
Subject Sample	# 1	0.123	22:55
Breath Volume		2.441	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:57
Subject Sample	# 2	0.123	22:58
Breath Volume		1.183	22:58
Air Blank		0.000	22:59
Control Test		0.081	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 12/20/2025 01:42:45

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.176	01:37
Breath Volume		1.757	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.174	01:41
Breath Volume		1.726	01:41
Air Blank		0.000	01:41
Control Test		0.082	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: CANCEL PAGAN

DATE/TIME OF TEST: 12/20/2025 18:45:47

Diagnostics Check		OK	18:39
Air Blank		0.000	18:40
Control Test		0.082	18:40
Air Blank		0.000	18:40
Subject Sample	# 1	0.074	18:41
Breath Volume		3.445	18:41
Air Blank		0.000	18:42
Air Blank		0.000	18:43
Subject Sample	# 2	0.073	18:44
Breath Volume		3.035	18:44
Air Blank		0.000	18:44
Control Test		0.082	18:45
Air Blank		0.000	18:45
Diagnostics Check		OK	18:45

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORRALES,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CORRALES,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GROSSI

DATE/TIME OF TEST: 12/21/2025 02:03:45

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.081	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.018	01:59
Breath Volume		3.929	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.017	02:02
Breath Volume		4.671	02:02
Air Blank		0.000	02:02
Control Test		0.082	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/21/2025 23:54:55

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.081	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.000	23:50
Breath Volume		3.554	23:50
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.000	23:53
Breath Volume		3.281	23:53
Air Blank		0.000	23:53
Control Test		0.081	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIDD,DAVID,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KIDD,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: SOLACHE

DATE/TIME OF TEST: 12/24/2025 00:46:43

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.081	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.177	00:42
Breath Volume		1.839	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.175	00:45
Breath Volume		3.027	00:45
Air Blank		0.000	00:45
Control Test		0.081	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,SGT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MONGE

DATE/TIME OF TEST: 12/24/2025 01:54:35

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.000	01:49
Breath Volume		1.480	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.000	01:53
Breath Volume		1.371	01:53
Air Blank		0.000	01:53
Control Test		0.082	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LAMPARELLI,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/25/2025 20:27:30

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.082	20:18
Air Blank		0.000	20:18
Subject Sample	# 1	0.193	20:21
Breath Volume		1.242	20:21
Air Blank		0.000	20:22
Air Blank		0.000	20:23
Subject Sample	# 2	0.190	20:25
Breath Volume		1.316	20:25
Air Blank		0.000	20:26
Control Test		0.081	20:26
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 12/28/2025 21:15:31

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.082	21:09
Air Blank		0.000	21:09
Subject Sample	# 1	0.261	21:10
Breath Volume		3.449	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:12
Subject Sample	# 2	0.246	21:13
Breath Volume		1.601	21:13
Air Blank		0.000	21:14
Control Test		0.082	21:14
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIDD,DAVID,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: THORTON

DATE/TIME OF TEST: 12/31/2025 00:40:03

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.082	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.105	00:35
Breath Volume		5.949	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.106	00:38
Breath Volume		5.539	00:38
Air Blank		0.000	00:39
Control Test		0.082	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:40

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: DOWNS

DATE/TIME OF TEST: 12/31/2025 03:10:41

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:58	
Control Test		0.082	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1	0.155	02:59	
Breath Volume		1.207	02:59	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 2		03:05	No Sample Provided
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 3	0.153	03:09	
Breath Volume		2.113	03:09	
Air Blank		0.000	03:09	
Control Test		0.082	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 01/23/2026 19:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 12/31/2025 23:22:58

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.082	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.055	23:18
Breath Volume		2.316	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:20
Subject Sample	# 2	0.056	23:21
Breath Volume		2.187	23:21
Air Blank		0.000	23:22
Control Test		0.083	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 01/23/2026 19:31

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VECCHIARELLI,N
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MUJIC

DATE/TIME OF TEST: 12/01/2025 00:18:59

Diagnostics Check		OK	00:16	
Air Blank		0.000	00:16	
Control Test		0.080	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 1		00:17	Subject Test Refused
Breath Volume		0.000	00:17	
Air Blank		0.000	00:18	
Control Test		0.080	00:18	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINA,C,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 12/04/2025 04:48:43

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.242	04:40
Breath Volume		3.414	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.211	04:44
Breath Volume		1.523	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 3	0.211	04:47
Breath Volume		3.710	04:47
Air Blank		0.000	04:47
Control Test		0.079	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 12/06/2025 00:03:14

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 12/29/2025 10:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/07/2025 19:39:00

Diagnostics Check		OK	19:36	
Air Blank		0.000	19:36	
Control Test		0.081	19:36	
Air Blank		0.000	19:37	
Subject Sample	# 1		19:37	Subject Test Refused
Breath Volume		0.000	19:37	
Air Blank		0.000	19:38	
Control Test		0.080	19:38	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:39	

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BOURET

DATE/TIME OF TEST: 12/10/2025 02:20:25

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.251	02:14
Breath Volume		1.683	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.232	02:18
Breath Volume		1.921	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OWENS,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DORE

DATE/TIME OF TEST: 12/13/2025 00:38:55

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:36	
Control Test		0.081	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GUIFARRA GONZALEZ

DATE/TIME OF TEST: 12/15/2025 00:12:57

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.349	00:07
Breath Volume		2.519	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.338	00:11
Breath Volume		2.160	00:11
Air Blank		0.000	00:11
Control Test		0.082	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: BURKHOLDER

DATE/TIME OF TEST: 12/18/2025 23:32:26

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.081	23:25
Air Blank		0.000	23:26
Subject Sample	# 1	0.094	23:27
Breath Volume		1.781	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:29
Subject Sample	# 2	0.098	23:30
Breath Volume		2.363	23:30
Air Blank		0.000	23:31
Control Test		0.078	23:31
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 12/20/2025 01:13:08

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:10	
Control Test		0.081	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:12	
Control Test		0.080	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 12/20/2025 03:57:32

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.107	03:52
Breath Volume		2.222	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.111	03:55
Breath Volume		1.773	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 12/29/2025 10:19

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 12/21/2025 05:16:08

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.081	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.130	05:10
Breath Volume		1.765	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.128	05:14
Breath Volume		1.570	05:14
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 12/29/2025 10:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: PERES MATTOS

DATE/TIME OF TEST: 12/21/2025 17:34:26

Diagnostics Check		OK	17:27
Air Blank		0.000	17:28
Control Test		0.081	17:28
Air Blank		0.000	17:29
Subject Sample	# 1	0.182	17:29
Breath Volume		2.312	17:29
Air Blank		0.000	17:30
Air Blank		0.000	17:32
Subject Sample	# 2	0.185	17:32
Breath Volume		2.484	17:32
Air Blank		0.000	17:33
Control Test		0.080	17:33
Air Blank		0.000	17:34
Diagnostics Check		OK	17:34

Data Download Date/Time: 12/29/2025 10:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: KAMEKA

DATE/TIME OF TEST: 12/27/2025 15:42:01

Diagnostics Check		OK	15:35
Air Blank		0.000	15:36
Control Test		0.079	15:36
Air Blank		0.000	15:37
Subject Sample	# 1	0.160	15:37
Breath Volume		1.769	15:37
Air Blank		0.000	15:38
Air Blank		0.000	15:39
Subject Sample	# 2	0.170	15:40
Breath Volume		2.023	15:40
Air Blank		0.000	15:41
Control Test		0.079	15:41
Air Blank		0.000	15:41
Diagnostics Check		OK	15:42

Data Download Date/Time: 12/29/2025 10:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: NAPIER

DATE/TIME OF TEST: 12/28/2025 00:04:27

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:03	
Control Test		0.079	00:03	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 12/29/2025 10:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: NAPIER

DATE/TIME OF TEST: 12/28/2025 00:21:48

Diagnostics Check		OK	00:18	
Air Blank		0.000	00:19	
Control Test		0.080	00:19	
Air Blank		0.000	00:20	
Subject Sample	# 1		00:20	Subject Test Refused
Breath Volume		0.000	00:20	
Air Blank		0.000	00:20	
Control Test		0.079	00:21	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 12/29/2025 10:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FLANDERS,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 12/28/2025 06:57:56

Diagnostics Check		OK	06:50
Air Blank		0.000	06:50
Control Test		0.080	06:50
Air Blank		0.000	06:51
Subject Sample	# 1	0.185	06:52
Breath Volume		1.339	06:52
Air Blank		0.000	06:53
Air Blank		0.000	06:55
Subject Sample	# 2	0.187	06:56
Breath Volume		1.582	06:56
Air Blank		0.000	06:57
Control Test		0.080	06:57
Air Blank		0.000	06:57
Diagnostics Check		OK	06:57

Data Download Date/Time: 12/29/2025 10:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 12/29/2025 05:09:45

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.080	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.115	05:04
Breath Volume		2.644	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.113	05:08
Breath Volume		2.312	05:08
Air Blank		0.000	05:08
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 12/29/2025 10:20

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: PONNELL,R
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 11/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/21/2025 17:45:04

Diagnostics Check		OK	17:38
Air Blank		0.000	17:39
Control Test		0.079	17:39
Air Blank		0.000	17:40
Subject Sample	# 1	0.154	17:40
Breath Volume		2.242	17:40
Air Blank		0.000	17:41
Air Blank		0.000	17:43
Subject Sample	# 2	0.160	17:43
Breath Volume		3.187	17:43
Air Blank		0.000	17:44
Control Test		0.078	17:44
Air Blank		0.000	17:44
Diagnostics Check		OK	17:45

Data Download Date/Time: 12/28/2025 18:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES,JESSE,E
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: BOYLE,MATTHEW
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CHIGLO

DATE/TIME OF TEST: 12/29/2025 22:19:17

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.081	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.187	22:13	
Breath Volume		3.164	22:13	
Air Blank		0.000	22:14	
Air Blank		0.000	22:16	
Subject Sample	# 2		22:18	Sequence Aborted
Breath Volume		0.164	22:18	
Air Blank		0.000	22:19	

Data Download Date/Time: 01/01/2026 20:14

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SEIN-CABAN,G
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: AHMAD

DATE/TIME OF TEST: 12/14/2025 04:33:56

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.138	04:29
Breath Volume		1.503	04:29
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.137	04:32
Breath Volume		2.195	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 12/31/2025 14:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MURRELL

DATE/TIME OF TEST: 12/15/2025 02:46:51

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.079	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.143	02:41
Breath Volume		1.113	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.154	02:45
Breath Volume		1.976	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 12/31/2025 14:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SYMONS,DEPUTY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: ALVES

DATE/TIME OF TEST: 12/17/2025 08:08:53

Diagnostics Check		OK	08:00
Air Blank		0.000	08:00
Control Test		0.079	08:01
Air Blank		0.000	08:01
Subject Sample	# 1	0.142	08:02
Breath Volume		1.593	08:02
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.136	08:07
Breath Volume		1.351	08:07
Air Blank		0.000	08:07
Control Test		0.079	08:08
Air Blank		0.000	08:08
Diagnostics Check		OK	08:08

Data Download Date/Time: 12/31/2025 14:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 12/23/2025 11:50:50

Diagnostics Check		OK	11:44
Air Blank		0.000	11:45
Control Test		0.078	11:45
Air Blank		0.000	11:45
Subject Sample	# 1	0.000	11:46
Breath Volume		1.636	11:46
Air Blank		0.000	11:46
Air Blank		0.000	11:48
Subject Sample	# 2	0.000	11:49
Breath Volume		1.464	11:49
Air Blank		0.000	11:49
Control Test		0.080	11:50
Air Blank		0.000	11:50
Diagnostics Check		OK	11:50

Data Download Date/Time: 12/31/2025 14:57

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DEVARASETTY

DATE/TIME OF TEST: 12/27/2025 03:09:34

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.097	03:05
Breath Volume		1.871	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.095	03:08
Breath Volume		2.000	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 12/31/2025 14:58

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,SOLOMON,
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: MALDONADO,SOLOMON
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 12/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/30/2025 11:28:52

Diagnostics Check		OK	11:23
Air Blank		0.000	11:23
Control Test		0.079	11:23
Air Blank		0.000	11:24
Subject Sample	# 1	0.000	11:24
Breath Volume		3.156	11:24
Air Blank		0.000	11:25
Air Blank		0.000	11:27
Subject Sample	# 2	0.000	11:27
Breath Volume		3.335	11:27
Air Blank		0.000	11:27
Control Test		0.080	11:28
Air Blank		0.000	11:28
Diagnostics Check		OK	11:28

Data Download Date/Time: 03/03/2026 09:45

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,SOLEY,G
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SKINNER,OWEN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MESTIZA MELCHOR

DATE/TIME OF TEST: 12/14/2025 01:02:35

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.122	00:58
Breath Volume		3.164	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.122	01:01
Breath Volume		4.054	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 02/01/2026 00:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SKINNER,OWEN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: SCHOTTENHEIMER

DATE/TIME OF TEST: 12/20/2025 21:14:43

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.081	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.084	21:10
Breath Volume		1.160	21:10
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.082	21:13
Breath Volume		1.234	21:13
Air Blank		0.000	21:13
Control Test		0.080	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 02/01/2026 00:59

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KOPF,NICHOLAS
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/02/2025 14:11:06

Diagnostics Check		OK	14:04
Air Blank		0.000	14:05
Control Test		0.079	14:05
Air Blank		0.000	14:05
Subject Sample	# 1	0.000	14:06
Breath Volume		1.949	14:06
Air Blank		0.000	14:07
Air Blank		0.000	14:09
Subject Sample	# 2	0.000	14:09
Breath Volume		1.964	14:09
Air Blank		0.000	14:10
Control Test		0.079	14:10
Air Blank		0.000	14:11
Diagnostics Check		OK	14:11

Data Download Date/Time: 12/18/2025 10:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: BAUMBARGER

DATE/TIME OF TEST: 12/05/2025 16:52:52

Diagnostics Check		OK	16:46
Air Blank		0.000	16:46
Control Test		0.079	16:47
Air Blank		0.000	16:47
Subject Sample	# 1	0.256	16:48
Breath Volume		3.281	16:48
Air Blank		0.000	16:49
Air Blank		0.000	16:50
Subject Sample	# 2	0.251	16:51
Breath Volume		2.867	16:51
Air Blank		0.000	16:51
Control Test		0.079	16:52
Air Blank		0.000	16:52
Diagnostics Check		OK	16:52

Data Download Date/Time: 12/18/2025 10:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: CLARK,RYAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LUSTER

DATE/TIME OF TEST: 12/06/2025 01:39:28

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:36	
Control Test		0.079	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 12/18/2025 10:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: CLARK,RYAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: DE OLIVEIRA DE SOUZA

DATE/TIME OF TEST: 12/13/2025 20:58:58

Diagnostics Check		OK	20:52
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:53
Subject Sample	# 1	0.152	20:54
Breath Volume		3.628	20:54
Air Blank		0.000	20:55
Air Blank		0.000	20:57
Subject Sample	# 2	0.153	20:57
Breath Volume		3.351	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:58
Diagnostics Check		OK	20:58

Data Download Date/Time: 12/18/2025 10:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HADDOCK,CLAY,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KOPF,NICHOLAS
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 12/15/2025 07:49:29

Diagnostics Check		OK	07:43
Air Blank		0.000	07:43
Control Test		0.080	07:44
Air Blank		0.000	07:44
Subject Sample	# 1	0.076	07:45
Breath Volume		2.621	07:45
Air Blank		0.000	07:45
Air Blank		0.000	07:47
Subject Sample	# 2	0.075	07:48
Breath Volume		2.386	07:48
Air Blank		0.000	07:48
Control Test		0.080	07:48
Air Blank		0.000	07:49
Diagnostics Check		OK	07:49

Data Download Date/Time: 12/18/2025 10:29

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: CLARK,RYAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 12/18/2025 21:39:42

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:28	
Control Test		0.079	21:28	
Air Blank		0.000	21:29	
Subject Sample	# 1	0.182	21:32	Volume Not Met
Breath Volume		0.914	21:32	
Air Blank		0.000	21:33	
Air Blank		0.000	21:35	
Subject Sample	# 2	0.170	21:38	Volume Not Met
Breath Volume		0.785	21:38	
Air Blank		0.000	21:38	
Control Test		0.079	21:39	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 01/20/2026 12:05

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 12/20/2025 01:28:14

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.134	01:21	Volume Not Met
Breath Volume		0.109	01:21	
Air Blank		0.000	01:21	
Air Blank		0.000	01:23	
Subject Sample	# 2	0.099	01:26	Volume Not Met
Breath Volume		0.058	01:26	
Air Blank		0.000	01:27	
Control Test		0.080	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 01/20/2026 12:05

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 12/20/2025 02:07:35

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.211	02:03
Breath Volume		1.511	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.213	02:06
Breath Volume		1.187	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 01/20/2026 12:05

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: VALLE

DATE/TIME OF TEST: 12/27/2025 00:03:20

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 01/20/2026 12:05

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 12/31/2025 23:17:16

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:14	
Control Test		0.080	23:15	
Air Blank		0.000	23:15	
Subject Sample	# 1		23:15	Subject Test Refused
Breath Volume		0.000	23:15	
Air Blank		0.000	23:16	
Control Test		0.080	23:16	
Air Blank		0.000	23:17	
Diagnostics Check		OK	23:17	

Data Download Date/Time: 01/20/2026 12:05

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/01/2025 03:24:27

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.077	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.080	03:15	Slope Not Level
Breath Volume		1.359	03:15	
Air Blank		0.000	03:16	
Air Blank		0.000	03:18	
Subject Sample	# 2	0.081	03:18	
Breath Volume		1.187	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 3	0.078	03:22	
Breath Volume		1.160	03:22	
Air Blank		0.000	03:23	
Control Test		0.077	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,PETER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: HETTICK

DATE/TIME OF TEST: 12/02/2025 02:47:19

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.077	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.132	02:42
Breath Volume		1.980	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.124	02:45
Breath Volume		1.843	02:45
Air Blank		0.000	02:46
Control Test		0.077	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: SERMON

DATE/TIME OF TEST: 12/03/2025 22:39:44

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.078	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.189	22:32
Breath Volume		1.230	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.190	22:38
Breath Volume		1.371	22:38
Air Blank		0.000	22:38
Control Test		0.077	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MCGEHEE

DATE/TIME OF TEST: 12/04/2025 05:12:25

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.078	05:06
Air Blank		0.000	05:07
Subject Sample	# 1	0.071	05:07
Breath Volume		1.519	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.066	05:10
Breath Volume		1.210	05:10
Air Blank		0.000	05:11
Control Test		0.078	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: VANHORN

DATE/TIME OF TEST: 12/06/2025 01:17:58

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.077	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.135	01:13
Breath Volume		1.796	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.130	01:16
Breath Volume		2.046	01:16
Air Blank		0.000	01:17
Control Test		0.077	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/06/2025 03:46:54

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.076	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.159	03:41
Breath Volume		1.519	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.150	03:45
Breath Volume		1.468	03:45
Air Blank		0.000	03:45
Control Test		0.077	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ISIC,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 12/07/2025 02:43:59

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.077	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.178	02:39
Breath Volume		2.304	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.178	02:42
Breath Volume		1.589	02:42
Air Blank		0.000	02:43
Control Test		0.077	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: UPTON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/07/2025 04:59:06

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.077	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.138	04:54
Breath Volume		1.160	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.158	04:57
Breath Volume		1.132	04:57
Air Blank		0.000	04:58
Control Test		0.076	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DOMINGO GODINEZ

DATE/TIME OF TEST: 12/08/2025 02:54:03

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.078	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.077	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 12/28/2025 20:01

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/08/2025 22:37:01

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:34	
Control Test		0.078	22:34	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:35	Subject Test Refused
Breath Volume		0.000	22:35	
Air Blank		0.000	22:36	
Control Test		0.078	22:36	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/09/2025 00:41:17

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.077	00:39	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:39	Subject Test Refused
Breath Volume		0.000	00:39	
Air Blank		0.000	00:40	
Control Test		0.078	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KOHLER,E
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: KREITZER

DATE/TIME OF TEST: 12/09/2025 04:43:07

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.078	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:42	
Control Test		0.077	04:42	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: AUDE MARK

DATE/TIME OF TEST: 12/09/2025 23:01:35

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.078	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.140	22:57
Breath Volume		2.304	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.137	23:00
Breath Volume		2.054	23:00
Air Blank		0.000	23:00
Control Test		0.076	23:00
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 12/11/2025 20:01:03

Diagnostics Check		OK	19:52
Air Blank		0.000	19:52
Control Test		0.078	19:53
Air Blank		0.000	19:53
Subject Sample	# 1	0.181	19:54
Breath Volume		2.105	19:54
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.167	19:59
Breath Volume		1.816	19:59
Air Blank		0.000	20:00
Control Test		0.077	20:00
Air Blank		0.000	20:00
Diagnostics Check		OK	20:01

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/12/2025 22:23:26

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.077	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.196	22:18
Breath Volume		1.625	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.186	22:21
Breath Volume		1.222	22:21
Air Blank		0.000	22:22
Control Test		0.076	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STARLING,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: CRENSHAW

DATE/TIME OF TEST: 12/13/2025 06:29:48

Diagnostics Check		OK	06:15
Air Blank		0.000	06:15
Control Test		0.078	06:16
Air Blank		0.000	06:16
Subject Sample	# 1	0.154	06:18
Breath Volume		1.429	06:18
Air Blank		0.000	06:19
Air Blank		0.000	06:21
Subject Sample	# 2	0.132	06:22
Breath Volume		1.238	06:22
Air Blank		0.000	06:23
Air Blank		0.000	06:25
Subject Sample	# 3	0.130	06:28
Breath Volume		2.113	06:28
Air Blank		0.000	06:28
Control Test		0.077	06:29
Air Blank		0.000	06:29
Diagnostics Check		OK	06:29

Data Download Date/Time: 12/28/2025 20:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHAW,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: NEGRON SALDANA

DATE/TIME OF TEST: 12/14/2025 00:42:08

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.077	00:40	
Air Blank		0.000	00:40	
Subject Sample	# 1		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:41	
Control Test		0.078	00:41	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/14/2025 04:39:18

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.078	04:37	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:38	
Control Test		0.077	04:38	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/14/2025 04:44:48

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:42	
Control Test		0.077	04:42	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:43	
Control Test		0.078	04:44	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HUGHES,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: CHIARLE

DATE/TIME OF TEST: 12/14/2025 07:39:46

Diagnostics Check		OK	07:32
Air Blank		0.000	07:33
Control Test		0.077	07:33
Air Blank		0.000	07:34
Subject Sample	# 1	0.036	07:35
Breath Volume		2.656	07:35
Air Blank		0.000	07:35
Air Blank		0.000	07:37
Subject Sample	# 2	0.035	07:38
Breath Volume		2.355	07:38
Air Blank		0.000	07:38
Control Test		0.076	07:39
Air Blank		0.000	07:39
Diagnostics Check		OK	07:39

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MINCEY

DATE/TIME OF TEST: 12/14/2025 20:20:05

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.078	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.000	20:15
Breath Volume		2.929	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.000	20:18
Breath Volume		2.511	20:18
Air Blank		0.000	20:19
Control Test		0.078	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GOODIN,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/14/2025 23:47:34

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:45	
Control Test		0.079	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:46	Subject Test Refused
Breath Volume		0.000	23:46	
Air Blank		0.000	23:46	
Control Test		0.079	23:46	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: SELZER

DATE/TIME OF TEST: 12/15/2025 01:50:23

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.068	01:45
Breath Volume		3.769	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.063	01:48
Breath Volume		3.386	01:48
Air Blank		0.000	01:49
Control Test		0.078	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/15/2025 23:06:12

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:03	
Control Test		0.078	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1		23:04	Subject Test Refused
Breath Volume		0.000	23:04	
Air Blank		0.000	23:05	
Control Test		0.079	23:05	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 12/16/2025 01:52:48

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.078	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.000	01:48
Breath Volume		2.226	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.000	01:51
Breath Volume		2.097	01:51
Air Blank		0.000	01:51
Control Test		0.078	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BADZ

DATE/TIME OF TEST: 12/16/2025 19:39:04

Diagnostics Check		OK	19:30	
Air Blank		0.000	19:31	
Control Test		0.078	19:31	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:35	No Sample Provided
Breath Volume		0.000	19:35	
Air Blank		0.000	19:35	
Air Blank		0.000	19:37	
Subject Sample	# 2		19:37	Subject Test Refused
Breath Volume		0.000	19:37	
Air Blank		0.000	19:38	
Control Test		0.077	19:38	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:39	

Data Download Date/Time: 12/28/2025 20:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: BUENO DIZ

DATE/TIME OF TEST: 12/17/2025 04:50:47

Diagnostics Check		OK	04:44
Air Blank		0.000	04:44
Control Test		0.078	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.194	04:46
Breath Volume		1.855	04:46
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.196	04:49
Breath Volume		2.210	04:49
Air Blank		0.000	04:49
Control Test		0.077	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/17/2025 21:58:57

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.077	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.188	21:54
Breath Volume		2.093	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:56
Subject Sample	# 2	0.183	21:57
Breath Volume		1.921	21:57
Air Blank		0.000	21:58
Control Test		0.077	21:58
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 12/28/2025 20:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOOD,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/17/2025 22:54:06

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:51	
Control Test		0.077	22:52	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:52	Subject Test Refused
Breath Volume		0.000	22:52	
Air Blank		0.000	22:53	
Control Test		0.077	22:53	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/19/2025 03:45:10

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:42	
Control Test		0.078	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:43	Subject Test Refused
Breath Volume		0.000	03:43	
Air Blank		0.000	03:44	
Control Test		0.078	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 12/28/2025 20:02

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 12/19/2025 22:18:51

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.078	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.089	22:14
Breath Volume		1.246	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.096	22:17
Breath Volume		1.242	22:17
Air Blank		0.000	22:17
Control Test		0.077	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: ZWICK

DATE/TIME OF TEST: 12/20/2025 00:26:28

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:24	
Control Test		0.079	00:24	
Air Blank		0.000	00:24	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Control Test		0.078	00:25	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HASTON

DATE/TIME OF TEST: 12/20/2025 01:05:04

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.077	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.150	01:00
Breath Volume		1.363	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.156	01:03
Breath Volume		1.476	01:03
Air Blank		0.000	01:04
Control Test		0.077	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:05

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LACERDA

DATE/TIME OF TEST: 12/20/2025 03:57:54

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.078	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.152	03:52
Breath Volume		1.152	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.169	03:56
Breath Volume		1.578	03:56
Air Blank		0.000	03:56
Control Test		0.078	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 12/20/2025 04:55:02

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:52	
Control Test		0.078	04:52	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:53	Subject Test Refused
Breath Volume		0.000	04:53	
Air Blank		0.000	04:54	
Control Test		0.079	04:54	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 12/20/2025 21:14:41

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:12	
Control Test		0.079	21:12	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:13	Subject Test Refused
Breath Volume		0.000	21:13	
Air Blank		0.000	21:13	
Control Test		0.078	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 12/21/2025 00:13:52

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.078	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.231	00:09
Breath Volume		1.507	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.234	00:12
Breath Volume		2.257	00:12
Air Blank		0.000	00:12
Control Test		0.077	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YATES,M,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAMASCEN,A
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 12/21/2025 17:14:07

Diagnostics Check		OK	17:11	
Air Blank		0.000	17:11	
Control Test		0.078	17:12	
Air Blank		0.000	17:12	
Subject Sample	# 1		17:12	Subject Test Refused
Breath Volume		0.000	17:12	
Air Blank		0.000	17:13	
Control Test		0.077	17:13	
Air Blank		0.000	17:14	
Diagnostics Check		OK	17:14	

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,D,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: ROBINSON-TUCKER

DATE/TIME OF TEST: 12/21/2025 20:29:11

Diagnostics Check		OK	20:18
Air Blank		0.000	20:19
Control Test		0.078	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.186	20:22
Breath Volume		1.882	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.181	20:27
Breath Volume		1.179	20:27
Air Blank		0.000	20:28
Control Test		0.077	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: FETAHOVIC

DATE/TIME OF TEST: 12/22/2025 01:27:46

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:25	
Control Test		0.078	01:25	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:26	
Control Test		0.078	01:27	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLS,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 12/23/2025 19:24:23

Diagnostics Check		OK	19:17
Air Blank		0.000	19:17
Control Test		0.077	19:17
Air Blank		0.000	19:18
Subject Sample	# 1	0.139	19:19
Breath Volume		1.449	19:19
Air Blank		0.000	19:20
Air Blank		0.000	19:22
Subject Sample	# 2	0.144	19:22
Breath Volume		2.664	19:22
Air Blank		0.000	19:23
Control Test		0.077	19:23
Air Blank		0.000	19:24
Diagnostics Check		OK	19:24

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/23/2025 22:09:31

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:07	
Control Test		0.078	22:07	
Air Blank		0.000	22:08	
Subject Sample	# 1		22:08	Subject Test Refused
Breath Volume		0.000	22:08	
Air Blank		0.000	22:08	
Control Test		0.078	22:08	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 12/25/2025 02:09:26

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.090	02:04
Breath Volume		1.636	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.096	02:07
Breath Volume		1.636	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: GILDBERG

DATE/TIME OF TEST: 12/25/2025 21:17:44

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.078	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.191	21:12
Breath Volume		1.750	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.194	21:16
Breath Volume		2.769	21:16
Air Blank		0.000	21:16
Control Test		0.077	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CROCKETT

DATE/TIME OF TEST: 12/26/2025 02:40:59

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.078	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.091	02:36
Breath Volume		1.894	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:38
Subject Sample	# 2	0.090	02:39
Breath Volume		1.171	02:39
Air Blank		0.000	02:40
Control Test		0.077	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TURPIN,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 12/26/2025 05:35:35

Diagnostics Check		OK	05:32	
Air Blank		0.000	05:33	
Control Test		0.079	05:33	
Air Blank		0.000	05:33	
Subject Sample	# 1		05:34	Subject Test Refused
Breath Volume		0.000	05:34	
Air Blank		0.000	05:34	
Control Test		0.078	05:34	
Air Blank		0.000	05:35	
Diagnostics Check		OK	05:35	

Data Download Date/Time: 12/28/2025 20:03

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 12/28/2025 00:42:21

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.078	00:39	
Air Blank		0.000	00:40	
Subject Sample	# 1		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:41	
Control Test		0.078	00:41	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 12/28/2025 20:04

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/28/2025 03:11:09

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.077	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.202	03:06
Breath Volume		1.746	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.203	03:09
Breath Volume		1.371	03:09
Air Blank		0.000	03:10
Control Test		0.077	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 12/28/2025 20:04

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 12/28/2025 06:15:36

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.078	06:09
Air Blank		0.000	06:09
Subject Sample	# 1	0.253	06:10
Breath Volume		3.347	06:10
Air Blank		0.000	06:11
Air Blank		0.000	06:13
Subject Sample	# 2	0.257	06:14
Breath Volume		2.941	06:14
Air Blank		0.000	06:14
Control Test		0.076	06:14
Air Blank		0.000	06:15
Diagnostics Check		OK	06:15

Data Download Date/Time: 12/28/2025 20:04

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEDERBERG,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 12/29/2025 02:10:41

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.169	02:06
Breath Volume		1.351	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.163	02:09
Breath Volume		1.378	02:09
Air Blank		0.000	02:09
Control Test		0.078	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 01/07/2026 10:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: GRILL

DATE/TIME OF TEST: 12/29/2025 21:05:14

Diagnostics Check		OK	21:02	
Air Blank		0.000	21:02	
Control Test		0.079	21:03	
Air Blank		0.000	21:03	
Subject Sample	# 1		21:03	Subject Test Refused
Breath Volume		0.000	21:03	
Air Blank		0.000	21:04	
Control Test		0.079	21:04	
Air Blank		0.000	21:05	
Diagnostics Check		OK	21:05	

Data Download Date/Time: 01/07/2026 10:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DICKSON,R
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: RUDOLPH

DATE/TIME OF TEST: 12/29/2025 23:41:51

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.079	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.107	23:37
Breath Volume		1.816	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.107	23:40
Breath Volume		1.429	23:40
Air Blank		0.000	23:40
Control Test		0.079	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 01/07/2026 10:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BARNWELL

DATE/TIME OF TEST: 12/31/2025 03:12:52

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:01	
Control Test		0.079	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1	0.047	03:05	Volume Not Met
Breath Volume		0.437	03:05	
Air Blank		0.000	03:05	
Air Blank		0.000	03:07	
Subject Sample	# 2	0.053	03:08	
Breath Volume		1.218	03:08	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 3	0.050	03:11	
Breath Volume		1.160	03:11	
Air Blank		0.000	03:11	
Control Test		0.079	03:12	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 01/07/2026 10:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A1
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: MOLLNOW

DATE/TIME OF TEST: 12/31/2025 19:50:58

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.079	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.149	19:46
Breath Volume		2.382	19:46
Air Blank		0.000	19:46
Air Blank		0.000	19:48
Subject Sample	# 2	0.153	19:49
Breath Volume		1.683	19:49
Air Blank		0.000	19:50
Control Test		0.079	19:50
Air Blank		0.000	19:50
Diagnostics Check		OK	19:50

Data Download Date/Time: 01/07/2026 10:40

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: DEMIRANDA LIMA

DATE/TIME OF TEST: 12/07/2025 00:15:06

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:08	
Control Test		0.079	00:08	
Air Blank		0.000	00:08	
Subject Sample	# 1	0.132	00:10	
Breath Volume		1.800	00:10	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.122	00:13	
Breath Volume		1.839	00:13	
Air Blank		0.000	00:13	
Control Test		0.079	00:14	
Air Blank		0.000	00:14	
Diagnostics Check		Fail	00:14	Analytical Stability Test
Air Blank		0.000	00:15	

Data Download Date/Time: 01/03/2026 12:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 12/07/2025 02:21:13

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.218	02:19	
Air Blank		0.000	02:20	
Control Test		0.082	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 01/03/2026 12:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/10/2025 19:42:18

Diagnostics Check		OK	19:30	
Air Blank		0.000	19:31	
Control Test		0.079	19:31	
Air Blank		0.000	19:32	
Subject Sample	# 1	0.110	19:35	Volume Not Met
Breath Volume		0.183	19:35	
Air Blank		0.000	19:35	
Air Blank		0.000	19:37	
Subject Sample	# 2		19:40	Subject Test Refused
Breath Volume		0.199	19:40	
Air Blank		0.000	19:40	
Control Test		0.077	19:41	
Air Blank		0.000	19:41	
Diagnostics Check		Fail	19:41	Analytical Stability Test
Air Blank		0.000	19:42	

Data Download Date/Time: 01/03/2026 12:24

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLLER,LOGAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: KOLLER,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: M610477671900

DATE/TIME OF TEST: 12/06/2025 03:31:37

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.082	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.189	03:26
Breath Volume		1.550	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.194	03:30
Breath Volume		1.832	03:30
Air Blank		0.000	03:30
Control Test		0.081	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 12/31/2025 11:37

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,DAVID,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JUDAH,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: RAMIREZ RAMIREZ

DATE/TIME OF TEST: 12/31/2025 16:47:01

Diagnostics Check		OK	16:40
Air Blank		0.000	16:41
Control Test		0.082	16:41
Air Blank		0.000	16:42
Subject Sample	# 1	0.220	16:42
Breath Volume		1.285	16:42
Air Blank		0.000	16:43
Air Blank		0.000	16:45
Subject Sample	# 2	0.223	16:45
Breath Volume		2.160	16:45
Air Blank		0.000	16:46
Control Test		0.081	16:46
Air Blank		0.000	16:46
Diagnostics Check		OK	16:47

Data Download Date/Time: 02/02/2026 10:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JOSHUA,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BIGGS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 12/22/2025 14:07:23

Diagnostics Check		OK	14:00
Air Blank		0.000	14:01
Control Test		0.080	14:01
Air Blank		0.000	14:01
Subject Sample	# 1	0.137	14:02
Breath Volume		2.289	14:02
Air Blank		0.000	14:03
Air Blank		0.000	14:05
Subject Sample	# 2	0.137	14:05
Breath Volume		2.593	14:05
Air Blank		0.000	14:06
Control Test		0.081	14:06
Air Blank		0.000	14:07
Diagnostics Check		OK	14:07

Data Download Date/Time: 01/19/2026 16:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ELMORE,TT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: WASHBURN

DATE/TIME OF TEST: 12/27/2025 04:05:20

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.156	03:59
Breath Volume		1.347	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.161	04:03
Breath Volume		1.828	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 01/19/2026 16:10

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUMMOND,ALEX,V
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ADAMS,PAUL
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 11/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 12/12/2025 20:07:46

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.083	20:00
Air Blank		0.000	20:01
Subject Sample	# 1	0.234	20:02
Breath Volume		1.820	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.233	20:06
Breath Volume		1.894	20:06
Air Blank		0.000	20:06
Control Test		0.081	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 12/23/2025 18:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAUENHOFER,TIMOTHY,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FRAUENHOFER,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: ALCANTARA PALESTINA

DATE/TIME OF TEST: 12/13/2025 08:02:51

Diagnostics Check		OK	07:56
Air Blank		0.000	07:56
Control Test		0.084	07:56
Air Blank		0.000	07:57
Subject Sample	# 1	0.000	07:58
Breath Volume		2.257	07:58
Air Blank		0.000	07:59
Air Blank		0.000	08:01
Subject Sample	# 2	0.000	08:01
Breath Volume		2.156	08:01
Air Blank		0.000	08:01
Control Test		0.082	08:02
Air Blank		0.000	08:02
Diagnostics Check		OK	08:02

Data Download Date/Time: 12/23/2025 18:34

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,HUNTER,R
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: WILLIAMS,CALIBER
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 11/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 12/23/2025 00:04:05

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.082	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.193	23:58
Breath Volume		1.746	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.199	00:02
Breath Volume		1.789	00:02
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 12/23/2025 18:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTA,BRITTNEY,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MCNABB,RENEE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/10/2025 21:58:57

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:46	
Control Test		0.079	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 1		21:50	No Sample Provided
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Air Blank		0.000	21:52	
Subject Sample	# 2	0.063	21:54	
Breath Volume		1.363	21:54	
Air Blank		0.000	21:55	
Air Blank		0.000	21:57	
Subject Sample	# 3	0.059	21:57	
Breath Volume		1.218	21:57	
Air Blank		0.000	21:58	
Control Test		0.076	21:58	
Air Blank		0.000	21:58	
Diagnostics Check		OK	21:58	

Data Download Date/Time: 12/22/2025 11:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTA,BRITTNEY,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MCNABB,RENEE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/12/2025 01:43:09

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.077	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.230	01:38
Breath Volume		1.257	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.235	01:41
Breath Volume		1.199	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 12/22/2025 11:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WOODIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: LANEY

DATE/TIME OF TEST: 12/12/2025 20:38:56

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.079	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.208	20:34
Breath Volume		2.667	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.193	20:37
Breath Volume		2.085	20:37
Air Blank		0.000	20:38
Control Test		0.078	20:38
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 12/22/2025 11:23

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTA,BRITTNEY,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HUGHES,ADIAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/25/2025 01:59:29

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.078	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.117	01:54
Breath Volume		1.910	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.115	01:57
Breath Volume		2.109	01:57
Air Blank		0.000	01:58
Control Test		0.077	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 01/30/2026 16:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RIVERA,JOSE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CONCANNON

DATE/TIME OF TEST: 12/27/2025 04:21:09

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.203	04:16
Breath Volume		2.093	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.207	04:19
Breath Volume		2.574	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 01/30/2026 16:30

Printed: 07 March 2026

COBRA ©1997-2014 CMI, Inc.