

CERTIFICATE OF ANALYSIS

Alcohol Countermeasure Systems
Alcohol Reference Solution
Lot No: 201404C

Expiry Date: April 08, 2016

This solution was analyzed using the direct injection gas-chromatographic procedure coupled with the internal standard technique commensurate with forensic alcohol methodology.

A screening (pre-mix) analysis of the distilled water used in the preparation of this solution indicated that there were no volatile impurities present.


The target analytical concentration was 0.0968 – 0.0992 gram/dL (wt/vol) ethyl alcohol in aqueous solution with an equivalent vapour alcohol concentration of 0.0080 - 0.081 g/210L when used in a breath alcohol simulator heated to $34 \pm 0.2^\circ\text{C}$.

The solution was found to have an analytical concentration of 0.0969 gram/dL (wt/vol) ethyl alcohol in aqueous solution with an equivalent vapour alcohol concentration of 0.081 g/210L when used in a breath alcohol simulator heated to $34 \pm 0.2^\circ\text{C}$.

The solution is manufactured from distilled/de-ionized water and absolute ethyl alcohol (USP Grade). This lot contains 1640 bottles (of 500 mL each) of Alcohol Reference Solution.



Date of Analysis: April 08, 2014


G. J. Kupferschmidt, B.Sc., M.Sc., MCIC, C.Chem
G. Kupferschmidt Consulting Services Ltd.