



Alcohol Countermeasure Systems Corp
60 International Boulevard
Toronto, Ontario M9W 6J2
CANADA
acs-corp.com

CERTIFICATE OF ANALYSIS

Alcohol Countermeasure Systems
Alcohol Reference Solution
Lot No: 201702B

Expiry Date: February 23, 2019

This solution was analyzed using the direct injection gas-chromatographic procedure coupled with the internal standard technique commensurate with forensic alcohol methodology.


A screening (pre-mix) analysis of the distilled water used in the preparation of this solution indicated that there were no volatile impurities present.

The target analytical concentration was 0.242 – 0.248 gram/dL (wt/vol) ethyl alcohol in aqueous solution with an equivalent vapour alcohol concentration of 0.200 - 0.205 gram/210L when used in a breath alcohol simulator heated to $34 \pm 0.2^{\circ}\text{C}$.

The solution was found to have an analytical concentration of 0.245 gram/dL (wt/vol) ethyl alcohol in aqueous solution with an equivalent vapour alcohol concentration of 0.202 g/210L when used in a breath alcohol simulator heated to $34 \pm 0.2^{\circ}\text{C}$.

The solution is manufactured from distilled/de-ionized water and absolute ethyl alcohol (USP Grade). This lot contains 2500 bottles (of 500 mL each) of Alcohol Reference Solution.

Date of Analysis: February 23, 2017


G.J. Kupferschmidt, B.Sc., M.Sc., M.C.I.C., C.Chem
G. Kupferschmidt Consulting Services Ltd.

