

Florida Department of Law Enforcement Alcohol Testing Program

CERTIFICATE OF ASSURANCE

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 201611B

Simulated Breath Alcohol Concentration (g/210L): 0.080 g/210L

Date of Approval: December 16th, 2016

Source: Alcohol Countermeasure Systems, Inc.

Date of Manufacture: November 15th, 2016

For Use Through: November 15th, 2018

Number of Bottles Manufactured: 2500

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.

Brett Kirkland

Signature

Program Manager

Title

Brett Kirkland

Name (Printed or Typed)

12/16/16

Date