



Florida Department of
Law Enforcement

Richard L. Swearingen
Commissioner

Criminal Justice Professionalism
Post Office Box 1489
Tallahassee, Florida 32302-1489
(850) 410-8600
www.fdle.state.fl.us

Rick Scott, *Governor*
Pam Bondi, *Attorney General*
Jeff Atwater, *Chief Financial Officer*
Adam Putnam, *Commissioner of Agriculture*

February 12, 2016

Mr. Christopher Wilson
Alcohol Countermeasure Systems, Inc.
60 International Boulevard
Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201601F

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201601F, as samples for approval on February 1, 2016. Gas chromatographic analyses of ten (10) bottles from this lot was completed on February 11, 2016. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

Lot 201601F is approved for use in Florida effective February 12, 2016.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1280.

Sincerely,

Patrick J. Murphy
Program Manager
Alcohol Testing Program



Florida Department of
Law Enforcement


Richard L. Swearingen
Commissioner

Criminal Justice Professionalism
Post Office Box 1489
Tallahassee, Florida 32302-1489
(850) 410-8600
www.fdle.state.fl.us

Rick Scott, *Governor*
Pam Bondi, *Attorney General*
Jeff Atwater, *Chief Financial Officer*
Adam Putnam, *Commissioner of Agriculture*

MEMORANDUM

TO: Department Inspectors and Agency Inspectors

FROM: Patrick Murphy, Program Manager 

DATE: February 12, 2016

SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 201601F

The following lot of alcohol reference solution has been approved for use in the State of Florida effective February 12, 2016:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
201601F	0.08 g/210L	1/26/2016	1/26/2018

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

If you have any questions concerning this approved lot, please contact me at (850) 617-1280.

PJM/pjm