



Florida Department of  
Law Enforcement

Richard L. Swearingen  
*Commissioner*

**Criminal Justice Professionalism**  
Post Office Box 1489  
Tallahassee, Florida 32302-1489  
(850) 410-8600  
[www.fdle.state.fl.us](http://www.fdle.state.fl.us)

Rick Scott, *Governor*  
Pam Bondi, *Attorney General*  
Jimmy Patronis, *Chief Financial Officer*  
Adam Putnam, *Commissioner of Agriculture*

October 5<sup>th</sup>, 2017

Mr. Christopher Wilson  
Alcohol Countermeasure Systems, Inc.  
60 International Boulevard  
Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201709A

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201709A, as samples for approval on October 3<sup>rd</sup>, 2017. Gas chromatographic analyses of ten (10) bottles from this lot was completed on October 4<sup>th</sup>, 2017. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

**Lot 201709A is approved for use in Florida effective October 5<sup>th</sup>, 2017.**

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

Brett Kirkland, Ph.D.  
Program Manager  
Alcohol Testing Program  
Florida Department of Law Enforcement

**MEMORANDUM**

TO: Department Inspectors and Agency Inspectors  
FROM: Brett Kirkland, Program Manager *BK*  
DATE: October 5<sup>th</sup>, 2017  
SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 201709A

---

The following lot of alcohol reference solution has been approved for use in the State of Florida effective October 5<sup>th</sup>, 2017:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
201709A	0.080 g/210L	9/12/2017	9/12/2019

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 2.14 of the ATP standard operating procedure, effective May 2017.

If you have any questions concerning this approved lot, please contact me at (850) 617-1275.

BHK/bhk