

March 7, 2018

Mr. Christopher Wilson  
Alcohol Countermeasure Systems, Inc.  
60 International Boulevard  
Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201802G

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201802G, as samples for approval on March 1<sup>st</sup>, 2018. Gas chromatographic analyses of ten (10) bottles from this lot was completed on March 5<sup>th</sup>, 2018. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

**Lot 201802G is approved for use in Florida effective March 7, 2018.**

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,



Brett Kirkland, Ph.D.  
Program Manager  
Alcohol Testing Program  
Florida Department of Law Enforcement

**MEMORANDUM**

TO: Department Inspectors and Agency Inspectors  
FROM: Brett Kirkland, Program Manager *BK*  
DATE: March 7<sup>th</sup>, 2018  
SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 201802G

---

The following lot of alcohol reference solution has been approved for use in the State of Florida effective March 7<sup>th</sup>, 2018:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
201802G	0.200 g/210L	2/22/2018	2/22/2020

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 4 of the ATP standard operating procedure, effective October 2017.

If you have any questions concerning this approved lot, please contact me at (850) 617-1290.

BHK