



ISO/IEC 17025:2005 Accredited Laboratory

Certificate of Analysis

Certificate ID: 8396
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 24515080A2
Expiration: 10/5/2017

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	208 ppm	+/- 0.002 or 2%	NDIR
Nitrogen	Balance	BAC whichever is greater	

RECEIVED

DEC 22 2015

FDLE
Alcohol Testing Program

*NIST Standard Reference Material
 Cylinder No. CC14290 / Job No. 09160202
 Certified 212.8 µmol/mol Ethanol in Nitrogen
 for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition
 and direct sunlight. Do not allow storage area to
 exceed 52 °C (125 °F).


 Specialty Gas Lab Tech

09/22/15
 Date



Distributed by:

CMI Inc.
 316 East Ninth Street
 Owensboro, KY 42303
 Phone 866-835-0690
www.alcoholtest.com