



RECEIVED

MAR 20 2017

FDLE  
Alcohol Testing Program

ISO/IEC 17025:2005 Accredited Laboratory

## Certificate of Analysis

**Certificate ID:** 9833  
**Part #:** BAC105L080T  
**Cylinder Size:** 105L  
**Lot Number:** 34416080A2  
**Expiration:** 2/5/2019


**0.080 BAC** (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	209 PPM	+/- 0.002 or 2%	NDIR
Nitrogen	BALANCE	BAC whichever is greater	

\*NIST Standard Reference Material  
Cylinder No. CCI03723 / Job No. 09160202  
Certified 212.8 µmol/mol Ethanol in Nitrogen  
for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

  
Specialty Gas Lab Tech

Date 12/29/16

Distributed by:

CMI Inc.  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

