

Florida Department of Law Enforcement

Alcohol Testing Program

CERTIFICATE OF ASSURANCE

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 202307C

Simulated Breath Alcohol Concentration (g/210L): 0.200

Date of Approval: August 18, 2023

Source: Alcohol Countermeasure Solutions

Date of Manufacture: July 12, 2023

For Use Through: July 12, 2025

Number of Bottles Manufactured: 2256

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.



Signature

Shayla Platt

Name (Printed or Typed)

Program Manager

Title

September 12, 2023

Date