

Criminal Justice Professionalism Post Office Box 1489 Tallahassee, Florida 32302-1489 (850) 410-8600 www.fdle.state.fl.us Ron DeSantis, *Governor* Ashley Moody, *Attorney General* Jimmy Patronis, *Chief Financial Officer* Nikki Fried, *Commissioner of Agriculture* 

July 6, 2022

Mr. Christopher Wilson Alcohol Countermeasure Systems, Inc. 60 International Boulevard Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 202206B

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 202206B, as samples for approval on June 22, 2022. Gas chromatographic analyses of ten (10) bottles from this lot was completed on July 1, 2022. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

## Lot 202206B is approved for use in Florida effective July 6, 2022.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

TJ Graham, D-ABFT-FA Program Manager Alcohol Testing Program Florida Department of Law Enforcement

## MEMORANDUM

TO:	Department Inspectors and Agency Inspectors	
FROM:	TJ Graham, Program Manager	
DATE:	July 6, 2022	
SUBJECT:	Approved Lot of Alcohol Reference Solution – Lot 202206B	

The following lot of alcohol reference solution has been approved for use in the State of Florida effective July 6, 2022:

Lot Number	Concentration	Manufacture Date	Expiration Date
202206B	0.080 g/210L	06/14/2022	06/14/2024

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 4 of the ATP technical standard operating procedure, effective February 2022.

If you have any questions concerning this approved lot, please contact me at (850) 617-1290.

Alsha