# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Agency** | SUMTER COUNTY SO | **Date** | 01/30/2016 | **Type** | Agency Inspection | | |
| **Inspector** | JACOBSON,JAMES,W | **Time** | 15:55:29 | **Software** | 8100.27 | | |
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28814080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/30/2016 15:56

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO **Date** 02/13/2016 **Type** Agency Inspection
**Inspector** FORINASH,GARY,B **Time** 19:31:26 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 28814080A3
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 11/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2016 19:28

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SUMTER COUNTY SO | **Date** | 03/07/2016 | **Type** | Department Inspection |
| **Inspector** | BELL,DANIELLE,M | **Time** | 17:14:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 09014080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 05/01/2016 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Test (Std Dev) | 0.0007 |
| 0.20 g/210L Test (Std Dev) | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 03/07/2016 17:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SUMTER COUNTY SO | **Date** | 03/28/2016 | **Type** | Agency Inspection |
| **Inspector** | JACOBSON,JAMES,W | **Time** | 13:13:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28814080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/28/2016 13:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO **Date** 04/24/2016 **Type** Agency Inspection
**Inspector** FORINASH,GARY,B **Time** 10:27:38 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 28814080A3
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 11/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/24/2016 18:36

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO **Date** 05/30/2016 **Type** Agency Inspection
**Inspector** FORINASH,GARY,B **Time** 12:34:55 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 28814080A3
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 11/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/30/2016 12:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO **Date** 06/28/2016 **Type** Agency Inspection
**Inspector** JACOBSON,JAMES,W **Time** 02:27:13 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 28814080A3
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 11/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2016 02:23

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SUMTER COUNTY SO | **Date** | 07/31/2016 | **Type** | Agency Inspection |
| **Inspector** | JACOBSON,JAMES,W | **Time** | 14:36:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 28814080A3 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/31/2016 14:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SUMTER COUNTY SO | **Date** | 08/28/2016 | **Type** | Agency Inspection |
| **Inspector** | FORINASH,GARY,B | **Time** | 21:14:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 28814080A3 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.075 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/28/2016 21:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO **Date** 09/25/2016 **Type** Agency Inspection
**Inspector** JACOBSON,JAMES,W **Time** 12:12:53 **Software** 8100.27

**0.05 Lot #** 201603D **0.08 Lot #** 201601F **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 03/08/2018 **Expiration** 01/26/2018 **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.073 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 1 | INT |
| Mouth Alcohol Test: Slope Not Met | Yes | Alcohol Free Test # 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Diagnostic Check (Post-Inspection): OK | No | 0.08 g/210L Test # 2 | 0.076 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| 0.20 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Dry Gas Std Test: OK | No | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.000 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:SIM NOT ATTACHED.

**Data Download Date/Time** 09/25/2016 12:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO **Date** 09/25/2016 **Type** Agency Inspection
**Inspector** JACOBSON,JAMES,W **Time** 12:49:42 **Software** 8100.27

**0.05 Lot #** 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 28814080A3
**Expiration** 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 11/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2016 12:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SUMTER COUNTY SO | **Date** | 10/22/2016 | **Type** | Agency Inspection |
| **Inspector** | FORINASH,GARY,B | **Time** | 11:04:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 20016080A3 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 09/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/22/2016 10:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SUMTER COUNTY SO | **Date** | 11/24/2016 | **Type** | Agency Inspection |
| **Inspector** | JACOBSON,JAMES,W | **Time** | 02:51:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 20016080A3 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 09/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/24/2016 02:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO **Date** 12/30/2016 **Type** Agency Inspection
**Inspector** FORINASH,GARY,B **Time** 08:03:12 **Software** 8100.27

**0.05 Lot #** 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A3
**Expiration** 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 09/05/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/30/2016 07:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 01/25/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 11:58:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/25/2016 12:00

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 02/16/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 14:06:12 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2016 14:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 03/18/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 10:35:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/18/2016 10:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 04/28/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 11:51:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2016 11:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 05/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 10:56:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/12/2016 10:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 06/29/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 11:55:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.075 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/29/2016 11:52

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 07/29/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 11:07:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.206 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.207 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.207 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/29/2016 11:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 08/22/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 09:38:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.083 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.206 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.083 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/22/2016 09:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO **Date** 08/26/2016 **Type** Department Inspection
**Inspector** MURPHY,PATRICK,J **Time** 12:37:15 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG612405
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 05/03/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.082 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.082 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.082 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 08/26/2016 12:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 09/27/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 15:10:24 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.053 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.084 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.207 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.084 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.208 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.084 |
| | | 0.20 g/210L Test # 3 | 0.207 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2016 15:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 10/19/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 21:38:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.053 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.208 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.207 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/19/2016 21:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 11/23/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 03:07:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/23/2016 03:03

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 21:21:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201611B | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 03/08/2018 | **Expiration** | 11/15/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.083 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2016 21:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 01/25/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 11:57:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/25/2016 12:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 02/16/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 14:02:52 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2016 14:08

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 03/18/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 10:32:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/18/2016 11:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 04/28/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 11:50:33 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2016 11:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 05/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 10:55:51 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/12/2016 11:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 06/29/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 10:43:24 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

STOPPED INSPECTION

**Data Download Date/Time** 06/29/2016 11:48

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 06/29/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 11:51:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/29/2016 11:48

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 07/29/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 11:12:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.208 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.209 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.209 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/29/2016 11:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 08/22/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 09:35:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.207 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/22/2016 09:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 09/27/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 15:07:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.208 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.209 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.209 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2016 15:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

**Agency** CITRUS COUNTY SO **Date** 10/13/2016 **Type** Department Inspection
**Inspector** MURPHY,PATRICK,J **Time** 13:42:39 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG619605
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 07/14/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.082 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 10/13/2016 13:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 10/19/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 21:35:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.207 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.207 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/19/2016 21:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 11/23/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 03:11:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.207 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/23/2016 03:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CITRUS COUNTY SO | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | ROBERSON,WILLIAM,E | **Time** | 21:18:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201611B | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 13115080A1 |
| **Expiration** | 03/08/2018 | **Expiration** | 11/15/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2016 21:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 01/28/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 22:53:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/28/2016 22:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 02/25/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 10:11:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/25/2016 10:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 03/08/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 15:41:16 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG605301
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 02/22/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0010 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 03/08/2016 15:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 03/31/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 15:04:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/31/2016 14:57

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 04/29/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 22:38:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/29/2016 22:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 05/31/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 16:03:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/31/2016 15:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 06/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 13:43:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

MOUTH ALC AT WRONG TIME.

**Data Download Date/Time** 06/27/2016 13:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 07/27/2016 **Type** Agency Inspection
**Inspector** CHESSHER,SANDRA,W **Time** 07:37:31 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201408B **0.08 Gas Lot #** 12716080A7
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 08/12/2016 **Expiration** 06/05/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/27/2016 07:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 08/29/2016 **Type** Agency Inspection
**Inspector** CHESSHER,SANDRA,W **Time** 10:40:12 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 12716080A7
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 06/05/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.182 |
| 0.20 g/210L Test # (Repeat) 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.190 |
| 0.20 g/210L Test # (Repeat) 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.20 g/210L Test # (Repeat) 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance/ SIM NOT PLUGGED IN.

**Data Download Date/Time** 08/29/2016 10:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 09/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 16:34:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 12716080A7 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2016 16:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 10/27/2016 | **Type** | Agency Inspection |
| **Inspector** | DESANTIS,JEFFERY,S | **Time** | 12:10:13 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 12716080A7 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/27/2016 12:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 11/30/2016 | **Type** | Agency Inspection |
| **Inspector** | DESANTIS,JEFFERY,S | **Time** | 11:32:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 12716080A7 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/30/2016 11:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY S.O. | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | BOWDEN,MATTHEW,C | **Time** | 10:54:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 12716080A7 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2016 11:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 01/28/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 22:55:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | AG402702 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 01/27/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:NOT ENOUGH SOULTION IN MOUTH.

**Data Download Date/Time** 01/28/2016 22:54

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 02/25/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 10:17:44 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/25/2016 10:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO **Date** 03/08/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 15:49:27 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG605301
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 02/22/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.084 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.084 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.083 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.083 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.084 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.083 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.083 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.083 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.083 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 03/08/2016 15:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 03/31/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 15:06:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/31/2016 15:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 04/30/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 20:11:38 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: RFI Detect. REMOVED CELL PHONE FROM AREA

**Data Download Date/Time** 04/30/2016 20:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 05/31/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 15:45:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/31/2016 15:42

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 06/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 13:26:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/27/2016 13:23

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 07/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 10:08:24 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/27/2016 10:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 08/29/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 10:20:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: TOO MUCH SOLUTION IN MOUTH. REPEATED.

**Data Download Date/Time** 08/29/2016 10:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 09/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 16:40:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2016 16:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO **Date** 10/27/2016 **Type** Agency Inspection
**Inspector** BOWDEN,MATTHEW,C **Time** 11:52:10 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 1766638
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 11/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance/HOSE DISCONNECTED.

**Data Download Date/Time** 10/27/2016 11:47

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 11/30/2016 | **Type** | Agency Inspection |
| **Inspector** | BOWDEN,MATTHEW,C | **Time** | 09:15:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/30/2016 09:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | BOWDEN,MATTHEW,C | **Time** | 11:06:35 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2016 11:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 01/27/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 13:57:32 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 03415080A1
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 03/05/2017

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/27/2016 14:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 02/16/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 03:52:09 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 03415080A1
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 03/05/2017

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2016 04:03

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 03/19/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 11:12:23 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 03415080A1
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 03/05/2017

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/19/2016 11:22

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 04/30/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 09:01:35 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 03415080A1
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 03/05/2017

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/30/2016 08:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 05/14/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 09:34:30 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 03415080A1
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 03/05/2017

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/14/2016 09:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LEESBURG P.D. | **Date** | 06/01/2016 | **Type** | Agency Inspection |
| **Inspector** | SOMMERSDORF,JOHN,G | **Time** | 23:58:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/01/2016 23:57

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 07/25/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 02:54:06 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 03415080A1
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 03/05/2017

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/25/2016 02:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LEESBURG P.D. | **Date** | 08/23/2016 | **Type** | Department Inspection |
| **Inspector** | BELL,DANIELLE,M | **Time** | 19:12:47 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG612405 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 05/03/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0017 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 08/23/2016 19:07

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 08/30/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 18:39:09 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 01316080A2
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 02/05/2018

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

POST DEPARTMENT INSPECTION. BACK IN SERVICE AT LEESBURG PD.

**Data Download Date/Time** 08/30/2016 18:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 09/27/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 08:39:50 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 01316080A2
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 02/05/2018

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2016 08:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LEESBURG P.D. | **Date** | 10/30/2016 | **Type** | Agency Inspection |
| **Inspector** | SOMMERSDORF,JOHN,G | **Time** | 11:36:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 01316080A2 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/05/2018 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/30/2016 11:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 11/14/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 02:24:09 **Software** 8100.27

**0.05 Lot #** 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 01316080A2
**Expiration** 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 02/05/2018

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/14/2016 03:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 12/15/2016 **Type** Agency Inspection
**Inspector** SOMMERSDORF,JOHN,G **Time** 04:21:50 **Software** 8100.27

**0.05 Lot #** 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 01316080A2
**Expiration** 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 02/05/2018

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.186 |
| 0.20 g/210L Test # (Repeat) 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.189 |
| 0.20 g/210L Test # (Repeat) 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.190 |
| 0.20 g/210L Test # (Repeat) 3 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 12/15/2016 04:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 01/25/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 10:07:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 04/08/2016 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/25/2016 10:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 02/08/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 11:02:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/08/2016 11:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 02/22/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 16:14:14 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG511701
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 04/27/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 02/22/2016 16:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 02/27/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 10:28:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/27/2016 10:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 03/30/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 11:59:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/30/2016 11:57

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 04/23/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 08:34:13 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TOO MUCH MOUTH ALCOHOL.

**Data Download Date/Time** 04/23/2016 08:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 05/31/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 04:33:33 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TO MUCH M/A

**Data Download Date/Time** 05/31/2016 04:32

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 06/27/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 21:39:44 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/27/2016 21:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 07/31/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 23:13:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # (Repeat) 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # (Repeat) 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

TOO MUCH M/A. ON DUTY RADIO TRIGGERED RFI.

**Data Download Date/Time** 07/31/2016 23:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 08/01/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 01:55:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 08/01/2016 01:52

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 09/30/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 10:08:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/30/2016 10:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 10/26/2016 | **Type** | Agency Inspection |
| **Inspector** | KEMENY,ANTHONY,M | **Time** | 04:30:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail. INTOXILYZER WAS CLEARED OF SURROUNDINGS

**Data Download Date/Time** 10/26/2016 04:24

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 11/23/2016 | **Type** | Agency Inspection |
| **Inspector** | KEMENY,ANTHONY,M | **Time** | 03:05:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 03/08/2018 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.067 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.078 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.078 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance FORGOT TO RECONNECT DGS.

**Data Download Date/Time** 11/23/2016 02:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MOUNT DORA P.D. | **Date** | 12/26/2016 | **Type** | Agency Inspection |
| **Inspector** | TAYLOR,KEITH,A | **Time** | 10:41:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1841081 |
| **Expiration** | 03/08/2018 | **Expiration** | 02/24/2017 | **Expiration** | 04/05/2018 | **Expiration** | 04/08/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/26/2016 10:41

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 01/06/2016 | **Type** | Agency Inspection |
| **Inspector** | JACOBS,JENNIFER,H | **Time** | 08:17:16 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

SKIPPED TO OPERATE INSTRUMENT

**Data Download Date/Time** 01/06/2016 08:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO **Date** 01/06/2016 **Type** Department Inspection
**Inspector** JACOBS,JENNIFER,H **Time** 11:04:15 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG511701
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 04/27/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.052 |
| 0.08 g/210L Test # 1 | 0.082 |
| 0.20 g/210L Test # 1 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.052 |
| 0.08 g/210L Test # 2 | 0.083 |
| 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.052 |
| 0.08 g/210L Test # 4 | 0.082 |
| 0.20 g/210L Test # 4 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.053 |
| 0.08 g/210L Test # 5 | 0.083 |
| 0.20 g/210L Test # 5 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.082 |
| 0.20 g/210L Test # 6 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.053 |
| 0.08 g/210L Test # 7 | 0.083 |
| 0.20 g/210L Test # 7 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.053 |
| 0.08 g/210L Test # 8 | 0.083 |
| 0.20 g/210L Test # 8 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.053 |
| 0.08 g/210L Test # 9 | 0.083 |
| 0.20 g/210L Test # 9 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.052 |
| 0.08 g/210L Test # 10 | 0.083 |
| 0.20 g/210L Test # 10 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

A F / M A:SAMPLE NOT LONG ENOUGH

**Data Download Date/Time** 01/06/2016 11:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 01/21/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 23:56:58 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1669440 |
| **Expiration** | 03/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/19/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 01/21/2016 23:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 02/17/2016 | **Type** | Agency Inspection |
| **Inspector** | MOORE,MICAH, | **Time** | 17:01:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1669940 |
| **Expiration** | 03/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/19/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 02/17/2016 16:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 03/24/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 22:37:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1669440 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/19/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 03/24/2016 22:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO **Date** 04/28/2016 **Type** Agency Inspection
**Inspector** RAKER,JEFFERY,A **Time** 10:39:52 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 83550
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 09/26/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.053 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 04/28/2016 10:32

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 05/27/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 12:17:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1669440 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/19/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 05/27/2016 12:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 06/23/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 13:17:47 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1669440 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/19/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 06/23/2016 13:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 07/21/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 12:37:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1669440 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/19/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 07/21/2016 12:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 08/26/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 12:34:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 14968 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/06/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 08/26/2016 12:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 09/26/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 16:56:51 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 14968 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/06/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 09/26/2016 16:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 10/25/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 12:11:04 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 14968 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/06/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 10/25/2016 12:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 11/22/2016 | **Type** | Agency Inspection |
| **Inspector** | RAKER,JEFFERY,A | **Time** | 12:48:49 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | INT |
| Date and/or Time Adjusted | No | Alcohol Free Test # (Repeat) 1 | INT |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | Yes | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Interferent DetectAmbient Fail. 00: Interferent Detect, Interferent Detect. Non-compliance: .

**Data Download Date/Time** 11/22/2016 12:41

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO **Date** 01/04/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 19:06:48 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 03415080A2
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 03/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C KIRK PRESENT ENTIRE INSPECTION

**Data Download Date/Time** 01/04/2016 19:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO **Date** 02/01/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 20:55:44 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 03415080A2
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 03/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI C KIRK PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 02/01/2016 21:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO **Date** 03/03/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 20:13:04 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 18015080A2
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 08/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK 35602 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 03/03/2016 20:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO **Date** 03/15/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 13:02:57 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG605301
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 02/22/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.081 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 03/15/2016 12:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 04/06/2016 | **Type** | Agency Inspection |
| **Inspector** | SKIPPER,ROGER,G | **Time** | 17:22:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK 35602 PRESENT ENTIRE INSPECTION

**Data Download Date/Time** 04/06/2016 17:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO **Date** 05/11/2016 **Type** Agency Inspection
**Inspector** BELL,TIMOTHY,R **Time** 11:50:12 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 18015080A2
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 08/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI SUPV. R SKIPPER PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 05/11/2016 11:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 06/07/2016 | **Type** | Agency Inspection |
| **Inspector** | BELL,TIMOTHY,R | **Time** | 13:38:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI SUPV. R. SKIPPER #246109 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 06/07/2016 13:36

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 07/12/2016 | **Type** | Agency Inspection |
| **Inspector** | SKIPPER,ROGER,G | **Time** | 18:36:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK #35602 PRESENT ENTIRE INSPECTION

**Data Download Date/Time** 07/13/2016 14:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 08/10/2016 | **Type** | Agency Inspection |
| **Inspector** | KIRK,CASEY,S | **Time** | 20:14:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI R. SKIPPER #246109 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 08/10/2016 20:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 09/07/2016 | **Type** | Agency Inspection |
| **Inspector** | SKIPPER,ROGER,G | **Time** | 18:49:44 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK #35602 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 09/07/2016 18:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 10/05/2016 | **Type** | Agency Inspection |
| **Inspector** | SKIPPER,ROGER,G | **Time** | 17:38:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK 3

**Data Download Date/Time** 10/05/2016 17:39

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO **Date** 11/07/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 20:44:59 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 18015080A2
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 08/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK #35602 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 11/07/2016 20:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 12/14/2016 | **Type** | Agency Inspection |
| **Inspector** | SKIPPER,ROGER,G | **Time** | 19:19:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK #35602 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 12/14/2016 19:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 02/01/2016 | **Type** | Agency Inspection |
| **Inspector** | SHANAHAN,JAKE,L | **Time** | 09:06:28 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASS AI TO OPERATE INSTRUMENTNon-compliance:NOT DETERMINED.

**Data Download Date/Time** 02/01/2016 11:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO **Date** 02/01/2016 **Type** Department Inspection
**Inspector** SHANAHAN,JAKE,L **Time** 11:36:59 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG511701
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 04/27/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.076 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.076 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.076 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

**Remarks:**

A F / M A:NOT ENOUGH MA.

**Data Download Date/Time** 02/01/2016 11:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 02/17/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 03:47:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/17/2016 03:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 03/24/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 20:45:45 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/24/2016 20:43

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 04/28/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 13:47:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2016 13:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 05/18/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 13:05:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.206 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.207 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.208 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/18/2016 13:00

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 06/26/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 09:08:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2016 00:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 07/25/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 16:45:06 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 04/05/2018 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/25/2016 16:42

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 08/17/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 13:37:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/17/2016 13:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 09/07/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 12:30:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/07/2016 12:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 10/13/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 18:22:53 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/13/2016 18:19

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 11/16/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 13:34:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.190 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.191 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/16/2016 13:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 15:02:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2016 14:57

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 01/27/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 13:00:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 04/08/2016 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/27/2016 12:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO **Date** 02/07/2016 **Type** Department Inspection
**Inspector** SHANAHAN,JAKE,L **Time** 12:19:38 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG511701
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 04/27/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.052 |
| 0.08 g/210L Test # 1 | 0.082 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.052 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.052 |
| 0.08 g/210L Test # 4 | 0.082 |
| 0.20 g/210L Test # 4 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.052 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.082 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.052 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.052 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.052 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 02/07/2016 12:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 02/17/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 05:00:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 31715080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 12/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/17/2016 04:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 03/24/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 19:59:26 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

TO LITTLE MOUTH ALCOHOL

**Data Download Date/Time** 03/24/2016 21:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 03/24/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 21:00:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 31715080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 12/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/24/2016 21:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 04/28/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 15:15:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 24515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2016 15:13

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 05/18/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 13:15:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06616080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.210 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.210 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.210 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/18/2016 13:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 06/26/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 09:20:17 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06616080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2016 00:56

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 07/25/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 16:52:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 06616080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 04/05/2018 | **Expiration** | 04/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/25/2016 16:48

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 08/17/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 13:17:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 06616080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 04/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/17/2016 13:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 09/07/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 12:42:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 20016080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 09/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/07/2016 12:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 10/13/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 18:51:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 20016080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 09/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/13/2016 18:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 11/16/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 13:42:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 20016080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 09/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/16/2016 13:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | BAKER,BRADFORD,B | **Time** | 14:57:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 20016080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 09/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2016 14:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO **Date** 01/04/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 20:42:41 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 12014080A1
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 06/01/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C KIRK PRESENT ENTIRE INSPECTION

**Data Download Date/Time** 01/04/2016 20:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO **Date** 02/01/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 21:09:54 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 12014080A1
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 06/01/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI C KIRK PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 02/01/2016 21:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO **Date** 02/23/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 12:50:31 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG511701
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 04/27/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.052 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.082 |
| 0.20 g/210L Test # 6 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.052 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.083 |
| 0.20 g/210L Test # 8 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.052 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.053 |
| 0.08 g/210L Test # 10 | 0.083 |
| 0.20 g/210L Test # 10 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Test (Std Dev) | 0.0009 |
| 0.20 g/210L Test (Std Dev) | 0.0010 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 02/23/2016 12:47

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO **Date** 03/03/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 18:38:23 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 12014080A1
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 06/01/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK 35602 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 03/03/2016 18:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 07/06/2016 | **Type** | Agency Inspection |
| **Inspector** | BELL,DANIELLE,M | **Time** | 15:50:33 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

**Data Download Date/Time** 07/14/2016 10:53

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO **Date** 07/14/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 14:04:03 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG612405
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 05/03/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.081 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0009 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 07/14/2016 13:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO **Date** 08/10/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 18:20:19 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 18015080A2
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 08/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK #35602 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 08/10/2016 18:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO **Date** 09/07/2016 **Type** Agency Inspection
**Inspector** SKIPPER,ROGER,G **Time** 20:19:41 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 18015080A2
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 08/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK #35602 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 09/07/2016 20:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 10/05/2016 | **Type** | Agency Inspection |
| **Inspector** | SKIPPER,ROGER,G | **Time** | 17:53:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI C. KIRK #35602 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 10/05/2016 17:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO **Date** 10/27/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 13:59:32 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG619605
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 07/14/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 10/27/2016 15:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 11/07/2016 | **Type** | Agency Inspection |
| **Inspector** | KIRK,CASEY,S | **Time** | 19:18:41 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI R. SKIPPER #246109 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 11/07/2016 19:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 12/14/2016 | **Type** | Agency Inspection |
| **Inspector** | KIRK,CASEY,S | **Time** | 19:29:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 18015080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI R. SKIPPER #246109 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 12/14/2016 19:25

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 01/28/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 22:00:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 12014080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/01/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/28/2016 21:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 02/23/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 12:18:13 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG511701
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 04/27/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 02/23/2016 12:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 02/25/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 10:23:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 12014080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/01/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/25/2016 10:20

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 03/31/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 22:29:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 12014080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/01/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/31/2016 22:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 04/29/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 22:27:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 12014080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/01/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/29/2016 22:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 05/31/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 15:55:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 12014080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 06/01/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/31/2016 15:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 06/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 13:37:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/27/2016 13:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 07/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 07:30:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:NOT ENOUGH SOLUTION IN MOUTH.

**Data Download Date/Time** 07/27/2016 07:25

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 08/29/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 10:15:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/29/2016 10:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 09/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CHESSHER,SANDRA,W | **Time** | 16:18:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2016 16:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 10/27/2016 | **Type** | Agency Inspection |
| **Inspector** | BOWDEN,MATTHEW,C | **Time** | 11:29:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1766638 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/27/2016 11:39

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 11/30/2016 **Type** Agency Inspection
**Inspector** BOWDEN,MATTHEW,C **Time** 08:56:27 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 12716080A7
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 06/05/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.TIMED OUT

**Data Download Date/Time** 11/30/2016 09:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | UMATILLA PD(LAKE SO) | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | BOWDEN,MATTHEW,C | **Time** | 11:01:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 12716080A7 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2016 10:56

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PANAMA CITY BEACH PD | **Date** | 01/27/2016 | **Type** | Agency Inspection |
| **Inspector** | GAILFOIL,MICHAEL,L | **Time** | 09:44:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1752856 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/13/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/27/2016 10:41

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD **Date** 02/19/2016 **Type** Agency Inspection
**Inspector** GAILFOIL,MICHAEL,L **Time** 09:25:03 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 1752856
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 10/13/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/19/2016 10:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD **Date** 03/28/2016 **Type** Agency Inspection
**Inspector** GAILFOIL,MICHAEL,L **Time** 09:31:03 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 1752856
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 10/13/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/28/2016 10:28

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PANAMA CITY BEACH PD | **Date** | 04/28/2016 | **Type** | Agency Inspection |
| **Inspector** | GAILFOIL,MICHAEL,L | **Time** | 09:16:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1752856 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/13/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2016 10:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PANAMA CITY BEACH PD | **Date** | 05/31/2016 | **Type** | Agency Inspection |
| **Inspector** | GAILFOIL,MICHAEL,L | **Time** | 09:46:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1752856 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/13/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/31/2016 11:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PANAMA CITY BEACH PD | **Date** | 06/17/2016 | **Type** | Agency Inspection |
| **Inspector** | GAILFOIL,MICHAEL,L | **Time** | 09:30:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1752856 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 10/13/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/17/2016 10:24

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD **Date** 07/26/2016 **Type** Agency Inspection
**Inspector** GAILFOIL,MICHAEL,L **Time** 09:50:35 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 1752856
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 10/13/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TAKEN OUT OF SERVICE FOR YEARLY FDLE INSPECTION

**Data Download Date/Time** 07/26/2016 10:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD **Date** 08/02/2016 **Type** Department Inspection
**Inspector** SUBER,PHILLIP,W **Time** 13:09:25 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG612405
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 05/03/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 08/02/2016 14:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PANAMA CITY BEACH PD | **Date** | 08/19/2016 | **Type** | Agency Inspection |
| **Inspector** | GAILFOIL,MICHAEL,L | **Time** | 09:51:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1752856 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 10/13/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

BACK IN SERVICE AFTER YEARLY FDLE INSPECTION

**Data Download Date/Time** 08/19/2016 10:47

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD **Date** 09/14/2016 **Type** Agency Inspection
**Inspector** GAILFOIL,MICHAEL,L **Time** 08:37:06 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 1861309
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 05/19/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/14/2016 09:32

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PANAMA CITY BEACH PD | **Date** | 10/26/2016 | **Type** | Agency Inspection |
| **Inspector** | GAILFOIL,MICHAEL,L | **Time** | 09:50:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1861309 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 05/19/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .CHANGED MA SOLUTION

**Data Download Date/Time** 10/26/2016 10:43

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD **Date** 11/29/2016 **Type** Agency Inspection
**Inspector** GAILFOIL,MICHAEL,L **Time** 09:36:38 **Software** 8100.27

**0.05 Lot #** 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 1861309
**Expiration** 03/08/2018 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 05/19/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI Detect.

**Data Download Date/Time** 11/29/2016 10:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PANAMA CITY BEACH PD | **Date** | 12/21/2016 | **Type** | Agency Inspection |
| **Inspector** | MOORE,JACOB,A | **Time** | 06:15:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1861309 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 05/19/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 12/21/2016 07:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

**Agency** LYNN HAVEN PD **Date** 01/29/2016 **Type** Agency Inspection
**Inspector** HAMILTON,PATRICK,T **Time** 02:01:56 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 29614080A2
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 12/05/2016

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # (Repeat) 1 | 0.082 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # (Repeat) 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # (Repeat) 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI Detect.

**Data Download Date/Time** 01/29/2016 02:57

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 02/29/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 16:58:55 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 3 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Range ExceededSequence Aborted. Non-compliance: .

**Data Download Date/Time** 02/29/2016 18:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 02/29/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 17:35:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 29614080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 12/05/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.053 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/29/2016 18:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

**Agency** LYNN HAVEN PD **Date** 03/29/2016 **Type** Agency Inspection
**Inspector** HAMILTON,PATRICK,T **Time** 14:10:36 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 29614080A2
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 12/05/2016

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.192 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/29/2016 15:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

**Agency** LYNN HAVEN PD **Date** 04/30/2016 **Type** Agency Inspection
**Inspector** HAMILTON,PATRICK,T **Time** 05:08:22 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 29614080A2
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 12/05/2016

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/30/2016 06:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 06/09/2016 | **Type** | Agency Inspection |
| **Inspector** | SUBER,PHILLIP,W | **Time** | 09:29:25 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

SKIPPED AI TO OPERATE INSTRUMENT; NOT A COMPLIANCE CHECK

**Data Download Date/Time** 06/09/2016 14:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

**Agency** LYNN HAVEN PD **Date** 06/09/2016 **Type** Department Inspection
**Inspector** SUBER,PHILLIP,W **Time** 13:55:36 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG600504
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 01/05/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.082 |
| 0.20 g/210L Test # 6 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 06/09/2016 14:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 06/25/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 10:04:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 29614080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 12/05/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.053 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.053 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range ExceededRange Exceeded.

**Data Download Date/Time** 06/25/2016 10:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 07/27/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 17:43:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 29614080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 04/05/2018 | **Expiration** | 12/05/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.189 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.200 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.189 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.190 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded. 20: Control Outside Tolerance.

**Data Download Date/Time** 07/27/2016 18:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 08/29/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 23:03:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 29614080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 04/05/2018 | **Expiration** | 12/05/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.189 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.193 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.190 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.190 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.191 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded. 20: Control Outside Tolerance.

**Data Download Date/Time** 08/29/2016 23:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 09/30/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 22:10:39 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 09216080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 05/05/2018 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.190 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.200 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.189 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded. 20: Control Outside Tolerance.

**Data Download Date/Time** 09/30/2016 23:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 10/29/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 08:26:53 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 09216080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 05/05/2018 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.192 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/29/2016 09:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 11/25/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 15:09:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | |

**Number of Simulators** 3 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.186 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.187 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.188 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.193 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.188 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.195 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 11/25/2016 16:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LYNN HAVEN PD | **Date** | 11/25/2016 | **Type** | Agency Inspection |
| **Inspector** | HAMILTON,PATRICK,T | **Time** | 16:11:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 09216080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 05/05/2018 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.083 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.176 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.193 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.083 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.192 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 11/25/2016 17:03

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 01/25/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 10:09:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Purge Fail TOO MUCH MA.

**Data Download Date/Time** 01/25/2016 10:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 02/17/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 09:03:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/17/2016 08:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 03/09/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 09:21:17 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/09/2016 09:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 04/13/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 08:57:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/13/2016 08:52

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 05/17/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 10:54:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 034315080A |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/17/2016 10:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 06/08/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 09:56:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

05: RFI Detect RADIO OUTSIDE.

**Data Download Date/Time** 06/08/2016 09:53

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 07/06/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 10:19:17 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/06/2016 10:13

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 08/12/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 10:09:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201501F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.050 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

05: RFI Detect RADIO ON

**Data Download Date/Time** 08/12/2016 10:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 09/16/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 09:23:45 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/28/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/16/2016 09:19

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 10/11/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 09:55:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 03415080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/11/2016 09:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 10/20/2016 | **Type** | Department Inspection |
| **Inspector** | PLATT,SHAYLA,D | **Time** | 11:06:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG619605 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 07/14/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 10/20/2016 10:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 11/11/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 09:31:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 24515080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 10/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/11/2016 09:24

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CAPE CORAL PD | **Date** | 12/06/2016 | **Type** | Agency Inspection |
| **Inspector** | CLAFLIN,DAVID,S | **Time** | 11:33:24 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 07515080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/06/2016 11:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SARASOTA COUNTY SO | **Date** | 01/27/2016 | **Type** | Agency Inspection |
| **Inspector** | FELTOVIC,DEP,M | **Time** | 08:52:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28814080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 01/27/2016 09:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SARASOTA COUNTY SO | **Date** | 02/23/2016 | **Type** | Agency Inspection |
| **Inspector** | FELTOVIC,DEP,M | **Time** | 09:40:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28814080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 02/23/2016 09:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SARASOTA COUNTY SO | **Date** | 03/25/2016 | **Type** | Agency Inspection |
| **Inspector** | FELTOVIC,DEP,M | **Time** | 10:30:31 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28814080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # 1 | 0.048 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TRANSPORT VAN RFI.

**Data Download Date/Time** 03/25/2016 11:07

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SARASOTA COUNTY SO | **Date** | 04/28/2016 | **Type** | Agency Inspection |
| **Inspector** | FELTOVIC,DEP,M | **Time** | 10:21:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28814080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 04/28/2016 10:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SARASOTA COUNTY SO | **Date** | 05/26/2016 | **Type** | Agency Inspection |
| **Inspector** | FELTOVIC,DEP,M | **Time** | 08:00:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 05/26/2016 07:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SARASOTA COUNTY SO | **Date** | 06/28/2016 | **Type** | Agency Inspection |
| **Inspector** | FELTOVIC,DEP,M | **Time** | 10:44:41 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 02/24/2017 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 06/28/2016 10:41

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SARASOTA COUNTY SO | **Date** | 07/29/2016 | **Type** | Agency Inspection |
| **Inspector** | FELTOVIC,DEP,M | **Time** | 10:59:04 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28515080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 07/29/2016 10:54

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SARASOTA COUNTY SO | **Date** | 08/30/2016 | **Type** | Agency Inspection |
| **Inspector** | FELTOVIC,DEP,M | **Time** | 08:36:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28515080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 08/30/2016 08:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO **Date** 09/12/2016 **Type** Department Inspection
**Inspector** SUBER,PHILLIP,W **Time** 15:34:37 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG612405
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 05/03/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0011 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

**Remarks:**

**Data Download Date/Time** 09/12/2016 15:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO **Date** 11/30/2016 **Type** Department Inspection
**Inspector** PLATT,SHAYLA,D **Time** 12:27:40 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG626605
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 09/22/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0007 |
| 0.20 g/210L Test (Std Dev) | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 11/30/2016 12:22

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 01/28/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 23:26:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | AG425202 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 09/09/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/29/2016 00:23

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 02/29/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 21:32:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG425202 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/09/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.120 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.125 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.126 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 02/29/2016 22:28

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 03/27/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 17:42:39 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 3 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Interferent Detect Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 3 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

Int Det: . Non-compliance: .

**Data Download Date/Time** 03/27/2016 18:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 03/27/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 17:48:49 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 3 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Range Exceeded. Non-compliance: .

**Data Download Date/Time** 03/27/2016 18:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD **Date** 03/28/2016 **Type** Agency Inspection
**Inspector** HOAGLAND,JAMES,J **Time** 09:39:31 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG425202
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 09/09/2016

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/28/2016 10:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 04/30/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 01:45:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG425202 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/09/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.159 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:DIDN'T MEET REQUIREMENT. 20: Control Outside Tolerance.

**Data Download Date/Time** 04/30/2016 02:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 05/30/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 20:15:12 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG425202 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/09/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.188 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.193 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.191 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.192 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: RFI DetectRange Exceeded. 20: Control Outside Tolerance.

**Data Download Date/Time** 05/30/2016 21:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD **Date** 06/30/2016 **Type** Agency Inspection
**Inspector** HOAGLAND,JAMES,J **Time** 14:34:49 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # (Repeat) 3 | 0.000 |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | Yes | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

00: Ambient Fail. Non-compliance: .

**Data Download Date/Time** 06/30/2016 16:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 06/30/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 15:21:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG425202 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/09/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/30/2016 16:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD **Date** 08/01/2016 **Type** Agency Inspection
**Inspector** HOAGLAND,JAMES,J **Time** 23:57:48 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG425202
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 09/09/2016

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.066 |
| 0.08 g/210L Test # (Repeat) 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.187 |
| 0.20 g/210L Test # (Repeat) 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.066 |
| 0.08 g/210L Test # (Repeat) 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.191 |
| 0.20 g/210L Test # (Repeat) 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.068 |
| 0.08 g/210L Test # (Repeat) 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.20 g/210L Test # (Repeat) 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance. 20: Control Outside Tolerance.

**Data Download Date/Time** 08/02/2016 00:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 09/30/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 07:40:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG425202 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 09/09/2016 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/30/2016 08:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 11/11/2016 | **Type** | Agency Inspection |
| **Inspector** | HOAGLAND,JAMES,J | **Time** | 12:11:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/02/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/11/2016 13:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BLOUNTSTOWN PD | **Date** | 12/06/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 08:56:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/02/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.191 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.191 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 12/06/2016 09:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD **Date** 12/08/2016 **Type** Department Inspection
**Inspector** MURPHY,PATRICK,J **Time** 12:16:42 **Software** 8100.27

**0.05 Lot #** 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG626605
**Expiration** 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 09/22/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.082 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.082 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.082 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 12/08/2016 13:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 01/09/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 14:55:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1737087 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/12/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK

**Data Download Date/Time** 01/13/2016 10:20

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 02/13/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 12:17:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1737087 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/12/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK

**Data Download Date/Time** 02/13/2016 12:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 03/10/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:29:01 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK

**Data Download Date/Time** 03/10/2016 08:25

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 04/28/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 09:24:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK

**Data Download Date/Time** 04/28/2016 09:19

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 05/13/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 06:23:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK

**Data Download Date/Time** 05/13/2016 06:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 06/21/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 23:47:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK

**Data Download Date/Time** 06/21/2016 23:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 07/05/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:02:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK

**Data Download Date/Time** 07/05/2016 07:57

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 08/17/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 23:35:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK-GOING OUT FOR DEPT INSP

**Data Download Date/Time** 08/17/2016 23:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO **Date** 08/22/2016 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 12:10:24 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG600504
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 01/05/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 08/22/2016 12:07

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 09/16/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 05:16:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | No | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time** 09/16/2016 05:13

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 10/14/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:37:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # (Repeat) 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Interferent Detect Test # 2 | 0.000 |
| | | Interferent Detect Test # (Repeat) 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det: .

**Data Download Date/Time** 10/14/2016 08:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 11/13/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 09:20:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 02/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK MAIN

**Data Download Date/Time** 11/13/2016 09:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 06:44:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 02/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK MAIN

**Data Download Date/Time** 12/28/2016 06:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 01/07/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 13:27:55 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV

**Data Download Date/Time** 01/07/2016 13:23

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 02/11/2016 | **Type** | Agency Inspection | | |
| **Inspector** | HUFF,JON,B | **Time** | 14:12:23 | **Software** | 8100.27 | | |
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV

**Data Download Date/Time** 02/11/2016 14:09

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 03/09/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 10:57:31 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV

**Data Download Date/Time** 03/09/2016 10:53

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 04/27/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 01:30:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2015 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV

**Data Download Date/Time** 04/27/2016 01:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 05/05/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 09:54:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV

**Data Download Date/Time** 05/05/2016 09:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 06/21/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 11:42:01 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV

**Data Download Date/Time** 06/21/2016 11:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 07/04/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 07:24:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.084 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.085 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV

**Data Download Date/Time** 07/04/2016 07:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 08/14/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 07:35:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV

**Data Download Date/Time** 08/14/2016 07:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO **Date** 08/22/2016 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 12:25:50 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG600504
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 01/05/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.082 |
| 0.20 g/210L Test # 4 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

**Remarks:**

**Data Download Date/Time** 08/22/2016 12:22

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 09/15/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 02:36:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV-IN SERVICE

**Data Download Date/Time** 09/15/2016 02:32

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 10/24/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 07:21:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV MAIN

**Data Download Date/Time** 10/24/2016 07:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 11/16/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 07:43:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 02/14/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 11/16/2016 07:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 03:39:31 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 02/14/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.085 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.085 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.085 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV MAIN

**Data Download Date/Time** 12/28/2016 03:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 01/07/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 14:42:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/07/2016 14:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 02/11/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 14:51:31 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.053 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE

**Data Download Date/Time** 02/11/2016 14:47

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 03/09/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 11:35:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE

**Data Download Date/Time** 03/09/2016 11:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 04/27/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 02:41:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2015 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE

**Data Download Date/Time** 04/27/2016 02:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 05/05/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 10:30:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE

**Data Download Date/Time** 05/05/2016 10:25

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 06/21/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 12:55:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE

**Data Download Date/Time** 06/21/2016 12:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO **Date** 07/04/2016 **Type** Agency Inspection
**Inspector** HUFF,JON,B **Time** 08:07:02 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 279870
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 02/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 07/04/2016 08:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 08/14/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:14:55 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE-IN SERVICE KV

**Data Download Date/Time** 08/14/2016 08:09

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 09/15/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 01:59:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV SPARE-GOING OUT FOR ANNUAL DEPT INSP

**Data Download Date/Time** 09/15/2016 01:53

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 09/19/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 10:08:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG600504 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.053 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.053 |
| | | 0.08 g/210L Test # 4 | 0.082 |
| | | 0.20 g/210L Test # 4 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.053 |
| | | 0.08 g/210L Test # 5 | 0.082 |
| | | 0.20 g/210L Test # 5 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.082 |
| 0.20 g/210L Test # 6 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.053 |
| 0.08 g/210L Test # 7 | 0.083 |
| 0.20 g/210L Test # 7 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.053 |
| 0.08 g/210L Test # 8 | 0.083 |
| 0.20 g/210L Test # 8 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.053 |
| 0.08 g/210L Test # 9 | 0.083 |
| 0.20 g/210L Test # 9 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.053 |
| 0.08 g/210L Test # 10 | 0.084 |
| 0.20 g/210L Test # 10 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0008 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 09/19/2016 10:03

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 10/24/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:38:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/14/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV SPARE

**Data Download Date/Time** 10/24/2016 08:32

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 11/16/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:28:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 02/14/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV SPARE

**Data Download Date/Time** 11/16/2016 08:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 01/06/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 06:54:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1816818 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/13/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 01/06/2016 06:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 02/10/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 14:07:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 091602 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 02/10/2016 14:09

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 03/09/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 07:44:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 091602 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.206 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.207 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 03/09/2016 07:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 04/26/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:22:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 091602 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.206 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 04/26/2016 08:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 05/04/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:30:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 091602 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.207 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.207 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 05/04/2016 08:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 06/21/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:31:38 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 091602 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 06/21/2016 08:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 07/03/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 07:54:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 091602 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 07/03/2016 07:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 08/15/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:17:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 091602 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW-GOING OUT FOR DEPT INSPECTION

**Data Download Date/Time** 08/15/2016 08:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 08/22/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 12:41:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG600504 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.206 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.207 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.208 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.207 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.207 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 08/22/2016 12:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 09/15/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 03:15:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 279870 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 02/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 09/15/2016 03:09

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 10/23/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:14:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 14968 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 08/06/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW

**Data Download Date/Time** 10/23/2016 08:08

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 11/11/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 08:14:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 14968 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/06/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW MAIN

**Data Download Date/Time** 11/11/2016 08:07

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 12/28/2016 | **Type** | Agency Inspection |
| **Inspector** | HUFF,JON,B | **Time** | 01:19:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 14968 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/06/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW MAIN

**Data Download Date/Time** 12/28/2016 01:13

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 01/29/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 13:00:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 12014080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/01/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/29/2016 12:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 02/26/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 13:58:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 12014080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/01/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/26/2016 13:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 03/29/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 11:54:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 12014080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/01/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/29/2016 11:54

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

**Agency** CORAL GABLES PD **Date** 04/29/2016 **Type** Agency Inspection
**Inspector** SMITH,DANIEL,A **Time** 11:40:13 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG515306
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 06/02/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/29/2016 11:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 05/31/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 13:38:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/02/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/31/2016 13:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 06/29/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 14:05:17 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/02/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/29/2016 14:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

**Agency** CORAL GABLES PD **Date** 07/29/2016 **Type** Agency Inspection
**Inspector** SMITH,DANIEL,A **Time** 14:48:10 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 07/14/2017 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.042 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.041 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 1 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.05 g/210L Test # 2 | 0.041 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.041 |
| Diagnostic Check (Post-Inspection): OK | No | Interferent Detect Test # 2 | INT |
| Alcohol Free Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.041 |
| 0.08 g/210L Test: OK | No | 0.05 g/210L Test # (Repeat) 3 | 0.041 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 3 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 07/29/2016 14:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 08/09/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 15:33:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG600504 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2016 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 08/09/2016 15:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 08/12/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 14:43:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/02/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.000 |
| 0.05 g/210L Test # (Repeat) 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.05 g/210L Test # (Repeat) 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance. 08: Control Outside Tolerance.

**Data Download Date/Time** 08/12/2016 14:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 09/23/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 12:57:33 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/02/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/23/2016 12:53

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 10/28/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 15:34:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 06/02/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.045 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.045 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 10/28/2016 15:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 11/28/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 14:03:13 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 4 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # (Repeat) 3 | 0.000 |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | Yes | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

00: Ambient Fail. Non-compliance: .

**Data Download Date/Time** 11/28/2016 14:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 11/28/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 14:38:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/02/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 11/28/2016 14:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CORAL GABLES PD | **Date** | 12/01/2016 | **Type** | Agency Inspection |
| **Inspector** | SMITH,DANIEL,A | **Time** | 13:33:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG515306 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 06/02/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/01/2016 13:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 01/28/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 15:06:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201403F | **0.20 Lot #** | 201403F | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 03/25/2016 | **Expiration** | 03/25/2016 | **Expiration** | 03/25/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.208 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.208 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.207 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC LOUREIRO AND OFC MESA PRESENT

**Data Download Date/Time** 01/28/2016 15:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 02/25/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 17:56:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 03/25/2016 | **Expiration** | 04/08/2016 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/25/2016 18:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 03/10/2016 | **Type** | Agency Inspection |
| **Inspector** | REYES,BARBARA,T | **Time** | 16:41:51 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 03/25/2016 | **Expiration** | 04/08/2016 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/10/2016 16:48

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 04/22/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 17:48:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/22/2016 18:03

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 05/12/2016 | **Type** | Agency Inspection |
| **Inspector** | MATTSON,CHRIS,J | **Time** | 19:15:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/12/2016 20:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 06/23/2016 | **Type** | Agency Inspection |
| **Inspector** | MATTSON,CHRIS,J | **Time** | 11:45:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range ExceededNo Sample Provided.

**Data Download Date/Time** 06/23/2016 13:39

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 07/22/2016 | **Type** | Agency Inspection |
| **Inspector** | MATTSON,CHRIS,J | **Time** | 17:51:04 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/22/2016 21:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 08/22/2016 | **Type** | Agency Inspection |
| **Inspector** | MATTSON,CHRIS,J | **Time** | 20:05:17 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/22/2016 20:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 09/22/2016 | **Type** | Agency Inspection |
| **Inspector** | MATTSON,CHRIS,J | **Time** | 09:51:47 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/22/2016 11:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 09/29/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 12:05:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG600504 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.082 |
| | | 0.20 g/210L Test # 4 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.082 |
| | | 0.20 g/210L Test # 5 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.206 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 09/29/2016 11:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 10/13/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANADEZ,ANGEL, | **Time** | 17:11:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/20/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/13/2016 17:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 11/29/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 17:03:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2016 16:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 12/29/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 19:16:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/29/2016 19:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

**Agency** MIAMI PD **Date** 01/28/2016 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 15:39:54 **Software** 8100.27

**0.05 Lot #** 201403F **0.08 Lot #** 201404C **0.20 Lot #** 201408B **0.08 Gas Lot #** 1789170
**Expiration** 03/25/2016 **Expiration** 04/08/2016 **Expiration** 08/12/2016 **Expiration** 01/14/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC MESA AND OFC LOUREIRO PRESENT

**Data Download Date/Time** 01/28/2016 15:42

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 02/25/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 18:23:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 03/25/2016 | **Expiration** | 04/08/2016 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/25/2016 18:39

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 03/10/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 16:57:33 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 03/25/2016 | **Expiration** | 04/08/2016 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:OPERATOR ERROR.

**Data Download Date/Time** 03/10/2016 17:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 04/22/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 18:33:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/22/2016 18:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 05/25/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 18:30:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/25/2016 19:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 06/23/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 17:14:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/23/2016 17:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 09/15/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 11:08:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG600504 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.052 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.052 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 09/15/2016 11:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 09/22/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 17:14:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 203498 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 12/10/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/22/2016 17:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 10/13/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 16:57:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 203498 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/20/2017 | **Expiration** | 12/10/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/13/2016 16:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 11/29/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 16:55:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 203498 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 12/10/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2016 16:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 12/29/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 19:06:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 203498 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 12/10/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/29/2016 19:23

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

**Agency** MIAMI PD **Date** 01/28/2016 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 15:26:46 **Software** 8100.27

**0.05 Lot #** 201403F **0.08 Lot #** 201404C **0.20 Lot #** 201408B **0.08 Gas Lot #** 1789170
**Expiration** 03/25/2016 **Expiration** 04/08/2016 **Expiration** 08/12/2016 **Expiration** 01/14/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.211 |
| 0.20 g/210L Test # (Repeat) 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.052 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.208 |
| 0.20 g/210L Test # (Repeat) 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.206 |
| 0.20 g/210L Test # (Repeat) 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:OPERATOR ERROR. 20: Control Outside Tolerance.

**Data Download Date/Time** 01/28/2016 15:36

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

**Agency** MIAMI PD **Date** 02/25/2016 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 18:38:59 **Software** 8100.27

**0.05 Lot #** 201403F **0.08 Lot #** 201404C **0.20 Lot #** 201408B **0.08 Gas Lot #** 1789170
**Expiration** 03/25/2016 **Expiration** 04/08/2016 **Expiration** 08/12/2016 **Expiration** 01/14/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08:HOSE TO THE DRY GAS WAS NOT CONNECTED TEST REPEATED .

**Data Download Date/Time** 02/25/2016 18:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 03/10/2016 | **Type** | Agency Inspection |
| **Inspector** | REYES,BARBARA,T | **Time** | 16:49:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 03/25/2016 | **Expiration** | 04/08/2016 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/10/2016 16:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 04/22/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 18:46:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.083 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.083 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/22/2016 18:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 05/25/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 18:56:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.

**Data Download Date/Time** 05/25/2016 19:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 06/23/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 17:30:53 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/23/2016 17:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 07/28/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 19:52:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 07/28/2016 19:48

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 08/11/2016 | **Type** | Agency Inspection |
| **Inspector** | REYES,BARBARA,T | **Time** | 12:55:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1789170 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/14/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.083 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.083 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.206 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC. A. FERNANDEZ

**Data Download Date/Time** 08/11/2016 13:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 08/19/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 09:59:35 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG600504 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.052 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.052 |
| 0.08 g/210L Test # 6 | 0.082 |
| 0.20 g/210L Test # 6 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.052 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.052 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.052 |
| 0.08 g/210L Test # 10 | 0.082 |
| 0.20 g/210L Test # 10 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 08/19/2016 09:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 09/22/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 17:24:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 203498 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 12/10/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/22/2016 17:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 10/13/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 16:54:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 203498 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/20/2017 | **Expiration** | 12/10/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/13/2016 16:54

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 11/29/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 16:44:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 203498 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 12/10/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2016 16:42

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 12/29/2016 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 19:55:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 203498 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 12/10/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/29/2016 19:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 01/08/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 09:22:41 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 07014080A2 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 05/01/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 01/08/2016 10:52

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 02/18/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 08:41:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 34015080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 02/18/2016 09:54

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 03/28/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 08:24:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 34015080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 03/28/2016 09:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 04/06/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 09:23:52 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 34015080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 04/06/2016 10:35

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 05/13/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 08:59:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 34015080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 05/13/2016 10:07

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 06/06/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 09:36:51 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 34015080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 06/06/2016 11:20

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 07/07/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 09:32:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 34015080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: RFI Detect.UNKNOWN CAUSE OF RFI.

**Data Download Date/Time** 07/07/2016 10:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 08/10/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 09:12:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 34015080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 04/05/2018 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 08/10/2016 10:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CALHOUN COUNTY SO | **Date** | 09/26/2016 | **Type** | Agency Inspection |
| **Inspector** | WILSON,STANLEY,T | **Time** | 09:31:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 34015080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 09/26/2016 10:48

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO **Date** 10/28/2016 **Type** Department Inspection
**Inspector** MURPHY,PATRICK,J **Time** 10:54:25 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** AG626605
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 09/22/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.076 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.076 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.076 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.076 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.076 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.076 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 10/28/2016 11:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 01/21/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 11:07:53 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/21/2016 11:08

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 02/11/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 13:07:17 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 02/11/2016 13:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

**Agency** HIALEAH PD **Date** 03/31/2016 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 10:15:23 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG519701
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 07/16/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 03/31/2016 10:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 04/19/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 12:30:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/19/2016 13:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

**Agency** HIALEAH PD **Date** 05/18/2016 **Type** Agency Inspection
**Inspector** MONTOYA,GENARO,N **Time** 11:48:49 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 06415080A1
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 04/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/18/2016 11:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 06/22/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 12:59:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/22/2016 12:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 07/14/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 12:45:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/14/2016 13:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 08/22/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 12:10:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/22/2016 12:08

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 09/21/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 11:46:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

MA.

**Data Download Date/Time** 09/21/2016 11:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 10/12/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 12:36:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.075 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2016 12:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 11/16/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 12:09:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.074 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.076 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.075 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.074 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.075 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

08:Control Outside Tolerance.

**Data Download Date/Time** 11/16/2016 12:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIALEAH PD | **Date** | 12/21/2016 | **Type** | Agency Inspection |
| **Inspector** | MONTOYA,GENARO,N | **Time** | 12:49:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 06415080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.075 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/21/2016 12:48

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOMESTEAD PD | **Date** | 01/28/2016 | **Type** | Agency Inspection |
| **Inspector** | NEWTON,JOSHUA, | **Time** | 14:05:24 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1673300 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/28/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/28/2016 14:36

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOMESTEAD PD | **Date** | 02/26/2016 | **Type** | Agency Inspection |
| **Inspector** | NEWTON,JOSHUA, | **Time** | 16:46:41 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1673300 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/28/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/29/2016 07:13

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOMESTEAD PD | **Date** | 03/29/2016 | **Type** | Agency Inspection |
| **Inspector** | NEWTON,JOSHUA, | **Time** | 09:41:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1673300 |
| **Expiration** | 11/25/2016 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 04/28/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/29/2016 09:56

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOMESTEAD PD | **Date** | 04/27/2016 | **Type** | Agency Inspection |
| **Inspector** | NEWTON,JOSHUA, | **Time** | 09:07:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1673300 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 04/28/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2016 00:53

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD **Date** 05/30/2016 **Type** Agency Inspection
**Inspector** NEWTON,JOSHUA, **Time** 15:52:53 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 1673300
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 04/28/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/04/2016 10:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD **Date** 06/28/2016 **Type** Agency Inspection
**Inspector** NEWTON,JOSHUA, **Time** 09:48:27 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 1673300
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 04/28/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2016 10:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD **Date** 07/27/2016 **Type** Agency Inspection
**Inspector** NEWTON,JOSHUA, **Time** 09:09:10 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 1763300
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 04/28/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2016 07:23

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD **Date** 08/11/2016 **Type** Agency Inspection
**Inspector** NEWTON,JOSHUA, **Time** 09:21:45 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 1673300
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 04/28/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/11/2016 10:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOMESTEAD PD | **Date** | 09/27/2016 | **Type** | Agency Inspection |
| **Inspector** | NEWTON,JOSHUA, | **Time** | 15:23:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1673300 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 04/28/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2016 16:08

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD **Date** 09/29/2016 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 12:16:44 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG600504
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 01/05/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 09/29/2016 12:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOMESTEAD PD | **Date** | 10/27/2016 | **Type** | Agency Inspection |
| **Inspector** | NEWTON,JOSHUA, | **Time** | 09:33:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1673300 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 04/28/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/27/2016 10:16

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOMESTEAD PD | **Date** | 11/28/2016 | **Type** | Agency Inspection |
| **Inspector** | NEWTON,JOSHUA, | **Time** | 18:55:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1673300 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 04/28/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/02/2016 08:32

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD **Date** 12/29/2016 **Type** Agency Inspection
**Inspector** NEWTON,JOSHUA, **Time** 09:19:21 **Software** 8100.27

**0.05 Lot #** 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 1673300
**Expiration** 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 04/28/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.032 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.051 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.040 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.040 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance.SIMULATOR NOT SEALED PROPERLY

**Data Download Date/Time** 12/30/2016 07:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 01/26/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 10:13:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/26/2016 10:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 02/23/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 10:16:39 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/23/2016 10:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 03/21/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 14:49:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/21/2016 14:57

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 04/27/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 12:58:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/27/2016 13:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 05/26/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 10:43:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/26/2016 10:43

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 06/14/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 10:59:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/14/2016 10:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 07/26/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 11:29:31 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/26/2016 11:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 08/25/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 10:06:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/25/2016 10:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 09/26/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 11:48:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/26/2016 11:47

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 10/10/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 10:48:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG600504 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 01/05/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 10/10/2016 10:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 10/26/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 10:48:33 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/26/2016 11:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 11/29/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 11:02:06 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2016 11:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI BEACH PD | **Date** | 12/20/2016 | **Type** | Agency Inspection |
| **Inspector** | MILLAN,KEVIN, | **Time** | 14:10:12 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 1826916 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 03/09/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI N. GUASTO (848) PRESENT

**Data Download Date/Time** 12/20/2016 14:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 01/05/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:38:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/05/2016 10:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 02/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:35:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/12/2016 09:32

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD **Date** 02/16/2016 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 10:38:50 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG507503
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 03/16/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 02/16/2016 10:36

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD **Date** 03/02/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 10:02:34 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 19014080A3
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.061 |
| 0.08 g/210L Test # (Repeat) 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.068 |
| 0.08 g/210L Test # (Repeat) 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.062 |
| 0.08 g/210L Test # (Repeat) 3 | 0.075 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance/CHANGE O RING IN SIM.

**Data Download Date/Time** 03/02/2016 10:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 04/07/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 12:20:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201411K | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | |
| **Expiration** | 11/25/2016 | **Expiration** | 11/25/2016 | **Expiration** | 05/12/2017 | **Expiration** | |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.198 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.20 g/210L Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | No | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: RFI Detect/PHONE REMOVED. Non-compliance:RFI/RESTART AI PER DAVID REYES FDLE.

**Data Download Date/Time** 04/17/2016 02:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 04/17/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 02:34:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI RESTARTED PER FDLE DAVID REYES

**Data Download Date/Time** 04/17/2016 02:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 05/04/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:37:17 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/04/2016 09:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 01/05/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:34:52 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/05/2016 10:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 02/05/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:35:35 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/05/2016 09:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 02/09/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 11:44:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG507503 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/16/2017 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.082 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.082 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 02/09/2016 11:42

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 02/11/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:31:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/11/2016 10:00

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 03/23/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:12:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/23/2016 10:09

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 04/20/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:15:47 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/20/2016 10:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 05/24/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:00:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/24/2016 09:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 06/03/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:05:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/03/2016 10:20

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD **Date** 07/22/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 09:28:22 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 19014080A3
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/22/2016 09:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 08/10/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:20:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/10/2016 10:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 09/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:12:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/12/2016 10:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 10/04/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 12:08:06 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/04/2016 12:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI DADE PD | **Date** | 11/17/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 13:17:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/17/2016 13:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD **Date** 12/06/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 09:50:27 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 22315080A2
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/06/2016 09:46

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD **Date** 01/07/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 09:16:47 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201408B **0.08 Gas Lot #** 19014080A3
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 08/12/2016 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.052 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/07/2016 09:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 01/08/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 13:54:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG507503 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 03/16/2017 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.052 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 01/08/2016 13:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 01/20/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:13:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/20/2016 10:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD **Date** 02/05/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 09:39:50 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201408B **0.08 Gas Lot #** 19014080A3
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 08/12/2016 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/05/2016 09:36

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 03/03/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:11:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/03/2016 09:07

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 04/28/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 08:43:45 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2016 12:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 05/04/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:33:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/04/2016 09:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 06/22/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:07:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/22/2016 09:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 07/08/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:31:51 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/08/2016 09:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 08/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 21:35:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/13/2016 00:24

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 09/02/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 08:42:47 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/02/2016 08:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 10/13/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:11:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/13/2016 10:09

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 11/09/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:57:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/09/2016 12:22

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD **Date** 12/07/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 08:56:53 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 22315080A2
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/07/2016 08:56

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 01/21/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:37:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/21/2016 09:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD **Date** 02/04/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 11:24:07 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201408B **0.08 Gas Lot #** 19014080A3
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 08/12/2016 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/04/2016 11:24

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 03/03/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:14:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/03/2016 12:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 03/07/2016 | **Type** | Department Inspection |
| **Inspector** | REYES-RIVERA,DAVID,E | **Time** | 13:39:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG519701 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 07/16/2017 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0007 |
| 0.20 g/210L Test (Std Dev) | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 03/07/2016 13:41

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 03/10/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:21:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/10/2016 09:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD **Date** 04/20/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 10:10:14 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 19014080A3
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/20/2016 10:08

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD **Date** 05/13/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 09:55:40 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 19014080A3
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/13/2016 09:54

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 06/24/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:43:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/24/2016 09:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 07/27/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:51:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/27/2016 09:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 08/03/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:47:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/03/2016 09:47

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 09/06/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 12:08:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/06/2016 12:07

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 10/10/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:12:06 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/10/2016 10:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 11/08/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:18:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/08/2016 11:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 12/08/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:36:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/08/2016 09:37

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 01/05/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:20:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/05/2016 09:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 02/04/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:24:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/04/2016 11:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD **Date** 02/09/2016 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 13:52:23 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG507503
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 03/16/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 02/09/2016 13:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 02/11/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:30:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/11/2016 10:03

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 03/24/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:21:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/24/2016 10:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 04/21/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:13:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/21/2016 11:07

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 05/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:20:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/12/2016 10:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 06/23/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:44:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/23/2016 10:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD **Date** 07/21/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 11:36:01 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 19014080A3
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.000 |
| 0.05 g/210L Test # (Repeat) 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.05 g/210L Test # (Repeat) 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.05 g/210L Test # (Repeat) 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance/SIM NOT CLOSED PROPERLY.

**Data Download Date/Time** 07/21/2016 11:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 08/18/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:51:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/18/2016 10:45

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 09/05/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 12:04:35 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/05/2016 11:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 10/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:06:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2016 11:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 11/11/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:32:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/11/2016 10:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 12/27/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:12:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.085 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.085 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.085 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/27/2016 11:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 01/05/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:23:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/05/2016 09:20

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD **Date** 01/21/2016 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 15:23:35 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG431502
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 11/11/2016

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 01/21/2016 15:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 01/28/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:27:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/28/2016 09:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 02/02/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 09:22:04 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/02/2016 09:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 03/24/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:21:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/24/2016 10:24

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD **Date** 04/21/2016 **Type** Agency Inspection
**Inspector** ESPINOZA,PABLO, **Time** 11:07:21 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 19014080A3
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 08/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/21/2016 11:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 05/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:13:35 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/12/2016 10:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 06/23/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:09:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/23/2016 11:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 07/21/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:25:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19014080A3 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 08/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/21/2016 11:20

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 08/18/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:55:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/18/2016 10:50

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 09/05/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 12:07:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/05/2016 12:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 10/12/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:13:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2016 11:13

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 11/11/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 10:27:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/11/2016 10:22

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 12/27/2016 | **Type** | Agency Inspection |
| **Inspector** | ESPINOZA,PABLO, | **Time** | 11:13:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 22315080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/27/2016 11:08

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 01/28/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 22:47:24 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201402A | **0.08 Gas Lot #** | 1878845 |
| **Expiration** | 03/25/2016 | **Expiration** | 04/08/2016 | **Expiration** | 02/11/2016 | **Expiration** | 06/22/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/28/2016 22:44

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 02/28/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 08:55:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201403F | **0.08 Lot #** | 201404C | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1878845 |
| **Expiration** | 03/25/2016 | **Expiration** | 04/08/2016 | **Expiration** | 05/12/2017 | **Expiration** | 06/22/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/28/2016 08:52

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 03/28/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 18:09:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1878845 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/22/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.143 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.198 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance/IMPROPER SEAL.

**Data Download Date/Time** 03/28/2016 18:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 04/29/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 22:35:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1878845 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/22/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/02/2016 08:19

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 05/31/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 23:04:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1878845 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/22/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

00: RFI Detect/OFC WALKED BY W/RADIO ON..

**Data Download Date/Time** 05/31/2016 23:00

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 06/23/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 13:23:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 1878845 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 06/22/2018 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/23/2016 13:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD **Date** 07/28/2016 **Type** Agency Inspection
**Inspector** MANN,CORA,L **Time** 13:20:18 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 1878845
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 06/22/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.093 |
| 0.08 g/210L Test # (Repeat) 1 | 0.082 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.087 |
| 0.08 g/210L Test # (Repeat) 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.085 |
| 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance: SOLUTIONS TOO HOT. .

**Data Download Date/Time** 07/28/2016 13:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD **Date** 08/31/2016 **Type** Agency Inspection
**Inspector** BAKER,KENNETH, **Time** 01:53:12 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 19615080A2
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/31/2016 01:47

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 09/29/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 12:06:58 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/29/2016 12:03

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 10/26/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 01:39:52 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/26/2016 01:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 11/29/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 13:51:38 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.189 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.194 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.192 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance/IMPROPER SEAL..

**Data Download Date/Time** 11/29/2016 13:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD **Date** 12/01/2016 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 10:04:50 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG605301
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 02/22/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 12/01/2016 09:58

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | NORTH MIAMI BEACH PD | **Date** | 12/02/2016 | **Type** | Agency Inspection |
| **Inspector** | MANN,CORA,L | **Time** | 18:54:38 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A2 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/02/2016 18:49

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 01/17/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 14:24:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 08/12/2016 | **Expiration** | 09/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/17/2016 14:32

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 02/26/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 21:34:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.152 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.198 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.173 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.182 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 02/26/2016 21:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 03/22/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 19:15:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201408B | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 08/12/2016 | **Expiration** | 09/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.206 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/22/2016 20:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 04/27/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 22:05:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 3 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/27/2016 22:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 06/09/2016 | **Type** | Agency Inspection |
| **Inspector** | SUBER,PHILLIP,W | **Time** | 10:07:53 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

SKIPPED AI TO OPERATE INSTRUMENT; NOT A COMPLIANCE CHECK

**Data Download Date/Time** 06/09/2016 14:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD **Date** 06/09/2016 **Type** Department Inspection
**Inspector** SUBER,PHILLIP,W **Time** 14:34:24 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG600504
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 01/05/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

**Remarks:**

**Data Download Date/Time** 06/09/2016 14:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 07/03/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 13:34:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/03/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/03/2016 13:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD **Date** 08/22/2016 **Type** Agency Inspection
**Inspector** CAMPBELL,WILLIAM,A **Time** 18:38:07 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 19615080A4
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.000 |
| 0.20 g/210L Test # (Repeat) 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.000 |
| 0.20 g/210L Test # (Repeat) 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.000 |
| 0.20 g/210L Test # (Repeat) 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IMPROPER CONNECTION TO INTOXILYZER RECONNECT RETEST.

**Data Download Date/Time** 08/22/2016 19:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 09/15/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 18:54:55 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/15/2016 19:42

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 10/26/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 16:46:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/26/2016 17:04

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 11/23/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 19:34:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/23/2016 20:12

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHATTAHOOCHEE PD | **Date** | 12/16/2016 | **Type** | Agency Inspection |
| **Inspector** | CAMPBELL,WILLIAM,A | **Time** | 19:37:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201603D | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 03/08/2018 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 09/05/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/16/2016 21:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PINELLAS COUNTY SO | **Date** | 01/30/2016 | **Type** | Agency Inspection |
| **Inspector** | DEANE,M, | **Time** | 20:07:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG505803 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 02/27/2017 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/30/2016 20:08

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

**Agency** PINELLAS COUNTY SO **Date** 02/22/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 12:44:14 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** AG511701
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 04/27/2017

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 02/22/2016 12:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

**Agency** PINELLAS COUNTY SO **Date** 08/01/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 20:40:53 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG612405
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 05/03/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.081 |
| 0.20 g/210L Test # 4 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.082 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.082 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 08/01/2016 20:36

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PINELLAS COUNTY SO | **Date** | 06/02/2016 | **Type** | Agency Inspection |
| **Inspector** | BELL,DANIELLE,M | **Time** | 12:48:03 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

**Data Download Date/Time** 06/22/2016 16:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

**Agency** PINELLAS COUNTY SO **Date** 07/08/2016 **Type** Department Inspection
**Inspector** BELL,DANIELLE,M **Time** 19:30:39 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** AG605301
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 02/22/2018

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 07/08/2016 19:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 01/21/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 12:29:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/21/2016 13:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 02/11/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 12:08:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/11/2016 13:03

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 03/22/2016 | **Type** | Agency Inspection | | |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 11:08:50 | **Software** | 8100.27 | | |
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/22/2016 12:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 04/15/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 12:18:53 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/15/2016 13:14

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 05/16/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 11:34:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/06/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | INT |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

05: Interferent Detect.

**Data Download Date/Time** 05/16/2016 12:36

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 06/21/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 15:45:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/06/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 06/21/2016 16:39

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 07/28/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 14:17:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2016 15:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 08/23/2016 **Type** Agency Inspection
**Inspector** MOLNAR,JARROD,T **Time** 16:10:03 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 28214080A1
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 11/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | 0.000 |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det:UNK.

**Data Download Date/Time** 08/23/2016 17:10

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 09/22/2016 | **Type** | Agency Inspection |
| **Inspector** | MOLNAR,JARROD,T | **Time** | 15:23:01 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/22/2016 16:27

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 09/28/2016 | **Type** | Department Inspection | | |
| **Inspector** | SUBER,PHILLIP,W | **Time** | 13:47:59 | **Software** | 8100.27 | | |
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG612405 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 05/03/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.053 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.053 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.053 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.052 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.053 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.053 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 09/28/2016 14:40

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 10/15/2016 | **Type** | Agency Inspection |
| **Inspector** | MOLNAR,JARROD,T | **Time** | 12:26:06 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 507129 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 08/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.083 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | No | 0.08 g/210L Dry Gas Std Test # 1 | 0.084 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.083 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.084 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.083 |
| | | 0.20 g/210L Test # 3 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.084 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/15/2016 13:19

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 10/16/2016 | **Type** | Agency Inspection |
| **Inspector** | MOLNAR,JARROD,T | **Time** | 10:39:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 507129 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 08/04/2019 |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | No | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

.007 ALCOHOL FREE SUBJECT TESTNon-compliance:.007 ALCOHOL FREE SUBJECT TEST.

**Data Download Date/Time** 10/16/2016 14:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 11/03/2016 | **Type** | Department Inspection | | |
| **Inspector** | MURPHY,PATRICK,J | **Time** | 12:37:34 | **Software** | 8100.27 | | |
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG619605 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 07/14/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 11/03/2016 13:30

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 11/16/2016 | **Type** | Agency Inspection |
| **Inspector** | MOLNAR,JARROD,T | **Time** | 12:04:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 507129 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/16/2016 13:00

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 12/21/2016 | **Type** | Agency Inspection |
| **Inspector** | MOLNAR,JARROD,T | **Time** | 15:43:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | 507129 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 08/04/2019 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/21/2016 16:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 01/21/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 12:23:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/21/2016 13:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 02/11/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 12:02:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.143 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.201 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 02/11/2016 13:00

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 03/22/2016 **Type** Agency Inspection
**Inspector** GOLEY,ROBERT,A **Time** 10:58:36 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 28214080A1
**Expiration** 07/14/2017 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 11/05/2016

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/22/2016 12:11

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 04/15/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 12:08:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/15/2016 13:09

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 05/16/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 11:31:33 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/16/2016 12:31

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 06/21/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 15:35:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/21/2016 16:34

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 07/28/2016 | **Type** | Agency Inspection |
| **Inspector** | GOLEY,ROBERT,A | **Time** | 14:02:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2016 15:02

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 08/23/2016 | **Type** | Agency Inspection |
| **Inspector** | MOLNAR,JARROD,T | **Time** | 16:04:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 28214080A1 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 11/05/2016 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.

**Data Download Date/Time** 08/23/2016 17:06

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 09/22/2016 **Type** Agency Inspection
**Inspector** MOLNAR,JARROD,T **Time** 15:15:34 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 28214080A1
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 11/05/2016

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/22/2016 16:25

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 09/28/2016 | **Type** | Department Inspection | | |
| **Inspector** | SUBER,PHILLIP,W | **Time** | 13:04:28 | **Software** | 8100.27 | | |
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201604C | **0.08 Gas Lot #** | AG612405 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 04/05/2018 | **Expiration** | 05/03/2018 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.075 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.075 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.075 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.074 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.074 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.074 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.075 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.074 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.074 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

08: Control Outside Tolerance IN NEED OF CAL ADJUST. Non-compliance:UNABLE TO PERFORM OPT BENCH CAL. SEND TO REP.

**Data Download Date/Time** 09/28/2016 14:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 01/25/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 13:43:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 01/25/2016 13:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 02/25/2016 **Type** Agency Inspection
**Inspector** JOHNSTON,ROBERT,B **Time** 14:09:59 **Software** 8100.27

**0.05 Lot #** 201411K **0.08 Lot #** 201502G **0.20 Lot #** 201505A **0.08 Gas Lot #** 19615080A4
**Expiration** 11/25/2016 **Expiration** 02/24/2017 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.20 g/210L Test # (Repeat) 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.171 |
| 0.20 g/210L Test # (Repeat) 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.20 g/210L Test # (Repeat) 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance HOSE KINK OUT OF SERVICE

**Data Download Date/Time** 02/25/2016 14:59

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 03/01/2016 | **Type** | Department Inspection | | |
| **Inspector** | JACOBS,JENNIFER,H | **Time** | 11:15:14 | **Software** | 8100.27 | | |
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG511701 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/27/2017 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 03/01/2016 11:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 03/02/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 15:32:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT BACK IN SERVICE

**Data Download Date/Time** 03/02/2016 15:55

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 04/12/2016 **Type** Agency Inspection
**Inspector** JOHNSTON,ROBERT,B **Time** 15:39:13 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 19615080A4
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 04/12/2016 16:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 05/20/2016 **Type** Agency Inspection
**Inspector** JOHNSTON,ROBERT,B **Time** 13:00:36 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 19615080A4
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 05/20/2016 13:22

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 06/24/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 15:34:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 06/24/2016 16:05

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 07/27/2016 **Type** Agency Inspection
**Inspector** JOHNSTON,ROBERT,B **Time** 13:19:01 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 19615080A4
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 07/27/2016 13:29

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 08/26/2016 **Type** Agency Inspection
**Inspector** JOHNSTON,ROBERT,B **Time** 14:14:21 **Software** 8100.27

**0.05 Lot #** 201507A **0.08 Lot #** 201601F **0.20 Lot #** 201505A **0.08 Gas Lot #** 19615080A4
**Expiration** 07/14/2017 **Expiration** 01/26/2018 **Expiration** 05/12/2017 **Expiration** 09/05/2017

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.074 |
| 0.08 g/210L Test # (Repeat) 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.189 |
| 0.20 g/210L Test # (Repeat) 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.08 g/210L Test # (Repeat) 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.190 |
| 0.20 g/210L Test # (Repeat) 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test # 3 | 0.075 |
| 0.08 g/210L Test # (Repeat) 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.191 |
| 0.20 g/210L Test # (Repeat) 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance. 20: Control Outside Tolerance.INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 08/26/2016 14:21

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 09/27/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 15:42:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 09/27/2016 16:13

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 10/20/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 14:54:33 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 10/20/2016 15:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 11/29/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 12:52:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.018 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.018 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.021 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 11/29/2016 13:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 01/25/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 13:46:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 11/26/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 01/25/2016 13:52

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 02/25/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 13:57:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201411K | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 11/25/2016 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE. OUT OF SERVICE DEPT INSPECTION

**Data Download Date/Time** 02/25/2016 14:38

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 03/01/2016 | **Type** | Department Inspection | | |
| **Inspector** | JACOBS,JENNIFER,H | **Time** | 11:34:17 | **Software** | 8100.27 | | |
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201502G | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | AG511701 |
| **Expiration** | 07/14/2017 | **Expiration** | 02/24/2017 | **Expiration** | 05/12/2017 | **Expiration** | 04/27/2017 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.079 |
| | | 0.20 g/210L Test # 5 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 03/01/2016 11:26

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 03/02/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 15:22:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 03/02/2016 15:51

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 04/12/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 15:32:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 04/12/2016 16:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 05/20/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 12:53:51 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 05/20/2016 13:18

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 06/24/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 15:37:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 06/24/2016 16:01

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 07/27/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 13:13:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 07/27/2016 13:33

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 08/26/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 14:17:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 08/26/2016 14:17

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 09/27/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 15:37:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 09/27/2016 16:15

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 10/20/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 15:01:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 201507A | **0.08 Lot #** | 201601F | **0.20 Lot #** | 201505A | **0.08 Gas Lot #** | 19615080A4 |
| **Expiration** | 07/14/2017 | **Expiration** | 01/26/2018 | **Expiration** | 05/12/2017 | **Expiration** | 09/05/2017 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 10/20/2016 15:42

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FWC | **Date** | 11/29/2016 | **Type** | Agency Inspection |
| **Inspector** | JOHNSTON,ROBERT,B | **Time** | 12:05:49 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 4 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Range ExceededAmbient FailAmbient FailAmbient Fa. AMBIENT FAIL DUE MOUTH ALCOHOL IN TUBENon-compliance:MOUTH ALCOHOL IN TUBE CAUSING AMBIENT FAIL.

**Data Download Date/Time** 11/29/2016 13:23

**Printed: 27 January 2017**

COBRA ©1997-2014 CMI, Inc.