

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

Agency: FDLE

Date/Time: 11/12/2018 12:25:36

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/16/2018 15:22

Air Blank	0.000	12:24
Control Test	0.079	12:25
Air Blank	0.000	12:25

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

Agency: FDLE

Date/Time: 02/24/2018 06:11:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/05/2018 21:11

Air Blank	0.000	06:11
Control Test	0.078	06:11
Air Blank	0.000	06:11

Agency: FDLE

Date/Time: 02/24/2018 06:13:21

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/05/2018 21:11

Air Blank		06:12
Air Blank	0.000	06:13

Agency: FDLE

Date/Time: 02/24/2018 06:14:49

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/05/2018 21:11

Air Blank		06:13
Air Blank	0.000	06:14

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

Agency: FDLE

Date/Time: 02/24/2018 06:16:39

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/05/2018 21:11

Air Blank	0.000	06:15
ITP Check	0.076	06:16
Air Blank	0.000	06:16

Agency: FDLE

Date/Time: 03/01/2018 16:17:14

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/05/2018 21:11

Air Blank	0.000	16:15
ITP Check	0.078	16:16
Air Blank		16:16
Air Blank	0.000	16:17

Agency: FDLE

Date/Time: 03/01/2018 16:18:44

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/05/2018 21:11

Air Blank	0.000	16:17
ITP Check	0.078	16:18
Air Blank	0.000	16:18

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

Agency: FDLE

Date/Time: 04/02/2018 14:03:21

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/03/2018 19:30

Air Blank	0.000	14:02
Control Test	0.084	14:02
Air Blank	0.000	14:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

Agency: FDLE

Date/Time: 12/20/2018 12:09:33

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/20/2018 12:59

Air Blank	0.000	12:08
Control Test	0.079	12:09
Air Blank	0.000	12:09

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

Agency: ALACHUA COUNTY SO

Date/Time: 07/30/2018 10:10:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/30/2018 10:52

Air Blank	0.000	10:08
Control Test	0.193	10:09
Air Blank	0.000	10:10

Agency: ALACHUA COUNTY SO

Date/Time: 07/30/2018 10:12:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/30/2018 10:52

Air Blank	0.000	10:11
Control Test	0.076	10:11
Air Blank	0.000	10:12

Agency: ALACHUA COUNTY SO

Date/Time: 08/27/2018 09:52:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/27/2018 10:32

Air Blank	0.000	09:51
Control Test	0.048	09:51
Air Blank	0.000	09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

Agency: ALACHUA COUNTY SO

Date/Time: 09/24/2018 09:08:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2018 09:46

Air Blank	0.000	09:06
Control Test	0.077	09:07
Air Blank	0.000	09:08

Agency: ALACHUA COUNTY SO

Date/Time: 11/27/2018 09:30:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/27/2018 10:10

Air Blank	0.000	09:29
Control Test	0.078	09:30
Air Blank	0.000	09:30

Agency: ALACHUA COUNTY SO

Date/Time: 11/27/2018 09:32:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/27/2018 10:10

Air Blank	0.000	09:31
Control Test	0.196	09:32
Air Blank	0.000	09:32

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

Agency: OKALOOSA COUNTY SO

Date/Time: 12/14/2018 13:03:15

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/14/2018 15:43

Air Blank	0.000	13:02
Control Test	0.078	13:02
Air Blank	0.000	13:03

Agency: OKALOOSA COUNTY SO

Date/Time: 12/14/2018 13:50:22

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/14/2018 15:43

Air Blank	0.000	13:49
Control Test	0.079	13:49
Air Blank	0.000	13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

Agency: VALPARAISO PD

Date/Time: 05/21/2018 13:31:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/21/2018 15:33

Air Blank	0.000	13:30
Control Test	0.000	13:30
Air Blank	0.000	13:31

Agency: VALPARAISO PD

Date/Time: 10/26/2018 00:54:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/26/2018 02:48

Air Blank	0.000	00:53
Control Test	0.047	00:54
Air Blank	0.000	00:54

Agency: VALPARAISO PD

Date/Time: 12/04/2018 14:38:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2018 17:16

Air Blank	0.000	14:37
Control Test	0.048	14:37
Air Blank	0.000	14:38

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

Agency: VALPARAISO PD

Date/Time: 12/04/2018 14:45:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2018 17:16

Air Blank	0.000	14:44
Control Test	0.079	14:45
Air Blank	0.000	14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

Agency: ALACHUA COUNTY SO

Date/Time: 03/27/2018 09:08:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2018 09:45

Air Blank	0.000	09:07
Control Test	0.048	09:07
Air Blank	0.000	09:08

Agency: ALACHUA COUNTY SO

Date/Time: 03/27/2018 09:10:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2018 09:45

Air Blank	0.000	09:09
Control Test	0.078	09:09
Air Blank	0.000	09:10

Agency: ALACHUA COUNTY SO

Date/Time: 03/27/2018 09:12:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2018 09:45

Air Blank	0.000	09:11
Control Test	0.199	09:11
Air Blank	0.000	09:12

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

Agency: ALACHUA COUNTY SO

Date/Time: 04/23/2018 08:57:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 09:41

Air Blank	0.000	08:56
Control Test	0.000	08:56
Air Blank	0.000	08:57

Agency: ALACHUA COUNTY SO

Date/Time: 04/23/2018 08:59:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 09:41

Air Blank	0.000	08:58
Control Test	0.049	08:58
Air Blank	0.000	08:59

Agency: ALACHUA COUNTY SO

Date/Time: 04/23/2018 09:01:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 09:41

Air Blank	0.000	09:00
Control Test	0.079	09:00
Air Blank	0.000	09:01

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

Agency: ALACHUA COUNTY SO

Date/Time: 04/23/2018 09:03:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 09:41

Air Blank	0.000	09:02
Control Test	0.200	09:03
Air Blank	0.000	09:03

Agency: ALACHUA COUNTY SO

Date/Time: 06/25/2018 14:41:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/25/2018 15:26

Air Blank	0.000	14:40
Control Test	0.000	14:40
Air Blank	0.000	14:41

Agency: ALACHUA COUNTY SO

Date/Time: 06/25/2018 14:43:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/25/2018 15:26

Air Blank	0.000	14:42
Control Test	0.049	14:42
Air Blank	0.000	14:43

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

Agency: ALACHUA COUNTY SO

Date/Time: 06/25/2018 14:45:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/25/2018 15:26

Air Blank	0.000	14:44
Control Test	0.077	14:44
Air Blank	0.000	14:45

Agency: ALACHUA COUNTY SO

Date/Time: 06/25/2018 14:47:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/25/2018 15:26

Air Blank	0.000	14:46
Control Test	0.198	14:46
Air Blank	0.000	14:47

Agency: ALACHUA COUNTY SO

Date/Time: 07/30/2018 10:12:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/30/2018 10:48

Air Blank	0.000	10:11
Control Test	0.000	10:12
Air Blank	0.000	10:12

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

Agency: ALACHUA COUNTY SO

Date/Time: 07/30/2018 10:14:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/30/2018 10:48

Air Blank	0.000	10:13
Control Test	0.048	10:14
Air Blank	0.000	10:14

Agency: ALACHUA COUNTY SO

Date/Time: 08/27/2018 09:48:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/27/2018 10:31

Air Blank	0.000	09:47
Control Test	0.199	09:47
Air Blank	0.000	09:48

Agency: ALACHUA COUNTY SO

Date/Time: 08/27/2018 09:51:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/27/2018 10:31

Air Blank	0.000	09:50
Control Test	0.077	09:50
Air Blank	0.000	09:51

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

Agency: ALACHUA COUNTY SO

Date/Time: 08/27/2018 09:53:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/27/2018 10:31

Air Blank	0.000	09:52
Control Test	0.000	09:52
Air Blank	0.000	09:53

Agency: ALACHUA COUNTY SO

Date/Time: 09/24/2018 09:06:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2018 09:43

Air Blank	0.000	09:05
Control Test	0.000	09:05
Air Blank	0.000	09:06

Agency: ALACHUA COUNTY SO

Date/Time: 09/24/2018 09:08:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2018 09:43

Air Blank	0.000	09:07
Control Test	0.048	09:07
Air Blank	0.000	09:08

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

Agency: ALACHUA COUNTY SO

Date/Time: 11/27/2018 09:29:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/27/2018 10:08

Air Blank	0.000	09:27
Control Test	0.000	09:28
Air Blank	0.000	09:29

Agency: ALACHUA COUNTY SO

Date/Time: 11/27/2018 09:31:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/27/2018 10:08

Air Blank	0.000	09:30
Control Test	0.049	09:30
Air Blank	0.000	09:31

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

Agency: HAMILTON COUNTY SO
Date/Time: 11/30/2018 18:45:03
Standard Type: Dry Gas
Event Source: Local
Data Download Date/Time: 12/02/2018 03:44

Air Blank	0.000	18:44
Control Test	0.075	18:44
Air Blank	0.000	18:45

Agency: HAMILTON COUNTY SO
Date/Time: 12/05/2018 17:45:26
Standard Type: Dry Gas
Event Source: Local
Data Download Date/Time: 12/06/2018 19:43

Air Blank	0.000	17:44
Control Test	0.078	17:44
Air Blank	0.000	17:45

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 21:58:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/03/2019 15:51

Air Blank	0.000	21:57
Control Test	0.078	21:57
Air Blank	0.000	21:58

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 22:03:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2019 15:51

Air Blank	0.000	22:02
Control Test	0.052	22:03
Air Blank	0.000	22:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

Agency: COLUMBIA COUNTY SO

Date/Time: 04/07/2018 19:27:12

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/30/2018 01:28

Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:27

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

Agency: FHP

Date/Time: 05/08/2018 19:09:58

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 05/09/2018 12:14

Air Blank	0.000	19:09
ITP Check	0.080	19:09
Air Blank	0.000	19:09

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

Agency: GULF BREEZE PD

Date/Time: 03/29/2018 11:47:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/16/2018 10:46

Air Blank	0.000	11:46
Control Test	0.134	11:47
Air Blank	0.000	11:47

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

Agency: ST LUCIE COUNTY SO

Date/Time: 03/29/2018 08:32:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 07:36

Air Blank	0.000	08:31
Control Test	0.068	08:31
Air Blank	0.000	08:32

Agency: ST LUCIE COUNTY SO

Date/Time: 06/15/2018 08:54:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/12/2018 08:45

Air Blank	0.000	08:53
Control Test	0.072	08:54
Air Blank	0.000	08:54

Agency: ST LUCIE COUNTY SO

Date/Time: 06/15/2018 08:57:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/12/2018 08:45

Air Blank	0.000	08:56
Control Test	0.164	08:57
Air Blank	0.000	08:57

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

Agency: ST LUCIE COUNTY SO

Date/Time: 06/29/2018 19:09:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/12/2018 08:45

Air Blank	0.000	19:08
Control Test	0.083	19:08
Air Blank	0.000	19:09

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH P.D.

Date/Time: 12/24/2018 11:59:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/09/2019 13:56

Air Blank	0.000	11:58
Control Test	0.000	11:59
Air Blank	0.000	11:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

Agency: ST LUCIE COUNTY SO

Date/Time: 03/29/2018 07:38:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/29/2018 08:24

Air Blank	0.000	07:36
Control Test	0.000	07:37
Air Blank	0.000	07:38

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

Agency: ST LUCIE COUNTY SO

Date/Time: 04/04/2018 18:27:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 07:38

Air Blank	0.000	18:26
Control Test	0.050	18:27
Air Blank	0.000	18:27

Agency: ST LUCIE COUNTY SO

Date/Time: 04/04/2018 18:30:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 07:38

Air Blank	0.000	18:29
Control Test	0.197	18:29
Air Blank	0.000	18:30

Agency: ST LUCIE COUNTY SO

Date/Time: 04/04/2018 20:03:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 07:38

Air Blank	0.000	20:02
Control Test	0.202	20:02
Air Blank	0.000	20:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

Agency: ST LUCIE COUNTY SO

Date/Time: 04/04/2018 20:05:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/23/2018 07:38

Air Blank	0.000	20:04
Control Test	0.077	20:05
Air Blank	0.000	20:05

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

Agency: MARION COUNTY SO

Date/Time: 08/27/2018 12:13:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/27/2018 13:18

Air Blank	0.000	12:12
Control Test	0.050	12:12
Air Blank	0.000	12:13

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

Agency: HERNANDO COUNTY
S.O.

Date/Time: 02/17/2018 11:00:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/17/2018 13:05

Air Blank	0.000	10:59
Control Test	0.047	11:00
Air Blank	0.000	11:00

Agency: HERNANDO COUNTY
S.O.

Date/Time: 07/22/2018 09:48:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/22/2018 11:37

Air Blank	0.000	09:47
Control Test	0.000	09:48
Air Blank	0.000	09:48

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency: CITRUS COUNTY SO

Date/Time: 05/30/2018 07:42:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2018 10:15

Air Blank	0.000	07:40
Control Test	0.166	07:41
Air Blank	0.000	07:42

Agency: CITRUS COUNTY SO

Date/Time: 05/30/2018 07:44:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2018 10:15

Air Blank	0.000	07:43
Control Test	0.191	07:44
Air Blank	0.000	07:44

Agency: CITRUS COUNTY SO

Date/Time: 05/30/2018 07:47:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2018 10:15

Air Blank	0.000	07:46
Control Test	0.196	07:46
Air Blank	0.000	07:47

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency: CITRUS COUNTY SO

Date/Time: 12/18/2018 12:46:40

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/31/2018 10:11

Air Blank	0.000	12:45
Control Test	0.080	12:46
Air Blank	0.000	12:46

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency: LAKE COUNTY SO

Date/Time: 07/11/2018 22:33:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/15/2018 23:57

Air Blank	0.000	22:32
Control Test	0.076	22:32
Air Blank	0.000	22:33

Agency: LAKE COUNTY SO

Date/Time: 07/11/2018 22:34:51

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/15/2018 23:57

Air Blank	0.000	22:34
Control Test	0.075	22:34
Air Blank	0.000	22:34

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 06/26/2018 13:22:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 14:24

Air Blank	0.000	13:21
Control Test	0.000	13:22
Air Blank	0.000	13:22

Agency: MARTIN COUNTY SO

Date/Time: 06/26/2018 13:25:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 14:24

Air Blank	0.000	13:24
Control Test	0.000	13:24
Air Blank	0.000	13:25

Agency: MARTIN COUNTY SO

Date/Time: 06/26/2018 13:27:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 14:24

Air Blank	0.000	13:26
Control Test	0.048	13:27
Air Blank	0.000	13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 10:04:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 16:47

Air Blank	0.000	10:03
Control Test	0.079	10:04
Air Blank	0.000	10:04

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 10:06:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 16:47

Air Blank	0.000	10:05
Control Test	0.079	10:06
Air Blank	0.000	10:06

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 10:14:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 16:47

Air Blank	0.000	10:12
Control Test	0.049	10:13
Air Blank	0.000	10:14

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 10:16:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 16:47

Air Blank	0.000	10:14
Control Test	0.050	10:15
Air Blank	0.000	10:16

Agency: MARTIN COUNTY SO

Date/Time: 09/17/2018 14:16:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/17/2018 15:27

Air Blank	0.000	14:15
Control Test	0.000	14:15
Air Blank	0.000	14:16

Agency: MARTIN COUNTY SO

Date/Time: 09/17/2018 14:18:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/17/2018 15:27

Air Blank	0.000	14:17
Control Test	0.000	14:18
Air Blank	0.000	14:18

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 09/17/2018 14:20:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/17/2018 15:27

Air Blank	0.000	14:19
Control Test	0.000	14:20
Air Blank	0.000	14:20

Agency: MARTIN COUNTY SO

Date/Time: 09/17/2018 14:22:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/17/2018 15:27

Air Blank	0.000	14:21
Control Test	0.050	14:22
Air Blank	0.000	14:22

Agency: MARTIN COUNTY SO

Date/Time: 09/17/2018 14:24:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/17/2018 15:27

Air Blank	0.000	14:23
Control Test	0.078	14:23
Air Blank	0.000	14:24

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 09/17/2018 14:26:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/17/2018 15:27

Air Blank	0.000	14:25
Control Test	0.194	14:25
Air Blank	0.000	14:26

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH
PD

Date/Time: 12/02/2018 14:57:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/02/2018 18:57

Air Blank	0.000	14:56
Control Test	0.049	14:56
Air Blank	0.000	14:57

Agency: PANAMA CITY BEACH
PD

Date/Time: 12/02/2018 14:59:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/02/2018 18:57

Air Blank	0.000	14:58
Control Test	0.015	14:59
Air Blank	0.000	14:59

Agency: PANAMA CITY BEACH
PD

Date/Time: 12/02/2018 15:02:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/02/2018 18:57

Air Blank	0.000	15:00
Control Test	0.196	15:01
Air Blank	0.000	15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH
PD

Date/Time: 12/02/2018 15:04:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/02/2018 18:57

Air Blank	0.000	15:02
Control Test	0.014	15:03
Air Blank	0.000	15:04

Agency: PANAMA CITY BEACH
PD

Date/Time: 12/02/2018 15:05:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/02/2018 18:57

Air Blank	0.000	15:04
Control Test	0.013	15:05
Air Blank	0.000	15:05

Agency: PANAMA CITY BEACH
PD

Date/Time: 12/02/2018 15:10:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/02/2018 18:57

Air Blank	0.000	15:09
Control Test	0.012	15:10
Air Blank	0.000	15:10

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency: BLOUNTSTOWN PD

Date/Time: 10/08/2018 15:06:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/08/2018 18:06

Air Blank	0.000	15:04
Control Test	0.000	15:05
Air Blank	0.000	15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency: HOMESTEAD PD

Date/Time: 04/25/2018 09:30:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/29/2018 10:37

Air Blank	0.000	09:29
Control Test	0.000	09:29
Air Blank	0.000	09:30

Agency: HOMESTEAD PD

Date/Time: 04/25/2018 09:32:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/29/2018 10:37

Air Blank	0.000	09:31
Control Test	0.051	09:32
Air Blank	0.000	09:32

Agency: HOMESTEAD PD

Date/Time: 04/25/2018 09:35:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/29/2018 10:37

Air Blank	0.000	09:33
Control Test	0.081	09:34
Air Blank	0.000	09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency: HOMESTEAD PD

Date/Time: 06/20/2018 09:24:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/20/2018 10:45

Air Blank	0.000	09:22
Control Test	0.047	09:23
Air Blank	0.000	09:24

Agency: HOMESTEAD PD

Date/Time: 07/10/2018 08:29:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2018 09:48

Air Blank	0.000	08:28
Control Test	0.178	08:29
Air Blank	0.000	08:29

Agency: HOMESTEAD PD

Date/Time: 10/23/2018 12:03:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/23/2018 13:20

Air Blank	0.000	12:02
Control Test	0.000	12:03
Air Blank	0.000	12:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency: HOMESTEAD PD

Date/Time: 11/27/2018 13:56:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/27/2018 15:12

Air Blank	0.000	13:54
Control Test	0.050	13:55
Air Blank	0.000	13:56

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency: MIAMI-DADE PD

Date/Time: 03/08/2018 11:18:19

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/22/2018 11:44

Air Blank	0.000	11:17
Control Test	0.080	11:17
Air Blank	0.000	11:18

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency: CHATTAHOOCHEE PD

Date/Time: 12/17/2018 14:07:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/17/2018 15:41

Air Blank	0.000	14:06
Control Test	0.000	14:06
Air Blank	0.000	14:07

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency: FWC

Date/Time: 07/27/2018 15:03:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2018 16:45

Air Blank	0.000	15:02
Control Test	0.000	15:02
Air Blank	0.000	15:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

Agency: COLLIER COUNTY SO

Date/Time: 05/01/2018 07:58:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/01/2018 10:02

Air Blank	0.000	07:57
Control Test	0.000	07:57
Air Blank	0.000	07:58

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

Agency: GLADES COUNTY S.O.

Date/Time: 02/05/2018 16:46:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/05/2018 18:22

Air Blank	0.000	16:45
Control Test	0.075	16:45
Air Blank	0.000	16:46

Agency: GLADES COUNTY S.O.

Date/Time: 02/05/2018 16:48:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/05/2018 18:22

Air Blank	0.000	16:46
Control Test	0.186	16:47
Air Blank	0.000	16:48

Agency: GLADES COUNTY S.O.

Date/Time: 02/28/2018 10:05:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/05/2018 08:12

Air Blank	0.000	10:04
Control Test	0.000	10:05
Air Blank	0.000	10:05

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

Agency: GLADES COUNTY S.O.

Date/Time: 02/28/2018 10:07:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/05/2018 08:12

Air Blank	0.000	10:06
Control Test	0.000	10:07
Air Blank	0.000	10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

Agency: GLADES COUNTY S.O.

Date/Time: 04/30/2018 18:58:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/11/2018 18:23

Air Blank	0.000	18:57
Control Test	0.000	18:58
Air Blank	0.000	18:58

Agency: GLADES COUNTY S.O.

Date/Time: 04/30/2018 19:03:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/11/2018 18:23

Air Blank	0.000	19:02
Control Test	0.028	19:03
Air Blank	0.000	19:03

Agency: GLADES COUNTY S.O.

Date/Time: 05/26/2018 12:38:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/11/2018 18:23

Air Blank	0.000	12:36
Control Test	0.182	12:37
Air Blank	0.000	12:38

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

Agency: GLADES COUNTY S.O.

Date/Time: 05/26/2018 12:40:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/11/2018 18:23

Air Blank	0.000	12:39
Control Test	0.130	12:40
Air Blank	0.000	12:40

Agency: GLADES COUNTY S.O.

Date/Time: 05/26/2018 12:44:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/11/2018 18:23

Air Blank	0.000	12:43
Control Test	0.185	12:43
Air Blank	0.000	12:44

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

Agency: HENDRY COUNTY S.O.

Date/Time: 04/19/2018 01:47:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/09/2018 22:13

Air Blank	0.000	01:46
Control Test	0.051	01:47
Air Blank	0.000	01:47

Agency: HENDRY COUNTY S.O.

Date/Time: 04/19/2018 01:52:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/09/2018 22:13

Air Blank	0.000	01:51
Control Test	0.080	01:52
Air Blank	0.000	01:52

Agency: HENDRY COUNTY S.O.

Date/Time: 06/25/2018 20:30:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/25/2018 22:26

Air Blank	0.000	20:29
Control Test	0.000	20:29
Air Blank	0.000	20:30

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

Agency: HENDRY COUNTY S.O.

Date/Time: 06/25/2018 20:38:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/25/2018 22:26

Air Blank	0.000	20:37
Control Test	0.050	20:37
Air Blank	0.000	20:38

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

Agency: JEFFERSON COUNTY SO

Date/Time: 06/29/2018 08:58:30

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/29/2018 12:15

Air Blank	0.000	08:57
Control Test	0.078	08:58
Air Blank	0.000	08:58

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

Agency: COCOA BEACH P.D.

Date/Time: 10/28/2018 21:13:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/28/2018 23:59

Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 03/09/2018 14:32:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2018 15:40

Air Blank	0.000	14:31
Control Test	0.000	14:31
Air Blank	0.000	14:32

Agency: TITUSVILLE P.D.

Date/Time: 04/30/2018 13:56:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2018 15:58

Air Blank	0.000	13:55
Control Test	0.000	13:56
Air Blank	0.000	13:56

Agency: TITUSVILLE P.D.

Date/Time: 06/29/2018 19:12:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2018 20:18

Air Blank	0.000	19:11
Control Test	0.046	19:12
Air Blank	0.000	19:12

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 06/29/2018 19:15:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2018 20:18

Air Blank	0.000	19:14
Control Test	0.000	19:15
Air Blank	0.000	19:15

Agency: TITUSVILLE P.D.

Date/Time: 06/29/2018 19:18:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2018 20:18

Air Blank	0.000	19:17
Control Test	0.000	19:18
Air Blank	0.000	19:18

Agency: TITUSVILLE P.D.

Date/Time: 06/29/2018 19:21:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2018 20:18

Air Blank	0.000	19:20
Control Test	0.000	19:21
Air Blank	0.000	19:21

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 03/30/2018 12:00:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2018 14:51

Air Blank	0.000	11:59
Control Test	0.193	11:59
Air Blank	0.000	12:00

Agency: INDIALANTIC P.D.

Date/Time: 03/30/2018 12:03:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2018 14:51

Air Blank	0.000	12:01
Control Test	0.049	12:02
Air Blank	0.000	12:03

Agency: INDIALANTIC P.D.

Date/Time: 03/30/2018 12:37:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2018 14:51

Air Blank	0.000	12:36
Control Test	0.049	12:36
Air Blank	0.000	12:37

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 03/30/2018 12:39:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2018 14:51

Air Blank	0.000	12:38
Control Test	0.078	12:39
Air Blank	0.000	12:39

Agency: INDIALANTIC P.D.

Date/Time: 04/08/2018 11:53:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/08/2018 13:34

Air Blank	0.000	11:52
Control Test	0.197	11:53
Air Blank	0.000	11:53

Agency: INDIALANTIC P.D.

Date/Time: 04/08/2018 11:57:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/08/2018 13:34

Air Blank	0.000	11:55
Control Test	0.078	11:56
Air Blank	0.000	11:57

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 04/08/2018 11:59:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/08/2018 13:34

Air Blank	0.000	11:58
Control Test	0.048	11:59
Air Blank	0.000	11:59

Agency: INDIALANTIC P.D.

Date/Time: 05/25/2018 03:09:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/25/2018 07:46

Air Blank	0.000	03:08
Control Test	0.199	03:08
Air Blank	0.000	03:09

Agency: INDIALANTIC P.D.

Date/Time: 05/25/2018 03:11:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/25/2018 07:46

Air Blank	0.000	03:10
Control Test	0.077	03:10
Air Blank	0.000	03:11

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 05/25/2018 03:13:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/25/2018 07:46

Air Blank	0.000	03:12
Control Test	0.048	03:12
Air Blank	0.000	03:13

Agency: INDIALANTIC P.D.

Date/Time: 06/29/2018 22:25:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/30/2018 03:03

Air Blank	0.000	22:24
Control Test	0.197	22:25
Air Blank	0.000	22:25

Agency: INDIALANTIC P.D.

Date/Time: 06/29/2018 22:29:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/30/2018 03:03

Air Blank	0.000	22:27
Control Test	0.077	22:28
Air Blank	0.000	22:29

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 06/29/2018 22:32:39

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/30/2018 03:03

Air Blank	0.000	22:31
Control Test	0.080	22:32
Air Blank	0.000	22:32

Agency: INDIALANTIC P.D.

Date/Time: 06/29/2018 22:34:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/30/2018 03:04

Air Blank	0.000	22:33
Control Test	0.048	22:34
Air Blank	0.000	22:34

Agency: INDIALANTIC P.D.

Date/Time: 06/29/2018 22:40:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/30/2018 03:04

Air Blank	0.000	22:38
Control Test	0.047	22:39
Air Blank	0.000	22:40

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 07/28/2018 04:49:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 07:06

Air Blank	0.000	04:47
Control Test	0.047	04:48
Air Blank	0.000	04:49

Agency: INDIALANTIC P.D.

Date/Time: 07/28/2018 04:53:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 07:06

Air Blank	0.000	04:51
Control Test	0.048	04:52
Air Blank	0.000	04:53

Agency: INDIALANTIC P.D.

Date/Time: 07/28/2018 05:16:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 07:06

Air Blank	0.000	05:15
Control Test	0.048	05:16
Air Blank	0.000	05:16

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 07/28/2018 05:27:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 07:06

Air Blank	0.000	05:26
Control Test	0.197	05:27
Air Blank	0.000	05:27

Agency: INDIALANTIC P.D.

Date/Time: 08/24/2018 00:56:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2018 06:10

Air Blank	0.000	00:55
Control Test	0.199	00:56
Air Blank	0.000	00:56

Agency: INDIALANTIC P.D.

Date/Time: 08/24/2018 00:58:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2018 06:10

Air Blank	0.000	00:57
Control Test	0.078	00:58
Air Blank	0.000	00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 08/24/2018 01:01:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2018 06:10

Air Blank	0.000	00:59
Control Test	0.047	01:00
Air Blank	0.000	01:01

Agency: INDIALANTIC P.D.

Date/Time: 09/29/2018 12:32:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/02/2018 16:50

Air Blank	0.000	12:30
Control Test	0.048	12:31
Air Blank	0.000	12:32

Agency: INDIALANTIC P.D.

Date/Time: 09/29/2018 12:34:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/02/2018 16:50

Air Blank	0.000	12:32
Control Test	0.047	12:33
Air Blank	0.000	12:34

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 10/25/2018 12:00:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/25/2018 13:54

Air Blank	0.000	11:58
Control Test	0.047	11:59
Air Blank	0.000	12:00

Agency: INDIALANTIC P.D.

Date/Time: 10/25/2018 12:02:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/25/2018 13:54

Air Blank	0.000	12:01
Control Test	0.078	12:01
Air Blank	0.000	12:02

Agency: INDIALANTIC P.D.

Date/Time: 11/25/2018 14:14:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/25/2018 16:23

Air Blank	0.000	14:13
Control Test	0.047	14:13
Air Blank	0.000	14:14

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 11/25/2018 14:16:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/25/2018 16:24

Air Blank	0.000	14:15
Control Test	0.047	14:15
Air Blank	0.000	14:16

Agency: INDIALANTIC P.D.

Date/Time: 11/25/2018 14:18:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/25/2018 16:24

Air Blank	0.000	14:17
Control Test	0.078	14:18
Air Blank	0.000	14:18

Agency: INDIALANTIC P.D.

Date/Time: 12/29/2018 13:24:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/31/2018 13:34

Air Blank	0.000	13:23
Control Test	0.047	13:24
Air Blank	0.000	13:24

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

Agency: JACKSONVILLE SO

Date/Time: 01/23/2018 09:49:44

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/01/2018 13:32

Air Blank	0.000	09:48
Control Test	0.078	09:49
Air Blank	0.000	09:49

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 02/01/2018 15:14:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2018 15:54

Air Blank	0.000	15:12
Control Test	0.000	15:13
Air Blank	0.000	15:14

Agency: PINELLAS PARK PD

Date/Time: 04/06/2018 09:55:03

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/06/2018 10:31

Air Blank	0.000	09:54
Control Test	0.078	09:54
Air Blank	0.000	09:55

Agency: PINELLAS PARK PD

Date/Time: 04/06/2018 09:57:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/06/2018 10:31

Air Blank	0.000	09:56
Control Test	0.079	09:56
Air Blank	0.000	09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 04/06/2018 09:59:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/06/2018 10:31

Air Blank	0.000	09:57
Control Test	0.000	09:58
Air Blank	0.000	09:59

Agency: PINELLAS PARK PD

Date/Time: 05/04/2018 11:03:18

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/04/2018 11:39

Air Blank	0.000	11:02
Control Test	0.078	11:02
Air Blank	0.000	11:03

Agency: PINELLAS PARK PD

Date/Time: 05/04/2018 11:05:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2018 11:39

Air Blank	0.000	11:04
Control Test	0.000	11:04
Air Blank	0.000	11:05

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 07/13/2018 10:45:51

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/13/2018 12:11

Air Blank	0.000	10:45
Control Test	0.079	10:45
Air Blank	0.000	10:45

Agency: PINELLAS PARK PD

Date/Time: 07/13/2018 10:47:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2018 12:11

Air Blank	0.000	10:46
Control Test	0.000	10:47
Air Blank	0.000	10:47

Agency: PINELLAS PARK PD

Date/Time: 07/13/2018 10:49:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2018 12:11

Air Blank	0.000	10:48
Control Test	0.028	10:49
Air Blank	0.000	10:49

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 07/13/2018 10:54:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2018 12:11

Air Blank	0.000	10:52
Control Test	0.049	10:53
Air Blank	0.000	10:54

Agency: PINELLAS PARK PD

Date/Time: 07/13/2018 10:56:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2018 12:11

Air Blank	0.000	10:54
Control Test	0.078	10:55
Air Blank	0.000	10:56

Agency: PINELLAS PARK PD

Date/Time: 07/13/2018 10:57:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2018 12:11

Air Blank	0.000	10:56
Control Test	0.198	10:57
Air Blank	0.000	10:57

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 07/13/2018 11:27:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2018 12:11

Air Blank	0.000	11:26
Control Test	0.000	11:26
Air Blank	0.000	11:27

Agency: PINELLAS PARK PD

Date/Time: 08/08/2018 08:36:44

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/08/2018 09:24

Air Blank	0.000	08:35
Control Test	0.079	08:36
Air Blank	0.000	08:36

Agency: PINELLAS PARK PD

Date/Time: 08/08/2018 08:38:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2018 09:24

Air Blank	0.000	08:37
Control Test	0.000	08:38
Air Blank	0.000	08:38

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 08/08/2018 08:40:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2018 09:24

Air Blank	0.000	08:39
Control Test	0.000	08:39
Air Blank	0.000	08:40

Agency: PINELLAS PARK PD

Date/Time: 08/08/2018 08:42:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2018 09:24

Air Blank	0.000	08:40
Control Test	0.049	08:41
Air Blank	0.000	08:42

Agency: PINELLAS PARK PD

Date/Time: 08/08/2018 08:43:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2018 09:24

Air Blank	0.000	08:42
Control Test	0.078	08:43
Air Blank	0.000	08:43

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 08/08/2018 08:45:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2018 09:24

Air Blank	0.000	08:44
Control Test	0.195	08:45
Air Blank	0.000	08:45

Agency: PINELLAS PARK PD

Date/Time: 09/12/2018 09:56:05

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/12/2018 11:17

Air Blank	0.000	09:55
Control Test	0.079	09:55
Air Blank	0.000	09:56

Agency: PINELLAS PARK PD

Date/Time: 09/12/2018 09:58:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2018 11:18

Air Blank	0.000	09:57
Control Test	0.000	09:57
Air Blank	0.000	09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 09/12/2018 10:00:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2018 11:18

Air Blank	0.000	09:58
Control Test	0.000	09:59
Air Blank	0.000	10:00

Agency: PINELLAS PARK PD

Date/Time: 09/12/2018 10:02:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2018 11:18

Air Blank	0.000	10:00
Control Test	0.048	10:01
Air Blank	0.000	10:02

Agency: PINELLAS PARK PD

Date/Time: 09/12/2018 10:04:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2018 11:18

Air Blank	0.000	10:02
Control Test	0.078	10:03
Air Blank	0.000	10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 09/12/2018 10:06:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2018 11:18

Air Blank	0.000	10:04
Control Test	0.198	10:05
Air Blank	0.000	10:06

Agency: PINELLAS PARK PD

Date/Time: 10/09/2018 07:50:10

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/09/2018 08:52

Air Blank	0.000	07:49
Control Test	0.076	07:49
Air Blank	0.000	07:50

Agency: PINELLAS PARK PD

Date/Time: 10/09/2018 07:52:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2018 08:52

Air Blank	0.000	07:50
Control Test	0.000	07:51
Air Blank	0.000	07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 10/09/2018 07:53:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2018 08:52

Air Blank	0.000	07:52
Control Test	0.000	07:53
Air Blank	0.000	07:53

Agency: PINELLAS PARK PD

Date/Time: 10/09/2018 07:55:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2018 08:52

Air Blank	0.000	07:54
Control Test	0.048	07:55
Air Blank	0.000	07:55

Agency: PINELLAS PARK PD

Date/Time: 10/09/2018 07:57:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2018 08:52

Air Blank	0.000	07:56
Control Test	0.077	07:57
Air Blank	0.000	07:57

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 10/09/2018 07:59:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2018 08:52

Air Blank	0.000	07:58
Control Test	0.196	07:59
Air Blank	0.000	07:59

Agency: PINELLAS PARK PD

Date/Time: 11/02/2018 11:26:47

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/02/2018 12:19

Air Blank	0.000	11:25
Control Test	0.079	11:26
Air Blank	0.000	11:26

Agency: PINELLAS PARK PD

Date/Time: 11/02/2018 11:28:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/02/2018 12:19

Air Blank	0.000	11:27
Control Test	0.000	11:28
Air Blank	0.000	11:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 11/02/2018 11:30:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/02/2018 12:19

Air Blank	0.000	11:29
Control Test	0.000	11:30
Air Blank	0.000	11:30

Agency: PINELLAS PARK PD

Date/Time: 11/02/2018 11:32:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/02/2018 12:19

Air Blank	0.000	11:31
Control Test	0.049	11:32
Air Blank	0.000	11:32

Agency: PINELLAS PARK PD

Date/Time: 11/02/2018 11:35:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/02/2018 12:19

Air Blank	0.000	11:34
Control Test	0.077	11:35
Air Blank	0.000	11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 11/02/2018 11:38:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/02/2018 12:19

Air Blank	0.000	11:36
Control Test	0.196	11:37
Air Blank	0.000	11:38

Agency: PINELLAS PARK PD

Date/Time: 12/20/2018 07:53:48

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/20/2018 08:41

Air Blank	0.000	07:52
Control Test	0.078	07:53
Air Blank	0.000	07:53

Agency: PINELLAS PARK PD

Date/Time: 12/20/2018 07:57:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 08:41

Air Blank	0.000	07:56
Control Test	0.000	07:57
Air Blank	0.000	07:57

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 12/20/2018 07:59:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 08:41

Air Blank	0.000	07:58
Control Test	0.000	07:59
Air Blank	0.000	07:59

Agency: PINELLAS PARK PD

Date/Time: 12/20/2018 08:01:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 08:41

Air Blank	0.000	08:00
Control Test	0.049	08:01
Air Blank	0.000	08:01

Agency: PINELLAS PARK PD

Date/Time: 12/20/2018 08:03:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 08:41

Air Blank	0.000	08:02
Control Test	0.077	08:02
Air Blank	0.000	08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 12/20/2018 08:05:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 08:41

Air Blank	0.000	08:04
Control Test	0.196	08:04
Air Blank	0.000	08:05

Agency: PINELLAS PARK PD

Date/Time: 01/04/2019 09:05:20

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/04/2019 09:58

Air Blank	0.000	09:04
Control Test	0.078	09:04
Air Blank	0.000	09:05

Agency: PINELLAS PARK PD

Date/Time: 01/04/2019 09:07:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2019 09:58

Air Blank	0.000	09:06
Control Test	0.000	09:06
Air Blank	0.000	09:07

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 01/04/2019 09:09:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2019 09:58

Air Blank	0.000	09:08
Control Test	0.000	09:08
Air Blank	0.000	09:09

Agency: PINELLAS PARK PD

Date/Time: 01/04/2019 09:11:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2019 09:58

Air Blank	0.000	09:10
Control Test	0.049	09:10
Air Blank	0.000	09:11

Agency: PINELLAS PARK PD

Date/Time: 01/04/2019 09:13:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2019 09:58

Air Blank	0.000	09:12
Control Test	0.078	09:12
Air Blank	0.000	09:13

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 01/04/2019 09:15:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2019 09:58

Air Blank	0.000	09:14
Control Test	0.198	09:14
Air Blank	0.000	09:15

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/27/2018 03:01:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2018 04:46

Air Blank	0.000	03:00
Control Test	0.193	03:01
Air Blank	0.000	03:01

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/27/2018 03:04:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2018 04:46

Air Blank	0.000	03:03
Control Test	0.046	03:04
Air Blank	0.000	03:04

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/27/2018 03:07:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2018 04:46

Air Blank	0.000	03:05
Control Test	0.000	03:06
Air Blank	0.000	03:07

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/27/2018 03:09:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2018 04:46

Air Blank	0.000	03:08
Control Test	0.074	03:08
Air Blank	0.000	03:09

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 21:03:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 21:46

Air Blank	0.000	21:02
Control Test	0.194	21:02
Air Blank	0.000	21:03

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 21:05:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 21:46

Air Blank	0.000	21:04
Control Test	0.076	21:05
Air Blank	0.000	21:05

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 21:07:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 21:46

Air Blank	0.000	21:06
Control Test	0.048	21:07
Air Blank	0.000	21:07

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 21:10:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 21:46

Air Blank	0.000	21:08
Control Test	0.000	21:09
Air Blank	0.000	21:10

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/26/2018 21:52:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/26/2018 23:15

Air Blank	0.000	21:51
Control Test	0.197	21:51
Air Blank	0.000	21:52

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/26/2018 21:54:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/26/2018 23:15

Air Blank	0.000	21:53
Control Test	0.000	21:54
Air Blank	0.000	21:54

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/26/2018 21:56:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/26/2018 23:15

Air Blank	0.000	21:55
Control Test	0.048	21:56
Air Blank	0.000	21:56

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/26/2018 21:59:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/26/2018 23:15

Air Blank	0.000	21:58
Control Test	0.078	21:58
Air Blank	0.000	21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/26/2018 21:59:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/26/2018 23:36

Air Blank	0.000	21:58
Control Test	0.197	21:59
Air Blank	0.000	21:59

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/26/2018 22:02:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/26/2018 23:36

Air Blank	0.000	22:00
Control Test	0.077	22:01
Air Blank	0.000	22:02

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/26/2018 22:04:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/26/2018 23:36

Air Blank	0.000	22:03
Control Test	0.047	22:03
Air Blank	0.000	22:04

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/26/2018 22:06:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/26/2018 23:37

Air Blank	0.000	22:05
Control Test	0.000	22:06
Air Blank	0.000	22:06

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

Agency: LAKE ALFRED PD

Date/Time: 01/31/2018 09:57:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2018 13:13

Air Blank	0.000	09:55
Control Test	0.000	09:56
Air Blank	0.000	09:57

Agency: LAKE ALFRED PD

Date/Time: 07/19/2018 08:55:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/19/2018 10:10

Air Blank	0.000	08:54
Control Test	0.048	08:54
Air Blank	0.000	08:55

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 02/17/2018 06:47:43

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/21/2018 11:51

Air Blank	0.000	06:46
Control Test	0.079	06:47
Air Blank	0.000	06:47

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/27/2018 02:49:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2018 03:14

Air Blank	0.000	02:47
Control Test	0.195	02:48
Air Blank	0.000	02:49

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/27/2018 02:57:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2018 03:14

Air Blank	0.000	02:56
Control Test	0.046	02:56
Air Blank	0.000	02:57

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/27/2018 02:59:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2018 03:14

Air Blank	0.000	02:58
Control Test	0.076	02:59
Air Blank	0.000	02:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/27/2018 03:02:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2018 03:14

Air Blank	0.000	03:01
Control Test	0.000	03:01
Air Blank	0.000	03:02

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/24/2018 00:39:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/24/2018 01:31

Air Blank	0.000	00:37
Control Test	0.078	00:38
Air Blank	0.000	00:39

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/24/2018 00:41:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/24/2018 01:31

Air Blank	0.000	00:40
Control Test	0.195	00:40
Air Blank	0.000	00:41

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/24/2018 00:43:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/24/2018 01:31

Air Blank	0.000	00:42
Control Test	0.000	00:43
Air Blank	0.000	00:43

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/24/2018 00:46:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/24/2018 01:31

Air Blank	0.000	00:44
Control Test	0.048	00:45
Air Blank	0.000	00:46

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/24/2018 00:47:32

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/24/2018 01:31

Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/11/2018 00:37:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2018 01:33

Air Blank	0.000	00:36
Control Test	0.199	00:37
Air Blank	0.000	00:37

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/11/2018 00:40:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2018 01:33

Air Blank	0.000	00:39
Control Test	0.049	00:39
Air Blank	0.000	00:40

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/11/2018 00:42:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2018 01:33

Air Blank	0.000	00:41
Control Test	0.077	00:42
Air Blank	0.000	00:42

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/11/2018 00:45:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2018 01:33

Air Blank	0.000	00:44
Control Test	0.000	00:44
Air Blank	0.000	00:45

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/11/2018 00:49:47

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/11/2018 01:33

Air Blank	0.000	00:48
Control Test	0.079	00:49
Air Blank	0.000	00:49

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 20:48:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 22:26

Air Blank	0.000	20:47
Control Test	0.194	20:48
Air Blank	0.000	20:48

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 20:51:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 22:26

Air Blank	0.000	20:50
Control Test	0.075	20:50
Air Blank	0.000	20:51

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 20:53:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 22:26

Air Blank	0.000	20:52
Control Test	0.000	20:53
Air Blank	0.000	20:53

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 20:55:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 22:26

Air Blank	0.000	20:54
Control Test	0.000	20:55
Air Blank	0.000	20:55

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/28/2018 20:57:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/28/2018 22:26

Air Blank	0.000	20:56
Control Test	0.048	20:57
Air Blank	0.000	20:57

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/26/2018 21:49:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/26/2018 22:37

Air Blank	0.000	21:48
Control Test	0.198	21:49
Air Blank	0.000	21:49

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/26/2018 21:52:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/26/2018 22:37

Air Blank	0.000	21:51
Control Test	0.000	21:52
Air Blank	0.000	21:52

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/26/2018 21:55:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/26/2018 22:37

Air Blank	0.000	21:53
Control Test	0.047	21:54
Air Blank	0.000	21:55

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/26/2018 21:57:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/26/2018 22:37

Air Blank	0.000	21:56
Control Test	0.077	21:56
Air Blank	0.000	21:57

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/26/2018 21:57:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/26/2018 22:48

Air Blank	0.000	21:56
Control Test	0.195	21:57
Air Blank	0.000	21:57

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/26/2018 22:00:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/26/2018 22:48

Air Blank	0.000	21:59
Control Test	0.076	21:59
Air Blank	0.000	22:00

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/26/2018 22:02:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/26/2018 22:48

Air Blank	0.000	22:01
Control Test	0.046	22:02
Air Blank	0.000	22:02

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/26/2018 22:04:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/26/2018 22:48

Air Blank	0.000	22:03
Control Test	0.000	22:04
Air Blank	0.000	22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 01/20/2018 18:45:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/20/2018 19:41

Air Blank	0.000	18:44
Control Test	0.195	18:44
Air Blank	0.000	18:45

Agency: POLK COUNTY SO

Date/Time: 01/20/2018 18:47:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/20/2018 19:41

Air Blank	0.000	18:45
Control Test	0.077	18:46
Air Blank	0.000	18:47

Agency: POLK COUNTY SO

Date/Time: 02/28/2018 12:50:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2018 13:46

Air Blank	0.000	12:48
Control Test	0.000	12:49
Air Blank	0.000	12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 02/28/2018 12:52:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2018 13:46

Air Blank	0.000	12:50
Control Test	0.079	12:51
Air Blank	0.000	12:52

Agency: POLK COUNTY SO

Date/Time: 03/26/2018 10:21:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/26/2018 11:28

Air Blank	0.000	10:19
Control Test	0.078	10:20
Air Blank	0.000	10:21

Agency: POLK COUNTY SO

Date/Time: 03/26/2018 10:22:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/26/2018 11:28

Air Blank	0.000	10:21
Control Test	0.196	10:22
Air Blank	0.000	10:22

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 04/30/2018 10:02:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2018 11:26

Air Blank	0.000	10:01
Control Test	0.197	10:01
Air Blank	0.000	10:02

Agency: POLK COUNTY SO

Date/Time: 04/30/2018 10:04:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2018 11:26

Air Blank	0.000	10:02
Control Test	0.048	10:03
Air Blank	0.000	10:04

Agency: POLK COUNTY SO

Date/Time: 06/27/2018 10:23:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 14:56

Air Blank	0.000	10:22
Control Test	0.197	10:23
Air Blank	0.000	10:23

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 06/27/2018 10:36:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 14:56

Air Blank	0.000	10:35
Control Test	0.029	10:36
Air Blank	0.000	10:36

Agency: POLK COUNTY SO

Date/Time: 06/27/2018 12:50:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 14:56

Air Blank	0.000	12:49
Control Test	0.000	12:50
Air Blank	0.000	12:50

Agency: POLK COUNTY SO

Date/Time: 07/27/2018 06:36:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2018 09:05

Air Blank	0.000	06:35
Control Test	0.050	06:36
Air Blank	0.000	06:36

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 09:40:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 14:15

Air Blank	0.000	09:39
Control Test	0.000	09:40
Air Blank	0.000	09:40

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 11:05:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 14:15

Air Blank	0.000	11:04
Control Test	0.078	11:05
Air Blank	0.000	11:05

Agency: POLK COUNTY SO

Date/Time: 09/28/2018 09:45:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/28/2018 11:49

Air Blank	0.000	09:44
Control Test	0.079	09:45
Air Blank	0.000	09:45

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 10/27/2018 10:58:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 13:51

Air Blank	0.000	10:57
Control Test	0.079	10:57
Air Blank	0.000	10:58

Agency: POLK COUNTY SO

Date/Time: 11/28/2018 06:16:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2018 09:00

Air Blank	0.000	06:15
Control Test	0.196	06:16
Air Blank	0.000	06:16

Agency: POLK COUNTY SO

Date/Time: 12/28/2018 06:48:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/28/2018 09:00

Air Blank	0.000	06:46
Control Test	0.000	06:47
Air Blank	0.000	06:48

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency: GULFPORT PD

Date/Time: 08/31/2018 01:33:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 03:17

Air Blank	0.000	01:32
Control Test	0.059	01:33
Air Blank	0.000	01:33

Agency: GULFPORT PD

Date/Time: 08/31/2018 01:57:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 03:17

Air Blank	0.000	01:56
Control Test	0.078	01:57
Air Blank	0.000	01:57

Agency: GULFPORT PD

Date/Time: 12/30/2018 20:57:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/30/2018 22:19

Air Blank	0.000	20:56
Control Test	0.189	20:56
Air Blank	0.000	20:57

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency: GULFPORT PD

Date/Time: 12/30/2018 20:59:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/30/2018 22:19

Air Blank	0.000	20:58
Control Test	0.190	20:58
Air Blank	0.000	20:59

Agency: GULFPORT PD

Date/Time: 12/30/2018 21:01:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/30/2018 22:19

Air Blank	0.000	20:59
Control Test	0.189	21:00
Air Blank	0.000	21:01

Agency: GULFPORT PD

Date/Time: 12/30/2018 21:01:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/30/2018 22:20

Air Blank		21:01
Air Blank	0.000	21:01

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency: GULFPORT PD

Date/Time: 12/30/2018 21:05:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/30/2018 22:20

Air Blank	0.000	21:03
Control Test	0.189	21:04
Air Blank	0.000	21:05

Agency: GULFPORT PD

Date/Time: 01/01/2019 17:11:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/07/2019 12:48

Air Blank	0.000	17:10
Control Test	0.074	17:11
Air Blank	0.000	17:11

Agency: GULFPORT PD

Date/Time: 01/01/2019 17:13:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/07/2019 12:48

Air Blank	0.000	17:12
Control Test	0.193	17:13
Air Blank	0.000	17:13

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency: GULFPORT PD

Date/Time: 01/11/2019 03:14:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2019 04:04

Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:14

Agency: GULFPORT PD

Date/Time: 01/11/2019 03:16:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2019 04:04

Air Blank	0.000	03:15
Control Test	0.205	03:16
Air Blank	0.000	03:16

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency: HOLLYWOOD PD

Date/Time: 02/26/2018 10:46:48

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/26/2018 11:33

Air Blank	0.000	10:46
Control Test	0.075	10:46
Air Blank	0.000	10:46

Agency: HOLLYWOOD PD

Date/Time: 02/26/2018 10:48:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/26/2018 11:33

Air Blank	0.000	10:47
Control Test	0.076	10:47
Air Blank	0.000	10:48

Agency: HOLLYWOOD PD

Date/Time: 03/30/2018 10:22:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2018 11:23

Air Blank	0.000	10:20
Control Test	0.000	10:21
Air Blank	0.000	10:22

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency: HOLLYWOOD PD

Date/Time: 06/26/2018 09:53:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 11:09

Air Blank	0.000	09:51
Control Test	0.000	09:52
Air Blank	0.000	09:53

Agency: HOLLYWOOD PD

Date/Time: 10/29/2018 19:54:57

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/29/2018 20:51

Air Blank	0.000	19:54
Control Test	0.079	19:54
Air Blank	0.000	19:54

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency: TARPON SPRINGS PD

Date/Time: 08/17/2018 17:32:20

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/20/2018 14:27

Air Blank	0.000	17:31
Control Test	0.075	17:31
Air Blank	0.000	17:32

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency: PINELLAS COUNTY SO

Date/Time: 02/28/2018 21:25:59

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/22/2018 20:45

Air Blank	0.000	21:25
ITP Check	0.082	21:25
Air Blank	0.000	21:25

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency: BROWARD COUNTY SO

Date/Time: 07/13/2018 10:22:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2018 11:26

Air Blank	0.000	10:21
Control Test	0.000	10:22
Air Blank	0.000	10:22

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 02/24/2018 21:27:05

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/14/2018 17:13

Air Blank	0.000	21:26
Control Test	0.072	21:26
Air Blank	0.000	21:27

Agency: CLEARWATER PD

Date/Time: 02/24/2018 21:28:52

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/14/2018 17:13

Air Blank	0.000	21:27
Control Test	0.078	21:28
Air Blank	0.000	21:28

Agency: CLEARWATER PD

Date/Time: 03/02/2018 21:05:53

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/14/2018 17:13

Air Blank	0.000	21:05
Control Test	0.078	21:05
Air Blank	0.000	21:05

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency: HARDEE COUNTY SO

Date/Time: 06/26/2018 15:53:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/26/2018 16:59

Air Blank	0.000	15:52
Control Test	0.075	15:52
Air Blank	0.000	15:53

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 01/22/2018 11:02:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/22/2018 11:42

Air Blank	0.000	11:02
Control Test	0.080	11:02
Air Blank	0.000	11:02

Agency: LARGO PD

Date/Time: 01/22/2018 11:07:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/22/2018 11:42

Air Blank	0.000	11:06
Control Test	0.080	11:06
Air Blank	0.000	11:07

Agency: LARGO PD

Date/Time: 03/19/2018 19:41:40

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/19/2018 20:15

Air Blank	0.000	19:40
Control Test	0.080	19:41
Air Blank	0.000	19:41

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 08/20/2018 10:51:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/06/2018 14:20

Air Blank	0.000	10:51
Control Test	0.080	10:51
Air Blank	0.000	10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 03/26/2018 10:21:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/26/2018 11:32

Air Blank	0.000	10:20
Control Test	0.048	10:21
Air Blank	0.000	10:21

Agency: POLK COUNTY SO

Date/Time: 03/26/2018 10:23:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/26/2018 11:32

Air Blank	0.000	10:22
Control Test	0.000	10:23
Air Blank	0.000	10:23

Agency: POLK COUNTY SO

Date/Time: 04/30/2018 10:02:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2018 11:20

Air Blank	0.000	10:00
Control Test	0.000	10:01
Air Blank	0.000	10:02

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 04/30/2018 10:04:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2018 11:20

Air Blank	0.000	10:02
Control Test	0.080	10:03
Air Blank	0.000	10:04

Agency: POLK COUNTY SO

Date/Time: 05/22/2018 14:29:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/22/2018 16:44

Air Blank	0.000	14:28
Control Test	0.079	14:29
Air Blank	0.000	14:29

Agency: POLK COUNTY SO

Date/Time: 05/22/2018 14:31:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/22/2018 16:44

Air Blank	0.000	14:30
Control Test	0.048	14:31
Air Blank	0.000	14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 06/27/2018 10:20:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 15:33

Air Blank	0.000	10:19
Control Test	0.000	10:20
Air Blank	0.000	10:20

Agency: POLK COUNTY SO

Date/Time: 06/27/2018 10:27:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 15:33

Air Blank	0.000	10:25
Control Test	0.028	10:26
Air Blank		10:26
Air Blank		10:27

Agency: POLK COUNTY SO

Date/Time: 06/27/2018 11:42:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 15:33

Air Blank	0.000	11:40
Control Test	0.014	11:41
Air Blank		11:41
Air Blank	0.000	11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 07/27/2018 06:27:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2018 10:08

Air Blank	0.000	06:26
Control Test	0.000	06:27
Air Blank	0.000	06:27

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 09:48:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 13:52

Air Blank	0.000	09:46
Control Test	0.196	09:47
Air Blank		09:47
Air Blank		09:48

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 09:58:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 13:52

Air Blank	0.000	09:56
Control Test	0.041	09:56
Air Blank		09:57
Air Blank		09:58

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 10:32:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 13:52

Air Blank	0.000	10:31
Control Test	0.000	10:31
Air Blank	0.000	10:32

Agency: POLK COUNTY SO

Date/Time: 09/28/2018 09:45:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/28/2018 11:48

Air Blank	0.000	09:43
Control Test	0.000	09:44
Air Blank	0.000	09:45

Agency: POLK COUNTY SO

Date/Time: 10/27/2018 11:01:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 13:24

Air Blank	0.000	10:59
Control Test	0.199	11:00
Air Blank	0.000	11:01

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 11/28/2018 06:14:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2018 09:07

Air Blank	0.000	06:13
Control Test	0.000	06:14
Air Blank	0.000	06:14

Agency: POLK COUNTY SO

Date/Time: 12/28/2018 06:50:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/28/2018 08:46

Air Blank	0.000	06:49
Control Test	0.079	06:50
Air Blank	0.000	06:50

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 01/25/2018 14:05:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/25/2018 15:02

Air Blank	0.000	14:04
Control Test	0.076	14:04
Air Blank	0.000	14:05

Agency: PLANTATION PD

Date/Time: 01/25/2018 14:07:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2018 15:02

Air Blank	0.000	14:06
Control Test	0.049	14:07
Air Blank	0.000	14:07

Agency: PLANTATION PD

Date/Time: 01/25/2018 14:09:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2018 15:02

Air Blank	0.000	14:08
Control Test	0.079	14:09
Air Blank	0.000	14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 01/25/2018 14:14:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2018 15:02

Air Blank	0.000	14:13
Control Test	0.198	14:13
Air Blank	0.000	14:14

Agency: PLANTATION PD

Date/Time: 02/24/2018 13:37:01

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/24/2018 14:28

Air Blank	0.000	13:36
Control Test	0.077	13:36
Air Blank	0.000	13:37

Agency: PLANTATION PD

Date/Time: 02/24/2018 13:39:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/24/2018 14:28

Air Blank	0.000	13:38
Control Test	0.050	13:38
Air Blank	0.000	13:39

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 02/24/2018 13:41:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/24/2018 14:28

Air Blank	0.000	13:40
Control Test	0.079	13:40
Air Blank	0.000	13:41

Agency: PLANTATION PD

Date/Time: 02/24/2018 13:43:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/24/2018 14:28

Air Blank	0.000	13:42
Control Test	0.199	13:43
Air Blank	0.000	13:43

Agency: PLANTATION PD

Date/Time: 03/22/2018 18:59:15

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/22/2018 19:47

Air Blank	0.000	18:58
Control Test	0.077	18:58
Air Blank	0.000	18:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 03/22/2018 19:01:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/22/2018 19:47

Air Blank	0.000	19:00
Control Test	0.051	19:00
Air Blank	0.000	19:01

Agency: PLANTATION PD

Date/Time: 03/22/2018 19:03:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/22/2018 19:47

Air Blank	0.000	19:02
Control Test	0.080	19:02
Air Blank	0.000	19:03

Agency: PLANTATION PD

Date/Time: 03/22/2018 19:04:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/22/2018 19:47

Air Blank		19:03
Air Blank		19:04

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 03/22/2018 19:06:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/22/2018 19:47

Air Blank	0.000	19:05
Control Test	0.198	19:06
Air Blank	0.000	19:06

Agency: PLANTATION PD

Date/Time: 04/19/2018 10:22:25

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/19/2018 11:10

Air Blank	0.000	10:21
Control Test	0.077	10:21
Air Blank	0.000	10:22

Agency: PLANTATION PD

Date/Time: 04/19/2018 10:24:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 11:10

Air Blank	0.000	10:23
Control Test	0.051	10:24
Air Blank	0.000	10:24

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 04/19/2018 10:26:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 11:10

Air Blank	0.000	10:25
Control Test	0.079	10:26
Air Blank	0.000	10:26

Agency: PLANTATION PD

Date/Time: 04/19/2018 10:28:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 11:10

Air Blank	0.000	10:27
Control Test	0.197	10:28
Air Blank	0.000	10:28

Agency: PLANTATION PD

Date/Time: 05/26/2018 13:30:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/26/2018 23:46

Air Blank	0.000	13:29
Control Test	0.075	13:29
Air Blank	0.000	13:30

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 05/26/2018 13:32:07

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/26/2018 23:46

Air Blank	0.000	13:31
Control Test	0.075	13:31
Air Blank	0.000	13:32

Agency: PLANTATION PD

Date/Time: 05/26/2018 13:34:27

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/26/2018 23:46

Air Blank	0.000	13:33
Control Test	0.075	13:33
Air Blank	0.000	13:34

Agency: PLANTATION PD

Date/Time: 05/26/2018 13:37:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/26/2018 23:46

Air Blank	0.000	13:36
Control Test	0.050	13:36
Air Blank	0.000	13:37

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 05/26/2018 13:39:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/26/2018 23:46

Air Blank	0.000	13:38
Control Test	0.081	13:38
Air Blank	0.000	13:39

Agency: PLANTATION PD

Date/Time: 05/26/2018 13:43:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/26/2018 23:46

Air Blank	0.000	13:41
Control Test	0.200	13:42
Air Blank	0.000	13:43

Agency: PLANTATION PD

Date/Time: 05/26/2018 13:45:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/26/2018 23:46

Air Blank	0.000	13:43
Control Test	0.081	13:44
Air Blank	0.000	13:45

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 06/23/2018 13:50:59

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/23/2018 14:44

Air Blank	0.000	13:50
Control Test	0.079	13:50
Air Blank	0.000	13:50

Agency: PLANTATION PD

Date/Time: 06/23/2018 13:53:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/23/2018 14:44

Air Blank	0.000	13:52
Control Test	0.051	13:52
Air Blank	0.000	13:53

Agency: PLANTATION PD

Date/Time: 06/23/2018 13:57:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/23/2018 14:44

Air Blank	0.000	13:55
Control Test	0.079	13:56
Air Blank	0.000	13:57

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 06/23/2018 13:59:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/23/2018 14:44

Air Blank	0.000	13:58
Control Test	0.199	13:59
Air Blank	0.000	13:59

Agency: PLANTATION PD

Date/Time: 07/26/2018 14:11:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/26/2018 15:06

Air Blank	0.000	14:10
Control Test	0.079	14:10
Air Blank	0.000	14:11

Agency: PLANTATION PD

Date/Time: 07/26/2018 14:15:05

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/26/2018 15:06

Air Blank	0.000	14:14
Control Test	0.000	14:14
Air Blank	0.000	14:15

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 07/26/2018 14:16:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2018 15:06

Air Blank	0.000	14:15
Control Test	0.050	14:16
Air Blank	0.000	14:16

Agency: PLANTATION PD

Date/Time: 07/26/2018 14:18:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2018 15:06

Air Blank	0.000	14:17
Control Test	0.079	14:18
Air Blank	0.000	14:18

Agency: PLANTATION PD

Date/Time: 07/26/2018 14:21:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2018 15:06

Air Blank	0.000	14:20
Control Test	0.197	14:20
Air Blank	0.000	14:21

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 08/29/2018 19:27:28

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/29/2018 20:26

Air Blank	0.000	19:26
Control Test	0.079	19:26
Air Blank	0.000	19:27

Agency: PLANTATION PD

Date/Time: 08/29/2018 19:30:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 20:26

Air Blank	0.000	19:29
Control Test	0.049	19:30
Air Blank	0.000	19:30

Agency: PLANTATION PD

Date/Time: 08/29/2018 19:33:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 20:26

Air Blank	0.000	19:32
Control Test	0.081	19:33
Air Blank	0.000	19:33

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 08/29/2018 19:36:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 20:26

Air Blank	0.000	19:35
Control Test	0.196	19:36
Air Blank	0.000	19:36

Agency: PLANTATION PD

Date/Time: 09/20/2018 20:25:29

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/20/2018 21:15

Air Blank	0.000	20:24
Control Test	0.078	20:24
Air Blank	0.000	20:25

Agency: PLANTATION PD

Date/Time: 09/20/2018 20:28:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/20/2018 21:15

Air Blank	0.000	20:27
Control Test	0.000	20:27
Air Blank	0.000	20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 10/27/2018 13:38:25

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/27/2018 14:38

Air Blank	0.000	13:37
Control Test	0.079	13:37
Air Blank	0.000	13:38

Agency: PLANTATION PD

Date/Time: 10/27/2018 13:40:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 14:38

Air Blank	0.000	13:39
Control Test	0.000	13:39
Air Blank	0.000	13:40

Agency: PLANTATION PD

Date/Time: 10/27/2018 13:42:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 14:38

Air Blank	0.000	13:41
Control Test	0.049	13:41
Air Blank	0.000	13:42

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 10/27/2018 13:43:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 14:38

Air Blank		13:42
Air Blank	0.000	13:43

Agency: PLANTATION PD

Date/Time: 10/27/2018 13:44:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 14:38

Air Blank		13:43
Air Blank	0.000	13:44

Agency: PLANTATION PD

Date/Time: 10/27/2018 13:47:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 14:38

Air Blank	0.000	13:45
Control Test	0.079	13:46
Air Blank	0.000	13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 10/27/2018 13:49:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 14:38

Air Blank	0.000	13:48
Control Test	0.195	13:48
Air Blank	0.000	13:49

Agency: PLANTATION PD

Date/Time: 11/24/2018 13:31:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/24/2018 14:28

Air Blank	0.000	13:30
Control Test	0.078	13:30
Air Blank	0.000	13:31

Agency: PLANTATION PD

Date/Time: 11/24/2018 13:33:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/24/2018 14:28

Air Blank	0.000	13:32
Control Test	0.049	13:33
Air Blank	0.000	13:33

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 11/24/2018 13:35:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/24/2018 14:28

Air Blank	0.000	13:34
Control Test	0.079	13:35
Air Blank	0.000	13:35

Agency: PLANTATION PD

Date/Time: 11/24/2018 13:38:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/24/2018 14:28

Air Blank	0.000	13:37
Control Test	0.195	13:37
Air Blank	0.000	13:38

Agency: PLANTATION PD

Date/Time: 12/26/2018 13:52:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/26/2018 14:51

Air Blank	0.000	13:51
Control Test	0.049	13:51
Air Blank	0.000	13:52

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 12/26/2018 13:54:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/26/2018 14:51

Air Blank	0.000	13:53
Control Test	0.079	13:54
Air Blank	0.000	13:54

Agency: PLANTATION PD

Date/Time: 12/26/2018 13:57:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/26/2018 14:51

Air Blank	0.000	13:55
Control Test	0.195	13:56
Air Blank	0.000	13:57

Agency: PLANTATION PD

Date/Time: 12/26/2018 13:59:37

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/26/2018 14:51

Air Blank	0.000	13:58
Control Test	0.078	13:59
Air Blank	0.000	13:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency: JACKSONVILLE SO

Date/Time: 01/05/2018 14:42:00

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/19/2018 14:27

Air Blank	0.000	14:41
Control Test	0.083	14:41
Air Blank	0.000	14:42

Agency: JACKSONVILLE SO

Date/Time: 03/23/2018 14:52:52

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/02/2018 14:40

Air Blank	0.000	14:52
Control Test	0.081	14:52
Air Blank	0.000	14:52

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FL HIGHWAY PATROL

Date/Time: 06/19/2018 14:54:19

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/06/2018 19:04

Air Blank	0.000	14:53
ITP Check	0.081	14:53
Air Blank	0.000	14:54

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency: DELAND P.D.

Date/Time: 06/25/2018 11:45:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/26/2018 13:40

Air Blank	0.000	11:45
Control Test	0.076	11:45
Air Blank	0.000	11:45

Agency: DELAND P.D.

Date/Time: 06/25/2018 11:48:04

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/26/2018 13:40

Air Blank	0.000	11:47
Control Test	0.076	11:47
Air Blank	0.000	11:48

Agency: DELAND P.D.

Date/Time: 06/25/2018 11:49:42

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/26/2018 13:40

Air Blank	0.000	11:48
Control Test	0.077	11:49
Air Blank	0.000	11:49

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency: DELAND P.D.

Date/Time: 06/25/2018 12:05:59

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/26/2018 13:40

Air Blank	0.000	12:05
Control Test	0.076	12:05
Air Blank	0.000	12:05

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 18:40:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	18:39
Control Test	0.000	18:40
Air Blank	0.000	18:40

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 18:42:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	18:41
Control Test	0.000	18:42
Air Blank	0.000	18:42

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 18:45:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	18:44
Control Test	0.050	18:44
Air Blank	0.000	18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 18:47:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	18:45
Control Test	0.000	18:46
Air Blank	0.000	18:47

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:04:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:03
Control Test	0.000	19:03
Air Blank	0.000	19:04

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:08:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:07
Control Test	0.041	19:08
Air Blank	0.000	19:08

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:18:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:17
Control Test	0.045	19:18
Air Blank	0.000	19:18

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:20:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:19
Control Test	0.044	19:20
Air Blank	0.000	19:20

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:22:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:21
Control Test	0.000	19:22
Air Blank	0.000	19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:25:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:24
Control Test	0.044	19:25
Air Blank	0.000	19:25

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:44:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:43
Control Test	0.000	19:44
Air Blank	0.000	19:44

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:46:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:45
Control Test	0.051	19:45
Air Blank	0.000	19:46

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:47:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank		19:46
Air Blank	0.000	19:47

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:48:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:47
Control Test	0.051	19:48
Air Blank	0.000	19:48

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:49:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank		19:49
Air Blank	0.000	19:49

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:50:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank		19:50
Air Blank	0.000	19:50

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:51:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank		19:51
Air Blank	0.000	19:51

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:53:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:52
Control Test	0.051	19:53
Air Blank	0.000	19:53

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 19:57:43

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank		19:57
Air Blank	0.000	19:57

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 20:00:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	19:59
Control Test	0.047	19:59
Air Blank	0.000	20:00

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 20:54:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	20:53
Control Test	0.077	20:54
Air Blank	0.000	20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2018 20:57:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/03/2018 19:17

Air Blank	0.000	20:55
Control Test	0.080	20:56
Air Blank	0.000	20:57

Agency: FL HIGHWAY PATROL

Date/Time: 07/06/2018 17:24:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/06/2018 18:19

Air Blank	0.000	17:23
Control Test	0.048	17:24
Air Blank	0.000	17:24

Agency: FL HIGHWAY PATROL

Date/Time: 12/24/2018 09:21:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/24/2018 10:31

Air Blank	0.000	09:20
Control Test	0.081	09:20
Air Blank	0.000	09:21

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP D ORLANDO

Date/Time: 05/25/2018 00:59:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2018 22:18

Air Blank	0.000	00:58
Control Test	0.048	00:59
Air Blank	0.000	00:59

Agency: FHP TROOP D ORLANDO

Date/Time: 05/25/2018 01:01:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2018 22:18

Air Blank	0.000	01:00
Control Test	0.188	01:01
Air Blank	0.000	01:01

Agency: FHP TROOP D ORLANDO

Date/Time: 05/25/2018 01:03:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2018 22:18

Air Blank	0.000	01:02
Control Test	0.190	01:02
Air Blank	0.000	01:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP D ORLANDO

Date/Time: 05/25/2018 01:05:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2018 22:18

Air Blank	0.000	01:04
Control Test	0.077	01:05
Air Blank	0.000	01:05

Agency: FHP TROOP D ORLANDO

Date/Time: 05/25/2018 01:12:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2018 22:18

Air Blank	0.000	01:11
Control Test	0.186	01:12
Air Blank	0.000	01:12

Agency: FHP TROOP D ORLANDO

Date/Time: 05/25/2018 01:16:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2018 22:18

Air Blank	0.000	01:15
Control Test	0.190	01:15
Air Blank	0.000	01:16

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP D ORLANDO

Date/Time: 05/25/2018 01:20:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2018 22:18

Air Blank	0.000	01:19
Control Test	0.192	01:20
Air Blank	0.000	01:20

Agency: FHP TROOP D ORLANDO

Date/Time: 05/25/2018 01:26:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2018 22:18

Air Blank	0.000	01:24
Control Test	0.194	01:25
Air Blank	0.000	01:26

Agency: FHP TROOP D ORLANDO

Date/Time: 09/25/2018 14:26:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2018 16:21

Air Blank	0.000	14:25
Control Test	0.045	14:25
Air Blank	0.000	14:26

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP D ORLANDO

Date/Time: 09/25/2018 14:28:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2018 16:21

Air Blank	0.000	14:27
Control Test	0.191	14:27
Air Blank	0.000	14:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 05/22/2018 14:26:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/22/2018 16:49

Air Blank	0.000	14:25
Control Test	0.000	14:25
Air Blank	0.000	14:26

Agency: POLK COUNTY SO

Date/Time: 05/22/2018 14:28:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/22/2018 16:49

Air Blank	0.000	14:27
Control Test	0.180	14:28
Air Blank	0.000	14:28

Agency: POLK COUNTY SO

Date/Time: 06/27/2018 11:41:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 14:59

Air Blank	0.000	11:40
Control Test	0.022	11:40
Air Blank	0.000	11:41

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 07/27/2018 06:33:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2018 08:58

Air Blank	0.000	06:32
Control Test	0.000	06:33
Air Blank	0.000	06:33

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 09:42:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 14:00

Air Blank		09:41
Air Blank	0.000	09:42

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 09:45:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 14:00

Air Blank	0.000	09:44
Control Test	0.036	09:44
Air Blank	0.000	09:45

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 10:33:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 14:00

Air Blank	0.000	10:31
Control Test	0.064	10:32
Air Blank	0.000	10:33

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 10:58:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 14:00

Air Blank	0.000	10:56
Control Test	0.060	10:57
Air Blank	0.000	10:58

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 11:34:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 14:00

Air Blank		11:33
Air Blank	0.000	11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 09/28/2018 09:44:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/28/2018 11:55

Air Blank	0.000	09:43
Control Test	0.000	09:44
Air Blank	0.000	09:44

Agency: POLK COUNTY SO

Date/Time: 10/27/2018 11:00:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 13:39

Air Blank	0.000	10:59
Control Test	0.197	10:59
Air Blank	0.000	11:00

Agency: POLK COUNTY SO

Date/Time: 11/28/2018 06:12:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2018 09:31

Air Blank	0.000	06:10
Control Test	0.000	06:11
Air Blank	0.000	06:12

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 12/28/2018 06:41:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/28/2018 09:13

Air Blank	0.000	06:40
Control Test	0.000	06:40
Air Blank	0.000	06:41

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

Agency: MANATEE COUNTY SO

Date/Time: 05/29/2018 16:39:05

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/18/2018 09:27

Air Blank	0.000	16:38
Control Test	0.081	16:38
Air Blank	0.000	16:39

Agency: MANATEE COUNTY SO

Date/Time: 05/29/2018 16:40:48

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/18/2018 09:27

Air Blank	0.000	16:39
ITP Check	0.080	16:40
Air Blank	0.000	16:40

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

Agency: NEW SMYRNA BEACH
PD

Date/Time: 09/03/2018 08:48:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/03/2018 09:31

Air Blank	0.000	08:47
Control Test	0.000	08:48
Air Blank	0.000	08:48

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: LADY LAKE P.D.

Date/Time: 06/25/2018 08:32:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/03/2018 12:45

Air Blank	0.000	08:31
Control Test	0.075	08:32
Air Blank	0.000	08:32

Agency: LADY LAKE P.D.

Date/Time: 11/26/2018 07:57:50

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/21/2018 11:13

Air Blank	0.000	07:57
Control Test	0.079	07:57
Air Blank	0.000	07:57

Agency: LADY LAKE P.D.

Date/Time: 11/26/2018 07:59:14

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/21/2018 11:13

Air Blank	0.000	07:58
Control Test	0.079	07:58
Air Blank	0.000	07:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 13:50:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 14:55

Air Blank	0.000	13:49
Control Test	0.000	13:49
Air Blank	0.000	13:50

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 13:52:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 14:55

Air Blank	0.000	13:51
Control Test	0.000	13:52
Air Blank	0.000	13:52

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 13:54:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 14:55

Air Blank	0.000	13:53
Control Test	0.050	13:54
Air Blank	0.000	13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 13:57:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 14:55

Air Blank	0.000	13:55
Control Test	0.077	13:56
Air Blank	0.000	13:57

Agency: MARTIN COUNTY SO

Date/Time: 02/02/2018 14:01:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/02/2018 14:56

Air Blank	0.000	14:00
Control Test	0.201	14:01
Air Blank	0.000	14:01

Agency: MARTIN COUNTY SO

Date/Time: 10/24/2018 11:03:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/24/2018 12:06

Air Blank	0.000	11:02
Control Test	0.000	11:03
Air Blank	0.000	11:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 10/24/2018 11:05:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/24/2018 12:06

Air Blank	0.000	11:04
Control Test	0.000	11:05
Air Blank	0.000	11:05

Agency: MARTIN COUNTY SO

Date/Time: 10/24/2018 11:07:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/24/2018 12:06

Air Blank	0.000	11:06
Control Test	0.050	11:07
Air Blank	0.000	11:07

Agency: MARTIN COUNTY SO

Date/Time: 10/24/2018 11:09:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/24/2018 12:06

Air Blank	0.000	11:08
Control Test	0.077	11:09
Air Blank	0.000	11:09

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 10/24/2018 11:12:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/24/2018 12:06

Air Blank	0.000	11:10
Control Test	0.198	11:11
Air Blank	0.000	11:12

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 08/01/2018 07:28:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/01/2018 08:41

Air Blank	0.000	07:27
Control Test	0.000	07:28
Air Blank	0.000	07:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency: PANAMA CITY PD

Date/Time: 10/26/2018 11:22:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/26/2018 14:48

Air Blank	0.000	11:21
Control Test	0.000	11:22
Air Blank	0.000	11:22

Agency: PANAMA CITY PD

Date/Time: 10/26/2018 11:24:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/26/2018 14:49

Air Blank	0.000	11:23
Control Test	0.000	11:24
Air Blank	0.000	11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency: HENDRY COUNTY S.O.

Date/Time: 01/23/2018 22:08:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/23/2018 22:49

Air Blank	0.000	22:07
Control Test	0.000	22:07
Air Blank	0.000	22:08

Agency: HENDRY COUNTY S.O.

Date/Time: 01/23/2018 22:10:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/23/2018 22:49

Air Blank	0.000	22:09
Control Test	0.049	22:10
Air Blank	0.000	22:10

Agency: HENDRY COUNTY S.O.

Date/Time: 07/27/2018 21:15:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2018 22:12

Air Blank	0.000	21:13
Control Test	0.000	21:14
Air Blank	0.000	21:15

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency: HENDRY COUNTY S.O.

Date/Time: 08/22/2018 20:32:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/22/2018 21:53

Air Blank	0.000	20:30
Control Test	0.000	20:31
Air Blank	0.000	20:32

Agency: HENDRY COUNTY S.O.

Date/Time: 09/05/2018 21:26:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/05/2018 23:42

Air Blank	0.000	21:25
Control Test	0.078	21:26
Air Blank	0.000	21:26

Agency: HENDRY COUNTY S.O.

Date/Time: 10/01/2018 00:03:17

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/18/2018 23:15

Air Blank	0.000	00:02
Control Test	0.082	00:02
Air Blank	0.000	00:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency: HENDRY COUNTY S.O.

Date/Time: 11/28/2018 23:21:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/29/2018 00:47

Air Blank	0.000	23:19
Control Test	0.000	23:20
Air Blank	0.000	23:21

Agency: HENDRY COUNTY S.O.

Date/Time: 12/26/2018 20:59:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/26/2018 22:09

Air Blank	0.000	20:58
Control Test	0.000	20:59
Air Blank	0.000	20:59

Agency: HENDRY COUNTY S.O.

Date/Time: 01/06/2019 21:03:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/06/2019 22:08

Air Blank	0.000	21:02
Control Test	0.000	21:02
Air Blank	0.000	21:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency: HENDRY COUNTY S.O.

Date/Time: 01/06/2019 21:06:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/06/2019 22:08

Air Blank	0.000	21:04
Control Test	0.049	21:05
Air Blank	0.000	21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 01/30/2018 16:57:56

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/30/2018 17:39

Air Blank	0.000	16:57
Control Test	0.078	16:57
Air Blank	0.000	16:57

Agency: AUBURNDALE PD

Date/Time: 01/30/2018 17:00:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/30/2018 17:39

Air Blank	0.000	16:58
Control Test	0.050	16:59
Air Blank	0.000	17:00

Agency: AUBURNDALE PD

Date/Time: 01/30/2018 17:01:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/30/2018 17:39

Air Blank	0.000	17:00
Control Test	0.080	17:01
Air Blank	0.000	17:01

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 01/30/2018 17:03:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/30/2018 17:39

Air Blank	0.000	17:02
Control Test	0.199	17:03
Air Blank	0.000	17:03

Agency: AUBURNDALE PD

Date/Time: 03/20/2018 01:30:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2018 02:08

Air Blank	0.000	01:29
Control Test	0.048	01:29
Air Blank	0.000	01:30

Agency: AUBURNDALE PD

Date/Time: 03/20/2018 01:32:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2018 02:08

Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:32

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 03/20/2018 01:34:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2018 02:08

Air Blank	0.000	01:32
Control Test	0.200	01:33
Air Blank	0.000	01:34

Agency: AUBURNDALE PD

Date/Time: 04/12/2018 01:46:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/12/2018 02:27

Air Blank	0.000	01:45
Control Test	0.050	01:45
Air Blank	0.000	01:46

Agency: AUBURNDALE PD

Date/Time: 04/12/2018 01:48:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/12/2018 02:28

Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 04/12/2018 01:50:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/12/2018 02:28

Air Blank	0.000	01:48
Control Test	0.200	01:49
Air Blank	0.000	01:50

Agency: AUBURNDALE PD

Date/Time: 05/18/2018 21:23:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2018 22:21

Air Blank	0.000	21:21
Control Test	0.049	21:22
Air Blank	0.000	21:23

Agency: AUBURNDALE PD

Date/Time: 05/18/2018 21:25:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2018 22:21

Air Blank	0.000	21:23
Control Test	0.091	21:24
Air Blank	0.000	21:25

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 05/18/2018 21:27:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2018 22:21

Air Blank	0.000	21:25
Control Test	0.199	21:26
Air Blank	0.000	21:27

Agency: AUBURNDALE PD

Date/Time: 06/26/2018 02:26:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 03:17

Air Blank	0.000	02:25
Control Test	0.049	02:25
Air Blank	0.000	02:26

Agency: AUBURNDALE PD

Date/Time: 06/26/2018 02:28:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 03:17

Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 06/26/2018 02:30:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 03:17

Air Blank	0.000	02:29
Control Test	0.198	02:29
Air Blank	0.000	02:30

Agency: AUBURNDALE PD

Date/Time: 07/26/2018 10:13:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2018 11:05

Air Blank	0.000	10:12
Control Test	0.198	10:13
Air Blank	0.000	10:13

Agency: AUBURNDALE PD

Date/Time: 07/26/2018 10:15:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2018 11:05

Air Blank	0.000	10:14
Control Test	0.077	10:15
Air Blank	0.000	10:15

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 07/26/2018 10:17:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2018 11:05

Air Blank	0.000	10:16
Control Test	0.048	10:17
Air Blank	0.000	10:17

Agency: AUBURNDALE PD

Date/Time: 07/26/2018 10:20:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2018 11:05

Air Blank	0.000	10:19
Control Test	0.000	10:19
Air Blank	0.000	10:20

Agency: AUBURNDALE PD

Date/Time: 07/26/2018 10:23:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/26/2018 11:05

Air Blank	0.000	10:22
Control Test	0.079	10:22
Air Blank	0.000	10:23

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 08/25/2018 01:22:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/25/2018 02:09

Air Blank	0.000	01:21
Control Test	0.048	01:22
Air Blank	0.000	01:22

Agency: AUBURNDALE PD

Date/Time: 08/25/2018 01:24:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/25/2018 02:09

Air Blank	0.000	01:23
Control Test	0.077	01:24
Air Blank	0.000	01:24

Agency: AUBURNDALE PD

Date/Time: 08/25/2018 01:26:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/25/2018 02:09

Air Blank	0.000	01:25
Control Test	0.198	01:25
Air Blank	0.000	01:26

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 09/24/2018 00:15:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2018 01:05

Air Blank	0.000	00:14
Control Test	0.047	00:14
Air Blank	0.000	00:15

Agency: AUBURNDALE PD

Date/Time: 09/24/2018 00:17:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2018 01:05

Air Blank	0.000	00:16
Control Test	0.074	00:16
Air Blank	0.000	00:17

Agency: AUBURNDALE PD

Date/Time: 09/24/2018 00:19:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2018 01:05

Air Blank	0.000	00:18
Control Test	0.199	00:19
Air Blank	0.000	00:19

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 09/24/2018 00:21:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2018 01:05

Air Blank	0.000	00:20
Control Test	0.077	00:21
Air Blank	0.000	00:21

Agency: AUBURNDALE PD

Date/Time: 10/25/2018 09:09:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/25/2018 09:59

Air Blank	0.000	09:07
Control Test	0.000	09:08
Air Blank	0.000	09:09

Agency: AUBURNDALE PD

Date/Time: 10/25/2018 09:11:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/25/2018 09:59

Air Blank	0.000	09:10
Control Test	0.078	09:10
Air Blank	0.000	09:11

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 10/25/2018 09:13:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/25/2018 09:59

Air Blank	0.000	09:12
Control Test	0.000	09:12
Air Blank	0.000	09:13

Agency: AUBURNDALE PD

Date/Time: 10/25/2018 09:15:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/25/2018 09:59

Air Blank	0.000	09:13
Control Test	0.048	09:14
Air Blank	0.000	09:15

Agency: AUBURNDALE PD

Date/Time: 10/25/2018 09:17:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/25/2018 09:59

Air Blank	0.000	09:16
Control Test	0.199	09:16
Air Blank	0.000	09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 10/25/2018 09:18:51

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/25/2018 09:59

Air Blank	0.000	09:18
Control Test	0.079	09:18
Air Blank	0.000	09:18

Agency: AUBURNDALE PD

Date/Time: 11/21/2018 12:41:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/21/2018 13:49

Air Blank	0.000	12:41
Control Test	0.079	12:41
Air Blank	0.000	12:41

Agency: AUBURNDALE PD

Date/Time: 11/21/2018 12:43:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/21/2018 13:49

Air Blank	0.000	12:42
Control Test	0.048	12:43
Air Blank	0.000	12:43

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 11/21/2018 12:45:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/21/2018 13:49

Air Blank	0.000	12:44
Control Test	0.079	12:45
Air Blank	0.000	12:45

Agency: AUBURNDALE PD

Date/Time: 12/20/2018 09:22:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 10:11

Air Blank	0.000	09:21
Control Test	0.046	09:22
Air Blank	0.000	09:22

Agency: AUBURNDALE PD

Date/Time: 12/20/2018 09:24:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 10:11

Air Blank	0.000	09:23
Control Test	0.077	09:24
Air Blank	0.000	09:24

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 12/20/2018 09:27:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 10:11

Air Blank	0.000	09:25
Control Test	0.197	09:26
Air Blank	0.000	09:27

Agency: AUBURNDALE PD

Date/Time: 12/20/2018 09:28:41

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/20/2018 10:11

Air Blank	0.000	09:27
Control Test	0.079	09:28
Air Blank	0.000	09:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 01/20/2018 18:44:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/20/2018 19:39

Air Blank	0.000	18:43
Control Test	0.000	18:44
Air Blank	0.000	18:44

Agency: POLK COUNTY SO

Date/Time: 01/20/2018 18:46:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/20/2018 19:39

Air Blank	0.000	18:45
Control Test	0.048	18:46
Air Blank	0.000	18:46

Agency: POLK COUNTY SO

Date/Time: 02/28/2018 12:49:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2018 13:38

Air Blank	0.000	12:47
Control Test	0.195	12:48
Air Blank	0.000	12:49

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 02/28/2018 12:51:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2018 13:38

Air Blank	0.000	12:49
Control Test	0.047	12:50
Air Blank	0.000	12:51

Agency: POLK COUNTY SO

Date/Time: 07/27/2018 06:33:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2018 10:12

Air Blank	0.000	06:31
Control Test	0.078	06:32
Air Blank	0.000	06:33

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 09:44:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 13:58

Air Blank	0.000	09:43
Control Test	0.077	09:43
Air Blank	0.000	09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 08/29/2018 11:07:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 13:58

Air Blank	0.000	11:06
Control Test	0.196	11:06
Air Blank	0.000	11:07

Agency: POLK COUNTY SO

Date/Time: 09/28/2018 09:48:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/28/2018 11:53

Air Blank	0.000	09:47
Control Test	0.078	09:48
Air Blank	0.000	09:48

Agency: POLK COUNTY SO

Date/Time: 10/27/2018 10:59:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/27/2018 13:42

Air Blank	0.000	10:57
Control Test	0.079	10:58
Air Blank	0.000	10:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 11/28/2018 06:09:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2018 09:13

Air Blank	0.000	06:08
Control Test	0.000	06:09
Air Blank	0.000	06:09

Agency: POLK COUNTY SO

Date/Time: 12/28/2018 06:43:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/28/2018 09:06

Air Blank	0.000	06:42
Control Test	0.078	06:43
Air Blank	0.000	06:43

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency: BROWARD COUNTY SO

Date/Time: 09/20/2018 19:27:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/20/2018 20:18

Air Blank	0.000	19:26
Control Test	0.048	19:27
Air Blank	0.000	19:27

Agency: BROWARD COUNTY SO

Date/Time: 09/20/2018 19:29:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/20/2018 20:18

Air Blank	0.000	19:28
Control Test	0.077	19:29
Air Blank	0.000	19:29

Agency: BROWARD COUNTY SO

Date/Time: 09/20/2018 19:31:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/20/2018 20:18

Air Blank		19:30
Air Blank		19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency: BROWARD COUNTY SO

Date/Time: 09/20/2018 19:33:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/20/2018 20:18

Air Blank	0.000	19:31
Control Test	0.197	19:32
Air Blank	0.000	19:33

Agency: BROWARD COUNTY SO

Date/Time: 09/20/2018 19:35:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/20/2018 20:18

Air Blank	0.000	19:34
Control Test	0.000	19:35
Air Blank	0.000	19:35

Agency: BROWARD COUNTY SO

Date/Time: 09/20/2018 19:37:18

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/20/2018 20:18

Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency: ROCKLEDGE P.D.

Date/Time: 08/27/2018 07:54:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/27/2018 09:17

Air Blank	0.000	07:53
Control Test	0.052	07:53
Air Blank	0.000	07:54

Agency: ROCKLEDGE P.D.

Date/Time: 08/27/2018 07:57:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/27/2018 09:17

Air Blank	0.000	07:56
Control Test	0.052	07:56
Air Blank	0.000	07:57

Agency: ROCKLEDGE P.D.

Date/Time: 08/27/2018 07:59:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/27/2018 09:17

Air Blank	0.000	07:58
Control Test	0.053	07:59
Air Blank	0.000	07:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency: INDIAN HARBOUR PD

Date/Time: 04/26/2018 09:34:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/26/2018 18:47

Air Blank	0.000	09:33
Control Test	0.000	09:33
Air Blank	0.000	09:34

Agency: INDIAN HARBOUR PD

Date/Time: 05/28/2018 13:20:47

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/29/2018 05:54

Air Blank	0.000	13:19
Control Test	0.081	13:20
Air Blank	0.000	13:20

Agency: INDIAN HARBOUR PD

Date/Time: 05/28/2018 13:22:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/29/2018 05:54

Air Blank	0.000	13:21
Control Test	0.000	13:22
Air Blank	0.000	13:22

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency: INDIAN HARBOUR PD

Date/Time: 05/28/2018 13:24:03

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 05/29/2018 05:54

Air Blank	0.000	13:23
ITP Check	0.083	13:23
Air Blank	0.000	13:24

Agency: INDIAN HARBOUR PD

Date/Time: 05/29/2018 05:44:32

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 05/29/2018 05:54

Air Blank	0.000	05:43
ITP Check	0.078	05:44
Air Blank	0.000	05:44

Agency: INDIAN HARBOUR PD

Date/Time: 05/29/2018 05:46:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/29/2018 05:54

Air Blank	0.000	05:45
Control Test	0.000	05:45
Air Blank	0.000	05:46

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency: INDIAN HARBOUR PD

Date/Time: 10/20/2018 06:29:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/20/2018 08:52

Air Blank	0.000	06:28
Control Test	0.000	06:29
Air Blank	0.000	06:29

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency: PALM BAY P.D.

Date/Time: 04/03/2018 16:17:32

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/25/2018 20:33

Air Blank	0.000	16:16
Control Test	0.080	16:17
Air Blank	0.000	16:17

Agency: PALM BAY P.D.

Date/Time: 12/07/2018 00:30:50

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/29/2018 22:01

Air Blank	0.000	00:30
Control Test	0.082	00:30
Air Blank	0.000	00:30

Agency: PALM BAY P.D.

Date/Time: 12/08/2018 14:56:45

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/29/2018 22:01

Air Blank	0.000	14:55
Control Test	0.081	14:56
Air Blank	0.000	14:56

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency: PALM BAY P.D.

Date/Time: 12/08/2018 14:58:27

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 12/29/2018 22:01

Air Blank	0.000	14:57
ITP Check	0.081	14:58
Air Blank	0.000	14:58

Agency: PALM BAY P.D.

Date/Time: 12/08/2018 15:02:21

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 12/29/2018 22:01

Air Blank	0.000	15:01
ITP Check	0.079	15:01
Air Blank	0.000	15:02

Agency: PALM BAY P.D.

Date/Time: 12/08/2018 15:06:36

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 12/29/2018 22:01

Air Blank	0.000	15:05
ITP Check	0.080	15:06
Air Blank	0.000	15:06

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency: PINELLAS COUNTY SO

Date/Time: 10/26/2018 19:26:21

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/29/2018 01:10

Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26

Agency: PINELLAS COUNTY SO

Date/Time: 10/26/2018 19:28:41

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 11/29/2018 01:10

Air Blank	0.000	19:27
ITP Check	0.078	19:28
Air Blank	0.000	19:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency: JACKSONVILLE SO

Date/Time: 04/17/2018 13:04:43

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/01/2018 14:51

Air Blank	0.000	13:03
Control Test	0.081	13:04
Air Blank	0.000	13:04

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency: JACKSONVILLE SO

Date/Time: 04/06/2018 13:59:52

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/17/2018 13:09

Air Blank	0.000	13:59
Control Test	0.079	13:59
Air Blank	0.000	13:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

Agency: HIGH SPRINGS PD

Date/Time: 03/04/2018 12:17:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/04/2018 19:09

Air Blank	0.000	12:16
Control Test	0.000	12:17
Air Blank	0.000	12:17

Agency: HIGH SPRINGS PD

Date/Time: 07/12/2018 11:26:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/12/2018 16:00

Air Blank	0.000	11:25
Control Test	0.000	11:26
Air Blank	0.000	11:26

Agency: HIGH SPRINGS PD

Date/Time: 07/12/2018 11:28:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/12/2018 16:00

Air Blank	0.000	11:27
Control Test	0.051	11:27
Air Blank	0.000	11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

Agency: HIGH SPRINGS PD

Date/Time: 08/11/2018 03:29:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/11/2018 04:36

Air Blank	0.000	03:28
Control Test	0.000	03:28
Air Blank	0.000	03:29

Agency: HIGH SPRINGS PD

Date/Time: 08/11/2018 03:33:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/11/2018 04:36

Air Blank	0.000	03:32
Control Test	0.050	03:33
Air Blank	0.000	03:33

Agency: HIGH SPRINGS PD

Date/Time: 08/11/2018 03:34:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/11/2018 04:36

Air Blank		03:34
Air Blank	0.000	03:34

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

Agency: HIGH SPRINGS PD

Date/Time: 08/11/2018 03:37:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/11/2018 04:36

Air Blank	0.000	03:35
Control Test	0.000	03:36
Air Blank	0.000	03:37

Agency: HIGH SPRINGS PD

Date/Time: 11/13/2018 02:26:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/13/2018 03:40

Air Blank	0.000	02:26
Control Test	0.000	02:26
Air Blank	0.000	02:26

Agency: HIGH SPRINGS PD

Date/Time: 11/13/2018 02:28:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/13/2018 03:40

Air Blank	0.000	02:27
Control Test	0.000	02:28
Air Blank	0.000	02:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

Agency: HIGH SPRINGS PD

Date/Time: 11/13/2018 02:30:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/13/2018 03:40

Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30

Agency: HIGH SPRINGS PD

Date/Time: 11/13/2018 02:34:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/13/2018 03:40

Air Blank	0.000	02:33
Control Test	0.050	02:33
Air Blank	0.000	02:34

Agency: HIGH SPRINGS PD

Date/Time: 11/13/2018 02:36:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/13/2018 03:40

Air Blank	0.000	02:35
Control Test	0.000	02:36
Air Blank	0.000	02:36

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

Agency: BRADFORD COUNTY SO

Date/Time: 01/21/2018 13:01:10

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/21/2018 13:56

Air Blank	0.000	13:00
Control Test	0.081	13:00
Air Blank	0.000	13:01

Agency: BRADFORD COUNTY SO

Date/Time: 01/21/2018 13:16:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/21/2018 13:56

Air Blank	0.000	13:15
Control Test	0.000	13:15
Air Blank	0.000	13:16

Agency: BRADFORD COUNTY SO

Date/Time: 01/21/2018 13:18:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/21/2018 13:56

Air Blank	0.000	13:16
Control Test	0.016	13:17
Air Blank	0.000	13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

Agency: BRADFORD COUNTY SO

Date/Time: 01/21/2018 13:20:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/21/2018 13:56

Air Blank	0.000	13:19
Control Test	0.048	13:19
Air Blank	0.000	13:20

Agency: BRADFORD COUNTY SO

Date/Time: 01/21/2018 13:22:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/21/2018 13:56

Air Blank	0.000	13:20
Control Test	0.077	13:21
Air Blank	0.000	13:22

Agency: BRADFORD COUNTY SO

Date/Time: 01/21/2018 13:24:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/21/2018 13:56

Air Blank	0.000	13:23
Control Test	0.193	13:24
Air Blank	0.000	13:24

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency: BRADFORD COUNTY SO

Date/Time: 01/21/2018 13:00:17

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/21/2018 14:03

Air Blank	0.000	12:59
Control Test	0.080	12:59
Air Blank	0.000	13:00

Agency: BRADFORD COUNTY SO

Date/Time: 07/18/2018 09:17:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/18/2018 10:36

Air Blank	0.000	09:16
Control Test	0.193	09:16
Air Blank	0.000	09:17

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

Agency: LEVY COUNTY SO

Date/Time: 05/06/2018 15:37:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/06/2018 15:44

Air Blank	0.000	15:36
Control Test	0.046	15:37
Air Blank	0.000	15:37

Agency: LEVY COUNTY SO

Date/Time: 07/27/2018 19:58:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 19:41

Air Blank	0.000	19:57
Control Test	0.046	19:58
Air Blank	0.000	19:58

Agency: LEVY COUNTY SO

Date/Time: 11/29/2018 20:32:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/29/2018 21:26

Air Blank	0.000	20:31
Control Test	0.000	20:31
Air Blank	0.000	20:32

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 19:52:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/03/2019 16:12

Air Blank	0.000	19:51
Control Test	0.071	19:51
Air Blank	0.000	19:52

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 19:54:17

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/03/2019 16:12

Air Blank	0.000	19:53
Control Test	0.070	19:53
Air Blank	0.000	19:54

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 21:49:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2019 16:12

Air Blank	0.000	21:48
Control Test	0.082	21:48
Air Blank	0.000	21:49

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 21:53:45

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/03/2019 16:12

Air Blank	0.000	21:52
Control Test	0.074	21:53
Air Blank	0.000	21:53

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 22:55:14

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/03/2019 16:12

Air Blank	0.000	22:54
Control Test	0.075	22:54
Air Blank	0.000	22:55

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 22:57:17

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/03/2019 16:12

Air Blank	0.000	22:56
Control Test	0.074	22:56
Air Blank	0.000	22:57

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 20:58:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/03/2019 16:04

Air Blank	0.000	20:58
Control Test	0.074	20:58
Air Blank	0.000	20:58

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 21:03:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2019 16:04

Air Blank	0.000	21:01
Control Test	0.198	21:02
Air Blank	0.000	21:03

Agency: SANTA ROSA COUNTY
SO

Date/Time: 12/31/2018 21:04:38

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/03/2019 16:04

Air Blank	0.000	21:03
Control Test	0.074	21:04
Air Blank	0.000	21:04

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

Agency: BAY COUNTY SO

Date/Time: 05/17/2018 16:09:19

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/30/2018 10:11

Air Blank	0.000	16:08
Control Test	0.081	16:08
Air Blank	0.000	16:09

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

Agency: INDIAN RIVER CO. SO

Date/Time: 06/30/2018 05:44:19

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/13/2018 05:20

Air Blank	0.000	05:43
Control Test	0.076	05:43
Air Blank	0.000	05:44

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

Agency: DESOTO COUNTY SO

Date/Time: 04/16/2018 07:59:29

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/24/2018 09:30

Air Blank	0.000	07:58
Control Test	0.079	07:59
Air Blank	0.000	07:59

Agency: DESOTO COUNTY SO

Date/Time: 04/16/2018 08:00:54

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/24/2018 09:30

Air Blank	0.000	08:00
Control Test	0.078	08:00
Air Blank	0.000	08:00

Agency: DESOTO COUNTY SO

Date/Time: 04/16/2018 08:02:14

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/24/2018 09:30

Air Blank	0.000	08:01
Control Test	0.078	08:01
Air Blank	0.000	08:02

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

Agency: SARASOTA COUNTY SO
Date/Time: 05/26/2018 06:59:16
Standard Type: Dry Gas
Event Source: Local
Data Download Date/Time: 05/31/2018 07:38

Air Blank	0.000	06:58
Control Test	0.079	06:58
Air Blank	0.000	06:59

Agency: SARASOTA COUNTY SO
Date/Time: 05/26/2018 07:08:01
Standard Type: ITP
Event Source: Local
Data Download Date/Time: 05/31/2018 07:38

Air Blank	0.000	07:07
ITP Check	0.082	07:07
Air Blank	0.000	07:08

Agency: SARASOTA COUNTY SO
Date/Time: 05/26/2018 07:11:04
Standard Type: Dry Gas
Event Source: Local
Data Download Date/Time: 05/31/2018 07:38

Air Blank	0.000	07:10
Control Test	0.080	07:10
Air Blank	0.000	07:11

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency: EVERGLADES NAT PARK

Date/Time: 08/29/2018 10:12:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2018 13:41

Air Blank	0.000	10:11
Control Test	0.078	10:12
Air Blank	0.000	10:12

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/22/2018 00:20:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/26/2018 07:19

Air Blank	0.000	00:19
Control Test	0.000	00:20
Air Blank	0.000	00:20

Agency: CHARLOTTE COUNTY
SO

Date/Time: 02/22/2018 19:36:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/22/2018 23:49

Air Blank	0.000	19:34
Control Test	0.000	19:35
Air Blank	0.000	19:36

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 03/15/2018 13:22:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/15/2018 14:38

Air Blank	0.000	13:20
Control Test	0.000	13:21
Air Blank	0.000	13:22

Agency: MARTIN COUNTY SO

Date/Time: 03/15/2018 13:28:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/15/2018 14:38

Air Blank	0.000	13:27
Control Test	0.000	13:28
Air Blank	0.000	13:28

Agency: MARTIN COUNTY SO

Date/Time: 03/15/2018 13:30:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/15/2018 14:38

Air Blank	0.000	13:29
Control Test	0.007	13:30
Air Blank	0.000	13:30

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 03/15/2018 13:33:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/15/2018 14:38

Air Blank	0.000	13:31
Control Test	0.052	13:32
Air Blank	0.000	13:33

Agency: MARTIN COUNTY SO

Date/Time: 03/15/2018 13:37:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/15/2018 14:38

Air Blank	0.000	13:36
Control Test	0.079	13:36
Air Blank	0.000	13:37

Agency: MARTIN COUNTY SO

Date/Time: 03/15/2018 13:39:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/15/2018 14:38

Air Blank	0.000	13:38
Control Test	0.199	13:39
Air Blank	0.000	13:39

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency: ST PETERSBURG PD

Date/Time: 11/28/2018 07:38:43

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/28/2018 08:18

Air Blank	0.000	07:37
Control Test	0.080	07:38
Air Blank	0.000	07:38

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

Agency: MICCOSUKEE PD

Date/Time: 11/30/2018 08:58:22

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/30/2018 11:08

Air Blank	0.000	08:57
Control Test	0.079	08:57
Air Blank	0.000	08:58

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency: ALACHUA PD

Date/Time: 12/29/2018 03:36:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/29/2018 04:36

Air Blank	0.000	03:35
Control Test	0.197	03:36
Air Blank	0.000	03:36

Agency: ALACHUA PD

Date/Time: 12/29/2018 03:40:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/29/2018 04:36

Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency: PALM BAY P.D.

Date/Time: 03/28/2018 20:40:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/28/2018 21:23

Air Blank		20:39
Air Blank	0.000	20:40

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 01/05/2018 09:45:58

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/05/2018 10:52

Air Blank	0.000	09:45
Control Test	0.081	09:45
Air Blank	0.000	09:45

Agency: PINELLAS PARK P.D.

Date/Time: 01/05/2018 09:47:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/05/2018 10:52

Air Blank	0.000	09:46
Control Test	0.000	09:47
Air Blank	0.000	09:47

Agency: PINELLAS PARK P.D.

Date/Time: 01/05/2018 09:51:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/05/2018 10:52

Air Blank	0.000	09:49
Control Test	0.000	09:50
Air Blank	0.000	09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 01/05/2018 09:53:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/05/2018 10:52

Air Blank	0.000	09:51
Control Test	0.048	09:52
Air Blank	0.000	09:53

Agency: PINELLAS PARK P.D.

Date/Time: 01/05/2018 09:55:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/05/2018 10:52

Air Blank	0.000	09:53
Control Test	0.078	09:54
Air Blank	0.000	09:55

Agency: PINELLAS PARK P.D.

Date/Time: 01/05/2018 09:57:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/05/2018 10:52

Air Blank	0.000	09:55
Control Test	0.206	09:56
Air Blank		09:57

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 01/05/2018 10:27:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/05/2018 10:52

Air Blank	0.000	10:25
Control Test	0.193	10:26
Air Blank	0.000	10:27

Agency: PINELLAS PARK P.D.

Date/Time: 02/01/2018 13:42:27

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/01/2018 14:37

Air Blank	0.000	13:41
Control Test	0.081	13:41
Air Blank	0.000	13:42

Agency: PINELLAS PARK P.D.

Date/Time: 02/01/2018 13:45:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2018 14:37

Air Blank	0.000	13:43
Control Test	0.000	13:43
Air Blank		13:44
Air Blank		13:45

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 02/01/2018 13:59:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2018 14:37

Air Blank	0.000	13:58
Control Test	0.000	13:59
Air Blank	0.000	13:59

Agency: PINELLAS PARK P.D.

Date/Time: 02/01/2018 14:01:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2018 14:37

Air Blank	0.000	14:00
Control Test	0.000	14:00
Air Blank	0.000	14:01

Agency: PINELLAS PARK P.D.

Date/Time: 02/01/2018 14:03:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2018 14:37

Air Blank	0.000	14:02
Control Test	0.047	14:02
Air Blank	0.000	14:03

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 02/01/2018 14:05:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2018 14:37

Air Blank	0.000	14:04
Control Test	0.076	14:04
Air Blank	0.000	14:05

Agency: PINELLAS PARK P.D.

Date/Time: 02/01/2018 14:07:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2018 14:37

Air Blank	0.000	14:06
Control Test	0.192	14:06
Air Blank	0.000	14:07

Agency: PINELLAS PARK P.D.

Date/Time: 02/13/2018 09:41:24

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/13/2018 10:34

Air Blank	0.000	09:40
Control Test	0.079	09:40
Air Blank	0.000	09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 02/13/2018 09:43:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2018 10:34

Air Blank	0.000	09:42
Control Test	0.000	09:42
Air Blank	0.000	09:43

Agency: PINELLAS PARK P.D.

Date/Time: 02/13/2018 09:45:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2018 10:34

Air Blank	0.000	09:44
Control Test	0.000	09:44
Air Blank	0.000	09:45

Agency: PINELLAS PARK P.D.

Date/Time: 02/13/2018 09:47:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2018 10:34

Air Blank	0.000	09:46
Control Test	0.048	09:46
Air Blank	0.000	09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 02/13/2018 09:57:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2018 10:34

Air Blank	0.000	09:55
Control Test	0.077	09:56
Air Blank	0.000	09:57

Agency: PINELLAS PARK P.D.

Date/Time: 02/13/2018 09:59:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2018 10:35

Air Blank	0.000	09:57
Control Test	0.193	09:58
Air Blank	0.000	09:59

Agency: PINELLAS PARK P.D.

Date/Time: 03/02/2018 07:12:32

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/02/2018 07:58

Air Blank	0.000	07:11
Control Test	0.079	07:12
Air Blank	0.000	07:12

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 03/02/2018 07:14:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/02/2018 07:58

Air Blank	0.000	07:13
Control Test	0.000	07:13
Air Blank	0.000	07:14

Agency: PINELLAS PARK P.D.

Date/Time: 03/02/2018 07:16:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/02/2018 07:58

Air Blank	0.000	07:14
Control Test	0.000	07:15
Air Blank	0.000	07:16

Agency: PINELLAS PARK P.D.

Date/Time: 03/02/2018 07:17:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/02/2018 07:58

Air Blank	0.000	07:16
Control Test	0.048	07:17
Air Blank	0.000	07:17

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 03/02/2018 07:18:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/02/2018 07:58

Air Blank		07:18
Air Blank	0.000	07:18

Agency: PINELLAS PARK P.D.

Date/Time: 03/02/2018 07:20:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/02/2018 07:58

Air Blank	0.000	07:19
Control Test	0.077	07:20
Air Blank	0.000	07:20

Agency: PINELLAS PARK P.D.

Date/Time: 03/02/2018 07:22:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/02/2018 07:58

Air Blank	0.000	07:21
Control Test	0.193	07:22
Air Blank	0.000	07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 04/06/2018 08:28:58

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/06/2018 09:18

Air Blank	0.000	08:28
Control Test	0.078	08:28
Air Blank	0.000	08:28

Agency: PINELLAS PARK P.D.

Date/Time: 04/06/2018 08:30:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/06/2018 09:18

Air Blank	0.000	08:29
Control Test	0.000	08:30
Air Blank	0.000	08:30

Agency: PINELLAS PARK P.D.

Date/Time: 04/06/2018 08:32:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/06/2018 09:18

Air Blank	0.000	08:31
Control Test	0.000	08:32
Air Blank	0.000	08:32

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 04/06/2018 08:34:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/06/2018 09:18

Air Blank	0.000	08:33
Control Test	0.047	08:34
Air Blank	0.000	08:34

Agency: PINELLAS PARK P.D.

Date/Time: 04/06/2018 08:36:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/06/2018 09:18

Air Blank	0.000	08:35
Control Test	0.076	08:36
Air Blank	0.000	08:36

Agency: PINELLAS PARK P.D.

Date/Time: 04/06/2018 08:39:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/06/2018 09:18

Air Blank	0.000	08:38
Control Test	0.195	08:38
Air Blank	0.000	08:39

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 05/04/2018 09:38:16

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/04/2018 10:35

Air Blank	0.000	09:37
Control Test	0.079	09:37
Air Blank	0.000	09:38

Agency: PINELLAS PARK P.D.

Date/Time: 05/04/2018 09:40:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2018 10:35

Air Blank	0.000	09:39
Control Test	0.000	09:40
Air Blank	0.000	09:40

Agency: PINELLAS PARK P.D.

Date/Time: 05/04/2018 09:42:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2018 10:35

Air Blank	0.000	09:41
Control Test	0.000	09:42
Air Blank	0.000	09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 05/04/2018 09:44:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2018 10:35

Air Blank	0.000	09:43
Control Test	0.048	09:43
Air Blank	0.000	09:44

Agency: PINELLAS PARK P.D.

Date/Time: 05/04/2018 09:46:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2018 10:35

Air Blank	0.000	09:45
Control Test	0.076	09:46
Air Blank	0.000	09:46

Agency: PINELLAS PARK P.D.

Date/Time: 05/04/2018 09:48:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2018 10:35

Air Blank	0.000	09:47
Control Test	0.192	09:47
Air Blank	0.000	09:48

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 05/04/2018 09:49:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2018 10:35

Air Blank	0.000	09:48
Control Test	0.195	09:49
Air Blank	0.000	09:49

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 12:03:33

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	12:02
Control Test	0.079	12:03
Air Blank	0.000	12:03

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 12:05:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	12:04
Control Test	0.000	12:05
Air Blank	0.000	12:05

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 12:08:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	12:06
Control Test	0.000	12:07
Air Blank	0.000	12:08

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 12:10:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	12:08
Control Test	0.050	12:09
Air Blank	0.000	12:10

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 12:11:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank		12:10
Air Blank	0.000	12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 12:12:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	12:11
Control Test	0.049	12:12
Air Blank	0.000	12:12

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 12:14:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	12:13
Control Test	0.077	12:14
Air Blank	0.000	12:14

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 12:16:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	12:15
Control Test	0.195	12:16
Air Blank	0.000	12:16

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 13:21:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	13:19
Control Test	0.047	13:20
Air Blank	0.000	13:21

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2018 13:25:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 13:37

Air Blank	0.000	13:24
Control Test	0.000	13:24
Air Blank	0.000	13:25

Agency: PINELLAS PARK P.D.

Date/Time: 06/20/2018 13:02:40

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/20/2018 13:55

Air Blank	0.000	13:01
Control Test	0.079	13:02
Air Blank	0.000	13:02

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 06/20/2018 13:04:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/20/2018 13:55

Air Blank	0.000	13:03
Control Test	0.000	13:03
Air Blank	0.000	13:04

Agency: PINELLAS PARK P.D.

Date/Time: 06/20/2018 13:06:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/20/2018 13:55

Air Blank	0.000	13:05
Control Test	0.000	13:05
Air Blank	0.000	13:06

Agency: PINELLAS PARK P.D.

Date/Time: 06/20/2018 13:08:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/20/2018 13:55

Air Blank	0.000	13:06
Control Test	0.048	13:07
Air Blank	0.000	13:08

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 06/20/2018 13:10:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/20/2018 13:55

Air Blank	0.000	13:08
Control Test	0.077	13:09
Air Blank	0.000	13:10

Agency: PINELLAS PARK P.D.

Date/Time: 06/20/2018 13:11:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/20/2018 13:55

Air Blank	0.000	13:10
Control Test	0.195	13:11
Air Blank	0.000	13:11

Agency: PINELLAS PARK P.D.

Date/Time: 12/20/2018 13:25:35

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/20/2018 14:16

Air Blank	0.000	13:24
Control Test	0.078	13:25
Air Blank	0.000	13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 12/20/2018 13:27:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 14:16

Air Blank	0.000	13:26
Control Test	0.000	13:27
Air Blank	0.000	13:27

Agency: PINELLAS PARK P.D.

Date/Time: 12/20/2018 13:29:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 14:16

Air Blank	0.000	13:28
Control Test	0.000	13:29
Air Blank	0.000	13:29

Agency: PINELLAS PARK P.D.

Date/Time: 12/20/2018 13:31:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 14:16

Air Blank	0.000	13:30
Control Test	0.047	13:31
Air Blank	0.000	13:31

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 12/20/2018 13:34:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 14:16

Air Blank	0.000	13:32
Control Test	0.077	13:33
Air Blank	0.000	13:34

Agency: PINELLAS PARK P.D.

Date/Time: 12/20/2018 13:36:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/20/2018 14:16

Air Blank	0.000	13:34
Control Test	0.196	13:35
Air Blank	0.000	13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency: NAS JACKSONVILLE PD

Date/Time: 01/13/2018 05:18:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2018 05:59

Air Blank	0.000	05:17
Control Test	0.000	05:18
Air Blank	0.000	05:18

Agency: NAS JACKSONVILLE PD

Date/Time: 11/18/2018 03:52:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/18/2018 07:18

Air Blank	0.000	03:50
Control Test	0.000	03:51
Air Blank	0.000	03:52

Agency: NAS JACKSONVILLE PD

Date/Time: 11/18/2018 03:54:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/18/2018 07:18

Air Blank	0.000	03:53
Control Test	0.000	03:54
Air Blank	0.000	03:54

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

Agency: SURFSIDE PD

Date/Time: 03/05/2018 12:44:17

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/05/2018 14:54

Air Blank	0.000	12:43
Control Test	0.000	12:43
Air Blank	0.000	12:44

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency: FHP TROOP L

Date/Time: 05/11/2018 15:06:47

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/11/2018 15:43

Air Blank	0.000	15:05
Control Test	0.081	15:06
Air Blank	0.000	15:06

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency: FL HIGHWAY PATROL

Date/Time: 01/30/2018 16:59:13

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/30/2018 18:04

Air Blank	0.000	16:58
Control Test	0.081	16:58
Air Blank	0.000	16:59

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency: HERNANDO COUNTY SO

Date/Time: 11/21/2018 18:31:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/21/2018 20:31

Air Blank	0.000	18:30
Control Test	0.000	18:30
Air Blank	0.000	18:31

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency: PINELLAS COUNTY SO

Date/Time: 04/14/2018 00:05:34

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/28/2018 03:02

Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05

Agency: PINELLAS COUNTY SO

Date/Time: 04/14/2018 00:07:42

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 04/28/2018 03:02

Air Blank	0.000	00:06
ITP Check	0.081	00:07
Air Blank	0.000	00:07

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency: DAYTONA BEACH PD

Date/Time: 05/01/2018 16:34:04

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/29/2018 17:09

Air Blank	0.000	16:33
Control Test	0.081	16:33
Air Blank	0.000	16:34

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency: FERNANDINA BEACH PD

Date/Time: 01/03/2018 01:19:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2018 01:55

Air Blank	0.000	01:18
Control Test	0.000	01:18
Air Blank	0.000	01:19

Agency: FERNANDINA BEACH PD

Date/Time: 08/08/2018 16:36:42

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/09/2018 09:57

Air Blank	0.000	16:35
Control Test	0.080	16:36
Air Blank	0.000	16:36

Agency: FERNANDINA BEACH PD

Date/Time: 08/08/2018 16:39:46

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 08/09/2018 09:57

Air Blank	0.000	16:38
ITP Check	0.082	16:39
Air Blank	0.000	16:39

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency: FERNANDINA BEACH PD

Date/Time: 08/08/2018 16:42:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/09/2018 09:57

Air Blank	0.000	16:41
Control Test	0.051	16:42
Air Blank	0.000	16:42

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

Agency: MARTIN COUNTY SO

Date/Time: 08/12/2018 21:06:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/12/2018 22:17

Air Blank	0.000	21:05
Control Test	0.000	21:06
Air Blank	0.000	21:06

Agency: MARTIN COUNTY SO

Date/Time: 08/12/2018 21:19:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/12/2018 22:17

Air Blank	0.000	21:18
Control Test	0.000	21:19
Air Blank	0.000	21:19

Agency: MARTIN COUNTY SO

Date/Time: 08/12/2018 21:21:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/12/2018 22:17

Air Blank	0.000	21:20
Control Test	0.048	21:21
Air Blank	0.000	21:21

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

Agency: MARTIN COUNTY SO

Date/Time: 08/12/2018 21:23:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/12/2018 22:17

Air Blank	0.000	21:22
Control Test	0.075	21:23
Air Blank	0.000	21:23

Agency: MARTIN COUNTY SO

Date/Time: 08/12/2018 21:26:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/12/2018 22:17

Air Blank	0.000	21:24
Control Test	0.195	21:25
Air Blank	0.000	21:26

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

Agency: MARTIN COUNTY SO

Date/Time: 04/27/2018 08:40:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/27/2018 09:37

Air Blank	0.000	08:39
Control Test	0.076	08:40
Air Blank	0.000	08:40

Agency: MARTIN COUNTY SO

Date/Time: 05/27/2018 16:07:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/27/2018 16:55

Air Blank	0.000	16:05
Control Test	0.000	16:06
Air Blank	0.000	16:07

Agency: MARTIN COUNTY SO

Date/Time: 05/27/2018 16:09:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/27/2018 16:55

Air Blank	0.000	16:08
Control Test	0.000	16:08
Air Blank	0.000	16:09

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

Agency: MARTIN COUNTY SO

Date/Time: 05/27/2018 16:11:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/27/2018 16:55

Air Blank	0.000	16:10
Control Test	0.048	16:11
Air Blank	0.000	16:11

Agency: MARTIN COUNTY SO

Date/Time: 05/27/2018 16:13:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/27/2018 16:55

Air Blank	0.000	16:12
Control Test	0.076	16:13
Air Blank	0.000	16:13

Agency: MARTIN COUNTY SO

Date/Time: 05/27/2018 16:16:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/27/2018 16:55

Air Blank	0.000	16:15
Control Test	0.192	16:15
Air Blank	0.000	16:16

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

Agency: MARTIN COUNTY SO

Date/Time: 06/26/2018 13:30:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 15:51

Air Blank	0.000	13:29
Control Test	0.078	13:30
Air Blank	0.000	13:30

Agency: MARTIN COUNTY SO

Date/Time: 06/26/2018 13:32:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/26/2018 15:51

Air Blank	0.000	13:31
Control Test	0.193	13:32
Air Blank	0.000	13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 04/27/2018 08:32:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/27/2018 09:28

Air Blank	0.000	08:31
Control Test	0.000	08:32
Air Blank	0.000	08:32

Agency: MARTIN COUNTY SO

Date/Time: 04/27/2018 08:35:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/27/2018 09:28

Air Blank	0.000	08:33
Control Test	0.000	08:34
Air Blank	0.000	08:35

Agency: MARTIN COUNTY SO

Date/Time: 04/27/2018 08:37:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/27/2018 09:29

Air Blank	0.000	08:36
Control Test	0.048	08:36
Air Blank	0.000	08:37

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 04/27/2018 08:39:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/27/2018 09:29

Air Blank	0.000	08:38
Control Test	0.193	08:39
Air Blank	0.000	08:39

Agency: MARTIN COUNTY SO

Date/Time: 04/27/2018 08:41:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/27/2018 09:29

Air Blank	0.000	08:40
Control Test	0.076	08:41
Air Blank	0.000	08:41

Agency: MARTIN COUNTY SO

Date/Time: 07/21/2018 00:22:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2018 01:36

Air Blank	0.000	00:21
Control Test	0.000	00:22
Air Blank	0.000	00:22

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 07/21/2018 00:25:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2018 01:36

Air Blank	0.000	00:23
Control Test	0.000	00:24
Air Blank	0.000	00:25

Agency: MARTIN COUNTY SO

Date/Time: 07/21/2018 00:27:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2018 01:36

Air Blank	0.000	00:25
Control Test	0.000	00:26
Air Blank	0.000	00:27

Agency: MARTIN COUNTY SO

Date/Time: 07/21/2018 00:28:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2018 01:36

Air Blank	0.000	00:27
Control Test	0.048	00:28
Air Blank	0.000	00:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 07/21/2018 00:30:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2018 01:36

Air Blank	0.000	00:29
Control Test	0.074	00:30
Air Blank	0.000	00:30

Agency: MARTIN COUNTY SO

Date/Time: 07/21/2018 00:32:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2018 01:36

Air Blank	0.000	00:31
Control Test	0.076	00:32
Air Blank	0.000	00:32

Agency: MARTIN COUNTY SO

Date/Time: 07/21/2018 00:34:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2018 01:36

Air Blank	0.000	00:33
Control Test	0.191	00:33
Air Blank	0.000	00:34

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 11/29/2018 14:39:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/29/2018 15:50

Air Blank	0.000	14:38
Control Test	0.000	14:38
Air Blank	0.000	14:39

Agency: MARTIN COUNTY SO

Date/Time: 11/29/2018 14:41:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/29/2018 15:50

Air Blank	0.000	14:40
Control Test	0.000	14:41
Air Blank	0.000	14:41

Agency: MARTIN COUNTY SO

Date/Time: 11/29/2018 14:43:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/29/2018 15:50

Air Blank	0.000	14:42
Control Test	0.048	14:42
Air Blank	0.000	14:43

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 11/29/2018 14:45:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/29/2018 15:50

Air Blank	0.000	14:44
Control Test	0.076	14:45
Air Blank	0.000	14:45

Agency: MARTIN COUNTY SO

Date/Time: 11/29/2018 14:47:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/29/2018 15:50

Air Blank	0.000	14:46
Control Test	0.194	14:46
Air Blank	0.000	14:47

Agency: MARTIN COUNTY SO

Date/Time: 12/19/2018 21:08:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/19/2018 22:30

Air Blank	0.000	21:07
Control Test	0.000	21:08
Air Blank	0.000	21:08

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 12/19/2018 21:10:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/19/2018 22:30

Air Blank	0.000	21:09
Control Test	0.000	21:10
Air Blank	0.000	21:10

Agency: MARTIN COUNTY SO

Date/Time: 12/19/2018 21:12:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/19/2018 22:30

Air Blank	0.000	21:11
Control Test	0.048	21:12
Air Blank	0.000	21:12

Agency: MARTIN COUNTY SO

Date/Time: 12/19/2018 21:15:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/19/2018 22:30

Air Blank	0.000	21:13
Control Test	0.076	21:14
Air Blank	0.000	21:15

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 12/19/2018 21:17:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/19/2018 22:30

Air Blank	0.000	21:16
Control Test	0.194	21:17
Air Blank	0.000	21:17

Agency: MARTIN COUNTY SO

Date/Time: 01/13/2019 21:59:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2019 23:12

Air Blank	0.000	21:58
Control Test	0.000	21:59
Air Blank	0.000	21:59

Agency: MARTIN COUNTY SO

Date/Time: 01/13/2019 22:01:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2019 23:12

Air Blank	0.000	22:00
Control Test	0.000	22:01
Air Blank	0.000	22:01

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 01/13/2019 22:03:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2019 23:12

Air Blank	0.000	22:02
Control Test	0.048	22:03
Air Blank	0.000	22:03

Agency: MARTIN COUNTY SO

Date/Time: 01/13/2019 22:05:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2019 23:12

Air Blank	0.000	22:04
Control Test	0.076	22:05
Air Blank	0.000	22:05

Agency: MARTIN COUNTY SO

Date/Time: 01/13/2019 22:07:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2019 23:12

Air Blank	0.000	22:06
Control Test	0.192	22:07
Air Blank	0.000	22:07

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

Agency: HOMESTEAD PD

Date/Time: 12/19/2018 08:41:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/19/2018 09:49

Air Blank	0.000	08:40
Control Test	0.046	08:40
Air Blank	0.000	08:41

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

Agency: OCALA POLICE DEPT

Date/Time: 01/09/2019 08:25:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/09/2019 10:50

Air Blank	0.000	08:24
Control Test	0.049	08:24
Air Blank	0.000	08:25

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006221

Agency: LAKE CITY PD

Date/Time: 06/08/2018 13:53:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2018 15:06

Air Blank	0.000	13:52
Control Test	0.000	13:53
Air Blank	0.000	13:53

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

Agency: BREVARD COUNTY S.O.

Date/Time: 09/20/2018 10:56:05

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/20/2018 14:24

Air Blank		10:55
Air Blank	0.000	10:56

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

Agency: PALM BEACH CO SO

Date/Time: 06/12/2018 08:19:10

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/22/2018 12:53

Air Blank	0.000	08:18
Control Test	0.078	08:18
Air Blank	0.000	08:19

Agency: PALM BEACH CO SO

Date/Time: 06/12/2018 08:20:58

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/22/2018 12:53

Air Blank	0.000	08:20
Control Test	0.078	08:20
Air Blank	0.000	08:20

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

Agency: PALM BEACH CO SO

Date/Time: 06/12/2018 08:32:03

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/22/2018 12:43

Air Blank	0.000	08:31
Control Test	0.064	08:31
Air Blank	0.000	08:32

Agency: PALM BEACH CO SO

Date/Time: 06/12/2018 08:35:28

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/22/2018 12:43

Air Blank	0.000	08:34
Control Test	0.070	08:34
Air Blank	0.000	08:35

Agency: PALM BEACH CO SO

Date/Time: 06/12/2018 08:37:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/22/2018 12:43

Air Blank	0.000	08:36
Control Test	0.073	08:36
Air Blank	0.000	08:37

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 07/18/2018 19:24:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank		19:23
Air Blank	0.000	19:24

Agency: FL HIGHWAY PATROL

Date/Time: 07/31/2018 19:48:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank	0.000	19:47
Control Test	0.046	19:47
Air Blank	0.000	19:48

Agency: FL HIGHWAY PATROL

Date/Time: 07/31/2018 19:50:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank	0.000	19:49
Control Test	0.074	19:50
Air Blank	0.000	19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 07/31/2018 19:52:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank	0.000	19:51
Control Test	0.186	19:52
Air Blank	0.000	19:52

Agency: FL HIGHWAY PATROL

Date/Time: 07/31/2018 19:54:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank	0.000	19:53
Control Test	0.018	19:54
Air Blank	0.000	19:54

Agency: FL HIGHWAY PATROL

Date/Time: 07/31/2018 20:41:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank	0.000	20:40
Control Test	0.046	20:41
Air Blank	0.000	20:41

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 07/31/2018 20:43:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank		20:42
Air Blank		20:43

Agency: FL HIGHWAY PATROL

Date/Time: 07/31/2018 20:46:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank	0.000	20:45
Control Test	0.076	20:45
Air Blank	0.000	20:46

Agency: FL HIGHWAY PATROL

Date/Time: 07/31/2018 20:48:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 21:54

Air Blank	0.000	20:47
Control Test	0.193	20:47
Air Blank	0.000	20:48

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

Agency: HENDRY COUNTY SO

Date/Time: 04/18/2018 22:10:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	22:09
Control Test	0.073	22:09
Air Blank	0.000	22:10

Agency: HENDRY COUNTY SO

Date/Time: 04/18/2018 22:19:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	22:17
Control Test	0.070	22:18
Air Blank	0.000	22:19

Agency: HENDRY COUNTY SO

Date/Time: 04/18/2018 22:51:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	22:50
Control Test	0.074	22:50
Air Blank	0.000	22:51

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

Agency: HENDRY COUNTY SO

Date/Time: 04/18/2018 23:08:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08

Agency: HENDRY COUNTY SO

Date/Time: 04/18/2018 23:17:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	23:15
Control Test	0.179	23:16
Air Blank	0.000	23:17

Agency: HENDRY COUNTY SO

Date/Time: 04/18/2018 23:34:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	23:33
Control Test	0.192	23:34
Air Blank	0.000	23:34

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

Agency: HENDRY COUNTY SO

Date/Time: 04/19/2018 00:22:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	00:21
Control Test	0.042	00:22
Air Blank	0.000	00:22

Agency: HENDRY COUNTY SO

Date/Time: 04/19/2018 00:36:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	00:35
Control Test	0.049	00:36
Air Blank	0.000	00:36

Agency: HENDRY COUNTY SO

Date/Time: 04/19/2018 00:47:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	00:45
Control Test	0.073	00:46
Air Blank	0.000	00:47

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

Agency: HENDRY COUNTY SO

Date/Time: 04/19/2018 01:01:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	01:00
Control Test	0.077	01:00
Air Blank	0.000	01:01

Agency: HENDRY COUNTY SO

Date/Time: 04/19/2018 01:18:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	01:17
Control Test	0.212	01:18
Air Blank	0.000	01:18

Agency: HENDRY COUNTY SO

Date/Time: 04/19/2018 01:22:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2018 02:19

Air Blank	0.000	01:21
Control Test	0.208	01:21
Air Blank	0.000	01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

Agency: FL HIGHWAY PATROL

Date/Time: 01/31/2018 12:31:15

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/31/2018 14:29

Air Blank	0.000	12:30
Control Test	0.082	12:30
Air Blank	0.000	12:31

Agency: FL HIGHWAY PATROL

Date/Time: 01/31/2018 12:47:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2018 14:29

Air Blank	0.000	12:46
Control Test	0.000	12:46
Air Blank	0.000	12:47

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2018 16:48:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2018 17:41

Air Blank	0.000	16:47
Control Test	0.000	16:48
Air Blank	0.000	16:48

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 01/30/2018 13:47:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/30/2018 14:40

Air Blank	0.000	13:46
Control Test	0.046	13:47
Air Blank	0.000	13:47

Agency: BOCA RATON PD

Date/Time: 07/27/2018 08:07:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2018 08:53

Air Blank	0.000	08:05
Control Test	0.076	08:06
Air Blank	0.000	08:07

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

Agency: FHP

Date/Time: 08/31/2018 17:06:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 21:48

Air Blank	0.000	17:05
Control Test	0.047	17:05
Air Blank	0.000	17:06

Agency: FHP

Date/Time: 08/31/2018 17:09:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 21:48

Air Blank	0.000	17:07
Control Test	0.192	17:08
Air Blank	0.000	17:09

Agency: FHP

Date/Time: 08/31/2018 17:11:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 21:48

Air Blank	0.000	17:10
Control Test	0.073	17:10
Air Blank	0.000	17:11

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

Agency: FHP

Date/Time: 08/31/2018 17:17:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 21:48

Air Blank	0.000	17:16
Control Test	0.074	17:17
Air Blank	0.000	17:17

Agency: FHP

Date/Time: 08/31/2018 17:20:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 21:48

Air Blank	0.000	17:18
Control Test	0.074	17:19
Air Blank	0.000	17:20

Agency: FHP

Date/Time: 08/31/2018 17:25:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 21:48

Air Blank	0.000	17:24
Control Test	0.075	17:24
Air Blank	0.000	17:25

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

Agency: FHP

Date/Time: 08/31/2018 18:24:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 21:48

Air Blank	0.000	18:23
Control Test	0.047	18:24
Air Blank	0.000	18:24

Agency: FHP

Date/Time: 08/31/2018 18:26:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/31/2018 21:48

Air Blank	0.000	18:25
Control Test	0.076	18:26
Air Blank	0.000	18:26

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

Agency: FHP

Date/Time: 07/31/2018 16:01:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 17:58

Air Blank	0.000	15:59
Control Test	0.195	16:00
Air Blank	0.000	16:01

Agency: FHP

Date/Time: 07/31/2018 16:04:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 17:58

Air Blank	0.000	16:02
Control Test	0.075	16:03
Air Blank	0.000	16:04

Agency: FHP

Date/Time: 07/31/2018 16:06:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 17:58

Air Blank	0.000	16:05
Control Test	0.022	16:06
Air Blank	0.000	16:06

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

Agency: FHP

Date/Time: 09/25/2018 14:27:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2018 16:14

Air Blank	0.000	14:26
Control Test	0.073	14:26
Air Blank	0.000	14:27

Agency: FHP

Date/Time: 09/25/2018 14:30:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2018 16:14

Air Blank	0.000	14:28
Control Test	0.045	14:29
Air Blank	0.000	14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Agency: FL HIGHWAY PATROL

Date/Time: 04/20/2018 03:40:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/26/2018 01:11

Air Blank	0.000	03:39
Control Test	0.046	03:40
Air Blank	0.000	03:40

Agency: FL HIGHWAY PATROL

Date/Time: 04/20/2018 03:42:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/26/2018 01:11

Air Blank	0.000	03:41
Control Test	0.075	03:41
Air Blank	0.000	03:42

Agency: FL HIGHWAY PATROL

Date/Time: 06/27/2018 13:42:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 16:03

Air Blank	0.000	13:41
Control Test	0.047	13:42
Air Blank	0.000	13:42

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Agency: FL HIGHWAY PATROL

Date/Time: 06/27/2018 13:44:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 16:03

Air Blank	0.000	13:43
Control Test	0.074	13:44
Air Blank	0.000	13:44

Agency: FL HIGHWAY PATROL

Date/Time: 06/27/2018 13:47:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 16:03

Air Blank	0.000	13:45
Control Test	0.194	13:46
Air Blank	0.000	13:47

Agency: FL HIGHWAY PATROL

Date/Time: 06/27/2018 13:51:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 16:03

Air Blank	0.000	13:50
Control Test	0.074	13:50
Air Blank	0.000	13:51

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Agency: FL HIGHWAY PATROL

Date/Time: 06/27/2018 13:56:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/27/2018 16:03

Air Blank	0.000	13:55
Control Test	0.075	13:55
Air Blank	0.000	13:56

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

Agency: FL HIGHWAY PATROL

Date/Time: 11/05/2018 10:18:37

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/05/2018 15:40

Air Blank	0.000	10:17
Control Test	0.079	10:18
Air Blank	0.000	10:18

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

Agency: FHP

Date/Time: 02/26/2018 21:14:11

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/26/2018 22:59

Air Blank	0.000	21:13
ITP Check	0.077	21:13
Air Blank	0.000	21:14

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

Agency: FHP TROOP G

Date/Time: 07/31/2018 20:47:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 22:33

Air Blank	0.000	20:46
Control Test	0.047	20:47
Air Blank	0.000	20:47

Agency: FHP TROOP G

Date/Time: 07/31/2018 20:49:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 22:33

Air Blank	0.000	20:48
Control Test	0.077	20:49
Air Blank	0.000	20:49

Agency: FHP TROOP G

Date/Time: 07/31/2018 20:51:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2018 22:33

Air Blank	0.000	20:50
Control Test	0.193	20:50
Air Blank	0.000	20:51

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

Agency: FHP TROOP G

Date/Time: 03/10/2018 00:37:48

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/03/2018 19:35

Air Blank	0.000	00:36
Control Test	0.038	00:37
Air Blank	0.000	00:37

Agency: FHP TROOP G

Date/Time: 11/21/2018 19:19:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/21/2018 20:00

Air Blank	0.000	19:18
Control Test	0.195	19:18
Air Blank	0.000	19:19

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

Agency: FL HIGHWAY PATROL

Date/Time: 05/01/2018 11:10:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/02/2018 16:43

Air Blank	0.000	11:09
Control Test	0.000	11:10
Air Blank	0.000	11:10

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

Agency: FL HIGHWAY PATROL

Date/Time: 05/01/2018 11:04:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/01/2018 12:01

Air Blank	0.000	11:03
Control Test	0.000	11:03
Air Blank	0.000	11:04

Agency: FL HIGHWAY PATROL

Date/Time: 06/05/2018 09:26:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/05/2018 11:02

Air Blank	0.000	09:25
Control Test	0.191	09:26
Air Blank	0.000	09:26

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 02/18/2018 13:28:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	13:27
Control Test	0.075	13:27
Air Blank	0.000	13:28

Agency: FT WALTON BEACH PD

Date/Time: 02/18/2018 13:31:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	13:30
Control Test	0.075	13:31
Air Blank	0.000	13:31

Agency: FT WALTON BEACH PD

Date/Time: 02/21/2018 11:47:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	11:46
Control Test	0.075	11:47
Air Blank	0.000	11:47

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 02/21/2018 11:52:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	11:51
Control Test	0.000	11:52
Air Blank	0.000	11:52

Agency: FT WALTON BEACH PD

Date/Time: 02/21/2018 12:06:10

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	12:05
ITP Check	0.082	12:05
Air Blank	0.000	12:06

Agency: FT WALTON BEACH PD

Date/Time: 02/21/2018 12:12:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	12:11
Control Test	0.046	12:11
Air Blank	0.000	12:12

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 02/21/2018 12:15:09

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	12:14
Control Test	0.079	12:14
Air Blank	0.000	12:15

Agency: FT WALTON BEACH PD

Date/Time: 02/21/2018 12:19:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	12:18
Control Test	0.000	12:19
Air Blank	0.000	12:19

Agency: FT WALTON BEACH PD

Date/Time: 02/21/2018 12:23:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2018 14:53

Air Blank	0.000	12:21
Control Test	0.197	12:22
Air Blank	0.000	12:23

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 03/31/2018 16:57:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/04/2018 13:49

Air Blank	0.000	16:56
Control Test	0.046	16:56
Air Blank	0.000	16:57

Agency: FT WALTON BEACH PD

Date/Time: 03/31/2018 16:59:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/04/2018 13:49

Air Blank	0.000	16:58
Control Test	0.046	16:59
Air Blank	0.000	16:59

Agency: FT WALTON BEACH PD

Date/Time: 03/31/2018 17:06:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/04/2018 13:49

Air Blank	0.000	17:05
Control Test	0.073	17:06
Air Blank	0.000	17:06

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 03/31/2018 17:28:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/04/2018 13:49

Air Blank	0.000	17:27
Control Test	0.072	17:27
Air Blank	0.000	17:28

Agency: FT WALTON BEACH PD

Date/Time: 03/31/2018 17:31:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/04/2018 13:49

Air Blank		17:30
Air Blank	0.000	17:31

Agency: FT WALTON BEACH PD

Date/Time: 03/31/2018 17:33:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/04/2018 13:49

Air Blank	0.000	17:31
Control Test	0.048	17:32
Air Blank	0.000	17:33

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 03/31/2018 17:35:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/04/2018 13:49

Air Blank	0.000	17:34
Control Test	0.195	17:35
Air Blank	0.000	17:35

Agency: FT WALTON BEACH PD

Date/Time: 03/31/2018 17:42:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/04/2018 13:49

Air Blank	0.000	17:40
Control Test	0.074	17:41
Air Blank	0.000	17:42

Agency: FT WALTON BEACH PD

Date/Time: 04/28/2018 08:30:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/02/2018 02:53

Air Blank	0.000	08:29
Control Test	0.000	08:30
Air Blank	0.000	08:30

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 04/28/2018 08:47:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/02/2018 02:53

Air Blank	0.000	08:45
Control Test	0.193	08:46
Air Blank	0.000	08:47

Agency: FT WALTON BEACH PD

Date/Time: 04/28/2018 09:03:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/02/2018 02:53

Air Blank	0.000	09:02
Control Test	0.074	09:02
Air Blank	0.000	09:03

Agency: FT WALTON BEACH PD

Date/Time: 05/31/2018 14:40:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/03/2018 05:21

Air Blank	0.000	14:39
Control Test	0.072	14:39
Air Blank	0.000	14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 05/31/2018 15:43:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/03/2018 05:21

Air Blank	0.000	15:42
Control Test	0.074	15:43
Air Blank	0.000	15:43

Agency: FT WALTON BEACH PD

Date/Time: 06/22/2018 21:18:19

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	21:17
ITP Check	0.081	21:17
Air Blank	0.000	21:18

Agency: FT WALTON BEACH PD

Date/Time: 06/22/2018 21:20:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	21:18
Control Test	0.000	21:19
Air Blank	0.000	21:20

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 06/22/2018 21:24:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank		21:24
Air Blank	0.000	21:24

Agency: FT WALTON BEACH PD

Date/Time: 06/24/2018 02:33:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	02:32
Control Test	0.000	02:33
Air Blank	0.000	02:33

Agency: FT WALTON BEACH PD

Date/Time: 06/24/2018 02:46:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	02:45
Control Test	0.071	02:45
Air Blank	0.000	02:46

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 06/24/2018 02:58:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	02:57
Control Test	0.000	02:58
Air Blank	0.000	02:58

Agency: FT WALTON BEACH PD

Date/Time: 06/24/2018 03:01:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	03:00
Control Test	0.000	03:01
Air Blank	0.000	03:01

Agency: FT WALTON BEACH PD

Date/Time: 06/24/2018 03:34:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	03:33
Control Test	0.044	03:34
Air Blank	0.000	03:34

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

Agency: FT WALTON BEACH PD

Date/Time: 06/24/2018 03:36:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	03:35
Control Test	0.044	03:36
Air Blank	0.000	03:36

Agency: FT WALTON BEACH PD

Date/Time: 06/24/2018 03:43:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/01/2018 19:25

Air Blank	0.000	03:42
Control Test	0.075	03:42
Air Blank	0.000	03:43

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

Agency: FORT LAUDERDALE PD

Date/Time: 10/17/2018 12:37:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/17/2018 13:43

Air Blank	0.000	12:36
Control Test	0.047	12:37
Air Blank	0.000	12:37

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

Agency: BROWARD COUNTY SO

Date/Time: 11/27/2018 14:28:01

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/27/2018 15:18

Air Blank	0.000	14:27
Control Test	0.080	14:27
Air Blank	0.000	14:28

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

Agency: CMI INC.

Date/Time: 02/20/2018 08:33:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/26/2018 11:01

Air Blank		08:32
Air Blank	0.000	08:33

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

Agency: COLLIER COUNTY SO

Date/Time: 07/03/2018 06:50:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/03/2018 08:36

Air Blank	0.000	06:49
Control Test	0.000	06:50
Air Blank	0.000	06:50

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

Agency: COLLIER COUNTY SO

Date/Time: 12/10/2018 11:18:31

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 01/02/2019 10:46

Air Blank	0.000	11:17
ITP Check	0.082	11:18
Air Blank	0.000	11:18

Agency: COLLIER COUNTY SO

Date/Time: 12/10/2018 11:20:03

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 01/02/2019 10:46

Air Blank	0.000	11:19
ITP Check	0.082	11:19
Air Blank	0.000	11:20

Agency: COLLIER COUNTY SO

Date/Time: 12/10/2018 11:21:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/02/2019 10:46

Air Blank	0.000	11:20
Control Test	0.081	11:20
Air Blank	0.000	11:21

**Florida Department of Law Enforcement
Alcohol Testing Program**

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

Agency: CHARLOTTE COUNTY
SO

Date/Time: 08/29/2018 10:18:01

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/20/2018 14:11

Air Blank	0.000	10:17
Control Test	0.000	10:17
Air Blank		10:17
Air Blank	0.000	10:18
