Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency	COLLIER COUNTY SO	**Date**	01/02/2018	**Type**	Agency Inspection
Inspector	NEIDERHISER,JOHN,E	**Time**	10:35:45	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	17316080A3
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	08/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT SPEC OPS

Data Download Date/Time 01/02/2018 11:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency	COLLIER COUNTY SO	**Date**	02/06/2018	**Type**	Agency Inspection
Inspector	NEIDERHISER,JOHN,E	**Time**	12:36:33	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	03417080A2
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT SPEC OPS

Data Download Date/Time 02/06/2018 12:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency	COLLIER COUNTY SO	**Date**	10/08/2018	**Type**	Agency Inspection
Inspector	REYES-RIVERA,DAVID,E	**Time**	08:23:22	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT DETERMINED AI NOT CONDUCTED

Data Download Date/Time 12/03/2018 09:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency COLLIER COUNTY SO **Date** 10/10/2018 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 10:20:32 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.083
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.084
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.084
0.20 g/210L Test # 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.084
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.084
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0009
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 12/03/2018 09:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	**Date**	01/30/2018	**Type**	Agency Inspection
Inspector	ABRAHAM,ERIC,D	**Time**	11:15:49	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	366849
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2018 12:34

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	**Date**	02/28/2018	**Type**	Agency Inspection
Inspector	ABRAHAM,ERIC,D	**Time**	14:47:42	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	366849
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2018 15:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	**Date**	03/30/2018	**Type**	Agency Inspection
Inspector	ABRAHAM,ERIC,D	**Time**	14:49:36	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	366849
Expiration	04/05/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/30/2018 15:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency PANAMA CITY PD **Date** 04/12/2018 **Type** Agency Inspection
Inspector ABRAHAM,ERIC,D **Time** 00:34:48 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 366849
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/12/2018 02:41

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency PANAMA CITY PD **Date** 05/28/2018 **Type** Agency Inspection
Inspector ABRAHAM,ERIC,D **Time** 09:37:37 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 366849
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # (Repeat) 1	0.076
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # (Repeat) 2	0.076
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # (Repeat) 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

08: Ambient Fail.

Data Download Date/Time 05/28/2018 10:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency PANAMA CITY PD **Date** 06/26/2018 **Type** Agency Inspection
Inspector ABRAHAM,ERIC,D **Time** 05:12:41 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 366849
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/26/2018 06:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency PANAMA CITY PD **Date** 07/06/2018 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 10:25:26 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.171
0.20 g/210L Test # (Repeat) 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.194
0.20 g/210L Test # (Repeat) 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.194
0.20 g/210L Test # (Repeat) 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.197
0.20 g/210L Test # (Repeat) 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.198
0.20 g/210L Test # (Repeat) 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.198
0.20 g/210L Test # (Repeat) 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.20 g/210L Test # (Repeat) 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.20 g/210L Test # (Repeat) 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.198
0.20 g/210L Test # (Repeat) 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.198
0.20 g/210L Test # (Repeat) 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0083
0.20 g/210L Test (Std Dev) (Repeat)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

20: Control Outside Tolerance SIM LEAK.

Data Download Date/Time 07/06/2018 11:35

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency PANAMA CITY PD **Date** 07/15/2018 **Type** Agency Inspection
Inspector ABRAHAM,ERIC,D **Time** 06:41:43 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 366849
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/15/2018 21:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	**Date**	08/26/2018	**Type**	Agency Inspection
Inspector	TURNER,JONATHAN,R	**Time**	20:26:58	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # 2	
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: RFI Detect. Non-compliance: .

Data Download Date/Time 08/28/2018 21:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	**Date**	08/26/2018	**Type**	Agency Inspection
Inspector	TURNER,JONATHAN,R	**Time**	20:31:05	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted. Non-compliance: .

Data Download Date/Time 08/28/2018 21:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	**Date**	08/26/2018	**Type**	Agency Inspection
Inspector	TURNER,JONATHAN,R	**Time**	21:07:35	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	366849
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2018 21:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency PANAMA CITY PD **Date** 09/30/2018 **Type** Agency Inspection
Inspector TURNER,JONATHAN,R **Time** 17:46:21 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.049
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.049
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.050
		0.08 g/210L Test # (Repeat) 3	0.049
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 10/03/2018 21:29

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	**Date**	09/30/2018	**Type**	Agency Inspection
Inspector	TURNER,JONATHAN,R	**Time**	18:40:46	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	366849
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/22/2020	**Expiration**	03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/03/2018 21:29

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	**Date**	10/26/2018	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	12:05:33	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	366849
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	03/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 10/26/2018 14:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency PANAMA CITY PD **Date** 11/30/2018 **Type** Agency Inspection
Inspector ABRAHAM,ERIC,D **Time** 07:06:43 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 366849
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/02/2018 09:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency PANAMA CITY PD **Date** 12/29/2018 **Type** Agency Inspection
Inspector ABRAHAM,ERIC,D **Time** 08:42:33 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 366849
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/30/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/30/2018 10:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	01/30/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	19:38:45	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	09117080A2
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.028
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 01/30/2018 20:44

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	02/27/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	20:11:02	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2018 21:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	03/27/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	20:59:48	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2018 22:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency NAS PENSACOLA PD **Date** 04/29/2018 **Type** Agency Inspection
Inspector HOGUE,CHARLES,D **Time** 20:06:11 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 03/07/2019 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.046
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.047
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.048
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.046
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 04/29/2018 21:55

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	04/29/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	20:54:11	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.189
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.193
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.193
		0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.20 g/210L Test # (Repeat) 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect. 20: Control Outside Tolerance.

Data Download Date/Time 04/29/2018 21:55

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	05/30/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	19:29:26	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2018 20:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency NAS PENSACOLA PD **Date** 06/29/2018 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 10:21:23 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 06/29/2018 11:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	07/31/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	19:46:38	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.056
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

05: Improper Sample.

Data Download Date/Time 07/31/2018 20:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	08/30/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	19:53:52	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/30/2018 21:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	09/28/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	20:06:35	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/28/2018 21:12

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	10/28/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	20:24:24	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/28/2018 21:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	11/29/2018	**Type**	Agency Inspection		
Inspector	HOGUE,CHARLES,D	**Time**	21:17:12	**Software**	8100.27		
0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2018 22:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency	NAS PENSACOLA PD	**Date**	12/27/2018	**Type**	Agency Inspection
Inspector	HOGUE,CHARLES,D	**Time**	19:41:30	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	09117080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/27/2018 20:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency	HENDRY COUNTY S.O.	**Date**	01/23/2018	**Type**	Agency Inspection
Inspector	NORWOOD,CHRISTOPHER,J	**Time**	22:54:35	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	83551
Expiration	03/08/2018	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/23/2018 22:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY S.O. **Date** 02/23/2018 **Type** Agency Inspection
Inspector NORWOOD,CHRISTOPHER,J **Time** 20:09:03 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 83551
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/23/2018 20:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency	HENDRY COUNTY S.O.	**Date**	03/24/2018	**Type**	Agency Inspection
Inspector	NORWOOD,CHRISTOPHER,J	**Time**	00:35:36	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	83551
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/24/2018 00:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency	HENDRY COUNTY S.O.	**Date**	04/18/2018	**Type**	Agency Inspection
Inspector	NORWOOD,CHRISTOPHER,J	**Time**	23:53:07	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	83551
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.082
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 04/18/2018 23:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY S.O. **Date** 04/25/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 10:46:45 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 04/25/2018 10:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency	HENDRY COUNTY S.O.	**Date**	05/09/2018	**Type**	Agency Inspection
Inspector	NORWOOD,CHRISTOPHER,J	**Time**	23:49:44	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	83551
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/09/2018 23:48

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY S.O. **Date** 06/25/2018 **Type** Agency Inspection
Inspector NORWOOD,CHRISTOPHER,J **Time** 21:16:40 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 83551
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/25/2018 21:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY S.O. **Date** 07/27/2018 **Type** Agency Inspection
Inspector NORWOOD,CHRISTOPHER,J **Time** 22:09:08 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 83551
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/27/2018 22:12

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency	HENDRY COUNTY S.O.	**Date**	08/22/2018	**Type**	Agency Inspection
Inspector	NORWOOD,CHRISTOPHER,J	**Time**	21:55:32	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	83551
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:DID NOT USE ENOUGH MOUTH ALCOHOL.

Data Download Date/Time 08/22/2018 21:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY S.O. **Date** 09/05/2018 **Type** Agency Inspection
Inspector NORWOOD,CHRISTOPHER,J **Time** 23:29:34 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 83551
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/02/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2018 00:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency	HENDRY COUNTY S.O.	**Date**	10/18/2018	**Type**	Agency Inspection
Inspector	NORWOOD,CHRISTOPHER,J	**Time**	23:17:20	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	437738
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/18/2018 23:15

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY S.O. **Date** 11/29/2018 **Type** Agency Inspection
Inspector NORWOOD,CHRISTOPHER,J **Time** 00:47:33 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 437738
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2018 00:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY S.O. **Date** 12/26/2018 **Type** Agency Inspection
Inspector NORWOOD,CHRISTOPHER,J **Time** 22:03:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 437738
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/26/2018 22:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY S.O. **Date** 01/06/2019 **Type** Agency Inspection
Inspector NORWOOD,CHRISTOPHER,J **Time** 22:03:23 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 437738
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/06/2019 22:08

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency LEE COUNTY SO **Date** 02/06/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 12:21:07 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG715202
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/06/2018 12:19

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency LEE COUNTY SO **Date** 02/14/2018 **Type** Agency Inspection
Inspector CLAFLIN,DAVID,S **Time** 10:08:50 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/14/2018 10:08

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency	LEE COUNTY SO	**Date**	03/07/2018	**Type**	Agency Inspection
Inspector	CLAFLIN,DAVID,S	**Time**	13:00:31	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	29017080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/07/2018 13:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency	TYNDALL AFB	**Date**	01/03/2018	**Type**	Agency Inspection
Inspector	LORD,LANNY,A	**Time**	11:29:59	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	32316080A3
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/03/2018 12:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency	TYNDALL AFB	**Date**	02/09/2018	**Type**	Agency Inspection
Inspector	TRACHTE,JOHN,E	**Time**	12:55:35	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	32316080A3
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	04/05/2018	**Expiration**	01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.199
		0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det: RFI Detect. REM RF DEVICE

Data Download Date/Time 02/09/2018 13:57

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency	TYNDALL AFB	**Date**	03/02/2018	**Type**	Agency Inspection
Inspector	LORD,LANNY,A	**Time**	09:16:13	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	32316080A3
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.050
		0.08 g/210L Test # (Repeat) 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance PUT ON CORRECT SOLUTIONS . PUT IN CORRECT SOLUTION

Data Download Date/Time 03/02/2018 10:19

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency	TYNDALL AFB	**Date**	04/04/2018	**Type**	Agency Inspection
Inspector	TRACHTE,JOHN,E	**Time**	09:07:23	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 04/04/2018 10:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency	TYNDALL AFB	**Date**	04/04/2018	**Type**	Agency Inspection
Inspector	TRACHTE,JOHN,E	**Time**	09:51:59	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	32316080A3
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/04/2018 10:57

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB **Date** 05/10/2018 **Type** Agency Inspection
Inspector TRACHTE,JOHN,E **Time** 14:16:38 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/10/2018 15:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB **Date** 06/22/2018 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 08:43:06 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/22/2018 16:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB **Date** 07/11/2018 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 16:28:46 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2018 17:34

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB **Date** 08/01/2018 **Type** Agency Inspection
Inspector TRACHTE,JOHN,E **Time** 09:07:58 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: .RETESTED

Data Download Date/Time 08/01/2018 10:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency	TYNDALL AFB	**Date**	09/07/2018	**Type**	Agency Inspection
Inspector	LORD,LANNY,A	**Time**	08:54:04	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	32316080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: .

Data Download Date/Time 09/07/2018 10:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB **Date** 10/05/2018 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 11:32:08 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 10/05/2018 12:37

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency	TYNDALL AFB	**Date**	11/06/2018	**Type**	Agency Inspection
Inspector	BARGE,SUSAN,M	**Time**	13:45:22	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	770784
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 11/06/2018 14:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB **Date** 12/04/2018 **Type** Agency Inspection
Inspector HERRIGES,JASON,J **Time** 09:30:19 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 770784
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 12/04/2018 10:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB **Date** 01/07/2019 **Type** Agency Inspection
Inspector LORD,LANNY,A **Time** 10:02:07 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 770784
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/07/2019 11:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency	AUBURNDALE PD	**Date**	01/30/2018	**Type**	Agency Inspection
Inspector	OROS,DAVID,A	**Time**	17:37:48	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	03417080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2018 17:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency	AUBURNDALE PD	**Date**	03/06/2018	**Type**	Agency Inspection
Inspector	BELL,DANIELLE,M	**Time**	11:07:21	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

Data Download Date/Time 03/14/2018 09:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency	AUBURNDALE PD	**Date**	03/14/2018	**Type**	Department Inspection		
Inspector	BELL,DANIELLE,M	**Time**	14:13:29	**Software**	8100.27		
0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG715202
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/14/2018 14:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency	AUBURNDALE PD	**Date**	03/20/2018	**Type**	Agency Inspection
Inspector	LOKIETEK,THOMAS,J	**Time**	02:06:51	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	03417080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/20/2018 02:08

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency	AUBURNDALE PD	**Date**	04/12/2018	**Type**	Agency Inspection
Inspector	LOKIETEK,THOMAS,J	**Time**	02:26:16	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	03417080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/12/2018 02:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency	AUBURNDALE PD	**Date**	05/18/2018	**Type**	Agency Inspection
Inspector	LOKIETEK,THOMAS,J	**Time**	22:18:09	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	03417080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.072
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 05/18/2018 22:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency AUBURNDALE PD **Date** 06/26/2018 **Type** Agency Inspection
Inspector LOKIETEK,THOMAS,J **Time** 03:09:09 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/26/2018 03:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency AUBURNDALE PD **Date** 07/26/2018 **Type** Agency Inspection
Inspector OROS,DAVID,A **Time** 11:00:22 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2018 11:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency AUBURNDALE PD **Date** 08/25/2018 **Type** Agency Inspection
Inspector LOKIETEK,THOMAS,J **Time** 02:05:18 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/25/2018 02:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency AUBURNDALE PD **Date** 09/24/2018 **Type** Agency Inspection
Inspector LOKIETEK,THOMAS,J **Time** 01:00:30 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/24/2018 01:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency AUBURNDALE PD **Date** 10/25/2018 **Type** Agency Inspection
Inspector OROS,DAVID,A **Time** 09:54:24 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2018 09:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency AUBURNDALE PD **Date** 11/21/2018 **Type** Agency Inspection
Inspector LOKIETEK,THOMAS,J **Time** 13:30:24 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/21/2018 13:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency AUBURNDALE PD **Date** 12/20/2018 **Type** Agency Inspection
Inspector OROS,DAVID,A **Time** 10:04:26 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/20/2018 10:11

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	01/19/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	16:26:56	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	20016080A1
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/19/2018 16:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	02/02/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	17:02:39	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12317080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2018 17:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	03/09/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	16:32:07	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12317080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/09/2018 16:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	04/13/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	15:39:29	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12317080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/13/2018 15:41

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency PALM BEACH CO SO **Date** 05/15/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 13:23:01 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.083
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 05/15/2018 13:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	06/22/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	15:00:20	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A5
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/22/2018 15:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	07/13/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:18:44	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A5
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/13/2018 12:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	08/10/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:10:52	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A5
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/10/2018 12:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	09/07/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:48:57	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A5
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/07/2018 12:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	10/19/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:34:24	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A5
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/19/2018 12:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	11/23/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:41:08	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	22817080A5
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/23/2018 12:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

Agency	PALM BEACH CO SO	**Date**	12/14/2018	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	13:08:40	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	22817080A5
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/14/2018 13:19

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	**Date**	01/29/2018	**Type**	Agency Inspection
Inspector	SECORD,RYAN,T	**Time**	09:39:28	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	152169
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	10/30/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2018 09:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	**Date**	03/01/2018	**Type**	Agency Inspection
Inspector	SECORD,RYAN,T	**Time**	07:07:44	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	152169
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	10/30/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/01/2018 07:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	**Date**	04/29/2018	**Type**	Agency Inspection
Inspector	SECORD,RYAN,T	**Time**	08:06:52	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	152169
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/30/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 04/29/2018 08:00

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	**Date**	05/30/2018	**Type**	Agency Inspection
Inspector	SECORD,RYAN,T	**Time**	08:57:54	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	152169
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/30/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2018 08:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	**Date**	06/23/2018	**Type**	Agency Inspection
Inspector	SECORD,RYAN,T	**Time**	09:42:23	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	152169
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/30/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/23/2018 09:38

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency WEST PALM BEACH PD **Date** 08/24/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 11:33:00 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 08/24/2018 11:38

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	**Date**	09/19/2018	**Type**	Agency Inspection		
Inspector	SECORD,RYAN,T	**Time**	11:51:09	**Software**	8100.27		
0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	152169
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/30/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/19/2018 11:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency WEST PALM BEACH PD **Date** 10/28/2018 **Type** Agency Inspection
Inspector SECORD,RYAN,T **Time** 08:43:14 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 152169
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/28/2018 08:57

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	**Date**	11/29/2018	**Type**	Agency Inspection
Inspector	SECORD,RYAN,T	**Time**	07:36:32	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	24818080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2018 07:44

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	**Date**	12/31/2018	**Type**	Agency Inspection
Inspector	SECORD,RYAN,T	**Time**	07:20:04	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	24818080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/31/2018 07:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency	POLK COUNTY SO	**Date**	01/20/2018	**Type**	Agency Inspection
Inspector	WIEGERT,KIMBRA,L	**Time**	19:27:00	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	20016080A1
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail FAN NOT ON, ROOM CLEARED.

Data Download Date/Time 01/20/2018 19:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 02/28/2018 **Type** Agency Inspection
Inspector WIEGERT,KIMBRA,L **Time** 13:31:20 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail FAN NOT ON.

Data Download Date/Time 02/28/2018 13:38

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 03/26/2018 **Type** Agency Inspection
Inspector WIEGERT,KIMBRA,L **Time** 11:17:57 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 2281708OA4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/26/2018 11:29

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 04/16/2018 **Type** Department Inspection
Inspector BELL,DANIELLE,M **Time** 11:55:23 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 04/16/2018 12:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 04/30/2018 **Type** Agency Inspection
Inspector HAMMONDS,TEAQULIA,L **Time** 11:04:55 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SGT WIEGERT PRESENT

Data Download Date/Time 04/30/2018 11:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 05/22/2018 **Type** Agency Inspection
Inspector HAMMONDS,TEAQULIA,L **Time** 15:35:44 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:TO MUCH MOUTH ALCOHOL.SGT. WIEGERT PRESENT

Data Download Date/Time 05/22/2018 16:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 06/27/2018 **Type** Agency Inspection
Inspector HAMMONDS,TEAQULIA,L **Time** 14:01:46 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2018 15:40

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency	POLK COUNTY SO	**Date**	07/27/2018	**Type**	Agency Inspection
Inspector	HAMMONDS,TEAQULIA,L	**Time**	07:40:04	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/27/2018 10:12

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency	POLK COUNTY SO	**Date**	08/29/2018	**Type**	Agency Inspection
Inspector	HAMMONDS,TEAQULIA,L	**Time**	12:52:23	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	22817080A4
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2018 13:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 09/28/2018 **Type** Agency Inspection
Inspector HAMMONDS,TEAQULIA,L **Time** 10:41:18 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/28/2018 11:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 10/27/2018 **Type** Agency Inspection
Inspector HAMMONDS,TEAQULIA,L **Time** 12:01:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/27/2018 13:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 11/28/2018 **Type** Agency Inspection
Inspector HAMMONDS,TEAQULIA,L **Time** 07:28:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

A F / M A:COUGHING DURING BLOW.

Data Download Date/Time 11/28/2018 09:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency POLK COUNTY SO **Date** 12/28/2018 **Type** Agency Inspection
Inspector HAMMONDS,TEAQULIA,L **Time** 07:45:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/28/2018 09:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	01/04/2018	**Type**	Agency Inspection
Inspector	CHESSHER,PITCHER,/	**Time**	18:01:02	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12716080A7
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/04/2018 17:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	02/01/2018	**Type**	Agency Inspection
Inspector	PITCHER,CHESSHER,/	**Time**	19:05:14	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12716080A7
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2018 19:34

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	03/09/2018	**Type**	Agency Inspection
Inspector	CHESSHER,PITCHER,/	**Time**	18:36:10	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12716080A7
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/09/2018 18:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	04/07/2018	**Type**	Agency Inspection
Inspector	PITCHER,CHESSHER,/	**Time**	22:38:22	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12716080A7
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/07/2018 22:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	04/11/2018	**Type**	Department Inspection		
Inspector	BELL,DANIELLE,M	**Time**	16:18:19	**Software**	8100.27		
0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805701
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/11/2018 16:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency LAKE COUNTY SO **Date** 04/26/2018 **Type** Agency Inspection
Inspector CHESSHER,BINDER, **Time** 22:04:55 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A7
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.073
0.08 g/210L Test # (Repeat) 1	0.078
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.075
0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.076
0.08 g/210L Test # (Repeat) 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed. 08: Control Outside Tolerance.

Data Download Date/Time 04/26/2018 22:08

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency LAKE COUNTY SO **Date** 05/03/2018 **Type** Agency Inspection
Inspector PITCHER,BINDER,/ **Time** 18:08:22 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A7
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/03/2018 18:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	06/07/2018	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	22:09:44	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12716080A7
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	06/05/2018

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

DRY GASS EXPIRED.

Data Download Date/Time 06/07/2018 23:07

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	06/07/2018	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	22:53:45	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	814181
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/07/2018 23:07

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	07/06/2018	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:38:10	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	814181
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/06/2018 23:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	08/15/2018	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:21:42	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	814181
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/15/2018 23:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	09/04/2018	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	12:39:04	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	814181
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/04/2018 12:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	10/18/2018	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	20:45:40	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	814181
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded/RINSED MOUTH .

Data Download Date/Time 10/18/2018 20:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	11/15/2018	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	20:27:24	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/15/2018 20:35

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency LAKE COUNTY SO **Date** 12/06/2018 **Type** Agency Inspection
Inspector BINDER,PITCHER,/ **Time** 00:33:37 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 814181
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/06/2018 00:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency	LAKE COUNTY SO	**Date**	01/01/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	20:55:21	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/01/2019 21:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 01/31/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 23:46:29 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 04216080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/03/2018 21:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 02/28/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 20:35:55 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 04216080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.08 g/210L Test # (Repeat) 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.075
0.08 g/210L Test # (Repeat) 2	0.076
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.072
0.08 g/210L Test # (Repeat) 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

SMALL AIR LEAK IN TUBING/FIX.

Data Download Date/Time 02/28/2018 20:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 03/04/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 23:54:19 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 04216080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/05/2018 00:08

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 04/29/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 02:22:23 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 906364
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.027
0.05 g/210L Test # (Repeat) 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.05 g/210L Test # (Repeat) 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

LOWER TUBING WAS LEAKING/REPLACED AFTER 1ST READING..

Data Download Date/Time 04/29/2018 02:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 05/22/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 05:19:51 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 906364
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/22/2018 05:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 06/20/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 02:38:22 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 906364
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2018 02:40

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 08/03/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 20:34:44 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 906364
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.074
0.08 g/210L Test # (Repeat) 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.08 g/210L Test # (Repeat) 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

SIM JAR WAS NOT TIGHT, CORRECTED AFTER 1ST SAMPLE.

Data Download Date/Time 08/03/2018 20:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 08/31/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 22:21:11 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 906364
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	No
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.048
0.08 g/210L Test # (Repeat) 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.048
0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.08 g/210L Test # (Repeat) 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

SIMULATOR JAR NOT TURNED TIGHT, AND I WAS CALLED AWAY .

Data Download Date/Time 08/31/2018 22:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 09/26/2018 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 12:02:08 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/26/2018 12:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 10/01/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 21:50:11 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 906364
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.000
0.05 g/210L Test # (Repeat) 1	0.045
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.041
0.05 g/210L Test # (Repeat) 2	0.045
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.039
0.05 g/210L Test # (Repeat) 3	0.046
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

GASKET ON SIM WAS CUT/AIR LEAK.

Data Download Date/Time 10/01/2018 22:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 11/28/2018 **Type** Agency Inspection
Inspector MURPHEY SR,DAVID,G **Time** 04:34:41 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 906364
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/28/2018 04:44

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency	DAYTONA BEACH SHORES	**Date**	12/09/2018	**Type**	Agency Inspection
Inspector	MURPHEY SR,DAVID,G	**Time**	14:25:52	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	906364
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/09/2018 14:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency	EDGEWATER P.D.	**Date**	01/21/2018	**Type**	Agency Inspection
Inspector	GINTZ,BRIAN,W	**Time**	10:05:56	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	765462
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/21/2018 10:07

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 02/18/2018 **Type** Agency Inspection
Inspector GINTZ,BRIAN,W **Time** 09:23:03 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 765462
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2018 09:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 03/31/2018 **Type** Agency Inspection
Inspector GINTZ,BRIAN,W **Time** 09:01:55 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 765462
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/31/2018 09:07

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 04/27/2018 **Type** Agency Inspection
Inspector GINTZ,BRIAN,W **Time** 03:41:20 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 765462
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/27/2018 03:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 05/30/2018 **Type** Agency Inspection
Inspector GINTZ,BRIAN,W **Time** 23:22:07 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 765462
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2018 23:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 06/30/2018 **Type** Agency Inspection
Inspector GINTZ,BRIAN,W **Time** 14:28:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201109A **0.20 Lot #** 201102B **0.08 Gas Lot #** 765462
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2018 14:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency	EDGEWATER P.D.	**Date**	06/30/2018	**Type**	Agency Inspection
Inspector	GABOURY,SCOTT,E	**Time**	19:18:24	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	765462
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2018 19:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 07/30/2018 **Type** Agency Inspection
Inspector GABOURY,SCOTT,E **Time** 02:56:54 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 765462
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/18/2018 23:41

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 08/29/2018 **Type** Agency Inspection
Inspector GINTZ,BRIAN,W **Time** 01:40:35 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 765462
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2018 01:44

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency	EDGEWATER P.D.	**Date**	09/06/2018	**Type**	Department Inspection
Inspector	MURPHY,PATRICK,J	**Time**	11:52:55	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805701
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 09/06/2018 12:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 09/10/2018 **Type** Agency Inspection
Inspector GABOURY,SCOTT,E **Time** 05:58:16 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 765462
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/10/2018 06:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency	EDGEWATER P.D.	**Date**	10/21/2018	**Type**	Agency Inspection
Inspector	GINTZ,BRIAN,W	**Time**	13:48:57	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	765462
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

NOT MIXED.

Data Download Date/Time 10/21/2018 13:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER P.D. **Date** 11/18/2018 **Type** Agency Inspection
Inspector GINTZ,BRIAN,W **Time** 09:40:25 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 765462
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/18/2018 09:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency	EDGEWATER P.D.	**Date**	12/24/2018	**Type**	Agency Inspection
Inspector	GINTZ,BRIAN,W	**Time**	10:46:58	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	765462
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	04/13/2020

Number of Simulators 2 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.185
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.188
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.190
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 12/24/2018 10:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency	PUTNAM COUNTY SO	**Date**	07/23/2018	**Type**	Agency Inspection
Inspector	SHANAHAN,JAKE,L	**Time**	07:30:33	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 07/23/2018 10:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency PUTNAM COUNTY SO **Date** 07/23/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 10:01:18 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 07/23/2018 10:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency	PUTNAM COUNTY SO	**Date**	11/07/2018	**Type**	Agency Inspection
Inspector	ROBERTS,KENNETH,L	**Time**	15:56:37	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Data Download Date/Time 11/07/2018 16:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency	PUTNAM COUNTY SO	**Date**	11/07/2018	**Type**	Agency Inspection
Inspector	ROBERTS,KENNETH,L	**Time**	16:49:24	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	29017080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/07/2018 16:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency	PUTNAM COUNTY SO	**Date**	12/10/2018	**Type**	Agency Inspection
Inspector	ROBERTS,KENNETH,L	**Time**	20:26:54	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	29017080A3
Expiration	03/07/2019	**Expiration**	12/15/2018	**Expiration**	02/23/2019	**Expiration**	12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/10/2018 20:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency	PUTNAM COUNTY SO	**Date**	01/09/2019	**Type**	Agency Inspection
Inspector	ROBERTS,KENNETH,L	**Time**	14:39:14	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201808E	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	29017080A3
Expiration	03/07/2019	**Expiration**	08/22/2020	**Expiration**	02/23/2019	**Expiration**	12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed. 00: Ambient Fail.

Data Download Date/Time 01/09/2019 15:00

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 01/31/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 13:56:52 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 609878
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2018 08:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 02/28/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 15:48:52 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 609878
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	0.000
Interferent Detect Test # (Repeat) 3	INT

Remarks:

SIMULATOR NOT CONNECTED.

Data Download Date/Time 02/28/2018 15:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 03/30/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 18:48:51 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 609878
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.181
0.20 g/210L Test # (Repeat) 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.183
0.20 g/210L Test # (Repeat) 2	0.192
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.184
0.20 g/210L Test # (Repeat) 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 03/30/2018 18:55

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency	PORT ORANGE P.D.	**Date**	04/30/2018	**Type**	Agency Inspection
Inspector	HARRISON,WILLIAM,C	**Time**	12:38:57	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	609878
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det: .

Data Download Date/Time 04/30/2018 12:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 05/31/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 15:41:53 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 609878
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2018 15:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 06/29/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 15:26:48 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 609878
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/06/2018 08:46

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 07/31/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 13:40:24 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 609878
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/11/2018 01:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 08/31/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 18:17:33 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 609878
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/05/2018 10:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 09/30/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 16:06:13 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 609878
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/05/2018 18:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 10/30/2018 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 10:58:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 609878
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/30/2018 11:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency PORT ORANGE P.D. **Date** 11/14/2018 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 12:40:33 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 11/14/2018 12:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 01/02/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 02:19:49 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 1824329
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/09/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/02/2018 02:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 02/03/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 04:30:56 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 1824329
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/09/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/03/2018 04:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency	SOUTH DAYTONA P.D.	**Date**	03/04/2018	**Type**	Agency Inspection
Inspector	MELODY,FRANCIS,D	**Time**	19:50:02	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	894572
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	01/17/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/04/2018 19:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 04/14/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 20:50:33 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 894572
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/17/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/14/2018 20:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 05/17/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 23:05:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 894572
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/17/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/17/2018 23:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 07/02/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 14:34:53 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 937240
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/07/2021

Number of Simulators 4 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.000
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.072
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.000
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.000
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:DRY GAS OUTSIDE CONTROL NORMS.

Data Download Date/Time 07/02/2018 14:37

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 07/27/2018 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 10:39:47 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 07/27/2018 10:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 09/18/2018 **Type** Agency Inspection
Inspector MURPHY,PATRICK,J **Time** 12:38:04 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SKIPPED AI TO OPERATE INSTRUMENT

Data Download Date/Time 09/23/2018 12:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 09/23/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 12:07:43 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 980558
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/09/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 09/23/2018 12:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 10/27/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 04:17:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 980558
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/09/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 10/27/2018 04:29

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 11/26/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 07:17:59 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 980558
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/09/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/26/2018 07:23

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA P.D. **Date** 12/16/2018 **Type** Agency Inspection
Inspector MELODY,FRANCIS,D **Time** 18:10:39 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 980558
Expiration 02/06/2018 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/09/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/16/2018 18:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	01/10/2018	**Type**	Agency Inspection
Inspector	WASHINGTON,YVETTE,D	**Time**	13:38:46	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	572920
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER

Data Download Date/Time 01/10/2018 13:41

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency BROWARD COUNTY SO **Date** 01/17/2018 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 14:51:07 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG626604
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/17/2018 14:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency BROWARD COUNTY SO **Date** 02/07/2018 **Type** Agency Inspection
Inspector WASHINGTON,YVETTE,D **Time** 13:46:49 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 572920
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

A FRAZIER OBSERVED INSPECTION

Data Download Date/Time 02/07/2018 13:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	03/01/2018	**Type**	Agency Inspection
Inspector	WASHINGTON,YVETTE,D	**Time**	20:00:15	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

UNKNOWN REASONNon-compliance:UNK DIAGNOSTIC FAILURE.

Data Download Date/Time 03/01/2018 20:36

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	03/01/2018	**Type**	Agency Inspection
Inspector	WASHINGTON,YVETTE,D	**Time**	20:36:03	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	572920
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A. FRAZIER PRESENT & OBSERVED

Data Download Date/Time 03/01/2018 20:36

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	04/26/2018	**Type**	Agency Inspection
Inspector	WASHINGTON,YVETTE,D	**Time**	14:15:41	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	572920
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER PRESENT & OBSERVED

Data Download Date/Time 04/26/2018 14:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	05/10/2018	**Type**	Agency Inspection
Inspector	WASHINGTON,YVETTE,D	**Time**	17:35:06	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	572920
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER OBSERVED INSPECTIONS

Data Download Date/Time 05/10/2018 17:35

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	06/21/2018	**Type**	Agency Inspection
Inspector	WASHINGTON,YVETTE,D	**Time**	12:31:13	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	572920
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A. FRAZIER PRESENT & OBSERVED

Data Download Date/Time 06/21/2018 12:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	07/13/2018	**Type**	Agency Inspection
Inspector	WASHINGTON,YVETTE,D	**Time**	15:13:02	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	572920
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:ALCOHOL RESIDUE.A FRAZIER OBS INSPECTION

Data Download Date/Time 07/13/2018 15:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency BROWARD COUNTY SO **Date** 08/29/2018 **Type** Agency Inspection
Inspector WASHINGTON,YVETTE,D **Time** 14:32:08 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 572920
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER OBSERVED

Data Download Date/Time 08/29/2018 14:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	09/20/2018	**Type**	Agency Inspection
Inspector	CAMPBELL,STUART,	**Time**	20:16:27	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	572920
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	08/29/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.

Data Download Date/Time 09/20/2018 20:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency BROWARD COUNTY SO **Date** 10/17/2018 **Type** Agency Inspection
Inspector WASHINGTON,YVETTE,D **Time** 17:23:08 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 572920
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER PRESENT & OBSERVED

Data Download Date/Time 10/17/2018 17:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001250

Agency	BROWARD COUNTY SO	**Date**	12/26/2018	**Type**	Agency Inspection
Inspector	WASHINGTON,YVETTE,D	**Time**	15:12:29	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	572920
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/13/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER OBSERVED INSPECTION

Data Download Date/Time 12/26/2018 15:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 01/11/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 20:11:59 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A5
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/11/2018 21:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 02/28/2018 **Type** Agency Inspection
Inspector WAGNER,ROBERT,F **Time** 17:36:10 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A5
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/05/2018 09:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 03/29/2018 **Type** Agency Inspection
Inspector MOORE,KIMBERLY,N **Time** 17:18:36 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A5
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2018 19:07

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency	BREVARD COUNTY S.O.	**Date**	05/10/2018	**Type**	Agency Inspection
Inspector	PLATT,SHAYLA,D	**Time**	14:13:23	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 05/15/2018 11:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 05/15/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 11:13:33 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 05/15/2018 11:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 05/23/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 18:11:50 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/23/2018 18:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 06/13/2018 **Type** Agency Inspection
Inspector GARRITY,THYOMAS,A **Time** 12:34:13 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/13/2018 12:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 07/26/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 17:43:25 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2018 19:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency	BREVARD COUNTY S.O.	**Date**	08/23/2018	**Type**	Agency Inspection
Inspector	GARRITY,THOMAS,A	**Time**	20:47:21	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	19817080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/23/2018 20:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 09/28/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 18:38:08 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/28/2018 18:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 10/25/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 18:30:31 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 19817080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2018 18:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 11/30/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 15:10:13 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 19817080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/30/2018 15:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency	BREVARD COUNTY S.O.	**Date**	12/27/2018	**Type**	Agency Inspection
Inspector	GARRITY,THOMAS,A	**Time**	16:53:41	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/27/2018 17:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY S.O. **Date** 01/09/2019 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 16:58:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 19817080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/09/2019 17:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	01/17/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:48:59	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	686331
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

KV INTOX

Data Download Date/Time 01/17/2018 14:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	02/02/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	15:05:20	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	686331
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

KV INTOX

Data Download Date/Time 02/02/2018 15:08

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	03/02/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	15:01:58	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	686331
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TAKEN OUT OF SERVICE IN KV DUE TO FDLE DEPT INSPECTION

Data Download Date/Time 03/02/2018 15:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency MONROE COUNTY S.O. **Date** 03/07/2018 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 10:10:11 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG626605
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/07/2018 10:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	04/04/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	15:17:16	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	686331
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE

Data Download Date/Time 04/04/2018 15:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	05/04/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:18:59	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	823252
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE

Data Download Date/Time 05/04/2018 14:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	06/08/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:15:00	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	686331
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 06/08/2018 14:19

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	07/13/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	13:19:14	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	686331
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 07/13/2018 13:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	08/01/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:40:10	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	686331
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 08/01/2018 14:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency MONROE COUNTY S.O. **Date** 09/05/2018 **Type** Agency Inspection
Inspector KEITH,JASON,T **Time** 15:31:55 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 686331
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 09/05/2018 15:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	10/03/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:50:24	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	686331
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	01/31/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 10/03/2018 14:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	11/07/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:05:44	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	823252
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 11/07/2018 14:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	12/19/2018	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	12:25:35	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	823252
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 12/19/2018 12:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	01/02/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	16:59:50	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE. PK. PRE-EVIDENTIARY AI COMPLETED.

Data Download Date/Time 01/02/2019 17:07

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency	MONROE COUNTY S.O.	**Date**	01/09/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:26:53	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

PK. SPARE. POST-EVIDENTIARY AI COMPLETED.

Data Download Date/Time 01/09/2019 14:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 01/17/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:11:16 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A1
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2018 11:12

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 02/13/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:59:25 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A1
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2018 11:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 03/27/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 13:30:25 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2018 13:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 04/17/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:02:34 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 12:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 05/08/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 09:56:21 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2018 10:03

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 06/12/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:17:10 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/12/2018 12:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 07/19/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:17:43 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2018 10:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 08/28/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:04:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2018 11:12

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 09/20/2018 **Type** Agency Inspection
Inspector MILLS,ARMAND,G **Time** 11:24:48 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/20/2018 11:44

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 10/15/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 09:34:40 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 10/15/2018 09:37

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency	ORANGE COUNTY S.O.	**Date**	10/24/2018	**Type**	Agency Inspection
Inspector	MILLS,ARMAND,G	**Time**	14:13:48	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13917080A3
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/24/2018 14:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency	ORANGE COUNTY S.O.	**Date**	11/07/2018	**Type**	Agency Inspection
Inspector	MILLS,ARMAND,G	**Time**	14:49:12	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13917080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS PREPARED BY AI MELVILLE

Data Download Date/Time 11/07/2018 14:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency ORANGE COUNTY S.O. **Date** 12/05/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:35:23 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2018 12:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency	ORANGE COUNTY S.O.	**Date**	01/17/2018	**Type**	Agency Inspection
Inspector	MELVILLE,KELLY,M	**Time**	09:58:15	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	26316080A1
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2018 09:57

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 02/13/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:01:24 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A1
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2018 12:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 03/27/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:31:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A1
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2018 10:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 04/17/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:11:43 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 11:20

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 05/08/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:30:08 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2018 12:34

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 06/12/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:27:49 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/12/2018 14:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 07/19/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:21:52 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2018 10:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency	ORANGE COUNTY S.O.	**Date**	08/28/2018	**Type**	Agency Inspection
Inspector	MELVILLE,KELLY,M	**Time**	11:08:50	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	13917080A3
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2018 11:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 09/20/2018 **Type** Agency Inspection
Inspector MILLS,ARMAND,G **Time** 11:30:59 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/20/2018 11:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 10/16/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 10:37:05 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 10/16/2018 10:41

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 10/24/2018 **Type** Agency Inspection
Inspector MILLS,ARMAND,G **Time** 14:17:42 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/24/2018 14:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency	ORANGE COUNTY S.O.	**Date**	11/07/2018	**Type**	Agency Inspection
Inspector	MILLS,ARMAND,G	**Time**	14:49:36	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13917080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS PREPARED BY AI MELVILLE

Data Download Date/Time 11/07/2018 14:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency ORANGE COUNTY S.O. **Date** 12/05/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:06:56 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2018 11:15

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency	ORANGE COUNTY S.O.	**Date**	01/17/2018	**Type**	Agency Inspection
Inspector	MELVILLE,KELLY,M	**Time**	12:02:15	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	26316080A1
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2018 12:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 02/13/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:04:38 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A1
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2018 11:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 03/27/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:49:59 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2018 11:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 04/17/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:04:37 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 11:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 05/08/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:01:18 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2018 10:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 06/12/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 09:53:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/12/2018 09:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 07/19/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 13:07:56 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2018 13:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 08/08/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 11:13:23 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 08/08/2018 11:19

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 08/28/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:17:57 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2018 12:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 09/17/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:15:28 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/17/2018 10:20

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 10/09/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:12:07 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/09/2018 11:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 11/07/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:46:39 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded.

Data Download Date/Time 11/07/2018 12:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY S.O. **Date** 12/05/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:40:46 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2018 12:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency	ORANGE COUNTY S.O.	**Date**	03/08/2018	**Type**	Agency Inspection
Inspector	SHANAHAN,JAKE,L	**Time**	07:57:04	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 03/08/2018 10:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 03/08/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 10:22:40 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG715202
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/08/2018 10:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency	ORANGE COUNTY S.O.	**Date**	04/17/2018	**Type**	Agency Inspection
Inspector	MELVILLE,KELLY,M	**Time**	09:49:37	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	26316080A1
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 09:55

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 05/08/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:18:49 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 201702B
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/23/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2018 11:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 06/12/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 09:48:32 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/12/2018 10:20

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 07/19/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 13:12:46 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2018 13:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 08/28/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 09:56:05 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 13917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2018 10:00

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 09/17/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:20:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 26316080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/17/2018 11:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 10/09/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:05:45 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 26316080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/09/2018 10:11

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 11/07/2018 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:31:50 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13917080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/07/2018 11:38

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency	ORANGE COUNTY S.O.	**Date**	12/05/2018	**Type**	Agency Inspection
Inspector	MELVILLE,KELLY,M	**Time**	10:51:18	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13917080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2018 11:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency	PINELLAS COUNTY SO	**Date**	03/13/2018	**Type**	Agency Inspection
Inspector	BELL,DANIELLE,M	**Time**	11:29:30	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

Data Download Date/Time 03/13/2018 15:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency PINELLAS COUNTY SO **Date** 03/13/2018 **Type** Department Inspection
Inspector BELL,DANIELLE,M **Time** 13:52:34 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG715202
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.083
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.083
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 03/13/2018 15:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency	PINELLAS COUNTY SO	**Date**	01/08/2019	**Type**	Agency Inspection
Inspector	GRAHAM,THOMAS,J	**Time**	12:30:29	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

Data Download Date/Time 01/08/2019 15:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency PINELLAS COUNTY SO **Date** 01/08/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 14:56:57 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0007
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/08/2019 15:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency	COCOA P.D.	**Date**	01/16/2018	**Type**	Agency Inspection
Inspector	FRAZIER,JAMES,H	**Time**	21:40:26	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	06616080A2
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	04/05/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2018 21:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency	COCOA P.D.	**Date**	02/24/2018	**Type**	Agency Inspection
Inspector	FRAZIER,JAMES,H	**Time**	20:14:22	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	06616080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	04/05/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/24/2018 20:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency	COCOA P.D.	**Date**	03/29/2018	**Type**	Agency Inspection
Inspector	FRAZIER,JAMES,H	**Time**	22:25:02	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	06616080A2
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	04/05/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2018 22:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency COCOA P.D. **Date** 04/10/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 13:36:56 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/10/2018 13:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency	COCOA P.D.	**Date**	04/26/2018	**Type**	Agency Inspection
Inspector	FRAZIER,JAMES,H	**Time**	21:55:13	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	02618080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	03/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/26/2018 22:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency COCOA P.D. **Date** 05/26/2018 **Type** Agency Inspection
Inspector FRAZIER,JAMES,H **Time** 20:18:10 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 02618080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/26/2018 20:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency COCOA P.D. **Date** 10/23/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 15:02:05 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 10/23/2018 15:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency COCOA P.D. **Date** 10/25/2018 **Type** Agency Inspection
Inspector FRAZIER,JAMES,H **Time** 21:18:35 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 02618080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2018 21:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency COCOA P.D. **Date** 11/28/2018 **Type** Agency Inspection
Inspector FRAZIER,JAMES,H **Time** 20:55:52 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 02618080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:GAVE READING.

Data Download Date/Time 11/28/2018 21:03

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency COCOA P.D. **Date** 12/19/2018 **Type** Agency Inspection
Inspector FRAZIER,JAMES,H **Time** 21:22:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 02618080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/19/2018 21:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 01/31/2018 **Type** Agency Inspection
Inspector GOMEZ,IVETTE, **Time** 13:12:52 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 579920
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.182
0.20 g/210L Test # (Repeat) 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.194
0.20 g/210L Test # (Repeat) 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.20 g/210L Test # (Repeat) 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AIR LEAK SIM CAP.

Data Download Date/Time 01/31/2018 13:12

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency	ROCKLEDGE P.D.	**Date**	03/08/2018	**Type**	Agency Inspection
Inspector	PLATT,SHAYLA,D	**Time**	12:28:53	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 03/08/2018 15:37

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 03/08/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 15:39:02 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG626604
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/08/2018 15:37

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency	ROCKLEDGE P.D.	**Date**	03/12/2018	**Type**	Agency Inspection
Inspector	KENNEDY,PATRICK,W	**Time**	04:41:36	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	572920
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 03/12/2018 05:00

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency	ROCKLEDGE P.D.	**Date**	04/30/2018	**Type**	Agency Inspection
Inspector	KENNEDY,PATRICK,W	**Time**	12:47:17	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	572920
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/20/2018 09:35

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 05/20/2018 **Type** Agency Inspection
Inspector KENNEDY,PATRICK,W **Time** 10:25:30 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 572920
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/20/2018 10:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 06/30/2018 **Type** Agency Inspection
Inspector KENNEDY,PATRICK,W **Time** 17:23:06 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 572920
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2018 17:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency	ROCKLEDGE P.D.	**Date**	07/31/2018	**Type**	Agency Inspection
Inspector	GOMEZ,IVETTE,	**Time**	08:14:06	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	572920
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 07/31/2018 08:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 08/27/2018 **Type** Agency Inspection
Inspector GOMEZ,IVETTE, **Time** 07:41:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.043
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.047
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.045
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.052
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SIMULATOR ISSUE WITH AIR LEAK. SIMULATOR ISSUE, WILL TAKE OFF LINE TO CORRECT PROBLEM.

Data Download Date/Time 08/27/2018 09:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 08/27/2018 **Type** Agency Inspection
Inspector GOMEZ,IVETTE, **Time** 09:12:04 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 572920
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/27/2018 09:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 09/30/2018 **Type** Agency Inspection
Inspector GOMEZ,IVETTE, **Time** 10:04:05 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 572920
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/30/2018 10:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 10/23/2018 **Type** Agency Inspection
Inspector GOMEZ,IVETTE, **Time** 04:42:19 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 572920
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.072
0.08 g/210L Test # (Repeat) 1	0.078
0.20 g/210L Test # 1	0.128
0.20 g/210L Test # (Repeat) 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.075
0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test # 2	0.139
0.20 g/210L Test # (Repeat) 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.08 g/210L Test # (Repeat) 3	0.078
0.20 g/210L Test # 3	0.139
0.20 g/210L Test # (Repeat) 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance SEAL NOT TIGHT. HEATING ISSUE.

Data Download Date/Time 10/23/2018 04:46

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency	ROCKLEDGE P.D.	**Date**	10/29/2018	**Type**	Agency Inspection
Inspector	KENNEDY,PATRICK,W	**Time**	10:02:54	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	572920
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/29/2018 10:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 11/29/2018 **Type** Agency Inspection
Inspector GOMEZ,IVETTE, **Time** 05:26:02 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 572920
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.084
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.054
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.054
		0.08 g/210L Test # 3	0.084
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2018 05:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency ROCKLEDGE P.D. **Date** 12/28/2018 **Type** Agency Inspection
Inspector GOMEZ,IVETTE, **Time** 04:58:31 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 572920
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 09/13/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

MOUTH ALCOHOL PERFORM FIRST BY ACCIDENT

Data Download Date/Time 12/28/2018 05:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency	MELBOURNE P.D.	**Date**	01/30/2018	**Type**	Agency Inspection
Inspector	WHITRIGHT,MARK,D	**Time**	13:57:35	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	712031
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/21/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2018 13:57

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency	MELBOURNE P.D.	**Date**	02/28/2018	**Type**	Agency Inspection
Inspector	WHITRIGHT,MARK,D	**Time**	16:17:38	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	712031
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/21/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2018 16:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency	MELBOURNE P.D.	**Date**	03/29/2018	**Type**	Agency Inspection
Inspector	WHITRIGHT,MARK,D	**Time**	15:57:44	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	814183
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	07/26/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2018 15:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. **Date** 04/30/2018 **Type** Agency Inspection
Inspector WHITRIGHT,MARK,D **Time** 10:38:21 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 814183
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/26/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/30/2018 10:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. **Date** 05/31/2018 **Type** Agency Inspection
Inspector WHITRIGHT,MARK,D **Time** 08:08:45 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 814183
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/26/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2018 08:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency	MELBOURNE P.D.	**Date**	06/28/2018	**Type**	Agency Inspection
Inspector	WHITRIGHT,MARK,D	**Time**	08:14:14	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	814183
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	07/26/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/28/2018 08:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. **Date** 07/25/2018 **Type** Agency Inspection
Inspector WHITRIGHT,MARK,D **Time** 10:06:30 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 814183
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/26/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/25/2018 10:15

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. **Date** 08/23/2018 **Type** Agency Inspection
Inspector WHITRIGHT,MARK,D **Time** 15:45:04 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 814183
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/26/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/23/2018 15:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency	MELBOURNE P.D.	**Date**	09/11/2018	**Type**	Agency Inspection
Inspector	WHITRIGHT,MARK,D	**Time**	15:00:22	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814183
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/26/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/11/2018 15:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. **Date** 09/20/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 13:42:40 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 09/20/2018 13:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency	MELBOURNE P.D.	**Date**	09/25/2018	**Type**	Agency Inspection
Inspector	WHITRIGHT,MARK,D	**Time**	08:08:57	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814183
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/26/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

RETURN TO SERVICE AI FOLLOWING 2018 DI

Data Download Date/Time 09/25/2018 08:12

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. **Date** 10/31/2018 **Type** Agency Inspection
Inspector WHITRIGHT,MARK,D **Time** 12:10:20 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 906364
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2018 12:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency	MELBOURNE P.D.	**Date**	11/27/2018	**Type**	Agency Inspection
Inspector	WHITRIGHT,MARK,D	**Time**	13:15:26	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	906364
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	02/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/27/2018 13:19

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency	MELBOURNE P.D.	**Date**	12/27/2018	**Type**	Agency Inspection
Inspector	WHITRIGHT,MARK,D	**Time**	11:13:16	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	906364
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	02/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/27/2018 11:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	01/28/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	01:25:03	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	03417080A1
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/28/2018 01:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	02/24/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	05:06:55	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	03417080A1
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

EVERYTHING T4

Data Download Date/Time 02/24/2018 07:06

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	03/25/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	20:05:14	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	03417080A1
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:VALVE NOT TIGHTENED.

Data Download Date/Time 03/25/2018 20:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency INDIAN HARBOUR PD **Date** 04/26/2018 **Type** Agency Inspection
Inspector WIMMER,MATTHEW,S **Time** 09:25:20 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:VALVE NOT TIGHTENED, . Non-compliance:NOT DETECTING INTERFERENT / CHANGING O RING .

Data Download Date/Time 04/26/2018 18:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency INDIAN HARBOUR PD **Date** 04/26/2018 **Type** Agency Inspection
Inspector WIMMER,MATTHEW,S **Time** 10:45:05 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded TOO MUCH MOUTH ALCOHOL.

Data Download Date/Time 04/26/2018 18:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency INDIAN HARBOUR PD **Date** 05/28/2018 **Type** Agency Inspection
Inspector WIMMER,MATTHEW,S **Time** 14:26:18 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test # 2	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 3	0.000
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Ambient Fail, Ambient Fail. Non-compliance: AMBIENT FAIL READING TWICE.

Data Download Date/Time 05/29/2018 05:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency INDIAN HARBOUR PD **Date** 06/07/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 11:27:05 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 06/07/2018 11:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	06/16/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	08:44:46	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	03417080A1
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/18/2018 09:29

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	07/15/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	19:28:11	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	03417080A1
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/15/2018 19:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	08/17/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	03:34:50	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range ExceededAmbient FailRange ExceededRange Ex. Non-compliance:RANGE EXCEEDED MOUTH ALCOHOL.

Data Download Date/Time 08/17/2018 04:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	08/17/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	04:32:07	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	03417080A1
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: .

Data Download Date/Time 08/17/2018 04:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	09/22/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	19:35:50	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	03417080A1
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/22/2018 19:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency	INDIAN HARBOUR PD	**Date**	10/20/2018	**Type**	Agency Inspection
Inspector	WIMMER,MATTHEW,S	**Time**	08:45:11	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	03417080A1
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NO REMARKS

Data Download Date/Time 10/20/2018 08:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency INDIAN HARBOUR PD **Date** 11/07/2018 **Type** Agency Inspection
Inspector WIMMER,MATTHEW,S **Time** 08:14:08 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 03417080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/07/2018 08:40

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency INDIAN HARBOUR PD **Date** 12/20/2018 **Type** Agency Inspection
Inspector WIMMER,MATTHEW,S **Time** 08:06:00 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 02/06/2020 **Expiration** 06/22/2020 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

08: FAILED TO PLUG IN GAS CYLINDER . FAILED TO PLUG IN GAS CYLINDER

Data Download Date/Time 12/20/2018 08:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency	BREVARD COUNTY S.O.	**Date**	01/11/2018	**Type**	Agency Inspection
Inspector	GARRITY,THOMAS,A	**Time**	20:17:58	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	12716080A5
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/11/2018 21:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 02/28/2018 **Type** Agency Inspection
Inspector WAGNER,ROBERT,F **Time** 16:53:16 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A5
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.044
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance SEAL.

Data Download Date/Time 03/05/2018 09:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 03/29/2018 **Type** Agency Inspection
Inspector MOORE,KIMBERLY,N **Time** 17:16:44 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A5
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	No
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2018 19:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 03/29/2018 **Type** Agency Inspection
Inspector MOORE,KIMBERLY,N **Time** 18:57:18 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A5
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2018 19:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 04/26/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 17:16:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A5
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/26/2018 17:20

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 05/23/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 18:04:43 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

05: SIMLATOR NOT CONNECTED PROPERLY .

Data Download Date/Time 05/23/2018 18:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 06/13/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 12:39:41 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/13/2018 12:48

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 06/26/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 14:46:58 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.083
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.083
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.083
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.083
0.08 g/210L Dry Gas Std Test # (Repeat) 4	0.083
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.083
0.08 g/210L Dry Gas Std Test # (Repeat) 5	0.083
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.083
0.08 g/210L Dry Gas Std Test # (Repeat) 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.083
0.08 g/210L Dry Gas Std Test # (Repeat) 7	0.083
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.086
0.08 g/210L Dry Gas Std Test # (Repeat) 8	0.083
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.086
0.08 g/210L Dry Gas Std Test # (Repeat) 9	0.083
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.086
0.08 g/210L Dry Gas Std Test # (Repeat) 10	0.083
Interferent Detect Test # 10	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0015
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0014
0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat)	0.0003

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 06/26/2018 14:55

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 07/26/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 17:38:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2018 20:07

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 08/23/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 21:19:39 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/23/2018 21:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 09/28/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 18:51:53 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/28/2018 19:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency	BREVARD COUNTY S.O.	**Date**	10/25/2018	**Type**	Agency Inspection
Inspector	GARRITY,THOMAS,A	**Time**	18:55:50	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.

Data Download Date/Time 10/25/2018 19:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 11/30/2018 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 15:31:36 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 19817080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/30/2018 15:40

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency	BREVARD COUNTY S.O.	**Date**	12/27/2018	**Type**	Agency Inspection
Inspector	GARRITY,THOMAS,A	**Time**	17:07:12	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/27/2018 17:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency BREVARD COUNTY S.O. **Date** 01/09/2019 **Type** Agency Inspection
Inspector GARRITY,THOMAS,A **Time** 17:09:19 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201802B **0.20 Lot #** 201802G **0.08 Gas Lot #** 19817080A3
Expiration 02/06/2020 **Expiration** 02/06/2020 **Expiration** 02/22/2020 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/09/2019 17:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency	PALM BAY P.D.	**Date**	01/30/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	14:48:30	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Ambient Fail.

Data Download Date/Time 01/30/2018 14:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 02/13/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 13:06:17 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG708807
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.087
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.085
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.084
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.084
0.08 g/210L Dry Gas Std Test # (Repeat) 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 5	0.081
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.082
0.08 g/210L Dry Gas Std Test # (Repeat) 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 10	0.081
Interferent Detect Test # 10	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0021
0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat)	0.0000

Remarks:

08: Control Outside Tolerance.- CHANGED CYLINDER & RETESTED

Data Download Date/Time 02/13/2018 13:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency	PALM BAY P.D.	**Date**	02/26/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	04:40:05	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2018 04:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency	PALM BAY P.D.	**Date**	03/28/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	21:28:34	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2018 21:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 04/25/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 20:29:29 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2018 20:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 05/25/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 18:35:14 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/25/2018 18:38

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 06/27/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 18:32:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2018 18:38

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency	PALM BAY P.D.	**Date**	07/26/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	19:38:45	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2018 19:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 08/28/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 13:20:00 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2018 13:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 09/27/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 16:09:40 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/27/2018 16:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 10/24/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 13:24:03 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/24/2018 13:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 11/26/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 02:57:48 **Software** 8100.27

0.05 Lot # 201808B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/26/2018 03:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency PALM BAY P.D. **Date** 12/29/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 21:54:36 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 24818080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/29/2018 22:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency	PALM BAY P.D.	**Date**	01/30/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	14:40:52	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2018 14:48

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency PALM BAY P.D. **Date** 02/26/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 04:47:31 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2018 04:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency	PALM BAY P.D.	**Date**	03/28/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	21:22:39	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2018 21:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency	PALM BAY P.D.	**Date**	04/25/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	20:19:55	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2018 20:23

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency PALM BAY P.D. **Date** 05/25/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 18:38:49 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/25/2018 18:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency PALM BAY P.D. **Date** 06/27/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 18:42:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2018 18:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency PALM BAY P.D. **Date** 07/26/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 19:42:26 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2018 19:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency	PALM BAY P.D.	**Date**	08/28/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	13:26:48	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2018 13:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency PALM BAY P.D. **Date** 09/07/2018 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 09:32:58 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/07/2018 09:37

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency PALM BAY P.D. **Date** 09/27/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 16:17:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 22817080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/27/2018 16:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency	PALM BAY P.D.	**Date**	10/24/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	13:13:34	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/24/2018 13:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency	PALM BAY P.D.	**Date**	11/26/2018	**Type**	Agency Inspection
Inspector	LUTZ,EDWIN,J	**Time**	02:47:51	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	24818080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/26/2018 02:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency PALM BAY P.D. **Date** 12/29/2018 **Type** Agency Inspection
Inspector LUTZ,EDWIN,J **Time** 21:50:13 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 24818080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/29/2018 21:55

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency SEMINOLE COUNTY S.O. **Date** 01/26/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 22:13:28 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 282638
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2018 00:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency	SEMINOLE COUNTY S.O.	**Date**	02/24/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	00:52:01	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	282638
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2016	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2018 03:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency	SEMINOLE COUNTY S.O.	**Date**	03/07/2018	**Type**	Agency Inspection		
Inspector	BETHAM,KEITH,D	**Time**	00:01:26	**Software**	8100.27		
0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	282638
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/11/2018 22:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency	SEMINOLE COUNTY S.O.	**Date**	04/28/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	00:39:35	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/29/2018 01:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency SEMINOLE COUNTY S.O. **Date** 05/27/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 01:04:54 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 282638
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/29/2018 00:57

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency	SEMINOLE COUNTY S.O.	**Date**	06/29/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	23:10:15	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/01/2018 21:34

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency SEMINOLE COUNTY S.O. **Date** 07/29/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 02:29:30 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 282638
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/29/2018 21:29

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency	SEMINOLE COUNTY S.O.	**Date**	08/31/2018	**Type**	Agency Inspection
Inspector	BEHAM,KEITH,D	**Time**	01:01:26	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/02/2018 20:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency	SEMINOLE COUNTY S.O.	**Date**	09/22/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	23:52:35	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/26/2018 20:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency SEMINOLE COUNTY S.O. **Date** 10/19/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 06:21:24 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 282638
Expiration 02/26/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/27/2018 04:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency SEMINOLE COUNTY S.O. **Date** 11/15/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 11:16:59 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.204
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.204
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.204
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 11/15/2018 11:20

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency SEMINOLE COUNTY S.O. **Date** 11/29/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 23:59:18 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 282638
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/04/2018 01:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency SEMINOLE COUNTY S.O. **Date** 12/28/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 01:32:24 **Software** 8100.27

0.05 Lot # 20802B **0.08 Lot #** 20709 **0.20 Lot #** 201702B **0.08 Gas Lot #** 282638
Expiration 02/06/2020 **Expiration** 09/22/2019 **Expiration** 02/23/2019 **Expiration** 02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/31/2018 22:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency	SEMINOLE COUNTY S.O.	**Date**	01/26/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	22:05:17	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	282638
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2018 01:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency	SEMINOLE COUNTY S.O.	**Date**	02/24/2018	**Type**	Agency Inspection
Inspector	BETHAM,KETH,D	**Time**	00:42:47	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	282638
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2018 03:40

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency	SEMINOLE COUNTY S.O.	**Date**	03/06/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	23:51:41	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	282638
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/11/2018 22:20

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency	SEMINOLE COUNTY S.O.	**Date**	04/28/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	00:41:11	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/29/2018 01:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency	SEMINOLE COUNTY S.O.	**Date**	05/27/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	00:47:48	**Software**	8100.27

0.05 Lot #	2018020620	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	11/15/2018	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 05/29/2018 01:00

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency	SEMINOLE COUNTY S.O.	**Date**	06/29/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	23:05:24	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/01/2018 21:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency	SEMINOLE COUNTY S.O.	**Date**	07/29/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	01:35:21	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/29/2018 21:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency	SEMINOLE COUNTY S.O.	**Date**	08/31/2018	**Type**	Agency Inspection
Inspector	BETHAM,KEITH,D	**Time**	00:56:30	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	282638
Expiration	02/06/2020	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	02/15/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect/RERUN.

Data Download Date/Time 09/03/2018 23:37

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency SEMINOLE COUNTY S.O. **Date** 09/22/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 23:20:41 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 915310
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 09/26/2018 20:57

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency SEMINOLE COUNTY S.O. **Date** 10/26/2018 **Type** Agency Inspection
Inspector BEHAM,KEITH,D **Time** 00:36:03 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 915310
Expiration 02/26/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/27/2018 04:23

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency SEMINOLE COUNTY S.O. **Date** 11/14/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 09:09:02 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/24/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.083
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 11/14/2018 09:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency SEMINOLE COUNTY S.O. **Date** 11/29/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 23:54:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 915310
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/02/2018 21:48

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency SEMINOLE COUNTY S.O. **Date** 12/28/2018 **Type** Agency Inspection
Inspector BETHAM,KEITH,D **Time** 01:34:41 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 915310
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/31/2018 22:47

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency	PINELLAS COUNTY SO	**Date**	01/25/2018	**Type**	Agency Inspection
Inspector	DEANE,M,	**Time**	21:43:56	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	AG629802
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	10/24/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/25/2018 21:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 02/08/2018 **Type** Department Inspection
Inspector BELL,DANIELLE,M **Time** 11:56:34 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG708807
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/08/2018 11:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 02/16/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 22:31:06 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** AG629802
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 10/24/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2018 22:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 03/22/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 21:55:00 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** AG629802
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 10/24/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/22/2018 21:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 04/28/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 02:19:52 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG629802
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/24/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/28/2018 02:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 05/31/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 01:12:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2018 01:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency	PINELLAS COUNTY SO	**Date**	06/28/2018	**Type**	Agency Inspection
Inspector	DEANE,M,	**Time**	21:18:02	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	AG728303
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/10/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/28/2018 21:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 07/25/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 23:55:09 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.189
0.20 g/210L Test # (Repeat) 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.194
0.20 g/210L Test # (Repeat) 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.194
0.20 g/210L Test # (Repeat) 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside ToleranceUNK. PROB. REPEAT OK.

Data Download Date/Time 07/26/2018 00:00

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 08/02/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 22:10:33 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/02/2018 22:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 09/12/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 23:15:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.189
0.20 g/210L Test # (Repeat) 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.193
0.20 g/210L Test # (Repeat) 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.193
0.20 g/210L Test # (Repeat) 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance UNK PROB REPEAT OK.

Data Download Date/Time 09/12/2018 23:21

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 10/25/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 00:38:45 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2018 00:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 11/28/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 22:36:07 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/28/2018 22:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency PINELLAS COUNTY SO **Date** 12/26/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 20:08:49 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG733803
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/04/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.192
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance UNK. PROB. REPEAT OK..

Data Download Date/Time 12/26/2018 20:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency	PINELLAS COUNTY SO	**Date**	01/26/2018	**Type**	Agency Inspection
Inspector	DEANE,M,	**Time**	00:31:36	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	AG629802
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	04/05/2018	**Expiration**	10/24/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/26/2018 00:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 02/08/2018 **Type** Department Inspection
Inspector BELL,DANIELLE,M **Time** 12:11:22 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG708807
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/08/2018 12:15

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 02/16/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 20:56:06 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** AG629802
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 10/24/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample ProvidedREST ROOM BREAK PEPEAT OK.

Data Download Date/Time 02/16/2018 20:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 03/22/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 20:12:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** AG629802
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 10/24/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.189
0.20 g/210L Test # (Repeat) 1	0.190
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.189
0.20 g/210L Test # (Repeat) 2	0.191
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.191
0.20 g/210L Test # (Repeat) 3	0.191
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside ToleranceUNK.PROB.REPEAT OK.

Data Download Date/Time 03/22/2018 20:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency	PINELLAS COUNTY SO	**Date**	04/27/2018	**Type**	Agency Inspection
Inspector	DEANE,M,	**Time**	22:20:20	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	AG728303
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	10/10/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/27/2018 22:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 05/30/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 20:19:59 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.074
0.08 g/210L Test # (Repeat) 1	0.076
0.20 g/210L Test # 1	0.190
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test # 2	0.192
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.08 g/210L Test # (Repeat) 3	0.077
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside ToleranceUNK.PROB.REPEAT OK.

Data Download Date/Time 05/30/2018 20:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 06/28/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 23:41:03 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/28/2018 23:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 07/25/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 21:18:56 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # (Repeat) 1	0.077
0.20 g/210L Test # 1	0.189
0.20 g/210L Test # (Repeat) 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test # 2	0.190
0.20 g/210L Test # (Repeat) 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # (Repeat) 3	0.078
0.20 g/210L Test # 3	0.192
0.20 g/210L Test # (Repeat) 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

08: RFI DetectMICROWAVE IN USE IN KITCHEN. 20: Control Outside ToleranceUNK. PROB. REPEAT OK.

Data Download Date/Time 07/25/2018 21:23

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 08/02/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 21:38:10 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.188
0.20 g/210L Test # (Repeat) 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.20 g/210L Test # (Repeat) 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.193
0.20 g/210L Test # (Repeat) 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance UNK. PROB. REPEAT OK.

Data Download Date/Time 08/02/2018 21:40

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 09/12/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 21:43:46 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/12/2018 21:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 10/24/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 23:55:45 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/24/2018 23:57

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency	PINELLAS COUNTY SO	**Date**	11/29/2018	**Type**	Agency Inspection		
Inspector	DEANE,M,	**Time**	01:03:13	**Software**	8100.27		
0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	AG815602
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2018 01:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency PINELLAS COUNTY SO **Date** 12/26/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 22:09:36 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/26/2018 22:15

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 01/30/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:34:18 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 19817080A2
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2018 11:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 02/05/2018 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 13:06:12 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG626605
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/05/2018 13:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 02/07/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:51:23 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 19817080A2
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2018 09:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 03/29/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:03:29 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2018 09:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 04/03/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:28:07 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/03/2018 09:44

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 05/24/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:18:27 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/24/2018 09:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 06/28/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 08:59:54 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/28/2018 09:03

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 07/19/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:57:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2018 11:05

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 08/31/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:01:16 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/31/2018 12:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 09/10/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:09:23 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/10/2018 10:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency	HIALEAH GARDENS PD	**Date**	10/04/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:12:11	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	19817080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/04/2018 10:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 11/30/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:51:25 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/30/2018 11:07

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency	HIALEAH GARDENS PD	**Date**	12/05/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:31:55	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	19817080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2018 10:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency HIALEAH GARDENS PD **Date** 01/10/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:38:25 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/10/2019 12:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency	MIAMI-DADE PD	**Date**	09/12/2018	**Type**	Agency Inspection
Inspector	REYES-RIVERA,DAVID,E	**Time**	06:15:52	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COMPLIANCE NOT DETERMINED AI NOT CONDUCTED

Data Download Date/Time 09/12/2018 14:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency	MIAMI-DADE PD	**Date**	09/12/2018	**Type**	Department Inspection
Inspector	REYES-RIVERA,DAVID,E	**Time**	14:38:17	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805701
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/12/2018 14:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency	MIAMI-DADE PD	**Date**	10/05/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	12:42:36	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	N	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/05/2018 13:11

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency MIAMI-DADE PD **Date** 11/09/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:25:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/09/2018 09:37

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency MIAMI-DADE PD **Date** 12/07/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:01:30 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/07/2018 09:33

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency MIAMI-DADE PD **Date** 01/11/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:44:24 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/11/2019 11:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency MIAMI-DADE PD **Date** 01/04/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:11:28 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NOT ENOUGH MA

Data Download Date/Time 01/04/2018 12:12

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency MIAMI-DADE PD **Date** 02/01/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:19:02 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2018 13:36

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	02/05/2018	**Type**	Department Inspection
Inspector	REYES-RIVERA,DAVID,E	**Time**	13:19:50	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG626605
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.052
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.052
0.08 g/210L Test # 4	0.083
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.052
0.08 g/210L Test # 5	0.083
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.082
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.052
0.08 g/210L Test # 7	0.084
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.084
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0009
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/05/2018 13:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	02/07/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:48:19	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201709A	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	20016080A2
Expiration	03/08/2018	**Expiration**	09/12/2019	**Expiration**	04/05/2018	**Expiration**	09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2018 09:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	03/09/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:52:41	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201604C	**0.08 Gas Lot #**	20016080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	04/05/2018	**Expiration**	09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/09/2018 11:03

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency MIAMI-DADE PD **Date** 04/19/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:38:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/19/2018 09:48

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency MIAMI-DADE PD **Date** 05/25/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:59:03 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/25/2018 10:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency MIAMI-DADE PD **Date** 06/01/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:49:20 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/01/2018 11:08

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	07/26/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:25:38	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	20016080A2
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2018 11:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	08/14/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:33:01	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	20016080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/14/2018 10:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency MIAMI-DADE PD **Date** 09/07/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:19:05 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/07/2018 10:23

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	10/19/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:40:25	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/19/2018 09:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	11/12/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:20:35	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/12/2018 13:38

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	12/13/2018	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:10:03	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/13/2018 10:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency	MIAMI-DADE PD	**Date**	01/04/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:24:28	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/04/2019 10:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	03/20/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	11:48:47	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

PLACED INTO SERV 3/20/2018

Data Download Date/Time 03/21/2018 09:39

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	04/02/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:42:06	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/02/2018 14:42

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	04/17/2018	**Type**	Agency Inspection
Inspector	POTTER,M,R	**Time**	12:45:02	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 12:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	05/01/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:40:41	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/01/2018 14:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	05/30/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	15:06:29	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2018 15:11

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	06/19/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:40:59	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/19/2018 15:02

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	07/05/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:57:27	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/06/2018 12:22

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	07/18/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:37:08	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/18/2018 14:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	08/17/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	11:09:45	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/17/2018 11:23

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency JACKSONVILLE SO **Date** 09/06/2018 **Type** Agency Inspection
Inspector POTTER,MICHAEL,R **Time** 14:16:30 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 05218080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM SERV FOR ANNUAL PM 9.6.2018

Data Download Date/Time 09/06/2018 14:20

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency JACKSONVILLE SO **Date** 10/15/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 09:22:22 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 10/15/2018 09:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	10/17/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	11:26:15	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

PLACED INTO SERV 10.17.2018

Data Download Date/Time 10/17/2018 11:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	11/06/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	12:03:52	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/06/2018 12:14

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

Agency	JACKSONVILLE SO	**Date**	12/20/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	15:27:28	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2018 08:34

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	01/19/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:43:51	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201601F	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/08/2018	**Expiration**	01/26/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/19/2018 14:43

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	02/01/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	13:43:51	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2018 13:49

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	02/15/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	13:19:32	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect, SOURCE UNIDENTIFIED, RE-RAN, OK

Data Download Date/Time 02/15/2018 13:20

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	02/28/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	13:50:23	**Software**	8100.27

0.05 Lot #	201603D	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/08/2018	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2018 14:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	03/20/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	12:23:23	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/21/2018 09:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	04/02/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:57:51	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/02/2018 15:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	04/17/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	13:05:26	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 13:09

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	05/01/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:56:37	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/01/2018 15:08

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	05/30/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	15:25:47	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2018 15:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	06/19/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:56:24	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/19/2018 15:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	07/05/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	15:13:27	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/06/2018 12:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	07/18/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:58:30	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM SERV FOR ANNUAL PM 7/18/2018

Data Download Date/Time 07/18/2018 15:04

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency JACKSONVILLE SO **Date** 08/29/2018 **Type** Department Inspection
Inspector SHANAHAN,JAKE,L **Time** 09:58:35 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/29/2018 10:03

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	09/06/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	12:20:28	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

PLACED INTO SERV 9.6.2018

Data Download Date/Time 09/06/2018 12:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	09/25/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:10:16	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/25/2018 14:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	10/17/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	15:10:04	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/17/2018 15:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	11/06/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	11:52:44	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/06/2018 11:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency	JACKSONVILLE SO	**Date**	12/20/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	15:17:56	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2018 08:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	03/15/2018	**Type**	Department Inspection
Inspector	SHANAHAN,JAKE,L	**Time**	10:30:51	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG715202
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.053
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0012
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/15/2018 10:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	05/30/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	13:57:05	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

PLACED INTO SERV. 05/30/2018

Data Download Date/Time 05/30/2018 14:00

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	06/19/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	14:31:37	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	22817080A4
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/19/2018 14:54

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	08/29/2018	**Type**	Agency Inspection
Inspector	SHANAHAN,JAKE,L	**Time**	07:52:44	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 10/16/2018 10:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	10/16/2018	**Type**	Agency Inspection
Inspector	SHANAHAN,JAKE,L	**Time**	07:27:19	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 10/16/2018 10:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	10/16/2018	**Type**	Agency Inspection
Inspector	SHANAHAN,JAKE,L	**Time**	08:32:56	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 10/16/2018 10:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	10/16/2018	**Type**	Department Inspection
Inspector	SHANAHAN,JAKE,L	**Time**	10:49:34	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805701
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.076
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 10/16/2018 10:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	10/31/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	11:52:22	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201611B	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	17817080A3
Expiration	03/07/2019	**Expiration**	11/15/2018	**Expiration**	02/23/2019	**Expiration**	08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

PLACES INTO SERV @ NAS MAYPORT PD 10.31.2018

Data Download Date/Time 10/31/2018 12:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency	JACKSONVILLE SO	**Date**	11/29/2018	**Type**	Agency Inspection
Inspector	POTTER,MICHAEL,R	**Time**	15:19:01	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	05218080A3
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM SERV 11.29.18 PLACED IN RESERVE

Data Download Date/Time 11/29/2018 15:24

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.