

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 01/23/2018
Time 13:50:29

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/23/2018 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 01/23/2018 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 13:52:59 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 01/23/2018 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,M

Date 02/14/2018
Time 22:49:40

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/14/2018 22:49

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,M

Date 03/22/2018
Time 14:59:10

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/25/2018 00:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 04/25/2018
Time 13:57:15

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 04/25/2018 14:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 05/29/2018
Time 17:11:29

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/29/2018 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO	Date 07/03/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 12:59:52	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0009 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0017
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 07/03/2018 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector BIXON,BARRY,

Date 07/11/2018
Time 14:55:25

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2018 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 08/03/2018
Time 14:17:03

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/03/2018 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 09/26/2018
Time 21:51:17

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/26/2018 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 10/10/2018
Time 18:36:58

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/10/2018 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 11/26/2018
Time 19:07:28

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/26/2018 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 12/19/2018
Time 18:17:44

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/19/2018 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 01/09/2019
Time 20:43:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 17817080A2
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/09/2019 20:52

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 01/26/2018
Time 09:00:02

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 1878845
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/22/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 504 PSI

Data Download Date/Time 01/26/2018 10:25

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 02/28/2018
Time 07:54:50

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 1878845
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/22/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 500 PSI

Data Download Date/Time 02/28/2018 08:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 03/19/2018
Time 15:19:30

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 1878845
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 06/22/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

SIMULATOR NOT SEALED PROPERLY FOR .05 TEST

Data Download Date/Time 03/19/2018 15:29

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI	Date 04/04/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 11:14:22	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 04/04/2018 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 04/27/2018
Time 08:23:28

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 712031
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/21/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 950 PSI

Data Download Date/Time 04/27/2018 08:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 05/29/2018
Time 07:24:40

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/21/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 1000 PSI

Data Download Date/Time 05/29/2018 11:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 06/29/2018
Time 09:14:39

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 1000 SPI

Data Download Date/Time 06/29/2018 11:28

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 07/31/2018
Time 12:06:02

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 900 PSI

Data Download Date/Time 08/03/2018 06:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 08/30/2018
Time 08:28:54

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 900 PSI

Data Download Date/Time 08/30/2018 10:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI	Date 10/01/2018	Type Agency Inspection	
Inspector ROCA,LUIS,C	Time 13:10:00	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201611B	0.20 Lot # 201802G	0.08 Gas Lot # 712031
Expiration 02/06/2020	Expiration 11/15/2018	Expiration 02/22/2020	Expiration 02/21/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

THIS INSTR WAS MISPLACED IN SEPT AI NOT DONE TPR ROCA

Data Download Date/Time 10/01/2018 13:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI	Date 10/08/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 11:43:39	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 10/08/2018 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI	Date 10/31/2018	Type Agency Inspection	
Inspector ADDISON, RAYMOND, E	Time 00:48:41	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 712031
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 02/21/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

ALCO SOL. 2016C; ACET SOL. 2017A

Data Download Date/Time 10/31/2018 01:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 11/29/2018 **Type** Agency Inspection
Inspector ADDISON,RAYMOND,E **Time** 02:59:05 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

ALCO. SOL. 12/01/2018; ACET. SOL. 02/20/2019

Data Download Date/Time 11/29/2018 04:15

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 12/30/2018
Time 13:48:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 750 PSI

Data Download Date/Time 01/02/2019 06:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI	Date 01/02/2019	Type Agency Inspection	
Inspector ROCA,LUIS,C	Time 08:39:01	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 712031
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 02/21/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 750 PSI

Data Download Date/Time 01/02/2019 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP MIAMI	Date 01/07/2019	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 14:46:33	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/07/2019 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 01/03/2018 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 17:51:32 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 01/03/2018 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO
Inspector SKIPPER, ROGER, G

Date 02/13/2018
Time 17:45:58

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 02/13/2018 17:46

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO	Date 03/05/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 15:46:12	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/05/2018 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO	Date 04/03/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 13:20:52	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805702
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0013
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 04/03/2018 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 05/03/2018 **Type** Agency Inspection
Inspector GRANNIS, RONALD, R **Time** 17:16:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI R. SKIPPER #246109 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 05/03/2018 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO	Date 06/19/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 19:10:00	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 03417080A1
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 03/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 06/20/2018 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 07/12/2018 **Type** Agency Inspection
Inspector GRANNIS, RONALD, R **Time** 19:10:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

AI R. SKIPPER #246109 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 07/12/2018 19:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 08/07/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 13:36:17 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 03417080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T.BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 08/07/2018 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 09/06/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 14:28:10 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 03417080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 09/06/2018 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 10/04/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 11:34:35 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 03417080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 10/04/2018 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO	Date 11/01/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 12:06:05	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 03417080A1
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 03/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 11/01/2018 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 12/18/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 17:35:34 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 03417080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 12/18/2018 17:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 01/10/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 17:42:30 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 28818080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 01/10/2019 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 01/03/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 17:59:48 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 11117080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 01/03/2018 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO
Inspector SKIPPER, ROGER, G

Date 02/13/2018
Time 17:42:49

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 02/13/2018 17:44

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO	Date 03/05/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 15:57:08	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.204
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/05/2018 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 04/05/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 17:14:54 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 04/05/2018 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO
Inspector SKIPPER, ROGER, G

Date 05/03/2018
Time 18:36:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 05/03/2018 18:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO	Date 06/19/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 18:55:00	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 03417080A1
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 03/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 06/19/2018 18:58

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO	Date 07/12/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 17:34:39	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 31517080A1
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 01/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 07/12/2018 17:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 08/07/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 13:29:59 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 08/07/2018 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 09/06/2018 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 14:32:39 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 09/06/2018 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 10/04/2018 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 11:39:01 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 10/04/2018 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO	Date 11/01/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 12:02:09	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 31517080A1
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 01/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 11/01/2018 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 12/18/2018 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 17:29:33 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 12/18/2018 17:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 01/10/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 19:37:43 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 01/10/2019 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency	FL HIGHWAY PATROL	Date	04/12/2018	Type	Agency Inspection
Inspector	REYES-RIVERA,DAVID,E	Time	06:20:20	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED

Data Download Date/Time 04/12/2018 14:16

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency FHP TROOP L	Date 04/12/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 14:14:21	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/12/2018 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency FHP TROOP L	Date 05/11/2018	Type Agency Inspection	
Inspector WASHINGTON,YVETTE,D	Time 15:42:49	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 875088
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 12/12/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: TOO MUCH MOUTH ALCOHOL

Data Download Date/Time 05/11/2018 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency FHP TROOP L	Date 07/13/2018	Type Agency Inspection	
Inspector WASHINGTON,YVETTE,D	Time 13:06:36	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 875088
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 12/12/2020

Number of Simulators 5	Compliance No
-------------------------------	----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.069
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.000
		Interferent Detect Test # 3	INT

Remarks:

08: ContR OUTS, CYCL TO BE CHANGED & NEW INSP COMPL. . Non-compliance:CYCL NEEDS TO BE CHANGED.

Data Download Date/Time 07/13/2018 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency FHP TROOP L **Date** 07/13/2018 **Type** Agency Inspection
Inspector WASHINGTON,YVETTE,D **Time** 13:39:42 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 875088
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/12/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

A. FRAZIER PRESENT & OBSERVED

Data Download Date/Time 07/13/2018 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency FHP TROOP L **Date** 08/29/2018 **Type** Agency Inspection
Inspector WASHINGTON,YVETTE,D **Time** 14:36:15 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 875088
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/12/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER OBSERVED

Data Download Date/Time 08/29/2018 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency FHP TROOP L **Date** 09/19/2018 **Type** Agency Inspection
Inspector WASHINGTON,YVETTE,D **Time** 14:14:52 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 875088
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/12/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER

Data Download Date/Time 09/19/2018 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 01/18/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:52:11 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 14816080A3
Expiration 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 07/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect:LEFT CELL BY INTOX..

Data Download Date/Time 01/18/2018 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 02/05/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 10:07:51 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 14816080A3
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 07/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 02/05/2018 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 03/05/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 10:45:35 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 14816080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 03/05/2018 11:13

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 04/02/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:50:33 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** \

Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 04/02/2018 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 05/16/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:42:08 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 14816080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 05/16/2018 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 06/05/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:51:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 14816080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 06/05/2018 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 07/19/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 12:03:20 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 07/19/2018 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL	Date 08/09/2018	Type Agency Inspection	
Inspector WILSON,STANLEY,T	Time 09:47:09	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 29017080A3
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 12/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 08/09/2018 09:56

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 09/12/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 10:20:51 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 09/12/2018 10:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 10/25/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:50:57 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 10/25/2018 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL	Date 11/14/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 13:04:55	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

Data Download Date/Time 11/14/2018 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 11/26/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 12:10:19 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 11/26/2018 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 12/03/2018 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 10:27:24 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 12/03/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 01/03/2019 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 10:29:18 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 06717080A6
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 01/03/2019 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency	FL HIGHWAY PATROL	Date	01/18/2018	Type	Agency Inspection
Inspector	SHANAHAN,JAKE,L	Time	07:18:25	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 01/18/2018 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL	Date 01/18/2018	Type Department Inspection
Inspector SHANAHAN, JAKE, L	Time 10:29:29	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG708807
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 01/18/2018 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 01/21/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 16:30:55 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 26316080A2
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 01/21/2018 17:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 02/18/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 20:26:10 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 26316080A2
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2018 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 03/03/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 15:28:15 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/03/2018 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 04/06/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 18:01:59 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/06/2018 18:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 05/12/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 15:52:00 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/12/2018 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 05/12/2018 **Type** Agency Inspection
Inspector PALESE,ANTHONY,W **Time** 15:53:12 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Data Download Date/Time 05/12/2018 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL	Date 05/30/2018	Type Department Inspection
Inspector SHANAHAN, JAKE, L	Time 09:37:43	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805702
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 05/30/2018 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL	Date 06/10/2018	Type Agency Inspection	
Inspector PALESE, ANTHONY, W	Time 22:55:06	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 26316080A2
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 11/05/2018
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/10/2018 23:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 07/23/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 17:55:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/23/2018 18:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 08/07/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 13:13:51 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/17/2018 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 09/25/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 15:03:11 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range ExceededRFI Detect.

Data Download Date/Time 09/25/2018 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 10/19/2018 **Type** Agency Inspection
Inspector CABRAL,JONATHAN,M **Time** 18:25:18 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/19/2018 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 11/29/2018 **Type** Agency Inspection
Inspector CABRAL,JONATHAN,M **Time** 18:15:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2018 18:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 12/28/2018 **Type** Agency Inspection
Inspector CABRAL,JONATHAN,M **Time** 01:25:42 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.

Data Download Date/Time 12/28/2018 01:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 01/04/2019 **Type** Agency Inspection
Inspector CABRAL,JONATHAN,M **Time** 17:38:04 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/04/2019 17:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 01/26/2018
Time 09:17:45

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 579

Data Download Date/Time 01/26/2018 10:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 02/28/2018
Time 07:41:56

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 550 PSI

Data Download Date/Time 02/28/2018 08:03

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 03/19/2018
Time 15:25:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 550 PSI

Data Download Date/Time 03/19/2018 15:32

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 04/27/2018
Time 07:04:24

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 450 PSI

Data Download Date/Time 04/27/2018 07:53

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 05/29/2018
Time 07:45:24

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 500 PSI

Data Download Date/Time 05/29/2018 11:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI	Date 06/29/2018	Type Agency Inspection	
Inspector ROCA,LUIS,C	Time 10:03:53	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 163437
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 10/30/2018

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # (Repeat) 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

RFI DURING .08 TEST DUE TO CELL PHONE.

Data Download Date/Time 06/29/2018 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 07/31/2018
Time 06:18:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 163437
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 450 PSI

Data Download Date/Time 08/03/2018 06:27

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 08/30/2018
Time 08:04:45

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 163437
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 400 PSI

Data Download Date/Time 08/30/2018 10:23

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI	Date 09/04/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 14:11:17	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 09/04/2018 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 09/30/2018
Time 08:00:18

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201802G **0.08 Gas Lot #** 796576
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/22/2020 **Expiration** 06/14/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 1000 PSI

Data Download Date/Time 09/30/2018 10:41

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI	Date 10/31/2018	Type Agency Inspection
Inspector ADDISON, RAYMOND, E	Time 00:44:31	Software 8100.27

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 796576
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 06/14/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

ALCO SOL. 2016C; ACET SOL. 2017A

Data Download Date/Time 10/31/2018 01:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 11/29/2018 **Type** Agency Inspection
Inspector ADDISON, RAYMOND, E **Time** 02:37:35 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 796576
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/14/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

ALCO. SOL. 12/01/18; ACET. SOL. 02/02/2019

Data Download Date/Time 11/29/2018 03:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 12/30/2018
Time 13:16:19

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 796576
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/14/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 1000 PSI

Data Download Date/Time 01/02/2019 07:03

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI
Inspector ROCA,LUIS,C

Date 01/02/2019
Time 09:05:50

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 796576
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/14/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TPR ROCA 1000 PSI

Data Download Date/Time 01/02/2019 09:51

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency FL HIGHWAY PATROL **Date** 01/22/2018 **Type** Agency Inspection
Inspector LEDFORD,TAYLOR,M **Time** 09:49:28 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 840875
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/19/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

WITH MSO DEPUTY CLINE PRESENT

Data Download Date/Time 01/22/2018 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency FL HIGHWAY PATROL **Date** 02/26/2018 **Type** Agency Inspection
Inspector LEDFORD,TAYLOR,M **Time** 09:45:24 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 840875
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/19/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

DEPUTY CLINE PRESENT

Data Download Date/Time 02/26/2018 09:50

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency FL HIGHWAY PATROL **Date** 03/26/2018 **Type** Agency Inspection
Inspector WATSON,TAYLOR, **Time** 09:12:53 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 840875
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/19/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TROOPER TAYLOR LEDFORD ACTUAL INSPECTOR

Data Download Date/Time 03/26/2018 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency FL HIGHWAY PATROL	Date 04/23/2018	Type Agency Inspection	
Inspector CLINE,GLENN,	Time 09:44:36	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 840875
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 09/19/2020
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 04/23/2018 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency FL HIGHWAY PATROL **Date** 07/23/2018 **Type** Agency Inspection
Inspector LEDFORD,TAYLOR,M **Time** 12:01:14 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 840875
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/19/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

DEP CLINE PRESENT

Data Download Date/Time 07/23/2018 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency FL HIGHWAY PATROL **Date** 09/24/2018 **Type** Agency Inspection
Inspector LEDFORD,TAYLOR, **Time** 09:22:15 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 840875
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/19/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

TPR. BIZOT PRESENT

Data Download Date/Time 09/24/2018 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency	FL HIGHWAY PATROL	Date	10/22/2018	Type	Agency Inspection		
Inspector	LEDFOORD,TAYLOR,M	Time	09:26:25	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201611B	0.20 Lot #	201702B	0.08 Gas Lot #	840875
Expiration	03/07/2019	Expiration	11/15/2018	Expiration	02/23/2019	Expiration	09/19/2020
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

WITH MSO DEPUTY KLINE

Data Download Date/Time 10/22/2018 09:29

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency	FL HIGHWAY PATROL	Date	12/04/2018	Type	Agency Inspection
Inspector	PLATT,SHAYLA,D	Time	09:20:23	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 12/04/2018 12:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency FL HIGHWAY PATROL	Date 12/04/2018	Type Department Inspection
Inspector PLATT,SHAYLA,D	Time 11:42:48	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 12/04/2018 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 01/30/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 18:03:40 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 437738
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 01/30/2018 18:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL	Date 02/09/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 11:54:03	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 02/09/2018 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 04/03/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 17:43:34 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 437738
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 04/03/2018 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 05/11/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 04:25:15 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 03/07/2019 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.025
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. SIMULATOR ORING ISSUE.

Data Download Date/Time 05/11/2018 05:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 05/11/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 04:57:52 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 437738
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 05/11/2018 05:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 06/26/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 17:58:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.068
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.071
		0.08 g/210L Test # (Repeat) 3	0.069
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. BAD ORING ON SIMULATOR.

Data Download Date/Time 06/26/2018 18:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 06/26/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 18:29:37 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 437738
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 06/26/2018 18:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 07/06/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 18:21:13 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 437738
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 07/06/2018 18:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 08/28/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 16:01:54 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 437738
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 08/28/2018 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL	Date 10/01/2018	Type Agency Inspection	
Inspector MCKENZIE, WILLIAM, S	Time 12:41:31	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 437738
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 06/08/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.08 g/210L Test # (Repeat) 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

O-RING ISSUE, RESEALED

Data Download Date/Time 10/01/2018 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 11/27/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 18:20:46 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 437738
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 11/27/2018 18:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 12/24/2018 **Type** Agency Inspection
Inspector MCKENZIE, WILLIAM, S **Time** 10:22:05 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 437738
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/08/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 12/24/2018 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 01/30/2018 **Type** Agency Inspection
Inspector CAMPBELL,JOSHUA,C **Time** 13:50:48 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2018 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL	Date 02/22/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 13:32:32	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/22/2018 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 03/29/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 17:42:21 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2018 17:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 04/29/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 11:36:54 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/29/2018 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 05/27/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 17:11:48 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/27/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 06/25/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 19:35:36 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 06/25/2018 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 07/22/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 12:49:13 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/23/2018 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 08/26/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 18:38:35 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI DetectRFI DetectRFI DetectRFI Detect.

Data Download Date/Time 08/29/2018 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 09/27/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 19:40:11 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/27/2018 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 10/29/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 17:28:50 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/29/2018 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL	Date 11/27/2018	Type Agency Inspection	
Inspector HERNANDEZ, ALAIN,	Time 19:22:24	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 477627
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed. 20: RFI Detect.

Data Download Date/Time 11/28/2018 16:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 12/30/2018 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 20:21:28 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/02/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency	FL HIGHWAY PATROL	Date	01/18/2018	Type	Agency Inspection
Inspector	SHANAHAN,JAKE,L	Time	07:28:16	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 01/18/2018 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL	Date 01/18/2018	Type Department Inspection
Inspector SHANAHAN, JAKE, L	Time 10:00:43	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG619605
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 07/14/2018

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

Data Download Date/Time 01/18/2018 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 01/21/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 16:33:41 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/21/2018 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 02/18/2018 **Type** Agency Inspection
Inspector PALESE, ANTHONY, W **Time** 20:23:55 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 02/18/2018 **Type** Agency Inspection
Inspector PALESE,ANTHONY,W **Time** 20:28:58 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Non-compliance: .

Data Download Date/Time 02/18/2018 20:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL	Date 03/03/2018	Type Agency Inspection	
Inspector PALESE, ANTHONY, W	Time 16:49:06	Software 8100.27	

0.05 Lot # 201603D	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 20016080A2
Expiration 03/08/2018	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: RFI Detect KEYED UP RADIO IN CLOSE PROX TO INTOX.

Data Download Date/Time 03/03/2018 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency	FL HIGHWAY PATROL	Date	01/03/2019	Type	Agency Inspection
Inspector	GRAHAM,THOMAS,J	Time	08:15:31	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance Yes
-----------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

Data Download Date/Time 01/03/2019 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL	Date 01/03/2019	Type Department Inspection
Inspector GRAHAM, THOMAS, J	Time 11:47:29	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/03/2019 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 01/31/2018 **Type** Agency Inspection
Inspector KNOWLES, CONSTANCE, J **Time** 13:58:15 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det: INT NOT DETECTED, TIGHTENED LID. INT STILL NOT DET. Non-compliance: INT NOT BEING DETECTED ON MULTIPLE INSPS .

Data Download Date/Time 01/31/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency	FL HIGHWAY PATROL	Date	03/26/2018	Type	Agency Inspection
Inspector	MURPHY,PATRICK,J	Time	12:21:48	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SKIPPED AI TO OPERATE INSTRUMENT

Data Download Date/Time 03/27/2018 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL	Date 03/27/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 11:39:23	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/27/2018 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 07/29/2018 **Type** Agency Inspection
Inspector DAVIS,BRIAN,O **Time** 15:29:23 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 390314
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 04/19/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/29/2018 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 08/24/2018 **Type** Agency Inspection
Inspector KNOWLES, CONSTANCE, J **Time** 11:21:42 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 390314
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/19/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/24/2018 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 09/27/2018 **Type** Agency Inspection
Inspector DAVIS,BRIAN,O **Time** 21:50:25 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 390314
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/19/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/27/2018 23:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL	Date 10/30/2018	Type Agency Inspection
Inspector KNOWLES, CONSTANCE, J	Time 20:48:07	Software 8100.27

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 390314
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 04/19/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.103
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.196
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # 2	0.108
		0.20 g/210L Test # (Repeat) 2	0.198
		0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.110
		0.20 g/210L Test # (Repeat) 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance TIGHTENED LID. 20: Control Outside Tolerance RESEALED LID, NO AIR LEAKS.

Data Download Date/Time 11/02/2018 17:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 11/30/2018 **Type** Agency Inspection
Inspector KNOWLES, CONSTANCE, J **Time** 17:06:45 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 390314
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 04/19/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/30/2018 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 01/08/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 19:31:29 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2018 19:31

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 02/01/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 19:35:08 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2018 19:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O.	Date 02/20/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 21:25:31	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance	. FAILED TO CONNECT DRY GAS LINE. CONNECTED, RETESTED, IN COM
-------------------------------	---

Data Download Date/Time 02/20/2018 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 03/26/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 22:11:31 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/26/2018 22:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O.	Date 07/03/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 10:22:39	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 07/03/2018 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 08/08/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 21:04:20 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/08/2018 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 09/06/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, **Time** 23:37:13 **Software** 8100.27
JACOB,

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2018 23:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 10/09/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, **Time** 23:35:40 **Software** 8100.27
JACOB,

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/09/2018 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 11/06/2018 **Type** Agency Inspection
Inspector LYONS,BEEKMAN, **Time** 19:37:30 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/06/2018 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O.	Date 12/03/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 03:25:56	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 021802G	0.08 Gas Lot # 646645
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.190
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.189
		0.20 g/210L Test # (Repeat) 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 12/03/2018 03:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 01/07/2019 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 00:12:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/07/2019 00:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O.	Date 01/08/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 19:44:53	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

08: RFI Detect PORTABLE RADIO TURNED ON DURING INSPECTION.

Data Download Date/Time 01/08/2018 19:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 02/01/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 19:39:39 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2018 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O.
Inspector LYONS,DANIEL,E

Date 02/20/2018
Time 19:27:35

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/20/2018 19:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O.	Date 03/14/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 14:57:38	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/14/2018 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 03/26/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 21:26:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/26/2018 21:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 04/17/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 18:55:37 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 05/03/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 19:04:11 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/03/2018 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O.	Date 06/05/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 22:37:15	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08:DID NOT ATTACHED DRY GAS HOSE	. ATTACHED/RETESTED
----------------------------------	---------------------

Data Download Date/Time 06/05/2018 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 06/18/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 19:52:41 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/18/2018 19:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 07/02/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL, BEEKMAN, JACOB, **Time** 23:06:58 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/02/2018 23:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O.	Date 08/08/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 21:24:15	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05:Control Outside Tolerance/SIM NOT CONNECTED/SIM CONNECT . SIM CONNECTED/RETESTED

Data Download Date/Time 08/08/2018 21:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 09/06/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, **Time** 23:41:24 **Software** 8100.27
JACOB,

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2018 23:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 10/09/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, **Time** 23:44:38 **Software** 8100.27
JACOB,

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/09/2018 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O.	Date 11/06/2018	Type Agency Inspection	
Inspector LYONS,BEEKMAN,	Time 19:44:19	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 646645
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 12/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.000
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance/HOSE NOT CON/RECONNECTED.

Data Download Date/Time 11/06/2018 21:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 12/03/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 03:29:41 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

LOOSE LID ON SIMULATOR, LID SECURED, RE-TESTED

Data Download Date/Time 12/03/2018 03:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O.	Date 01/06/2019	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 22:16:14	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 646645
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.124
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.198
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # 2	0.197
		0.20 g/210L Test # (Repeat) 2	0.198
		0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

05:TIGHTEN CONNECTIONS/SIM LID/RETESTED RETESTED/WITHIN	20:LID NOT TIGHT ON SIM/LID SECURED/RETESTED
--	--

Data Download Date/Time 01/06/2019 22:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O.	Date 01/14/2019	Type Department Inspection
Inspector GRAHAM, THOMAS, J	Time 10:58:01	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/14/2019 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 01/23/2018 **Type** Agency Inspection
Inspector FRANZ,GORDON,L **Time** 09:15:16 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/23/2018 09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 02/07/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 12:34:02	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG708807
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.083
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 02/07/2018 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 02/26/2018 **Type** Agency Inspection
Inspector FRANZ,GORDON,L **Time** 10:35:41 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2018 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 03/19/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 09:22:17	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 17817080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/19/2018 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 04/30/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 09:50:49	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 17817080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Purge Fail.

Data Download Date/Time 04/30/2018 10:45

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 05/30/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 11:23:08	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 17817080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2018 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 06/25/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 10:20:27	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 17817080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/25/2018 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 07/16/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 10:05:48	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 17817080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/16/2018 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 08/20/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 10:54:41	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 17817080A3
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 08/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 08/20/2018 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 08/20/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 15:14:41	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 17817080A3
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/20/2018 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 09/24/2018 **Type** Agency Inspection
Inspector FRANZ,GORDON,L **Time** 11:33:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 17817080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/24/2018 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 10/15/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 13:19:17	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 17817080A3
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/15/2018 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD	Date 11/06/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 12:12:02	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 17817080A3
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.190
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/06/2018 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 12/19/2018 **Type** Agency Inspection
Inspector FRANZ,GORDON,L **Time** 13:19:00 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.036
Minimum Sample Volume Check: OK		Interferent Detect Test # 1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test # 2	0.040
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 3	0.042
Alcohol Free Test: OK	Yes	Interferent Detect Test # 3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Ambient Fail. Non-compliance: .

Data Download Date/Time 12/19/2018 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 01/14/2019 **Type** Agency Inspection
Inspector FRANZ,GORDON,L **Time** 11:05:13 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 17817080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2019 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 01/08/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 19:24:08 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2018 19:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O.	Date 02/01/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 19:10:24	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 10/30/2018

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 02/01/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O.	Date 02/08/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 14:37:06	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG708807
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/08/2018 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 02/20/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 21:11:35 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/20/2018 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 03/26/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 22:03:06 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/26/2018 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 04/17/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 18:41:22 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 05/03/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 20:04:25 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/03/2018 20:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 06/05/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 22:25:42 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/05/2018 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 06/18/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 19:59:35 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/18/2018 20:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 07/02/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL, BEEKMAN, JACOB, **Time** 22:52:36 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/02/2018 22:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 08/08/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 21:12:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/08/2018 21:18

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 09/06/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, **Time** 21:40:29 **Software** 8100.27
JACOB,

0.05 Lot # 201703A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 03/07/2019 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.046
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.046
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:BAD 0.05 SOL 1ST TEST/2ND SOL UP TO TEMP .

Data Download Date/Time 09/06/2018 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency	OSCEOLA COUNTY S.O.	Date	09/06/2018	Type	Agency Inspection
Inspector	LYONS, DANIEL,BEEKMAN, JACOB,	Time	21:49:25	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	No
-----------------------------	-------------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence AbortedSequence AbortedSequence Aborted. Non-compliance:RETEST IN ERROR DURING PRINT PROCESS.

Data Download Date/Time 09/06/2018 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 09/06/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL, BEEKMAN, JACOB, **Time** 23:22:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2018 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 10/09/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, **Time** 23:29:42 **Software** 8100.27
JACOB,

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/09/2018 23:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 11/06/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, **Time** 18:25:23 **Software** 8100.27
JACOB,

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/06/2018 18:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 12/03/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 03:34:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/03/2018 03:51

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 01/07/2019 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 00:02:13 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/07/2019 00:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency HURLBURT FIELD PD	Date 01/04/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 11:53:41	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 01/04/2018 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency HURLBURT FIELD PD	Date 01/19/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 12:06:52	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG708807
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/19/2018 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency HURLBURT FIELD PD	Date 03/23/2018	Type Agency Inspection	
Inspector MCCOY,RICHARD,G	Time 15:28:44	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 869968
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 12/06/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.259
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.238
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.225
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 03/27/2018 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency HURLBURT FIELD PD	Date 06/05/2018	Type Agency Inspection	
Inspector SCHLAGER, JONATHAN, D	Time 18:30:16	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 869968
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 12/06/2020

Number of Simulators 4	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.20 g/210L Test # (Repeat) 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

20: LOOSE CYLINDER.

Data Download Date/Time 06/05/2018 19:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency HURLBURT FIELD PD	Date 07/17/2018	Type Agency Inspection
Inspector SCHLAGER, JONATHAN, D	Time 14:15:48	Software 8100.27

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 869968
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 12/06/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.191
		0.20 g/210L Test # (Repeat) 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A : . 20: Control Outside Tolerance.

Data Download Date/Time 07/17/2018 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency	OKALOOSA COUNTY SO	Date	04/25/2018	Type	Agency Inspection
Inspector	MURPHY,PATRICK,J	Time	13:32:34	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance Yes
-----------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SKIPPED AI TO OPERATE INSTRUMENT

Data Download Date/Time 04/26/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO	Date 04/26/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 11:41:05	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805702
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/26/2018 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 05/03/2018 **Type** Agency Inspection
Inspector JOHNSON, DAVID, L **Time** 11:24:23 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 666491
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/03/2018 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 06/06/2018 **Type** Agency Inspection
Inspector JOHNSON, DAVID, L **Time** 07:48:34 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 666491
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/06/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 07/04/2018 **Type** Agency Inspection
Inspector JOHNSON,DAVID,L **Time** 08:02:04 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 666491
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/04/2018 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 08/02/2018 **Type** Agency Inspection
Inspector JOHNSON, DAVID, L **Time** 07:51:04 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 666491
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/02/2018 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 09/27/2018 **Type** Agency Inspection
Inspector JOHNSON,DAVID,L **Time** 08:02:27 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 666491
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/27/2018 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO	Date 10/25/2018	Type Agency Inspection	
Inspector JOHNSON, DAVID, L	Time 07:37:37	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 666491
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 12/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.CLEARED AREA REPERFORMED TEST CHECKED GOOD

Data Download Date/Time 10/25/2018 08:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 11/01/2018 **Type** Agency Inspection
Inspector JOHNSON, DAVID, L **Time** 08:21:58 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 666491
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/01/2018 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 12/06/2018 **Type** Agency Inspection
Inspector JOHNSON,DAVID,L **Time** 07:57:06 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 666491
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/30/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/06/2018 09:04

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 01/29/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 09:10:07 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 31517080A1
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2018 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 02/28/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 15:23:31 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 31517080A1
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.191
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/01/2018 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 03/30/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 15:07:24 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 31517080A1
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/05/2018 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 04/26/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 10:49:38 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/26/2018 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD	Date 05/31/2018	Type Agency Inspection	
Inspector SCHNEIDER, ANDREW, G	Time 16:19:16	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 31517080A1
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 01/05/2020
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.191
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance LOOSE SIM SEAL.

Data Download Date/Time 05/31/2018 17:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD	Date 06/28/2018	Type Agency Inspection	
Inspector SCHNEIDER, ANDREW, G	Time 16:06:48	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 31517080A1
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 01/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

AF STILL ATTACHED DURING ACET INTF..

Data Download Date/Time 07/02/2018 17:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 07/31/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 15:06:16 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/31/2018 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 08/30/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 15:52:39 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/10/2018 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 09/27/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 15:06:09 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.073
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.073
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.073
		0.08 g/210L Test # (Repeat) 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:BAD SEAL (RETESTING).

Data Download Date/Time 10/04/2018 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 09/27/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 15:37:37 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/04/2018 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD	Date 10/31/2018	Type Agency Inspection	
Inspector SCHNEIDER, ANDREW, G	Time 16:12:21	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201808E	0.20 Lot # 201802G	0.08 Gas Lot # 31517080A1
Expiration 02/06/2020	Expiration 08/22/2020	Expiration 02/22/2020	Expiration 01/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.074
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Improper Sample: ADJUSTED SEAL RETEST.

Data Download Date/Time 11/05/2018 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 11/28/2018 **Type** Agency Inspection
Inspector SCHNEIDER, ANDREW, G **Time** 10:26:06 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/28/2018 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD	Date 11/30/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 10:36:09	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 11/30/2018 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO	Date 01/31/2018	Type Agency Inspection	
Inspector FELTOVIC,DEP,M	Time 07:51:52	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 09216080A2
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 05/05/2018
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 01/31/2018 07:48

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 02/23/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 08:01:27 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 09216080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 05/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 02/23/2018 07:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO	Date 03/08/2018	Type Department Inspection
Inspector PLATT,SHAYLA,D	Time 13:34:23	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG626604
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/08/2018 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 03/29/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 08:45:47 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 09216080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 05/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 03/29/2018 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency	SARASOTA COUNTY SO	Date	04/28/2018	Type	Agency Inspection		
Inspector	FELTOVIC,DEP,M	Time	07:47:58	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201611B	0.20 Lot #	201702B	0.08 Gas Lot #	09216080A2
Expiration	03/07/2019	Expiration	11/15/2018	Expiration	02/23/2019	Expiration	05/05/2018
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 04/28/2018 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 05/31/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 08:19:55 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 05/31/2018 08:20

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency	SARASOTA COUNTY SO	Date	06/29/2018	Type	Agency Inspection		
Inspector	FELTOVIC,DEP,M	Time	08:59:50	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201611B	0.20 Lot #	201702B	0.08 Gas Lot #	19817080A2
Expiration	03/07/2019	Expiration	11/15/2018	Expiration	02/23/2019	Expiration	09/05/2019
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 06/29/2018 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 07/30/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 14:57:20 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 07/30/2018 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 08/21/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 09:31:32 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 08/21/2018 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 09/30/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 09:29:45 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 09/30/2018 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 10/31/2018 **Type** Agency Inspection
Inspector CLINE,GLENN, **Time** 10:09:39 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

SGT SPURLOCK PRESENT

Data Download Date/Time 10/31/2018 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 11/27/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 07:51:50 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 11/27/2018 07:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 12/28/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 09:13:19 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 19817080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 12/28/2018 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 01/16/2018
Time 13:24:11

Type Agency Inspection
Software 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 32316080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2018 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD	Date 02/07/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 10:06:47	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG626604
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.083
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0007
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/07/2018 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 02/13/2018
Time 12:07:28

Type Agency Inspection
Software 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 32316080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2018 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 03/01/2018
Time 10:42:17

Type Agency Inspection
Software 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 32316080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/01/2018 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD	Date 04/30/2018	Type Agency Inspection	
Inspector DE JESUS, RAY,	Time 11:46:46	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 32316080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 01/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.043
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.044
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance TIGHTENED CAP ON SIMULATOR .

Data Download Date/Time 04/30/2018 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DEJESUS,RAY,

Date 05/01/2018
Time 11:14:31

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/01/2018 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency	NORTH MIAMI PD	Date	06/18/2018	Type	Agency Inspection		
Inspector	DE JESUS,RAY,N	Time	13:24:14	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201709A	0.20 Lot #	201702B	0.08 Gas Lot #	32316080A3
Expiration	03/07/2019	Expiration	09/12/2019	Expiration	02/23/2019	Expiration	01/05/2019
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/18/2018 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 07/11/2018
Time 10:10:35

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2018 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 08/13/2018
Time 11:58:37

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/13/2018 12:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD	Date 09/20/2018	Type Agency Inspection	
Inspector DE JESUS, RAY,	Time 11:36:08	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 32316080A3
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 01/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.073
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.073
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance-TIGHTENED SIMULATOR.

Data Download Date/Time 09/20/2018 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 10/18/2018
Time 10:54:03

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/18/2018 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD	Date 11/08/2018	Type Agency Inspection	
Inspector DE JESUS, RAY,	Time 11:36:40	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 32316080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 01/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.08 g/210L Test # (Repeat) 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: RFI Detect-CLEANING CREW IN HOLDING CELLS.

Data Download Date/Time 11/08/2018 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 12/10/2018
Time 14:11:59

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/10/2018 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 01/02/2019
Time 09:58:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 32316080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 01/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/02/2019 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD	Date 01/14/2019	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 09:01:25	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.204
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.204
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.204
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/14/2019 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO
Inspector SKIPPER, ROGER, G

Date 01/03/2018
Time 19:31:12

Type Agency Inspection
Software 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 01/03/2018 19:31

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency	HILLSBOROUGH CO SO	Date	02/13/2018	Type	Agency Inspection
Inspector	SKIPPER,ROGER,G	Time	18:31:59	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	5	Compliance	No
-----------------------------	---	-------------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # 2	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test # 3	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A/F SIM TEST STARTED W/O SIM CONNECTED. RESTARTED INSPECT. Non-compliance: AI R. GRANNIS #256794 PRESENT AND OBSERVED..

Data Download Date/Time 02/14/2018 09:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 02/13/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 19:13:11	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 03417080A1
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 03/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 02/14/2018 09:10

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 03/05/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 17:19:34	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 03417080A1
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 03/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 03/05/2018 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 04/05/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 18:49:58	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 03417080A1
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 03/05/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 04/05/2018 18:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 04/16/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 11:37:12	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805702
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 04/16/2018 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 05/03/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 17:01:20 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 03417080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 03/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 05/03/2018 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 06/19/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 19:03:55	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 31517080A1
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 01/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 06/19/2018 19:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 07/12/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 19:05:53	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 31517080A1
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 01/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 07/12/2018 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 08/07/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 13:40:10	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 31517080A1
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 01/05/2020

Number of Simulators 4	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 08/07/2018 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 09/06/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 12:58:30	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 31517080A1
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 01/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 09/06/2018 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO	Date 10/04/2018	Type Agency Inspection	
Inspector SKIPPER, ROGER, G	Time 13:28:57	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 31517080A1
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 10/04/2018 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 11/01/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 10:28:44 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 11/01/2018 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 12/18/2018 **Type** Agency Inspection
Inspector SKIPPER, ROGER, G **Time** 17:48:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 12/18/2018 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 01/10/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 17:48:36 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 01/10/2019 17:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency PATRICK AFB **Date** 01/24/2018 **Type** Agency Inspection
Inspector OUVERSON,JESSE,T **Time** 13:34:14 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 14968
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/06/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/24/2018 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency PATRICK AFB **Date** 03/02/2018 **Type** Agency Inspection
Inspector OUVERSON,JESSE,T **Time** 14:26:53 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 14968
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/06/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/02/2018 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency PATRICK AFB	Date 04/06/2018	Type Agency Inspection	
Inspector OUVERSON,JESSE,T	Time 13:40:16	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 14968
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 08/06/2018

Number of Simulators 4	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

08: TUBE TO CALIBRATION INLET DISCONNECTED. RE-CONNECTED .

Data Download Date/Time 04/06/2018 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency PATRICK AFB **Date** 05/29/2018 **Type** Agency Inspection
Inspector OUVERSON,JESSE,T **Time** 14:06:29 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 14968
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/06/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/05/2018 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency PATRICK AFB
Inspector OUVERSON,JESSE,T

Date 07/11/2018
Time 11:40:52

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 14968
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/06/2018

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2018 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency PATRICK AFB
Inspector OUVERSON,JESSE,T

Date 08/13/2018
Time 12:56:54

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 894572
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/21/2021

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/13/2018 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency PATRICK AFB
Inspector OUVERSON,JESSE,T

Date 09/24/2018
Time 09:42:57

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 894572
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/17/2021

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/24/2018 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency	PATRICK AFB	Date	12/04/2018	Type	Agency Inspection
Inspector	PLATT,SHAYLA,D	Time	09:17:20	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed. BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 12/07/2018 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency PATRICK AFB	Date 12/07/2018	Type Department Inspection
Inspector PLATT,SHAYLA,D	Time 11:28:36	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 12/07/2018 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FWC	Date 01/17/2018	Type Agency Inspection	
Inspector PLUSSA, KYLE, A	Time 11:27:21	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 279870
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 02/04/2019

Number of Simulators 5	Compliance Yes
------------------------	----------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.015
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.018
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.019
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance. AI COMPLETED AND PASSED.

Data Download Date/Time 01/18/2018 08:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC	Date 01/18/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 08:18:11	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG626604
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.084
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.207
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.083
		0.20 g/210L Test # 5	0.207
		0.08 g/210L Dry Gas Std Test # 5	0.082
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.207
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.207
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.206
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.205
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.205
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/18/2018 08:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 02/02/2018 **Type** Agency Inspection
Inspector PLUSSA,KYLE,A **Time** 14:03:08 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 279870
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/04/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A.I. COMPLETED AND PASSED.

Data Download Date/Time 02/02/2018 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 03/15/2018 **Type** Agency Inspection
Inspector PLUSSA,KYLE,A **Time** 12:45:11 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 279870
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/04/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A.I. COMPLETED AND PASSED.

Data Download Date/Time 04/05/2018 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 04/05/2018 **Type** Agency Inspection
Inspector PLUSSA,KYLE,A **Time** 13:41:25 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 279870
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 02/04/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A.I. COMPLETED AND PASSED.

Data Download Date/Time 04/05/2018 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 05/31/2018 **Type** Agency Inspection
Inspector BLANCO,JUAN,S **Time** 11:40:34 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 279870
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/04/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2018 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 06/27/2018 **Type** Agency Inspection
Inspector BLANCO,JUAN,S **Time** 19:44:41 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 279870
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/04/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI

Data Download Date/Time 06/27/2018 20:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC	Date 07/31/2018	Type Agency Inspection	
Inspector BLANCO,JUAN,S	Time 17:59:38	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 279870
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 02/04/2019
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI

Data Download Date/Time 07/31/2018 18:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 08/20/2018 **Type** Agency Inspection
Inspector BLANCO,JUAN,S **Time** 22:50:31 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 279870
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/04/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI

Data Download Date/Time 08/20/2018 22:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 09/26/2018 **Type** Agency Inspection
Inspector BLANCO,JUAN,S **Time** 20:43:22 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 279870
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/04/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI

Data Download Date/Time 09/26/2018 20:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency	FISH & WILDLIFE CC	Date	10/29/2018	Type	Agency Inspection		
Inspector	BLANCO,JUAN,S	Time	18:50:45	Software	8100.27		
0.05 Lot #	201802B	0.08 Lot #	201709A	0.20 Lot #	201802G	0.08 Gas Lot #	279870
Expiration	02/06/2020	Expiration	09/12/2019	Expiration	02/22/2020	Expiration	02/04/2019
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI

Data Download Date/Time 10/29/2018 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency	FISH & WILDLIFE CC	Date	11/28/2018	Type	Agency Inspection		
Inspector	BLANCO,JUAN,S	Time	16:56:23	Software	8100.27		
0.05 Lot #	201802B	0.08 Lot #	201709A	0.20 Lot #	201802G	0.08 Gas Lot #	279870
Expiration	02/06/2020	Expiration	09/12/2019	Expiration	02/22/2020	Expiration	02/04/2019
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI

Data Download Date/Time 11/28/2018 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 12/28/2018 **Type** Agency Inspection
Inspector BLANCO,JUAN,S **Time** 13:47:44 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 279870
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/04/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI

Data Download Date/Time 12/28/2018 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 01/15/2018 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 16:14:27 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 17316080A2
Expiration 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 08/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/15/2018 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO	Date 01/22/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 14:32:15	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG708807
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.082
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.082
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.083
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 01/22/2018 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO	Date 04/24/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 13:12:20	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805702
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 04/24/2018 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO	Date 04/27/2018	Type Agency Inspection	
Inspector COLLITO, BRAD, R	Time 04:56:26	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 34917080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

A F / M A: Range Exceeded Range Exceeded. Int Det: RFI Detect.

Data Download Date/Time 04/27/2018 05:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 05/13/2018 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 08:50:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: SAMPLE INTERRUPTED.

Data Download Date/Time 05/13/2018 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 06/23/2018 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 08:25:00 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded.

Data Download Date/Time 06/23/2018 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 07/22/2018 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 11:38:42 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/22/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 08/19/2018 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 15:26:03 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded.

Data Download Date/Time 08/19/2018 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 09/15/2018 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 11:11:24 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/15/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO	Date 10/19/2018	Type Agency Inspection	
Inspector COLLITO, BRAD, R	Time 20:18:26	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 34917080A3
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance TUBE DISCONNECTED.

Data Download Date/Time 10/19/2018 20:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 11/21/2018 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 20:03:33 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/21/2018 20:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 12/10/2018 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 20:20:40 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/10/2018 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 01/09/2019 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 19:23:30 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/09/2019 19:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO	Date 01/25/2018	Type Agency Inspection	
Inspector DEANE,M,	Time 23:50:32	Software 8100.27	

0.05 Lot # 201603D	0.08 Lot # 201611B	0.20 Lot # 201604C	0.08 Gas Lot # AG629802
Expiration 03/08/2018	Expiration 11/15/2018	Expiration 04/05/2018	Expiration 10/24/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside ToleranceUNK. PROB. REPEAT OK.

Data Download Date/Time 01/25/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 02/16/2018
Time 23:49:26

Type Agency Inspection
Software 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** AG629802
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 10/24/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2018 23:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO	Date 03/14/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 14:28:20	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/14/2018 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 03/22/2018
Time 23:47:16

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** AG629802
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 10/24/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/22/2018 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 04/28/2018
Time 02:59:12

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/28/2018 03:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 05/30/2018
Time 21:03:41

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2018 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 06/28/2018
Time 20:45:53

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/28/2018 20:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 07/25/2018
Time 20:37:23

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/25/2018 20:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO	Date 08/02/2018	Type Agency Inspection	
Inspector DEANE, M,	Time 20:54:46	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # AG728303
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 10/10/2019

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance UNK. PROB. REPEAT OK.

Data Download Date/Time 08/02/2018 21:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO	Date 09/12/2018	Type Agency Inspection	
Inspector DEANE,M,	Time 21:06:50	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # AG728303
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 10/10/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance UNK PROB REPEAT OK.

Data Download Date/Time 09/12/2018 21:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 10/25/2018 **Type** Agency Inspection
Inspector DEANE,M, **Time** 01:13:55 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2018 01:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 11/28/2018
Time 23:48:17

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2018 00:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 12/26/2018
Time 21:08:17

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG728303
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/10/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/26/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT
Inspector GRAY,LIN,

Date 01/29/2018
Time 14:59:39

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** AG619605
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/14/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail/ALCOHOL NEAR INSTRUMENT WHEN PURGING.

Data Download Date/Time 01/29/2018 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT
Inspector GRAY,LIN,

Date 02/26/2018
Time 15:20:40

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** AG619605
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/14/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS STARTED AT 1315 AND INSPECTION BEGAN AT 1415.

Data Download Date/Time 02/26/2018 15:23

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT
Inspector GRAY,LIN,

Date 03/27/2018
Time 15:12:53

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** AG619605
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/14/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS STARTED AT 1100 AND INSPECTION BEGAN AT 1330

Data Download Date/Time 03/27/2018 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT
Inspector GRAY,LIN,

Date 04/23/2018
Time 15:11:05

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** AG619605
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/14/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.191
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS STARTED AT 1100 AND INSPECTION BEGAN AT 1400.

Data Download Date/Time 04/23/2018 15:18

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT	Date 05/25/2018	Type Agency Inspection	
Inspector GRAY,LIN,	Time 17:33:50	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # AG619605
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 07/14/2018

Number of Simulators 4	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS STARTED AT 1000 AND INSPECTION BEGAN AT 1600.

Data Download Date/Time 05/25/2018 17:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT
Inspector GRAY,LIN,

Date 06/29/2018
Time 11:29:18

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** AG619605
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 07/14/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS STARTED AT 0900 AND INSPECTION BEGAN AT 1000.

Data Download Date/Time 06/29/2018 11:52

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT	Date 07/06/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 11:03:32	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805702
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0008 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0015
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 07/06/2018 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT
Inspector GRAY,LIN,

Date 07/25/2018
Time 11:18:56

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS STARTED AT 0830 AND INSPECTION BEGAN AT 1030.

Data Download Date/Time 07/25/2018 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT	Date 08/29/2018	Type Agency Inspection	
Inspector GRAY,LIN,	Time 11:45:19	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # AG815602
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 06/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS STARTED AT 0800 AND INSPECTION BEGAT AT 1045.

Data Download Date/Time 08/29/2018 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT	Date 09/28/2018	Type Agency Inspection	
Inspector GRAY,LIN,	Time 17:53:22	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # AG815602
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 06/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect. BILL'S LIFE VEST.

Data Download Date/Time 10/03/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT	Date 10/17/2018	Type Agency Inspection	
Inspector GRAY,LIN,	Time 19:16:04	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # AG815602
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 06/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.204
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.204
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # (Repeat) 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

20: RFI Detect/CELL PHONE RANG CLOSE TO INSTRUMENT..

Data Download Date/Time 10/17/2018 20:20

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT	Date 11/28/2018	Type Agency Inspection	
Inspector RICKETSON, JAY, R	Time 15:10:48	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # AG815602
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 06/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SIMULATORS STARTED AT 0830 AND INSPECTION BEGAN AT 1415.

Data Download Date/Time 11/28/2018 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 12/28/2018 **Type** Agency Inspection
Inspector CEPHUS,JAMIN,I **Time** 23:09:37 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: . SAMPLE NOT .

Data Download Date/Time 01/02/2019 18:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT
Inspector CEPHUS,JAMIN,I

Date 12/29/2018
Time 00:06:38

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** AG815602
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

ALL 5 SIMS STARTED AT 1940 ALL ARE WORKING PROPERLY

Data Download Date/Time 01/02/2019 18:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD	Date 01/17/2018	Type Agency Inspection	
Inspector WEISKOPF,MICHAEL,D	Time 08:58:37	Software 8100.27	
0.05 Lot # 201603D	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 26316080A2
Expiration 03/08/2018	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 11/05/2018
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/17/2018 09:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 02/05/2018 **Type** Agency Inspection
Inspector WEISKOPF,MICHAEL,D **Time** 10:38:46 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/05/2018 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD	Date 03/06/2018	Type Agency Inspection	
Inspector WEISKOPF,MICHAEL,D	Time 08:58:12	Software 8100.27	

0.05 Lot # 201603D	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 26316080A2
Expiration 03/08/2018	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 11/05/2018

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INCOMPLIANCE

Data Download Date/Time 03/06/2018 08:59

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD	Date 03/19/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 12:31:28	Software 8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators 5	Average Standard Deviation	Compliance No
-------------------------------	-----------------------------------	----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	INT
Minimum Sample Volume Check: OK	Yes	Interferent Detect Test # (Repeat) 1	INT
Barometric Pressure Sensor Check: OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	INT
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	INT
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	INT
0.05 g/210L Test: OK	No	Alcohol Free Test # 4	0.000
0.08 g/210L Test: OK	No	Interferent Detect Test # 4	INT
0.20 g/210L Test: OK	No	Interferent Detect Test # (Repeat) 4	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 5	0.000
		Interferent Detect Test # 5	0.000
		Interferent Detect Test # (Repeat) 5	INT
		Alcohol Free Test # 6	0.000
		Interferent Detect Test # 6	0.000
		Interferent Detect Test # (Repeat) 6	INT
		Alcohol Free Test # 7	0.000
		Interferent Detect Test # 7	0.000
		Interferent Detect Test # (Repeat) 7	0.000
		Alcohol Free Test # 8	0.000
		Interferent Detect Test # 8	0.000
		Interferent Detect Test # (Repeat) 8	INT
		Alcohol Free Test # 9	0.000
		Interferent Detect Test # 9	0.000
		Interferent Detect Test # (Repeat) 9	INT
		Alcohol Free Test # 10	0.000
		Interferent Detect Test # 10	0.000
		Interferent Detect Test # (Repeat) 10	INT

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Remarks:

Int Det: ADJUSTED SIM, RETESTING, UNKNOWN ISSUE. WILL RECAL. Non-compliance: UNKNOWN INT ISSUE. WILL RECAL ADJUST..

Data Download Date/Time 03/19/2018 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD	Date 07/03/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 10:41:42	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 07/03/2018 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 07/11/2018 **Type** Agency Inspection
Inspector WEISKOPF,MICHAEL,D **Time** 10:10:37 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INCOMPLIANCE

Data Download Date/Time 07/11/2018 10:17

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 08/21/2018 **Type** Agency Inspection
Inspector WEISKOPF,MICHAEL,D **Time** 09:09:44 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 08/21/2018 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 09/10/2018 **Type** Agency Inspection
Inspector WEISKOPF,MICHAEL,D **Time** 09:42:49 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 09/10/2018 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD	Date 10/15/2018	Type Agency Inspection	
Inspector WEISKOPF,MICHAEL,D	Time 08:33:12	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 26316080A2
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 11/05/2018

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 10/15/2018 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 11/20/2018 **Type** Agency Inspection
Inspector WEISKOPF,MICHAEL,D **Time** 09:55:02 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 11/20/2018 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 12/24/2018 **Type** Agency Inspection
Inspector WEISKOPF,MICHAEL,D **Time** 09:05:49 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 12/24/2018 09:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency	PORT ORANGE PD	Date	09/25/2018	Type	Agency Inspection
Inspector	MURPHY,PATRICK,J	Time	12:02:01	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SKIPPED AI TO OPERATE INSTRUMENT

Data Download Date/Time 09/26/2018 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD	Date 09/26/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 11:37:24	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 09/26/2018 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 10/30/2018 **Type** Agency Inspection
Inspector HARRISON, WILLIAM, C **Time** 11:08:15 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 609878
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/30/2018 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD	Date 11/30/2018	Type Agency Inspection	
Inspector HARRISON, WILLIAM, C	Time 17:52:50	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 609878
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 11/01/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.05 g/210L Test # (Repeat) 3	0.045
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 11/30/2018 17:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD	Date 12/31/2018	Type Agency Inspection	
Inspector HARRISON, WILLIAM, C	Time 16:21:45	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201808E	0.20 Lot # 201802G	0.08 Gas Lot # 609878
Expiration 02/06/2020	Expiration 08/22/2020	Expiration 02/22/2020	Expiration 11/01/2019

Number of Simulators 4	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 12/31/2018 16:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 01/31/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 01:10:27 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 01/31/2018 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency	OCALA POLICE DEPT	Date	03/15/2018	Type	Agency Inspection
Inspector	SHANAHAN,JAKE,L	Time	08:15:33	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 03/15/2018 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency Ocala Police Dept	Date 03/15/2018	Type Department Inspection
Inspector SHANAHAN, JAKE, L	Time 10:20:32	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.052
		0.08 g/210L Test # 4	0.084
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.052
		0.08 g/210L Test # 5	0.083
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.052
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.053
0.08 g/210L Test # 10	0.084
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/15/2018 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 03/29/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 02:21:24 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 04/01/2018 02:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency	OCALA POLICE DEPT	Date	04/20/2018	Type	Agency Inspection		
Inspector	JOEDICKE, WILLIAM, R	Time	18:09:02	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201611B	0.20 Lot #	201702B	0.08 Gas Lot #	26316080A2
Expiration	03/07/2019	Expiration	11/15/2018	Expiration	02/23/2019	Expiration	11/05/2018
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 05/03/2018 21:30

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 05/31/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 19:39:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 06/07/2018 04:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency	OCALA POLICE DEPT	Date	06/30/2018	Type	Agency Inspection		
Inspector	JOEDICKE, WILLIAM, R	Time	02:07:04	Software	8100.27		
0.05 Lot #	201802B	0.08 Lot #	201611B	0.20 Lot #	201702B	0.08 Gas Lot #	26316080A2
Expiration	02/06/2020	Expiration	11/15/2018	Expiration	02/23/2019	Expiration	11/05/2018
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 06/30/2018 18:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 07/25/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 09:44:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 07/25/2018 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 08/29/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 15:09:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 02/06/2020 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 08/31/2018 22:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 09/30/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 21:55:05 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded TOO MUCH MA.

Data Download Date/Time 10/05/2018 18:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 10/31/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 21:17:00 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 26316080A2
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 11/08/2018 22:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 11/30/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 22:46:09 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 24818080A2
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 01/02/2019 06:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 12/31/2018 **Type** Agency Inspection
Inspector JOEDICKE, WILLIAM, R **Time** 17:03:17 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 24818080A2
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded TOO MUCH MA.

Data Download Date/Time 01/02/2019 06:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency Ocala Police Dept	Date 01/09/2019	Type Department Inspection
Inspector GRAHAM, THOMAS, J	Time 12:16:02	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

--

Data Download Date/Time 01/09/2019 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 01/10/2018
Time 14:03:36

Type Agency Inspection
Software 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 09216080A2
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 05/05/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/10/2018 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 02/14/2018
Time 12:13:00

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 09216080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 05/05/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/14/2018 12:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 03/07/2018
Time 11:20:57

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 09216080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 05/05/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/07/2018 11:26

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD	Date 04/25/2018	Type Department Inspection
Inspector PLATT,SHAYLA,D	Time 11:02:46	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/25/2018 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 04/27/2018
Time 11:03:34

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 23016080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/27/2018 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 05/09/2018
Time 14:55:52

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 2901708A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/09/2018 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency	FORT MYERS PD	Date	06/06/2018	Type	Agency Inspection		
Inspector	CLAFLIN,DAVID,S	Time	10:40:22	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201709A	0.20 Lot #	201702B	0.08 Gas Lot #	29017080A2
Expiration	03/07/2019	Expiration	09/12/2019	Expiration	02/23/2019	Expiration	12/05/2019
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/06/2018 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 07/18/2018
Time 13:18:48

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 10818080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/18/2018 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 08/01/2018 **Type** Agency Inspection
Inspector CLAFLIN, DAVID, S **Time** 12:07:17 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 10818080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/20/2020 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/01/2018 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 09/05/2018
Time 09:44:54

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 10818080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/05/2018 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 10/03/2018 **Type** Agency Inspection
Inspector CLAFLIN, DAVID, S **Time** 12:16:39 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 10818080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/03/2018 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 11/14/2018
Time 12:47:08

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 10818080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2020

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/14/2018 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 12/12/2018 **Type** Agency Inspection
Inspector CLAFLIN, DAVID, S **Time** 12:59:10 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 10818080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/12/2018 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 01/09/2019 **Type** Agency Inspection
Inspector CLAFLIN, DAVID, S **Time** 14:55:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 10818080A4
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/09/2019 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 01/17/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 17:47:06 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A6
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2018 17:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 02/19/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 16:45:51 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A6
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2018 16:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD	Date 03/06/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 15:07:03	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.052
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

--

Data Download Date/Time 03/06/2018 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 03/14/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 17:13:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A6
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/14/2018 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD	Date 04/16/2018	Type Agency Inspection	
Inspector BALDWIN, DEAN,	Time 16:51:16	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 12716080A6
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 06/05/2018
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/16/2018 16:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 05/08/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 17:23:19 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A6
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2018 17:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 06/20/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 11:18:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2018 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 07/18/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 12:35:31 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/18/2018 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 08/16/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 17:22:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/16/2018 17:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 09/17/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 16:37:35 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/17/2018 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD	Date 10/18/2018	Type Agency Inspection	
Inspector BALDWIN, DEAN, -	Time 16:44:51	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 29017080A2
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 12/05/2019

Number of Simulators 4	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance FAILED TO CONNECT HOSE.

Data Download Date/Time 10/18/2018 16:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 11/12/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 16:34:52 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/12/2018 16:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 12/11/2018 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 16:41:45 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 29017080A2
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/11/2018 16:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 01/28/2018 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 15:14:30 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	INT
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Interferent Detect. Non-compliance:INTERFERT DURING ALCOHOL FREE.

Data Download Date/Time 01/28/2018 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 01/28/2018 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 16:17:56 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 796576
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 06/14/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/28/2018 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 02/25/2018 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 22:46:42 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 796576
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 06/14/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/25/2018 23:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 03/28/2018 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 16:54:16 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 796576
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/14/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2018 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 04/27/2018 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 14:02:47 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 796576
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/14/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/27/2018 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 05/29/2018 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 17:04:46 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 796576
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/14/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/29/2018 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 06/27/2018 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 13:29:53 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 796576
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/14/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2018 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 07/27/2018 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 14:54:05 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 796576
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/14/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/27/2018 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD	Date 08/22/2018	Type Agency Inspection	
Inspector DUNNE,JUSTIN,R	Time 21:10:00	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 796576
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 06/14/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 08/22/2018 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD	Date 08/28/2018	Type Department Inspection
Inspector MURPHY,PATRICK,J	Time 12:38:31	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 08/28/2018 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 01/03/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 01:56:38 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 17316080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/03/2018 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 02/15/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 16:32:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 17316080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.208
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/15/2018 16:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 03/13/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 18:00:00 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.210
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/13/2018 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD	Date 04/02/2018	Type Department Inspection
Inspector SHANAHAN, JAKE, L	Time 12:14:30	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805702
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.206
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.083
		0.20 g/210L Test # 5	0.206
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.206
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.205
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.206
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.206
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.206
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/02/2018 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 04/05/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 16:06:35 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/05/2018 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 05/15/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 15:44:09 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/15/2018 15:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 06/13/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 14:41:48 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 17316080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/13/2018 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 07/09/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 07:25:06 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 06717080A7
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/09/2018 07:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 08/08/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 22:09:26 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 06717080A7
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/09/2018 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 09/25/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 15:35:23 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 06717080A7
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/25/2018 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 10/23/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 12:05:18 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 06717080A7
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2018 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 11/13/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 16:05:04 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 06717080A7
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/13/2018 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 12/05/2018 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 06:59:46 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 06717080A7
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2018 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 01/04/2019 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 12:02:29 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 06717080A7
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/04/2019 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 01/12/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:34:04 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/12/2018 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 02/08/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:25:23 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/08/2018 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD	Date 02/13/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 10:10:39	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG626605
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 09/22/2018

Number of Simulators 5 **Average Standard Deviation** 0.0009 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0010
0.20 g/210L Test (Std Dev)	0.0017
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/13/2018 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 02/19/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:30:26 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2018 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD	Date 03/16/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 09:54:15	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201604C	0.08 Gas Lot # 20016080A2
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 04/05/2018	Expiration 09/05/2018
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/16/2018 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 04/06/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:18:06 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/06/2018 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 05/02/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:06:28 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/02/2018 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 06/01/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:49:34 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/01/2018 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 07/04/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:28:30 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/04/2018 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 08/14/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:40:38 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/14/2018 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD	Date 09/06/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 13:01:06	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 34917080A4
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 02/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.177
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.177
		0.20 g/210L Test # (Repeat) 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance/TIGHTENED SIM/ORING CHANGED

Data Download Date/Time 09/06/2018 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 09/10/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:02:34 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/10/2018 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD	Date 09/18/2018	Type Department Inspection
Inspector REYES-RIVERA,DAVID,E	Time 08:57:11	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 05/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 09/18/2018 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 10/04/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:16:10 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/04/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 11/06/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:54:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/06/2018 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 12/13/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:03:10 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/13/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 01/24/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:32:40 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201601F **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 01/26/2018 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/24/2018 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 02/19/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:40:54 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2018 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 03/16/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:55:19 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/16/2018 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 04/17/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:24:32 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 05/15/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:08:03 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/15/2018 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 06/13/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:41:48 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/13/2018 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 07/19/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:22:28 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2018 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD	Date 07/24/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 08:05:28	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 07/24/2018 08:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD	Date 08/07/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 13:17:31	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 20016080A2
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 09/05/2018

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Interferent Detect/CONNECTING HOSE DETECTED HAND SMELL .

Data Download Date/Time 08/07/2018 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 09/06/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:08:44 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2018 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 10/18/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:13:13 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/18/2018 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 11/09/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:30:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/09/2018 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD	Date 12/07/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 09:04:38	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 34917080A4
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 02/05/2020
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/07/2018 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 01/11/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:26:14 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/11/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 09/12/2018 **Type** Agency Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 06:16:36 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COMPLIANCE NOT DETERMINED AI NOT CONDUCTED

Data Download Date/Time 09/12/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD	Date 09/12/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 14:51:19	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 09/12/2018 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD	Date 10/05/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 13:24:17	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 34917080A4
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 02/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.186
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.197
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Alcohol Free Test: OK	Yes	Interferent Detect Test # 1	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.074
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.076
		0.20 g/210L Test # 2	0.188
		0.20 g/210L Test # (Repeat) 2	0.198
		0.08 g/210L Dry Gas Std Test # 2	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.073
		0.08 g/210L Test # (Repeat) 3	0.076
		0.20 g/210L Test # 3	0.189
		0.20 g/210L Test # (Repeat) 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance/CHANGED SIM. 20: Control Outside Tolerance/TIGHTENED SIM. 08: Control Outside Tolerance/FORGOT TO ATTACH HOSE.

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Data Download Date/Time 10/08/2018 13:34

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD	Date 11/30/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 12:07:10	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 34917080A4
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 02/05/2020

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance/FORGOT TO ATTACH HOSE.

Data Download Date/Time 12/05/2018 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD	Date 12/05/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 10:25:01	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 34917080A4
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 02/05/2020
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2018 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 01/10/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:34:56 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/10/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD	Date 01/04/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 10:16:53	Software 8100.27	
0.05 Lot # 201603D	0.08 Lot # 201601F	0.20 Lot # 201604C	0.08 Gas Lot # 20016080A2
Expiration 03/08/2018	Expiration 01/26/2018	Expiration 04/05/2018	Expiration 09/05/2018
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/04/2018 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 02/08/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:22:21 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/08/2018 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/08/2018 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 03/07/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:30:07 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201604C **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 04/05/2018 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/07/2018 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 04/17/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:39:10 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD	Date 04/30/2018	Type Department Inspection
Inspector REYES-RIVERA, DAVID, E	Time 10:28:03	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715502
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.052
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.205
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.052
		0.08 g/210L Test # 5	0.084
		0.20 g/210L Test # 5	0.205
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.084
0.20 g/210L Test # 6	0.205
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.052
0.08 g/210L Test # 7	0.084
0.20 g/210L Test # 7	0.205
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.084
0.20 g/210L Test # 8	0.205
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.084
0.20 g/210L Test # 9	0.205
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.053
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.205
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/30/2018 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 05/03/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:36:45 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/03/2018 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 06/01/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:56:49 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/01/2018 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 07/12/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:36:02 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 20016080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/12/2018 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD	Date 08/14/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 09:42:43	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 20016080A2
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 09/05/2018
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/14/2018 09:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 09/06/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:17:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2018 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 09/10/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:05:38 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/10/2018 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD	Date 09/19/2018	Type Department Inspection
Inspector REYES-RIVERA,DAVID,E	Time 10:44:03	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.206
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.083
		0.20 g/210L Test # 5	0.206
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.205
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.052
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.206
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.084
0.20 g/210L Test # 8	0.206
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.084
0.20 g/210L Test # 9	0.206
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.084
0.20 g/210L Test # 10	0.206
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0010
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 09/19/2018 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD	Date 10/05/2018	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 12:37:30	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot # 34917080A4
Expiration 03/07/2019	Expiration 09/12/2019	Expiration 02/23/2019	Expiration 02/05/2020

Number of Simulators 5	Compliance Yes
------------------------	----------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:NOT ENOUGH MA

Data Download Date/Time 10/05/2018 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 11/06/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:45:52 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/06/2018 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 12/07/2018 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 08:55:26 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/07/2018 09:35

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency	FORT MYERS PD	Date	03/12/2018	Type	Agency Inspection
Inspector	PLATT,SHAYLA,D	Time	09:35:25	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 03/13/2018 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD	Date 03/13/2018	Type Department Inspection
Inspector PLATT,SHAYLA,D	Time 10:18:39	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG715202
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 06/01/2019

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/13/2018 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 04/27/2018 **Type** Agency Inspection
Inspector CLAFLIN, DAVID, S **Time** 12:33:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 23016080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.075
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: TOO MUCH M/A.

Data Download Date/Time 04/27/2018 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 05/09/2018 **Type** Agency Inspection
Inspector CLAFLIN, DAVID, S **Time** 13:47:32 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 26316080A1
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/09/2018 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD	Date 05/23/2018	Type Department Inspection
Inspector PLATT,SHAYLA,D	Time 10:31:47	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/23/2018 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 06/06/2018
Time 12:00:17

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/08/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 07/18/2018
Time 10:09:47

Type Agency Inspection
Software 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 29017080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/18/2018 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 08/01/2018
Time 10:24:18

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 29017080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/20/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/01/2018 10:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 09/05/2018
Time 11:01:57

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 26316080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/05/2018 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 10/03/2018
Time 11:36:13

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 26316080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/03/2018 11:40

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 11/14/2018
Time 11:38:19

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 29017080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/14/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 12/12/2018
Time 11:48:50

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 29017080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/12/2018 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD
Inspector CLAFLIN, DAVID, S

Date 01/09/2019
Time 11:05:14

Type Agency Inspection
Software 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 29017080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/09/2019 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 01/23/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 08:35:06	Software 8100.27	
0.05 Lot # 201603D	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 17817080A3
Expiration 03/08/2018	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A : . 05: Control Outside Tolerance.

Data Download Date/Time 01/23/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 03/19/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 11:59:29	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG708807
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/19/2018 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD **Date** 04/30/2018 **Type** Agency Inspection
Inspector FRANZ,GORDON,L **Time** 11:43:59 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/30/2018 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD **Date** 05/30/2018 **Type** Agency Inspection
Inspector FRANZ,GORDON,L **Time** 10:42:35 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 17817080A3
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2018 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 06/25/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 10:56:34	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 17817080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/25/2018 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 07/16/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 08:56:33	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 17817080A3
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/16/2018 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 08/20/2018	Type Agency Inspection
Inspector FRANZ,GORDON,L	Time 08:59:20	Software 8100.27

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot #	0.08 Gas Lot #
Expiration 02/06/2020	Expiration 09/12/2019	Expiration	Expiration

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.074
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.071
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.075
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.074
		0.08 g/210L Test # (Repeat) 3	0.075
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 08/20/2018 09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 08/20/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 16:36:31	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 17817080A3
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/20/2018 16:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 09/24/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 10:24:31	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 17817080A3
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/24/2018 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 10/15/2018	Type Agency Inspection	
Inspector FRANZ,GORDON,L	Time 13:59:59	Software 8100.27	
0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot # 17817080A3
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration 08/05/2019
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/15/2018 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD	Date 11/06/2018	Type Agency Inspection
Inspector FRANZ, GORDON, L	Time 11:21:44	Software 8100.27

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201802G	0.08 Gas Lot #
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/22/2020	Expiration

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.186
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.190
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.187
0.20 g/210L Test: OK	No	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.189
		0.20 g/210L Test # (Repeat) 3	0.185
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 11/06/2018 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency LAKELAND PD **Date** 12/19/2018 **Type** Agency Inspection
Inspector FRANZ,GORDON,L **Time** 13:50:14 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.029
Minimum Sample Volume Check: OK		Interferent Detect Test # 1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test # 2	0.034
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 3	0.044
Alcohol Free Test: OK	Yes	Interferent Detect Test # 3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Ambient Fail. Non-compliance: .

Data Download Date/Time 12/19/2018 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency	OSCEOLA COUNTY SO	Date	01/08/2018	Type	Agency Inspection		
Inspector	LYONS,DANIEL,E	Time	19:19:30	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201611B	0.20 Lot #	201702B	0.08 Gas Lot #	163437
Expiration	03/07/2019	Expiration	11/15/2018	Expiration	02/23/2019	Expiration	10/30/2018
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2018 19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO	Date 01/22/2018	Type Department Inspection
Inspector BELL,DANIELLE,M	Time 13:57:55	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG708807
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 01/22/2018 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency	OSCEOLA COUNTY SO	Date	02/01/2018	Type	Agency Inspection		
Inspector	LYONS,DANIEL,E	Time	19:45:40	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201611B	0.20 Lot #	201702B	0.08 Gas Lot #	163437
Expiration	03/07/2019	Expiration	11/15/2018	Expiration	02/23/2019	Expiration	10/30/2018
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2018 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO	Date 02/20/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 19:14:43	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 10/30/2018

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.076
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 02/20/2018 19:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO	Date 03/26/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 21:04:16	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # 2	0.193
		0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.075
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det: . 08: Control Outside Tolerance.

Data Download Date/Time 03/26/2018 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 04/17/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 18:50:23 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2018 18:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO	Date 05/03/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 20:12:39	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 10/30/2018
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/03/2018 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency	OSCEOLA COUNTY SO	Date	06/05/2018	Type	Agency Inspection
Inspector	LYONS,DANIEL,E	Time	20:37:11	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	5	Compliance	Yes
-----------------------------	---	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:SIM NOT HOOKED UP.

Data Download Date/Time 06/05/2018 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 06/05/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 21:28:28 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/05/2018 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO	Date 06/18/2018	Type Agency Inspection	
Inspector LYONS,DANIEL,E	Time 20:05:14	Software 8100.27	
0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 163437
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 10/30/2018
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.044
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.044
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance. RESET SIM LID/CONN RETESTED

Data Download Date/Time 06/18/2018 20:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 07/02/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, **Time** 21:39:55 **Software** 8100.27
JACOB,

0.05 Lot # 201802B **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 02/06/2018 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/02/2018 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 08/08/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 21:18:55 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Ambient Fail.

Data Download Date/Time 08/08/2018 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 09/06/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, JACOB, **Time** 23:46:58 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 163437
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/30/2018

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2018 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 10/09/2018 **Type** Agency Inspection
Inspector LYONS, DANIEL,BEEKMAN, JACOB, **Time** 23:40:13 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/09/2018 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 11/06/2018 **Type** Agency Inspection
Inspector LYONS,BEEKMAN, **Time** 19:31:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2018 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/06/2018 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 12/03/2018 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 01:54:26 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/03/2018 02:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 01/07/2019 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 00:09:37 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 646645
Expiration 02/06/2018 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/07/2019 00:14

Printed: 14 January 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 01/31/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 09:31:42 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 09216080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 05/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 01/31/2018 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 02/23/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 08:48:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 09216080A2
Expiration 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 05/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 02/23/2018 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO	Date 03/29/2018	Type Agency Inspection	
Inspector FELTOVIC,DEP,M	Time 08:37:39	Software 8100.27	

0.05 Lot # 201703A	0.08 Lot # 201611B	0.20 Lot # 201702B	0.08 Gas Lot # 09216080A2
Expiration 03/07/2019	Expiration 11/15/2018	Expiration 02/23/2019	Expiration 05/05/2018

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 03/29/2018 08:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency	SARASOTA COUNTY SO	Date	08/10/2018	Type	Agency Inspection
Inspector	PLATT,SHAYLA,D	Time	09:16:17	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 08/10/2018 12:01

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO	Date 08/10/2018	Type Department Inspection
Inspector PLATT,SHAYLA,D	Time 11:52:25	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805702
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/10/2018 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency	SARASOTA COUNTY SO	Date	08/21/2018	Type	Agency Inspection		
Inspector	FELTOVIC,DEP,M	Time	09:36:24	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201709A	0.20 Lot #	201702B	0.08 Gas Lot #	19817080A2
Expiration	03/07/2019	Expiration	09/12/2019	Expiration	02/23/2019	Expiration	09/05/2019
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 08/21/2018 09:36

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 09/30/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 07:50:20 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 19817080A2
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 09/30/2018 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency	SARASOTA COUNTY SO	Date	09/30/2018	Type	Agency Inspection		
Inspector	FELTOVIC,DEP,M	Time	09:23:14	Software	8100.27		
0.05 Lot #	201703A	0.08 Lot #	201709A	0.20 Lot #	201702B	0.08 Gas Lot #	19817080A2
Expiration	03/07/2019	Expiration	09/12/2019	Expiration	02/23/2019	Expiration	09/05/2019
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 09/30/2018 09:26

Printed: 14 January 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO	Date 12/07/2018	Type Department Inspection
Inspector PLATT,SHAYLA,D	Time 11:42:43	Software 8100.27

0.05 Lot # 201707D	0.08 Lot # 201707E	0.20 Lot # 201707C	0.08 Gas Lot # AG805701
Expiration 07/25/2019	Expiration 07/25/2019	Expiration 07/24/2019	Expiration 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 12/07/2018 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 12/28/2018 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 09:07:33 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 19817080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 12/28/2018 09:12