Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency SUMTER COUNTY SO **Date** 01/27/2019 **Type** Agency Inspection
Inspector JACOBSON,JAMES,W **Time** 14:58:10 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.074
0.08 g/210L Test # (Repeat) 1	0.076
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.075
0.08 g/210L Test # (Repeat) 2	0.076
0.20 g/210L Test # 2	0.192
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.08 g/210L Test # (Repeat) 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/27/2019 15:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency SUMTER COUNTY SO **Date** 02/26/2019 **Type** Agency Inspection
Inspector FORINASH,GARY,B **Time** 19:33:35 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/26/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2019 19:40

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency SUMTER COUNTY SO **Date** 03/30/2019 **Type** Agency Inspection
Inspector FORINASH,GARY,B **Time** 20:00:38 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/30/2019 20:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency SUMTER COUNTY SO **Date** 04/27/2019 **Type** Agency Inspection
Inspector JACOBSON,JAMES,W **Time** 11:55:30 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/30/2019 17:05

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency SUMTER COUNTY SO **Date** 05/21/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 13:58:48 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/21/2019 14:05

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency SUMTER COUNTY SO **Date** 10/31/2019 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 14:27:11 **Software** 8100.27

0.05 Lot # 201905B **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG831804
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 10/31/2019 14:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency	SUMTER COUNTY SO	**Date**	11/19/2019	**Type**	Agency Inspection
Inspector	GEOGHEGAN,ASHLEY,M	**Time**	16:00:26	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201908B	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	22419080A2
Expiration	11/13/2020	**Expiration**	08/07/2021	**Expiration**	02/20/2021	**Expiration**	10/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2019 16:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency SUMTER COUNTY SO **Date** 12/25/2019 **Type** Agency Inspection
Inspector FORINASH,GARY,B **Time** 11:08:36 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201908B **0.20 Lot #** 201902C **0.08 Gas Lot #** 22419080A2
Expiration 11/13/2020 **Expiration** 08/07/2021 **Expiration** 02/20/2021 **Expiration** 10/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/26/2019 10:35

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 01/31/2019 **Type** Agency Inspection
Inspector ROBERSON,WILLIAM,E **Time** 10:11:33 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2019 10:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 02/28/2019 **Type** Agency Inspection
Inspector ROBERSON,WILLIAM,E **Time** 09:22:54 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 03/07/2019 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2019 09:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 03/29/2019 **Type** Agency Inspection
Inspector ROBERSON,WILLIAM,E **Time** 07:10:16 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2019 07:19

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 04/30/2019 **Type** Agency Inspection
Inspector ROBERSON,WILLIAM,E **Time** 10:11:44 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A1
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/01/2019 07:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 05/20/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 12:51:42 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/20/2019 13:02

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 06/30/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 20:56:36 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range Exceeded.

Data Download Date/Time 07/06/2019 19:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 06/30/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 21:01:07 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 07/06/2019 19:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency	CITRUS COUNTY SO	**Date**	06/30/2019	**Type**	Agency Inspection
Inspector	BOWERMASTER,MARY,E	**Time**	21:15:46	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 07/06/2019 19:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 06/30/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 22:10:33 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/06/2019 19:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 07/30/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 16:08:27 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 07/31/2019 12:06

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 07/30/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 17:10:45 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** **Expiration**

Number of Simulators 3 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.022
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.050
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test # 2	0.031
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 3	0.050
0.08 g/210L Test: OK	No	0.08 g/210L Test # 3	0.032
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

08: Control Outside Tolerance. WRONGREFERENCESOLUTION.

Data Download Date/Time 07/31/2019 12:06

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 07/31/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 11:57:37 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/31/2019 12:06

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 08/25/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 19:13:11 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2019 12:55

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency	CITRUS COUNTY SO	**Date**	12/04/2019	**Type**	Agency Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	09:48:31	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BY-PASSED AI.COMPLIANCE NOT DETERMINED.

Data Download Date/Time 12/04/2019 14:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency CITRUS COUNTY SO **Date** 12/04/2019 **Type** Department Inspection
Inspector HAUGHEY,MICHAEL,D **Time** 14:29:06 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG916501
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 12/04/2019 14:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency CITRUS COUNTY SO **Date** 01/31/2019 **Type** Agency Inspection
Inspector ROBERSON,WILLIAM,E **Time** 10:08:37 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2019 10:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency CITRUS COUNTY SO **Date** 02/28/2019 **Type** Agency Inspection
Inspector ROBERSON,WILLIAM,E **Time** 09:20:45 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 03/07/2019 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.052
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2019 09:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency CITRUS COUNTY SO **Date** 03/29/2019 **Type** Agency Inspection
Inspector ROBERSON,WILLIAM,E **Time** 07:16:21 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.084
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2019 07:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency CITRUS COUNTY SO **Date** 04/30/2019 **Type** Agency Inspection
Inspector ROBERSON,WILLIAM,E **Time** 10:16:58 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/01/2019 06:47

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency CITRUS COUNTY SO **Date** 05/20/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 13:02:15 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/20/2019 13:13

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency CITRUS COUNTY SO **Date** 06/30/2019 **Type** Agency Inspection
Inspector BOWERMASTER,MARY,E **Time** 21:02:33 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 10/25/2019 11:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency	CITRUS COUNTY SO	**Date**	06/30/2019	**Type**	Agency Inspection
Inspector	BOWERMASTER,MARY,E	**Time**	22:04:38	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 3 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	INT
Minimum Sample Volume Check: OK		Alcohol Free Test # 2	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 2	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 3	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 3	INT
Interferent Detect Test: Interferent Detect	Yes		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	Yes		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 10/25/2019 11:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency	CITRUS COUNTY SO	**Date**	07/31/2019	**Type**	Agency Inspection
Inspector	PLATT,SHAYLA,D	**Time**	09:55:24	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 10/25/2019 11:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency	CITRUS COUNTY SO	**Date**	12/04/2019	**Type**	Agency Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	09:50:49	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BY-PASSED AI.COMPLIANCE NOT DETERMINED.

Data Download Date/Time 12/04/2019 14:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency CITRUS COUNTY SO **Date** 12/04/2019 **Type** Department Inspection
Inspector HAUGHEY,MICHAEL,D **Time** 14:18:18 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG916501
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 12/04/2019 14:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency CITRUS COUNTY SO **Date** 12/05/2019 **Type** Department Inspection
Inspector HAUGHEY,MICHAEL,D **Time** 14:37:37 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG916501
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 12/05/2019 14:46

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency LAKE COUNTY S.O. **Date** 01/01/2019 **Type** Agency Inspection
Inspector BINDER,PITCHER,/ **Time** 20:52:14 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 814181
Expiration 02/06/2020 **Expiration** 09/12/2919 **Expiration** 02/22/2020 **Expiration** 07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/01/2019 21:05

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency LAKE COUNTY S.O. **Date** 02/13/2019 **Type** Agency Inspection
Inspector PITCHER,BINDER,/ **Time** 20:23:29 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 814181
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2019 20:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency LAKE COUNTY S.O. **Date** 03/14/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 09:08:06 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/14/2019 09:13

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency	LAKE COUNTY S.O.	**Date**	04/10/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:27:05	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1099238
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	03/01/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.088
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.085
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.084
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance/ RE-SEATED DRY GAS BOTTLE .

Data Download Date/Time 04/10/2019 23:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency LAKE COUNTY S.O. **Date** 05/02/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 11:24:42 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/02/2019 11:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency LAKE COUNTY S.O. **Date** 05/16/2019 **Type** Agency Inspection
Inspector PITCHER,BINDER,/ **Time** 00:21:39 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 970670
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.000
0.20 g/210L Test # (Repeat) 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.193
0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.20 g/210L Test # (Repeat) 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance/CONNECTED HOSE FROM SIMULATOR.

Data Download Date/Time 05/16/2019 16:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency	LAKE COUNTY S.O.	**Date**	06/19/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	23:13:10	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.20 g/210L Test # (Repeat) 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance/CHECKED TEMP/RECONNECTED SIM.

Data Download Date/Time 06/20/2019 00:10

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency LAKE COUNTY S.O. **Date** 07/03/2019 **Type** Agency Inspection
Inspector PITCHER,BINDER,/ **Time** 20:55:54 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.175
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.181
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.20 g/210L Test: OK	No	0.20 g/210L Test # 2	0.178
0.08 g/210L Dry Gas Std Test: OK	No	0.20 g/210L Test # (Repeat) 2	0.183
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.182
		0.20 g/210L Test # (Repeat) 3	0.183
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance/CHECKED TEMP. AND CONNECTIONS. 20: Control Outside Tolerance/CHECK TEMP. AND CONNECTIONS, . Non-compliance:0.20 OUT OF COMPLIANCE.

Data Download Date/Time 07/03/2019 23:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency	LAKE COUNTY S.O.	**Date**	07/03/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	21:46:27	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.180
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.179
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.178
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.180
0.20 g/210L Test: OK	No	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.178
		0.20 g/210L Test # (Repeat) 3	0.180
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance;REPLACED SIM CONNECTORS, Cont. Non-compliance:.20 CONTROL TEST OUT OF TOLERANCE
.

Data Download Date/Time 07/03/2019 23:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency	LAKE COUNTY S.O.	**Date**	07/24/2019	**Type**	Department Inspection
Inspector	GRAHAM,THOMAS,J	**Time**	10:36:31	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

Data Download Date/Time 07/24/2019 10:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency LAKE COUNTY S.O. **Date** 08/23/2019 **Type** Agency Inspection
Inspector BINDER,PITCHER,/ **Time** 15:16:41 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 970670
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 08/23/2019 17:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency	LAKE COUNTY S.O.	**Date**	09/25/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	21:59:41	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:REMIXED SOLUTION,REPLACED SIMULATOR. Non-compliance:BAD SIMULATOR JAR, REPLACED.

Data Download Date/Time 09/26/2019 01:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency	LAKE COUNTY S.O.	**Date**	09/25/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:08:51	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/26/2019 01:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency	LAKE COUNTY S.O.	**Date**	10/23/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:10:41	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	970670
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2019 23:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency LAKE COUNTY S.O. **Date** 11/09/2019 **Type** Agency Inspection
Inspector BINDER,PITCHER,/ **Time** 01:19:12 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 970670
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/09/2019 01:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

Agency	LAKE COUNTY S.O.	**Date**	12/05/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	18:26:20	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	970670
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2019 19:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency LAKE COUNTY SO **Date** 01/01/2019 **Type** Agency Inspection
Inspector BINDER,PITCHER,/ **Time** 21:02:31 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 970670
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/01/2019 21:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	02/13/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	20:05:40	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2019 20:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	03/06/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	20:20:35	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/06/2019 20:25

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	04/10/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:40:53	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/10/2019 23:47

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	05/16/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	00:21:54	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/16/2019 16:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	06/19/2019	**Type**	Agency Inspection		
Inspector	PITCHER,BINDER,/	**Time**	23:50:26	**Software**	8100.27		
0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2019 00:02

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	07/03/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	22:02:43	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/03/2019 22:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency LAKE COUNTY SO **Date** 08/23/2019 **Type** Agency Inspection
Inspector BINDER,PITCHER,/ **Time** 15:24:33 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 1099238
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/01/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/23/2019 15:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	09/19/2019	**Type**	Department Inspection
Inspector	PLATT,SHAYLA,D	**Time**	14:25:43	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG916501
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.076
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.076
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.076
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 09/19/2019 14:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	09/25/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:24:16	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1099238
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	03/01/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/26/2019 01:48

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	10/23/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:09:10	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1099238
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/01/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2019 23:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	11/09/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	01:30:34	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1099238
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/01/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/09/2019 01:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency	LAKE COUNTY SO	**Date**	12/05/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	18:23:42	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1099238
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/01/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2019 19:10

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 01/20/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 00:25:00 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.044
0.05 g/210L Test # (Repeat) 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.044
0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.045
0.05 g/210L Test # (Repeat) 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.REPLACE ORING,RETIGHTEN JAR

Data Download Date/Time 01/20/2019 00:35

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency	LEESBURG P.D.	**Date**	02/21/2019	**Type**	Agency Inspection
Inspector	SOMMERSDORF,JOHN,G	**Time**	23:06:52	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:LOOSE SIMULATOR JAR.

Data Download Date/Time 02/21/2019 23:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 03/16/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 12:05:42 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/16/2019 12:17

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 04/13/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 10:32:06 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/13/2019 10:46

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 05/15/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 15:00:31 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/15/2019 15:17

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 06/06/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 18:40:20 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/07/2019 07:16

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 07/26/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 16:16:26 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2019 16:31

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 08/13/2019 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 14:04:03 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG916501
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/3021 **Expiration** 06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 08/13/2019 14:26

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 08/21/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 23:43:06 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/21/2019 23:59

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 09/16/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 01:28:08 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/16/2019 01:40

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 10/17/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 04:18:16 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/17/2019 04:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency LEESBURG P.D. **Date** 11/23/2019 **Type** Agency Inspection
Inspector SOMMERSDORF,JOHN,G **Time** 08:57:07 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/23/2019 09:13

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency	LEESBURG P.D.	**Date**	12/25/2019	**Type**	Agency Inspection
Inspector	SOMMERSDORF,JOHN,G	**Time**	11:00:53	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	34917080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.073
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.074
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance CHECKED ORINGS TIGHTENED JAR . CHECKED ORINGS,TIGHTENED JAR

Data Download Date/Time 12/30/2019 15:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency	MOUNT DORA P.D.	**Date**	01/04/2019	**Type**	Agency Inspection
Inspector	DANCEL,KAREN,J	**Time**	00:29:12	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	583457
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	10/20/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/04/2019 00:52

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency MOUNT DORA P.D. **Date** 01/15/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 14:08:21 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 01/15/2019 14:12

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency	MOUNT DORA P.D.	**Date**	02/28/2019	**Type**	Agency Inspection
Inspector	DANCEL,KAREN,J	**Time**	23:11:53	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	583457
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	10/20/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.074
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.075
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.074
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.075
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.074
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 02/28/2019 23:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency MOUNT DORA P.D. **Date** 04/26/2019 **Type** Agency Inspection
Inspector DANCEL,KAREN,J **Time** 04:09:34 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 1099238
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 03/01/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/17/2019 11:04

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency MOUNT DORA P.D. **Date** 05/14/2019 **Type** Agency Inspection
Inspector DANCEL,KAREN,J **Time** 11:17:06 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.180
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.180
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.182
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.180
0.20 g/210L Test: OK	No	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.183
		0.20 g/210L Test # (Repeat) 3	0.180
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:INSTRUMENT BEING SENT TO FDLE FOR INSPECTION.

Data Download Date/Time 05/17/2019 11:04

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency MOUNT DORA P.D. **Date** 05/17/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 10:53:34 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/17/2019 11:04

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency	MOUNT DORA P.D.	**Date**	06/01/2019	**Type**	Agency Inspection
Inspector	DANCEL,KAREN,J	**Time**	12:42:43	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1099238
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	03/01/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/14/2019 14:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency	MOUNT DORA P.D.	**Date**	08/14/2019	**Type**	Agency Inspection
Inspector	MURPHY,PATRICK,J	**Time**	14:10:56	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

SKIPPED AI TO OPERATE INSTRUMENT

Data Download Date/Time 08/14/2019 14:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Agency MOUNT DORA P.D. **Date** 08/15/2019 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 11:59:49 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG916501
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/15/2019 12:07

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency	MARION COUNTY SO	**Date**	01/14/2019	**Type**	Agency Inspection
Inspector	RAKER,JEFFERY,A	**Time**	09:50:13	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	875088
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	12/12/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 01/14/2019 09:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 03/01/2019 **Type** Department Inspection
Inspector KIRKLAND,BRETT,H **Time** 12:21:03 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 03/01/2019 12:24

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency	MARION COUNTY SO	**Date**	03/26/2019	**Type**	Agency Inspection
Inspector	RAKER,JEFFERY,A	**Time**	23:23:42	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	875088
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	12/12/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE

Data Download Date/Time 03/26/2019 23:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 04/19/2019 **Type** Agency Inspection
Inspector RATH,ROBERT,S **Time** 17:05:11 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 875088
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/12/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE

Data Download Date/Time 04/19/2019 17:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 05/22/2019 **Type** Agency Inspection
Inspector RATH,ROBERT,S **Time** 16:16:09 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 875088
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/12/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE

Data Download Date/Time 05/22/2019 16:26

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 06/14/2019 **Type** Agency Inspection
Inspector RATH,ROBERT,S **Time** 10:12:54 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 1031121
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/12/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE.

Data Download Date/Time 06/14/2019 10:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 07/22/2019 **Type** Agency Inspection
Inspector RATH,ROBERT,S **Time** 14:12:42 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 875088
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/12/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE.

Data Download Date/Time 07/22/2019 14:19

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 08/15/2019 **Type** Agency Inspection
Inspector RATH,ROBERT,S **Time** 10:16:46 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 181692
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 03/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE.

Data Download Date/Time 08/15/2019 10:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency	MARION COUNTY SO	**Date**	09/16/2019	**Type**	Agency Inspection
Inspector	RATH,ROBERT,S	**Time**	17:27:00	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	181692
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	03/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE.

Data Download Date/Time 09/16/2019 18:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 10/03/2019 **Type** Agency Inspection
Inspector FORINASH,GARY,B **Time** 23:29:27 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE

Data Download Date/Time 10/04/2019 14:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 10/26/2019 **Type** Agency Inspection
Inspector FORINASH,GARY,B **Time** 09:52:31 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

WITHIN COMPLIANCE

Data Download Date/Time 10/30/2019 10:04

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency MARION COUNTY SO **Date** 11/19/2019 **Type** Agency Inspection
Inspector GEOGHEGAN,ASHLEY,M **Time** 13:58:42 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201908B **0.20 Lot #** 201902C **0.08 Gas Lot #** 13518080A5
Expiration 11/13/2020 **Expiration** 08/07/2021 **Expiration** 02/20/2021 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/19/2019 14:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency	MARION COUNTY SO	**Date**	12/25/2019	**Type**	Agency Inspection
Inspector	FORINASH,GARY,B	**Time**	11:05:32	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201908B	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	13518080A5
Expiration	11/13/2020	**Expiration**	08/07/2021	**Expiration**	02/20/2021	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/26/2019 10:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency HILLSBOROUGH CO SO **Date** 01/10/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 19:31:27 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 31517080A1
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 01/10/2019 19:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency	HILLSBOROUGH CO SO	**Date**	02/04/2019	**Type**	Agency Inspection
Inspector	SKIPPER,ROGER,G	**Time**	17:25:32	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	31517080A1
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 02/04/2019 17:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency HILLSBOROUGH CO SO **Date** 02/27/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 09:01:27 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/27/2019 09:07

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency	HILLSBOROUGH CO SO	**Date**	03/13/2019	**Type**	Agency Inspection
Inspector	SKIPPER,ROGER,G	**Time**	16:30:48	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	31517080A1
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 03/13/2019 16:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency	HILLSBOROUGH CO SO	**Date**	04/04/2019	**Type**	Agency Inspection
Inspector	SKIPPER,ROGER,G	**Time**	17:02:52	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	31517080A1
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 04/04/2019 17:10

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency	HILLSBOROUGH CO SO	**Date**	05/02/2019	**Type**	Agency Inspection
Inspector	SKIPPER,ROGER,G	**Time**	18:01:57	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	31517080A1
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	01/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 05/06/2019 13:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency	HILLSBOROUGH CO SO	**Date**	06/05/2019	**Type**	Agency Inspection
Inspector	SKIPPER,ROGER,G	**Time**	18:18:48	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	28818080A1
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 06/11/2019 10:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency HILLSBOROUGH CO SO **Date** 07/08/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 18:15:46 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 28818080A1
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 07/10/2019 14:10

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency HILLSBOROUGH CO SO **Date** 08/05/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 16:23:24 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 28818080A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 08/06/2019 14:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency HILLSBOROUGH CO SO **Date** 09/11/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 16:06:32 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 2881808A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 09/11/2019 16:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency HILLSBOROUGH CO SO **Date** 10/09/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 18:14:18 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 28818080A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 10/09/2019 19:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency	HILLSBOROUGH CO SO	**Date**	11/06/2019	**Type**	Agency Inspection
Inspector	SKIPPER,ROGER,G	**Time**	16:03:41	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	28818080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 11/06/2019 16:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency	HILLSBOROUGH CO SO	**Date**	12/04/2019	**Type**	Agency Inspection
Inspector	SKIPPER,ROGER,G	**Time**	18:49:41	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	28818080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 12/04/2019 19:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency MARTIN COUNTY SO **Date** 01/13/2019 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 22:53:19 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 06717080A7
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/13/2019 23:03

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	02/09/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	19:49:37	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06717080A7
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/09/2019 19:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	02/20/2019	**Type**	Department Inspection
Inspector	PLATT,SHAYLA,D	**Time**	10:58:24	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805701
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/20/2019 11:07

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency MARTIN COUNTY SO **Date** 03/28/2019 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 23:31:03 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 06717080A7
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2019 23:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	04/26/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	00:47:20	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06717080A7
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 04/26/2019 01:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	05/27/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	22:51:48	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/27/2019 23:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	06/25/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	00:40:07	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/25/2019 00:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	07/30/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	17:53:38	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/30/2019 21:45

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	08/25/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	21:15:31	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/25/2019 21:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	09/09/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	00:36:04	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/09/2019 01:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	10/04/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	17:11:55	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/04/2019 17:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency MARTIN COUNTY SO **Date** 11/04/2019 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 00:04:00 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 05218080A4
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/04/2019 02:03

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency	MARTIN COUNTY SO	**Date**	12/16/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	00:14:44	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/16/2019 00:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	01/14/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	00:19:42	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2019 00:25

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	02/09/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	21:33:28	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/09/2019 21:45

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	02/20/2019	**Type**	Department Inspection
Inspector	PLATT,SHAYLA,D	**Time**	10:44:52	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805701
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/20/2019 10:54

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	03/29/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	00:52:22	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2019 01:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	04/26/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	00:58:21	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/26/2019 01:10

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	05/28/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	00:26:07	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/28/2019 00:40

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	06/24/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	23:08:52	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/24/2019 23:16

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	07/30/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	17:37:49	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/30/2019 17:47

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	08/25/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	21:05:45	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/25/2019 22:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	09/08/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	21:56:28	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/08/2019 22:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	10/04/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	17:09:49	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/04/2019 17:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	11/03/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	21:30:55	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/03/2019 22:40

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency	MARTIN COUNTY SO	**Date**	12/15/2019	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	22:47:57	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/15/2019 23:06

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency HILLSBOROUGH CO SO **Date** 03/20/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 08:44:27 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/20/2019 08:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency HILLSBOROUGH CO SO **Date** 04/04/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 15:26:02 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 28818080A1
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 04/04/2019 15:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency HILLSBOROUGH CO SO **Date** 05/02/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 18:16:11 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 28818080A1
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 05/06/2019 14:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency HILLSBOROUGH CO SO **Date** 06/05/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 16:20:56 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 28818080A1
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 06/05/2019 16:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency HILLSBOROUGH CO SO **Date** 07/08/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 16:26:21 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 28818080A1
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 07/10/2019 09:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency HILLSBOROUGH CO SO **Date** 08/05/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 18:08:47 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 28818080A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 08/06/2019 14:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency HILLSBOROUGH CO SO **Date** 09/11/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 16:17:44 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 28818080A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 09/11/2019 16:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency	HILLSBOROUGH CO SO	**Date**	10/09/2019	**Type**	Agency Inspection
Inspector	SKIPPER,ROGER,G	**Time**	18:18:37	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	28818080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT ENTIRE INSPECTION.

Data Download Date/Time 10/10/2019 14:35

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency HILLSBOROUGH CO SO **Date** 11/06/2019 **Type** Agency Inspection
Inspector SKIPPER,ROGER,G **Time** 16:09:30 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 28818080A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: UNK. AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 11/06/2019 17:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	01/22/2019	**Type**	Agency Inspection
Inspector	BELL,DANIELLE,M	**Time**	11:00:34	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

Data Download Date/Time 01/22/2018 13:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency UMATILLA PD(LAKE SO) **Date** 01/22/2019 **Type** Department Inspection
Inspector BELL,DANIELLE,M **Time** 13:25:21 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG708807
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 03/29/2019

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.083
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.084
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.083
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.083
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/22/2018 13:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	01/01/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	20:50:32	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34416080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	02/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/01/2019 20:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	02/13/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	20:11:11	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2019 20:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	03/06/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	20:08:30	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/06/2019 20:16

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency UMATILLA PD(LAKE SO) **Date** 04/10/2019 **Type** Agency Inspection
Inspector BINDER,PITCHER,/ **Time** 23:25:29 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 970670
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/10/2019 23:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency UMATILLA PD(LAKE SO) **Date** 05/16/2019 **Type** Agency Inspection
Inspector PITCHER,BINDER,/ **Time** 00:19:36 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 970670
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det: RFI Detect/RADIO TRANSMISSION IN ROOM.

Data Download Date/Time 05/16/2019 00:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	06/19/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	23:26:01	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/19/2019 23:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	07/03/2019	**Type**	Agency Inspection
Inspector	PITCHER,BINDER,/	**Time**	21:04:28	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/03/2019 21:26

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	08/23/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	15:14:47	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 08/23/2019 16:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	09/25/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:11:27	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/25/2019 23:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency UMATILLA PD(LAKE SO) **Date** 10/10/2019 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 13:43:28 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG916501
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 10/10/2019 14:30

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	10/23/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	23:12:04	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	814181
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2019 23:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	11/09/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	01:24:18	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	970670
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/09/2019 01:59

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

Agency	UMATILLA PD(LAKE SO)	**Date**	12/05/2019	**Type**	Agency Inspection
Inspector	BINDER,PITCHER,/	**Time**	18:13:08	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	970670
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2019 19:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency PANAMA CITY BEACH PD **Date** 01/04/2019 **Type** Agency Inspection
Inspector CHERI,DIANE,M **Time** 08:43:21 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 906364
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:TOO MUCH MA.

Data Download Date/Time 01/04/2019 10:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency PANAMA CITY BEACH PD **Date** 02/12/2019 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 10:31:43 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/12/2019 11:42

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency	PANAMA CITY BEACH PD	**Date**	03/03/2019	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	18:41:27	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 03/03/2019 20:52

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency PANAMA CITY BEACH PD **Date** 04/03/2019 **Type** Agency Inspection
Inspector CHERI,DIANE,M **Time** 19:29:05 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 989917
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 04/03/2019 21:47

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency	PANAMA CITY BEACH PD	**Date**	05/07/2019	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	07:26:27	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

TIME WAS 10 MINS SLOW

Data Download Date/Time 05/07/2019 09:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency	PANAMA CITY BEACH PD	**Date**	06/03/2019	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	08:10:48	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

CHANGED SIM HEAD

Data Download Date/Time 06/03/2019 10:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency	PANAMA CITY BEACH PD	**Date**	07/03/2019	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	16:56:41	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 07/03/2019 19:24

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency	PANAMA CITY BEACH PD	**Date**	08/04/2019	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	18:53:25	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 08/04/2019 21:01

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency	PANAMA CITY BEACH PD	**Date**	09/01/2019	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	16:51:05	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 09/04/2019 10:26

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency	PANAMA CITY BEACH PD	**Date**	10/01/2019	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	12:22:11	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 10/01/2019 14:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency PANAMA CITY BEACH PD **Date** 11/03/2019 **Type** Agency Inspection
Inspector CHERI,DIANE,M **Time** 14:24:00 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 989917
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 11/03/2019 19:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency	PANAMA CITY BEACH PD	**Date**	12/02/2019	**Type**	Agency Inspection
Inspector	CHERI,DIANE,M	**Time**	07:27:48	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	989917
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 12/02/2019 10:50

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency LYNN HAVEN PD **Date** 02/13/2019 **Type** Agency Inspection
Inspector MAYFIELD,MATTHEW,E **Time** 14:08:37 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 02/14/2019 13:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency LYNN HAVEN PD **Date** 02/13/2019 **Type** Agency Inspection
Inspector MAYFIELD,MATTHEW,E **Time** 15:09:58 **Software** 8100.27

0.05 Lot # 20180B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 05218080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 02/14/2019 13:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency LYNN HAVEN PD **Date** 02/22/2019 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 11:11:41 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/22/2019 12:16

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency	LYNN HAVEN PD	**Date**	03/18/2019	**Type**	Agency Inspection
Inspector	BARGE,SUSAN,M	**Time**	15:46:28	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	05218080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 03/18/2019 16:59

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency LYNN HAVEN PD **Date** 04/10/2019 **Type** Agency Inspection
Inspector BARGE,SUSAN,M **Time** 10:48:03 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 05218080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 05/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 04/10/2019 11:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency	CAPE CORAL PD	**Date**	06/04/2019	**Type**	Department Inspection
Inspector	GRAHAM,THOMAS,J	**Time**	10:40:19	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG831804
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 06/04/2019 10:46

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency	CAPE CORAL PD	**Date**	06/26/2019	**Type**	Agency Inspection
Inspector	CLAFLIN,DAVID,S	**Time**	11:51:28	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	28818080A1
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.034
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.051
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.034
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.051
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.034
		0.05 g/210L Test # (Repeat) 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance LOOSE CONECTION.

Data Download Date/Time 06/26/2019 11:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency	CAPE CORAL PD	**Date**	07/19/2019	**Type**	Agency Inspection
Inspector	CLAFLIN,DAVID,S	**Time**	09:52:27	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	17618080A1
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2019 10:01

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency	CAPE CORAL PD	**Date**	12/04/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	12:09:52	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG916501
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	06/14/2021

Number of Simulators 5 **Average Standard Deviation** **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.082
0.08 g/210L Dry Gas Std Test # (Repeat) 1	INT
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	INT
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.203
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.202
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.052
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.202
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.203
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0007
0.20 g/210L Test (Std Dev)	0.0005

Remarks:

08: Interferent Detect, Interferent Detect AIR OUT, CAL ADJ.

Data Download Date/Time 12/04/2019 12:17

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency	CAPE CORAL PD	**Date**	12/05/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	14:27:56	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG916501
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 12/05/2019 14:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency	CAPE CORAL PD	**Date**	12/13/2019	**Type**	Agency Inspection
Inspector	BARKLEY,JEFFREY,G	**Time**	07:08:44	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201908B	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	03519080A5
Expiration	11/13/2020	**Expiration**	08/07/2021	**Expiration**	02/20/2021	**Expiration**	04/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/13/2019 07:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	01/30/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	07:49:56	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 01/30/2019 08:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	02/27/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	09:29:01	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 02/27/2019 09:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	03/27/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	08:10:12	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 03/27/2019 08:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	04/16/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	08:05:04	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 04/16/2019 08:13

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency SARASOTA COUNTY SO **Date** 04/19/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 13:10:41 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 04/19/2019 13:19

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	05/28/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	08:44:40	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 05/28/2019 10:16

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	06/26/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	08:18:05	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 06/26/2019 08:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	07/31/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	15:48:46	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

OUTSIDE RADIO TRAFFIC RFI.

Data Download Date/Time 07/31/2019 16:04

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	08/19/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	16:25:03	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	19817080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT NORTH JAIL

Data Download Date/Time 08/19/2019 16:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	09/25/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	07:33:55	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	17919080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 09/25/2019 07:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	10/15/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	10:20:28	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	17919080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 10/15/2019 10:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	11/26/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	09:00:04	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	17919080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	08/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 11/26/2019 09:12

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

Agency	SARASOTA COUNTY SO	**Date**	12/18/2019	**Type**	Agency Inspection
Inspector	FELTOVIC,DEP,M	**Time**	10:56:41	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	17919080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	08/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 12/18/2019 11:05

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 02/14/2019 **Type** Agency Inspection
Inspector BARGE,SUSAN,M **Time** 16:10:27 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 770784
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 02/14/2019 17:24

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 02/19/2019 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 12:17:17 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.075
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.076
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.076
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.076
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.076
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/19/2019 13:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 02/22/2019 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 11:29:53 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/22/2019 12:35

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency	BLOUNTSTOWN PD	**Date**	03/18/2019	**Type**	Agency Inspection
Inspector	BARGE,SUSAN,M	**Time**	15:45:54	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	770784
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 03/18/2019 16:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 04/10/2019 **Type** Agency Inspection
Inspector BARGE,SUSAN,M **Time** 11:10:22 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 770784
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 04/10/2019 12:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 05/05/2019 **Type** Agency Inspection
Inspector BARGE,SUSAN,M **Time** 14:50:09 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 770784
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 05/05/2019 15:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency	BLOUNTSTOWN PD	**Date**	06/03/2019	**Type**	Agency Inspection
Inspector	BARGE,SUSAN,M	**Time**	13:52:36	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	770784
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 06/03/2019 15:19

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 07/26/2019 **Type** Agency Inspection
Inspector RUBEL,NEIL,F **Time** 02:22:10 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 770784
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 07/26/2019 03:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 08/05/2019 **Type** Agency Inspection
Inspector RUBEL,NEIL,F **Time** 13:39:17 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 770784
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/05/2019 15:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 09/04/2019 **Type** Agency Inspection
Inspector RUBEL,NEIL,F **Time** 00:19:05 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 770784
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/04/2019 01:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency BLOUNTSTOWN PD **Date** 10/01/2019 **Type** Agency Inspection
Inspector RUBEL,NEIL,F **Time** 02:09:59 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 770784
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/01/2019 03:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency	BLOUNTSTOWN PD	**Date**	11/08/2019	**Type**	Agency Inspection
Inspector	BARGE,SUSAN,M	**Time**	15:18:24	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	770784
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 11/08/2019 16:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency	BLOUNTSTOWN PD	**Date**	12/11/2019	**Type**	Agency Inspection
Inspector	BARGE,SUSAN,M	**Time**	14:59:22	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	770784
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	04/27/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 12/11/2019 16:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency MONROE COUNTY SO **Date** 01/02/2019 **Type** Agency Inspection
Inspector KEITH,JASON,T **Time** 11:00:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 823252
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE. PRE-EVIDENTIARY AL COMPLETED.

Data Download Date/Time 01/02/2019 11:08

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency MONROE COUNTY SO **Date** 01/10/2019 **Type** Agency Inspection
Inspector KEITH,JASON,T **Time** 13:28:35 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 915310
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 03/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.058
0.05 g/210L Test # (Repeat) 1	0.053
0.08 g/210L Test # 1	0.083
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.056
0.05 g/210L Test # (Repeat) 2	0.053
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.055
0.05 g/210L Test # (Repeat) 3	0.052
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 01/10/2019 13:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency MONROE COUNTY SO **Date** 01/17/2019 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 10:00:27 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.206
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.206
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.205
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.206
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.205
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.205
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.206
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.205
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/17/2019 10:05

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency MONROE COUNTY SO **Date** 01/23/2019 **Type** Agency Inspection
Inspector KEITH,JASON,T **Time** 17:00:52 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 823252
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE. POST-FDLE DI AI COMPLETED.

Data Download Date/Time 01/23/2019 17:06

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	02/06/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	07:55:03	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 02/06/2019 07:59

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	03/21/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	11:56:49	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	823252
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX. PSI 241

Data Download Date/Time 03/21/2019 12:02

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency MONROE COUNTY SO **Date** 04/04/2019 **Type** Agency Inspection
Inspector KEITH,JASON,T **Time** 10:27:37 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 823252
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 04/04/2019 10:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	05/01/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	13:31:19	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	915310
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	03/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 05/01/2019 13:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	06/20/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	11:29:27	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 06/20/2019 11:35

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	07/11/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	13:14:18	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	823252
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 07/11/2019 13:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	08/07/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	15:38:08	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1012059
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 08/07/2019 15:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	09/18/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	10:42:37	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1012059
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 09/18/2019 10:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	10/02/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	13:35:47	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1012059
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 10/02/2019 13:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	11/20/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	11:16:21	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 11/20/2019 11:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency	MONROE COUNTY SO	**Date**	12/05/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:29:08	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1012059
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 12/05/2019 14:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	01/02/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	17:05:46	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE. PRE-FDLE DI AI COMPLETED.

Data Download Date/Time 01/02/2019 17:13

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency MONROE COUNTY SO **Date** 01/17/2019 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 10:23:07 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.083
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.083
0.20 g/210L Test # 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.083
0.20 g/210L Test # 4	0.204
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.083
0.20 g/210L Test # 5	0.204
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.204
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.083
0.20 g/210L Test # 7	0.204
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.083
0.20 g/210L Test # 8	0.204
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/17/2019 10:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	01/23/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	08:15:25	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

POST-FDLE DI AI COMPLETED.

Data Download Date/Time 01/23/2019 08:19

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	02/20/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	14:57:03	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.209
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.209
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.209
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 02/20/2019 15:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	03/21/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	08:43:07	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	823252
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SARE INTOX. PSI 244

Data Download Date/Time 03/21/2019 08:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency MONROE COUNTY SO **Date** 04/04/2019 **Type** Agency Inspection
Inspector KEITH,JASON,T **Time** 07:36:47 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 823252
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 04/04/2019 07:50

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency MONROE COUNTY SO **Date** 05/02/2019 **Type** Agency Inspection
Inspector KEITH,JASON,T **Time** 09:34:56 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 823252
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 05/02/2019 09:57

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	06/06/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	13:46:43	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1012059
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 06/07/2019 10:06

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	07/10/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	16:16:04	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	823252
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	10/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 07/10/2019 16:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	08/07/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	12:13:48	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1012059
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	09/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.209
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.209
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.208
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 08/07/2019 12:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	09/11/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	12:24:39	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1060911
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/11/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.208
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.209
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 09/12/2019 15:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	10/02/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	10:28:42	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1060911
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/11/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 10/02/2019 10:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	11/20/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	13:58:15	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201908B	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1060911
Expiration	11/13/2020	**Expiration**	08/07/2021	**Expiration**	02/20/2021	**Expiration**	12/11/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 11/20/2019 14:07

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency	MONROE COUNTY SO	**Date**	12/05/2019	**Type**	Agency Inspection
Inspector	KEITH,JASON,T	**Time**	11:32:07	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201908B	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1060911
Expiration	11/13/2020	**Expiration**	08/07/2021	**Expiration**	02/20/2021	**Expiration**	12/11/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SPARE INTOX

Data Download Date/Time 12/05/2019 11:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency CORAL GABLES PD **Date** 01/29/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 10:58:08 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 01/29/2019 11:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	02/27/2019	**Type**	Agency Inspection
Inspector	PASTOR,JOSE,F	**Time**	12:11:18	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	620051
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

20: RFI Detect-OFC. PASTOR SMART WATCH WENT OFF .

Data Download Date/Time 02/27/2019 13:12

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency CORAL GABLES PD **Date** 03/27/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 10:13:59 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2019 12:06

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency CORAL GABLES PD **Date** 04/30/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 14:41:12 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/30/2019 14:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	05/31/2019	**Type**	Agency Inspection
Inspector	SMITH,DANIEL,A	**Time**	08:59:15	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201709A	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	620051
Expiration	11/13/2020	**Expiration**	09/12/2019	**Expiration**	02/20/2021	**Expiration**	11/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2019 09:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	06/28/2019	**Type**	Agency Inspection
Inspector	SMITH,DANIEL,A	**Time**	11:17:45	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201709A	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	620051
Expiration	11/13/2020	**Expiration**	09/12/2019	**Expiration**	02/20/2021	**Expiration**	11/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 06/28/2019 11:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	07/29/2019	**Type**	Agency Inspection
Inspector	SMITH,DANIEL,A	**Time**	10:37:12	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201709A	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	620051
Expiration	11/13/2020	**Expiration**	09/12/2019	**Expiration**	02/20/2021	**Expiration**	11/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/29/2019 10:54

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	08/28/2019	**Type**	Agency Inspection
Inspector	SMITH,DANIEL,A	**Time**	11:36:17	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201709A	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	620051
Expiration	11/13/2020	**Expiration**	09/12/2019	**Expiration**	02/20/2021	**Expiration**	11/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/05/2019 10:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	09/25/2019	**Type**	Agency Inspection
Inspector	SMITH,DANIEL,A	**Time**	07:27:56	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	620051
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	11/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 09/25/2019 07:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	10/22/2019	**Type**	Agency Inspection
Inspector	SMITH,DANIEL,A	**Time**	12:07:31	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	620051
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	11/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/24/2019 11:47

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	10/24/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	11:34:01	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 10/24/2019 11:47

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	11/19/2019	**Type**	Agency Inspection
Inspector	SMITH,DANIEL,A	**Time**	11:29:19	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	32518080AI
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	02/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	INT
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Interferent Detect.

Data Download Date/Time 11/19/2019 11:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

Agency	CORAL GABLES PD	**Date**	12/26/2019	**Type**	Agency Inspection
Inspector	SMITH,DANIEL,A	**Time**	11:22:40	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	32518080AI
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	02/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/26/2019 11:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD **Date** 01/29/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 11:40:25 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 640976
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2019 12:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency	MIAMI PD	**Date**	02/27/2019	**Type**	Agency Inspection
Inspector	PASTOR,JOSE,F	**Time**	13:21:15	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	640976
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2019 15:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD **Date** 03/27/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 10:49:51 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 640976
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/10/2019 07:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD **Date** 04/29/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE, **Time** 13:30:44 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 640976
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

OFC. FERNANDEZ IBM 28394

Data Download Date/Time 05/02/2019 09:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD **Date** 05/31/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 11:51:48 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 640976
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2019 13:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency	MIAMI PD	**Date**	06/27/2019	**Type**	Agency Inspection
Inspector	FERNANDEZ,ANGEL,	**Time**	14:27:50	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	640976
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2019 15:53

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD **Date** 07/26/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 14:04:16 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 640976
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2019 14:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD **Date** 08/27/2019 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 15:29:00 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 640976
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/03/2019 12:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD **Date** 09/30/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 09:42:11 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.072
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.073
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.075
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.074
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.075
		Interferent Detect Test # 3	INT

Remarks:

MALFUNCTIONING SIMULATOR SN#G13061 REPLACED, Control Outsid. Non-compliance: .

Data Download Date/Time 10/03/2019 12:03

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD **Date** 09/30/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 10:57:43 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201808E **0.08 Gas Lot #** 640976
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 08/22/2020 **Expiration** 11/21/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/03/2019 12:03

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency	MIAMI PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	PASTOR,JOSE,F	**Time**	10:56:05	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

BROKEN BEAKER.

Data Download Date/Time 11/04/2019 07:19

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency	MIAMI PD	**Date**	11/18/2019	**Type**	Agency Inspection
Inspector	FERNANDEZ,ANGEL,	**Time**	13:09:30	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/18/2019 14:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency	MIAMI PD	**Date**	12/10/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	11:54:02	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG916501
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.076
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 12/10/2019 12:08

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency	MIAMI PD	**Date**	12/23/2019	**Type**	Agency Inspection
Inspector	FERNANDEZ,ANGEL,	**Time**	18:32:32	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/23/2019 19:51

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 01/29/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 11:49:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2019 12:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 02/27/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 14:58:05 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2019 15:52

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 03/27/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 10:46:49 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.

Data Download Date/Time 03/28/2019 20:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 04/29/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE, **Time** 13:39:27 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

OFC FERNANDEZ IBM 28394 PRESENT

Data Download Date/Time 05/02/2019 09:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 05/31/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 11:27:45 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 05/31/2019 12:10

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 06/27/2019 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 14:09:07 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2019 16:16

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 07/26/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 14:19:39 **Software** 8100.27

0.05 Lot # 2018B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MOUTH ALCOHOL NOT ENOUHG TO RINSE.

Data Download Date/Time 07/26/2019 14:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 08/27/2019 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 15:48:21 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2019 14:59

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency	MIAMI PD	**Date**	09/30/2019	**Type**	Agency Inspection
Inspector	PASTOR,JOSE,F	**Time**	11:19:33	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	620051
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/03/2019 13:04

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency	MIAMI PD	**Date**	10/10/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	12:02:32	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.000
0.08 g/210L Test # (Repeat) 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.000
0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.000
0.08 g/210L Test # (Repeat) 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # (Repeat) 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.000
0.08 g/210L Test # (Repeat) 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # (Repeat) 5	0.000
0.05 g/210L Test # 5	0.049

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

0.08 g/210L Test # 5	0.000
0.08 g/210L Test # (Repeat) 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT
Alcohol Free Test # (Repeat) 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.000
0.08 g/210L Test # (Repeat) 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # (Repeat) 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.000
0.08 g/210L Test # (Repeat) 7	0.078
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # (Repeat) 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.000
0.08 g/210L Test # (Repeat) 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # (Repeat) 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.000
0.08 g/210L Test # (Repeat) 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # (Repeat) 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.000
0.08 g/210L Test # (Repeat) 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev) (Repeat)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

00: Ambient Fail MA STILL IN AIR; VENTED 08: Control SIM Outside Tolerance SIM LOST POWER.

Data Download Date/Time 10/10/2019 13:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency	MIAMI PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	PASTOR,JOSE,F	**Time**	11:26:24	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/04/2019 07:42

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 11/18/2019 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 13:50:17 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201808E **0.08 Gas Lot #** 989917
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 08/22/2020 **Expiration** 07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

.20 HAS WRONG LOT NUMBER WRITTEN IN REPORT

Data Download Date/Time 11/18/2019 14:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency MIAMI PD **Date** 12/23/2019 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 18:54:25 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 989917
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/23/2019 19:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 01/29/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 11:16:48 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.049
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.079
Alcohol Free Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	Yes	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	Yes	0.05 g/210L Test # 3	0.049
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2019 12:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 01/29/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 12:08:17 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2019 12:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 02/27/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 17:48:57 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2019 17:55

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 03/27/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 11:01:44 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/31/2019 19:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 04/29/2019 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 13:17:10 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

OFC. PASTOR IBM 5533 PRESENT

Data Download Date/Time 05/02/2019 09:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 05/31/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 11:38:48 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2019 12:50

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 06/30/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 09:57:08 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2019 10:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 07/26/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 14:16:11 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2019 14:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency	MIAMI PD	**Date**	08/01/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	15:15:35	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/01/2019 16:08

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency	MIAMI PD	**Date**	08/27/2019	**Type**	Agency Inspection
Inspector	FERNANDEZ,ANGEL,	**Time**	17:00:26	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	620051
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/03/2019 11:57

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 09/30/2019 **Type** Agency Inspection
Inspector PASTOR,JOSE,F **Time** 10:14:59 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 620051
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 11/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/03/2019 12:17

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency	MIAMI PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	PASTOR,JOSE,F	**Time**	11:10:00	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	08/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: .

Data Download Date/Time 11/04/2019 07:51

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency MIAMI PD **Date** 11/18/2019 **Type** Agency Inspection
Inspector FERNANDEZ,ANGEL, **Time** 13:42:23 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 989917
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/18/2019 14:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency	MIAMI PD	**Date**	12/23/2019	**Type**	Agency Inspection
Inspector	FERNANDEZ,ANGEL,	**Time**	18:48:08	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	989917
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/25/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/23/2019 19:46

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 01/03/2019 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:09:18 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 06717080A6
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 01/03/2019 11:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 02/06/2019 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 08:34:32 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 06717080A6
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 02/06/2019 09:48

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 03/18/2019 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:48:22 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 06717080A6
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 03/18/2019 11:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 04/16/2019 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:30:24 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 06717080A6
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 05/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 04/16/2019 09:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 05/06/2019 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:55:26 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 29017080A3
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 05/06/2019 12:26

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 05/16/2019 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 11:57:10 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 05/16/2019 12:04

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency	CALHOUN COUNTY SO	**Date**	06/03/2019	**Type**	Agency Inspection
Inspector	WILSON,STANLEY,T	**Time**	10:12:50	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	29017080A3
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 06/03/2019 11:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 07/08/2019 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:15:39 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 29017080A3
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 07/08/2019 12:54

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 08/12/2019 **Type** Agency Inspection
Inspector WILSON,STANLEY,T **Time** 09:44:40 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 29017080A3
Expiration 11/13/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 08/13/2019 07:54

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency	CALHOUN COUNTY SO	**Date**	09/02/2019	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	10:38:36	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	29017080A3
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 09/03/2019 09:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency	CALHOUN COUNTY SO	**Date**	10/01/2019	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	12:30:37	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	29017080A3
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 10/01/2019 14:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency	CALHOUN COUNTY SO	**Date**	11/04/2019	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	10:59:10	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	29017080A3
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 11/10/2019 23:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency CALHOUN COUNTY SO **Date** 12/12/2019 **Type** Agency Inspection
Inspector RUBEL,NEIL,F **Time** 00:05:44 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 970670
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/12/2019 01:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency HIALEAH PD **Date** 01/16/2019 **Type** Agency Inspection
Inspector MONTOYA,GENARO,N **Time** 11:24:08 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2019 11:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency HIALEAH PD **Date** 02/13/2019 **Type** Agency Inspection
Inspector MONTOYA,GENARO,N **Time** 11:16:23 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2019 11:25

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency HIALEAH PD **Date** 03/26/2019 **Type** Agency Inspection
Inspector MONTOYA,GENARO,N **Time** 10:47:48 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/26/2019 10:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency	HIALEAH PD	**Date**	04/16/2019	**Type**	Agency Inspection
Inspector	MONTOYA,GENARO,N	**Time**	11:12:57	**Software**	8100.27

0.05 Lot #	201802G	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	712031
Expiration	02/22/2018	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/16/2019 11:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency HIALEAH PD **Date** 05/23/2019 **Type** Agency Inspection
Inspector MONTOYA,GENARO,N **Time** 10:59:55 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/23/2019 11:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency HIALEAH PD **Date** 06/26/2019 **Type** Agency Inspection
Inspector MONTOYA,GENARO,N **Time** 10:33:13 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 712031
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.075
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/26/2019 10:45

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency	HIALEAH PD	**Date**	07/25/2019	**Type**	Agency Inspection
Inspector	MONTOYA,GENARO,N	**Time**	10:02:32	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	712031
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.075
		Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.074
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.075
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail / VENTILATED ROOM. 08: Control Outside Tolerance. LOW READING, REPEAT..

Data Download Date/Time 07/25/2019 10:16

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency	HIALEAH PD	**Date**	08/28/2019	**Type**	Agency Inspection
Inspector	MONTOYA,GENARO,N	**Time**	11:20:45	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	712031
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.075
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2019 11:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency	HIALEAH PD	**Date**	09/19/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	11:43:07	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 09/19/2019 12:01

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency	HIALEAH PD	**Date**	09/26/2019	**Type**	Agency Inspection
Inspector	MONTOYA,GENARO,N	**Time**	11:51:32	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	712031
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	02/21/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.074
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.075
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.074
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.075
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.075
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.075
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance. LOW READING.

Data Download Date/Time 09/26/2019 12:07

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency	HIALEAH PD	**Date**	10/22/2019	**Type**	Agency Inspection
Inspector	MONTOYA,GENARO,N	**Time**	10:59:08	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	712031
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	02/21/2020

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.074
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.073
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.074
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.074
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.074
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.074
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance / LOW READING , Control Outsi. Non-compliance:LOW READIND ON GAS STANDARD.

Data Download Date/Time 11/26/2019 15:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency	HIALEAH PD	**Date**	11/26/2019	**Type**	Agency Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	09:30:43	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BY-PASSED AI.COMPLIANCE NOT DETERMINED.

Data Download Date/Time 11/26/2019 15:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency	HIALEAH PD	**Date**	11/26/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	14:34:35	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 11/26/2019 15:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency HIALEAH PD **Date** 12/18/2019 **Type** Agency Inspection
Inspector MONTOYA,GENARO,N **Time** 12:16:52 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 712031
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 02/21/2020

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.075
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.075
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.074
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.074
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.075
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.074
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance / ONE LOW READING, Control Ou. Non-compliance:DRY GAS STANDARD LOW READING.

Data Download Date/Time 01/29/2020 12:46

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 01/14/2019 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 11:35:44 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/14/2019 11:42

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency	HOMESTEAD PD	**Date**	01/14/2019	**Type**	Agency Inspection
Inspector	PEREZ,RAUL,	**Time**	17:43:02	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	937240
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	04/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 01/14/2019 17:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 02/21/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 10:15:00 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 937240
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 04/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 02/21/2019 10:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 03/29/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 14:00:06 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 937240
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 04/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.190
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.191
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 03/29/2019 14:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 04/27/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 13:09:58 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 937240
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 04/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 04/27/2019 13:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 05/28/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 16:30:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 937240
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 04/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 05/28/2019 16:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency	HOMESTEAD PD	**Date**	06/30/2019	**Type**	Agency Inspection
Inspector	PEREZ,RAUL,	**Time**	21:21:53	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	937240
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	04/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 06/30/2019 21:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 07/30/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 08:12:07 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 1114319
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 03/27/2022

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 07/30/2019 08:24

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 08/30/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 03:00:16 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 1114319
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 03/27/2022

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.208
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.208
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 08/30/2019 03:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 09/26/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 09:49:05 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 1114319
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 03/27/2022

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 09/26/2019 10:06

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 10/31/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 11:00:03 **Software** 8100.27

0.05 Lot # 201811E **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 1114319
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 03/27/2022

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 10/31/2019 11:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency HOMESTEAD PD **Date** 11/22/2019 **Type** Agency Inspection
Inspector PEREZ,RAUL, **Time** 03:36:55 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 1114319
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 03/27/2022

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.207
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.208
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 11/22/2019 03:52

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency	HOMESTEAD PD	**Date**	12/31/2019	**Type**	Agency Inspection
Inspector	PEREZ,RAUL,	**Time**	16:34:53	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1114319
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/27/2022

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 12/31/2019 16:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	01/28/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	12:06:01	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/28/2019 12:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	02/27/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	13:13:23	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2019 13:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency MIAMI BEACH PD **Date** 03/07/2019 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 12:21:28 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/07/2019 12:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	03/27/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	09:25:36	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	2O1808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2019 09:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	04/26/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	11:14:47	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.085
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/26/2019 11:25

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	05/29/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	12:10:24	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/29/2019 12:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	06/28/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	12:27:45	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/28/2019 12:40

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	07/31/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	09:34:51	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 07/31/2019 09:50

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	08/31/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	14:41:26	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	814181
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	07/24/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/03/2019 15:51

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	09/30/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	09:33:06	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1175968
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	07/30/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI P.SOCARRAS #501 PRESENT

Data Download Date/Time 09/30/2019 09:47

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	10/29/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	13:54:43	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1175968
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	07/30/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Improper Sample.BLEW AT WRONG TIME

Data Download Date/Time 10/30/2019 11:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	11:26:57	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1175968
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	07/30/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI P. SOCARRAS (#501) PRESENT

Data Download Date/Time 11/15/2019 11:54

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

Agency	MIAMI BEACH PD	**Date**	12/26/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	11:51:41	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	1175968
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	07/30/2022

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI P.SOCARRAS #501 PRESENT

Data Download Date/Time 12/26/2019 14:35

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency	MIAMI-DADE PD	**Date**	01/29/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:25:23	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2019 09:51

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency MIAMI-DADE PD **Date** 02/07/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:20:14 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/08/2019 13:50

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency	MIAMI-DADE PD	**Date**	02/18/2019	**Type**	Department Inspection
Inspector	REYES-RIVERA,DAVID,E	**Time**	14:03:32	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805701
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0013 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.054
0.08 g/210L Test # 3	0.083
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.084
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.054
0.08 g/210L Test # 6	0.083
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0017
0.08 g/210L Test (Std Dev)	0.0017
0.20 g/210L Test (Std Dev)	0.0012
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0008

Remarks:

Data Download Date/Time 02/18/2019 14:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency MIAMI-DADE PD **Date** 08/20/2019 **Type** Department Inspection
Inspector HAUGHEY,MICHAEL,D **Time** 14:11:00 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG831804
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 08/20/2019 14:30

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency MIAMI-DADE PD **Date** 08/29/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:34:58 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2019 12:55

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency MIAMI-DADE PD **Date** 09/02/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:39:56 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.078
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance/FORGOT TO ATTACH HOSE.

Data Download Date/Time 09/03/2019 09:53

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency	MIAMI-DADE PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	11:48:28	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI LEY #6259 & MYRTIL #5383 PRESENT

Data Download Date/Time 10/31/2019 02:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency	MIAMI-DADE PD	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	13:47:21	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL#5383 PRESENT

Data Download Date/Time 11/15/2019 14:02

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency	MIAMI-DADE PD	**Date**	12/26/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	12:45:16	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL 5383 PRESENT

Data Download Date/Time 12/26/2019 12:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency MIAMI DADE PD **Date** 01/10/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:56:46 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail/CLEARED AREA.

Data Download Date/Time 01/10/2019 10:03

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	01/15/2019	**Type**	Department Inspection
Inspector	REYES-RIVERA,DAVID,E	**Time**	14:08:13	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805701
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/15/2019 14:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency MIAMI DADE PD **Date** 01/21/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:43:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/21/2019 11:45

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	02/21/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:45:17	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/21/2019 10:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	03/07/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:20:17	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/07/2019 09:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency MIAMI DADE PD **Date** 04/23/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:24:33 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2019 13:30

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	05/01/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	11:29:34	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/01/2019 12:04

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency MIAMI DADE PD **Date** 06/20/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 14:03:45 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2019 14:35

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	07/02/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:52:00	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/02/2019 11:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	08/09/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:34:03	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/09/2019 11:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	09/02/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	13:40:01	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/03/2019 09:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	11:50:22	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI LEY #6259 & MYRTIL #5383 PRESENT

Data Download Date/Time 10/31/2019 02:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	13:57:43	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL# 5383 PRESENT

Data Download Date/Time 11/15/2019 14:16

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency	MIAMI DADE PD	**Date**	12/26/2019	**Type**	Agency Inspection
Inspector	MILAN,KEVIN,	**Time**	12:40:22	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL 5383 PRESENT

Data Download Date/Time 12/26/2019 12:55

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency MIAMI-DADE PD **Date** 01/04/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:56:01 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2019 13:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	02/01/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:38:47	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2019 10:03

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency MIAMI-DADE PD **Date** 02/05/2019 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 08:14:03 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/05/2019 08:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	02/08/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:08:53	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/08/2019 11:13

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	03/22/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:29:16	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/22/2019 10:15

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency MIAMI-DADE PD **Date** 04/09/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:39:31 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/10/2019 08:59

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency MIAMI-DADE PD **Date** 05/15/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:03:50 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/16/2019 13:03

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	06/14/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:49:06	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/14/2019 11:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	07/10/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:55:27	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/10/2019 11:13

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	08/08/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:11:35	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Interferent Detect/STRONG PERFUME DETECTED ON PERSON .

Data Download Date/Time 08/08/2019 09:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	09/04/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:24:35	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/04/2019 09:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	13:01:09	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI LEY #6259 AND MYRTIL #5383 PRESENT

Data Download Date/Time 10/31/2019 01:17

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	MILLIAN,KEVIN,	**Time**	11:52:17	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL#5383

Data Download Date/Time 11/15/2019 12:19

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency	MIAMI-DADE PD	**Date**	12/26/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	12:29:11	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.AI MYRTIL5383

Data Download Date/Time 12/26/2019 12:42

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency MIAMI-DADE PD **Date** 01/31/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:53:59 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2019 13:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency	MIAMI-DADE PD	**Date**	02/07/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:24:35	**Software**	8100.27

0.05 Lot #	201703A	**0.08 Lot #**	201709A	**0.20 Lot #**	201702B	**0.08 Gas Lot #**	34917080A4
Expiration	03/07/2019	**Expiration**	09/12/2019	**Expiration**	02/23/2019	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/08/2019 07:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency	MIAMI-DADE PD	**Date**	02/19/2019	**Type**	Department Inspection
Inspector	REYES-RIVERA,DAVID,E	**Time**	08:13:42	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG805702
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/19/2019 08:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency	MIAMI-DADE PD	**Date**	08/20/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	13:15:06	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

Data Download Date/Time 08/20/2019 13:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency MIAMI-DADE PD **Date** 08/29/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:34:41 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
Interferent Detect Test # 3	INT

Remarks:

A F / M A:NOT ENOUGH MA. 08: Control Outside Tolerance/FORGOT TO ATTACH HOSE.

Data Download Date/Time 08/29/2019 13:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency MIAMI-DADE PD **Date** 09/02/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:32:13 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/03/2019 13:48

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency	MIAMI-DADE PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	MILLIAN,KEVIN,	**Time**	13:46:14	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI LEY #6259 & MYRTIL #5383 PRESENT

Data Download Date/Time 10/31/2019 03:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency MIAMI-DADE PD **Date** 11/15/2019 **Type** Agency Inspection
Inspector MILLAN,KEVIN, **Time** 12:09:26 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 06419080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance,HOSE NOT CONNECTED AI MYRTIL

Data Download Date/Time 11/15/2019 12:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

Agency	MIAMI-DADE PD	**Date**	12/27/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	11:19:57	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL #5383 PRESENT

Data Download Date/Time 12/27/2019 11:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency MIAMI-DADE PD **Date** 01/04/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:01:26 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2019 13:08

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency MIAMI-DADE PD **Date** 02/18/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 12:15:30 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2019 12:26

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency MIAMI-DADE PD **Date** 02/25/2019 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 13:47:20 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805702
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/25/2019 13:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency	MIAMI-DADE PD	**Date**	03/06/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:54:26	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/06/2019 10:01

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency MIAMI-DADE PD **Date** 04/15/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:26:59 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/16/2019 13:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency MIAMI-DADE PD **Date** 05/15/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:00:58 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/16/2019 13:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency MIAMI-DADE PD **Date** 06/12/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:59:21 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/12/2019 11:25

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency MIAMI-DADE PD **Date** 07/12/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:12:15 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.085
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/12/2019 13:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency	MIAMI-DADE PD	**Date**	08/02/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	10:49:43	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/02/2019 14:21

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency	MIAMI-DADE PD	**Date**	09/13/2019	**Type**	Agency Inspection
Inspector	ESPINOZA,PABLO,	**Time**	09:47:48	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/17/2019 08:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency	MIAMI-DADE PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	11:36:11	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL #5383 AND LEY #6259 PRESENT

Data Download Date/Time 10/31/2019 03:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency	MIAMI-DADE PD	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	13:09:15	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL #5383 PRESENT

Data Download Date/Time 11/15/2019 13:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency	MIAMI-DADE PD	**Date**	12/26/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	12:23:41	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	34917080A4
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL 5383 PRESENT

Data Download Date/Time 12/26/2019 12:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency MIAMI-DADE PD **Date** 01/10/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 09:51:17 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A4
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/10/2019 10:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency MIAMI-DADE PD **Date** 08/20/2019 **Type** Department Inspection
Inspector HAUGHEY,MICHAEL,D **Time** 13:58:21 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG831804
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.08 g/210L Test # (Repeat) 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.08 g/210L Test # (Repeat) 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.08 g/210L Test # (Repeat) 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.077
0.08 g/210L Test # (Repeat) 4	0.080
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

0.08 g/210L Test # 5	0.076
0.08 g/210L Test # (Repeat) 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.076
0.08 g/210L Test # (Repeat) 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.075
0.08 g/210L Test # (Repeat) 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.074
0.08 g/210L Test # (Repeat) 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.074
0.08 g/210L Test # (Repeat) 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.073
0.08 g/210L Test # (Repeat) 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0020
0.08 g/210L Test (Std Dev) (Repeat)	0.0003
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

08: Control Outside Tolerance SIM LOST POWER. RECONNECTED .

Data Download Date/Time 08/20/2019 14:19

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency MIAMI-DADE PD **Date** 08/29/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 11:15:07 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2019 13:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency MIAMI-DADE PD **Date** 09/02/2019 **Type** Agency Inspection
Inspector ESPINOZA,PABLO, **Time** 10:14:35 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 22817080A4
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 10/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance/FORGOT TO ATTACH HOSE.

Data Download Date/Time 09/03/2019 09:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency	MIAMI-DADE PD	**Date**	10/30/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	14:30:32	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI LEY #6259 & MYRTIL #5383 PRESENT

Data Download Date/Time 10/31/2019 01:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency MIAMI-DADE PD **Date** 11/15/2019 **Type** Agency Inspection
Inspector MILLAN,KEVIN, **Time** 13:59:36 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 06419080A2
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance HOSE NOT CONNECTED.AI MYRTIL

Data Download Date/Time 11/15/2019 14:12

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency	MIAMI-DADE PD	**Date**	12/26/2019	**Type**	Agency Inspection
Inspector	MILLAN,KEVIN,	**Time**	15:01:48	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	06419080A2
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	05/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI MYRTIL 5383 PRESENT

Data Download Date/Time 12/26/2019 15:12

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency NORTH MIAMI BEACH PD **Date** 01/30/2019 **Type** Agency Inspection
Inspector MANN,CORA,L **Time** 18:15:43 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 12317080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2019 18:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency NORTH MIAMI BEACH PD **Date** 02/24/2019 **Type** Agency Inspection
Inspector MANN,CORA,L **Time** 20:49:29 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 12317080A3
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/24/2019 20:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency NORTH MIAMI BEACH PD **Date** 03/25/2019 **Type** Agency Inspection
Inspector MANN,CORA,L **Time** 21:21:15 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 12317080A3
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/25/2019 21:28

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency NORTH MIAMI BEACH PD **Date** 03/29/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 14:27:03 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/29/2019 14:32

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency	NORTH MIAMI BEACH PD	**Date**	04/03/2019	**Type**	Agency Inspection
Inspector	MANN,CORA,L	**Time**	11:26:31	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	12317080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/03/2019 11:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency	NORTH MIAMI BEACH PD	**Date**	05/23/2019	**Type**	Agency Inspection
Inspector	MANN,CORA,L	**Time**	02:01:32	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	12317080A3
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	07/05/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/23/2019 02:12

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency	NORTH MIAMI BEACH PD	**Date**	05/23/2019	**Type**	Agency Inspection
Inspector	MANN,CORA,L	**Time**	02:05:47	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 05/23/2019 02:12

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency	NORTH MIAMI BEACH PD	**Date**	06/29/2019	**Type**	Agency Inspection
Inspector	BAKER,KENNETH,	**Time**	17:35:29	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/29/2019 17:53

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency NORTH MIAMI BEACH PD **Date** 07/30/2019 **Type** Agency Inspection
Inspector BAKER,KENNETH, **Time** 22:03:04 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.061
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.064
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.046
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.062
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.066
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.046
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.063
		0.08 g/210L Test # (Repeat) 3	0.068
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 07/31/2019 23:25

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency	NORTH MIAMI BEACH PD	**Date**	07/31/2019	**Type**	Agency Inspection
Inspector	BAKER,KENNETH,	**Time**	23:18:22	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/31/2019 23:25

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency	NORTH MIAMI BEACH PD	**Date**	08/30/2019	**Type**	Agency Inspection
Inspector	MANN,CORA,L	**Time**	13:28:45	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.187
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.190
		0.20 g/210L Test # (Repeat) 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance/IMPROPER SEAL.

Data Download Date/Time 08/30/2019 13:42

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency	NORTH MIAMI BEACH PD	**Date**	09/30/2019	**Type**	Agency Inspection
Inspector	MANN,CORA,L	**Time**	06:03:10	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/30/2019 06:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency NORTH MIAMI BEACH PD **Date** 10/28/2019 **Type** Agency Inspection
Inspector MANN,CORA,L **Time** 06:18:03 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded BLEW OUT OF SEQUENCE.

Data Download Date/Time 10/28/2019 06:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency NORTH MIAMI BEACH PD **Date** 11/25/2019 **Type** Agency Inspection
Inspector MANN,CORA,L **Time** 05:38:10 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A4
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/25/2019 05:48

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency	NORTH MIAMI BEACH PD	**Date**	12/31/2019	**Type**	Agency Inspection
Inspector	MANN,CORA,L	**Time**	19:25:21	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	34917080A4
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	02/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/31/2019 19:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency	CHATTAHOOCHEE PD	**Date**	01/10/2019	**Type**	Agency Inspection
Inspector	OBRIAN,FREDERICK,C	**Time**	12:28:10	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:TIGHTEN SIM

Data Download Date/Time 01/10/2019 12:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency	CHATTAHOOCHEE PD	**Date**	04/23/2019	**Type**	Department Inspection
Inspector	GRAHAM,THOMAS,J	**Time**	13:43:22	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG831804
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 04/23/2019 13:47

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency	CHATTAHOOCHEE PD	**Date**	04/25/2019	**Type**	Agency Inspection
Inspector	OBRIAN,FREDERICK,C	**Time**	17:38:02	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 04/25/2019 18:02

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency CHATTAHOOCHEE PD **Date** 05/29/2019 **Type** Agency Inspection
Inspector OBRIAN,FREDERICK,C **Time** 13:51:45 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 970670
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.188
0.20 g/210L Test # (Repeat) 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.189
0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.189
0.20 g/210L Test # (Repeat) 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	0.000
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:TIGHTEN SIM CAP. 20: Control Outside Tolerance.

Data Download Date/Time 06/05/2019 11:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency	CHATTAHOOCHEE PD	**Date**	06/25/2019	**Type**	Agency Inspection
Inspector	OBRIAN,FREDERICK,C	**Time**	15:09:32	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.174
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.170
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.174
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.175
0.20 g/210L Test: OK	No	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.175
		0.20 g/210L Test # (Repeat) 3	0.178
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 06/25/2019 17:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency	CHATTAHOOCHEE PD	**Date**	06/25/2019	**Type**	Agency Inspection
Inspector	OBRIAN,FREDERICK,C	**Time**	17:11:51	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.054
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.054
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/25/2019 17:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency	CHATTAHOOCHEE PD	**Date**	10/23/2019	**Type**	Agency Inspection
Inspector	OBRIAN,FREDERICK,C	**Time**	13:23:26	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.085
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.085
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.085
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2019 17:01

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency	CHATTAHOOCHEE PD	**Date**	11/29/2019	**Type**	Agency Inspection
Inspector	OBRIAN,FREDERICK,C	**Time**	13:31:52	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.079
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.080
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: . 05: Control Outside Tolerance.

Data Download Date/Time 11/29/2019 13:37

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

Agency	CHATTAHOOCHEE PD	**Date**	12/30/2019	**Type**	Agency Inspection
Inspector	OBRIAN,FREDERICK,C	**Time**	12:02:16	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	970670
Expiration	11/13/2029	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/07/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/30/2019 12:14

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

Agency	PINELLAS COUNTY SO	**Date**	01/08/2019	**Type**	Agency Inspection
Inspector	GRAHAM,THOMAS,J	**Time**	12:37:44	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

Data Download Date/Time 01/08/2019 15:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

Agency PINELLAS COUNTY SO **Date** 01/08/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 15:17:53 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/08/2019 15:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

Agency	PINELLAS COUNTY SO	**Date**	01/18/2019	**Type**	Agency Inspection
Inspector	GRAHAM,THOMAS,J	**Time**	07:14:51	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

Data Download Date/Time 01/18/2019 09:53

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

Agency PINELLAS COUNTY SO **Date** 01/18/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 09:47:37 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 01/18/2019 09:53

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

Agency	PINELLAS COUNTY SO	**Date**	01/08/2019	**Type**	Agency Inspection
Inspector	GRAHAM,THOMAS,J	**Time**	12:36:41	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

Data Download Date/Time 01/08/2019 15:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

Agency PINELLAS COUNTY SO **Date** 01/08/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 15:06:11 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/08/2019 15:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency	FWC	**Date**	01/22/2019	**Type**	Agency Inspection
Inspector	MOLNAR,JARROD,T	**Time**	13:26:06	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	962587
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2019 14:31

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency	FWC	**Date**	02/25/2019	**Type**	Agency Inspection
Inspector	MOLNAR,JARROD,T	**Time**	13:53:04	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	962587
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/25/2019 15:05

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency	FWC	**Date**	03/24/2019	**Type**	Agency Inspection
Inspector	BASFORD,NICOLE,M	**Time**	14:46:00	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	962587
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/24/2019 15:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 04/16/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 12:32:20 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

READ MOUTH ALCOHOL.

Data Download Date/Time 04/16/2019 13:41

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 05/13/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 19:35:06 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 02/06/2020 **Expiration** 02/22/2020 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 05/13/2019 20:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 06/13/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 12:28:30 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 06/13/2019 13:36

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 07/17/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 13:21:05 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 07/17/2019 14:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 08/22/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 12:44:02 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 08/22/2019 13:54

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 09/11/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 15:09:23 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 09/11/2019 16:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 09/12/2019 **Type** Department Inspection
Inspector MURPHY,PATRICK,J **Time** 11:42:44 **Software** 8100.27

0.05 Lot # 201905A **0.08 Lot #** 201905B **0.20 Lot #** 201904D **0.08 Gas Lot #** AG916501
Expiration 05/14/2021 **Expiration** 05/14/2021 **Expiration** 04/30/2021 **Expiration** 06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	INT
0.08 g/210L Test # (Repeat) 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # (Repeat) 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # (Repeat) 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # (Repeat) 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.081

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # (Repeat) 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # (Repeat) 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # (Repeat) 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # (Repeat) 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # (Repeat) 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev) (Repeat)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

08: Interferent Detect.BLEW SIM TUBES CLEAN

Data Download Date/Time 09/12/2019 12:52

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 09/13/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 13:44:11 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 09/13/2019 14:59

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 10/17/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 13:45:49 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 10/17/2019 14:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 11/27/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 10:49:02 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 11/27/2019 11:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency FWC **Date** 12/15/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 15:07:51 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201908B **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587
Expiration 11/13/2020 **Expiration** 08/07/2021 **Expiration** 02/22/2020 **Expiration** 06/04/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 12/16/2019 10:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 01/22/2019 **Type** Agency Inspection
Inspector MOLNAR,JARROD,T **Time** 13:13:31 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2019 14:22

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency	FWC	**Date**	02/25/2019	**Type**	Agency Inspection
Inspector	MOLNAR,JARROD,T	**Time**	13:42:48	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	477627
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/25/2019 14:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency	FWC	**Date**	03/24/2019	**Type**	Agency Inspection
Inspector	BASFORD,NICOLE,M	**Time**	14:45:00	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	477627
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/24/2019 16:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 04/16/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 12:26:13 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

READ MOUTH ALCOHOL.

Data Download Date/Time 04/16/2019 13:39

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency	FWC	**Date**	05/13/2019	**Type**	Agency Inspection
Inspector	BASFORD,NICOLE,M	**Time**	19:40:33	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	477627
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	06/29/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 05/13/2019 20:51

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 06/13/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 11:36:14 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 1060911
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/21/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 06/13/2019 12:45

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency	FWC	**Date**	07/17/2019	**Type**	Agency Inspection
Inspector	BASFORD,NICOLE,M	**Time**	13:25:18	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	1060911
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	12/21/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 07/17/2019 14:44

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 08/22/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 12:50:28 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 1060911
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/21/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.082
Interferent Detect Test # 3	INT

Remarks:

FORGOT TO CONNECT.

Data Download Date/Time 08/22/2019 14:00

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 09/11/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 15:02:15 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 1060911
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/21/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 09/11/2019 16:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency	FWC	**Date**	09/12/2019	**Type**	Department Inspection
Inspector	MURPHY,PATRICK,J	**Time**	11:11:19	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG916501
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	06/14/2021

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/12/2019 12:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 09/13/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 13:40:38 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 1060911
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/21/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 09/13/2019 15:05

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 10/17/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 13:38:28 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 1060911
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/21/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 10/17/2019 14:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 11/27/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 10:51:50 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 1060911
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 12/21/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 11/27/2019 12:26

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency FWC **Date** 12/15/2019 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 14:57:24 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201908B **0.20 Lot #** 201802G **0.08 Gas Lot #** 1060911
Expiration 11/13/2020 **Expiration** 08/07/2021 **Expiration** 02/22/2020 **Expiration** 12/21/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 12/19/2019 17:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	03/24/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	19:12:33	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RAN OUT OF MOUTH ALCOHOL SOLUTION. Non-compliance: .

Data Download Date/Time 03/27/2019 09:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	03/24/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	20:28:53	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	09117080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	INT
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: Interferent Detect.

Data Download Date/Time 03/27/2019 09:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	03/28/2019	**Type**	Department Inspection		
Inspector	KIRKLAND,BRETT,H	**Time**	13:11:49	**Software**	8100.27		
0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG831804
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	11/14/2020

Number of Simulators 4 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/28/2019 14:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	04/08/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	19:53:09	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	09117080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT BACK IN SERVICE

Data Download Date/Time 04/29/2019 13:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	05/02/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	13:39:53	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	]	**0.08 Gas Lot #**	09117080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 05/02/2019 16:23

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	06/20/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	12:04:37	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT IN COMPLIANCE

Data Download Date/Time 06/28/2019 08:49

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency FWC **Date** 07/30/2019 **Type** Agency Inspection
Inspector JOHNSTON,ROBERT,B **Time** 15:55:23 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 00919080A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det: .CORECTED LEVELS OF ACETONE AND INSTRUMENT PASSED

Data Download Date/Time 08/26/2019 09:51

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	08/28/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	19:41:15	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT IN COMPLIANCE

Data Download Date/Time 10/28/2019 09:42

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	10/22/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	11:48:32	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Improper Sample.

Data Download Date/Time 10/28/2019 09:42

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	GILMORE,RICHARD,L	**Time**	11:27:47	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: .

Data Download Date/Time 11/15/2019 13:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

Agency	FWC	**Date**	12/14/2019	**Type**	Agency Inspection
Inspector	GILMORE,RICHARD,L	**Time**	09:38:18	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/14/2019 10:02

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency FWC **Date** 03/24/2019 **Type** Agency Inspection
Inspector JOHNSTON,ROBERT,B **Time** 19:14:10 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

RAN OUT OF MOUTH ALCOHOL SOLUTION.

Data Download Date/Time 03/27/2019 09:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency FWC **Date** 03/24/2019 **Type** Agency Inspection
Inspector JOHNSTON,ROBERT,B **Time** 20:39:36 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT READY FOR DEPT INSPECTION

Data Download Date/Time 03/27/2019 09:11

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency	FWC	**Date**	03/28/2019	**Type**	Department Inspection
Inspector	KIRKLAND,BRETT,H	**Time**	13:00:56	**Software**	8100.27

0.05 Lot #	201707D	**0.08 Lot #**	201707E	**0.20 Lot #**	201707C	**0.08 Gas Lot #**	AG831804
Expiration	07/25/2019	**Expiration**	07/25/2019	**Expiration**	07/24/2019	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.082
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.082
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/28/2019 13:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency	FWC	**Date**	04/08/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	19:57:30	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	09117080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT BACK IN SERVICE

Data Download Date/Time 04/29/2019 13:27

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency	FWC	**Date**	05/02/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	13:36:55	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	09117080A2
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	06/05/2019

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 05/02/2019 16:26

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency	FWC	**Date**	06/20/2019	**Type**	Agency Inspection		
Inspector	JOHNSTON,ROBERT,B	**Time**	11:52:40	**Software**	8100.27		
0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	0919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT IN COMPLIANCE

Data Download Date/Time 06/28/2019 08:25

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency FWC **Date** 07/30/2019 **Type** Agency Inspection
Inspector JOHNSTON,ROBERT,B **Time** 15:51:51 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 00919080A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT IN COMPLIANCE

Data Download Date/Time 08/26/2019 10:13

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency	FWC	**Date**	08/28/2019	**Type**	Agency Inspection
Inspector	JOHNSTON,ROBERT,B	**Time**	19:37:48	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT IN COMPLIANCE

Data Download Date/Time 10/28/2019 09:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency	FWC	**Date**	10/22/2019	**Type**	Agency Inspection		
Inspector	JOHNSTON,ROBERT,B	**Time**	11:47:19	**Software**	8100.27		
0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

INSTRUMENT IN COMPLIANCE

Data Download Date/Time 10/28/2019 09:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency	FWC	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	GILMORE,RICHARD,L	**Time**	11:24:38	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	INT
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:NEW SIM WASNT WARM,BLOWN FUSE. Non-compliance:INTERFERENT NO DETECTED ALL 3 TEST.

Data Download Date/Time 11/15/2019 13:38

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

Agency FWC **Date** 12/14/2019 **Type** Agency Inspection
Inspector GILMORE,RICHARD,L **Time** 09:30:10 **Software** 8100.27

0.05 Lot # 201811C **0.08 Lot #** 201808E **0.20 Lot #** 201902C **0.08 Gas Lot #** 00919080A1
Expiration 11/13/2020 **Expiration** 08/22/2020 **Expiration** 02/20/2021 **Expiration** 03/05/2021

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/14/2019 09:58

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency FFWCC TAMPA **Date** 01/02/2019 **Type** Agency Inspection
Inspector PULASKI,DAMON,J **Time** 12:47:32 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/02/2019 13:07

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	02/13/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	20:59:38	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13518080A5
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/13/2019 22:01

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency FFWCC TAMPA **Date** 03/18/2019 **Type** Agency Inspection
Inspector PULASKI,DAMON,J **Time** 10:34:53 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/18/2019 10:46

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	04/24/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	19:07:13	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13518080A5
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/24/2019 19:40

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	05/20/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	21:40:07	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13518080A5
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 05/21/2019 19:33

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency FFWCC TAMPA **Date** 06/21/2019 **Type** Agency Inspection
Inspector PULASKI,DAMON,J **Time** 13:40:03 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** 201808E **0.20 Lot #** 201802G **0.08 Gas Lot #** 13518080A5
Expiration 02/06/2020 **Expiration** 08/22/2020 **Expiration** 02/22/2020 **Expiration** 08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 06/25/2019 10:30

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	07/26/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	19:23:13	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13518080A5
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 07/26/2019 22:57

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	08/24/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	12:49:32	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	13518080A5
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 08/24/2019 13:30

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	09/12/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	14:31:12	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	13518080A5
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 09/13/2019 09:12

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	09/19/2019	**Type**	Department Inspection
Inspector	HAUGHEY,MICHAEL,D	**Time**	15:11:25	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/19/2019 15:43

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	10/04/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	22:18:25	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 10/04/2019 22:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	20:12:37	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 11/15/2019 20:24

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency	FFWCC TAMPA	**Date**	12/13/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	12:59:34	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	00919080A1
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	03/05/2021

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 12/13/2019 13:34

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Agency	FFWCC TAMPA	**Date**	01/02/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	12:10:10	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201709A	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13518080A5
Expiration	02/06/2020	**Expiration**	09/12/2019	**Expiration**	02/22/2020	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/02/2019 12:56

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Agency	FFWCC TAMPA	**Date**	02/13/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	21:43:09	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13518080A5
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/13/2019 21:50

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Agency	FFWCC TAMPA	**Date**	03/18/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	11:57:32	**Software**	8100.27

0.05 Lot #	201802B	**0.08 Lot #**	201808E	**0.20 Lot #**	201802G	**0.08 Gas Lot #**	13518080A5
Expiration	02/06/2020	**Expiration**	08/22/2020	**Expiration**	02/22/2020	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/18/2019 12:09

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Agency	FFWCC TAMPA	**Date**	08/29/2019	**Type**	Department Inspection
Inspector	PLATT,SHAYLA,D	**Time**	13:45:22	**Software**	8100.27

0.05 Lot #	201905A	**0.08 Lot #**	201905B	**0.20 Lot #**	201904D	**0.08 Gas Lot #**	AG831804
Expiration	05/14/2021	**Expiration**	05/14/2021	**Expiration**	04/30/2021	**Expiration**	11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 08/29/2019 13:51

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Agency	FFWCC TAMPA	**Date**	09/12/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	13:45:29	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	13518080A5
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 09/13/2019 09:20

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Agency	FFWCC TAMPA	**Date**	10/04/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	21:40:15	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	13518080A5
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 10/04/2019 21:57

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Agency	FFWCC TAMPA	**Date**	11/15/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	19:04:11	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	13518080A5
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/15/2019 19:18

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

Agency	FFWCC TAMPA	**Date**	12/13/2019	**Type**	Agency Inspection
Inspector	PULASKI,DAMON,J	**Time**	12:22:48	**Software**	8100.27

0.05 Lot #	201811C	**0.08 Lot #**	201808E	**0.20 Lot #**	201902C	**0.08 Gas Lot #**	13518080A5
Expiration	11/13/2020	**Expiration**	08/22/2020	**Expiration**	02/20/2021	**Expiration**	08/05/2020

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 12/13/2019 13:29

Printed: 04 February 2020

COBRA ©1997-2014 CMI, Inc.