

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

Agency: FDLE

Date/Time: 01/17/2020 10:55:34

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/22/2020 11:25

Air Blank		10:55
Air Blank	0.000	10:55

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

Agency: FDLE

Date/Time: 02/12/2020 17:45:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/04/2020 12:44

Air Blank	0.000	17:44
Control Test	0.193	17:44
Air Blank	0.000	17:45

---

Agency: FDLE

Date/Time: 02/12/2020 17:49:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/04/2020 12:45

Air Blank	0.000	17:47
Control Test	0.073	17:48
Air Blank	0.000	17:49

---

Agency: FDLE

Date/Time: 02/12/2020 18:03:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/04/2020 12:45

Air Blank	0.000	18:02
Control Test	0.075	18:03
Air Blank	0.000	18:03

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 02/20/2020 16:07:00

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/27/2020 11:56

Air Blank	0.000	16:06
Control Test	0.081	16:06
Air Blank	0.000	16:07

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

Agency: MARION COUNTY SO

Date/Time: 02/24/2020 15:39:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 14:25

Air Blank	0.000	15:38
Control Test	0.000	15:39
Air Blank	0.000	15:39

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 01/19/2020 14:04:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 00:17

Air Blank	0.000	14:03
Control Test	0.000	14:04
Air Blank	0.000	14:04

---

Agency: LEESBURG P.D.

Date/Time: 01/19/2020 14:06:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 00:17

Air Blank	0.000	14:05
Control Test	0.000	14:06
Air Blank	0.000	14:06

---

Agency: LEESBURG P.D.

Date/Time: 01/19/2020 14:08:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 00:17

Air Blank	0.000	14:07
Control Test	0.000	14:08
Air Blank	0.000	14:08

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 01/22/2020 09:21:31

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/27/2020 14:31

Air Blank		09:21
Air Blank	0.000	09:21

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

Agency: LEE COUNTY SO

Date/Time: 03/18/2020 07:52:16

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/08/2020 10:59

Air Blank	0.000	07:51
Control Test	0.081	07:51
Air Blank	0.000	07:52

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

Agency: COLLIER COUNTY SO

Date/Time: 02/14/2020 13:59:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/24/2020 10:35

Air Blank	0.000	13:58
Control Test	0.081	13:58
Air Blank	0.000	13:59

---

Agency: COLLIER COUNTY SO

Date/Time: 02/14/2020 14:01:24

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/24/2020 10:35

Air Blank	0.000	14:00
ITP Check	0.080	14:00
Air Blank	0.000	14:01

---

Agency: COLLIER COUNTY SO

Date/Time: 02/24/2020 09:05:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/24/2020 10:35

Air Blank	0.000	09:05
Control Test	0.082	09:05
Air Blank	0.000	09:05

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

Agency: GLADES COUNTY S.O.

Date/Time: 01/21/2020 18:56:12

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/23/2020 23:38

Air Blank	0.000	18:55
Control Test	0.078	18:55
Air Blank	0.000	18:56

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

Agency: GLADES COUNTY S.O.

Date/Time: 01/21/2020 18:54:12

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/23/2020 23:57

Air Blank	0.000	18:53
Control Test	0.076	18:53
Air Blank	0.000	18:54

---

Agency: GLADES COUNTY S.O.

Date/Time: 01/21/2020 18:57:37

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/23/2020 23:57

Air Blank	0.000	18:56
Control Test	0.076	18:57
Air Blank	0.000	18:57

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

Agency: CLEWISTON P.D.

Date/Time: 03/21/2020 23:06:33

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/07/2020 13:34

Air Blank	0.000	23:05
Control Test	0.085	23:06
Air Blank	0.000	23:06

---

Agency: CLEWISTON P.D.

Date/Time: 03/23/2020 06:17:57

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/07/2020 13:34

Air Blank	0.000	06:17
Control Test	0.079	06:17
Air Blank	0.000	06:17

---

Agency: CLEWISTON P.D.

Date/Time: 03/23/2020 06:19:22

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 04/07/2020 13:34

Air Blank	0.000	06:18
ITP Check	0.079	06:18
Air Blank	0.000	06:19

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 02/26/2020 20:56:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/26/2020 22:42

Air Blank	0.000	20:55
Control Test	0.082	20:55
Air Blank	0.000	20:56

---

Agency: TITUSVILLE P.D.

Date/Time: 03/08/2020 16:09:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/08/2020 17:31

Air Blank	0.000	16:08
Control Test	0.082	16:08
Air Blank	0.000	16:09

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

Agency: PINELLAS COUNTY SO

Date/Time: 01/31/2020 00:48:09

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/05/2020 16:52

Air Blank	0.000	00:47
Control Test	0.076	00:47
Air Blank	0.000	00:48

---

Agency: PINELLAS COUNTY SO

Date/Time: 01/31/2020 00:51:32

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/05/2020 16:52

Air Blank	0.000	00:50
ITP Check	0.077	00:51
Air Blank	0.000	00:51

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

Agency: PINELLAS COUNTY SO

Date/Time: 03/29/2020 11:50:40

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/30/2020 09:28

Air Blank		11:50
Air Blank	0.000	11:50

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 01/28/2020 11:08:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/28/2020 13:22

Air Blank	0.000	11:07
Control Test	0.000	11:08
Air Blank	0.000	11:08

---

Agency: INDIALANTIC P.D.

Date/Time: 01/28/2020 11:10:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/28/2020 13:22

Air Blank	0.000	11:09
Control Test	0.047	11:10
Air Blank	0.000	11:10

---

Agency: INDIALANTIC P.D.

Date/Time: 02/27/2020 13:33:27

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/27/2020 16:16

Air Blank	0.000	13:32
Control Test	0.081	13:32
Air Blank	0.000	13:33

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 02/27/2020 14:03:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/27/2020 16:16

Air Blank	0.000	14:02
Control Test	0.047	14:03
Air Blank	0.000	14:03

---

Agency: INDIALANTIC P.D.

Date/Time: 03/17/2020 13:36:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/17/2020 16:11

Air Blank	0.000	13:35
Control Test	0.045	13:35
Air Blank	0.000	13:36

---

Agency: INDIALANTIC P.D.

Date/Time: 03/17/2020 13:50:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/17/2020 16:11

Air Blank	0.000	13:49
Control Test	0.046	13:50
Air Blank	0.000	13:50

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

Agency: PALM BAY P.D.

Date/Time: 01/15/2020 22:56:13

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/28/2020 17:13

Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

Agency: JACKSONVILLE SO

Date/Time: 03/12/2020 14:52:39

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/25/2020 05:52

Air Blank	0.000	14:51
Control Test	0.078	14:52
Air Blank	0.000	14:52

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:11:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:05

Air Blank	0.000	05:10
Control Test	0.049	05:10
Air Blank	0.000	05:11

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:14:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:05

Air Blank	0.000	05:12
Control Test	0.199	05:13
Air Blank	0.000	05:14

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:16:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:05

Air Blank	0.000	05:15
Control Test	0.077	05:16
Air Blank	0.000	05:16

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:19:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:05

Air Blank	0.000	05:17
Control Test	0.000	05:18
Air Blank	0.000	05:19

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 01:09:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:05

Air Blank	0.000	01:08
Control Test	0.047	01:08
Air Blank	0.000	01:09

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 01:11:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:05

Air Blank	0.000	01:10
Control Test	0.195	01:11
Air Blank	0.000	01:11

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 01:14:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:05

Air Blank	0.000	01:12
Control Test	0.074	01:13
Air Blank	0.000	01:14

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 01:15:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:05

Air Blank	0.000	01:14
Control Test	0.076	01:15
Air Blank	0.000	01:15

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 01:18:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:06

Air Blank	0.000	01:16
Control Test	0.000	01:17
Air Blank	0.000	01:18

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/18/2020 06:01:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/18/2020 07:18

Air Blank	0.000	05:59
Control Test	0.048	06:00
Air Blank	0.000	06:01

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/18/2020 06:03:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/18/2020 07:19

Air Blank	0.000	06:02
Control Test	0.199	06:02
Air Blank	0.000	06:03

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/18/2020 06:05:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/18/2020 07:19

Air Blank	0.000	06:04
Control Test	0.077	06:05
Air Blank	0.000	06:05

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/18/2020 06:07:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/18/2020 07:19

Air Blank	0.000	06:06
Control Test	0.000	06:07
Air Blank	0.000	06:07

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 01/23/2020 13:29:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/23/2020 15:54

Air Blank	0.000	13:28
Control Test	0.076	13:29
Air Blank	0.000	13:29

---

Agency: NASSAU CO SO

Date/Time: 01/23/2020 13:31:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/23/2020 15:54

Air Blank	0.000	13:30
Control Test	0.047	13:30
Air Blank	0.000	13:31

---

Agency: NASSAU CO SO

Date/Time: 01/23/2020 13:35:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/23/2020 15:54

Air Blank	0.000	13:34
Control Test	0.077	13:34
Air Blank	0.000	13:35

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:07:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:37

Air Blank	0.000	05:06
Control Test	0.049	05:07
Air Blank	0.000	05:07

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:10:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:37

Air Blank	0.000	05:09
Control Test	0.198	05:09
Air Blank	0.000	05:10

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:12:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:37

Air Blank	0.000	05:11
Control Test	0.077	05:12
Air Blank	0.000	05:12

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:15:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:37

Air Blank	0.000	05:14
Control Test	0.000	05:15
Air Blank	0.000	05:15

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/13/2020 05:17:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/13/2020 06:37

Air Blank	0.000	05:16
Control Test	0.000	05:17
Air Blank	0.000	05:17

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 02:06:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:59

Air Blank	0.000	02:05
Control Test	0.000	02:06
Air Blank	0.000	02:06

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 02:09:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:59

Air Blank	0.000	02:07
Control Test	0.046	02:08
Air Blank	0.000	02:09

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 02:11:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:59

Air Blank	0.000	02:10
Control Test	0.076	02:11
Air Blank	0.000	02:11

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 02:13:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:59

Air Blank	0.000	02:12
Control Test	0.190	02:13
Air Blank	0.000	02:13

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 02:14:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:59

Air Blank		02:14
Air Blank	0.000	02:14

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/13/2020 02:16:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2020 02:59

Air Blank	0.000	02:15
Control Test	0.191	02:16
Air Blank	0.000	02:16

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/18/2020 05:57:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/18/2020 06:42

Air Blank	0.000	05:56
Control Test	0.050	05:57
Air Blank	0.000	05:57

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/18/2020 06:00:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/18/2020 06:42

Air Blank	0.000	05:59
Control Test	0.080	06:00
Air Blank	0.000	06:00

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/18/2020 06:03:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/18/2020 06:42

Air Blank	0.000	06:01
Control Test	0.000	06:02
Air Blank	0.000	06:03

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/18/2020 06:05:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/18/2020 06:42

Air Blank	0.000	06:04
Control Test	0.198	06:04
Air Blank	0.000	06:05

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 01/24/2020 05:53:37

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/27/2020 09:05

Air Blank	0.000	05:52
Control Test	0.081	05:53
Air Blank	0.000	05:53

---

Agency: POLK COUNTY SO

Date/Time: 01/24/2020 05:55:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2020 09:05

Air Blank	0.000	05:54
Control Test	0.000	05:55
Air Blank	0.000	05:55

---

Agency: POLK COUNTY SO

Date/Time: 01/24/2020 11:36:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2020 09:05

Air Blank	0.000	11:35
Control Test	0.000	11:36
Air Blank	0.000	11:36

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 02/24/2020 17:29:08

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/25/2020 08:44

Air Blank	0.000	17:28
Control Test	0.081	17:28
Air Blank	0.000	17:29

---

Agency: POLK COUNTY SO

Date/Time: 02/24/2020 17:31:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 08:44

Air Blank	0.000	17:29
Control Test	0.000	17:30
Air Blank	0.000	17:31

---

Agency: POLK COUNTY SO

Date/Time: 02/24/2020 17:32:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 08:45

Air Blank	0.000	17:31
Control Test	0.196	17:32
Air Blank	0.000	17:32

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 02/24/2020 17:34:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 08:45

Air Blank	0.000	17:33
Control Test	0.081	17:34
Air Blank	0.000	17:34

---

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:31:15

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/27/2020 11:19

Air Blank	0.000	09:30
Control Test	0.081	09:30
Air Blank	0.000	09:31

---

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:33:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2020 11:19

Air Blank	0.000	09:32
Control Test	0.196	09:33
Air Blank	0.000	09:33

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:36:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2020 11:19

Air Blank	0.000	09:35
Control Test	0.000	09:36
Air Blank	0.000	09:36

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency: HOLLYWOOD PD

Date/Time: 03/11/2020 10:47:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2020 11:59

Air Blank	0.000	10:46
Control Test	0.000	10:47
Air Blank	0.000	10:47

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency: PINELLAS COUNTY SO

Date/Time: 01/30/2020 23:52:03

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/05/2020 17:06

Air Blank	0.000	23:51
Control Test	0.054	23:51
Air Blank	0.000	23:52

---

Agency: PINELLAS COUNTY SO

Date/Time: 01/30/2020 23:54:07

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/05/2020 17:06

Air Blank	0.000	23:53
ITP Check	0.082	23:53
Air Blank	0.000	23:54

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 02/26/2020 20:03:33

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/27/2020 10:20

Air Blank	0.000	20:02
Control Test	0.080	20:03
Air Blank	0.000	20:03

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 01/24/2020 06:12:45

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/27/2020 09:22

Air Blank	0.000	06:11
Control Test	0.075	06:12
Air Blank	0.000	06:12

---

Agency: POLK COUNTY SO

Date/Time: 01/24/2020 06:17:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2020 09:22

Air Blank	0.000	06:16
Control Test	0.079	06:17
Air Blank	0.000	06:17

---

Agency: POLK COUNTY SO

Date/Time: 01/24/2020 11:41:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2020 09:22

Air Blank	0.000	11:40
Control Test	0.200	11:41
Air Blank	0.000	11:41

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 02/24/2020 17:46:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/25/2020 10:11

Air Blank	0.000	17:45
Control Test	0.077	17:45
Air Blank	0.000	17:46

---

Agency: POLK COUNTY SO

Date/Time: 02/24/2020 17:48:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 10:11

Air Blank	0.000	17:47
Control Test	0.046	17:48
Air Blank	0.000	17:48

---

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:51:30

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/27/2020 12:33

Air Blank	0.000	09:50
Control Test	0.075	09:50
Air Blank	0.000	09:51

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:53:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2020 12:33

Air Blank	0.000	09:52
Control Test	0.197	09:52
Air Blank	0.000	09:53

---

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:55:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2020 12:33

Air Blank	0.000	09:54
Control Test	0.000	09:54
Air Blank	0.000	09:55

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 01/29/2020 15:00:54

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/30/2020 13:39

Air Blank	0.000	15:00
Control Test	0.076	15:00
Air Blank	0.000	15:00

---

Agency: PLANTATION PD

Date/Time: 01/29/2020 15:03:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/30/2020 13:39

Air Blank	0.000	15:02
Control Test	0.050	15:03
Air Blank	0.000	15:03

---

Agency: PLANTATION PD

Date/Time: 01/29/2020 15:05:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/30/2020 13:39

Air Blank	0.000	15:04
Control Test	0.080	15:05
Air Blank	0.000	15:05

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 01/29/2020 15:08:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/30/2020 13:39

Air Blank	0.000	15:07
Control Test	0.197	15:07
Air Blank	0.000	15:08

---

Agency: PLANTATION PD

Date/Time: 02/20/2020 19:15:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2020 20:16

Air Blank	0.000	19:13
Control Test	0.000	19:14
Air Blank	0.000	19:15

---

Agency: PLANTATION PD

Date/Time: 02/20/2020 19:17:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2020 20:16

Air Blank	0.000	19:16
Control Test	0.050	19:16
Air Blank	0.000	19:17

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 02/20/2020 19:19:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2020 20:16

Air Blank	0.000	19:18
Control Test	0.079	19:19
Air Blank	0.000	19:19

---

Agency: PLANTATION PD

Date/Time: 02/20/2020 19:22:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2020 20:16

Air Blank	0.000	19:20
Control Test	0.198	19:21
Air Blank	0.000	19:22

---

Agency: PLANTATION PD

Date/Time: 02/20/2020 19:23:46

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/20/2020 20:17

Air Blank	0.000	19:22
Control Test	0.074	19:23
Air Blank	0.000	19:23

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency: PLANTATION PD

Date/Time: 02/20/2020 19:25:57

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/20/2020 20:17

Air Blank	0.000	19:25
Control Test	0.075	19:25
Air Blank	0.000	19:25

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency: JACKSONVILLE SO

Date/Time: 03/12/2020 09:44:16

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/12/2020 11:35

Air Blank	0.000	09:43
Control Test	0.079	09:43
Air Blank	0.000	09:44

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

Agency: CLERMONT P.D.

Date/Time: 03/25/2020 10:46:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2020 13:12

Air Blank	0.000	10:45
Control Test	0.048	10:46
Air Blank	0.000	10:46

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 01/28/2020 15:10:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/28/2020 16:00

Air Blank	0.000	15:09
Control Test	0.047	15:10
Air Blank	0.000	15:10

---

Agency: AUBURNDALE PD

Date/Time: 01/28/2020 15:12:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/28/2020 16:00

Air Blank	0.000	15:11
Control Test	0.077	15:12
Air Blank	0.000	15:12

---

Agency: AUBURNDALE PD

Date/Time: 01/28/2020 15:14:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/28/2020 16:00

Air Blank	0.000	15:13
Control Test	0.195	15:14
Air Blank	0.000	15:14

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 02/26/2020 10:03:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2020 11:05

Air Blank	0.000	10:02
Control Test	0.047	10:02
Air Blank	0.000	10:03

---

Agency: AUBURNDALE PD

Date/Time: 02/26/2020 10:07:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2020 11:05

Air Blank	0.000	10:06
Control Test	0.078	10:07
Air Blank	0.000	10:07

---

Agency: AUBURNDALE PD

Date/Time: 02/26/2020 10:09:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2020 11:05

Air Blank	0.000	10:08
Control Test	0.199	10:09
Air Blank	0.000	10:09

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 02/26/2020 10:15:59

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/26/2020 11:05

Air Blank	0.000	10:15
Control Test	0.080	10:15
Air Blank	0.000	10:15

---

Agency: AUBURNDALE PD

Date/Time: 03/26/2020 07:34:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2020 16:11

Air Blank	0.000	07:33
Control Test	0.047	07:34
Air Blank	0.000	07:34

---

Agency: AUBURNDALE PD

Date/Time: 03/26/2020 07:36:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2020 16:11

Air Blank	0.000	07:35
Control Test	0.077	07:36
Air Blank	0.000	07:36

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 03/26/2020 07:38:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/30/2020 16:11

Air Blank	0.000	07:37
Control Test	0.198	07:38
Air Blank	0.000	07:38

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 01/24/2020 06:04:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2020 09:11

Air Blank	0.000	06:02
Control Test	0.200	06:03
Air Blank	0.000	06:04

---

Agency: POLK COUNTY SO

Date/Time: 01/24/2020 11:38:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2020 09:11

Air Blank	0.000	11:37
Control Test	0.048	11:38
Air Blank	0.000	11:38

---

Agency: POLK COUNTY SO

Date/Time: 02/24/2020 17:41:22

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/25/2020 09:48

Air Blank	0.000	17:40
Control Test	0.079	17:40
Air Blank	0.000	17:41

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 02/24/2020 17:43:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/25/2020 09:48

Air Blank	0.000	17:42
Control Test	0.000	17:43
Air Blank	0.000	17:43

---

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:46:29

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/27/2020 12:05

Air Blank	0.000	09:45
Control Test	0.078	09:45
Air Blank	0.000	09:46

---

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:48:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2020 12:05

Air Blank	0.000	09:47
Control Test	0.049	09:47
Air Blank	0.000	09:48

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 03/25/2020 09:50:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/27/2020 12:05

Air Blank	0.000	09:49
Control Test	0.077	09:50
Air Blank	0.000	09:50

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency: JACKSONVILLE SO

Date/Time: 01/30/2020 10:33:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/30/2020 11:37

Air Blank	0.000	10:31
Control Test	0.186	10:32
Air Blank	0.000	10:33

---

Agency: JACKSONVILLE SO

Date/Time: 03/12/2020 14:53:41

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/25/2020 08:07

Air Blank	0.000	14:52
Control Test	0.079	14:53
Air Blank	0.000	14:53

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

Agency: HIGH SPRINGS PD

Date/Time: 02/28/2020 04:54:08

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/28/2020 05:56

Air Blank	0.000	04:53
Control Test	0.000	04:53
Air Blank	0.000	04:54

---

Agency: HIGH SPRINGS PD

Date/Time: 02/28/2020 04:55:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2020 05:56

Air Blank	0.000	04:54
Control Test	0.000	04:55
Air Blank	0.000	04:55

---

Agency: HIGH SPRINGS PD

Date/Time: 02/28/2020 04:57:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2020 05:56

Air Blank	0.000	04:56
Control Test	0.047	04:57
Air Blank	0.000	04:57

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

Agency: UNION COUNTY SO

Date/Time: 03/04/2020 12:00:24

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/04/2020 12:20

Air Blank	0.000	11:59
ITP Check	0.085	11:59
Air Blank	0.000	12:00

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 01/04/2020 12:56:43

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 01/06/2020 23:35

Air Blank	0.000	12:55
ITP Check	0.080	12:56
Air Blank	0.000	12:56

---

Agency: WALTON COUNTY SO

Date/Time: 01/04/2020 12:58:10

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/06/2020 23:35

Air Blank	0.000	12:57
Control Test	0.080	12:57
Air Blank	0.000	12:58

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

Agency: WALTON COUNTY SO

Date/Time: 01/04/2020 13:10:45

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/07/2020 00:47

Air Blank	0.000	13:09
Control Test	0.080	13:10
Air Blank	0.000	13:10

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: EUSTIS P.D.

Date/Time: 01/21/2020 03:38:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/21/2020 04:56

Air Blank	0.000	03:37
Control Test	0.000	03:38
Air Blank	0.000	03:38

---

Agency: EUSTIS P.D.

Date/Time: 01/21/2020 03:41:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/21/2020 04:56

Air Blank	0.000	03:40
Control Test	0.000	03:40
Air Blank	0.000	03:41

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency: PINELLAS COUNTY SO

Date/Time: 01/31/2020 00:12:01

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/05/2020 16:39

Air Blank	0.000	00:11
Control Test	0.063	00:11
Air Blank	0.000	00:12

---

Agency: PINELLAS COUNTY SO

Date/Time: 01/31/2020 00:13:43

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/05/2020 16:39

Air Blank	0.000	00:12
ITP Check	0.080	00:13
Air Blank	0.000	00:13

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 03/06/2020 13:42:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/06/2020 15:29

Air Blank	0.000	13:41
Control Test	0.196	13:42
Air Blank	0.000	13:42

---

Agency: MARTIN COUNTY SO

Date/Time: 03/06/2020 13:44:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/06/2020 15:29

Air Blank	0.000	13:43
Control Test	0.197	13:44
Air Blank	0.000	13:44

---

Agency: MARTIN COUNTY SO

Date/Time: 03/06/2020 13:47:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/06/2020 15:29

Air Blank	0.000	13:45
Control Test	0.077	13:46
Air Blank	0.000	13:47

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 03/06/2020 13:49:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/06/2020 15:29

Air Blank	0.000	13:48
Control Test	0.047	13:48
Air Blank	0.000	13:49

---

Agency: MARTIN COUNTY SO

Date/Time: 03/06/2020 13:51:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/06/2020 15:29

Air Blank	0.000	13:50
Control Test	0.000	13:50
Air Blank	0.000	13:51

---

Agency: MARTIN COUNTY SO

Date/Time: 03/06/2020 13:54:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/06/2020 15:29

Air Blank	0.000	13:52
Control Test	0.000	13:53
Air Blank	0.000	13:54

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency: CAPE CORAL PD

Date/Time: 01/23/2020 11:42:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2020 12:14

Air Blank	0.000	11:41
Control Test	0.000	11:41
Air Blank	0.000	11:42

---

Agency: CAPE CORAL PD

Date/Time: 01/27/2020 09:24:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2020 12:14

Air Blank	0.000	09:23
Control Test	0.000	09:23
Air Blank	0.000	09:24

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency: ST PETERSBURG PD

Date/Time: 03/09/2020 06:46:19

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/10/2020 15:33

Air Blank	0.000	06:45
Control Test	0.082	06:45
Air Blank	0.000	06:46

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

Agency: MARTIN COUNTY SO

Date/Time: 02/07/2020 15:25:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2020 18:33

Air Blank	0.000	15:24
Control Test	0.193	15:24
Air Blank	0.000	15:25

---

Agency: MARTIN COUNTY SO

Date/Time: 02/07/2020 15:27:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2020 18:33

Air Blank	0.000	15:26
Control Test	0.077	15:27
Air Blank	0.000	15:27

---

Agency: MARTIN COUNTY SO

Date/Time: 02/07/2020 15:30:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2020 18:33

Air Blank	0.000	15:29
Control Test	0.048	15:29
Air Blank	0.000	15:30

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

Agency: MARTIN COUNTY SO

Date/Time: 02/07/2020 15:32:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2020 18:33

Air Blank	0.000	15:31
Control Test	0.000	15:32
Air Blank	0.000	15:32

---

Agency: MARTIN COUNTY SO

Date/Time: 02/07/2020 15:34:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2020 18:33

Air Blank	0.000	15:33
Control Test	0.000	15:33
Air Blank	0.000	15:34

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

Agency: MARTIN COUNTY SO

Date/Time: 01/28/2020 19:52:02

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/07/2020 17:39

Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:52

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 01/11/2020 00:31:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2020 01:40

Air Blank	0.000	00:30
Control Test	0.193	00:30
Air Blank	0.000	00:31

---

Agency: MARTIN COUNTY SO

Date/Time: 01/11/2020 00:33:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2020 01:40

Air Blank	0.000	00:31
Control Test	0.076	00:32
Air Blank	0.000	00:33

---

Agency: MARTIN COUNTY SO

Date/Time: 01/11/2020 00:34:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2020 01:40

Air Blank	0.000	00:33
Control Test	0.045	00:34
Air Blank	0.000	00:34

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 01/11/2020 00:36:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2020 01:40

Air Blank	0.000	00:35
Control Test	0.000	00:36
Air Blank	0.000	00:36

---

Agency: MARTIN COUNTY SO

Date/Time: 01/11/2020 00:37:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2020 01:40

Air Blank		00:37
Air Blank		00:37

---

Agency: MARTIN COUNTY SO

Date/Time: 01/11/2020 00:39:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2020 01:40

Air Blank	0.000	00:38
Control Test	0.000	00:39
Air Blank	0.000	00:39

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

Agency: HOMESTEAD PD

Date/Time: 01/29/2020 13:17:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/29/2020 15:35

Air Blank	0.000	13:16
Control Test	0.047	13:17
Air Blank	0.000	13:17

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

Agency: ESCAMBIA COUNTY SO

Date/Time: 02/14/2020 19:25:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2020 21:30

Air Blank	0.000	19:23
Control Test	0.000	19:24
Air Blank	0.000	19:25

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

Agency: FHP TROOP G

Date/Time: 03/11/2020 17:55:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2020 19:04

Air Blank	0.000	17:54
Control Test	0.075	17:55
Air Blank	0.000	17:55

---

Agency: FHP TROOP G

Date/Time: 03/11/2020 17:58:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2020 19:04

Air Blank	0.000	17:56
Control Test	0.048	17:57
Air Blank	0.000	17:58

---

Agency: FHP TROOP G

Date/Time: 03/11/2020 18:00:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2020 19:04

Air Blank	0.000	17:58
Control Test	0.188	17:59
Air Blank	0.000	18:00

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

Agency: FHP TROOP G

Date/Time: 03/11/2020 18:10:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/11/2020 19:04

Air Blank	0.000	18:08
Control Test	0.190	18:09
Air Blank	0.000	18:10

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

Agency: COLLIER COUNTY SO

Date/Time: 02/27/2020 12:21:19

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/15/2020 13:54

Air Blank	0.000	12:20
Control Test	0.082	12:20
Air Blank	0.000	12:21

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

Agency: MARION COUNTY SO

Date/Time: 01/28/2020 14:06:39

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/28/2020 15:19

Air Blank	0.000	14:05
Control Test	0.080	14:06
Air Blank	0.000	14:06

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 01/06/2020 21:00:57

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/06/2020 23:39

Air Blank	0.000	21:00
Control Test	0.078	21:00
Air Blank	0.000	21:00

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

Agency: TREASURE ISLAND PD

Date/Time: 03/20/2020 11:17:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2020 13:06

Air Blank	0.000	11:16
Control Test	0.052	11:17
Air Blank	0.000	11:17

---

Agency: TREASURE ISLAND PD

Date/Time: 03/20/2020 11:21:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2020 13:06

Air Blank	0.000	11:20
Control Test	0.051	11:20
Air Blank	0.000	11:21

---

Agency: TREASURE ISLAND PD

Date/Time: 03/20/2020 11:25:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2020 13:06

Air Blank	0.000	11:24
Control Test	0.078	11:25
Air Blank	0.000	11:25

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

Agency: TREASURE ISLAND PD

Date/Time: 03/20/2020 11:28:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2020 13:06

Air Blank	0.000	11:27
Control Test	0.105	11:28
Air Blank	0.000	11:28

---

Agency: TREASURE ISLAND PD

Date/Time: 03/20/2020 11:31:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2020 13:06

Air Blank	0.000	11:30
Control Test	0.107	11:30
Air Blank	0.000	11:31

---

Agency: TREASURE ISLAND PD

Date/Time: 03/20/2020 11:52:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/20/2020 13:06

Air Blank	0.000	11:51
Control Test	0.193	11:52
Air Blank	0.000	11:52

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Control Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

Agency: WALTON COUNTY SO

Date/Time: 03/06/2020 13:36:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/12/2020 06:56

Air Blank	0.000	13:35
Control Test	0.078	13:35
Air Blank	0.000	13:36

---

Agency: WALTON COUNTY SO

Date/Time: 03/06/2020 13:38:01

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 03/12/2020 06:56

Air Blank	0.000	13:37
ITP Check	0.080	13:37
Air Blank	0.000	13:38

---