

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

Agency: FDLE

Date/Time: 01/10/2020 20:16:01

Data Download Date/Time: 01/13/2020 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

Agency: FDLE

Date/Time: 02/14/2020 11:28:17

Data Download Date/Time: 03/04/2020 12:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 03/31/2020 12:19:57

Data Download Date/Time: 03/31/2020 14:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

Agency: FHP

Date/Time: 01/02/2020 04:42:53

Data Download Date/Time: 02/14/2020 19:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

Agency: HAMILTON COUNTY SO

Date/Time: 02/27/2020 16:29:37

Data Download Date/Time: 02/27/2020 18:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

Agency: DIXIE COUNTY SO

Date/Time: 03/31/2020 10:25:44

Data Download Date/Time: 03/31/2020 12:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: DIXIE COUNTY SO

Date/Time: 03/31/2020 11:35:01

Data Download Date/Time: 03/31/2020 12:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

Agency: MARION COUNTY SO

Date/Time: 02/24/2020 14:41:19

Data Download Date/Time: 02/25/2020 14:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency: BLOUNTSTOWN PD

Date/Time: 02/12/2020 01:40:56

Data Download Date/Time: 02/12/2020 03:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency: MIAMI PD

Date/Time: 02/19/2020 21:41:05

Data Download Date/Time: 02/28/2020 13:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

Agency: HOMESTEAD PD

Date/Time: 03/13/2020 15:51:54

Data Download Date/Time: 03/31/2020 17:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

Agency: FISH & WILDLIFE CC

Date/Time: 02/05/2020 14:28:09

Data Download Date/Time: 02/05/2020 16:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FISH & WILDLIFE CC

Date/Time: 02/05/2020 15:23:03

Data Download Date/Time: 02/05/2020 16:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 01/16/2020 20:05:56

Data Download Date/Time: 01/16/2020 21:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST JOE PD

Date/Time: 02/09/2020 20:10:43

Data Download Date/Time: 02/09/2020 20:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 03/17/2020 15:50:52

Data Download Date/Time: 03/17/2020 16:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

Agency: LEE COUNTY SO

Date/Time: 03/18/2020 07:50:21

Data Download Date/Time: 04/08/2020 10:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEE COUNTY SO

Date/Time: 03/24/2020 15:48:35

Data Download Date/Time: 04/08/2020 10:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

Agency: COLLIER COUNTY SO

Date/Time: 01/15/2020 13:37:53

Data Download Date/Time: 01/20/2020 16:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COLLIER COUNTY SO

Date/Time: 02/14/2020 13:57:18

Data Download Date/Time: 02/24/2020 10:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

Agency: OSCEOLA COUNTY SO

Date/Time: 03/30/2020 13:51:39

Data Download Date/Time: 03/31/2020 13:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 01/29/2020 18:16:13

Data Download Date/Time: 02/13/2020 16:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE P.D.

Date/Time: 02/26/2020 20:54:31

Data Download Date/Time: 02/26/2020 22:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 03/08/2020 16:07:59

Data Download Date/Time: 03/08/2020 17:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency: SEMINOLE COUNTY S.O.

Date/Time: 01/15/2020 11:16:26

Data Download Date/Time: 01/24/2020 16:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

Agency: PINELLAS COUNTY SO

Date/Time: 01/31/2020 00:45:52

Data Download Date/Time: 02/05/2020 16:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

Agency: INDIALANTIC P.D.

Date/Time: 02/12/2020 07:03:40

Data Download Date/Time: 02/25/2020 10:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: INDIALANTIC P.D.

Date/Time: 03/17/2020 08:05:06

Data Download Date/Time: 03/17/2020 16:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

Agency: COCONUT CREEK PD

Date/Time: 03/25/2020 00:02:56

Data Download Date/Time: 03/25/2020 15:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 03/21/2020 15:00:56

Data Download Date/Time: 03/21/2020 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency: PINELLAS COUNTY SO

Date/Time: 01/30/2020 23:49:48

Data Download Date/Time: 02/05/2020 17:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

Agency: FL HIGHWAY PATROL

Date/Time: 01/10/2020 12:06:31

Data Download Date/Time: 01/12/2020 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 01/10/2020 12:23:25

Data Download Date/Time: 01/12/2020 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

Agency: FHP TROOP E MIAMI

Date/Time: 01/23/2020 07:44:19

Data Download Date/Time: 01/27/2020 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP E MIAMI

Date/Time: 01/23/2020 07:46:00

Data Download Date/Time: 01/27/2020 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

Agency: FHP TROOP E MIAMI

Date/Time: 01/23/2020 07:48:54

Data Download Date/Time: 01/27/2020 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP E MIAMI

Date/Time: 01/23/2020 07:56:25

Data Download Date/Time: 01/27/2020 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency: FHP TROOP E MIAMI

Date/Time: 01/23/2020 07:37:54

Data Download Date/Time: 01/27/2020 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP E MIAMI

Date/Time: 01/23/2020 23:27:35

Data Download Date/Time: 01/27/2020 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

Agency: ST JOHNS COUNTY SO

Date/Time: 01/29/2020 14:34:30

Data Download Date/Time: 01/30/2020 08:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST JOHNS COUNTY SO

Date/Time: 01/29/2020 14:36:02

Data Download Date/Time: 01/30/2020 08:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

Agency: ST JOHNS COUNTY SO

Date/Time: 01/29/2020 14:39:16

Data Download Date/Time: 01/30/2020 08:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST JOHNS COUNTY SO

Date/Time: 01/29/2020 14:39:46

Data Download Date/Time: 01/30/2020 08:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

Agency: ST JOHNS COUNTY SO

Date/Time: 02/25/2020 10:46:01

Data Download Date/Time: 02/25/2020 11:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency: FLAGLER COUNTY SO

Date/Time: 03/30/2020 09:15:45

Data Download Date/Time: 03/30/2020 11:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

Agency: BROWARD COUNTY SO

Date/Time: 02/09/2020 23:20:06

Data Download Date/Time: 02/20/2020 14:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 01/15/2020 14:50:13

Data Download Date/Time: 01/21/2020 11:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

Agency: BREVARD COUNTY SO

Date/Time: 01/14/2020 08:39:15

Data Download Date/Time: 01/20/2020 11:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency: LEE COUNTY SO

Date/Time: 01/14/2020 15:40:07

Data Download Date/Time: 01/21/2020 10:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 01/28/2020 15:08:32

Data Download Date/Time: 01/28/2020 16:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: AUBURNDALE PD

Date/Time: 01/28/2020 15:53:49

Data Download Date/Time: 02/26/2020 11:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency: DAYTONA BEACH SHORES

Date/Time: 03/24/2020 23:23:22

Data Download Date/Time: 03/25/2020 06:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency: ROCKLEDGE P.D.

Date/Time: 02/03/2020 11:27:16

Data Download Date/Time: 02/03/2020 12:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency: SEMINOLE COUNTY S.O.

Date/Time: 01/14/2020 10:25:31

Data Download Date/Time: 01/20/2020 16:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

Agency: UNION COUNTY SO

Date/Time: 02/13/2020 09:34:30

Data Download Date/Time: 03/04/2020 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: UNION COUNTY SO

Date/Time: 02/13/2020 09:35:59

Data Download Date/Time: 03/04/2020 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

Agency: UNION COUNTY SO

Date/Time: 02/13/2020 09:37:27

Data Download Date/Time: 03/04/2020 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: UNION COUNTY SO

Date/Time: 02/13/2020 09:38:49

Data Download Date/Time: 03/04/2020 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

Agency: UNION COUNTY SO

Date/Time: 02/13/2020 09:42:08

Data Download Date/Time: 03/04/2020 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: UNION COUNTY SO

Date/Time: 02/13/2020 09:47:51

Data Download Date/Time: 03/04/2020 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

Agency: UNION COUNTY SO

Date/Time: 02/13/2020 10:35:48

Data Download Date/Time: 03/04/2020 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

Agency: UNION COUNTY SO

Date/Time: 02/13/2020 14:00:33

Data Download Date/Time: 03/04/2020 12:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 01/04/2020 12:37:58

Data Download Date/Time: 01/06/2020 23:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

Agency: WALTON COUNTY SO

Date/Time: 01/04/2020 13:09:17

Data Download Date/Time: 01/07/2020 00:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 01/27/2020 05:04:09

Data Download Date/Time: 02/27/2020 06:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

Agency: SARASOTA COUNTY SO

Date/Time: 03/26/2020 08:38:10

Data Download Date/Time: 03/26/2020 09:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency: LEVY COUNTY SO

Date/Time: 03/24/2020 11:03:31

Data Download Date/Time: 04/03/2020 13:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEVY COUNTY SO

Date/Time: 03/24/2020 11:04:34

Data Download Date/Time: 04/03/2020 13:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency: LEVY COUNTY SO

Date/Time: 03/24/2020 11:06:39

Data Download Date/Time: 04/03/2020 13:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency: PINELLAS COUNTY SO

Date/Time: 01/31/2020 00:10:11

Data Download Date/Time: 02/05/2020 16:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency: NAS MAYPORT

Date/Time: 03/09/2020 10:41:26

Data Download Date/Time: 03/30/2020 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: NAS MAYPORT

Date/Time: 03/10/2020 10:53:23

Data Download Date/Time: 03/30/2020 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency: CAPE CORAL PD

Date/Time: 01/23/2020 09:20:59

Data Download Date/Time: 01/27/2020 12:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency: OSCEOLA COUNTY S.O.

Date/Time: 03/30/2020 13:52:18

Data Download Date/Time: 03/31/2020 13:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency: MANATEE COUNTY SO

Date/Time: 01/14/2020 14:52:24

Data Download Date/Time: 01/16/2020 07:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MANATEE COUNTY SO

Date/Time: 01/15/2020 15:41:26

Data Download Date/Time: 01/16/2020 07:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency: MANATEE COUNTY SO

Date/Time: 02/14/2020 15:54:22

Data Download Date/Time: 02/15/2020 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency: HIALEAH PD

Date/Time: 03/26/2020 12:33:45

Data Download Date/Time: 03/26/2020 13:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

Agency: HIALEAH PD

Date/Time: 03/26/2020 11:25:07

Data Download Date/Time: 03/26/2020 13:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIALEAH PD

Date/Time: 03/26/2020 12:31:07

Data Download Date/Time: 03/26/2020 13:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency: FL HIGHWAY PATROL

Date/Time: 01/10/2020 12:05:52

Data Download Date/Time: 01/12/2020 16:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 01/10/2020 12:23:23

Data Download Date/Time: 01/12/2020 16:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency: PORT ORANGE PD

Date/Time: 02/27/2020 04:22:22

Data Download Date/Time: 02/27/2020 05:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

Agency: MANATEE COUNTY SO

Date/Time: 02/26/2020 13:37:30

Data Download Date/Time: 02/28/2020 09:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MANATEE COUNTY SO

Date/Time: 02/27/2020 10:56:44

Data Download Date/Time: 02/28/2020 09:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 03/21/2020 14:03:20

Data Download Date/Time: 03/21/2020 14:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 03/27/2020 21:24:56

Data Download Date/Time: 03/30/2020 11:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

Agency: FL HIGHWAY PATROL

Date/Time: 03/15/2020 01:06:42

Data Download Date/Time: 03/21/2020 13:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 03/15/2020 01:07:41

Data Download Date/Time: 03/21/2020 13:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

Agency: MANATEE COUNTY SO

Date/Time: 01/14/2020 14:52:45

Data Download Date/Time: 01/21/2020 07:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

Agency: MANATEE COUNTY SO

Date/Time: 03/24/2020 14:56:29

Data Download Date/Time: 03/25/2020 10:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

Agency: HOLMES BEACH PD

Date/Time: 01/13/2020 08:32:09

Data Download Date/Time: 01/15/2020 11:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HOLMES BEACH PD

Date/Time: 01/13/2020 08:35:00

Data Download Date/Time: 01/15/2020 11:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

Agency: MANATEE COUNTY SO

Date/Time: 01/15/2020 15:42:57

Data Download Date/Time: 01/27/2020 10:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MANATEE COUNTY SO

Date/Time: 01/22/2020 13:47:27

Data Download Date/Time: 01/27/2020 10:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

Agency: MANATEE COUNTY SO

Date/Time: 01/23/2020 15:32:24

Data Download Date/Time: 01/27/2020 10:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MANATEE COUNTY SO

Date/Time: 02/14/2020 16:56:42

Data Download Date/Time: 02/15/2020 08:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 12:47:41

Data Download Date/Time: 03/27/2020 09:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 12:50:07

Data Download Date/Time: 03/27/2020 09:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 12:51:20

Data Download Date/Time: 03/27/2020 09:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 13:01:01

Data Download Date/Time: 03/27/2020 09:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 13:01:52

Data Download Date/Time: 03/27/2020 09:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 13:03:23

Data Download Date/Time: 03/27/2020 09:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 13:04:52

Data Download Date/Time: 03/27/2020 09:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 13:07:06

Data Download Date/Time: 03/27/2020 09:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

Agency: FL HIGHWAY PATROL

Date/Time: 02/28/2020 13:08:31

Data Download Date/Time: 03/27/2020 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

Agency: FHP

Date/Time: 03/14/2020 00:54:38

Data Download Date/Time: 03/21/2020 13:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

Agency: COLLIER COUNTY SO

Date/Time: 02/06/2020 12:49:48

Data Download Date/Time: 03/15/2020 13:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

Agency: BROWARD COUNTY S.O.

Date/Time: 03/08/2020 04:12:11

Data Download Date/Time: 03/11/2020 12:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

Agency: PENSACOLA P.D.

Date/Time: 01/11/2020 18:59:31

Data Download Date/Time: 01/11/2020 20:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

Agency: MARION COUNTY SO

Date/Time: 01/28/2020 13:34:58

Data Download Date/Time: 01/28/2020 15:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

Agency: MARION COUNTY SO

Date/Time: 01/28/2020 13:35:40

Data Download Date/Time: 01/28/2020 15:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 02/03/2020 12:39:27

Data Download Date/Time: 02/05/2020 14:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

Agency: MARION COUNTY SO

Date/Time: 02/24/2020 14:38:48

Data Download Date/Time: 02/25/2020 14:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

Agency: WALTON COUNTY SO

Date/Time: 03/02/2020 10:31:32

Data Download Date/Time: 03/12/2020 06:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 03/02/2020 10:35:44

Data Download Date/Time: 03/12/2020 06:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

Agency: WALTON COUNTY SO

Date/Time: 03/06/2020 13:46:56

Data Download Date/Time: 03/12/2020 06:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK
