

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

<b>Agency</b> NASSAU COUNTY SO	<b>Date</b> 02/04/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PLATT,SHAYLA,D	<b>Time</b> 11:43:51	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 02/04/2020 14:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO                      **Date** 02/04/2020                      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 13:58:50                      **Software** 8100.27

**0.05 Lot #** 201905A                      **0.08 Lot #** 201905B                      **0.20 Lot #** 201904D                      **0.08 Gas Lot #** AG916501  
**Expiration** 05/14/2021                      **Expiration** 05/14/2021                      **Expiration** 04/30/2021                      **Expiration** 06/14/2021

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.047
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.047
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.047
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

---

**Data Download Date/Time** 02/04/2020 14:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

<b>Agency</b> NASSAU COUNTY SO	<b>Date</b> 02/24/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGGINBOTHAM,BRIAN,K	<b>Time</b> 11:35:25	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 30219080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 01/05/2022

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.043
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.045
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.043
		0.05 g/210L Test # (Repeat) 3	0.045
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 02/24/2020 12:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

<b>Agency</b> NASSAU COUNTY SO	<b>Date</b> 03/21/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> HIGGINBOTHAM, BRIAN, K	<b>Time</b> 13:13:52	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 30219080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 01/05/2022

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/21/2020 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

<b>Agency</b> UNIVERSITY OF FL PD	<b>Date</b> 01/31/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PETERSON, KYLE, A	<b>Time</b> 21:14:24	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 22419080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/31/2020 21:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

<b>Agency</b> UNIVERSITY OF FL PD	<b>Date</b> 02/29/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PETERSON, KYLE, A	<b>Time</b> 03:56:29	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 22419080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # (Repeat) 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.208
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det: Ambient Fail.

**Data Download Date/Time** 03/16/2020 02:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

<b>Agency</b> UNIVERSITY OF FL PD	<b>Date</b> 03/30/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PETERSON,KYLE,A	<b>Time</b> 00:01:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 22419080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # (Repeat) 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det: RFI Detect.

**Data Download Date/Time** 03/30/2020 00:09



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

<b>Agency</b> HIGH SPRINGS PD	<b>Date</b> 01/08/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> YAKUBSIN,JAMES,R	<b>Time</b> 06:32:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 17618080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

05: Improper Sample SIMULATOR ISSUE.

**Data Download Date/Time** 01/08/2020 06:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

<b>Agency</b> HIGH SPRINGS PD	<b>Date</b> 02/28/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> YAKUBSIN,JAMES,R	<b>Time</b> 05:34:48	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 17618080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/28/2020 05:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

<b>Agency</b> GILCHRIST COUNTY SO	<b>Date</b> 01/05/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD, BRIAN, A	<b>Time</b> 22:08:15	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1048227
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 11/15/2021

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/09/2020 09:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

<b>Agency</b> GILCHRIST COUNTY SO	<b>Date</b> 01/31/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD, BRIAN, A	<b>Time</b> 11:06:48	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1048227
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 11/15/2021

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/10/2020 06:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

<b>Agency</b> GILCHRIST COUNTY SO	<b>Date</b> 03/06/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LANGFORD, BRIAN, A	<b>Time</b> 10:11:20	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1048227
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 11/15/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/10/2020 12:31

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

<b>Agency</b> BAKER COUNTY SO	<b>Date</b> 01/19/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> CALEB COLLINS,DEREK WATSON, /	<b>Time</b> 09:34:43	<b>Software</b> 8100.27	

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24818080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/19/2020 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

<b>Agency</b> BAKER COUNTY SO	<b>Date</b> 02/28/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> COLLINS, CALEB,WATSON, DEREK,/	<b>Time</b> 15:33:07	<b>Software</b> 8100.27

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24818080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det:NO INT DET.

**Data Download Date/Time** 02/28/2020 16:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

<b>Agency</b> BAKER COUNTY SO	<b>Date</b> 03/22/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> COLLINS CALEB, WATSON DEREK, /	<b>Time</b> 16:57:02	<b>Software</b> 8100.27	

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24818080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/22/2020 17:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

<b>Agency</b> UNION COUNTY SO	<b>Date</b> 01/30/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MCDANIE, MARCUS, G	<b>Time</b> 10:50:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 823252
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/07/2020

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/30/2020 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

<b>Agency</b> UNION COUNTY SO	<b>Date</b> 02/28/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> MCDANIEL,MARCUS,G	<b>Time</b> 10:48:10	<b>Software</b> 8100.27	

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b> 11/13/2020	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 03/04/2020 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

<b>Agency</b> UNION COUNTY SO	<b>Date</b> 03/04/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MCDANIEL,MARCUS,G	<b>Time</b> 11:40:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 823252
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/07/2020

**Number of Simulators** 3

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 03/04/2020 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

**Agency** UNION COUNTY SO                      **Date** 01/30/2020                      **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G                  **Time** 10:47:05                              **Software** 8100.27

**0.05 Lot #** 201811C                      **0.08 Lot #** 201808E                      **0.20 Lot #** 201902C                      **0.08 Gas Lot #** 823252  
**Expiration** 11/13/2020                  **Expiration** 08/22/2020                  **Expiration** 02/20/2021                  **Expiration** 08/07/2020

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/30/2020 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

<b>Agency</b> UNION COUNTY SO	<b>Date</b> 02/28/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> MCDANIEL,MARCUS,G	<b>Time</b> 10:44:42	<b>Software</b> 8100.27	

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b> 11/13/2020	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 4

**Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 03/04/2020 12:07

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

**Agency** UNION COUNTY SO                      **Date** 03/04/2020                      **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G              **Time** 11:47:13                      **Software** 8100.27  
  
**0.05 Lot #** 201811C                      **0.08 Lot #** 201808E                      **0.20 Lot #** 201902C                      **0.08 Gas Lot #** 823252  
**Expiration** 11/13/2020              **Expiration** 08/22/2020              **Expiration** 02/20/2021              **Expiration** 08/07/2020  
  
**Number of Simulators** 3    **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance.
--------------------------------

Data Download Date/Time 03/04/2020 12:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

**Agency** BRADFORD COUNTY SO      **Date** 01/18/2020      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 14:33:45      **Software** 8100.27

**0.05 Lot #** 201811C      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** AG805701  
**Expiration** 11/13/2020      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 02/26/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.194
		0.20 g/210L Test # (Repeat) 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

20: Control Outside Tolerance-LOOSE SIM HOSE.

**Data Download Date/Time** 01/19/2020 10:38

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

<b>Agency</b> BRADFORD COUNTY SO	<b>Date</b> 02/01/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LEVY, JAYSON, I	<b>Time</b> 19:53:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG805701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/26/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/05/2020 14:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

<b>Agency</b> BRADFORD COUNTY SO	<b>Date</b> 03/31/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> LEVY, JAYSON, I	<b>Time</b> 11:48:14	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG000606
<b>Expiration</b> 11/13/2000	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 01/06/2022
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/31/2020 13:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

<b>Agency</b> BRADFORD COUNTY SO	<b>Date</b> 01/02/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LEVY, JAYSON, I	<b>Time</b> 21:22:36	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG805701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/26/2020

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/08/2020 07:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

**Agency** BRADFORD COUNTY SO      **Date** 02/15/2020      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 15:06:44      **Software** 8100.27

**0.05 Lot #** 201811C      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** AG805701  
**Expiration** 11/13/2020      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 02/26/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/16/2020 15:43

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

<b>Agency</b> LEVY COUNTY SO	<b>Date</b> 01/18/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> LEVY, JAYSON, I	<b>Time</b> 16:59:08	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG805701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/26/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/18/2020 17:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

<b>Agency</b> LEVY COUNTY SO	<b>Date</b> 02/15/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> LEVY, JAYSON, I	<b>Time</b> 18:09:39	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG805701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/26/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/15/2020 18:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

<b>Agency</b> LEVY COUNTY SO	<b>Date</b> 03/28/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LEVY, JAYSON, I	<b>Time</b> 09:25:31	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG000606
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 01/06/2022

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

00: Ambient Fail CLEARED / AIRED OUT ROOM.

**Data Download Date/Time** 03/28/2020 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

<b>Agency</b> SANTA ROSA COUNTY SO	<b>Date</b> 01/30/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> NELSON,DURON,A	<b>Time</b> 07:19:13	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1206004
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 09/20/2022
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/30/2020 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

<b>Agency</b> SANTA ROSA COUNTY SO	<b>Date</b> 02/28/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> NELSON,DURON,A	<b>Time</b> 18:38:31	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1206004
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 09/20/2022
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/28/2020 19:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

<b>Agency</b> SANTA ROSA COUNTY SO	<b>Date</b> 03/30/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> NELSON,DURON,A	<b>Time</b> 17:17:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1206004
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 09/20/2022

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/02/2020 17:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

<b>Agency</b> SANTA ROSA COUNTY SO	<b>Date</b> 01/30/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> NELSON,DURON,A	<b>Time</b> 07:16:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1206004
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 09/20/2022

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/30/2020 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

<b>Agency</b> SANTA ROSA COUNTY SO	<b>Date</b> 02/28/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> NELSON,DURON,A	<b>Time</b> 18:34:10	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1206004
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 09/20/2022
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/28/2020 19:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

<b>Agency</b> SANTA ROSA COUNTY SO	<b>Date</b> 03/30/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> NELSON,DURON,A	<b>Time</b> 17:30:45	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1206004
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 09/20/2022

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/02/2020 17:32

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

<b>Agency</b> MARIANNA PD	<b>Date</b> 01/02/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> BARGE,SUSAN,M	<b>Time</b> 00:09:17	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201311C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 970670
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/07/2021

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NA

**Data Download Date/Time** 01/02/2020 01:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

<b>Agency</b> MARIANNA PD	<b>Date</b> 02/14/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> BARGE,SUSAN,M	<b>Time</b> 18:29:46	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 970670
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/07/2021

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.189
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.190
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

20: Control Outside Tolerance.NOT SEEDED IN CAL INLET

**Data Download Date/Time** 02/14/2020 19:45

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD                                   **Date** 02/26/2020                                   **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                           **Time** 11:32:36                                   **Software** 8100.27

**0.05 Lot #** 201905A                   **0.08 Lot #** 201905B                   **0.20 Lot #** 201904D                   **0.08 Gas Lot #** AG931603  
**Expiration** 05/14/2021                   **Expiration** 05/14/2021                   **Expiration** 04/30/2021                   **Expiration** 11/12/2021

**Number of Simulators** 5                                   **Average Standard Deviation** 0.0003                                   **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

---

**Data Download Date/Time** 02/26/2020 12:41



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

<b>Agency</b> MARIANNA PD	<b>Date</b> 03/05/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> BARGE,SUSAN,M	<b>Time</b> 14:28:19	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 970670
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/07/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NA

**Data Download Date/Time** 03/05/2020 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD                      **Date** 01/16/2020                      **Type** Agency Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 13:40:24                      **Software** 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** \_\_\_\_\_ **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 01/22/2020 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD                      **Date** 01/22/2020                      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 10:44:02                      **Software** 8100.27

**0.05 Lot #** 201905A                      **0.08 Lot #** 201905B                      **0.20 Lot #** 201904D                      **0.08 Gas Lot #** AG916501  
**Expiration** 05/14/2021                      **Expiration** 05/14/2021                      **Expiration** 04/30/2021                      **Expiration** 06/14/2021

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

---

**Data Download Date/Time** 01/22/2020 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

<b>Agency</b> GULF BREEZE PD	<b>Date</b> 02/10/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> ROBERSON, THOMAS, C	<b>Time</b> 02:52:34	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1094494
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/19/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/10/2020 05:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

<b>Agency</b> GULF BREEZE PD	<b>Date</b> 03/03/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> ROBERSON, THOMAS, C	<b>Time</b> 01:28:49	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1094494
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/19/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 03/17/2020 06:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

<b>Agency</b> MADISON COUNTY SO	<b>Date</b> 01/06/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> HENING, FRED, C	<b>Time</b> 12:08:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 00919080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/06/2020 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

<b>Agency</b> MADISON COUNTY SO	<b>Date</b> 02/13/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> HENING, FRED, C	<b>Time</b> 11:02:04	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 00919080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2021

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.000
		0.08 g/210L Test # (Repeat) 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance SIMULATOR WAS NOT SECURED .

**Data Download Date/Time** 02/13/2020 11:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

<b>Agency</b> MADISON COUNTY SO	<b>Date</b> 03/10/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> HENING, FRED, C	<b>Time</b> 12:01:41	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 00919080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 03/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/10/2020 12:09

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 01/30/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLER,BRIAN,D	<b>Time</b> 12:32:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG816701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/16/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/30/2020 14:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 02/27/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> MILLER, BRIAN, D	<b>Time</b> 09:27:58	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG716701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/16/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2020 10:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 03/24/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> MILLER, BRIAN, D	<b>Time</b> 13:28:25	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG816701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/16/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/24/2020 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 01/30/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> MILLER,BRIAN,D	<b>Time</b> 13:19:15	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG923908
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/30/2020 14:55

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 02/27/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLER,BRIAN,D	<b>Time</b> 09:32:03	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG923908
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/27/2021

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2020 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

<b>Agency</b> BAY COUNTY SO	<b>Date</b> 03/24/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLER, BRIAN, D	<b>Time</b> 14:10:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG923908
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/27/2021

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/24/2020 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

<b>Agency</b> NICEVILLE PD	<b>Date</b> 01/10/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> CARRICO,KELLY,D	<b>Time</b> 10:06:46	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 869968
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 12/06/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/10/2020 11:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

<b>Agency</b> NICEVILLE PD	<b>Date</b> 02/21/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> CARRICO,KELLY,D	<b>Time</b> 15:54:16	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 869968
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 12/06/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/21/2020 17:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

<b>Agency</b> NICEVILLE PD	<b>Date</b> 03/25/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> CARRICO,KELLY,D	<b>Time</b> 10:00:56	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 869968
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 12/06/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/25/2020 11:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

<b>Agency</b> WALTON COUNTY SO	<b>Date</b> 01/06/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> GAYDOS, JONATHAN, A	<b>Time</b> 22:12:21	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 980558
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/09/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/06/2020 23:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO                      **Date** 01/22/2020                      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 11:07:57                      **Software** 8100.27

**0.05 Lot #** 201905A                      **0.08 Lot #** 201905B                      **0.20 Lot #** 201904D                      **0.08 Gas Lot #** AG916501  
**Expiration** 05/14/2021                      **Expiration** 05/14/2021                      **Expiration** 04/30/2021                      **Expiration** 06/14/2021

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 01/22/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

<b>Agency</b> WALTON COUNTY SO	<b>Date</b> 02/18/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> GAYDOS, JONATHAN, A	<b>Time</b> 22:56:55	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 980558
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/09/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2020 06:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

<b>Agency</b> WALTON COUNTY SO	<b>Date</b> 03/12/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GAYDOS, JONATHAN, A	<b>Time</b> 02:32:41	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C
<b>0.08 Gas Lot #</b> 980558		
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021
		<b>Expiration</b> 07/09/2021

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/12/2020 03:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

<b>Agency</b> WALTON COUNTY SO	<b>Date</b> 01/06/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> GAYDOS, JONATHAN, A	<b>Time</b> 22:00:41	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 181692
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/07/2020 00:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

<b>Agency</b> WALTON COUNTY SO	<b>Date</b> 02/19/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GAYDOS, JONATHAN, A	<b>Time</b> 02:30:03	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 181692
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2021

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2020 06:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

<b>Agency</b> WALTON COUNTY SO	<b>Date</b> 03/12/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GAYDOS, JONATHAN, A	<b>Time</b> 02:23:37	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 181692
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 03/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/12/2020 03:36

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

<b>Agency</b> WALTON COUNTY SO	<b>Date</b> 03/12/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GAYDOS, JONATHAN, A	<b>Time</b> 05:34:52	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 181692
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 03/05/2021

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/12/2020 06:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

<b>Agency</b> OKEECHOBEE CO SO	<b>Date</b> 01/08/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GARLAND, MARK, W	<b>Time</b> 10:30:46	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201802G
<b>0.08 Gas Lot #</b> 1076482		
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/22/2020
		<b>Expiration</b> 01/15/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/08/2020 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

<b>Agency</b> OKEECHOBEE CO SO	<b>Date</b> 02/04/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GARLAND, MARK, W	<b>Time</b> 11:42:10	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201802G
<b>0.08 Gas Lot #</b> 1076482		
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/22/2020
		<b>Expiration</b> 01/15/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/06/2020 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

<b>Agency</b> OKEECHOBEE CO SO	<b>Date</b> 03/05/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> GARLAND, MARK, W	<b>Time</b> 09:29:26	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1076482
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 01/15/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/05/2020 09:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

<b>Agency</b> OKEECHOBEE COUNTY SO	<b>Date</b> 01/08/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GARLAND, MARK, W	<b>Time</b> 10:15:10	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1076482
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/15/2022

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/08/2020 11:33

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

<b>Agency</b> OKEECHOBEE COUNTY SO	<b>Date</b> 02/04/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GARLAND, MARK, W	<b>Time</b> 11:32:39	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1076482
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/15/2022

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/06/2020 17:39

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

<b>Agency</b> OKEECHOBEE COUNTY SO	<b>Date</b> 03/05/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> GARLAND, MARK, W	<b>Time</b> 09:24:14	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C
<b>0.08 Gas Lot #</b> 1076482		
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021
		<b>Expiration</b> 01/15/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/05/2020 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

<b>Agency</b> OKEECHOBEE P.D.	<b>Date</b> 01/03/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> MARRERO,RAUL,E	<b>Time</b> 14:02:17	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 840875
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 09/19/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/07/2020 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 01/31/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> HARRIS, IAN,	<b>Time</b> 19:04:03	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 03519080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 04/05/2021
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

00: RFI Detect.CELL PX NEAR INTOX, MOVED AND RETESTED

**Data Download Date/Time** 01/31/2020 19:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 02/27/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> HARRIS, IAN,	<b>Time</b> 15:39:49	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 201910D	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 03519080A4
<b>Expiration</b> 10/22/2021	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 04/05/2021

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.000
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

20: Control Outside Tolerance, SIMULATOR NOT PLUGGED IN.

**Data Download Date/Time** 02/27/2020 16:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

<b>Agency</b> PORT ST LUCIE PD	<b>Date</b> 03/31/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> HARRIS, IAN,	<b>Time</b> 15:51:07	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 201910D	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 03519080A4
<b>Expiration</b> 10/22/2021	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 04/05/2021

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: RFI Detect.

**Data Download Date/Time** 03/31/2020 16:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO      **Date** 01/03/2020      **Type** Agency Inspection  
**Inspector** KANTORSKI, MARK, A      **Time** 05:17:48      **Software** 8100.27

**0.05 Lot #** 201811C      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 1167545  
**Expiration** 11/13/2020      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 07/02/2022

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/03/2020 05:44

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

<b>Agency</b> INDIAN RIVER CO. SO	<b>Date</b> 02/07/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> KANTORSKI, MARK, A	<b>Time</b> 08:30:49	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1167545
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 07/02/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/07/2020 08:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

<b>Agency</b> INDIAN RIVER CO. SO	<b>Date</b> 03/06/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> KANTORSKI, MARK, A	<b>Time</b> 04:03:00	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1167545
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 07/02/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance DRY GAS CONNECTOR NOT ATTACHD

**Data Download Date/Time** 03/06/2020 04:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

<b>Agency</b> INDIAN RIVER CO. SO	<b>Date</b> 01/03/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> KANTORSKI, MARK, A	<b>Time</b> 05:10:37	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1167545
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 07/02/2022

  

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/03/2020 05:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

<b>Agency</b> INDIAN RIVER CO. SO	<b>Date</b> 02/07/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> KANTORSKI, MARK, A	<b>Time</b> 08:21:35	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C
<b>0.08 Gas Lot #</b> 1167545		
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021
		<b>Expiration</b> 07/02/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/07/2020 08:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

<b>Agency</b> INDIAN RIVER CO. SO	<b>Date</b> 03/06/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> KANTORSKI, MARK, A	<b>Time</b> 03:54:16	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1167545
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 07/02/2022

  

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/06/2020 04:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

<b>Agency</b> DESOTO COUNTY SO	<b>Date</b> 01/30/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PITTS,JOSHUA,E	<b>Time</b> 09:05:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24119080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 11/05/2021

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: MOUTH ALCOHOL DID NOT SHOW SLOPE NOT MET .00: Ambient Fail. MOVED INTOX 8000 OVER ON COUNTER FIXED AF

**Data Download Date/Time** 02/03/2020 09:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

<b>Agency</b> DESOTO COUNTY SO	<b>Date</b> 02/22/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PITTS,JOSHUA,E	<b>Time</b> 20:48:49	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: Range Exceeded DIDNT BLOW PROPERLY . Non-compliance: AI WAS NOT COMPLETED WILL DUE AGAIN  
.

**Data Download Date/Time** 02/26/2020 00:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

Agency DESOTO COUNTY SO	Date 02/22/2020	Type Agency Inspection	
Inspector PITTS,JOSHUA,E	Time 20:56:15	Software 8100.27	

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators 5 Compliance No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

RANGE EXCEED ON MA DUE TO MOUTH ALCOHOL. REPEAT AFTER MA . REPEATED AFTER MA CAUSED .013 WHEN REPEATING TEST Non-compliance: CONTINUES TO SHOW RANGE EXCEEDED DURING MA .

Data Download Date/Time 02/26/2020 00:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO                      **Date** 03/19/2020                      **Type** Department Inspection  
**Inspector** HAUGHEY, MICHAEL, D                      **Time** 12:06:27                      **Software** 8100.27

**0.05 Lot #** 201905A                      **0.08 Lot #** 201905B                      **0.20 Lot #** 201904D                      **0.08 Gas Lot #** AG931603  
**Expiration** 05/14/2021                      **Expiration** 05/14/2021                      **Expiration** 04/30/2021                      **Expiration** 11/12/2021

**Number of Simulators** 5                      **Average Standard Deviation** 0.0005                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 03/19/2020 13:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

<b>Agency</b> DESOTO COUNTY SO	<b>Date</b> 01/30/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PITTS,JOSHUA,E	<b>Time</b> 08:57:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24119080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 11/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/03/2020 09:54

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

<b>Agency</b> DESOTO COUNTY SO	<b>Date</b> 02/22/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PITTS,JOSHUA,E	<b>Time</b> 21:13:33	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24119080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 11/05/2021

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/26/2020 00:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

<b>Agency</b> DESOTO COUNTY SO	<b>Date</b> 03/30/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> PITTS,JOSHUA,E	<b>Time</b> 14:54:01	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24119080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 11/05/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: RFI Detect. DUE TO PASSING RADIO

**Data Download Date/Time** 03/30/2020 15:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

**Agency** SARASOTA COUNTY SO      **Date** 01/09/2020      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D      **Time** 14:05:32      **Software** 8100.27

**0.05 Lot #** 201905A      **0.08 Lot #** 201905B      **0.20 Lot #** 201904D      **0.08 Gas Lot #** AG916501  
**Expiration** 05/14/2021      **Expiration** 05/14/2021      **Expiration** 04/30/2021      **Expiration** 06/14/2021

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

---

**Data Download Date/Time** 01/09/2020 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 01/21/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 08:08:03	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 17919080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 01/21/2020 08:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 02/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 06:04:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 17919080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 02/26/2020 06:12

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 03/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 07:46:49	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 17919080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 03/26/2020 09:12

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

<b>Agency</b> EUSTIS P.D.	<b>Date</b> 01/21/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> CARNEY, DAVID, M	<b>Time</b> 04:39:12	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1031121
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/12/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/21/2020 04:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

<b>Agency</b> EUSTIS P.D.	<b>Date</b> 02/24/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> CARNEY, DAVID, M	<b>Time</b> 03:54:33	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1031121
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/12/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/26/2020 11:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

<b>Agency</b> EUSTIS P.D.	<b>Date</b> 03/31/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> CARNEY, DAVID, M	<b>Time</b> 03:04:26	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1031121
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/12/2021

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

00: RFI Detect.

**Data Download Date/Time** 04/01/2020 00:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 01/21/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 09:47:53	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	No	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

RFI DEP RADIO OUTSIDE ROOM

**Data Download Date/Time** 01/21/2020 09:55

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 02/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 06:13:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 02/26/2020 06:20

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 03/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 07:57:30	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 03/26/2020 08:20

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 01/21/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 07:56:48	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 17919080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/05/2021

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 01/21/2020 08:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 02/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 08:53:34	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 02/26/2020 09:02



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 03/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 09:28:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> No
-------------------------------	----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.075
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAILIN COMPLAINCE SCRIBNER ERROR.

**Data Download Date/Time** 03/26/2020 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 01/21/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 09:32:50	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 01/21/2020 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 02/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 06:05:18	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 02/26/2020 06:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

<b>Agency</b> SARASOTA COUNTY SO	<b>Date</b> 03/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> FELTOVIC,DEP,M	<b>Time</b> 08:40:43	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

CHECK SIMULATOR.

**Data Download Date/Time** 03/26/2020 09:36

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

<b>Agency</b> EVERGLADES NAT PARK	<b>Date</b> 01/24/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 09:15:47	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 30219080A5
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 01/05/2022

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: Improper Sample.AI MYRTIL 5383 PRESENT

**Data Download Date/Time** 01/24/2020 09:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

<b>Agency</b> EVERGLADES NAT PARK	<b>Date</b> 02/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 12:29:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 30219080A5
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 01/05/2022

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

AI MYRTIL# 5383 PRESENT

**Data Download Date/Time** 02/26/2020 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

<b>Agency</b> EVERGLADES NAT PARK	<b>Date</b> 03/25/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 13:46:52	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 30219080A5
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 01/05/2022

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

AI MYRTIL#5383 PRESENT

**Data Download Date/Time** 03/25/2020 15:00

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/15/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 02:30:54	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1175968
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 07/30/2022
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/15/2020 02:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/29/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 21:59:39	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1175968
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 07/30/2022
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.203
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.043
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 02/29/2020 22:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/31/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 19:36:04	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 1175968
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 07/30/2022
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/31/2020 19:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

**Agency** CHARLOTTE COUNTY SO      **Date** 03/19/2020      **Type** Department Inspection  
**Inspector** HAUGHEY, MICHAEL, D      **Time** 12:16:37      **Software** 8100.27

**0.05 Lot #** 201905A      **0.08 Lot #** 201905B      **0.20 Lot #** 201904D      **0.08 Gas Lot #** AG931603  
**Expiration** 05/14/2021      **Expiration** 05/14/2021      **Expiration** 04/30/2021      **Expiration** 11/12/2021

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

Data Download Date/Time 03/19/2020 16:07

Printed: 27 April 2020

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

<b>Agency</b> GAINESVILLE PD	<b>Date</b> 01/29/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> STEMEN, AARON, M	<b>Time</b> 01:29:43	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 24818080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 10/05/2020

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 01/29/2020 01:43

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

<b>Agency</b> GAINESVILLE PD	<b>Date</b> 02/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> STEMEN, AARON, M	<b>Time</b> 01:34:09	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24818080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/05/2020

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/26/2020 01:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

<b>Agency</b> GAINESVILLE PD	<b>Date</b> 03/28/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> STEMEN, AARON, M	<b>Time</b> 00:27:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201910D	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 24818080A2
<b>Expiration</b> 10/22/2021	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 10/05/2020

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/30/2020 03:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO	Date 01/02/2020	Type Agency Inspection	
Inspector LEVY, JAYSON, I	Time 21:21:47	Software 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG805701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/26/2020
<b>Number of Simulators</b> 5	<b>Compliance</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

Data Download Date/Time 04/03/2020 13:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

<b>Agency</b> LEVY COUNTY SO	<b>Date</b> 01/18/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LEVY, JAYSON, I	<b>Time</b> 18:25:18	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG805701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/26/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/03/2020 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

<b>Agency</b> LEVY COUNTY SO	<b>Date</b> 02/01/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> LEVY, JAYSON, I	<b>Time</b> 19:51:28	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG805701
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/26/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.155
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.156
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.158
		0.20 g/210L Test # (Repeat) 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

20: Control Outside Tolerance SIM NOT TIGHT.

**Data Download Date/Time** 04/03/2020 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

<b>Agency</b> PINELLAS COUNTY SO	<b>Date</b> 01/11/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> DEANE, M,	<b>Time</b> 20:48:41	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG831804
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 11/14/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/11/2020 21:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

**Agency** PINELLAS COUNTY SO                      **Date** 02/05/2020                      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 16:30:20                      **Software** 8100.27

**0.05 Lot #** 201905A                      **0.08 Lot #** 201905B                      **0.20 Lot #** 201904D                      **0.08 Gas Lot #** AG931603  
**Expiration** 05/14/2021                      **Expiration** 05/14/2021                      **Expiration** 04/30/2021                      **Expiration** 11/12/2021

**Number of Simulators** 5                      **Average Standard Deviation** 0.0000                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

**Remarks:**

**Data Download Date/Time** 02/05/2020 16:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

<b>Agency</b> PINELLAS COUNTY SO	<b>Date</b> 02/15/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> DEANE, M,	<b>Time</b> 21:08:41	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG916303
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/12/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/15/2020 21:42

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

<b>Agency</b> PINELLAS COUNTY SO	<b>Date</b> 03/06/2020	<b>Type</b> Agency Inspection	
<b>Inspector</b> DEANE, M,	<b>Time</b> 21:29:52	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> AG815602
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/06/2020 21:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

<b>Agency</b> NAS MAYPORT	<b>Date</b> 01/21/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LIGHTSEY,ROBERT,D	<b>Time</b> 12:29:06	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 00919080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 03/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/21/2020 12:39



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

<b>Agency</b> NAS MAYPORT	<b>Date</b> 02/11/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> LIGHTSEY,ROBERT,D	<b>Time</b> 12:40:07	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 00919080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 03/05/2021

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/11/2020 13:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

<b>Agency</b> PINELLAS COUNTY SO	<b>Date</b> 01/10/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> SOTO,ISRAEL,	<b>Time</b> 08:13:04	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 01/16/2020 14:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

<b>Agency</b> PINELLAS COUNTY SO	<b>Date</b> 01/16/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> PLATT,SHAYLA,D	<b>Time</b> 10:25:56	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 01/16/2020 14:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

**Agency** PINELLAS COUNTY SO                      **Date** 01/16/2020                      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 12:28:33                      **Software** 8100.27

**0.05 Lot #** 201905A                      **0.08 Lot #** 201905B                      **0.20 Lot #** 201904D                      **0.08 Gas Lot #** AG916501  
**Expiration** 05/14/2021                      **Expiration** 05/14/2021                      **Expiration** 04/30/2021                      **Expiration** 06/14/2021

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.000
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.000
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.000
		0.20 g/210L Test # (Repeat) 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.000
		0.20 g/210L Test # (Repeat) 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.000
0.20 g/210L Test # (Repeat) 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.000
0.20 g/210L Test # (Repeat) 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.000
0.20 g/210L Test # (Repeat) 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.000
0.20 g/210L Test # (Repeat) 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.000
0.20 g/210L Test # (Repeat) 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.000
0.20 g/210L Test # (Repeat) 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT

Printed: 27 April 2020

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev) (Repeat)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

20: Control Outside Tolerance REPLACED SIMULATOR.

---

**Data Download Date/Time** 01/16/2020 14:11

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 01/07/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 11:01:37	<b>Software</b> 8100.27

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b> 11/13/2020	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.044
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.045
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

05: Control Outside Tolerance NOTIFIED DI PLATT VIA PHONENon-compliance: .

**Data Download Date/Time** 01/07/2020 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 01/07/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 11:46:16	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

REPEATED AI PER DI PLATT W/DIFF. SIMULATOR AND ARS FOR .05

**Data Download Date/Time** 01/07/2020 12:39

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

**Agency** ORANGE COUNTY S.O.                      **Date** 01/30/2020                      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 11:33:12                      **Software** 8100.27

**0.05 Lot #** 201905A                      **0.08 Lot #** 201905B                      **0.20 Lot #** 201904D                      **0.08 Gas Lot #** AG916501  
**Expiration** 05/14/2021                      **Expiration** 05/14/2021                      **Expiration** 04/30/2021                      **Expiration** 06/14/2021

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.204
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.204
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.204
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.205
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.204
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

Data Download Date/Time 01/30/2020 11:42

Printed: 27 April 2020

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 01/07/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 14:40:37	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 06519080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 05/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/07/2020 14:50

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 02/11/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 11:46:00	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 06419080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 05/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/11/2020 11:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 03/11/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 09:49:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 06419080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 05/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/11/2020 10:00

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 01/07/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 13:40:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 06419080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 05/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/07/2020 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 02/11/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 13:11:14	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 06419080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 05/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/11/2020 15:17

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 03/11/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 11:11:45	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 06419080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 05/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/11/2020 11:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

**Agency** MARTIN COUNTY SO                      **Date** 01/11/2020                      **Type** Agency Inspection  
**Inspector** THOMAS, CHRISTOPHER, L           **Time** 02:49:59                      **Software** 8100.27

**0.05 Lot #** 201802B                      **0.08 Lot #** 201808E                      **0.20 Lot #** 201902C                      **0.08 Gas Lot #** 05218080A4  
**Expiration** 02/06/2020                      **Expiration** 08/22/2020                      **Expiration** 02/20/2021                      **Expiration** 05/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/11/2020 03:00

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

**Agency** MARTIN COUNTY SO                      **Date** 01/27/2020                      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 11:41:47                      **Software** 8100.27

**0.05 Lot #** 201905A                      **0.08 Lot #** 201905B                      **0.20 Lot #** 201904D                      **0.08 Gas Lot #** AG916501  
**Expiration** 05/14/2021                      **Expiration** 05/14/2021                      **Expiration** 04/30/2021                      **Expiration** 06/14/2021

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 01/27/2020 11:49

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

**Agency** MARTIN COUNTY SO                      **Date** 02/07/2020                      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L              **Time** 16:23:49                      **Software** 8100.27

**0.05 Lot #** 201811C                      **0.08 Lot #** 201908B                      **0.20 Lot #** 201902C                      **0.08 Gas Lot #** 05218080A4  
**Expiration** 11/13/2020                      **Expiration** 08/07/2021                      **Expiration** 02/20/2021                      **Expiration** 05/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/07/2020 18:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

<b>Agency</b> MARTIN COUNTY SO	<b>Date</b> 03/06/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> THOMAS, CHRISTOPHER, L	<b>Time</b> 14:51:44	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 05218080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 05/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/06/2020 15:29

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

<b>Agency</b> BISCAYNE NAT PARK	<b>Date</b> 01/24/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN,KEVIN,	<b>Time</b> 09:50:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 34917080A4
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 02/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

AI MYRTIL #5383 PRESENT

**Data Download Date/Time** 01/24/2020 10:00

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

<b>Agency</b> BISCAYNE NAT PARK	<b>Date</b> 02/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN, KEVIN,	<b>Time</b> 13:39:33	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 17919080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.078
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed. A F / M A: . 08: COT FORGOT CONNECT HOSEceAI MYRTIL #5383 PRESENT

**Data Download Date/Time** 02/26/2020 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

<b>Agency</b> BISCAYNE NAT PARK	<b>Date</b> 03/25/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLAN, KEVIN,	<b>Time</b> 13:46:18	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 17919080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 08/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

AI MYRTIL#5383 PRESENT

**Data Download Date/Time** 03/25/2020 13:50

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 01/26/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> HARWOOD, CHARLES, M	<b>Time</b> 00:18:07	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 08819080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/27/2020 11:13

**Printed: 27 April 2020**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 02/28/2020	<b>Type</b> Agency Inspection
<b>Inspector</b> HARWOOD, CHARLES, M	<b>Time</b> 11:16:45	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201908B	<b>0.20 Lot #</b> 201902C	<b>0.08 Gas Lot #</b> 08819080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/07/2021	<b>Expiration</b> 02/20/2021	<b>Expiration</b> 06/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/28/2020 11:28