

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

Agency: FDLE

Date/Time: 03/29/2021 05:31:25

Data Download Date/Time: 03/31/2021 05:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE

Date/Time: 04/28/2021 07:33:12

Data Download Date/Time: 04/28/2021 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

Agency: FDLE

Date/Time: 05/27/2021 08:59:34

Data Download Date/Time: 05/27/2021 10:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE

Date/Time: 06/11/2021 08:02:16

Data Download Date/Time: 06/11/2021 08:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

Agency: FDLE

Date/Time: 06/11/2021 08:04:44

Data Download Date/Time: 06/11/2021 08:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

Agency: FDLE

Date/Time: 07/09/2021 15:07:02

Data Download Date/Time: 07/21/2021 14:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000230

Agency: FDLE TRAINING

Date/Time: 01/28/2021 07:52:08

Data Download Date/Time: 01/28/2021 08:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE TRAINING

Date/Time: 02/22/2021 08:50:52

Data Download Date/Time: 02/22/2021 09:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000230

Agency: FDLE TRAINING

Date/Time: 02/23/2021 09:59:40

Data Download Date/Time: 02/23/2021 10:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE TRAINING

Date/Time: 02/23/2021 14:08:12

Data Download Date/Time: 02/23/2021 14:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000230

Agency: FDLE TRAINING

Date/Time: 02/24/2021 09:06:36

Data Download Date/Time: 02/24/2021 09:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE TRAINING

Date/Time: 02/24/2021 11:51:04

Data Download Date/Time: 02/24/2021 12:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000230

Agency: FDLE TRAINING

Date/Time: 02/24/2021 14:11:35

Data Download Date/Time: 02/24/2021 14:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE TRAINING

Date/Time: 02/24/2021 16:26:06

Data Download Date/Time: 02/25/2021 10:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000230

Agency: FDLE TRAINING

Date/Time: 02/24/2021 16:44:52

Data Download Date/Time: 02/26/2021 11:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FDLE TRAINING

Date/Time: 02/24/2021 21:53:36

Data Download Date/Time: 02/26/2021 15:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000230

Agency: FDLE TRAINING

Date/Time: 02/24/2021 23:43:38

Data Download Date/Time: 02/26/2021 15:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

Agency: COLUMBIA COUNTY SO

Date/Time: 03/10/2021 10:27:07

Data Download Date/Time: 03/18/2021 12:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: COLUMBIA COUNTY SO

Date/Time: 03/10/2021 10:27:47

Data Download Date/Time: 03/18/2021 12:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

Agency: COLUMBIA COUNTY SO

Date/Time: 03/10/2021 10:41:46

Data Download Date/Time: 07/02/2021 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 07/30/2021 09:44:39

Data Download Date/Time: 07/30/2021 12:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: NAS JACKSONVILLE PD

Date/Time: 07/30/2021 09:45:28

Data Download Date/Time: 07/30/2021 12:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

Agency: NAS JACKSONVILLE PD

Date/Time: 07/30/2021 09:56:08

Data Download Date/Time: 07/30/2021 12:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: NAS JACKSONVILLE PD

Date/Time: 07/30/2021 10:05:10

Data Download Date/Time: 07/30/2021 12:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

Agency: FHP TROOP A

Date/Time: 01/19/2021 13:45:30

Data Download Date/Time: 03/29/2021 12:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP A

Date/Time: 01/19/2021 13:46:38

Data Download Date/Time: 03/29/2021 12:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

Agency: FHP TROOP A

Date/Time: 01/19/2021 14:03:13

Data Download Date/Time: 03/29/2021 12:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP A

Date/Time: 03/29/2021 10:34:59

Data Download Date/Time: 03/29/2021 12:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 02/16/2021 15:34:01

Data Download Date/Time: 02/16/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GULF COUNTY SO

Date/Time: 03/20/2021 15:51:17

Data Download Date/Time: 03/20/2021 17:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 05/11/2021 17:31:30

Data Download Date/Time: 05/14/2021 12:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GULF COUNTY SO

Date/Time: 06/07/2021 15:31:37

Data Download Date/Time: 06/07/2021 17:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 07/08/2021 14:37:32

Data Download Date/Time: 07/08/2021 15:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GULF COUNTY SO

Date/Time: 08/30/2021 06:31:45

Data Download Date/Time: 09/03/2021 09:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 10/03/2021 06:37:35

Data Download Date/Time: 10/03/2021 08:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GULF COUNTY SO

Date/Time: 11/17/2021 22:40:26

Data Download Date/Time: 11/19/2021 12:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 12/16/2021 06:16:43

Data Download Date/Time: 12/16/2021 07:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

Agency: HAMILTON COUNTY SO

Date/Time: 10/03/2021 16:16:07

Data Download Date/Time: 10/03/2021 17:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

Agency: DIXIE COUNTY SO

Date/Time: 07/01/2021 12:19:11

Data Download Date/Time: 07/28/2021 08:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

Agency: DIXIE COUNTY SO

Date/Time: 07/22/2021 10:38:42

Data Download Date/Time: 08/09/2021 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

Agency: HERNANDO COUNTY S.O.

Date/Time: 09/15/2021 08:00:59

Data Download Date/Time: 09/16/2021 08:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

Agency: MARION COUNTY SO

Date/Time: 03/29/2021 05:43:21

Data Download Date/Time: 03/31/2021 05:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 04/28/2021 07:33:23

Data Download Date/Time: 04/28/2021 11:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

Agency: MARION COUNTY SO

Date/Time: 05/27/2021 09:00:52

Data Download Date/Time: 06/01/2021 14:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 06/11/2021 09:51:23

Data Download Date/Time: 06/11/2021 10:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

Agency: MARION COUNTY SO

Date/Time: 07/01/2021 10:25:21

Data Download Date/Time: 07/26/2021 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 11/04/2021 10:18:29

Data Download Date/Time: 11/04/2021 11:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

Agency: SUMTER COUNTY SO

Date/Time: 04/27/2021 11:07:19

Data Download Date/Time: 04/27/2021 15:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SUMTER COUNTY SO

Date/Time: 06/11/2021 09:47:40

Data Download Date/Time: 06/11/2021 10:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

Agency: SUMTER COUNTY SO

Date/Time: 06/11/2021 09:49:06

Data Download Date/Time: 06/11/2021 10:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SUMTER COUNTY SO

Date/Time: 07/01/2021 10:15:16

Data Download Date/Time: 07/26/2021 11:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency: SUMTER CO SO

Date/Time: 12/22/2021 14:17:17

Data Download Date/Time: 12/22/2021 15:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency: CITRUS COUNTY SO

Date/Time: 01/28/2021 11:28:17

Data Download Date/Time: 09/24/2021 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CITRUS COUNTY SO

Date/Time: 03/01/2021 14:38:37

Data Download Date/Time: 09/24/2021 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency: CITRUS COUNTY SO

Date/Time: 07/17/2021 16:02:58

Data Download Date/Time: 09/24/2021 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CITRUS COUNTY SO

Date/Time: 07/17/2021 16:07:46

Data Download Date/Time: 09/24/2021 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency: CITRUS COUNTY SO

Date/Time: 07/31/2021 15:03:21

Data Download Date/Time: 09/24/2021 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CITRUS COUNTY SO

Date/Time: 07/31/2021 15:10:05

Data Download Date/Time: 09/24/2021 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 01/29/2021 03:19:35

Data Download Date/Time: 02/17/2021 09:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CITRUS COUNTY SO

Date/Time: 02/26/2021 14:00:45

Data Download Date/Time: 06/02/2021 15:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 02/26/2021 15:21:05

Data Download Date/Time: 06/02/2021 15:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CITRUS COUNTY SO

Date/Time: 06/02/2021 15:12:32

Data Download Date/Time: 06/03/2021 15:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency: LAKE COUNTY SO

Date/Time: 03/05/2021 11:16:32

Data Download Date/Time: 03/05/2021 14:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH PD

Date/Time: 11/10/2021 15:38:39

Data Download Date/Time: 11/10/2021 16:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PANAMA CITY BEACH PD

Date/Time: 11/23/2021 10:12:11

Data Download Date/Time: 11/23/2021 11:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency: MIAMI PD

Date/Time: 01/28/2021 12:56:58

Data Download Date/Time: 02/02/2021 13:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

Agency: FFWCC SOUTH REGION

Date/Time: 01/30/2021 21:23:32

Data Download Date/Time: 02/02/2021 19:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

Agency: MANATEE COUNTY SO

Date/Time: 02/01/2021 10:59:29

Data Download Date/Time: 02/02/2021 09:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MANATEE COUNTY SO

Date/Time: 06/10/2021 08:22:39

Data Download Date/Time: 06/10/2021 08:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 01/28/2021 10:39:44

Data Download Date/Time: 02/01/2021 11:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST JOE PD

Date/Time: 02/16/2021 15:36:00

Data Download Date/Time: 02/16/2021 16:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 03/20/2021 15:57:04

Data Download Date/Time: 04/30/2021 18:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST JOE PD

Date/Time: 04/27/2021 18:45:09

Data Download Date/Time: 04/30/2021 18:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 05/01/2021 15:14:04

Data Download Date/Time: 05/01/2021 21:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PORT ST JOE PD

Date/Time: 05/01/2021 15:25:58

Data Download Date/Time: 05/01/2021 21:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 11/17/2021 22:40:14

Data Download Date/Time: 11/19/2021 12:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST JOE PD

Date/Time: 12/16/2021 06:15:17

Data Download Date/Time: 12/16/2021 07:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

Agency: LEE COUNTY SO

Date/Time: 08/12/2021 09:59:47

Data Download Date/Time: 09/28/2021 10:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

Agency: LEE COUNTY SO

Date/Time: 06/21/2021 15:55:08

Data Download Date/Time: 06/22/2021 08:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

Agency: COLLIER COUNTY SO

Date/Time: 03/09/2021 13:45:30

Data Download Date/Time: 03/09/2021 14:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

Agency: COLLIER COUNTY SO

Date/Time: 05/04/2021 14:27:41

Data Download Date/Time: 05/06/2021 15:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

Agency: GLADES COUNTY S.O.

Date/Time: 06/12/2021 11:14:16

Data Download Date/Time: 06/16/2021 12:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

Agency: LEON COUNTY SO

Date/Time: 01/05/2021 14:00:56

Data Download Date/Time: 01/07/2021 12:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEON COUNTY SO

Date/Time: 01/05/2021 14:02:42

Data Download Date/Time: 01/07/2021 12:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

Agency: LEON COUNTY SO

Date/Time: 01/05/2021 14:03:23

Data Download Date/Time: 01/07/2021 12:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEON COUNTY SO

Date/Time: 01/05/2021 14:06:22

Data Download Date/Time: 01/07/2021 12:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

Agency: LEON COUNTY SO

Date/Time: 12/14/2021 11:02:53

Data Download Date/Time: 12/14/2021 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

Agency: LEON COUNTY SO

Date/Time: 09/03/2021 18:34:05

Data Download Date/Time: 10/11/2021 12:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

Agency: GADSDEN COUNTY SO

Date/Time: 04/19/2021 09:00:22

Data Download Date/Time: 10/19/2021 13:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

Agency: WAKULLA COUNTY SO

Date/Time: 07/28/2021 11:01:51

Data Download Date/Time: 11/10/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WAKULLA COUNTY SO

Date/Time: 07/28/2021 11:06:47

Data Download Date/Time: 11/10/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

Agency: WAKULLA COUNTY SO

Date/Time: 07/28/2021 11:21:53

Data Download Date/Time: 11/10/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: WAKULLA COUNTY SO

Date/Time: 07/29/2021 09:32:14

Data Download Date/Time: 11/10/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

Agency: OSCEOLA CORRECTIONS

Date/Time: 08/18/2021 14:23:40

Data Download Date/Time: 08/18/2021 14:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

Agency: COCOA BEACH P.D.

Date/Time: 01/29/2021 12:41:50

Data Download Date/Time: 02/03/2021 12:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COCOA BEACH P.D.

Date/Time: 04/09/2021 12:22:36

Data Download Date/Time: 05/07/2021 11:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

Agency: COCOA BEACH P.D.

Date/Time: 05/07/2021 10:43:07

Data Download Date/Time: 05/07/2021 11:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COCOA BEACH P.D.

Date/Time: 05/28/2021 08:29:34

Data Download Date/Time: 06/04/2021 09:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

Agency: COCOA BEACH P.D.

Date/Time: 05/28/2021 08:34:55

Data Download Date/Time: 06/04/2021 09:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: COCOA BEACH P.D.

Date/Time: 07/23/2021 10:07:29

Data Download Date/Time: 08/02/2021 10:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

Agency: COCOA BEACH P.D.

Date/Time: 07/23/2021 10:15:00

Data Download Date/Time: 08/02/2021 10:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 01/01/2021 11:16:02

Data Download Date/Time: 01/08/2021 13:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE P.D.

Date/Time: 01/01/2021 12:16:36

Data Download Date/Time: 01/08/2021 13:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 01/01/2021 12:17:06

Data Download Date/Time: 01/08/2021 13:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE P.D.

Date/Time: 01/01/2021 12:24:23

Data Download Date/Time: 01/08/2021 13:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 01/01/2021 12:38:50

Data Download Date/Time: 01/08/2021 13:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE P.D.

Date/Time: 03/03/2021 12:33:19

Data Download Date/Time: 03/03/2021 14:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE P.D.

Date/Time: 05/05/2021 18:09:10

Data Download Date/Time: 05/27/2021 19:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE P.D.

Date/Time: 05/27/2021 17:47:38

Data Download Date/Time: 05/27/2021 19:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency: SEMINOLE COUNTY S.O.

Date/Time: 12/07/2021 23:23:11

Data Download Date/Time: 12/09/2021 01:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SEMINOLE COUNTY S.O.

Date/Time: 12/08/2021 21:56:06

Data Download Date/Time: 12/09/2021 01:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency: SEMINOLE COUNTY S.O.

Date/Time: 12/08/2021 23:24:01

Data Download Date/Time: 12/09/2021 01:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

Agency: SEMINOLE COUNTY S.O.

Date/Time: 02/01/2021 09:56:57

Data Download Date/Time: 02/08/2021 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

Agency: HARDEE COUNTY SO

Date/Time: 04/06/2021 14:18:38

Data Download Date/Time: 06/29/2021 09:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HARDEE COUNTY SO

Date/Time: 04/30/2021 01:32:54

Data Download Date/Time: 06/29/2021 09:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

Agency: HARDEE COUNTY SO

Date/Time: 04/30/2021 01:40:10

Data Download Date/Time: 06/29/2021 09:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HARDEE COUNTY SO

Date/Time: 08/30/2021 12:23:14

Data Download Date/Time: 09/04/2021 11:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

Agency: BREVARD COUNTY S.O.

Date/Time: 07/01/2021 08:36:13

Data Download Date/Time: 07/06/2021 09:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 02/01/2021 14:34:59

Data Download Date/Time: 02/02/2021 13:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency: WEST MELBOURNE P.D.

Date/Time: 01/25/2021 10:54:55

Data Download Date/Time: 02/18/2021 15:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: WEST MELBOURNE P.D.

Date/Time: 02/02/2021 11:20:09

Data Download Date/Time: 02/18/2021 15:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency: WEST MELBOURNE P.D.

Date/Time: 02/18/2021 15:23:34

Data Download Date/Time: 02/18/2021 15:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: WEST MELBOURNE P.D.

Date/Time: 02/18/2021 15:28:05

Data Download Date/Time: 03/02/2021 05:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency: WEST MELBOURNE P.D.

Date/Time: 02/18/2021 16:04:54

Data Download Date/Time: 03/02/2021 05:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WEST MELBOURNE P.D.

Date/Time: 03/02/2021 05:11:51

Data Download Date/Time: 03/02/2021 05:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency: WEST MELBOURNE P.D.

Date/Time: 03/02/2021 05:15:28

Data Download Date/Time: 03/02/2021 05:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WEST MELBOURNE P.D.

Date/Time: 03/02/2021 05:16:50

Data Download Date/Time: 03/02/2021 05:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency: WEST MELBOURNE P.D.

Date/Time: 03/02/2021 05:17:35

Data Download Date/Time: 03/02/2021 05:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/23/2021 21:27:43

Data Download Date/Time: 01/29/2021 09:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/23/2021 23:10:42

Data Download Date/Time: 01/29/2021 09:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/24/2021 18:36:42

Data Download Date/Time: 01/29/2021 09:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/22/2021 05:18:47

Data Download Date/Time: 02/22/2021 06:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/25/2021 21:36:42

Data Download Date/Time: 03/25/2021 23:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/11/2021 18:59:46

Data Download Date/Time: 05/30/2021 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/11/2021 19:01:00

Data Download Date/Time: 05/30/2021 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/11/2021 19:02:05

Data Download Date/Time: 05/30/2021 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/11/2021 19:09:09

Data Download Date/Time: 05/30/2021 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/21/2021 03:02:10

Data Download Date/Time: 05/30/2021 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/21/2021 03:03:08

Data Download Date/Time: 05/30/2021 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/21/2021 03:07:04

Data Download Date/Time: 05/30/2021 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/21/2021 03:08:37

Data Download Date/Time: 05/30/2021 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/06/2021 23:41:49

Data Download Date/Time: 07/28/2021 11:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/07/2021 00:54:26

Data Download Date/Time: 07/28/2021 11:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/07/2021 00:56:34

Data Download Date/Time: 07/28/2021 11:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/20/2021 21:44:07

Data Download Date/Time: 07/28/2021 11:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/25/2021 20:03:36

Data Download Date/Time: 07/28/2021 11:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/26/2021 00:24:16

Data Download Date/Time: 07/28/2021 11:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/18/2021 21:22:42

Data Download Date/Time: 09/18/2021 22:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 04/22/2021 17:15:10

Data Download Date/Time: 04/26/2021 06:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 04/22/2021 18:07:05

Data Download Date/Time: 04/26/2021 06:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 10/27/2021 09:45:56

Data Download Date/Time: 10/27/2021 11:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 10/27/2021 09:48:28

Data Download Date/Time: 10/27/2021 11:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 10/27/2021 10:08:09

Data Download Date/Time: 10/27/2021 11:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 11/24/2021 04:31:20

Data Download Date/Time: 12/27/2021 10:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 12/27/2021 09:24:03

Data Download Date/Time: 12/27/2021 10:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency: CORAL SPRINGS PD

Date/Time: 10/27/2021 14:40:06

Data Download Date/Time: 10/27/2021 15:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Agency: ST PETERSBURG PD

Date/Time: 01/04/2021 10:09:23

Data Download Date/Time: 02/10/2021 17:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: ST PETERSBURG PD

Date/Time: 02/10/2021 11:07:36

Data Download Date/Time: 02/10/2021 17:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Agency: ST PETERSBURG PD

Date/Time: 05/11/2021 16:13:44

Data Download Date/Time: 05/12/2021 19:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: ST PETERSBURG PD

Date/Time: 08/23/2021 08:47:02

Data Download Date/Time: 08/24/2021 11:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Agency: ST PETERSBURG PD

Date/Time: 12/21/2021 16:19:13

Data Download Date/Time: 12/23/2021 08:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/22/2021 16:05:12

Data Download Date/Time: 03/23/2021 12:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/07/2021 00:54:38

Data Download Date/Time: 08/05/2021 09:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/07/2021 00:55:46

Data Download Date/Time: 08/05/2021 09:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/25/2021 20:23:34

Data Download Date/Time: 08/05/2021 09:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/18/2021 21:17:11

Data Download Date/Time: 09/18/2021 22:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/18/2021 21:24:09

Data Download Date/Time: 09/18/2021 22:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/18/2021 21:33:53

Data Download Date/Time: 09/18/2021 22:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 08/17/2021 14:43:30

Data Download Date/Time: 08/17/2021 15:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency: DAVIE PD

Date/Time: 07/12/2021 08:28:33

Data Download Date/Time: 07/12/2021 10:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency: FERNANDINA BEACH PD

Date/Time: 06/18/2021 11:13:57

Data Download Date/Time: 06/22/2021 10:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 01/27/2021 20:55:47

Data Download Date/Time: 02/03/2021 17:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 01/27/2021 21:15:59

Data Download Date/Time: 02/03/2021 17:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 05/29/2021 17:12:36

Data Download Date/Time: 05/30/2021 18:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency: JACKSONVILLE SO

Date/Time: 12/07/2021 09:42:56

Data Download Date/Time: 12/16/2021 10:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

Agency: FL HIGHWAY PATROL

Date/Time: 02/03/2021 14:28:32

Data Download Date/Time: 02/05/2021 14:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency: CLAY COUNTY SO

Date/Time: 03/21/2021 07:37:41

Data Download Date/Time: 03/24/2021 07:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 04/22/2021 18:01:45

Data Download Date/Time: 04/26/2021 14:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency: CLAY COUNTY SO

Date/Time: 07/24/2021 13:52:47

Data Download Date/Time: 07/28/2021 08:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 08/27/2021 08:07:12

Data Download Date/Time: 08/30/2021 09:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency: CLAY COUNTY SO

Date/Time: 08/27/2021 08:57:35

Data Download Date/Time: 08/30/2021 09:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 10/27/2021 09:29:39

Data Download Date/Time: 11/22/2021 10:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency: CLAY COUNTY SO

Date/Time: 10/27/2021 10:49:26

Data Download Date/Time: 11/22/2021 10:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 12/27/2021 09:49:44

Data Download Date/Time: 12/27/2021 10:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency: DADE CITY PD

Date/Time: 12/22/2021 10:05:45

Data Download Date/Time: 12/22/2021 10:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency: HARDEE COUNTY SO

Date/Time: 04/06/2021 14:18:28

Data Download Date/Time: 05/17/2021 12:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency: ST PETERSBURG PD

Date/Time: 05/24/2021 16:08:58

Data Download Date/Time: 12/16/2021 10:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST PETERSBURG PD

Date/Time: 06/29/2021 13:29:02

Data Download Date/Time: 12/16/2021 10:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency: ST PETERSBURG PD

Date/Time: 12/15/2021 13:27:54

Data Download Date/Time: 12/16/2021 10:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 04/12/2021 16:30:20

Data Download Date/Time: 04/13/2021 07:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LARGO PD

Date/Time: 05/12/2021 14:58:26

Data Download Date/Time: 05/13/2021 10:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 12/18/2021 03:10:28

Data Download Date/Time: 01/05/2022 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LARGO PD

Date/Time: 12/18/2021 03:11:39

Data Download Date/Time: 01/05/2022 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 12/18/2021 03:23:45

Data Download Date/Time: 01/05/2022 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency: FL HIGHWAY PATROL

Date/Time: 02/25/2021 11:08:42

Data Download Date/Time: 04/15/2021 14:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 10/03/2021 00:17:52

Data Download Date/Time: 10/03/2021 19:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

Agency: FHP MIAMI

Date/Time: 09/30/2021 08:14:31

Data Download Date/Time: 09/30/2021 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FL HIGHWAY PATROL

Date/Time: 06/03/2021 14:10:10

Data Download Date/Time: 06/04/2021 11:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Agency: FL HIGHWAY PATROL

Date/Time: 04/14/2021 15:26:10

Data Download Date/Time: 04/19/2021 09:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 09/28/2021 10:22:41

Data Download Date/Time: 09/28/2021 13:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP D ORLANDO

Date/Time: 06/21/2021 14:39:13

Data Download Date/Time: 12/20/2021 11:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 12/09/2021 15:30:30

Data Download Date/Time: 12/10/2021 08:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency: PUTNAM COUNTY SO

Date/Time: 06/04/2021 11:49:02

Data Download Date/Time: 06/04/2021 14:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PUTNAM COUNTY SO

Date/Time: 07/28/2021 16:05:13

Data Download Date/Time: 11/10/2021 15:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency: PUTNAM COUNTY SO

Date/Time: 07/28/2021 16:05:54

Data Download Date/Time: 11/10/2021 15:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PUTNAM COUNTY SO

Date/Time: 07/28/2021 16:08:43

Data Download Date/Time: 11/10/2021 15:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency: PUTNAM COUNTY SO

Date/Time: 10/29/2021 13:33:38

Data Download Date/Time: 11/10/2021 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PUTNAM COUNTY SO

Date/Time: 10/29/2021 13:36:35

Data Download Date/Time: 11/10/2021 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency: PUTNAM COUNTY SO

Date/Time: 10/29/2021 13:48:11

Data Download Date/Time: 11/10/2021 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PUTNAM COUNTY SO

Date/Time: 10/29/2021 14:02:19

Data Download Date/Time: 11/10/2021 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency: PUTNAM COUNTY SO

Date/Time: 10/29/2021 14:26:36

Data Download Date/Time: 11/10/2021 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PUTNAM COUNTY SO

Date/Time: 10/29/2021 14:31:57

Data Download Date/Time: 11/10/2021 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency: PUTNAM COUNTY SO

Date/Time: 10/29/2021 14:32:45

Data Download Date/Time: 11/10/2021 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PUTNAM COUNTY SO

Date/Time: 10/29/2021 14:33:52

Data Download Date/Time: 11/10/2021 16:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

Agency: ST JOHNS COUNTY SO

Date/Time: 07/10/2021 23:17:58

Data Download Date/Time: 07/29/2021 14:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST JOHNS COUNTY SO

Date/Time: 07/10/2021 23:18:49

Data Download Date/Time: 07/29/2021 14:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

Agency: ST JOHNS COUNTY SO

Date/Time: 08/08/2021 03:20:09

Data Download Date/Time: 08/27/2021 01:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency: FLAGLER COUNTY SO

Date/Time: 03/29/2021 21:10:56

Data Download Date/Time: 03/29/2021 21:53

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency: HAINES CITY PD

Date/Time: 04/28/2021 14:47:45

Data Download Date/Time: 04/29/2021 15:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

Agency: NEW SMYRNA BEACH PD

Date/Time: 02/28/2021 09:11:30

Data Download Date/Time: 03/02/2021 12:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: NEW SMYRNA BEACH PD

Date/Time: 09/04/2021 14:37:49

Data Download Date/Time: 09/22/2021 08:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

Agency: NEW SMYRNA BEACH PD

Date/Time: 09/18/2021 08:15:44

Data Download Date/Time: 09/22/2021 08:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: NEW SMYRNA BEACH PD

Date/Time: 09/18/2021 08:16:37

Data Download Date/Time: 09/22/2021 08:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

Agency: NEW SMYRNA BEACH PD

Date/Time: 09/22/2021 07:24:33

Data Download Date/Time: 09/22/2021 08:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: NEW SMYRNA BEACH PD

Date/Time: 11/13/2021 07:43:35

Data Download Date/Time: 11/13/2021 09:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 11/24/2021 16:02:42

Data Download Date/Time: 11/24/2021 17:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

Agency: BREVARD COUNTY SO

Date/Time: 11/17/2021 16:03:21

Data Download Date/Time: 11/24/2021 09:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 04/27/2021 15:01:42

Data Download Date/Time: 04/28/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COLLIER COUNTY SO

Date/Time: 04/28/2021 13:00:51

Data Download Date/Time: 04/28/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 04/28/2021 13:27:44

Data Download Date/Time: 04/28/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency: LEE COUNTY SO

Date/Time: 06/22/2021 09:21:09

Data Download Date/Time: 07/28/2021 11:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency: TYNDALL AFB

Date/Time: 07/28/2021 07:23:30

Data Download Date/Time: 07/28/2021 08:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 02/01/2021 15:07:45

Data Download Date/Time: 02/22/2021 17:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: AUBURNDALE PD

Date/Time: 05/20/2021 12:23:49

Data Download Date/Time: 05/20/2021 14:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 05/20/2021 14:27:01

Data Download Date/Time: 06/04/2021 11:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: AUBURNDALE PD

Date/Time: 06/29/2021 14:54:34

Data Download Date/Time: 06/30/2021 11:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency: WEST PALM BEACH PD

Date/Time: 04/10/2021 13:16:21

Data Download Date/Time: 05/22/2021 08:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: WEST PALM BEACH PD

Date/Time: 12/17/2021 09:13:14

Data Download Date/Time: 01/13/2022 09:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency: EDGEWATER P.D.

Date/Time: 02/01/2021 09:58:01

Data Download Date/Time: 02/01/2021 11:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency: PUTNAM COUNTY SO

Date/Time: 06/30/2021 11:58:54

Data Download Date/Time: 11/18/2021 16:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PUTNAM COUNTY SO

Date/Time: 06/30/2021 12:00:11

Data Download Date/Time: 11/18/2021 16:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

Agency: PORT ORANGE P.D.

Date/Time: 06/30/2021 13:30:31

Data Download Date/Time: 07/01/2021 17:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency: ROCKLEDGE P.D.

Date/Time: 02/10/2021 10:00:17

Data Download Date/Time: 02/10/2021 14:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ROCKLEDGE P.D.

Date/Time: 03/26/2021 18:15:57

Data Download Date/Time: 04/08/2021 13:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency: ROCKLEDGE P.D.

Date/Time: 04/01/2021 01:09:48

Data Download Date/Time: 04/08/2021 13:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ROCKLEDGE P.D.

Date/Time: 10/06/2021 14:12:13

Data Download Date/Time: 10/22/2021 11:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency: MELBOURNE P.D.

Date/Time: 03/30/2021 07:48:42

Data Download Date/Time: 03/31/2021 06:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MELBOURNE P.D.

Date/Time: 09/03/2021 09:10:11

Data Download Date/Time: 10/11/2021 07:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency: SEMINOLE COUNTY S.O.

Date/Time: 02/02/2021 12:09:38

Data Download Date/Time: 02/10/2021 11:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

Agency: PINELLAS COUNTY SO

Date/Time: 08/27/2021 01:03:33

Data Download Date/Time: 09/24/2021 04:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency: JACKSONVILLE SO

Date/Time: 10/26/2021 10:29:53

Data Download Date/Time: 10/26/2021 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: JACKSONVILLE SO

Date/Time: 10/26/2021 10:41:09

Data Download Date/Time: 10/26/2021 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

Agency: JACKSONVILLE SO

Date/Time: 03/24/2021 16:18:47

Data Download Date/Time: 03/25/2021 09:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

Agency: NASSAU COUNTY SO

Date/Time: 02/25/2021 19:22:22

Data Download Date/Time: 03/29/2021 14:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

Agency: GILCHRIST COUNTY SO

Date/Time: 02/28/2021 21:18:47

Data Download Date/Time: 03/01/2021 10:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GILCHRIST COUNTY SO

Date/Time: 07/31/2021 07:16:22

Data Download Date/Time: 08/06/2021 15:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

Agency: LEVY COUNTY SO

Date/Time: 03/01/2021 14:35:26

Data Download Date/Time: 03/02/2021 08:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 08/29/2021 09:59:36

Data Download Date/Time: 09/05/2021 03:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARIANNA PD

Date/Time: 09/30/2021 08:20:30

Data Download Date/Time: 10/11/2021 10:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

Agency: MADISON COUNTY SO

Date/Time: 02/17/2021 13:42:56

Data Download Date/Time: 02/19/2021 10:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

Agency: BAY COUNTY SO

Date/Time: 04/20/2021 10:03:46

Data Download Date/Time: 04/20/2021 14:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

Agency: BAY COUNTY SO

Date/Time: 12/15/2021 13:57:12

Data Download Date/Time: 12/16/2021 14:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 07/15/2021 20:54:28

Data Download Date/Time: 07/19/2021 19:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

Agency: OKEECHOBEE CO SO

Date/Time: 06/18/2021 03:08:46

Data Download Date/Time: 06/21/2021 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OKEECHOBEE CO SO

Date/Time: 10/28/2021 15:04:56

Data Download Date/Time: 11/01/2021 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

Agency: OKEECHOBEE COUNTY SO

Date/Time: 06/18/2021 03:10:57

Data Download Date/Time: 06/21/2021 07:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OKEECHOBEE COUNTY SO

Date/Time: 10/28/2021 15:06:30

Data Download Date/Time: 11/01/2021 15:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

Agency: OKEECHOBEE P.D.

Date/Time: 05/04/2021 16:28:40

Data Download Date/Time: 05/07/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OKEECHOBEE P.D.

Date/Time: 05/04/2021 16:33:25

Data Download Date/Time: 05/07/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

Agency: OKEECHOBEE P.D.

Date/Time: 05/04/2021 16:43:02

Data Download Date/Time: 05/07/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OKEECHOBEE P.D.

Date/Time: 05/04/2021 16:43:37

Data Download Date/Time: 05/07/2021 16:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

Agency: INDIAN RIVER CO. SO

Date/Time: 11/25/2021 06:44:17

Data Download Date/Time: 11/26/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

Agency: INDIAN RIVER CO. SO

Date/Time: 07/11/2021 09:39:11

Data Download Date/Time: 07/16/2021 17:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: INDIAN RIVER CO. SO

Date/Time: 07/11/2021 09:54:49

Data Download Date/Time: 07/16/2021 17:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: EUSTIS P.D.

Date/Time: 11/30/2021 09:49:40

Data Download Date/Time: 12/01/2021 12:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: EUSTIS P.D.

Date/Time: 11/30/2021 09:50:56

Data Download Date/Time: 12/01/2021 12:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

Agency: SARASOTA COUNTY SO

Date/Time: 09/15/2021 10:55:06

Data Download Date/Time: 09/15/2021 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY SO

Date/Time: 06/22/2021 09:46:18

Data Download Date/Time: 07/21/2021 06:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency: NAS MAYPORT

Date/Time: 06/17/2021 13:17:31

Data Download Date/Time: 07/19/2021 11:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

Agency: CLAY COUNTY SO

Date/Time: 01/22/2021 15:02:58

Data Download Date/Time: 01/28/2021 07:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 03/21/2021 11:03:05

Data Download Date/Time: 03/24/2021 07:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

Agency: CLAY COUNTY SO

Date/Time: 04/22/2021 18:22:07

Data Download Date/Time: 04/26/2021 11:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 08/27/2021 08:12:48

Data Download Date/Time: 08/27/2021 09:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency: TYNDALL AFB

Date/Time: 01/20/2021 08:03:43

Data Download Date/Time: 01/20/2021 11:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TYNDALL AFB

Date/Time: 01/20/2021 08:35:13

Data Download Date/Time: 01/20/2021 11:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency: TYNDALL AFB

Date/Time: 01/20/2021 08:47:25

Data Download Date/Time: 01/20/2021 11:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency: PUTNAM COUNTY SO

Date/Time: 07/29/2021 13:29:49

Data Download Date/Time: 08/11/2021 10:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PUTNAM COUNTY SO

Date/Time: 07/29/2021 13:30:51

Data Download Date/Time: 08/11/2021 10:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency: PUTNAM COUNTY SO

Date/Time: 08/11/2021 10:44:29

Data Download Date/Time: 09/08/2021 16:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PUTNAM COUNTY SO

Date/Time: 08/23/2021 02:23:12

Data Download Date/Time: 09/08/2021 16:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency: PUTNAM COUNTY SO

Date/Time: 08/28/2021 17:18:11

Data Download Date/Time: 09/08/2021 16:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PUTNAM COUNTY SO

Date/Time: 09/08/2021 20:46:04

Data Download Date/Time: 10/01/2021 15:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency: PUTNAM COUNTY SO

Date/Time: 12/20/2021 10:50:25

Data Download Date/Time: 12/20/2021 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Agency: TITUSVILLE PD

Date/Time: 12/20/2021 16:06:51

Data Download Date/Time: 12/27/2021 16:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 12/20/2021 17:26:51

Data Download Date/Time: 12/27/2021 16:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Agency: TITUSVILLE PD

Date/Time: 12/22/2021 12:24:20

Data Download Date/Time: 12/27/2021 16:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency: ST PETERSBURG PD

Date/Time: 06/01/2021 12:26:12

Data Download Date/Time: 06/26/2021 16:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: ST PETERSBURG PD

Date/Time: 06/01/2021 12:55:58

Data Download Date/Time: 06/26/2021 16:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 01/27/2021 20:54:28

Data Download Date/Time: 01/30/2021 19:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 05/29/2021 17:13:55

Data Download Date/Time: 06/01/2021 18:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 10/30/2021 17:37:58

Data Download Date/Time: 10/30/2021 18:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency: NAS JACKSONVILLE PD

Date/Time: 05/27/2021 02:59:27

Data Download Date/Time: 06/01/2021 22:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: NAS JACKSONVILLE PD

Date/Time: 07/30/2021 09:41:11

Data Download Date/Time: 07/30/2021 10:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency: NAS JACKSONVILLE PD

Date/Time: 10/24/2021 19:15:04

Data Download Date/Time: 01/03/2022 16:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency: NEW SMYRNA BEACH PD

Date/Time: 09/18/2021 08:22:17

Data Download Date/Time: 09/22/2021 08:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: NEW SMYRNA BEACH PD

Date/Time: 09/22/2021 07:40:28

Data Download Date/Time: 09/22/2021 08:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency: NEW SMYRNA BEACH PD

Date/Time: 11/13/2021 07:38:16

Data Download Date/Time: 11/13/2021 10:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: NEW SMYRNA BEACH PD

Date/Time: 11/14/2021 08:37:15

Data Download Date/Time: 12/20/2021 17:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency: NEW SMYRNA BEACH PD

Date/Time: 12/18/2021 12:51:27

Data Download Date/Time: 12/20/2021 17:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency: TALLAHASSEE PD

Date/Time: 08/24/2021 09:54:09

Data Download Date/Time: 08/27/2021 08:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TALLAHASSEE PD

Date/Time: 09/28/2021 09:24:56

Data Download Date/Time: 09/28/2021 14:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency: TALLAHASSEE PD

Date/Time: 10/20/2021 09:23:02

Data Download Date/Time: 10/20/2021 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency: FFWCC

Date/Time: 03/23/2021 09:30:20

Data Download Date/Time: 03/23/2021 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FFWCC

Date/Time: 03/23/2021 09:35:41

Data Download Date/Time: 03/23/2021 14:08

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency: FFWCC

Date/Time: 03/23/2021 09:37:22

Data Download Date/Time: 03/23/2021 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency: FHP TRP E MIAMI

Date/Time: 04/01/2021 03:04:24

Data Download Date/Time: 04/04/2021 23:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency: FHP TROOP K

Date/Time: 10/02/2021 09:37:48

Data Download Date/Time: 11/03/2021 23:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency: FL HIGHWAY PATROL

Date/Time: 01/03/2021 10:28:15

Data Download Date/Time: 11/30/2021 15:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 12/02/2021 11:05:53

Data Download Date/Time: 12/02/2021 13:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency: FL HIGHWAY PATROL

Date/Time: 10/26/2021 20:34:18

Data Download Date/Time: 11/29/2021 00:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency: FL HIGHWAY PATROL

Date/Time: 02/03/2021 14:34:54

Data Download Date/Time: 02/05/2021 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 03/30/2021 14:26:12

Data Download Date/Time: 04/02/2021 14:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency: FL HIGHWAY PATROL

Date/Time: 05/26/2021 11:35:11

Data Download Date/Time: 05/26/2021 12:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency: FL HIGHWAY PATROL

Date/Time: 06/25/2021 11:27:29

Data Download Date/Time: 06/25/2021 13:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 07/28/2021 07:49:58

Data Download Date/Time: 07/28/2021 12:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency: OSCEOLA COUNTY S.O.

Date/Time: 09/05/2021 11:15:03

Data Download Date/Time: 09/21/2021 17:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency: LAKELAND PD

Date/Time: 07/09/2021 08:57:58

Data Download Date/Time: 07/09/2021 10:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency: OSCEOLA COUNTY S.O.

Date/Time: 09/05/2021 12:22:24

Data Download Date/Time: 09/21/2021 17:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OSCEOLA COUNTY S.O.

Date/Time: 09/05/2021 12:32:00

Data Download Date/Time: 09/21/2021 17:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency: OSCEOLA COUNTY S.O.

Date/Time: 09/05/2021 12:36:30

Data Download Date/Time: 09/21/2021 17:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OSCEOLA COUNTY S.O.

Date/Time: 09/05/2021 12:54:38

Data Download Date/Time: 09/21/2021 17:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency: OKALOOSA COUNTY SO

Date/Time: 06/23/2021 07:21:33

Data Download Date/Time: 11/10/2021 15:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 01/30/2021 19:00:02

Data Download Date/Time: 02/01/2021 09:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 02/27/2021 08:26:05

Data Download Date/Time: 03/01/2021 09:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 03/29/2021 18:17:07

Data Download Date/Time: 03/29/2021 20:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 04/29/2021 11:56:52

Data Download Date/Time: 04/29/2021 13:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 05/30/2021 16:37:54

Data Download Date/Time: 05/30/2021 18:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 05/30/2021 16:48:20

Data Download Date/Time: 05/30/2021 18:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 06/30/2021 16:07:39

Data Download Date/Time: 07/06/2021 19:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 08/11/2021 17:50:18

Data Download Date/Time: 08/12/2021 07:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 08/24/2021 09:30:35

Data Download Date/Time: 08/25/2021 15:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 09/29/2021 11:10:12

Data Download Date/Time: 10/03/2021 16:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 10/03/2021 14:54:52

Data Download Date/Time: 10/03/2021 16:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CRESTVIEW PD

Date/Time: 10/03/2021 14:56:27

Data Download Date/Time: 10/03/2021 16:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 10/29/2021 15:41:41

Data Download Date/Time: 10/29/2021 17:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency: SARASOTA COUNTY SO

Date/Time: 07/30/2021 12:41:55

Data Download Date/Time: 08/05/2021 10:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency: HILLSBOROUGH CO SO

Date/Time: 04/28/2021 10:42:03

Data Download Date/Time: 04/28/2021 14:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HILLSBOROUGH CO SO

Date/Time: 05/05/2021 14:44:33

Data Download Date/Time: 05/11/2021 14:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency: HILLSBOROUGH CO SO

Date/Time: 05/06/2021 10:43:24

Data Download Date/Time: 05/11/2021 14:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency: HERNANDO COUNTY SO

Date/Time: 01/14/2021 17:54:17

Data Download Date/Time: 02/13/2021 10:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HERNANDO COUNTY SO

Date/Time: 07/12/2021 15:19:41

Data Download Date/Time: 07/13/2021 18:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency: HERNANDO COUNTY SO

Date/Time: 08/14/2021 10:01:30

Data Download Date/Time: 08/14/2021 11:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HERNANDO COUNTY SO

Date/Time: 08/14/2021 10:15:05

Data Download Date/Time: 08/14/2021 11:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency: HERNANDO COUNTY SO

Date/Time: 11/18/2021 11:45:44

Data Download Date/Time: 11/18/2021 14:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency: PINELLAS COUNTY SO

Date/Time: 11/10/2021 23:38:33

Data Download Date/Time: 11/24/2021 20:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency: ST PETERSBURG PD

Date/Time: 06/30/2021 09:21:43

Data Download Date/Time: 07/01/2021 16:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST PETERSBURG PD

Date/Time: 12/21/2021 16:17:43

Data Download Date/Time: 12/23/2021 08:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency: PORT ORANGE PD

Date/Time: 05/25/2021 01:56:01

Data Download Date/Time: 06/02/2021 20:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency: FORT MYERS PD

Date/Time: 01/27/2021 09:59:34

Data Download Date/Time: 01/27/2021 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FORT MYERS PD

Date/Time: 01/27/2021 10:00:22

Data Download Date/Time: 01/27/2021 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency: FORT MYERS PD

Date/Time: 01/27/2021 10:01:21

Data Download Date/Time: 01/27/2021 12:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency: FERNANDINA BEACH PD

Date/Time: 03/22/2021 14:12:42

Data Download Date/Time: 03/23/2021 09:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency: LAKELAND PD

Date/Time: 07/09/2021 12:44:35

Data Download Date/Time: 07/09/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency: OSCEOLA COUNTY SO

Date/Time: 09/05/2021 12:11:26

Data Download Date/Time: 09/21/2021 16:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OSCEOLA COUNTY SO

Date/Time: 09/05/2021 12:41:13

Data Download Date/Time: 09/21/2021 16:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency: OSCEOLA COUNTY SO

Date/Time: 09/05/2021 12:52:19

Data Download Date/Time: 09/21/2021 16:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency: SARASOTA COUNTY SO

Date/Time: 08/28/2021 11:38:48

Data Download Date/Time: 08/30/2021 21:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006220

Agency: FDLE TRAINER

Date/Time: 09/03/2021 10:29:33

Data Download Date/Time: 09/03/2021 10:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

Agency: BREVARD COUNTY S.O.

Date/Time: 05/27/2021 22:22:55

Data Download Date/Time: 07/23/2021 22:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

Agency: HENDRY COUNTY SO

Date/Time: 02/05/2021 09:54:29

Data Download Date/Time: 03/01/2021 09:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/22/2021 14:11:24

Data Download Date/Time: 01/28/2021 07:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 03/21/2021 07:39:37

Data Download Date/Time: 04/05/2021 10:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 04/22/2021 17:07:44

Data Download Date/Time: 04/26/2021 06:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 04/22/2021 17:09:20

Data Download Date/Time: 04/26/2021 06:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 04/22/2021 17:12:43

Data Download Date/Time: 04/26/2021 06:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 07/24/2021 10:06:10

Data Download Date/Time: 07/27/2021 09:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 08/27/2021 08:07:59

Data Download Date/Time: 08/27/2021 09:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 10/27/2021 09:35:23

Data Download Date/Time: 10/27/2021 11:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 10/27/2021 10:05:12

Data Download Date/Time: 10/27/2021 11:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 12/27/2021 09:19:25

Data Download Date/Time: 12/27/2021 10:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

Agency: VOLUSIA COUNTY S.O.

Date/Time: 01/25/2021 12:02:37

Data Download Date/Time: 02/08/2021 16:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: VOLUSIA COUNTY S.O.

Date/Time: 01/25/2021 12:03:35

Data Download Date/Time: 02/08/2021 16:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

Agency: VOLUSIA COUNTY S.O.

Date/Time: 01/25/2021 12:14:55

Data Download Date/Time: 02/08/2021 16:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: VOLUSIA COUNTY S.O.

Date/Time: 01/25/2021 12:19:00

Data Download Date/Time: 02/08/2021 16:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

Agency: GAINESVILLE PD

Date/Time: 10/04/2021 22:09:45

Data Download Date/Time: 10/29/2021 11:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

Agency: MIAMI PD

Date/Time: 04/20/2021 17:29:08

Data Download Date/Time: 05/03/2021 14:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MIAMI PD

Date/Time: 06/18/2021 19:25:42

Data Download Date/Time: 06/28/2021 16:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

Agency: MIAMI PD

Date/Time: 07/25/2021 15:21:53

Data Download Date/Time: 07/27/2021 11:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

Agency: MONROE COUNTY S.O.

Date/Time: 02/28/2021 10:44:02

Data Download Date/Time: 03/02/2021 14:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

Agency: FL HIGHWAY PATROL

Date/Time: 02/21/2021 01:09:45

Data Download Date/Time: 02/21/2021 02:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 02/21/2021 01:10:43

Data Download Date/Time: 02/21/2021 02:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	Fail
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

Agency: FL HIGHWAY PATROL

Date/Time: 06/25/2021 11:23:36

Data Download Date/Time: 06/25/2021 13:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 07/28/2021 07:49:40

Data Download Date/Time: 07/28/2021 12:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

Agency: FL HIGHWAY PATROL

Date/Time: 12/10/2021 11:00:53

Data Download Date/Time: 12/10/2021 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 12/10/2021 11:02:06

Data Download Date/Time: 12/10/2021 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

Agency: FL HIGHWAY PATROL

Date/Time: 12/10/2021 11:04:55

Data Download Date/Time: 12/10/2021 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

Agency: PALM BEACH CO SO

Date/Time: 09/11/2021 03:11:35

Data Download Date/Time: 09/13/2021 10:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

Agency: FLAGLER COUNTY SO

Date/Time: 05/01/2021 08:24:59

Data Download Date/Time: 05/03/2021 16:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 04/28/2021 09:11:27

Data Download Date/Time: 06/03/2021 14:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: BOCA RATON PD

Date/Time: 04/28/2021 09:14:10

Data Download Date/Time: 06/03/2021 14:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 04/28/2021 11:04:18

Data Download Date/Time: 06/03/2021 14:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 06/23/2021 15:01:36

Data Download Date/Time: 06/30/2021 13:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 07/27/2021 16:03:45

Data Download Date/Time: 07/27/2021 19:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

Agency: MANATEE COUNTY SO

Date/Time: 05/17/2021 10:10:59

Data Download Date/Time: 05/17/2021 11:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

Agency: FL HIGHWAY PATROL

Date/Time: 10/07/2021 21:51:43

Data Download Date/Time: 12/07/2021 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

Agency: FL HIGHWAY PATROL

Date/Time: 09/05/2021 14:06:29

Data Download Date/Time: 09/30/2021 11:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

Agency: FL HIGHWAY PATROL

Date/Time: 12/18/2021 20:12:51

Data Download Date/Time: 12/30/2021 05:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

Agency: MANATEE COUNTY SO

Date/Time: 06/18/2021 14:17:45

Data Download Date/Time: 06/30/2021 09:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

Agency: FHP TROOP G

Date/Time: 04/15/2021 14:24:44

Data Download Date/Time: 04/15/2021 19:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP G

Date/Time: 11/10/2021 16:56:29

Data Download Date/Time: 11/11/2021 10:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

Agency: FHP TROOP G

Date/Time: 11/11/2021 10:42:10

Data Download Date/Time: 11/11/2021 10:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

Agency: FL HIGHWAY PATROL

Date/Time: 07/21/2021 14:32:54

Data Download Date/Time: 08/17/2021 08:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 07/28/2021 08:34:53

Data Download Date/Time: 08/17/2021 08:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

Agency: FHP

Date/Time: 01/22/2021 15:25:46

Data Download Date/Time: 02/24/2021 14:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 06/25/2021 11:24:56

Data Download Date/Time: 06/25/2021 13:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

Agency: FHP

Date/Time: 07/28/2021 07:49:30

Data Download Date/Time: 07/29/2021 09:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 10/20/2021 14:52:29

Data Download Date/Time: 11/18/2021 10:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

Agency: TALLAHASSEE PD

Date/Time: 08/24/2021 10:15:12

Data Download Date/Time: 08/25/2021 10:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

Agency: COLLIER COUNTY SO

Date/Time: 03/18/2021 14:47:59

Data Download Date/Time: 03/19/2021 16:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

Agency: COLLIER COUNTY SO

Date/Time: 04/29/2021 16:09:15

Data Download Date/Time: 04/30/2021 15:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

Agency: HILLSBOROUGH CO SO

Date/Time: 04/27/2021 13:52:05

Data Download Date/Time: 04/28/2021 12:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HILLSBOROUGH CO SO

Date/Time: 05/25/2021 09:08:25

Data Download Date/Time: 05/25/2021 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

Agency: PENSACOLA P.D.

Date/Time: 05/28/2021 02:00:38

Data Download Date/Time: 06/09/2021 00:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

Agency: MONROE COUNTY SO

Date/Time: 04/28/2021 15:59:03

Data Download Date/Time: 04/30/2021 12:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MONROE COUNTY SO

Date/Time: 05/25/2021 11:36:03

Data Download Date/Time: 06/21/2021 14:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

Agency: MONROE COUNTY SO

Date/Time: 06/18/2021 09:30:15

Data Download Date/Time: 06/21/2021 14:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MONROE COUNTY SO

Date/Time: 11/07/2021 12:59:22

Data Download Date/Time: 11/07/2021 14:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

Agency: MONROE COUNTY SO

Date/Time: 11/07/2021 13:23:53

Data Download Date/Time: 11/07/2021 14:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MONROE COUNTY SO

Date/Time: 11/07/2021 13:24:43

Data Download Date/Time: 11/07/2021 14:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

Agency: MONROE COUNTY SO

Date/Time: 11/07/2021 13:25:43

Data Download Date/Time: 11/07/2021 14:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MONROE COUNTY SO

Date/Time: 11/07/2021 13:26:39

Data Download Date/Time: 11/07/2021 14:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

Agency: MONROE COUNTY SO

Date/Time: 11/07/2021 13:27:12

Data Download Date/Time: 11/07/2021 14:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

Agency: MARION COUNTY SO

Date/Time: 01/11/2021 13:17:47

Data Download Date/Time: 01/11/2021 14:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MARION COUNTY SO

Date/Time: 01/11/2021 14:50:28

Data Download Date/Time: 02/25/2021 11:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

Agency: MARION COUNTY SO

Date/Time: 09/17/2021 10:13:24

Data Download Date/Time: 09/23/2021 11:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 04/17/2021 11:44:57

Data Download Date/Time: 04/17/2021 13:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 04/17/2021 11:45:42

Data Download Date/Time: 04/17/2021 13:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 06/21/2021 11:12:09

Data Download Date/Time: 06/22/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 06/21/2021 11:26:15

Data Download Date/Time: 06/22/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 06/21/2021 12:00:32

Data Download Date/Time: 06/22/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 06/22/2021 07:20:48

Data Download Date/Time: 06/22/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 06/22/2021 08:24:47

Data Download Date/Time: 06/22/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 06/22/2021 09:13:51

Data Download Date/Time: 06/22/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 06/22/2021 10:09:22

Data Download Date/Time: 06/22/2021 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 06/22/2021 12:58:00

Data Download Date/Time: 07/19/2021 22:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 06/22/2021 13:07:32

Data Download Date/Time: 07/19/2021 22:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

Agency: PENSACOLA PD

Date/Time: 03/20/2021 18:34:30

Data Download Date/Time: 07/20/2021 08:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PENSACOLA PD

Date/Time: 07/14/2021 15:27:33

Data Download Date/Time: 07/20/2021 08:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

Agency: ST JOHNS SO

Date/Time: 09/25/2021 02:28:42

Data Download Date/Time: 12/02/2021 21:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

Agency: WALTON COUNTY SO

Date/Time: 07/15/2021 21:26:13

Data Download Date/Time: 07/19/2021 22:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 07/15/2021 21:42:50

Data Download Date/Time: 07/19/2021 22:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

Agency:

Date/Time: 03/23/2021 13:23:34

Data Download Date/Time: 04/05/2021 12:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency:

Date/Time: 03/24/2021 14:12:38

Data Download Date/Time: 04/05/2021 12:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

Agency:

Date/Time: 03/24/2021 15:38:29

Data Download Date/Time: 04/05/2021 17:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 08/31/2021 20:23:55

Data Download Date/Time: 10/07/2021 19:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

Agency: FHP

Date/Time: 08/31/2021 20:24:29

Data Download Date/Time: 10/07/2021 19:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FHP

Date/Time: 08/31/2021 20:30:03

Data Download Date/Time: 10/07/2021 19:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

Agency: FHP

Date/Time: 08/31/2021 20:44:26

Data Download Date/Time: 10/07/2021 19:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

Agency:

Date/Time: 03/18/2021 13:43:25

Data Download Date/Time: 04/02/2021 12:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 08/10/2021 14:05:01

Data Download Date/Time: 08/21/2021 02:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

Agency: FHP

Date/Time: 08/10/2021 14:05:56

Data Download Date/Time: 08/21/2021 02:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FHP

Date/Time: 08/10/2021 14:14:07

Data Download Date/Time: 08/21/2021 02:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

Agency: FHP

Date/Time: 08/10/2021 14:59:12

Data Download Date/Time: 08/21/2021 02:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 08/10/2021 15:02:11

Data Download Date/Time: 08/21/2021 02:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

Agency: FHP

Date/Time: 07/14/2021 16:35:14

Data Download Date/Time: 08/05/2021 11:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FHP

Date/Time: 07/14/2021 16:39:35

Data Download Date/Time: 08/05/2021 11:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

Agency: FHP

Date/Time: 07/14/2021 16:49:14

Data Download Date/Time: 08/05/2021 11:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

Agency:

Date/Time: 03/26/2021 10:16:35

Data Download Date/Time: 04/05/2021 16:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 10/01/2021 17:19:20

Data Download Date/Time: 10/07/2021 19:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

Agency: FHP

Date/Time: 10/01/2021 17:22:55

Data Download Date/Time: 10/07/2021 19:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

Agency:

Date/Time: 03/31/2021 08:46:31

Data Download Date/Time: 04/05/2021 17:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 10/01/2021 16:53:40

Data Download Date/Time: 10/07/2021 20:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

Agency: FHP

Date/Time: 10/01/2021 17:22:18

Data Download Date/Time: 10/07/2021 20:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 10/01/2021 17:30:14

Data Download Date/Time: 10/07/2021 20:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency: FHP

Date/Time: 07/07/2021 22:57:24

Data Download Date/Time: 08/25/2021 20:40

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 07/07/2021 22:58:23

Data Download Date/Time: 08/25/2021 20:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency: FHP

Date/Time: 07/08/2021 11:52:13

Data Download Date/Time: 08/25/2021 20:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

Agency: FHP

Date/Time: 08/30/2021 22:39:12

Data Download Date/Time: 09/02/2021 11:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

Agency: FHP

Date/Time: 09/21/2021 15:32:08

Data Download Date/Time: 01/05/2022 09:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FHP

Date/Time: 09/21/2021 15:34:11

Data Download Date/Time: 01/05/2022 09:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

Agency: FHP

Date/Time: 08/11/2021 11:00:14

Data Download Date/Time: 10/13/2021 21:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP

Date/Time: 08/11/2021 14:51:23

Data Download Date/Time: 10/13/2021 21:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

Agency: FHP

Date/Time: 09/14/2021 10:53:15

Data Download Date/Time: 09/27/2021 15:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

Agency: ORANGE CITY P.D.

Date/Time: 12/21/2021 04:59:19

Data Download Date/Time: 12/22/2021 16:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 10/31/2021 02:29:53

Data Download Date/Time: 11/22/2021 18:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 11/09/2021 00:56:54

Data Download Date/Time: 11/22/2021 18:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 11/09/2021 00:52:25

Data Download Date/Time: 11/22/2021 18:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK
