

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD  
**Inspector** SWINSON, CHRIS, J

**Date** 01/31/2022  
**Time** 10:24:24

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 06021080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 05/12/2022      **Expiration** 04/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/31/2022 10:38

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD  
**Inspector** SWINSON, CHRIS, J

**Date** 02/24/2022  
**Time** 13:30:54

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 06021080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 05/12/2022      **Expiration** 04/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/24/2022 13:32

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD      **Date** 03/31/2022      **Type** Agency Inspection  
**Inspector** SCHAUT,RICHARD,J      **Time** 15:24:06      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06021080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/01/2022 13:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | CORAL SPRINGS PD | <b>Date</b>       | 04/22/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SCHAUT,RICHARD,J | <b>Time</b>       | 18:20:03   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C          | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 06021080A2 |
| <b>Expiration</b>           | 01/12/2023       | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/05/2023 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 04/22/2022 18:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD  
**Inspector** SWINSON,CHRIS,J

**Date** 05/31/2022  
**Time** 13:37:12

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06021080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/31/2022 13:51

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | CORAL SPRINGS PD | <b>Date</b>       | 06/29/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SCHAUT,RICHARD,J | <b>Time</b>       | 18:23:58   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C          | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 06021080A2 |
| <b>Expiration</b>           | 01/12/2023       | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/05/2023 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/28/2022 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD  
**Inspector** SCHAUT,RICHARD,J

**Date** 07/28/2022  
**Time** 14:12:48

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06021080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/28/2022 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD  
**Inspector** SWINSON, CHRIS, J

**Date** 08/29/2022  
**Time** 14:06:10

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06021080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/29/2022 14:50

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD      **Date** 08/31/2022      **Type** Agency Inspection  
**Inspector** SCHAUT,RICHARD,J      **Time** 09:25:06      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06021080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/12/2022 11:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD      **Date** 09/12/2022      **Type** Department Inspection  
**Inspector** REYES-RIVERA,DAVID,E      **Time** 14:54:36      **Software** 8100.27

**0.05 Lot #** 202201C      **0.08 Lot #** 202201D      **0.20 Lot #** 202201E      **0.08 Gas Lot #** 00521080A2  
**Expiration** 01/11/2024      **Expiration** 01/18/2024      **Expiration** 01/18/2024      **Expiration** 02/05/2023

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.082  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.082  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.082  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.082  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0013 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 09/12/2022 15:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD      **Date** 09/21/2022      **Type** Agency Inspection  
**Inspector** SCHAUT,RICHARD,J      **Time** 11:30:36      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06021080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/23/2022 11:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | CORAL SPRINGS PD | <b>Date</b>       | 10/13/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SCHAUT,RICHARD,J | <b>Time</b>       | 15:09:26   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A          | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 06021080A2 |
| <b>Expiration</b>           | 12/07/2023       | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/05/2023 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 10/13/2022 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD      **Date** 11/08/2022      **Type** Agency Inspection  
**Inspector** SCHAUT,RICHARD,J      **Time** 15:21:54      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06021080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/08/2022 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

**Agency** CORAL SPRINGS PD      **Date** 12/27/2022      **Type** Agency Inspection  
**Inspector** SCHAUT,RICHARD,J      **Time** 11:03:07      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202206B      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06021080A2  
**Expiration** 12/07/2023      **Expiration** 06/14/2024      **Expiration** 06/22/2023      **Expiration** 04/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/27/2022 11:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

**Agency** ST PETERSBURG PD      **Date** 01/07/2022      **Type** Agency Inspection  
**Inspector** MCCULLOUGH,SEAN,J      **Time** 11:14:21      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 01/07/2022 11:31

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ST PETERSBURG PD  | <b>Date</b>       | 02/03/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MCCULLOUGH,SEAN,J | <b>Time</b>       | 10:35:00   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 08121080A1 |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 05/05/2023 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 02/03/2022 11:17

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

**Agency** NASSAU CO SO **Date** 01/21/2022 **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K **Time** 21:50:50 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2022

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/25/2022 15:42

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

**Agency** NASSAU CO SO **Date** 02/26/2022 **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K **Time** 11:12:09 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2022

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/01/2022 21:39

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

|                                  |                        |                                   |  |
|----------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> NASSAU CO SO       | <b>Date</b> 12/15/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> NICODEMO, PHIL, | <b>Time</b> 11:43:05   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG229803 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 10/25/2024   |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.077  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 12/15/2022 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

**Agency** NASSAU CO SO **Date** 12/21/2022 **Type** Agency Inspection  
**Inspector** KUTZ,THOMAS,E **Time** 18:27:12 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A LOOSE MOUTHPIECE. Non-compliance:LOOSE MOUTHPIECE.

**Data Download Date/Time** 12/24/2022 04:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

|                                |                              |                               |                                  |
|--------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> NASSAU CO SO     | <b>Date</b> 12/21/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> KUTZ,THOMAS,E | <b>Time</b> 19:25:29         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202112A      | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 26422080A1 |
| <b>Expiration</b> 12/07/2023   | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 11/05/2024     |
| <b>Number of Simulators</b> 4  | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/24/2022 04:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

**Agency** POLK COUNTY SO      **Date** 01/11/2022      **Type** Agency Inspection  
**Inspector** LUCAS,NATHANIEL,      **Time** 19:02:00      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/12/2022 08:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> POLK COUNTY SO      | <b>Date</b> 02/15/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUCAS,NATHANIEL, | <b>Time</b> 15:12:22         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C         | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 23621080A2 |
| <b>Expiration</b> 01/12/2023      | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2023     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/15/2022 16:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

**Agency** POLK COUNTY SO **Date** 03/21/2022 **Type** Agency Inspection  
**Inspector** LUCAS,NATHANIEL, **Time** 15:45:43 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

NO MOUTH ALCOHOL

**Data Download Date/Time** 03/22/2022 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

**Agency** POLK COUNTY SO      **Date** 03/21/2022      **Type** Agency Inspection  
**Inspector** LUCAS,NATHANIEL,      **Time** 16:35:47      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/22/2022 08:10

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

**Agency** POLK COUNTY SO **Date** 04/22/2022 **Type** Agency Inspection  
**Inspector** LUCAS,NATHANIEL, **Time** 14:44:21 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/22/2022 15:25

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

**Agency** POLK COUNTY SO      **Date** 05/11/2022      **Type** Agency Inspection  
**Inspector** JOHNSTON,CHRISTOPHER,T      **Time** 13:40:08      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/11/2022 16:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

**Agency** POLK COUNTY SO      **Date** 06/07/2022      **Type** Agency Inspection  
**Inspector** JOHNSTON,CHRISTOPHER,T      **Time** 11:56:31      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/07/2022 12:27

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

**Agency** POLK COUNTY SO      **Date** 07/15/2022      **Type** Agency Inspection  
**Inspector** LUCAS,NATHANIEL,      **Time** 10:38:41      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 07/15/2022 12:06

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

**Agency** POLK COUNTY SO      **Date** 08/08/2022      **Type** Agency Inspection  
**Inspector** JOHNSTON,CHRISTOPHER,T      **Time** 13:00:08      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/08/2022 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | POLK COUNTY SO   | <b>Date</b>       | 09/23/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LUCAS,NATHANIEL, | <b>Time</b>       | 10:04:42   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C          | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 23621080A2 |
| <b>Expiration</b>           | 01/12/2023       | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 11/05/2023 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/23/2022 10:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> POLK COUNTY SO       | <b>Date</b> 10/25/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 11:30:28   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0002 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 10/25/2022 12:22

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD Date 01/27/2022 Type Agency Inspection  
Inspector STRUNK,JOHN,D Time 10:31:09 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202108D 0.20 Lot # 202005A 0.08 Gas Lot # 1283438  
Expiration 01/12/2023 Expiration 08/17/2023 Expiration 05/12/2022 Expiration 02/25/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail-RESIDUAL FROM MA.

Data Download Date/Time 01/27/2022 10:33

Printed: 05 January 2023  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

**Agency** DAVIE PD  
**Inspector** STRUNK,JOHN,D

**Date** 02/24/2022  
**Time** 10:46:13

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1283438  
**Expiration** 01/13/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/25/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/24/2022 10:59

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD Date 03/28/2022 Type Agency Inspection  
Inspector STRUNK,JOHN,D Time 07:53:19 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 1283438  
Expiration 01/12/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 02/25/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/28/2022 08:04

Printed: 05 January 2023  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD Date 04/25/2022 Type Agency Inspection  
Inspector STRUNK,JOHN,D Time 07:51:43 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 1283438  
Expiration 01/12/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 02/25/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

DATE WAS INCORRECT

Data Download Date/Time 05/17/2022 08:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

**Agency** DAVIE PD  
**Inspector** STRUNK,JOHN,D

**Date** 05/31/2022  
**Time** 10:34:13

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/25/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/31/2022 10:46

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

**Agency** DAVIE PD  
**Inspector** STRUNK,JOHN,D

**Date** 06/29/2022  
**Time** 12:48:22

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/25/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

JUNE INSPECTION - OUT FOR ANNUAL

**Data Download Date/Time** 06/29/2022 14:32

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> DAVIE PD                  | <b>Date</b> 06/30/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 10:54:24   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 06/30/2022 11:10

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD Date 06/30/2022 Type Agency Inspection  
Inspector STRUNK,JOHN,D Time 14:21:49 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 1283438  
Expiration 01/12/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 02/25/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.BACK FROM ANNUAL INSPECTION

Data Download Date/Time 07/04/2022 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD Date 07/26/2022 Type Agency Inspection  
Inspector STRUNK,JOHN,D Time 10:16:48 Software 8100.27

0.05 Lot # 202112A 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 1283438  
Expiration 12/07/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 02/25/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 07/26/2022 15:05

Printed: 05 January 2023  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

**Agency** DAVIE PD  
**Inspector** STRUNK,JOHN,D

**Date** 08/29/2022  
**Time** 09:53:05

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1283438  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/25/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 08/29/2022 12:54

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD Date 09/19/2022 Type Agency Inspection  
Inspector STRUNK,JOHN,D Time 11:11:23 Software 8100.27

0.05 Lot # 202112A 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 1283438  
Expiration 12/07/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 02/25/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 09/19/2022 13:37

Printed: 05 January 2023  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

**Agency** DAVIE PD  
**Inspector** STRUNK,JOHN,D

**Date** 10/27/2022  
**Time** 08:52:42

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1283438  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/25/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 10/27/2022 10:16

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

**Agency** DAVIE PD  
**Inspector** STRUNK,JOHN,D

**Date** 11/28/2022  
**Time** 09:31:52

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1283438  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/25/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail-RESIDUAL FROM M/A.

**Data Download Date/Time** 11/28/2022 10:56

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD Date 12/26/2022 Type Agency Inspection  
Inspector STRUNK,JOHN,D Time 09:28:46 Software 8100.27

0.05 Lot # 202112A 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 1283438  
Expiration 12/07/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 02/25/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 12/26/2022 09:35

Printed: 05 January 2023  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

**Agency** LAKELAND PD **Date** 06/09/2022 **Type** Agency Inspection  
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:57:41 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT

**Data Download Date/Time** 06/09/2022 15:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> LAKELAND PD          | <b>Date</b> 06/09/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 14:35:37   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 06/09/2022 15:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> LAKELAND PD         | <b>Date</b> 06/27/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ALMEIDA,CAMILO,J | <b>Time</b> 13:02:25         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 2021112A        | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 11821080A2 |
| <b>Expiration</b> 12/07/2023      | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 06/05/2023     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/27/2022 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

**Agency** LAKELAND PD      **Date** 07/26/2022      **Type** Agency Inspection  
**Inspector** ALMEIDA,CAMILO,J      **Time** 18:32:55      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/26/2022 18:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> LAKELAND PD         | <b>Date</b> 08/24/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ALMEIDA,CAMILO,J | <b>Time</b> 13:40:43         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202112A         | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 11821080A2 |
| <b>Expiration</b> 12/07/2023      | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 06/05/2023     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/24/2022 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

**Agency** LAKELAND PD **Date** 09/14/2022 **Type** Agency Inspection  
**Inspector** ALMEIDA,CAMILO,J **Time** 11:00:49 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/15/2022 12:58

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

**Agency** LAKELAND PD **Date** 10/05/2022 **Type** Agency Inspection  
**Inspector** ALMEIDA,CAMILO,J **Time** 10:09:20 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/05/2022 10:56

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

**Agency** LAKELAND PD **Date** 11/09/2022 **Type** Agency Inspection  
**Inspector** ALMEIDA,CAMILO,J **Time** 13:25:49 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MA PROVIDED FOR AF ON ACCIDENT. AF PROVIDED, MET STANDARD

**Data Download Date/Time** 11/09/2022 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

**Agency** LAKELAND PD **Date** 12/07/2022 **Type** Agency Inspection  
**Inspector** ALMEIDA,CAMILO,J **Time** 12:56:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 11821080A2  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/07/2022 14:21

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

|                                    |                        |                               |
|------------------------------------|------------------------|-------------------------------|
| <b>Agency</b> POLK CSO             | <b>Date</b> 05/27/2022 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 08:49:50   | <b>Software</b> 8100.27       |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                       |
|-----------------------------|-----------------------|
| <b>Number of Simulators</b> | <b>Compliance</b> Yes |
|-----------------------------|-----------------------|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

**Data Download Date/Time** 05/27/2022 13:04

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> POLK CSO             | <b>Date</b> 05/27/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 11:46:57   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 05/27/2022 13:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

**Agency** FERNANDINA BEACH PD      **Date** 01/19/2022      **Type** Agency Inspection  
**Inspector** SMITH,JASON,E      **Time** 17:24:02      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 24620080A3  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 11/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/19/2022 17:54

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

**Agency** FERNANDINA BEACH PD      **Date** 02/23/2022      **Type** Agency Inspection  
**Inspector** SMITH,JASON,E      **Time** 11:31:28      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 24620080A3  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 11/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.REMOVE FROM USE FOR ANNUAL DEPT INSP

**Data Download Date/Time** 02/23/2022 11:34

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FERNANDINA BEACH PD | <b>Date</b> 03/29/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,     | <b>Time</b> 10:02:09   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/29/2022 10:03

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

**Agency** FERNANDINA BEACH PD      **Date** 04/06/2022      **Type** Agency Inspection  
**Inspector** SMITH,JASON,E      **Time** 13:02:13      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 24620080A3  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 11/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PLACED IN EVIDENTIARY USE

**Data Download Date/Time** 04/06/2022 13:05

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

**Agency** FERNANDINA BEACH PD      **Date** 05/09/2022      **Type** Agency Inspection  
**Inspector** SMITH,JASON,E      **Time** 11:51:24      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 24620080A3  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/09/2022 11:55

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

**Agency** FERNANDINA BEACH PD      **Date** 06/08/2022      **Type** Agency Inspection  
**Inspector** SMITH,JASON,E      **Time** 15:05:10      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 246200803  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/13/2022 13:21

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FERNANDINA BEACH PD | <b>Date</b> 06/14/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,     | <b>Time</b> 10:03:13   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 06/14/2022 10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD      **Date** 01/28/2022      **Type** Agency Inspection  
**Inspector** HEBSON,BRET,      **Time** 09:23:21      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration**      **Expiration**

**Number of Simulators** 2      **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.069  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1 | 0.072  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.048  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.072  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2 | 0.073  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.048  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Test # 3          | 0.072  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3 | 0.073  |
|                                             |        | Interferent Detect Test # 3   | INT    |

#### Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. OUTSIDE TOLERANCE FOR .08. FOUND MISSING O-RING ON SIMULATO.

**Data Download Date/Time** 01/28/2022 10:27

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> WINTER HAVEN PD | <b>Date</b> 01/28/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> HEBSON,BRET, | <b>Time</b> 14:54:11         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C     | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 08121080A1 |
| <b>Expiration</b> 01/12/2023  | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 05/05/2023     |
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.073  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

TIGHTEN SEALS AND WAITED 10 MIN

**Data Download Date/Time** 01/28/2022 14:56

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD  
**Inspector** HEBSON, BRET,

**Date** 02/28/2022  
**Time** 07:21:18

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/28/2022 10:01

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD  
**Inspector** HEBSON, BRET,

**Date** 03/24/2022  
**Time** 14:24:11

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/24/2022 14:39

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD      **Date** 04/29/2022      **Type** Agency Inspection  
**Inspector** HEBSON,BRET,      **Time** 14:20:45      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/29/2022 15:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD  
**Inspector** HEBSON, BRET,

**Date** 05/31/2022  
**Time** 07:15:24

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/31/2022 08:07

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD  
**Inspector** HEBSON, BRET,

**Date** 06/29/2022  
**Time** 08:54:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 06/29/2022 09:59

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> WINTER HAVEN PD        | <b>Date</b> 07/05/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW, TAYLOR, D | <b>Time</b> 15:43:56   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 07/05/2022 15:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD      **Date** 08/01/2022      **Type** Agency Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 09:58:58      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. AI NOT CONDUCTED.

**Data Download Date/Time** 08/01/2022 13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> WINTER HAVEN PD      | <b>Date</b> 08/01/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 12:58:20   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0005 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.083  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0013 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 08/01/2022 13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD      **Date** 08/08/2022      **Type** Agency Inspection  
**Inspector** WOMACK,ERIC,V      **Time** 15:58:55      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 4      **Compliance** No

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 2       | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 2 | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 3       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 3 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     |                             |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                             |        |
| Alcohol Free Test: OK                       | Yes    |                             |        |
| 0.05 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Test: OK                        | No     |                             |        |
| 0.20 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

Int Det:INCORRECT CONNECTION. Non-compliance:RESTART INSPECTION.

**Data Download Date/Time** 08/09/2022 08:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD  
**Inspector** WOMACK,ERIC,V

**Date** 08/08/2022  
**Time** 16:44:29

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/09/2022 08:05

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD  
**Inspector** WOMACK, ERIC,

**Date** 09/30/2022  
**Time** 09:21:22

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: SUBJECT WASHED MOUTH OUT.

**Data Download Date/Time** 09/30/2022 10:25

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD      **Date** 10/27/2022      **Type** Agency Inspection  
**Inspector** WOMACK,ERIC,      **Time** 10:49:33      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202206B      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 12/07/2023      **Expiration** 06/14/2024      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/27/2022 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD      **Date** 11/30/2022      **Type** Agency Inspection  
**Inspector** WOMACK,ERIC,      **Time** 14:40:34      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 08121080A1  
**Expiration** 12/07/2023      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 05/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/30/2022 15:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

**Agency** WINTER HAVEN PD      **Date** 12/29/2022      **Type** Agency Inspection  
**Inspector** WOMACK,ERIC,      **Time** 15:10:45      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 08121080A1  
**Expiration** 12/07/2023      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 05/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/29/2022 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> GULFPORT PD          | <b>Date</b> 06/28/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 11:08:36   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 06/28/2022 11:13

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

**Agency** GULFPORT PD **Date** 07/26/2022 **Type** Agency Inspection  
**Inspector** MAROTTA, MICHAEL, S **Time** 22:03:58 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 22521080A1  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail.

**Data Download Date/Time** 07/29/2022 09:23

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | GULFPORT PD       | <b>Date</b>       | 08/09/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MAROTTA,MICHAEL,S | <b>Time</b>       | 19:50:37   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 22521080A1 |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 10/05/2023 |
| <b>Number of Simulators</b> | 4                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/09/2022 20:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

**Agency** GULFPORT PD      **Date** 09/16/2022      **Type** Agency Inspection  
**Inspector** MAROTTA, MICHAEL, S      **Time** 21:50:39      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 22521080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 10/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/16/2022 21:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

**Agency** GULFPORT PD      **Date** 10/24/2022      **Type** Agency Inspection  
**Inspector** MAROTTA, MICHAEL, S      **Time** 02:49:22      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 22521080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 10/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/24/2022 17:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | GULFPORT PD       | <b>Date</b>       | 11/08/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MAROTTA,MICHAEL,S | <b>Time</b>       | 00:59:26   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 22521080A1 |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 10/05/2023 |
| <b>Number of Simulators</b> | 4                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.199  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.080  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.081  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.079  |
|                                             |        | Interferent Detect Test # 3          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

#### Remarks:

Int Det:SIM LID TIGHTENED.

Data Download Date/Time 11/09/2022 08:30

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

**Agency** GULFPORT PD      **Date** 12/23/2022      **Type** Agency Inspection  
**Inspector** MAROTTA, MICHAEL, S      **Time** 02:46:20      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 22521080A1  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 10/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/23/2022 19:42

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 01/05/2022  
**Time** 10:22:19

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1273831  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/31/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 01/05/2022 10:23

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 02/02/2022  
**Time** 11:20:38

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1273831  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/31/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO A. FRAIZER WAS PRESENT.

**Data Download Date/Time** 02/02/2022 11:22

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD      **Date** 02/02/2022      **Type** Agency Inspection  
**Inspector** DERIET,RAFAEL,      **Time** 10:41:43      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                  | Result |
|---------------------------------------------|--------|-----------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1 | 0.000  |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # 2 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 3 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        |                       |        |
| Alcohol Free Subject Test: 0.000            | Yes    |                       |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |                       |        |
| Interferent Detect Test: Interferent Detect | No     |                       |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                       |        |
| Alcohol Free Test: OK                       | Yes    |                       |        |
| 0.05 g/210L Test: OK                        | No     |                       |        |
| 0.08 g/210L Test: OK                        | No     |                       |        |
| 0.20 g/210L Test: OK                        | No     |                       |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                       |        |

#### Remarks:

UNK

**Data Download Date/Time** 02/02/2022 11:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 03/02/2022  
**Time** 11:27:32

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1273831  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/31/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO FRAZIER WAS PRESENT.

**Data Download Date/Time** 03/02/2022 12:08

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 04/13/2022  
**Time** 11:11:44

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1273831  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/31/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 04/13/2022 11:16

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 05/25/2022  
**Time** 10:59:50

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1273831  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/31/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 05/25/2022 11:52

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 06/22/2022  
**Time** 12:24:40

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402290667  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/07/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.191  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.191  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 06/22/2022 14:19

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 07/06/2022  
**Time** 12:42:11

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402290667  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/07/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO B. ARMSTRONG WAS PRESENT.

**Data Download Date/Time** 07/06/2022 13:14

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 08/03/2022  
**Time** 12:36:36

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402290667  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/07/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/03/2022 13:01

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

|                                       |                        |                                   |  |
|---------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> HOLLYWOOD PD            | <b>Date</b> 08/04/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> REYES-RIVERA,DAVID,E | <b>Time</b> 10:51:31   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

#### Remarks:

Data Download Date/Time 08/04/2022 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 08/05/2022  
**Time** 12:46:09

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402290677  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/07/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A:PERFORMED MOUTH ALCOHOL FIRST.

**Data Download Date/Time** 08/05/2022 13:01

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 09/01/2022  
**Time** 11:31:48

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402290677  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/07/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 09/01/2022 16:26

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 10/05/2022  
**Time** 11:03:45

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402290677  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/07/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail UNKNOWN REASON. TURNED FAN ON NO INCIDENTS

**Data Download Date/Time** 10/05/2022 11:12

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 11/02/2022  
**Time** 11:28:32

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402290677  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/07/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.191  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.191  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 11/02/2022 11:38

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

**Agency** HOLLYWOOD PD  
**Inspector** DERIET, RAFAEL,

**Date** 12/14/2022  
**Time** 11:07:39

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402290677  
**Expiration** 12/07/2023      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 12/07/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

BSO A. FRAZIER WAS PRESENT.

**Data Download Date/Time** 12/14/2022 12:44

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | NEW PORT RICHEY PD | <b>Date</b>       | 01/09/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | ENGEL,JASON,S      | <b>Time</b>       | 09:30:40   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C            | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 09820080A3 |
| <b>Expiration</b>           | 01/12/2023         | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 06/05/2022 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CHECKED TUBING

Data Download Date/Time 01/13/2022 10:36

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

**Agency** NEW PORT RICHEY PD      **Date** 02/06/2022      **Type** Agency Inspection  
**Inspector** ENGEL,JASON,S      **Time** 10:07:20      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 09820080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WIPED DOWN INSTRUMENT

**Data Download Date/Time** 02/06/2022 10:22

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

**Agency** NEW PORT RICHEY PD      **Date** 03/06/2022      **Type** Agency Inspection  
**Inspector** ENGEL,JASON,S      **Time** 10:19:00      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 09820080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CHECKED O-RINGS

**Data Download Date/Time** 03/06/2022 10:25

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

**Agency** NEW PORT RICHEY PD      **Date** 04/03/2022      **Type** Agency Inspection  
**Inspector** ENGEL,JASON,S      **Time** 08:13:26      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 4      **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Range Exceeded. Non-compliance: .

**Data Download Date/Time** 04/03/2022 09:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

**Agency** NEW PORT RICHEY PD      **Date** 04/03/2022      **Type** Agency Inspection  
**Inspector** ENGEL,JASON,S      **Time** 09:24:32      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 09820080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

2ND INSPECTION, 1ST USED TOO MUCH MOUTH ALC.

**Data Download Date/Time** 04/03/2022 09:28

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

**Agency** NEW PORT RICHEY PD      **Date** 05/14/2022      **Type** Agency Inspection  
**Inspector** ENGEL,JASON,S      **Time** 11:09:51      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 09820080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WIPED DOWN INS.

**Data Download Date/Time** 05/14/2022 11:35

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | NEW PORT RICHEY PD | <b>Date</b>       | 06/12/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | ENGEL,JASON,S      | <b>Time</b>       | 08:44:53   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A            | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 06022080A3 |
| <b>Expiration</b>           | 12/07/2023         | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CHECKED TUBING

Data Download Date/Time 06/15/2022 05:54

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | NEW PORT RICHEY PD | <b>Date</b>       | 07/19/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | ENGEL,JASON,S      | <b>Time</b>       | 00:11:58   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A            | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 06022080A3 |
| <b>Expiration</b>           | 12/07/2023         | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

REPLACED TUBING ORING AND RETURN PORT ORING

**Data Download Date/Time** 07/19/2022 00:27

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | NEW PORT RICHEY PD | <b>Date</b>       | 08/24/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | ENGEL,JASON,S      | <b>Time</b>       | 20:42:35   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A            | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 06022080A3 |
| <b>Expiration</b>           | 12/07/2023         | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WIPED DOWN INS.

**Data Download Date/Time** 09/01/2022 14:27

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> NEW PORT RICHEY PD   | <b>Date</b> 09/14/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 15:42:29   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 09/14/2022 15:45

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

**Agency** NEW PORT RICHEY PD      **Date** 09/18/2022      **Type** Agency Inspection  
**Inspector** ENGEL,JASON,S      **Time** 22:12:11      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06022080A3  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CHECKED TUBING

**Data Download Date/Time** 10/14/2022 13:18

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

**Agency** NEW PORT RICHEY PD      **Date** 10/19/2022      **Type** Agency Inspection  
**Inspector** ENGEL,JASON,S      **Time** 09:27:19      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06022080A3  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WIPED DOWN INSTRUMENT

**Data Download Date/Time** 10/19/2022 09:37

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

**Agency** NEW PORT RICHEY PD      **Date** 11/03/2022      **Type** Agency Inspection  
**Inspector** ENGEL,JASON,S      **Time** 10:24:50      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 06022080A3  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/05/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CHECKED ORINGS

**Data Download Date/Time** 11/03/2022 10:42

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | NEW PORT RICHEY PD | <b>Date</b>       | 12/11/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HAZIME,A,          | <b>Time</b>       | 10:16:16   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A            | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 06022080A3 |
| <b>Expiration</b>           | 12/07/2023         | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 04/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CLEANED INSTRUMENT; INSPECTED FOR CRACKS; ALL GOOD

**Data Download Date/Time** 12/14/2022 14:10

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 02/09/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 09:24:35      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 00521080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CLEANED AREA

**Data Download Date/Time** 02/09/2022 09:27

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 03/03/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 08:31:31      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AREA CLEANED

**Data Download Date/Time** 03/03/2022 08:35

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 04/29/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 14:00:44      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

Non-compliance:TIME OUT OF ADJUSTMENT ERROR ADJUSTING TIME.

**Data Download Date/Time** 05/02/2022 13:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 04/29/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 14:36:00      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.

**Data Download Date/Time** 05/02/2022 13:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 05/01/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 11:19:04      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CLEANED AREA

**Data Download Date/Time** 05/02/2022 13:41

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 06/08/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 16:29:04      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/09/2022 08:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

|                                 |                        |                                   |  |
|---------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> TARPON SPRINGS PD | <b>Date</b> 08/12/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> SOTO, ISRAEL,  | <b>Time</b> 12:35:37   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 08121080A1 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 08/12/2022 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 08/18/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 13:54:37      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Improper Sample. Non-compliance: IMPROPER SAMPE STARTED INSPECTION OVER.

**Data Download Date/Time** 08/18/2022 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 08/18/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 14:29:36      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail MOUTH ALCOHOL SAMPLE LEFT OPEN NEAR INST. .

**Data Download Date/Time** 08/18/2022 14:33

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 09/29/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 09:49:43      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CLEANED AREA

**Data Download Date/Time** 09/30/2022 12:59

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 10/12/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 09:18:50      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CLEANED AREA

**Data Download Date/Time** 10/17/2022 11:31

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 11/30/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 06:23:37      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AREA CLEANED

**Data Download Date/Time** 12/02/2022 09:18

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

**Agency** TARPON SPRINGS PD      **Date** 12/02/2022      **Type** Agency Inspection  
**Inspector** GASSEN,STEVE,      **Time** 08:54:33      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 00521080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CLEANED AREA

**Data Download Date/Time** 12/02/2022 09:18

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO     | <b>Date</b> 05/17/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW, TAYLOR, D | <b>Time</b> 11:44:59   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 05/17/2022 11:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 06/20/2022  
**Time** 10:08:06

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** AG115904  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/08/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/29/2022 10:35

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 07/13/2022  
**Time** 22:00:45

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** AG115904  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/08/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/15/2022 19:23

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 08/10/2022  
**Time** 20:10:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** AG115904  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/08/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/12/2022 20:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 09/21/2022  
**Time** 23:48:02

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** AG115904  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/08/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/22/2022 21:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 10/20/2022  
**Time** 01:52:55

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** AG115904  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/08/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/20/2022 02:00

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 11/03/2022  
**Time** 20:05:42

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** AG115904  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/08/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/04/2022 19:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 12/09/2022  
**Time** 20:37:02

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** AG115904  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/08/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.084  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/14/2022 11:26

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

**Agency** CITY OF MIRAMAR POL **Date** 11/15/2022 **Type** Agency Inspection  
**Inspector** REYES-RIVERA, DAVID, E **Time** 06:33:34 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

**Data Download Date/Time** 11/15/2022 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

|                                         |                        |                                   |  |
|-----------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> CITY OF MIRAMAR POL       | <b>Date</b> 11/15/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 08:58:18   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0001 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.077  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.077  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.077  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.077  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.077  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.076  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.077  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.076  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.077  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 11/15/2022 09:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

**Agency** CITY OF MIRAMAR POL **Date** 12/14/2022 **Type** Agency Inspection  
**Inspector** HOLLOWAY,DEVARIOUS,J **Time** 11:09:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1273831  
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/14/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A

**Data Download Date/Time** 12/14/2022 13:10

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD      **Date** 03/29/2022      **Type** Agency Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 09:49:26      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 03/29/2022 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> CLEARWATER PD        | <b>Date</b> 03/29/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 14:06:30   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/29/2022 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD      **Date** 05/26/2022      **Type** Agency Inspection  
**Inspector** HURT,JONATHAN,R      **Time** 01:21:03      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 35521080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 03/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

ACETONE LOT# 2021-C, M/A LOT # 2021-D

**Data Download Date/Time** 05/26/2022 01:23

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD      **Date** 06/18/2022      **Type** Agency Inspection  
**Inspector** HURT,JONATHAN,R      **Time** 00:29:15      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 35521080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 03/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

M/A LOT #2021-D, ACETONE LOT # 2021-C

**Data Download Date/Time** 06/18/2022 00:34

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD      **Date** 07/25/2022      **Type** Agency Inspection  
**Inspector** HURT,JONATHAN,R      **Time** 11:42:48      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 35521080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 03/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

ACETONE LOT # 2021-C, M/A LOT # 2021-D

**Data Download Date/Time** 07/27/2022 18:29

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> CLEARWATER PD        | <b>Date</b> 07/29/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 12:00:58   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 07/29/2022 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD      **Date** 08/08/2022      **Type** Agency Inspection  
**Inspector** BUIS,JUSTIN,M      **Time** 22:19:29      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 35521080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 03/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

ACETONE LOT # - 2021C M/A LOT # 2021D

**Data Download Date/Time** 08/08/2022 22:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD  
**Inspector** HURT, JONATHAN, R

**Date** 09/30/2022  
**Time** 20:55:47

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 35521080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 03/05/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

ACETONE LOT#2021C, M/A LOT#2021D

**Data Download Date/Time** 09/30/2022 21:03

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD  
**Inspector** HURT, JONATHAN, R

**Date** 10/30/2022  
**Time** 17:42:11

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 35521080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 03/05/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

M/A LOT #2021-D, ACETONE LOT #2021-C

**Data Download Date/Time** 10/30/2022 17:59

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD  
**Inspector** BUIS,JUSTIN,M

**Date** 11/28/2022  
**Time** 19:47:41

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 07622080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

M/A 2021-D ACETONE 2021-C

**Data Download Date/Time** 11/29/2022 23:19

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

**Agency** CLEARWATER PD  
**Inspector** BUIS,JUSTIN,M

**Date** 12/22/2022  
**Time** 16:13:09

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202206B      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 07622080A2  
**Expiration** 12/07/2023      **Expiration** 06/14/2024      **Expiration** 06/22/2023      **Expiration** 05/05/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

M/A 2021-D ACETONE 2021-C

**Data Download Date/Time** 12/22/2022 16:24

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO      **Date** 01/13/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 11:29:23      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/13/2022 11:34

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO **Date** 02/05/2022 **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D **Time** 15:18:24 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1 | 0.081  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.051  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.080  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2 | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.051  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Interferent Detect Test # 3   | INT    |

#### Remarks:

08: Ambient Fail, Ambient Fail.RMVD FROM EVED USE AT PTDF Non-compliance:AMB FAIL RETURNING TO CMI FOR INSPECTION.

**Data Download Date/Time** 02/05/2022 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> JACKSONVILLE SO | <b>Date</b> 06/08/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 12:45:38   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201E    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.082  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.082  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.052  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.206  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.082  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

#### Remarks:

Data Download Date/Time 06/08/2022 13:01

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO      **Date** 06/22/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 10:35:36      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 29821080A4  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 12/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PLCD INTO EVED USE AT PTDF 6-22-2022

**Data Download Date/Time** 06/22/2022 11:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

|                                  |                        |                                   |  |
|----------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO    | <b>Date</b> 09/02/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> NICODEMO, PHIL, | <b>Time</b> 11:28:47   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 08121080A1 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 09/02/2022 11:36

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO      **Date** 09/15/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 09:33:24      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PLACED INTO EVED USE PTDF 09/15/2022

**Data Download Date/Time** 09/16/2022 13:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO      **Date** 10/13/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 14:16:24      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/13/2022 14:36

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO      **Date** 10/14/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 11:45:12      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

REINSPECD AFTER CLN/UP OF BUILD MAINT ISSUE 10-14-2023

**Data Download Date/Time** 10/14/2022 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO      **Date** 10/26/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 11:23:47      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/26/2022 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO      **Date** 11/16/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 09:23:43      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 07622080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/16/2022 09:43

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

**Agency** JACKSONVILLE SO      **Date** 12/09/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 12:21:34      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 07622080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

REMOVED FROM EVIDENCE USE AT PTDF12-9-2022

**Data Download Date/Time** 12/12/2022 13:30

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 03/03/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LAMONT,GREGORY,S  | <b>Time</b>       | 14:06:51   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1433248    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 01/06/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.197  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.079  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.050  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.081  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.079  |
|                                             |        | Interferent Detect Test # 3          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

#### Remarks:

NO INT DET.

**Data Download Date/Time** 03/07/2022 11:31

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

**Agency** FL HIGHWAY PATROL **Date** 04/29/2022 **Type** Agency Inspection  
**Inspector** LAMONT,GREGORY,S **Time** 15:31:45 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 2       | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 2 | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 3       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 3 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     |                             |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                             |        |
| Alcohol Free Test: OK                       | Yes    |                             |        |
| 0.05 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Test: OK                        | No     |                             |        |
| 0.20 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

NO INT DETNO INT DET.

**Data Download Date/Time** 05/13/2022 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

**Agency** FL HIGHWAY PATROL      **Date** 04/29/2022      **Type** Agency Inspection  
**Inspector** LAMONT,GREGORY,S      **Time** 16:52:35      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1433248  
**Expiration** 01/22/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/06/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | No     | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/13/2022 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

**Agency** FL HIGHWAY PATROL      **Date** 05/13/2022      **Type** Agency Inspection  
**Inspector** LAMONT,GREGORY,S      **Time** 14:06:43      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/06/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/13/2022 14:28

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 06/14/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LAMONT,GREGORY,S  | <b>Time</b>       | 12:59:44   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1433248    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 01/06/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/14/2022 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 06/21/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,   | <b>Time</b> 12:01:30   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 06/21/2022 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 07/29/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LAMONT,GREGORY,S  | <b>Time</b>       | 14:07:00   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1433248    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 01/06/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/05/2022 11:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

**Agency** FL HIGHWAY PATROL **Date** 08/02/2022 **Type** Agency Inspection  
**Inspector** LAMONT,GREGORY,S **Time** 16:13:46 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: .

**Data Download Date/Time** 08/05/2022 11:50

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 09/30/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LAMONT,GREGORY,S  | <b>Time</b>       | 15:50:00   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1433248    |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 01/06/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/01/2022 12:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

**Agency** FL HIGHWAY PATROL **Date** 10/01/2022 **Type** Agency Inspection  
**Inspector** LAMONT,GREGORY,S **Time** 12:32:02 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433248  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/01/2022 12:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

**Agency** FL HIGHWAY PATROL **Date** 12/29/2022 **Type** Agency Inspection  
**Inspector** LAMONT,GREGORY,S **Time** 11:09:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248  
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: .

**Data Download Date/Time** 12/29/2022 12:17

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

**Agency** CLAY COUNTY SO      **Date** 01/27/2022      **Type** Agency Inspection  
**Inspector** PATRONE,RICHARD,L      **Time** 19:17:00      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 22620080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSP RETURN FOR ANNUAL INSP

**Data Download Date/Time** 01/27/2022 19:29

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD      **Date** 01/24/2022      **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN,      **Time** 16:18:23      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 25721080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/24/2022 16:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD      **Date** 02/21/2022      **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN,      **Time** 09:31:08      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 25721080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/24/2022 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD **Date** 03/29/2022 **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN, **Time** 16:40:22 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 25721080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail TO CLOSE TO MACHINE AFTER MOUTH ALCOHOL TE.

**Data Download Date/Time** 04/07/2022 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD      **Date** 04/20/2022      **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN,      **Time** 16:51:32      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 25721080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

OUT FOR DEPARTMENT INSPECTION

**Data Download Date/Time** 04/20/2022 16:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> DADE CITY PD         | <b>Date</b> 05/03/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 15:32:16   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.181  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | INT    |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.196  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # (Repeat) 4    | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # (Repeat) 5    | 0.195  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

|                                        |        |
|----------------------------------------|--------|
| 0.08 g/210L Dry Gas Std Test # 5       | 0.080  |
| Interferent Detect Test # 5            | INT    |
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # (Repeat) 6          | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # (Repeat) 7          | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # (Repeat) 8          | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # (Repeat) 9          | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # (Repeat) 10         | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev) (Repeat)    | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

20: Interferent Detect..

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Data Download Date/Time** 05/03/2022 15:37

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD      **Date** 05/10/2022      **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN,      **Time** 16:50:25      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 25721080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/12/2022 16:16

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD      **Date** 06/20/2022      **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN,      **Time** 09:27:49      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 25721080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/23/2022 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD      **Date** 07/15/2022      **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN,      **Time** 10:19:47      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 25721080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/15/2022 10:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD      **Date** 08/16/2022      **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN,      **Time** 16:21:47      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 25721080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: FORGOT MOUTH ALCOHOL.

**Data Download Date/Time** 08/16/2022 16:34

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

**Agency** DADE CITY PD      **Date** 09/15/2022      **Type** Agency Inspection  
**Inspector** BALDWIN, DEAN,      **Time** 16:00:47      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 25721080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/15/2022 16:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

|                                 |                              |                               |                                  |
|---------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> DADE CITY PD      | <b>Date</b> 10/18/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> BALDWIN, DEAN, | <b>Time</b> 16:51:17         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202112A       | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 25721080A2 |
| <b>Expiration</b> 12/07/2023    | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 11/05/2023     |
| <b>Number of Simulators</b> 4   | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/18/2022 16:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 01/07/2022  
**Time** 10:25:58

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1               | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2               | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08: RFI Detect.

**Data Download Date/Time** 01/07/2022 16:03

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 02/23/2022  
**Time** 19:48:25

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/24/2022 18:12

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 03/07/2022  
**Time** 05:36:28

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/07/2022 14:34

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PASCO COUNTY SO        | <b>Date</b> 03/07/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW, TAYLOR, D | <b>Time</b> 14:23:58   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 03/07/2022 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 03/07/2022  
**Time** 20:41:37

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/07/2022 20:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 04/01/2022  
**Time** 12:29:53

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/01/2022 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 05/16/2022  
**Time** 14:01:10

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/16/2022 14:02

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 06/23/2022  
**Time** 10:06:50

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/23/2022 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 07/19/2022  
**Time** 10:51:53

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/19/2022 10:53

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 08/22/2022  
**Time** 11:36:02

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/22/2022 11:39

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 09/27/2022  
**Time** 10:49:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/27/2022 10:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 10/14/2022  
**Time** 10:56:08

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/14/2022 10:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 11/29/2022  
**Time** 19:40:45

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.199  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.050  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.079  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08: Control Outside Tolerance HOSE NOT PLUGGED IN.

**Data Download Date/Time** 12/07/2022 18:16

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 12/03/2022  
**Time** 22:07:56

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/07/2022 18:16

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

**Agency** FHP TROOP E MIAMI **Date** 11/16/2022 **Type** Agency Inspection  
**Inspector** REYES-RIVERA, DAVID, E **Time** 09:09:07 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

**Data Download Date/Time** 11/16/2022 12:18

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

|                                         |                        |                                   |  |
|-----------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> FHP TROOP E MIAMI         | <b>Date</b> 11/16/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:12:28   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0005 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0015 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 11/16/2022 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

**Agency** ZEPHYRHILLS PD      **Date** 01/31/2022      **Type** Agency Inspection  
**Inspector** STARNES,AILISA,J      **Time** 09:22:00      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1392108OA2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/07/2022 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

**Agency** ZEPHYRHILLS PD      **Date** 11/02/2022      **Type** Agency Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 11:32:02      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED TO OPERATE INSTRUMENT. AI NOT CONDUCTED.

**Data Download Date/Time** 11/02/2022 15:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

|                             |                   |                                   |            |                   |                       |                       |            |
|-----------------------------|-------------------|-----------------------------------|------------|-------------------|-----------------------|-----------------------|------------|
| <b>Agency</b>               | ZEPHYRHILLS PD    | <b>Date</b>                       | 11/02/2022 | <b>Type</b>       | Department Inspection |                       |            |
| <b>Inspector</b>            | GUTSCHOW,TAYLOR,D | <b>Time</b>                       | 15:06:51   | <b>Software</b>   | 8100.27               |                       |            |
| <b>0.05 Lot #</b>           | 202201C           | <b>0.08 Lot #</b>                 | 202201D    | <b>0.20 Lot #</b> | 202201E               | <b>0.08 Gas Lot #</b> | 00521080A2 |
| <b>Expiration</b>           | 01/11/2024        | <b>Expiration</b>                 | 01/18/2024 | <b>Expiration</b> | 01/18/2024            | <b>Expiration</b>     | 02/05/2023 |
| <b>Number of Simulators</b> | 5                 | <b>Average Standard Deviation</b> | 0.0003     | <b>Compliance</b> | Yes                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 11/02/2022 15:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

**Agency** ZEPHYRHILLS PD  
**Inspector** STARNES,AILISA,J

**Date** 11/18/2022  
**Time** 12:13:01

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 13921080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 07/05/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/18/2022 12:20

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

**Agency** ZEPHYRHILLS PD  
**Inspector** STARNES,AILISA,J

**Date** 12/28/2022  
**Time** 09:13:49

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 13921080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 07/05/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

COUPLE AMB\*, PUR\*, RGE\*

**Data Download Date/Time** 12/28/2022 09:48

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

|                                       |                        |                                   |  |
|---------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> HARDEE COUNTY SO        | <b>Date</b> 08/02/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> REYES-RIVERA,DAVID,E | <b>Time</b> 10:21:07   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 08/02/2022 10:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

**Agency** HARDEE COUNTY SO **Date** 08/26/2022 **Type** Agency Inspection  
**Inspector** CARTWRIGHT JR,GARY,W **Time** 08:27:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477283  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/26/2022 09:56

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

**Agency** HARDEE COUNTY SO **Date** 09/24/2022 **Type** Agency Inspection  
**Inspector** CARTWRIGHT JR,GARY,W **Time** 11:25:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477283  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/24/2022 11:35

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

**Agency** HARDEE COUNTY SO **Date** 10/31/2022 **Type** Agency Inspection  
**Inspector** CARTWRIGHT JR,GARY,W **Time** 07:56:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477283  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.189  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.196  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20: Control Outside Tolerance.

**Data Download Date/Time** 10/31/2022 14:44

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

**Agency** HARDEE COUNTY SO **Date** 11/07/2022 **Type** Agency Inspection  
**Inspector** CARTWRIGHT JR,GARY,W **Time** 10:33:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477283  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/07/2022 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

**Agency** HARDEE COUNTY SO **Date** 12/30/2022 **Type** Agency Inspection  
**Inspector** CARTWRIGHT JR,GARY,W **Time** 11:55:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283  
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/03/2023 09:25

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ST PETERSBURG PD  | <b>Date</b>       | 01/07/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MCCULLOUGH,SEAN,J | <b>Time</b>       | 11:20:49   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 24620080A2 |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 11/05/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

Data Download Date/Time 01/07/2022 11:45

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ST PETERSBURG PD  | <b>Date</b>       | 02/03/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MCCULLOUGH,SEAN,J | <b>Time</b>       | 10:28:16   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 24620080A2 |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 11/05/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

Data Download Date/Time 02/03/2022 10:48

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

**Agency** ST PETERSBURG PD      **Date** 03/08/2022      **Type** Agency Inspection  
**Inspector** MCCULLOUGH,SEAN,J      **Time** 14:10:25      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 24620080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 03/08/2022 15:24

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> LARGO PD             | <b>Date</b> 01/05/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:54:55   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/05/2022 13:59

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD      **Date** 01/10/2022      **Type** Agency Inspection  
**Inspector** THORNSBROUGH,DREW,A      **Time** 10:32:18      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 07220080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 05/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/10/2022 11:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD **Date** 02/23/2022 **Type** Agency Inspection  
**Inspector** HASTINGS, KENNETH, **Time** 15:22:38 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 07220080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/23/2022 15:26

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

## Inspection Test Electronic Data

Printed: 05 January 2023  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD **Date** 04/06/2022 **Type** Agency Inspection  
**Inspector** THORNSBROUGH,DREW,A **Time** 08:45:10 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/06/2022 10:03

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD      **Date** 05/23/2022      **Type** Agency Inspection  
**Inspector** THORNSBROUGH,DREW,A      **Time** 10:38:28      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/23/2022 11:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD **Date** 06/20/2022 **Type** Agency Inspection  
**Inspector** HASTINGS, KENNETH, **Time** 14:46:50 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/20/2022 14:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD **Date** 07/06/2022 **Type** Agency Inspection  
**Inspector** THORNSBROUGH,DREW,A **Time** 08:22:03 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A:.204 MA.

**Data Download Date/Time** 07/06/2022 09:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD      **Date** 07/06/2022      **Type** Agency Inspection  
**Inspector** THORNSBROUGH,DREW,A      **Time** 09:20:09      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/06/2022 09:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD **Date** 08/17/2022 **Type** Agency Inspection  
**Inspector** HASTINGS, KENNETH, **Time** 10:34:31 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/17/2022 10:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD      **Date** 09/15/2022      **Type** Agency Inspection  
**Inspector** THORNSBROUGH,DREW,A      **Time** 08:51:37      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/15/2022 09:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD **Date** 10/12/2022 **Type** Agency Inspection  
**Inspector** HASTINGS, KENNETH, **Time** 12:20:00 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/12/2022 12:36

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD **Date** 11/10/2022 **Type** Agency Inspection  
**Inspector** THORNSBROUGH,DREW,A **Time** 12:52:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/10/2022 13:01

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

**Agency** LARGO PD **Date** 12/01/2022 **Type** Agency Inspection  
**Inspector** HASTINGS, KENNETH, **Time** 10:37:02 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/01/2022 10:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> POLK CSO             | <b>Date</b> 03/07/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 15:26:06   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0006      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0010 |

#### Remarks:

Data Download Date/Time 03/07/2022 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO Date 03/21/2022 Type Agency Inspection  
Inspector COKE,CLEVELAND,A Time 16:24:21 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 13921080A2  
Expiration 01/12/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 07/05/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: .NO MOUTH ALCOHOL

Data Download Date/Time 03/22/2022 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> POLK CSO            | <b>Date</b> 04/22/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> COKE,CLEVELAND,A | <b>Time</b> 12:58:35         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C         | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 13921080A2 |
| <b>Expiration</b> 01/12/2023      | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 07/05/2023     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO REMARKS

**Data Download Date/Time** 04/22/2022 15:07

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> POLK CSO            | <b>Date</b> 05/11/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> COKE,CLEVELAND,A | <b>Time</b> 10:19:28         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C         | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 13921080A2 |
| <b>Expiration</b> 01/12/2023      | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 07/05/2023     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.075  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Control Outside Tolerance.

**Data Download Date/Time** 05/11/2022 12:35

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

|                                    |                                          |                                   |                                  |
|------------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> POLK CSO             | <b>Date</b> 10/31/2022                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 12:29:25                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202201C          | <b>0.08 Lot #</b> 202201D                | <b>0.20 Lot #</b> 202201E         | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024       | <b>Expiration</b> 01/18/2024             | <b>Expiration</b> 01/18/2024      | <b>Expiration</b> 02/05/2023     |
| <b>Number of Simulators</b> 5      | <b>Average Standard Deviation</b> 0.0002 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.077  |
|                                             |        | 0.20 g/210L Test # 4             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.077  |
|                                             |        | 0.20 g/210L Test # 5             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.077  |
| 0.20 g/210L Test # 6                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.077  |
| 0.20 g/210L Test # 7                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.077  |
| 0.20 g/210L Test # 8                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.077  |
| 0.20 g/210L Test # 10                  | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 10/31/2022 12:35

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> POLK CSO            | <b>Date</b> 11/07/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> COKE,CLEVELAND,A | <b>Time</b> 15:26:44         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C         | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 15922080A3 |
| <b>Expiration</b> 01/12/2023      | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 08/05/2024     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 11/08/2022 09:23

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> POLK CSO        | <b>Date</b> 12/06/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> COLE,KEVIN,T | <b>Time</b> 11:37:31         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C     | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 15922080A3 |
| <b>Expiration</b> 01/12/2023  | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 08/05/2024     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/06/2022 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PASCO COUNTY SO  | <b>Date</b> 01/07/2022 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> NIXON, BARRY, | <b>Time</b> 10:28:15   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 14020080A1 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 07/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.162  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.197  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.157  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.198  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.160  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside Tolerance CHANGED SIMULATOR.

**Data Download Date/Time** 01/07/2022 15:57

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 02/23/2022  
**Time** 20:03:33

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/24/2022 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 03/07/2022  
**Time** 05:46:50

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration**      **Expiration**

**Number of Simulators** 5

**Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.073  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1 | 0.071  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.048  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.075  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2 | 0.072  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.048  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Test # 3          | 0.076  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3 | 0.072  |
|                                             |        | Interferent Detect Test # 3   | INT    |

#### Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 03/07/2022 14:56

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PASCO COUNTY SO        | <b>Date</b> 03/07/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW, TAYLOR, D | <b>Time</b> 14:38:55   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

#### Remarks:

Data Download Date/Time 03/07/2022 14:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 03/07/2022  
**Time** 20:43:56

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/07/2022 20:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 04/01/2022  
**Time** 12:41:27

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/01/2022 12:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 05/16/2022  
**Time** 14:11:36

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/16/2022 14:15

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 06/23/2022  
**Time** 10:17:23

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/23/2022 10:22

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 07/19/2022  
**Time** 11:00:26

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/19/2022 11:10

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 08/22/2022  
**Time** 11:20:11

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/22/2022 11:33

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 09/27/2022  
**Time** 10:33:28

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/27/2022 10:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 10/14/2022  
**Time** 10:40:45

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/14/2022 10:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 11/29/2022  
**Time** 19:21:33

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/07/2022 09:04

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

**Agency** PASCO COUNTY SO  
**Inspector** NIXON,BARRY,

**Date** 12/03/2022  
**Time** 21:46:05

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 11821080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/07/2022 09:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 01/12/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 12:13:41      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/12/2022 12:34

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 02/17/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 12:05:02      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/17/2022 12:35

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | JACKSONVILLE SO   | <b>Date</b>       | 03/23/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIGHTSEY,ROBERT,D | <b>Time</b>       | 13:17:14   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 08121080A1 |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 05/05/2023 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/24/2022 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 04/13/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 15:04:55      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

RMD FM MNS PLCD IN EVED USE AT PTDF 4-13-2022

**Data Download Date/Time** 04/15/2022 17:04

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 05/10/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 10:58:33      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/10/2022 11:54

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 05/30/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 09:19:05      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.190  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/01/2022 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 06/23/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 17:18:43      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

REMVD FROM EVED USE AT PTDF FOR P/M AN D/I ON 6-23-2022

**Data Download Date/Time** 06/23/2022 17:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 11/29/2022      **Type** Department Inspection  
**Inspector** NICODEMO,PHIL,      **Time** 14:00:13      **Software** 8100.27

**0.05 Lot #** 202201C      **0.08 Lot #** 202201D      **0.20 Lot #** 202201E      **0.08 Gas Lot #** AG229803  
**Expiration** 01/11/2024      **Expiration** 01/18/2024      **Expiration** 01/18/2024      **Expiration** 10/25/2024

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 11/29/2022 14:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 12/09/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 11:59:40      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 07622080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PLCED INTO EVED USE AT PTDF 12-09-2022

**Data Download Date/Time** 12/09/2022 12:08

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

**Agency** JACKSONVILLE SO      **Date** 12/28/2022      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 10:49:59      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 07622080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/29/2022 09:41

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

|                  |                   |             |            |                 |                   |
|------------------|-------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | FL HIGHWAY PATROL | <b>Date</b> | 02/02/2022 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | GUTSCHOW,TAYLOR,D | <b>Time</b> | 11:25:36   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 02/02/2022 15:20

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL    | <b>Date</b> 02/02/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 15:11:32   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 02/02/2022 15:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 02/16/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SMITH,WILLIAM,T   | <b>Time</b>       | 22:11:17   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1249794    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 12/13/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO ISSUES

Data Download Date/Time 02/16/2022 22:48

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

**Agency** FL HIGHWAY PATROL **Date** 03/22/2022 **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N **Time** 01:08:35 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | Yes    |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

Time-Date changed.AF TEST NOT ALCOHOL FREE  
Non-compliance: TEST ERROR ON ALCOHOL FREE TEST.

**Data Download Date/Time** 03/22/2022 02:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

**Agency** FL HIGHWAY PATROL **Date** 04/29/2022 **Type** Agency Inspection  
**Inspector** SMITH, WILLIAM, T **Time** 00:22:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1249794  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 12/13/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO ISSUES

**Data Download Date/Time** 04/29/2022 01:28

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 05/31/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SMITH,WILLIAM,T   | <b>Time</b>       | 23:33:46   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1506449    |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 06/10/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO ISSUES

**Data Download Date/Time** 08/30/2022 12:12

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

|                  |                   |             |            |                 |                   |
|------------------|-------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | FL HIGHWAY PATROL | <b>Date</b> | 09/01/2022 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | GUTSCHOW,TAYLOR,D | <b>Time</b> | 10:14:54   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

**Data Download Date/Time** 09/01/2022 13:56

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> FL HIGHWAY PATROL    | <b>Date</b> 09/01/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:44:05   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0001 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.077  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.077  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.077  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.077  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.077  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.077  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 09/01/2022 13:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

**Agency** FL HIGHWAY PATROL **Date** 09/25/2022 **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N **Time** 20:19:41 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 09/26/2022 08:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

**Agency** FL HIGHWAY PATROL **Date** 10/30/2022 **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N **Time** 18:33:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/03/2022 18:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

**Agency** FL HIGHWAY PATROL **Date** 11/30/2022 **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N **Time** 09:12:59 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE/ CASE CHECKED NO CRACKS

**Data Download Date/Time** 11/30/2022 09:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

**Agency** FL HIGHWAY PATROL **Date** 12/01/2022 **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N **Time** 16:58:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE/ CASE CHECKED NO CRACKS

**Data Download Date/Time** 12/05/2022 16:54

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

**Agency** FHP TROOP E MIAMI **Date** 11/16/2022 **Type** Agency Inspection  
**Inspector** REYES-RIVERA, DAVID, E **Time** 09:26:25 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

**Remarks:**

**Data Download Date/Time** 11/16/2022 12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

|                                         |                        |                                   |  |
|-----------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> FHP TROOP E MIAMI         | <b>Date</b> 11/16/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:26:00   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0005 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.191  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.191  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.193  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.193  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.194  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.194  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.193  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0012 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 11/16/2022 12:35

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 01/10/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,   | <b>Time</b> 10:34:32   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/10/2022 12:38

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

**Agency** FL HIGHWAY PATROL      **Date** 01/12/2022      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 13:15:43      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 04/21/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 01/13/2022 08:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

**Agency** FL HIGHWAY PATROL **Date** 02/10/2022 **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M **Time** 16:03:37 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/21/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 02/11/2022 17:20

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

**Agency** FL HIGHWAY PATROL  
**Inspector** RUBEL, NEIL, F

**Date** 04/04/2022  
**Time** 09:55:06

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/21/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 04/05/2022 10:54

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

**Agency** FL HIGHWAY PATROL **Date** 05/20/2022 **Type** Agency Inspection  
**Inspector** RUBEL, NEIL, F **Time** 13:52:53 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/21/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 05/26/2022 15:34

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

**Agency** FL HIGHWAY PATROL **Date** 06/03/2022 **Type** Agency Inspection  
**Inspector** RUBEL, NEIL, F **Time** 12:18:36 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/21/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 06/03/2022 13:27

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

|                                 |                              |                               |                               |
|---------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 07/11/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> RUBEL, NEIL, F | <b>Time</b> 15:51:12         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202101C       | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1482048 |
| <b>Expiration</b> 01/12/2023    | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 04/21/2024  |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 07/11/2022 16:59

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 08/09/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | RUBEL,NEIL,F      | <b>Time</b>       | 07:01:06   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1482048    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/21/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

Data Download Date/Time 08/09/2022 08:08

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

**Agency** FL HIGHWAY PATROL **Date** 09/17/2022 **Type** Agency Inspection  
**Inspector** ABRAHAM,ERIC,D **Time** 12:11:09 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/21/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/17/2022 13:17

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

|                                 |                              |                               |                               |
|---------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 10/11/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> BARGE,SUSAN,M  | <b>Time</b> 10:48:57         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202101C       | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1482048 |
| <b>Expiration</b> 01/12/2023    | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 04/21/2024  |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20: Control Outside Tolerance RAN IT WITH .08 SIM

**Data Download Date/Time** 10/11/2022 11:55

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 11/08/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | RUBEL,NEIL,F      | <b>Time</b>       | 08:54:32   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1482048    |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/21/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 11/08/2022 10:18

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

**Agency** FL HIGHWAY PATROL **Date** 12/01/2022 **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M **Time** 14:51:11 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | INT    |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.049  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | 0.050  |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.045  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | 0.049  |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

Int Det: WRONG SIM HOOKED UP. Non-compliance: WRONG SIM HOOKED UP.

**Data Download Date/Time** 12/01/2022 16:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

|                                 |                              |                               |                               |
|---------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 12/01/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> BARGE,SUSAN,M  | <b>Time</b> 15:48:28         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202112A       | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1482048 |
| <b>Expiration</b> 12/07/2023    | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 04/21/2024  |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 12/01/2022 16:53

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 01/28/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HITCHCOCK, WILLIAM, D | <b>Time</b>       | 20:28:28   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C               | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1206004    |
| <b>Expiration</b>           | 01/12/2023            | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 09/20/2022 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/28/2022 21:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

|                  |                   |             |            |                 |                   |
|------------------|-------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | FL HIGHWAY PATROL | <b>Date</b> | 03/01/2022 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | GUTSCHOW,TAYLOR,D | <b>Time</b> | 12:39:53   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 03/01/2022 15:45

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL    | <b>Date</b> 03/01/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 15:42:54   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/01/2022 15:45

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

**Agency** FL HIGHWAY PATROL **Date** 03/30/2022 **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D **Time** 03:51:20 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1206004  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/03/2022 04:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 04/21/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HITCHCOCK, WILLIAM, D | <b>Time</b>       | 01:01:14   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C               | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1206004    |
| <b>Expiration</b>           | 01/12/2023            | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 09/20/2022 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 04/21/2022 22:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 05/06/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HITCHCOCK, WILLIAM, D | <b>Time</b>       | 20:42:48   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C               | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1206004    |
| <b>Expiration</b>           | 01/12/2023            | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 09/20/2022 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/14/2022 00:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

**Agency** FL HIGHWAY PATROL **Date** 06/22/2022 **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D **Time** 00:10:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1206004  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/24/2022 13:48

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

**Agency** FL HIGHWAY PATROL **Date** 07/12/2022 **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D **Time** 01:13:29 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1206004  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/12/2022 01:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

|                                        |                        |                               |  |
|----------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> FL HIGHWAY PATROL        | <b>Date</b> 08/11/2022 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HITCHCOCK, WILLIAM, D | <b>Time</b> 01:27:56   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 202112A    | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B    | <b>0.08 Gas Lot #</b> 1206004 |
| <b>Expiration</b> 12/07/2023 | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023 | <b>Expiration</b> 09/20/2022  |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.043  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1        | 0.046  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1                 | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1                 | 0.197  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1          | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 2                | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2                 | 0.045  |
| 0.08 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2        | 0.048  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.078  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.20 g/210L Test # 2                 | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2     | 0.080  |
|                                             |        | Interferent Detect Test # 2          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.047  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3        | 0.047  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.081  |
|                                             |        | Interferent Detect Test # 3          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Int Det: BAD SEAL. TIGHTENED CAP. 05: Control Outside Tolerance LOW TEMP, WRAPPED WITH BLANKE.

**Data Download Date/Time** 08/12/2022 11:40

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

**Agency** FL HIGHWAY PATROL **Date** 09/27/2022 **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D **Time** 01:32:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.190  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/27/2022 01:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

**Agency** FL HIGHWAY PATROL **Date** 10/19/2022 **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D **Time** 02:14:51 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/21/2022 14:16

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

**Agency** FL HIGHWAY PATROL **Date** 11/17/2022 **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D **Time** 02:02:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.038  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.046  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.186  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.203  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Interferent Detect Test # 1      | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 2            | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2             | 0.039  |
| 0.08 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.046  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.20 g/210L Test # 2             | 0.196  |
|                                             |        | 0.20 g/210L Test # (Repeat) 2    | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.039  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: Control Outside Tolerance BAD SEAL. 20: Control Outside Tolerance LOW TEMP.

**Data Download Date/Time** 11/17/2022 02:11

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 12/06/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HITCHCOCK, WILLIAM, D | <b>Time</b>       | 01:52:15   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C               | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 402357111  |
| <b>Expiration</b>           | 01/12/2023            | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 02/10/2025 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.043  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.049  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.050  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: Control Outside Tolerance WRAPPED WITH TOWEL SIM WARMED.

**Data Download Date/Time** 12/11/2022 02:35

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 01/26/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SMITH,WILLIAM,T   | <b>Time</b>       | 20:39:46   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1249794    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 12/13/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO ISSUES

**Data Download Date/Time** 01/26/2022 21:16

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

**Agency** FL HIGHWAY PATROL      **Date** 02/16/2022      **Type** Agency Inspection  
**Inspector** SMITH, WILLIAM, T      **Time** 22:18:39      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1249794  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/13/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO ISSUES

**Data Download Date/Time** 02/16/2022 22:24

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

**Agency** FL HIGHWAY PATROL      **Date** 03/22/2022      **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N      **Time** 02:42:05      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1249794  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/13/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/22/2022 03:07

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 04/29/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SMITH,WILLIAM,T   | <b>Time</b>       | 00:12:54   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1249794    |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 12/13/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO ISSUES

**Data Download Date/Time** 04/29/2022 01:18

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 05/31/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SMITH,WILLIAM,T   | <b>Time</b>       | 23:02:57   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1249794    |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 12/13/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO ISSUES

**Data Download Date/Time** 06/03/2022 12:12

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 06/22/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SMITH,WILLIAM,T   | <b>Time</b>       | 23:51:32   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1249794    |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 12/13/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO ISSUES

**Data Download Date/Time** 06/23/2022 14:01

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 07/21/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DOBOSIEWICZ,ANTHONY,N | <b>Time</b>       | 00:54:42   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A               | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1249794    |
| <b>Expiration</b>           | 12/07/2023            | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 12/13/2022 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 07/21/2022 02:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 08/29/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DOBOSIEWICZ,ANTHONY,N | <b>Time</b>       | 02:08:43   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A               | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1249794    |
| <b>Expiration</b>           | 12/07/2023            | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 12/13/2022 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Range Exceeded.

Data Download Date/Time 09/14/2022 08:53

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> FL HIGHWAY PATROL    | <b>Date</b> 09/14/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 15:31:46   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0002 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

|  |
|--|
|  |
|--|

Data Download Date/Time 09/14/2022 15:32

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 09/25/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DOBOSIEWICZ,ANTHONY,N | <b>Time</b>       | 20:42:10   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A               | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 1249794    |
| <b>Expiration</b>           | 12/07/2023            | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 12/13/2022 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLINACE

**Data Download Date/Time** 09/26/2022 09:39

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 10/30/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DOBOSIEWICZ,ANTHONY,N | <b>Time</b>       | 18:02:43   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A               | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 2422080A1  |
| <b>Expiration</b>           | 12/07/2023            | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 02/05/2024 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 11/03/2022 18:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

**Agency** FL HIGHWAY PATROL **Date** 11/30/2022 **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N **Time** 06:58:59 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 2422080A1  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE/ NO CRACKS ON CASE

**Data Download Date/Time** 11/30/2022 07:32

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

**Agency** FL HIGHWAY PATROL **Date** 12/01/2022 **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N **Time** 17:51:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 2422080A1  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: No Sample Provided.

**Data Download Date/Time** 12/02/2022 16:54

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency DELAND P.D. Date 01/31/2022 Type Agency Inspection  
Inspector WALTER,CRAIG,K Time 19:18:27 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1433248  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 01/06/2024

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

Data Download Date/Time 02/04/2022 18:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

**Agency** DELAND P.D. **Date** 02/28/2022 **Type** Agency Inspection  
**Inspector** WALTER,CRAIG,K **Time** 19:33:12 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 02/28/2022 19:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

**Agency** DELAND P.D.  
**Inspector** WALTER,CRAIG,K

**Date** 03/30/2022  
**Time** 18:19:03

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/06/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 04/20/2022 09:16

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> DELAND P.D.     | <b>Date</b> 04/20/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 12:49:41   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 04/20/2022 12:50

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

**Agency** DELAND P.D.  
**Inspector** WALTER,CRAIG,K

**Date** 05/04/2022  
**Time** 21:04:24

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/06/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 05/04/2022 21:20

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

**Agency** DELAND P.D.  
**Inspector** WALTER,CRAIG,K

**Date** 05/30/2022  
**Time** 21:58:10

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/06/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 07/06/2022 03:16

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

**Agency** DELAND P.D.  
**Inspector** WALTER,CRAIG,K

**Date** 06/30/2022  
**Time** 16:11:44

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/06/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 07/06/2022 03:16

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

|                                 |                              |                               |                               |
|---------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> DELAND P.D.       | <b>Date</b> 07/31/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> WALTER,CRAIG,K | <b>Time</b> 21:57:04         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202101C       | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1433248 |
| <b>Expiration</b> 01/12/2023    | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 01/06/2024  |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.189  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.192  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.193  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.192  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20: Control Outside Tolerance.

**Data Download Date/Time** 07/31/2022 22:07

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

**Agency** DELAND P.D. **Date** 08/31/2022 **Type** Agency Inspection  
**Inspector** WALTER,CRAIG,K **Time** 15:16:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433248  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 08/31/2022 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

**Agency** DELAND P.D. **Date** 09/30/2022 **Type** Agency Inspection  
**Inspector** WALTER,CRAIG,K **Time** 20:41:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433248  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 09/30/2022 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

|                                 |                              |                               |                               |
|---------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> DELAND P.D.       | <b>Date</b> 10/31/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> WALTER,CRAIG,K | <b>Time</b> 15:33:23         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202112A       | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1433248 |
| <b>Expiration</b> 12/07/2023    | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 01/06/2024  |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.193  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Control Outside Tolerance. SIM HAD TO WARM UP MORE.

**Data Download Date/Time** 11/05/2022 12:20

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

|                                 |                              |                               |                               |
|---------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> DELAND P.D.       | <b>Date</b> 11/30/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> WALTER,CRAIG,K | <b>Time</b> 16:24:26         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202112A       | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1433248 |
| <b>Expiration</b> 12/07/2023    | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 01/06/2024  |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.178  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.194  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.188  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.190  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20: Control Outside Tolerance. COLD SIM. REPEATED WITH PASS

**Data Download Date/Time** 11/30/2022 16:32

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency DELAND P.D. Date 12/30/2022 Type Agency Inspection  
Inspector WALTER,CRAIG,K Time 18:07:23 Software 8100.27

0.05 Lot # 202112A 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 1433248  
Expiration 12/07/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 01/06/2024

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

Data Download Date/Time 01/03/2023 16:38

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 01/04/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,   | <b>Time</b> 09:40:52   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 01/04/2022 09:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL      **Date** 01/05/2022      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 14:38:18      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 03521080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 05/12/2022      **Expiration** 03/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 01/05/2022 14:41

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL **Date** 02/09/2022 **Type** Agency Inspection  
**Inspector** HENING,FRED,C **Time** 11:11:41 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 01/12/2023 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.042  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1   | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 2         | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2          | 0.042  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # (Repeat) 2 | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2   | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 3         | 0.000  |
| 0.05 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.041  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # (Repeat) 3 | 0.000  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                               |        |

#### Remarks:

05: Control Outside Tolerance DRY O RING ON SIMM, Control O. ISSUE WITH SIMMNon-compliance: UNKNOWN ISSUE W SIMM.

**Data Download Date/Time** 02/09/2022 15:02

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL      **Date** 02/09/2022      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 14:59:17      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 04120080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 03/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 02/09/2022 15:02

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL      **Date** 03/08/2022      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 14:34:44      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 34520080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 03/08/2022 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 04/26/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HENING,FRED,C     | <b>Time</b>       | 13:33:25   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 33321080A4 |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 01/05/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 04/26/2022 13:43

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL      **Date** 05/16/2022      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 12:35:13      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 33321080A4  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 05/16/2022 12:43

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 06/30/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HENING,FRED,C     | <b>Time</b>       | 11:23:49   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 34520080A1 |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 02/05/2023 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

Data Download Date/Time 07/07/2022 12:32

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL      **Date** 07/28/2022      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 09:49:55      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 07/28/2022 09:53

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL      **Date** 08/18/2022      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 10:05:54      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 08121080A1  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 05/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 08/18/2022 10:17

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL      **Date** 09/08/2022      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 11:49:28      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 34520080A1  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.045  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fai TOO MUCH ALCOHOL IN ROOM.

**Data Download Date/Time** 09/08/2022 12:01

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 10/10/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HENING,FRED,C     | <b>Time</b>       | 10:01:16   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 34520080A1 |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 02/05/2023 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

Data Download Date/Time 10/11/2022 11:07

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 11/21/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HENING,FRED,C     | <b>Time</b>       | 13:45:00   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 34520080A1 |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 02/05/2023 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 11/21/2022 14:01

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

**Agency** FL HIGHWAY PATROL **Date** 12/20/2022 **Type** Agency Inspection  
**Inspector** HENING,FRED,C **Time** 11:43:39 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 34520080A1  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 12/20/2022 12:14

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

**Agency** EGLIN AFB **Date** 01/24/2022 **Type** Agency Inspection  
**Inspector** TUSKES,ARTHUR,A **Time** 06:51:43 **Software** 8100.27

**0.05 Lot #** 2021101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/24/2022 08:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                                  |                              |                               |                               |
|----------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> EGLIN AFB          | <b>Date</b> 02/21/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> TUSKES,ARTHUR,A | <b>Time</b> 06:44:21         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202101C        | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1283438 |
| <b>Expiration</b> 01/12/2023     | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 02/25/2023  |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/21/2022 07:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                                  |                              |                               |                               |
|----------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> EGLIN AFB          | <b>Date</b> 03/25/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> TUSKES,ARTHUR,A | <b>Time</b> 06:41:30         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202101C        | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1283438 |
| <b>Expiration</b> 01/12/2023     | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 02/25/2023  |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/25/2022 07:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

**Agency** EGLIN AFB **Date** 04/26/2022 **Type** Agency Inspection  
**Inspector** TUSKES,ARTHUR,A **Time** 06:51:58 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/26/2022 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                                  |                              |                               |                               |
|----------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> EGLIN AFB          | <b>Date</b> 05/20/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> TUSKES,ARTHUR,A | <b>Time</b> 06:49:36         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202101C        | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1283438 |
| <b>Expiration</b> 01/12/2023     | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 02/25/2023  |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/25/2022 07:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                                  |                              |                               |                               |
|----------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> EGLIN AFB          | <b>Date</b> 06/28/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> TUSKES,ARTHUR,A | <b>Time</b> 06:20:46         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202101C        | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1283438 |
| <b>Expiration</b> 01/12/2023     | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 02/25/2023  |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/04/2022 11:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                               |                        |                               |
|-------------------------------|------------------------|-------------------------------|
| <b>Agency</b> EGLIN AFB       | <b>Date</b> 08/12/2022 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 07:11:16   | <b>Software</b> 8100.27       |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                       |
|-----------------------------|-----------------------|
| <b>Number of Simulators</b> | <b>Compliance</b> Yes |
|-----------------------------|-----------------------|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 08/12/2022 10:52

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                               |                        |                                   |  |
|-------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> EGLIN AFB       | <b>Date</b> 08/12/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 09:31:43   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 08121080A1 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 08/12/2022 10:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                                  |                              |                               |                               |
|----------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> EGLIN AFB          | <b>Date</b> 08/26/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> TUSKES,ARTHUR,A | <b>Time</b> 06:22:11         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202112A        | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1283438 |
| <b>Expiration</b> 12/07/2023     | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 02/25/2023  |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/29/2022 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

**Agency** EGLIN AFB **Date** 09/22/2022 **Type** Agency Inspection  
**Inspector** TUSKES,ARTHUR,A **Time** 06:42:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1283438  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/22/2022 08:13

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

|                                  |                              |                               |                               |
|----------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> EGLIN AFB          | <b>Date</b> 10/25/2022       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> TUSKES,ARTHUR,A | <b>Time</b> 06:55:26         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202112A        | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 1283438 |
| <b>Expiration</b> 12/07/2023     | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 02/25/2023  |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/25/2022 08:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

**Agency** EGLIN AFB **Date** 11/30/2022 **Type** Agency Inspection  
**Inspector** TUSKES,ARTHUR,A **Time** 06:08:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1283438  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/30/2022 07:25

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

**Agency** EGLIN AFB **Date** 12/26/2022 **Type** Agency Inspection  
**Inspector** TUSKES,ARTHUR,A **Time** 06:07:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1283438  
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.200  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.079  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.079  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08e FORGOT TO PLUG IN TUBE.

**Data Download Date/Time** 12/26/2022 07:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> PASCO COUNTY SO      | <b>Date</b> 11/30/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:57:10   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.077  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 11/30/2022 14:11

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                  |                   |             |            |                 |                   |
|------------------|-------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | FL HIGHWAY PATROL | <b>Date</b> | 01/18/2022 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | SOTO,ISRAEL,      | <b>Time</b> | 07:06:39   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 02/01/2022 13:41

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                  |                   |             |            |                 |                   |
|------------------|-------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | FL HIGHWAY PATROL | <b>Date</b> | 04/01/2022 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | SOTO,ISRAEL,      | <b>Time</b> | 07:24:34   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 04/01/2022 13:41

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 04/01/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,   | <b>Time</b> 12:27:49   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.082  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 04/01/2022 13:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

**Agency** FL HIGHWAY PATROL  
**Inspector** RUBEL, NEIL, F

**Date** 04/08/2022  
**Time** 14:32:26

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402328118  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/20/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 04/18/2022 14:28

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 05/05/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,   | <b>Time</b> 08:43:42   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0007      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0008 |
| 0.20 g/210L Test (Std Dev)             | 0.0011 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 05/05/2022 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

**Agency** FL HIGHWAY PATROL  
**Inspector** RUBEL, NEIL, F

**Date** 05/13/2022  
**Time** 07:58:38

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402328118  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/20/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 05/16/2022 10:38

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

**Agency** FL HIGHWAY PATROL  
**Inspector** RUBEL, NEIL, F

**Date** 06/14/2022  
**Time** 09:40:53

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402328118  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/20/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 06/14/2022 11:00

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

**Agency** FL HIGHWAY PATROL      **Date** 08/10/2022      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 16:02:47      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402328118  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/20/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 08/10/2022 17:16

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                                 |                        |                                   |  |
|---------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 10/18/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> SOTO,ISRAEL,   | <b>Time</b> 09:12:56   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 08121080A1 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0002 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1                      | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.081  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.082  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.081  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3               | INT    |
|                                             |        | Alcohol Free Test # 4                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 4                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 4                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 4                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4          | INT    |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.082  |
|                                             |        | Interferent Detect Test # 4               | INT    |
|                                             |        | Alcohol Free Test # 5                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 5                      | 0.049  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                                                    |        |
|----------------------------------------------------|--------|
| 0.08 g/210L Test # 5                               | 0.078  |
| 0.20 g/210L Test # 5                               | 0.198  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5          | 0.082  |
| Interferent Detect Test # 5                        | INT    |
| Alcohol Free Test # 6                              | 0.000  |
| 0.05 g/210L Test # 6                               | 0.048  |
| 0.08 g/210L Test # 6                               | 0.078  |
| 0.20 g/210L Test # 6                               | 0.198  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6          | 0.082  |
| Interferent Detect Test # 6                        | INT    |
| Alcohol Free Test # 7                              | 0.000  |
| 0.05 g/210L Test # 7                               | 0.048  |
| 0.08 g/210L Test # 7                               | 0.078  |
| 0.20 g/210L Test # 7                               | 0.199  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7          | 0.082  |
| Interferent Detect Test # 7                        | INT    |
| Alcohol Free Test # 8                              | 0.000  |
| 0.05 g/210L Test # 8                               | 0.048  |
| 0.08 g/210L Test # 8                               | 0.078  |
| 0.20 g/210L Test # 8                               | 0.198  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8          | 0.082  |
| Interferent Detect Test # 8                        | INT    |
| Alcohol Free Test # 9                              | 0.000  |
| 0.05 g/210L Test # 9                               | 0.048  |
| 0.08 g/210L Test # 9                               | 0.079  |
| 0.20 g/210L Test # 9                               | 0.199  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9          | 0.082  |
| Interferent Detect Test # 9                        | INT    |
| Alcohol Free Test # 10                             | 0.000  |
| 0.05 g/210L Test # 10                              | 0.048  |
| 0.08 g/210L Test # 10                              | 0.078  |
| 0.20 g/210L Test # 10                              | 0.198  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10         | 0.082  |
| Interferent Detect Test # 10                       | INT    |
| 0.05 g/210L Test (Std Dev)                         | 0.0003 |
| 0.08 g/210L Test (Std Dev)                         | 0.0003 |
| 0.20 g/210L Test (Std Dev)                         | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)<br>(Repeat) | 0.0000 |

**Remarks:**

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

08: Interferent Detect CLEARED AREA.

---

**Data Download Date/Time** 10/18/2022 11:04

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                                 |                              |                               |                                  |
|---------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 11/19/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> RUBEL, NEIL, F | <b>Time</b> 16:26:25         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202112A       | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 2402477284 |
| <b>Expiration</b> 12/07/2023    | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 06/27/2025     |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 11/19/2022 17:33

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 12/01/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | BARGE,SUSAN,M     | <b>Time</b>       | 14:54:00   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 402477284  |
| <b>Expiration</b>           | 12/07/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 06/27/2025 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 12/01/2022 15:59

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

**Agency** FL HIGHWAY PATROL      **Date** 01/19/2022      **Type** Agency Inspection  
**Inspector** BUDAY,JOSHUA,M      **Time** 11:53:10      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1185592  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/13/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION.

**Data Download Date/Time** 01/19/2022 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 01/21/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,   | <b>Time</b> 15:40:15   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 01/21/2022 15:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 02/28/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | BUDAY,JOSHUA,M    | <b>Time</b>       | 10:43:50   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1185592    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 08/13/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1               | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2                      | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2               | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.051  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08: Ambient Fail CLEARED AREA OK 2ND ATTEMPT.

**Data Download Date/Time** 02/28/2022 10:59

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 03/29/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | BUDAY,JOSHUA,M    | <b>Time</b>       | 12:55:27   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1185592    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 08/13/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION

**Data Download Date/Time** 03/29/2022 13:00

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 04/16/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | BUDAY,JOSHUA,M    | <b>Time</b>       | 00:45:25   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1185592    |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 08/13/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION

**Data Download Date/Time** 04/23/2022 02:47

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

**Agency** FL HIGHWAY PATROL **Date** 06/07/2022 **Type** Agency Inspection  
**Inspector** BUDAY,JOSHUA,M **Time** 11:16:46 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Sequence Aborted. Non-compliance: .

**Data Download Date/Time** 06/07/2022 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

**Agency** FL HIGHWAY PATROL      **Date** 06/07/2022      **Type** Agency Inspection  
**Inspector** BUDAY,JOSHUA,M      **Time** 11:51:37      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1185592  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/13/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION

**Data Download Date/Time** 06/07/2022 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

**Agency** FHP TROOP E MIAMI **Date** 11/16/2022 **Type** Agency Inspection  
**Inspector** REYES-RIVERA,DAVID,E **Time** 09:24:40 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | Yes    |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

Time-Date changed.COMPLIANCE NOT DETERMINED. AI NOT CONDUCT

**Data Download Date/Time** 11/16/2022 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

|                                         |                        |                                   |  |
|-----------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> FHP TROOP E MIAMI         | <b>Date</b> 11/16/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:31:19   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0001 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 11/16/2022 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 01/26/2022  
**Time** 09:51:12

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 01/26/2022 10:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 02/25/2022  
**Time** 09:58:05

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 02/25/2022 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 03/31/2022  
**Time** 09:40:53

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 03/31/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 04/27/2022  
**Time** 10:01:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 04/27/2022 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 05/23/2022  
**Time** 13:37:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 05/23/2022 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FHP TROOP E MIAMI         | <b>Date</b> 05/24/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:11:05   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 05/24/2022 12:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 05/25/2022  
**Time** 12:32:34

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 05/31/2022 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI **Date** 06/30/2022 **Type** Agency Inspection  
**Inspector** ADDISON, RAYMOND, E **Time** 07:12:44 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 06/22/2023 **Expiration** 08/03/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.189  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.198  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.191  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20:SIM STILL WARMING OFF.

**Data Download Date/Time** 07/05/2022 08:13

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 07/31/2022  
**Time** 12:05:44

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 07/31/2022 12:27

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 08/31/2022  
**Time** 20:15:14

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1356262  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 09/01/2022 02:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 09/01/2022  
**Time** 02:32:50

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1356262  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 09/01/2022 02:38

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 10/26/2022  
**Time** 08:44:03

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1356262  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 10/26/2022 08:50

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 11/30/2022  
**Time** 23:25:43

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1356262  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 12/01/2022 14:12

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 12/01/2022  
**Time** 02:08:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1356262  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 12/01/2022 14:12

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 01/26/2022  
**Time** 10:27:31

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1114319  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 03/27/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/26/2022 10:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 02/25/2022  
**Time** 10:08:44

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1114319  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 03/27/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 02/25/2022 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 03/31/2022  
**Time** 09:32:05

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/03/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/31/2022 09:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 04/27/2022  
**Time** 09:59:39

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 402310313  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/27/2022 10:20

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FHP TROOP E MIAMI         | <b>Date</b> 05/24/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:33:04   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1                      | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.077  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.077  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.077  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.079  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.076  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3               | INT    |
|                                             |        | Alcohol Free Test # 4                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 4                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 4                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 4                      | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4          | 0.076  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4               | INT    |
|                                             |        | Alcohol Free Test # 5                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 5                      | 0.049  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

|                                            |        |
|--------------------------------------------|--------|
| 0.08 g/210L Test # 5                       | 0.079  |
| 0.20 g/210L Test # 5                       | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 5           | 0.076  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5  | 0.077  |
| Interferent Detect Test # 5                | INT    |
| Alcohol Free Test # 6                      | 0.000  |
| 0.05 g/210L Test # 6                       | 0.049  |
| 0.08 g/210L Test # 6                       | 0.078  |
| 0.20 g/210L Test # 6                       | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6           | 0.075  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6  | 0.076  |
| Interferent Detect Test # 6                | INT    |
| Alcohol Free Test # 7                      | 0.000  |
| 0.05 g/210L Test # 7                       | 0.049  |
| 0.08 g/210L Test # 7                       | 0.078  |
| 0.20 g/210L Test # 7                       | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7           | 0.076  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7  | 0.076  |
| Interferent Detect Test # 7                | INT    |
| Alcohol Free Test # 8                      | 0.000  |
| 0.05 g/210L Test # 8                       | 0.049  |
| 0.08 g/210L Test # 8                       | 0.078  |
| 0.20 g/210L Test # 8                       | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8           | INT    |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8  | 0.076  |
| Interferent Detect Test # 8                | INT    |
| Alcohol Free Test # 9                      | 0.000  |
| 0.05 g/210L Test # 9                       | 0.049  |
| 0.08 g/210L Test # 9                       | 0.079  |
| 0.20 g/210L Test # 9                       | 0.198  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9  | 0.077  |
| Interferent Detect Test # 9                | INT    |
| Alcohol Free Test # 10                     | 0.000  |
| 0.05 g/210L Test # 10                      | 0.049  |
| 0.08 g/210L Test # 10                      | 0.078  |
| 0.20 g/210L Test # 10                      | 0.198  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.076  |
| Interferent Detect Test # 10               | INT    |
| 0.05 g/210L Test (Std Dev)                 | 0.0004 |
| 0.08 g/210L Test (Std Dev)                 | 0.0004 |

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

|                                                    |        |
|----------------------------------------------------|--------|
| 0.20 g/210L Test (Std Dev)                         | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)<br>(Repeat) | 0.0005 |

**Remarks:**

08: Interferent DetectUNKNOWN SOURCE.

---

**Data Download Date/Time** 05/24/2022 13:15

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 05/25/2022  
**Time** 12:27:48

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402310313  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 05/25/2022 12:34

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

|                                      |                              |                               |                                  |
|--------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> FHP TROOP E MIAMI      | <b>Date</b> 06/30/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ADDISON, RAYMOND, E | <b>Time</b> 07:30:01         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C            | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 3024023103 |
| <b>Expiration</b> 01/12/2023         | <b>Expiration</b> 07/12/2022 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 01/15/2025     |
| <b>Number of Simulators</b> 5        | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.08 GAS LOT#302402310313

**Data Download Date/Time** 07/05/2022 08:26

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 07/31/2022  
**Time** 11:56:32

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402310313  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/31/2022 12:00

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 08/31/2022  
**Time** 20:50:33

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402310313  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 09/01/2022 02:34

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 09/01/2022  
**Time** 02:24:34

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402310313  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | INT    |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.079  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Interferent Detect DISCONNECTED.

**Data Download Date/Time** 09/01/2022 02:34

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 10/26/2022  
**Time** 08:42:14

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402310313  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 10/26/2022 08:47

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 11/30/2022  
**Time** 23:15:36

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402310313  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail, NOUTH PIECE REMOVED.

**Data Download Date/Time** 12/01/2022 14:17

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

**Agency** FHP TROOP E MIAMI  
**Inspector** LORENZO, YOSNIEL,

**Date** 12/01/2022  
**Time** 01:59:02

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402310313  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.

**Data Download Date/Time** 12/01/2022 14:17

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

**Agency** FHP TROOP D ORLANDO **Date** 07/25/2022 **Type** Agency Inspection  
**Inspector** REYES-RIVERA,DAVID,E **Time** 09:05:16 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

**Data Download Date/Time** 07/25/2022 14:59

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FHP TROOP D ORLANDO       | <b>Date</b> 07/25/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 14:42:43   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.047  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.047  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.047  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.077  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

#### Remarks:

Data Download Date/Time 07/25/2022 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

|                             |                     |                   |            |                   |                   |                       |            |
|-----------------------------|---------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FHP TROOP D ORLANDO | <b>Date</b>       | 09/21/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | RIVERA,VICTOR,J     | <b>Time</b>       | 23:18:05   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A             | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 402477284  |
| <b>Expiration</b>           | 12/07/2023          | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 06/27/2025 |
| <b>Number of Simulators</b> | 5                   | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN NO ISSUES

**Data Download Date/Time** 09/23/2022 08:58

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FHP TROOP D ORLANDO   | <b>Date</b>       | 10/12/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | JEAN BAPTISTE,ALBENS, | <b>Time</b>       | 22:10:03   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202112A               | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 402477284  |
| <b>Expiration</b>           | 12/07/2023            | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 06/27/2025 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN NO ISSUES

**Data Download Date/Time** 10/13/2022 00:13

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

**Agency** FHP TROOP D ORLANDO      **Date** 11/09/2022      **Type** Agency Inspection  
**Inspector** RIVERA, VICTOR, J      **Time** 23:19:54      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402477284  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/27/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN NO ISSUES

**Data Download Date/Time** 11/14/2022 11:21

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

**Agency** FHP TROOP D ORLANDO      **Date** 12/21/2022      **Type** Agency Inspection  
**Inspector** JEAN BAPTISTE,ALBENS,      **Time** 23:45:38      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402477284  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 06/27/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN NO ISSUE

**Data Download Date/Time** 12/23/2022 08:17

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO **Date** 01/11/2022 **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A **Time** 13:59:32 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Range ExceededMOUTH ALCOHOL.

**Data Download Date/Time** 01/11/2022 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | POLK COUNTY SO   | <b>Date</b>       | 01/11/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | COKE,CLEVELAND,A | <b>Time</b>       | 14:42:32   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C          | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 13921080A2 |
| <b>Expiration</b>           | 01/12/2023       | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 07/05/2023 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MOUTH ALCOHOL

Data Download Date/Time 01/11/2022 16:04

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO      **Date** 02/15/2022      **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A      **Time** 14:29:11      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/16/2022 08:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

|                  |                  |             |            |                 |                   |
|------------------|------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | POLK COUNTY SO   | <b>Date</b> | 03/21/2022 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | COKE,CLEVELAND,A | <b>Time</b> | 15:43:52   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |     |
|-----------------------------|---|-------------------|-----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | Yes |
|-----------------------------|---|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: .

**Data Download Date/Time** 03/21/2022 16:53

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO      **Date** 03/21/2022      **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A      **Time** 16:27:57      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 23621080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO MOUTH ALCOHOL

**Data Download Date/Time** 03/21/2022 16:53

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | POLK COUNTY SO   | <b>Date</b>       | 04/22/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | COKE,CLEVELAND,A | <b>Time</b>       | 13:04:29   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C          | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 33321080A2 |
| <b>Expiration</b>           | 01/12/2023       | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 01/05/2024 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO REMARKS

**Data Download Date/Time** 04/22/2022 15:18

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO **Date** 04/22/2022 **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A **Time** 13:34:24 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

**Remarks:**

**Data Download Date/Time** 04/22/2022 15:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO Date 05/11/2022 Type Agency Inspection  
Inspector COKE,CLEVELAND,A Time 10:29:44 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202108D 0.20 Lot # 0.08 Gas Lot #  
Expiration 01/12/2023 Expiration 08/17/2023 Expiration Expiration

Number of Simulators 5 Compliance No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.066  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1 | 0.074  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.049  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.066  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2 | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.050  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Test # 3          | 0.067  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3 | 0.074  |
|                                             |        | Interferent Detect Test # 3   | INT    |

#### Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:LEK.

Data Download Date/Time 05/11/2022 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO      **Date** 05/11/2022      **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A      **Time** 11:19:32      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 33321080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.208  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.206  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SEAL LEAK FIXED SEAL IN COMPLIANCE

**Data Download Date/Time** 05/11/2022 12:03

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO      **Date** 06/07/2022      **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A      **Time** 10:28:39      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 33321080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/07/2022 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> POLK COUNTY SO      | <b>Date</b> 07/15/2022       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> COKE,CLEVELAND,A | <b>Time</b> 07:57:44         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C         | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B     | <b>0.08 Gas Lot #</b> 13921080A2 |
| <b>Expiration</b> 01/12/2023      | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023  | <b>Expiration</b> 07/05/2023     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/15/2022 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO      **Date** 08/08/2022      **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A      **Time** 11:08:03      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 33321080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 08/08/2022 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | POLK COUNTY SO   | <b>Date</b>       | 09/23/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | COKE,CLEVELAND,A | <b>Time</b>       | 09:06:56   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C          | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 33321080A2 |
| <b>Expiration</b>           | 01/12/2023       | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 01/05/2024 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

Data Download Date/Time 09/23/2022 10:43

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO      **Date** 10/24/2022      **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A      **Time** 09:31:30      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 33321080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 01/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 10/24/2022 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> POLK COUNTY SO       | <b>Date</b> 10/31/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:40:35   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0002 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 10/31/2022 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO      **Date** 11/07/2022      **Type** Agency Inspection  
**Inspector** COKE,CLEVELAND,A      **Time** 15:22:41      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 15922080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 11/08/2022 09:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

**Agency** POLK COUNTY SO  
**Inspector** COLE,KEVIN,T

**Date** 12/06/2022  
**Time** 11:33:59

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 15922080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 08/05/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/06/2022 12:09

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PUTNAM COUNTY SO      | <b>Date</b> 02/09/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 14:39:01   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/09/2022 14:41

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

**Agency** PUTNAM COUNTY SO      **Date** 02/22/2022      **Type** Agency Inspection  
**Inspector** SMITH,BRIAN,K      **Time** 11:28:54      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 28021080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 12/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/22/2022 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

**Agency** PUTNAM COUNTY SO **Date** 05/20/2022 **Type** Agency Inspection  
**Inspector** SMITH,BRIAN,K **Time** 15:25:51 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 01/12/2023 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1 | INT    |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 2       | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 2 | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 3       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 3 | INT    |
| Interferent Detect Test: Interferent Detect | Yes    |                             |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                             |        |
| Alcohol Free Test: OK                       | Yes    |                             |        |
| 0.05 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Test: OK                        | No     |                             |        |
| 0.20 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

05: Ambient Fail, Ambient Fail. AMBIENT FAIL.

**Data Download Date/Time** 06/06/2022 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

**Agency** PUTNAM COUNTY SO      **Date** 06/03/2022      **Type** Agency Inspection  
**Inspector** NICODEMO,PHIL,      **Time** 13:33:43      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

**Remarks:**

**Data Download Date/Time** 06/06/2022 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                |                        |                                   |
|--------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PUTNAM COUNTY SO | <b>Date</b> 06/06/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,  | <b>Time</b> 11:12:52   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0007 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1                      | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # (Repeat) 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1               | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2               | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3               | INT    |
|                                             |        | Alcohol Free Test # 4                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 4                      | 0.050  |
|                                             |        | 0.08 g/210L Test # 4                      | 0.081  |
|                                             |        | 0.20 g/210L Test # (Repeat) 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4               | INT    |
|                                             |        | Alcohol Free Test # 5                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 5                      | 0.050  |
|                                             |        | 0.08 g/210L Test # 5                      | 0.080  |
|                                             |        | 0.20 g/210L Test # (Repeat) 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5               | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                                 |        |
|-------------------------------------------------|--------|
| Alcohol Free Test # 6                           | 0.000  |
| 0.05 g/210L Test # 6                            | 0.050  |
| 0.08 g/210L Test # 6                            | 0.081  |
| 0.20 g/210L Test # (Repeat) 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6       | 0.081  |
| Interferent Detect Test # 6                     | INT    |
| Alcohol Free Test # 7                           | 0.000  |
| 0.05 g/210L Test # 7                            | 0.051  |
| 0.08 g/210L Test # 7                            | 0.082  |
| 0.20 g/210L Test # (Repeat) 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7       | 0.081  |
| Interferent Detect Test # 7                     | INT    |
| Alcohol Free Test # 8                           | 0.000  |
| 0.05 g/210L Test # 8                            | 0.051  |
| 0.08 g/210L Test # 8                            | 0.081  |
| 0.20 g/210L Test # (Repeat) 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8       | 0.081  |
| Interferent Detect Test # 8                     | INT    |
| Alcohol Free Test # 9                           | 0.000  |
| 0.05 g/210L Test # 9                            | 0.050  |
| 0.08 g/210L Test # 9                            | 0.081  |
| 0.20 g/210L Test # (Repeat) 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9       | 0.081  |
| Interferent Detect Test # 9                     | INT    |
| Alcohol Free Test # 10                          | 0.000  |
| 0.05 g/210L Test # 10                           | 0.049  |
| 0.08 g/210L Test # 10                           | 0.080  |
| 0.20 g/210L Test # (Repeat) 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10      | 0.081  |
| Interferent Detect Test # 10                    | INT    |
| 0.05 g/210L Test (Std Dev)                      | 0.0007 |
| 0.08 g/210L Test (Std Dev)                      | 0.0008 |
| 0.20 g/210L Test (Std Dev) (Repeat)             | 0.0012 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat) | 0.0003 |

#### Remarks:

20: Ambient Fail CLEARED AREA. 08: Ambient Fail CLEARED AREA.

Data Download Date/Time 06/06/2022 11:13

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PUTNAM COUNTY SO   | <b>Date</b> 09/15/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> NICODEMO, PHIL, | <b>Time</b> 11:45:17   | <b>Software</b> 8100.27           |

|                              |                              |                   |                       |
|------------------------------|------------------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> | <b>Expiration</b>     |

|                               |                                   |                      |
|-------------------------------|-----------------------------------|----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> | <b>Compliance</b> No |
|-------------------------------|-----------------------------------|----------------------|

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.046  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1          | 0.076  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.08 g/210L Test # (Repeat) 1 | 0.075  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.046  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.076  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2 | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.046  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Test # 3          | 0.075  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3 | 0.075  |
|                                             |        | Interferent Detect Test # 3   | INT    |
|                                             |        | Alcohol Free Test # 4         | 0.000  |
|                                             |        | 0.05 g/210L Test # 4          | 0.046  |
|                                             |        | 0.08 g/210L Test # 4          | 0.075  |
|                                             |        | 0.08 g/210L Test # (Repeat) 4 | 0.074  |
|                                             |        | Interferent Detect Test # 4   | INT    |
|                                             |        | Alcohol Free Test # 5         | 0.000  |
|                                             |        | 0.05 g/210L Test # 5          | 0.046  |
|                                             |        | 0.08 g/210L Test # 5          | 0.075  |
|                                             |        | 0.08 g/210L Test # (Repeat) 5 | 0.074  |
|                                             |        | Interferent Detect Test # 5   | INT    |
|                                             |        | Alcohol Free Test # 6         | 0.000  |
|                                             |        | 0.05 g/210L Test # 6          | 0.046  |
|                                             |        | 0.08 g/210L Test # 6          | 0.075  |
|                                             |        | 0.08 g/210L Test # (Repeat) 6 | 0.075  |
|                                             |        | Interferent Detect Test # 6   | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                     |        |
|-------------------------------------|--------|
| Alcohol Free Test # 7               | 0.000  |
| 0.05 g/210L Test # 7                | 0.046  |
| 0.08 g/210L Test # 7                | 0.075  |
| 0.08 g/210L Test # (Repeat) 7       | 0.075  |
| Interferent Detect Test # 7         | INT    |
| Alcohol Free Test # 8               | 0.000  |
| 0.05 g/210L Test # 8                | 0.046  |
| 0.08 g/210L Test # 8                | 0.075  |
| 0.08 g/210L Test # (Repeat) 8       | 0.074  |
| Interferent Detect Test # 8         | INT    |
| Alcohol Free Test # 9               | 0.000  |
| 0.05 g/210L Test # 9                | 0.046  |
| 0.08 g/210L Test # 9                | 0.075  |
| 0.08 g/210L Test # (Repeat) 9       | 0.075  |
| Interferent Detect Test # 9         | INT    |
| Alcohol Free Test # 10              | 0.000  |
| 0.05 g/210L Test # 10               | 0.046  |
| 0.08 g/210L Test # 10               | 0.074  |
| 0.08 g/210L Test # (Repeat) 10      | 0.074  |
| Interferent Detect Test # 10        | INT    |
| 0.05 g/210L Test (Std Dev)          | 0.0000 |
| 0.08 g/210L Test (Std Dev)          | 0.0005 |
| 0.08 g/210L Test (Std Dev) (Repeat) | 0.0005 |

#### Remarks:

08: Control Outside Tolerance,CAL ADJUSTMENT NEEDED Non-compliance: .

Data Download Date/Time 09/16/2022 10:25

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                  |                        |                                   |  |
|----------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> PUTNAM COUNTY SO   | <b>Date</b> 09/16/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> NICODEMO, PHIL, | <b>Time</b> 10:18:37   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 08121080A1 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 09/16/2022 10:25

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PUTNAM COUNTY SO | <b>Date</b> 10/25/2022 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SMITH,BRIAN,K | <b>Time</b> 11:18:34   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202112A    | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B    | <b>0.08 Gas Lot #</b> 28021080A2 |
| <b>Expiration</b> 12/07/2023 | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023 | <b>Expiration</b> 12/05/2023     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.042  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.048  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.073  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Test # (Repeat) 1    | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.20 g/210L Test # 1             | 0.179  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.199  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # 1      | INT    |
| 0.05 g/210L Test: OK                        | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| 0.08 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.20 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.049  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Test # 2             | 0.078  |
|                                             |        | 0.08 g/210L Test # (Repeat) 2    | 0.079  |
|                                             |        | 0.20 g/210L Test # 2             | 0.192  |
|                                             |        | 0.20 g/210L Test # (Repeat) 2    | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

00: Ambient Fail. LOOSE HOSE FITTING. LOOSE HOSE FITTING. LOOSE HOSE FITTING.

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

---

**Data Download Date/Time** 10/25/2022 11:23

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

**Agency** PUTNAM COUNTY SO  
**Inspector** SMITH,BRIAN,K

**Date** 11/22/2022  
**Time** 13:48:12

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 28201080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/22/2022 14:02

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PUTNAM COUNTY SO | <b>Date</b> 12/25/2022 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SMITH,BRIAN,K | <b>Time</b> 15:57:41   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202112A    | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202206C    | <b>0.08 Gas Lot #</b> 28021080A2 |
| <b>Expiration</b> 12/07/2023 | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/15/2024 | <b>Expiration</b> 12/05/2023     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.067  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.187  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.199  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Interferent Detect Test # 1      | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 2            | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.072  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.079  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.20 g/210L Test # 2             | 0.197  |
|                                             |        | 0.20 g/210L Test # (Repeat) 2    | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

08:SIMULATOR POWER SWITCH TURNED OFF. 20:SIMULATOR POWER SWITCH TURNED OFF.

**Data Download Date/Time** 12/28/2022 13:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PUTNAM COUNTY SO     | <b>Date</b> 05/31/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GRAHAM, THOMAS, J | <b>Time</b> 13:18:50   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 05/31/2022 13:20

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

**Agency** PUTNAM COUNTY SO  
**Inspector** SMITH,BRIAN,K

**Date** 08/29/2022  
**Time** 23:40:41

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 28021080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/30/2022 08:27

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

**Agency** PUTNAM COUNTY SO  
**Inspector** SMITH,BRIAN,K

**Date** 09/27/2022  
**Time** 08:43:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 28021080A2  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/27/2022 14:34

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

**Agency** PUTNAM COUNTY SO      **Date** 10/25/2022      **Type** Agency Inspection  
**Inspector** SMITH,BRIAN,K      **Time** 09:48:53      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 28021080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.188  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.198  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20:LOOSE HOSE FITTING

**Data Download Date/Time** 10/25/2022 09:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

**Agency** PUTNAM COUNTY SO      **Date** 11/22/2022      **Type** Agency Inspection  
**Inspector** SMITH,BRIAN,K      **Time** 13:55:24      **Software** 8100.27

**0.05 Lot #** 202112A      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 28021080A2  
**Expiration** 12/07/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 12/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | No     | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

TOO MUCH M/A SOLUTION.

**Data Download Date/Time** 11/22/2022 14:50

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 01/21/2022 **Type** Agency Inspection  
**Inspector** STONE,JIMMIE,N **Time** 11:03:43 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1206004  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/21/2022 11:10

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 02/28/2022 **Type** Agency Inspection  
**Inspector** STONE,JIMMIE,N **Time** 11:53:59 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1206004  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/28/2022 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 03/03/2022 **Type** Agency Inspection  
**Inspector** STONE,JIMMIE,N **Time** 11:24:17 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1206004  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/03/2022 11:29

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 04/27/2022 **Type** Agency Inspection  
**Inspector** STONE,JIMMIE,N **Time** 20:07:58 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1206004  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/27/2022 20:41

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 05/31/2022 **Type** Agency Inspection  
**Inspector** STONE,JIMMIE,N **Time** 20:22:05 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1206004  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/31/2022 20:25

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. Date 06/22/2022 Type Agency Inspection  
Inspector STONE,JIMMIE,N Time 19:13:39 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202108D 0.20 Lot # 202106B 0.08 Gas Lot # 1206004  
Expiration 01/12/2023 Expiration 08/17/2023 Expiration 06/22/2023 Expiration 09/20/2022

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 06/22/2022 20:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 07/14/2022 **Type** Agency Inspection  
**Inspector** STONE,JIMMIE,N **Time** 19:49:37 **Software** 8100.27

**0.05 Lot #** 1206004 **0.08 Lot #** 1206004 **0.20 Lot #** 1206004 **0.08 Gas Lot #** 1206004  
**Expiration** 09/20/2022 **Expiration** 09/20/2022 **Expiration** 09/20/2022 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/14/2022 20:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 08/30/2022 **Type** Agency Inspection  
**Inspector** RAY, RAY, D **Time** 19:54:12 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: .

**Data Download Date/Time** 08/30/2022 21:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 08/30/2022 **Type** Agency Inspection  
**Inspector** RAY, ANTHONY, D **Time** 21:06:35 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1206004  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/20/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/30/2022 21:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

|                                    |                        |                                   |  |
|------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> VOLUSIA COUNTY S.O.  | <b>Date</b> 09/14/2022 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 14:59:02   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> 00521080A2 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 02/05/2023     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 09/14/2022 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 09/23/2022 **Type** Agency Inspection  
**Inspector** RAY, ANTHONY, D **Time** 14:57:41 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CHECKING RESULTS

**Data Download Date/Time** 09/23/2022 15:05

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

|                  |                     |             |            |                 |                   |
|------------------|---------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | VOLUSIA COUNTY S.O. | <b>Date</b> | 10/28/2022 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | RAY,ANTHONY,D       | <b>Time</b> | 16:07:21   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |     |
|-----------------------------|---|-------------------|-----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | Yes |
|-----------------------------|---|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

TIMED OUT TOOK TO LONG TO PROVIDE SAMPLE.

**Data Download Date/Time** 10/28/2022 17:35

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 10/28/2022 **Type** Agency Inspection  
**Inspector** RAY, ANTHONY, D **Time** 17:30:41 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.076  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.048  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.076  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.048  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.077  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: USED .08 SOLUTION INSTEAD OF .05 SOLUTION BY MISTAKE .

**Data Download Date/Time** 10/28/2022 17:35

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 11/30/2022 **Type** Agency Inspection  
**Inspector** STONE,JIMMIE,N **Time** 19:30:42 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/01/2022 13:56

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

**Agency** VOLUSIA COUNTY S.O. **Date** 12/23/2022 **Type** Agency Inspection  
**Inspector** STONE,JIMMIE,N **Time** 12:34:49 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/23/2022 13:49

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> VOLUSIA COUNTY SO    | <b>Date</b> 07/05/2022 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 15:00:09   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202201C    | <b>0.08 Lot #</b> 202201D    | <b>0.20 Lot #</b> 202201E    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 01/11/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 01/18/2024 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 07/05/2022 15:05

Printed: 05 January 2023

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

**Agency** VOLUSIA COUNTY SO  
**Inspector** STONE,JIMMIE,N

**Date** 07/14/2022  
**Time** 20:13:37

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 402419647      **0.08 Lot #** 402419647      **0.20 Lot #** 402419647      **0.08 Gas Lot #** 402419647  
**Expiration** 04/14/2025      **Expiration** 04/14/2025      **Expiration** 04/14/2025      **Expiration** 04/14/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/14/2022 20:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | VOLUSIA COUNTY SO | <b>Date</b>       | 08/30/2022 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | RAY,ANTHONY,D     | <b>Time</b>       | 21:20:30   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C           | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202106B           | <b>0.08 Gas Lot #</b> | 402419647  |
| <b>Expiration</b>           | 01/12/2023        | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 06/22/2023        | <b>Expiration</b>     | 04/14/2025 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/30/2022 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

**Agency** VOLUSIA COUNTY SO  
**Inspector** RAY, ANTHONY, D

**Date** 08/30/2022  
**Time** 22:21:11

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 402419647      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402419647  
**Expiration** 04/14/2025      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/14/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

FORGOT MOUTH ALCH.

**Data Download Date/Time** 08/30/2022 22:29

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

**Agency** VOLUSIA COUNTY SO  
**Inspector** RAY, ANTHONY, D

**Date** 09/23/2022  
**Time** 14:52:18

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402419647  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/14/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/23/2022 14:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

**Agency** VOLUSIA COUNTY SO **Date** 10/28/2022 **Type** Agency Inspection  
**Inspector** RAY,ANTHONY,D **Time** 16:15:52 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

CELL PHONE NEXT TO MACHINE

**Data Download Date/Time** 10/28/2022 17:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

|                  |                   |             |            |                 |                   |
|------------------|-------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | VOLUSIA COUNTY SO | <b>Date</b> | 10/28/2022 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | RAY,ANTHONY,D     | <b>Time</b> | 16:26:17   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |     |
|-----------------------------|---|-------------------|-----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | Yes |
|-----------------------------|---|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Range Exceeded .

**Data Download Date/Time** 10/28/2022 17:32

**Printed: 05 January 2023**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

**Agency** VOLUSIA COUNTY SO  
**Inspector** RAY, ANTHONY, D

**Date** 10/28/2022  
**Time** 17:27:26

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402419647  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/14/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 10/28/2022 17:32

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

**Agency** VOLUSIA COUNTY SO  
**Inspector** STONE,JIMMIE,N

**Date** 11/30/2022  
**Time** 19:25:19

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 402419647  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/14/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/30/2022 19:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

**Agency** VOLUSIA COUNTY SO  
**Inspector** STONE,JIMMIE,N

**Date** 12/23/2022  
**Time** 12:42:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 404219647  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/14/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/23/2022 12:58

**Printed: 05 January 2023**  
COBRA ©1997-2014 CMI, Inc.