

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

Agency: FDLE

Date/Time: 08/17/2023 00:11:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/17/2023 02:38

Air Blank	0.000	00:10
Control Test	0.000	00:11
Air Blank	0.000	00:11

Agency: FDLE

Date/Time: 08/17/2023 00:14:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/17/2023 02:38

Air Blank	0.000	00:12
Control Test	0.007	00:13
Air Blank	0.000	00:14

Agency: FDLE

Date/Time: 08/17/2023 00:16:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/17/2023 02:38

Air Blank	0.000	00:15
Control Test	0.050	00:15
Air Blank	0.000	00:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

Agency: FDLE

Date/Time: 08/17/2023 00:18:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/17/2023 02:38

Air Blank	0.000	00:17
Control Test	0.079	00:18
Air Blank	0.000	00:18

Agency: FDLE

Date/Time: 08/17/2023 00:20:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/17/2023 02:38

Air Blank	0.000	00:19
Control Test	0.204	00:20
Air Blank	0.000	00:20

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

Agency: FDLE

Date/Time: 01/18/2023 09:48:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2023 14:02

Air Blank	0.000	09:47
Control Test	0.049	09:47
Air Blank	0.000	09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

Agency: OKALOOSA COUNTY SO

Date/Time: 09/28/2023 16:01:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/28/2023 17:56

Air Blank	0.000	16:00
Control Test	0.198	16:01
Air Blank	0.000	16:01

Agency: OKALOOSA COUNTY SO

Date/Time: 10/31/2023 08:48:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/31/2023 10:33

Air Blank	0.000	08:47
Control Test	0.200	08:48
Air Blank	0.000	08:48

Agency: OKALOOSA COUNTY SO

Date/Time: 11/29/2023 13:51:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/29/2023 15:32

Air Blank	0.000	13:49
Control Test	0.051	13:50
Air Blank	0.000	13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

Agency: OKALOOSA COUNTY SO

Date/Time: 12/01/2023 15:22:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/01/2023 17:10

Air Blank	0.000	15:20
Control Test	0.197	15:21
Air Blank	0.000	15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

Agency: VALPARAISO PD

Date/Time: 09/18/2023 23:35:26

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 09/19/2023 01:37

Air Blank	0.000	23:34
ITP Check	0.081	23:34
Air Blank	0.000	23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

Agency: FL HIGHWAY PATROL

Date/Time: 06/20/2023 13:47:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/20/2023 15:54

Air Blank	0.000	13:46
Control Test	0.000	13:47
Air Blank	0.000	13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 08:49:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:30

Air Blank	0.000	08:48
Control Test	0.189	08:49
Air Blank	0.000	08:49

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:15:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:30

Air Blank	0.000	09:14
Control Test	0.192	09:15
Air Blank	0.000	09:15

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:17:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:30

Air Blank	0.000	09:16
Control Test	0.075	09:17
Air Blank	0.000	09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:31:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:30

Air Blank	0.000	09:30
Control Test	0.076	09:31
Air Blank	0.000	09:31

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:33:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:30

Air Blank	0.000	09:32
Control Test	0.195	09:33
Air Blank	0.000	09:33

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:35:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:30

Air Blank	0.000	09:34
Control Test	0.046	09:34
Air Blank	0.000	09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:37:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:30

Air Blank	0.000	09:35
Control Test	0.073	09:36
Air Blank	0.000	09:37

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:40:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:31

Air Blank	0.000	09:39
Control Test	0.074	09:39
Air Blank	0.000	09:40

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:54:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:31

Air Blank	0.000	09:53
Control Test	0.079	09:54
Air Blank	0.000	09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 09:58:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:31

Air Blank	0.000	09:57
Control Test	0.193	09:58
Air Blank	0.000	09:58

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 10:03:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:31

Air Blank	0.000	10:02
Control Test	0.048	10:03
Air Blank	0.000	10:03

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 10:05:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:31

Air Blank	0.000	10:04
Control Test	0.000	10:05
Air Blank	0.000	10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 10:07:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:31

Air Blank	0.000	10:06
Control Test	0.000	10:07
Air Blank	0.000	10:07

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 10:11:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:31

Air Blank	0.000	10:09
Control Test	0.000	10:10
Air Blank	0.000	10:11

Agency: OKALOOSA COUNTY SO

Date/Time: 01/31/2023 10:23:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/31/2023 12:31

Air Blank	0.000	10:22
Control Test	0.000	10:23
Air Blank	0.000	10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 02/27/2023 15:16:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2023 08:14

Air Blank	0.000	15:15
Control Test	0.188	15:16
Air Blank	0.000	15:16

Agency: OKALOOSA COUNTY SO

Date/Time: 02/27/2023 15:19:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2023 08:15

Air Blank	0.000	15:18
Control Test	0.000	15:19
Air Blank	0.000	15:19

Agency: OKALOOSA COUNTY SO

Date/Time: 02/27/2023 15:21:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2023 08:15

Air Blank	0.000	15:20
Control Test	0.075	15:21
Air Blank	0.000	15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 02/27/2023 15:23:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2023 08:15

Air Blank	0.000	15:22
Control Test	0.046	15:23
Air Blank	0.000	15:23

Agency: OKALOOSA COUNTY SO

Date/Time: 03/30/2023 17:04:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/31/2023 10:35

Air Blank	0.000	17:03
Control Test	0.198	17:03
Air Blank	0.000	17:04

Agency: OKALOOSA COUNTY SO

Date/Time: 03/30/2023 17:06:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/31/2023 10:35

Air Blank	0.000	17:05
Control Test	0.077	17:05
Air Blank	0.000	17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 03/30/2023 17:08:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/31/2023 10:35

Air Blank	0.000	17:07
Control Test	0.000	17:07
Air Blank	0.000	17:08

Agency: OKALOOSA COUNTY SO

Date/Time: 04/24/2023 16:01:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/24/2023 18:05

Air Blank	0.000	15:59
Control Test	0.185	16:00
Air Blank	0.000	16:01

Agency: OKALOOSA COUNTY SO

Date/Time: 04/24/2023 16:08:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/24/2023 18:05

Air Blank	0.000	16:07
Control Test	0.045	16:08
Air Blank	0.000	16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 04/24/2023 16:11:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/24/2023 18:05

Air Blank	0.000	16:10
Control Test	0.000	16:11
Air Blank	0.000	16:11

Agency: OKALOOSA COUNTY SO

Date/Time: 04/24/2023 16:14:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/24/2023 18:05

Air Blank	0.000	16:13
Control Test	0.193	16:14
Air Blank	0.000	16:14

Agency: OKALOOSA COUNTY SO

Date/Time: 04/24/2023 16:16:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/24/2023 18:05

Air Blank	0.000	16:15
Control Test	0.078	16:16
Air Blank	0.000	16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 05/23/2023 13:38:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/23/2023 15:48

Air Blank	0.000	13:37
Control Test	0.182	13:37
Air Blank	0.000	13:38

Agency: OKALOOSA COUNTY SO

Date/Time: 05/23/2023 13:47:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/23/2023 15:48

Air Blank	0.000	13:45
Control Test	0.190	13:46
Air Blank	0.000	13:47

Agency: OKALOOSA COUNTY SO

Date/Time: 05/23/2023 14:01:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/23/2023 15:48

Air Blank	0.000	13:59
Control Test	0.199	14:00
Air Blank	0.000	14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 05/23/2023 14:03:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/23/2023 15:48

Air Blank	0.000	14:01
Control Test	0.000	14:02
Air Blank	0.000	14:03

Agency: OKALOOSA COUNTY SO

Date/Time: 06/29/2023 17:31:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2023 19:18

Air Blank	0.000	17:30
Control Test	0.196	17:31
Air Blank	0.000	17:31

Agency: OKALOOSA COUNTY SO

Date/Time: 06/29/2023 17:35:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/29/2023 19:18

Air Blank	0.000	17:33
Control Test	0.000	17:34
Air Blank	0.000	17:35

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 07/13/2023 13:01:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2023 14:57

Air Blank	0.000	13:00
Control Test	0.197	13:01
Air Blank	0.000	13:01

Agency: OKALOOSA COUNTY SO

Date/Time: 07/13/2023 13:14:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2023 14:57

Air Blank	0.000	13:13
Control Test	0.000	13:13
Air Blank	0.000	13:14

Agency: OKALOOSA COUNTY SO

Date/Time: 08/09/2023 14:01:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/30/2023 12:29

Air Blank	0.000	14:00
Control Test	0.000	14:01
Air Blank	0.000	14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

Agency: OKALOOSA COUNTY SO

Date/Time: 08/30/2023 09:08:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/30/2023 12:29

Air Blank	0.000	09:07
Control Test	0.194	09:08
Air Blank	0.000	09:08

Agency: OKALOOSA COUNTY SO

Date/Time: 08/30/2023 09:12:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/30/2023 12:29

Air Blank	0.000	09:11
Control Test	0.000	09:11
Air Blank	0.000	09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

Agency: SANTA ROSA COUNTY
SO

Date/Time: 06/08/2023 22:58:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/05/2023 18:34

Air Blank	0.000	22:57
Control Test	0.187	22:57
Air Blank	0.000	22:58

Agency: SANTA ROSA COUNTY
SO

Date/Time: 06/08/2023 23:00:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/05/2023 18:34

Air Blank	0.000	22:59
Control Test	0.189	22:59
Air Blank	0.000	23:00

Agency: SANTA ROSA COUNTY
SO

Date/Time: 06/08/2023 23:23:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/05/2023 18:34

Air Blank	0.000	23:21
Control Test	0.190	23:22
Air Blank	0.000	23:23

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

Agency: SANTA ROSA COUNTY
SO

Date/Time: 06/08/2023 23:24:53

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/05/2023 18:34

Air Blank	0.000	23:24
Control Test	0.077	23:24
Air Blank	0.000	23:24

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

Agency: COLUMBIA COUNTY SO
Date/Time: 03/04/2023 06:01:26
Standard Type: Dry Gas
Event Source: Local
Data Download Date/Time: 03/27/2023 08:42

Air Blank	0.000	06:00
Control Test	0.078	06:00
Air Blank	0.000	06:01

Agency: COLUMBIA COUNTY SO
Date/Time: 03/04/2023 06:02:48
Standard Type: Dry Gas
Event Source: Local
Data Download Date/Time: 03/27/2023 08:42

Air Blank	0.000	06:01
Control Test	0.077	06:02
Air Blank	0.000	06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

Agency: FHP

Date/Time: 06/21/2023 10:24:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/21/2023 13:00

Air Blank	0.000	10:23
Control Test	0.079	10:23
Air Blank	0.000	10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

Agency: FHP

Date/Time: 05/01/2023 09:57:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/01/2023 11:50

Air Blank	0.000	09:55
Control Test	0.048	09:56
Air Blank	0.000	09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

Agency: ST LUCIE COUNTY SO

Date/Time: 08/29/2023 18:31:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/27/2023 14:29

Air Blank	0.000	18:30
Control Test	0.049	18:31
Air Blank	0.000	18:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 02/25/2023 15:35:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2023 21:49

Air Blank	0.000	15:34
Control Test	0.079	15:35
Air Blank	0.000	15:35

Agency: GULF COUNTY SO

Date/Time: 05/17/2023 19:14:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/17/2023 21:01

Air Blank	0.000	19:13
Control Test	0.034	19:14
Air Blank	0.000	19:14

Agency: GULF COUNTY SO

Date/Time: 05/17/2023 19:17:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/17/2023 21:01

Air Blank	0.000	19:15
Control Test	0.000	19:16
Air Blank	0.000	19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 05/17/2023 19:19:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/17/2023 21:01

Air Blank	0.000	19:18
Control Test	0.079	19:18
Air Blank	0.000	19:19

Agency: GULF COUNTY SO

Date/Time: 05/17/2023 19:21:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/17/2023 21:01

Air Blank	0.000	19:20
Control Test	0.050	19:20
Air Blank	0.000	19:21

Agency: GULF COUNTY SO

Date/Time: 07/29/2023 19:52:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/29/2023 20:59

Air Blank	0.000	19:51
Control Test	0.050	19:51
Air Blank	0.000	19:52

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 07/29/2023 19:54:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/29/2023 20:59

Air Blank	0.000	19:53
Control Test	0.000	19:54
Air Blank	0.000	19:54

Agency: GULF COUNTY SO

Date/Time: 08/10/2023 15:46:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/10/2023 16:38

Air Blank	0.000	15:45
Control Test	0.201	15:46
Air Blank	0.000	15:46

Agency: GULF COUNTY SO

Date/Time: 09/07/2023 20:41:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/07/2023 21:30

Air Blank	0.000	20:40
Control Test	0.201	20:40
Air Blank	0.000	20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 10/10/2023 14:08:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 14:55

Air Blank	0.000	14:07
Control Test	0.201	14:08
Air Blank	0.000	14:08

Agency: GULF COUNTY SO

Date/Time: 11/28/2023 18:50:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2023 19:35

Air Blank	0.000	18:49
Control Test	0.000	18:50
Air Blank	0.000	18:50

Agency: GULF COUNTY SO

Date/Time: 11/28/2023 18:52:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2023 19:36

Air Blank	0.000	18:51
Control Test	0.049	18:52
Air Blank	0.000	18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 11/28/2023 18:54:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2023 19:36

Air Blank	18:53
Air Blank	18:54

Agency: GULF COUNTY SO

Date/Time: 11/28/2023 18:56:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2023 19:36

Air Blank	0.000	18:54
Control Test	0.198	18:55
Air Blank	0.000	18:56

Agency: GULF COUNTY SO

Date/Time: 12/04/2023 21:28:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 22:18

Air Blank	0.000	21:26
Control Test	0.079	21:27
Air Blank		21:27
Air Blank	0.000	21:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

Agency: HERNANDO COUNTY
S.O.

Date/Time: 12/25/2023 17:45:00

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 12/29/2023 08:04

Air Blank	0.000	17:44
ITP Check	0.078	17:44
Air Blank	0.000	17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

Agency: HERNANDO COUNTY
S.O.

Date/Time: 07/19/2023 17:57:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/19/2023 19:48

Air Blank	0.000	17:56
Control Test	0.000	17:56
Air Blank	0.000	17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 01/25/2023 15:02:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 15:59

Air Blank	0.000	15:00
Control Test	0.000	15:01
Air Blank	0.000	15:02

Agency: LEESBURG P.D.

Date/Time: 01/25/2023 15:04:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 15:59

Air Blank	0.000	15:03
Control Test	0.049	15:04
Air Blank	0.000	15:04

Agency: LEESBURG P.D.

Date/Time: 01/25/2023 15:07:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 15:59

Air Blank	0.000	15:06
Control Test	0.079	15:07
Air Blank	0.000	15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 01/25/2023 15:10:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 15:59

Air Blank	0.000	15:09
Control Test	0.197	15:09
Air Blank	0.000	15:10

Agency: LEESBURG P.D.

Date/Time: 02/21/2023 12:42:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2023 13:45

Air Blank	0.000	12:41
Control Test	0.000	12:42
Air Blank	0.000	12:42

Agency: LEESBURG P.D.

Date/Time: 02/21/2023 12:44:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2023 13:45

Air Blank	0.000	12:43
Control Test	0.048	12:44
Air Blank	0.000	12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 02/21/2023 12:46:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2023 13:45

Air Blank	0.000	12:45
Control Test	0.079	12:45
Air Blank	0.000	12:46

Agency: LEESBURG P.D.

Date/Time: 02/21/2023 12:47:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2023 13:46

Air Blank		12:46
Air Blank	0.000	12:47

Agency: LEESBURG P.D.

Date/Time: 02/21/2023 12:51:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2023 13:46

Air Blank	0.000	12:50
Control Test	0.201	12:51
Air Blank	0.000	12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 03/29/2023 09:16:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/29/2023 10:34

Air Blank	0.000	09:15
Control Test	0.049	09:15
Air Blank	0.000	09:16

Agency: LEESBURG P.D.

Date/Time: 03/29/2023 09:19:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/29/2023 10:34

Air Blank	0.000	09:18
Control Test	0.079	09:19
Air Blank	0.000	09:19

Agency: LEESBURG P.D.

Date/Time: 03/29/2023 09:21:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/29/2023 10:34

Air Blank	0.000	09:20
Control Test	0.201	09:21
Air Blank	0.000	09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 04/19/2023 08:59:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2023 13:00

Air Blank	0.000	08:57
Control Test	0.000	08:58
Air Blank	0.000	08:59

Agency: LEESBURG P.D.

Date/Time: 04/19/2023 09:00:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2023 13:00

Air Blank	0.000	08:59
Control Test	0.049	09:00
Air Blank	0.000	09:00

Agency: LEESBURG P.D.

Date/Time: 04/19/2023 09:03:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2023 13:00

Air Blank	0.000	09:02
Control Test	0.079	09:03
Air Blank	0.000	09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 04/19/2023 09:05:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2023 13:00

Air Blank	0.000	09:04
Control Test	0.200	09:04
Air Blank	0.000	09:05

Agency: LEESBURG P.D.

Date/Time: 05/11/2023 08:09:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/11/2023 14:06

Air Blank	0.000	08:07
Control Test	0.000	08:08
Air Blank	0.000	08:09

Agency: LEESBURG P.D.

Date/Time: 05/11/2023 08:11:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/11/2023 14:06

Air Blank	0.000	08:10
Control Test	0.049	08:10
Air Blank	0.000	08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 05/11/2023 08:12:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/11/2023 14:06

Air Blank	0.000	08:11
Control Test	0.080	08:12
Air Blank	0.000	08:12

Agency: LEESBURG P.D.

Date/Time: 05/11/2023 08:13:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/11/2023 14:06

Air Blank		08:13
Air Blank	0.000	08:13

Agency: LEESBURG P.D.

Date/Time: 05/11/2023 08:15:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/11/2023 14:07

Air Blank	0.000	08:14
Control Test	0.200	08:15
Air Blank	0.000	08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 06/28/2023 10:08:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/28/2023 11:07

Air Blank	0.000	10:07
Control Test	0.000	10:08
Air Blank	0.000	10:08

Agency: LEESBURG P.D.

Date/Time: 06/28/2023 10:10:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/28/2023 11:07

Air Blank	0.000	10:09
Control Test	0.050	10:10
Air Blank	0.000	10:10

Agency: LEESBURG P.D.

Date/Time: 06/28/2023 10:12:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/28/2023 11:07

Air Blank	0.000	10:11
Control Test	0.080	10:11
Air Blank	0.000	10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 06/28/2023 10:14:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/28/2023 11:07

Air Blank	0.000	10:12
Control Test	0.200	10:13
Air Blank	0.000	10:14

Agency: LEESBURG P.D.

Date/Time: 07/27/2023 13:18:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2023 14:14

Air Blank	0.000	13:17
Control Test	0.000	13:18
Air Blank	0.000	13:18

Agency: LEESBURG P.D.

Date/Time: 07/27/2023 13:20:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2023 14:14

Air Blank	0.000	13:19
Control Test	0.050	13:20
Air Blank	0.000	13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 07/27/2023 13:22:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2023 14:14

Air Blank	0.000	13:21
Control Test	0.081	13:21
Air Blank	0.000	13:22

Agency: LEESBURG P.D.

Date/Time: 07/27/2023 13:24:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2023 14:14

Air Blank	0.000	13:22
Control Test	0.200	13:23
Air Blank	0.000	13:24

Agency: LEESBURG P.D.

Date/Time: 08/23/2023 13:53:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/23/2023 14:51

Air Blank	0.000	13:52
Control Test	0.000	13:52
Air Blank	0.000	13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 08/23/2023 13:55:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/23/2023 14:51

Air Blank	0.000	13:53
Control Test	0.050	13:54
Air Blank	0.000	13:55

Agency: LEESBURG P.D.

Date/Time: 08/23/2023 13:56:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/23/2023 14:51

Air Blank	0.000	13:55
Control Test	0.079	13:56
Air Blank	0.000	13:56

Agency: LEESBURG P.D.

Date/Time: 08/23/2023 13:57:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/23/2023 14:52

Air Blank		13:57
Air Blank	0.000	13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 08/23/2023 13:58:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/23/2023 14:52

Air Blank		13:58
Air Blank	0.000	13:58

Agency: LEESBURG P.D.

Date/Time: 08/23/2023 14:00:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/23/2023 14:52

Air Blank	0.000	13:59
Control Test	0.201	14:00
Air Blank	0.000	14:00

Agency: LEESBURG P.D.

Date/Time: 09/27/2023 12:40:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/27/2023 13:57

Air Blank	0.000	12:39
Control Test	0.000	12:40
Air Blank	0.000	12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 09/27/2023 12:42:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/27/2023 13:57

Air Blank	0.000	12:41
Control Test	0.050	12:42
Air Blank	0.000	12:42

Agency: LEESBURG P.D.

Date/Time: 09/27/2023 12:44:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/27/2023 13:57

Air Blank	0.000	12:43
Control Test	0.080	12:43
Air Blank	0.000	12:44

Agency: LEESBURG P.D.

Date/Time: 09/27/2023 12:46:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/27/2023 13:57

Air Blank	0.000	12:45
Control Test	0.199	12:45
Air Blank	0.000	12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 10/06/2023 10:45:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/06/2023 14:08

Air Blank	0.000	10:44
Control Test	0.000	10:44
Air Blank	0.000	10:45

Agency: LEESBURG P.D.

Date/Time: 10/06/2023 10:47:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/06/2023 14:08

Air Blank	0.000	10:46
Control Test	0.047	10:47
Air Blank	0.000	10:47

Agency: LEESBURG P.D.

Date/Time: 10/06/2023 10:49:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/06/2023 14:08

Air Blank		10:48
Air Blank		10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 10/06/2023 10:51:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/06/2023 14:08

Air Blank	0.000	10:50
Control Test	0.080	10:50
Air Blank	0.000	10:51

Agency: LEESBURG P.D.

Date/Time: 10/06/2023 10:53:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/06/2023 14:08

Air Blank	0.000	10:52
Control Test	0.200	10:53
Air Blank	0.000	10:53

Agency: LEESBURG P.D.

Date/Time: 11/14/2023 07:39:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/21/2023 08:16

Air Blank	0.000	07:38
Control Test	0.000	07:38
Air Blank	0.000	07:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 11/14/2023 07:44:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/21/2023 08:16

Air Blank	0.000	07:42
Control Test	0.049	07:43
Air Blank	0.000	07:44

Agency: LEESBURG P.D.

Date/Time: 11/14/2023 07:46:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/21/2023 08:16

Air Blank	0.000	07:45
Control Test	0.079	07:45
Air Blank	0.000	07:46

Agency: LEESBURG P.D.

Date/Time: 11/14/2023 07:49:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/21/2023 08:16

Air Blank	0.000	07:47
Control Test	0.200	07:48
Air Blank	0.000	07:49

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 12/26/2023 08:52:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/26/2023 10:02

Air Blank	0.000	08:51
Control Test	0.000	08:52
Air Blank		08:52

Agency: LEESBURG P.D.

Date/Time: 12/26/2023 08:54:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/26/2023 10:02

Air Blank	0.000	08:53
Control Test	0.047	08:54
Air Blank	0.000	08:54

Agency: LEESBURG P.D.

Date/Time: 12/26/2023 08:56:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/26/2023 10:02

Air Blank	0.000	08:55
Control Test	0.078	08:56
Air Blank	0.000	08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 12/26/2023 09:01:32

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/26/2023 10:02

Air Blank	0.000	09:00
Control Test	0.000	09:01
Air Blank	0.000	09:01

Agency: LEESBURG P.D.

Date/Time: 12/26/2023 09:03:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/26/2023 10:02

Air Blank	0.000	09:02
Control Test	0.198	09:02
Air Blank	0.000	09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency: MARION COUNTY SO

Date/Time: 05/14/2023 14:43:30

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/24/2023 11:59

Air Blank	0.000	14:42
Control Test	0.081	14:42
Air Blank	0.000	14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency: HILLSBOROUGH CO SO

Date/Time: 06/08/2023 09:37:18

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/05/2023 10:04

Air Blank	0.000	09:36
Control Test	0.082	09:36
Air Blank	0.000	09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 02/11/2023 21:27:14

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/14/2023 14:37

Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27

Agency: MARTIN COUNTY SO

Date/Time: 02/14/2023 13:04:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 14:37

Air Blank	0.000	13:02
Control Test	0.199	13:03
Air Blank	0.000	13:04

Agency: MARTIN COUNTY SO

Date/Time: 02/14/2023 13:06:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 14:38

Air Blank	0.000	13:05
Control Test	0.079	13:05
Air Blank	0.000	13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 02/14/2023 13:08:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 14:38

Air Blank	0.000	13:07
Control Test	0.048	13:08
Air Blank	0.000	13:08

Agency: MARTIN COUNTY SO

Date/Time: 02/14/2023 13:10:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 14:38

Air Blank	0.000	13:09
Control Test	0.000	13:09
Air Blank	0.000	13:10

Agency: MARTIN COUNTY SO

Date/Time: 02/14/2023 13:12:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 14:38

Air Blank	0.000	13:11
Control Test	0.000	13:11
Air Blank	0.000	13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 08/08/2023 18:46:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2023 20:25

Air Blank	0.000	18:45
Control Test	0.200	18:46
Air Blank	0.000	18:46

Agency: MARTIN COUNTY SO

Date/Time: 08/08/2023 18:55:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2023 20:25

Air Blank	0.000	18:54
Control Test	0.078	18:54
Air Blank	0.000	18:55

Agency: MARTIN COUNTY SO

Date/Time: 08/08/2023 18:57:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2023 20:25

Air Blank	0.000	18:56
Control Test	0.049	18:57
Air Blank	0.000	18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 08/08/2023 19:00:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2023 20:25

Air Blank	0.000	18:59
Control Test	0.000	18:59
Air Blank	0.000	19:00

Agency: MARTIN COUNTY SO

Date/Time: 08/08/2023 19:02:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/08/2023 20:25

Air Blank	0.000	19:01
Control Test	0.000	19:01
Air Blank	0.000	19:02

Agency: MARTIN COUNTY SO

Date/Time: 09/08/2023 14:59:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/08/2023 16:47

Air Blank	0.000	14:57
Control Test	0.200	14:58
Air Blank	0.000	14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 09/08/2023 15:02:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/08/2023 16:48

Air Blank	0.000	15:00
Control Test	0.200	15:01
Air Blank	0.000	15:02

Agency: MARTIN COUNTY SO

Date/Time: 09/08/2023 15:04:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/08/2023 16:48

Air Blank	0.000	15:03
Control Test	0.079	15:04
Air Blank	0.000	15:04

Agency: MARTIN COUNTY SO

Date/Time: 09/08/2023 15:07:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/08/2023 16:48

Air Blank	0.000	15:05
Control Test	0.049	15:06
Air Blank	0.000	15:07

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

Agency: MARTIN COUNTY SO

Date/Time: 09/08/2023 15:09:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/08/2023 16:48

Air Blank	0.000	15:08
Control Test	0.000	15:08
Air Blank	0.000	15:09

Agency: MARTIN COUNTY SO

Date/Time: 09/08/2023 15:11:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/08/2023 16:48

Air Blank	0.000	15:10
Control Test	0.000	15:10
Air Blank	0.000	15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 03/09/2023 13:36:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 15:18

Air Blank	0.000	13:35
Control Test	0.199	13:36
Air Blank	0.000	13:36

Agency: MARTIN COUNTY SO

Date/Time: 03/09/2023 13:39:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 15:18

Air Blank	0.000	13:37
Control Test	0.079	13:38
Air Blank	0.000	13:39

Agency: MARTIN COUNTY SO

Date/Time: 03/09/2023 13:41:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 15:18

Air Blank	0.000	13:40
Control Test	0.049	13:40
Air Blank	0.000	13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 03/09/2023 13:43:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 15:18

Air Blank	0.000	13:42
Control Test	0.000	13:42
Air Blank	0.000	13:43

Agency: MARTIN COUNTY SO

Date/Time: 03/09/2023 13:45:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 15:18

Air Blank	0.000	13:44
Control Test	0.000	13:44
Air Blank	0.000	13:45

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:32:03

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/01/2023 19:33

Air Blank	0.000	16:31
Control Test	0.080	16:31
Air Blank	0.000	16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:33:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/01/2023 19:34

Air Blank	0.000	16:32
Control Test	0.000	16:33
Air Blank	0.000	16:33

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:36:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/01/2023 19:34

Air Blank	0.000	16:34
Control Test	0.000	16:35
Air Blank	0.000	16:36

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:41:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/01/2023 19:34

Air Blank	0.000	16:39
Control Test	0.200	16:40
Air Blank	0.000	16:41

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:43:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/01/2023 19:34

Air Blank	0.000	16:42
Control Test	0.079	16:42
Air Blank	0.000	16:43

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:45:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/01/2023 19:34

Air Blank	0.000	16:44
Control Test	0.048	16:44
Air Blank	0.000	16:45

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:47:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/01/2023 19:34

Air Blank	0.000	16:46
Control Test	0.000	16:47
Air Blank	0.000	16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:49:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/01/2023 19:34

Air Blank	0.000	16:48
Control Test	0.000	16:49
Air Blank	0.000	16:49

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH
PD

Date/Time: 04/07/2023 15:39:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/08/2023 16:03

Air Blank	0.000	15:38
Control Test	0.038	15:38
Air Blank	0.000	15:39

Agency: PANAMA CITY BEACH
PD

Date/Time: 04/07/2023 15:40:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/08/2023 16:03

Air Blank	0.000	15:39
Control Test	0.188	15:40
Air Blank	0.000	15:40

Agency: PANAMA CITY BEACH
PD

Date/Time: 11/06/2023 15:39:37

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/06/2023 17:32

Air Blank	0.000	15:38
Control Test	0.080	15:39
Air Blank	0.000	15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH
PD

Date/Time: 11/06/2023 15:42:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/06/2023 17:32

Air Blank	0.000	15:41
Control Test	0.000	15:42
Air Blank	0.000	15:42

Agency: PANAMA CITY BEACH
PD

Date/Time: 11/06/2023 15:44:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/06/2023 17:32

Air Blank	0.000	15:43
Control Test	0.049	15:43
Air Blank	0.000	15:44

Agency: PANAMA CITY BEACH
PD

Date/Time: 11/06/2023 15:46:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/06/2023 17:32

Air Blank	0.000	15:45
Control Test	0.079	15:46
Air Blank	0.000	15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH
PD

Date/Time: 11/06/2023 15:49:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/06/2023 17:32

Air Blank	0.000	15:47
Control Test	0.196	15:48
Air Blank	0.000	15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency: PANAMA CITY PD

Date/Time: 02/24/2023 16:48:43

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/23/2023 13:52

Air Blank		16:48
Air Blank	0.000	16:48

Agency: PANAMA CITY PD

Date/Time: 02/24/2023 16:50:15

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/23/2023 13:52

Air Blank	0.000	16:49
Control Test	0.077	16:49
Air Blank	0.000	16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

Agency: BLOUNTSTOWN PD

Date/Time: 11/28/2023 22:44:41

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/29/2023 00:37

Air Blank	0.000	22:43
Control Test	0.083	22:44
Air Blank	0.000	22:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 08:29:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:24

Air Blank	0.000	08:27
Control Test	0.047	08:28
Air Blank	0.000	08:29

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:10:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:24

Air Blank	0.000	10:09
Control Test	0.078	10:09
Air Blank	0.000	10:10

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:12:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:24

Air Blank	0.000	10:10
Control Test	0.082	10:11
Air Blank	0.000	10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:13:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:24

Air Blank	0.000	10:12
Control Test	0.049	10:13
Air Blank	0.000	10:13

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:15:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:24

Air Blank	0.000	10:14
Control Test	0.049	10:15
Air Blank	0.000	10:15

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:17:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:24

Air Blank	0.000	10:16
Control Test	0.050	10:17
Air Blank	0.000	10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:27:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:25

Air Blank	0.000	10:26
Control Test	0.201	10:27
Air Blank	0.000	10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency: MIAMI-DADE PD

Date/Time: 06/29/2023 01:20:44

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/06/2023 02:36

Air Blank	0.000	01:19
Control Test	0.079	01:20
Air Blank	0.000	01:20

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

Agency: MIAMI-DADE PD

Date/Time: 06/11/2023 01:15:19

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/06/2023 04:00

Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI-DADE PD

Date/Time: 07/29/2023 23:37:24

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/17/2023 02:41

Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37

Agency: MIAMI-DADE PD

Date/Time: 07/29/2023 23:38:50

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/17/2023 02:41

Air Blank	0.000	23:38
Control Test	0.077	23:38
Air Blank	0.000	23:38

Agency: MIAMI-DADE PD

Date/Time: 08/16/2023 23:40:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/17/2023 02:41

Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI-DADE PD

Date/Time: 09/01/2023 04:06:21

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/01/2023 05:49

Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06

Agency: MIAMI-DADE PD

Date/Time: 10/04/2023 23:19:09

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/05/2023 01:33

Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19

Agency: MIAMI-DADE PD

Date/Time: 10/04/2023 23:28:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/05/2023 01:33

Air Blank	0.000	23:27
Control Test	0.000	23:28
Air Blank	0.000	23:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI-DADE PD

Date/Time: 10/04/2023 23:30:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/05/2023 01:33

Air Blank	0.000	23:29
Control Test	0.000	23:30
Air Blank	0.000	23:30

Agency: MIAMI-DADE PD

Date/Time: 10/04/2023 23:35:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/05/2023 01:33

Air Blank	0.000	23:34
Control Test	0.050	23:34
Air Blank	0.000	23:35

Agency: MIAMI-DADE PD

Date/Time: 10/04/2023 23:38:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/05/2023 01:33

Air Blank	0.000	23:37
Control Test	0.080	23:38
Air Blank	0.000	23:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI-DADE PD

Date/Time: 10/04/2023 23:41:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/05/2023 01:33

Air Blank	0.000	23:40
Control Test	0.203	23:40
Air Blank	0.000	23:41

Agency: MIAMI-DADE PD

Date/Time: 11/03/2023 00:22:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/03/2023 02:21

Air Blank	0.000	00:21
Control Test	0.000	00:21
Air Blank	0.000	00:22

Agency: MIAMI-DADE PD

Date/Time: 11/03/2023 00:24:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/03/2023 02:21

Air Blank	0.000	00:23
Control Test	0.000	00:23
Air Blank	0.000	00:24

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI-DADE PD

Date/Time: 11/03/2023 00:26:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/03/2023 02:21

Air Blank	0.000	00:25
Control Test	0.050	00:26
Air Blank	0.000	00:26

Agency: MIAMI-DADE PD

Date/Time: 11/03/2023 00:28:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/03/2023 02:21

Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28

Agency: MIAMI-DADE PD

Date/Time: 11/03/2023 00:31:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/03/2023 02:21

Air Blank	0.000	00:29
Control Test	0.203	00:30
Air Blank	0.000	00:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI-DADE PD

Date/Time: 11/23/2023 22:57:36

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/01/2023 04:39

Air Blank	0.000	22:56
Control Test	0.000	22:57
Air Blank	0.000	22:57

Agency: MIAMI-DADE PD

Date/Time: 11/23/2023 22:59:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/01/2023 04:39

Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59

Agency: MIAMI-DADE PD

Date/Time: 11/23/2023 23:01:25

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/01/2023 04:39

Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI-DADE PD

Date/Time: 12/01/2023 02:14:25

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/01/2023 04:39

Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 01/16/2023 21:48:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2023 21:23

Air Blank	0.000	21:47
Control Test	0.000	21:48
Air Blank	0.000	21:48

Agency: PORT ST JOE PD

Date/Time: 01/16/2023 21:51:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2023 21:23

Air Blank	0.000	21:50
Control Test	0.049	21:50
Air Blank	0.000	21:51

Agency: PORT ST JOE PD

Date/Time: 02/25/2023 15:32:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2023 21:01

Air Blank	0.000	15:31
Control Test	0.000	15:32
Air Blank	0.000	15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 02/25/2023 15:35:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2023 21:01

Air Blank	0.000	15:34
Control Test	0.049	15:34
Air Blank	0.000	15:35

Agency: PORT ST JOE PD

Date/Time: 02/25/2023 15:37:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2023 21:01

Air Blank	0.000	15:36
Control Test	0.200	15:36
Air Blank	0.000	15:37

Agency: PORT ST JOE PD

Date/Time: 04/27/2023 15:59:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2023 18:46

Air Blank	0.000	15:58
Control Test	0.049	15:58
Air Blank	0.000	15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 04/27/2023 16:02:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2023 18:46

Air Blank	0.000	16:00
Control Test	0.000	16:01
Air Blank	0.000	16:02

Agency: PORT ST JOE PD

Date/Time: 04/27/2023 16:03:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2023 18:46

Air Blank	0.000	16:02
Control Test	0.079	16:03
Air Blank	0.000	16:03

Agency: PORT ST JOE PD

Date/Time: 04/27/2023 16:06:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/30/2023 18:47

Air Blank	0.000	16:04
Control Test	0.201	16:05
Air Blank	0.000	16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 05/17/2023 19:16:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 02:34

Air Blank	0.000	19:15
Control Test	0.047	19:16
Air Blank	0.000	19:16

Agency: PORT ST JOE PD

Date/Time: 05/17/2023 19:19:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 02:34

Air Blank	0.000	19:17
Control Test	0.197	19:18
Air Blank	0.000	19:19

Agency: PORT ST JOE PD

Date/Time: 06/30/2023 16:11:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/30/2023 17:25

Air Blank	0.000	16:09
Control Test	0.000	16:10
Air Blank	0.000	16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 06/30/2023 16:13:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/30/2023 17:25

Air Blank	0.000	16:12
Control Test	0.048	16:12
Air Blank	0.000	16:13

Agency: PORT ST JOE PD

Date/Time: 06/30/2023 16:15:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/30/2023 17:25

Air Blank	0.000	16:14
Control Test	0.079	16:14
Air Blank	0.000	16:15

Agency: PORT ST JOE PD

Date/Time: 06/30/2023 16:17:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/30/2023 17:25

Air Blank	0.000	16:16
Control Test	0.201	16:17
Air Blank	0.000	16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 07/29/2023 19:52:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/29/2023 21:02

Air Blank	0.000	19:50
Control Test	0.079	19:51
Air Blank	0.000	19:52

Agency: PORT ST JOE PD

Date/Time: 07/29/2023 19:54:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/29/2023 21:02

Air Blank	0.000	19:53
Control Test	0.199	19:54
Air Blank	0.000	19:54

Agency: PORT ST JOE PD

Date/Time: 08/10/2023 15:39:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/10/2023 16:32

Air Blank	0.000	15:38
Control Test	0.000	15:38
Air Blank	0.000	15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 08/10/2023 15:41:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/10/2023 16:32

Air Blank	0.000	15:40
Control Test	0.048	15:40
Air Blank	0.000	15:41

Agency: PORT ST JOE PD

Date/Time: 08/10/2023 15:43:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/10/2023 16:32

Air Blank	0.000	15:42
Control Test	0.080	15:43
Air Blank	0.000	15:43

Agency: PORT ST JOE PD

Date/Time: 09/07/2023 20:38:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/07/2023 21:28

Air Blank	0.000	20:37
Control Test	0.000	20:37
Air Blank	0.000	20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 09/07/2023 20:40:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/07/2023 21:28

Air Blank	0.000	20:39
Control Test	0.000	20:39
Air Blank	0.000	20:40

Agency: PORT ST JOE PD

Date/Time: 09/07/2023 20:42:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/07/2023 21:28

Air Blank	0.000	20:41
Control Test	0.050	20:41
Air Blank	0.000	20:42

Agency: PORT ST JOE PD

Date/Time: 09/07/2023 20:44:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/07/2023 21:28

Air Blank	0.000	20:42
Control Test	0.080	20:43
Air Blank	0.000	20:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 10/09/2023 00:08:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2023 01:08

Air Blank	0.000	00:07
Control Test	0.000	00:08
Air Blank	0.000	00:08

Agency: PORT ST JOE PD

Date/Time: 10/09/2023 00:10:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2023 01:08

Air Blank	0.000	00:09
Control Test	0.050	00:10
Air Blank	0.000	00:10

Agency: PORT ST JOE PD

Date/Time: 10/09/2023 00:13:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2023 01:08

Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 10/09/2023 00:14:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2023 01:08

Air Blank	0.000	00:14
Control Test	0.201	00:14
Air Blank	0.000	00:14

Agency: PORT ST JOE PD

Date/Time: 10/09/2023 00:16:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/09/2023 01:08

Air Blank	0.000	00:15
Control Test	0.203	00:16
Air Blank	0.000	00:16

Agency: PORT ST JOE PD

Date/Time: 10/10/2023 14:10:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 14:52

Air Blank	0.000	14:08
Control Test	0.000	14:09
Air Blank	0.000	14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 11/28/2023 18:51:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2023 19:40

Air Blank	0.000	18:50
Control Test	0.000	18:51
Air Blank	0.000	18:51

Agency: PORT ST JOE PD

Date/Time: 11/28/2023 18:53:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/28/2023 19:41

Air Blank	0.000	18:52
Control Test	0.074	18:53
Air Blank	0.000	18:53

Agency: PORT ST JOE PD

Date/Time: 12/04/2023 21:24:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 22:10

Air Blank	0.000	21:23
Control Test	0.000	21:24
Air Blank	0.000	21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 12/04/2023 21:26:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 22:10

Air Blank	0.000	21:25
Control Test	0.050	21:25
Air Blank	0.000	21:26

Agency: PORT ST JOE PD

Date/Time: 12/04/2023 21:28:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 22:10

Air Blank	0.000	21:27
Control Test	0.203	21:27
Air Blank	0.000	21:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

Agency: BREVARD COUNTY S.O.

Date/Time: 03/30/2023 09:00:10

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/18/2023 15:11

Air Blank	0.000	08:59
Control Test	0.000	08:59
Air Blank	0.000	09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

Agency: CHARLOTTE COUNTY
SO

Date/Time: 05/20/2023 19:34:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/17/2023 13:18

Air Blank	0.000	19:33
Control Test	0.000	19:34
Air Blank	0.000	19:34

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

Agency: LEON COUNTY SO

Date/Time: 02/08/2023 10:54:14

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/08/2023 11:55

Air Blank	0.000	10:53
Control Test	0.078	10:53
Air Blank	0.000	10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

Agency: PORT RICHEY PD

Date/Time: 06/18/2023 02:08:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/28/2023 20:08

Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:08

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 08/11/2023 14:38:27

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/14/2023 13:53

Air Blank	0.000	14:37
Control Test	0.076	14:37
Air Blank	0.000	14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

Agency: PINELLAS COUNTY SO

Date/Time: 02/03/2023 23:05:29

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/19/2023 03:13

Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05

Agency: PINELLAS COUNTY SO

Date/Time: 02/03/2023 23:14:32

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/19/2023 03:14

Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14

Agency: PINELLAS COUNTY SO

Date/Time: 02/03/2023 23:37:24

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/19/2023 03:14

Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

Agency: PINELLAS COUNTY SO

Date/Time: 02/03/2023 23:41:40

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/19/2023 03:14

Air Blank	0.000	23:40
Control Test	0.079	23:41
Air Blank	0.000	23:41

Agency: PINELLAS COUNTY SO

Date/Time: 12/29/2023 00:10:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/17/2024 13:02

Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:10

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 04/25/2023 08:22:14

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 05/25/2023 21:58

Air Blank	0.000	08:21
ITP Check	0.080	08:21
Air Blank	0.000	08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

Agency: PALM BAY P.D.

Date/Time: 08/08/2023 15:31:09

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/11/2023 12:15

Air Blank	0.000	15:30
Control Test	0.080	15:30
Air Blank	0.000	15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

Agency: BREVARD COUNTY S.O.

Date/Time: 11/15/2023 14:06:38

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/30/2023 10:31

Air Blank	0.000	14:05
Control Test	0.082	14:06
Air Blank	0.000	14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

Agency: JACKSONVILLE SO

Date/Time: 01/11/2023 07:07:22

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/02/2023 15:58

Air Blank	0.000	07:06
Control Test	0.084	07:06
Air Blank	0.000	07:07

Agency: JACKSONVILLE SO

Date/Time: 12/21/2023 10:59:03

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/28/2024 09:42

Air Blank	0.000	10:58
Control Test	0.079	10:58
Air Blank	0.000	10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 01/03/2023 14:29:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2023 15:25

Air Blank	0.000	14:28
Control Test	0.000	14:29
Air Blank	0.000	14:29

Agency: PINELLAS PARK PD

Date/Time: 01/03/2023 14:32:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2023 15:25

Air Blank	0.000	14:31
Control Test	0.000	14:31
Air Blank	0.000	14:32

Agency: PINELLAS PARK PD

Date/Time: 01/03/2023 14:34:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2023 15:25

Air Blank	0.000	14:33
Control Test	0.046	14:34
Air Blank	0.000	14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 01/03/2023 14:36:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2023 15:25

Air Blank	0.000	14:35
Control Test	0.076	14:36
Air Blank	0.000	14:36

Agency: PINELLAS PARK PD

Date/Time: 01/03/2023 14:39:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2023 15:25

Air Blank	0.000	14:38
Control Test	0.193	14:38
Air Blank	0.000	14:39

Agency: PINELLAS PARK PD

Date/Time: 02/20/2023 14:48:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 15:40

Air Blank	0.000	14:47
Control Test	0.000	14:48
Air Blank	0.000	14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 02/20/2023 14:50:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 15:40

Air Blank	0.000	14:49
Control Test	0.000	14:50
Air Blank	0.000	14:50

Agency: PINELLAS PARK PD

Date/Time: 02/20/2023 14:53:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 15:41

Air Blank	0.000	14:51
Control Test	0.045	14:52
Air Blank	0.000	14:53

Agency: PINELLAS PARK PD

Date/Time: 02/20/2023 14:55:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 15:41

Air Blank	0.000	14:54
Control Test	0.077	14:55
Air Blank	0.000	14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 02/20/2023 14:58:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 15:41

Air Blank	0.000	14:56
Control Test	0.194	14:57
Air Blank	0.000	14:58

Agency: PINELLAS PARK PD

Date/Time: 03/24/2023 11:23:52

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/24/2023 12:14

Air Blank	0.000	11:23
Control Test	0.078	11:23
Air Blank	0.000	11:23

Agency: PINELLAS PARK PD

Date/Time: 03/24/2023 11:26:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 12:14

Air Blank	0.000	11:24
Control Test	0.000	11:25
Air Blank	0.000	11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 03/24/2023 11:28:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 12:14

Air Blank	0.000	11:26
Control Test	0.000	11:27
Air Blank	0.000	11:28

Agency: PINELLAS PARK PD

Date/Time: 03/24/2023 11:30:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 12:14

Air Blank	0.000	11:29
Control Test	0.046	11:29
Air Blank	0.000	11:30

Agency: PINELLAS PARK PD

Date/Time: 03/24/2023 11:32:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 12:14

Air Blank	0.000	11:31
Control Test	0.076	11:31
Air Blank	0.000	11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 03/24/2023 11:34:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 12:15

Air Blank	0.000	11:33
Control Test	0.193	11:34
Air Blank	0.000	11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

Agency: JACKSONVILLE SO

Date/Time: 02/02/2023 11:28:19

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/15/2023 11:39

Air Blank	0.000	11:27
Control Test	0.081	11:27
Air Blank	0.000	11:28

Agency: JACKSONVILLE SO

Date/Time: 04/24/2023 06:35:35

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/24/2023 16:34

Air Blank	0.000	06:34
Control Test	0.082	06:35
Air Blank	0.000	06:35

Agency: JACKSONVILLE SO

Date/Time: 06/23/2023 09:31:00

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/20/2023 12:56

Air Blank	0.000	09:30
Control Test	0.082	09:30
Air Blank	0.000	09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 05/29/2023 18:56:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:06

Air Blank	0.000	18:55
Control Test	0.080	18:56
Air Blank	0.000	18:56

Agency: NASSAU CO SO

Date/Time: 05/29/2023 18:59:30

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:06

Air Blank	0.000	18:58
Control Test	0.078	18:58
Air Blank	0.000	18:59

Agency: NASSAU CO SO

Date/Time: 05/29/2023 19:04:29

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:06

Air Blank	0.000	19:03
Control Test	0.079	19:03
Air Blank	0.000	19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 05/29/2023 20:12:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:06

Air Blank	0.000	20:11
Control Test	0.079	20:12
Air Blank		20:12

Agency: NASSAU CO SO

Date/Time: 05/29/2023 20:15:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:06

Air Blank	0.000	20:15
Control Test	0.078	20:15
Air Blank		20:15

Agency: NASSAU CO SO

Date/Time: 05/30/2023 14:34:38

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:06

Air Blank	0.000	14:33
Control Test	0.079	14:34
Air Blank	0.000	14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 05/30/2023 14:45:33

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/27/2023 00:06

Air Blank	0.000	14:44
ITP Check	0.078	14:45
Air Blank	0.000	14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency: FERNANDINA BEACH PD

Date/Time: 03/13/2023 06:14:38

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/13/2023 08:58

Air Blank	0.000	06:13
Control Test	0.078	06:14
Air Blank	0.000	06:14

Agency: FERNANDINA BEACH PD

Date/Time: 03/13/2023 06:17:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/13/2023 08:58

Air Blank	0.000	06:16
Control Test	0.049	06:16
Air Blank	0.000	06:17

Agency: FERNANDINA BEACH PD

Date/Time: 03/13/2023 07:07:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/13/2023 08:58

Air Blank	0.000	07:06
Control Test	0.050	07:06
Air Blank	0.000	07:07

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency: HOLLYWOOD PD

Date/Time: 05/18/2023 13:43:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 15:10

Air Blank	0.000	13:42
Control Test	0.000	13:43
Air Blank	0.000	13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency: TARPON SPRINGS PD

Date/Time: 08/09/2023 14:23:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/15/2023 09:45

Air Blank	0.000	14:22
Control Test	0.076	14:22
Air Blank	0.000	14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency: JACKSONVILLE SO

Date/Time: 03/31/2023 15:32:59

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/21/2023 11:01

Air Blank	0.000	15:32
Control Test	0.084	15:32
Air Blank	0.000	15:32

Agency: JACKSONVILLE SO

Date/Time: 05/22/2023 06:03:07

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/24/2023 13:59

Air Blank	0.000	06:02
Control Test	0.082	06:02
Air Blank	0.000	06:03

Agency: JACKSONVILLE SO

Date/Time: 07/20/2023 10:10:57

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/17/2023 13:40

Air Blank	0.000	10:10
Control Test	0.082	10:10
Air Blank	0.000	10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 09/16/2023 10:05:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/16/2023 11:07

Air Blank	0.000	10:04
Control Test	0.196	10:04
Air Blank	0.000	10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 02/14/2023 12:48:17

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/14/2023 13:55

Air Blank	0.000	12:47
Control Test	0.000	12:47
Air Blank	0.000	12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 01/10/2023 09:02:04

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/10/2023 12:36

Air Blank		09:01
Air Blank	0.000	09:02

Agency: POLK CSO

Date/Time: 01/10/2023 09:03:25

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/10/2023 12:36

Air Blank	0.000	09:02
Control Test	0.077	09:02
Air Blank	0.000	09:03

Agency: POLK CSO

Date/Time: 01/10/2023 09:05:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:36

Air Blank	0.000	09:04
Control Test	0.045	09:05
Air Blank	0.000	09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 01/10/2023 09:08:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:36

Air Blank	0.000	09:07
Control Test	0.076	09:07
Air Blank	0.000	09:08

Agency: POLK CSO

Date/Time: 01/10/2023 09:10:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:37

Air Blank	0.000	09:09
Control Test	0.194	09:10
Air Blank	0.000	09:10

Agency: POLK CSO

Date/Time: 01/10/2023 09:26:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:37

Air Blank	0.000	09:24
Control Test	0.044	09:25
Air Blank	0.000	09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 01/10/2023 09:51:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:37

Air Blank	0.000	09:49
Control Test	0.045	09:50
Air Blank	0.000	09:51

Agency: POLK CSO

Date/Time: 01/10/2023 09:54:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:37

Air Blank	0.000	09:52
Control Test	0.045	09:53
Air Blank	0.000	09:54

Agency: POLK CSO

Date/Time: 02/14/2023 08:16:20

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/14/2023 09:15

Air Blank	0.000	08:15
Control Test	0.078	08:15
Air Blank	0.000	08:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 02/14/2023 08:19:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 09:15

Air Blank	0.000	08:18
Control Test	0.000	08:18
Air Blank	0.000	08:19

Agency: POLK CSO

Date/Time: 02/14/2023 08:21:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 09:15

Air Blank	0.000	08:20
Control Test	0.048	08:20
Air Blank	0.000	08:21

Agency: POLK CSO

Date/Time: 02/14/2023 08:23:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 09:15

Air Blank	0.000	08:22
Control Test	0.079	08:22
Air Blank	0.000	08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 02/14/2023 08:25:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2023 09:15

Air Blank	0.000	08:24
Control Test	0.195	08:25
Air Blank	0.000	08:25

Agency: POLK CSO

Date/Time: 03/14/2023 09:12:28

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/14/2023 11:04

Air Blank	0.000	09:11
Control Test	0.079	09:11
Air Blank	0.000	09:12

Agency: POLK CSO

Date/Time: 03/14/2023 09:14:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/14/2023 11:04

Air Blank	0.000	09:13
Control Test	0.047	09:13
Air Blank	0.000	09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 03/14/2023 09:16:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/14/2023 11:04

Air Blank	0.000	09:15
Control Test	0.077	09:16
Air Blank	0.000	09:16

Agency: POLK CSO

Date/Time: 03/14/2023 09:18:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/14/2023 11:04

Air Blank	0.000	09:17
Control Test	0.193	09:18
Air Blank	0.000	09:18

Agency: POLK CSO

Date/Time: 04/11/2023 09:47:27

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/11/2023 11:07

Air Blank	0.000	09:46
Control Test	0.079	09:46
Air Blank	0.000	09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 04/11/2023 09:49:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/11/2023 11:07

Air Blank	0.000	09:48
Control Test	0.047	09:49
Air Blank	0.000	09:49

Agency: POLK CSO

Date/Time: 04/11/2023 09:51:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/11/2023 11:08

Air Blank	0.000	09:50
Control Test	0.078	09:51
Air Blank	0.000	09:51

Agency: POLK CSO

Date/Time: 04/11/2023 09:53:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/11/2023 11:08

Air Blank	0.000	09:52
Control Test	0.193	09:53
Air Blank	0.000	09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 05/18/2023 09:03:56

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/18/2023 12:01

Air Blank	0.000	09:03
Control Test	0.081	09:03
Air Blank	0.000	09:03

Agency: POLK CSO

Date/Time: 05/18/2023 09:05:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 12:01

Air Blank	0.000	09:04
Control Test	0.047	09:05
Air Blank	0.000	09:05

Agency: POLK CSO

Date/Time: 05/18/2023 09:07:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 12:01

Air Blank	0.000	09:06
Control Test	0.078	09:07
Air Blank	0.000	09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 05/18/2023 09:10:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 12:01

Air Blank	0.000	09:08
Control Test	0.193	09:09
Air Blank	0.000	09:10

Agency: POLK CSO

Date/Time: 06/13/2023 09:37:16

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/13/2023 11:47

Air Blank	0.000	09:36
Control Test	0.081	09:36
Air Blank	0.000	09:37

Agency: POLK CSO

Date/Time: 06/13/2023 09:39:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/13/2023 11:47

Air Blank	0.000	09:37
Control Test	0.046	09:38
Air Blank	0.000	09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 06/13/2023 09:41:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/13/2023 11:47

Air Blank	0.000	09:39
Control Test	0.194	09:40
Air Blank	0.000	09:41

Agency: POLK CSO

Date/Time: 06/13/2023 09:42:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/13/2023 11:47

Air Blank	0.000	09:41
Control Test	0.078	09:42
Air Blank	0.000	09:42

Agency: POLK CSO

Date/Time: 07/13/2023 09:30:21

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/13/2023 10:43

Air Blank	0.000	09:29
Control Test	0.080	09:29
Air Blank	0.000	09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 07/13/2023 09:32:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2023 10:43

Air Blank	0.000	09:31
Control Test	0.047	09:31
Air Blank	0.000	09:32

Agency: POLK CSO

Date/Time: 07/13/2023 09:34:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2023 10:43

Air Blank	0.000	09:33
Control Test	0.195	09:34
Air Blank	0.000	09:34

Agency: POLK CSO

Date/Time: 07/13/2023 09:36:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2023 10:43

Air Blank	0.000	09:35
Control Test	0.078	09:36
Air Blank	0.000	09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 08/15/2023 09:04:13

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/15/2023 10:24

Air Blank	0.000	09:03
Control Test	0.080	09:03
Air Blank	0.000	09:04

Agency: POLK CSO

Date/Time: 08/15/2023 09:06:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/15/2023 10:24

Air Blank	0.000	09:04
Control Test	0.047	09:05
Air Blank	0.000	09:06

Agency: POLK CSO

Date/Time: 08/15/2023 09:08:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/15/2023 10:24

Air Blank	0.000	09:06
Control Test	0.192	09:07
Air Blank	0.000	09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 08/15/2023 09:09:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/15/2023 10:24

Air Blank	0.000	09:08
Control Test	0.078	09:09
Air Blank	0.000	09:09

Agency: POLK CSO

Date/Time: 09/12/2023 10:04:15

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/12/2023 11:14

Air Blank	0.000	10:03
Control Test	0.080	10:03
Air Blank	0.000	10:04

Agency: POLK CSO

Date/Time: 09/12/2023 10:06:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2023 11:14

Air Blank	0.000	10:05
Control Test	0.047	10:05
Air Blank	0.000	10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 09/12/2023 10:08:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2023 11:14

Air Blank	0.000	10:07
Control Test	0.078	10:07
Air Blank	0.000	10:08

Agency: POLK CSO

Date/Time: 09/12/2023 10:10:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2023 11:15

Air Blank	0.000	10:09
Control Test	0.193	10:09
Air Blank	0.000	10:10

Agency: POLK CSO

Date/Time: 10/10/2023 09:14:48

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/10/2023 11:17

Air Blank	0.000	09:13
Control Test	0.081	09:14
Air Blank	0.000	09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 10/10/2023 09:16:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:17

Air Blank	0.000	09:15
Control Test	0.078	09:16
Air Blank	0.000	09:16

Agency: POLK CSO

Date/Time: 10/10/2023 09:18:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:17

Air Blank	0.000	09:17
Control Test	0.047	09:18
Air Blank	0.000	09:18

Agency: POLK CSO

Date/Time: 10/10/2023 09:21:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:17

Air Blank	0.000	09:19
Control Test	0.193	09:20
Air Blank	0.000	09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 10/10/2023 10:02:54

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/10/2023 11:17

Air Blank	0.000	10:02
Control Test	0.081	10:02
Air Blank	0.000	10:02

Agency: POLK CSO

Date/Time: 10/10/2023 10:04:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:17

Air Blank	0.000	10:03
Control Test	0.047	10:04
Air Blank	0.000	10:04

Agency: POLK CSO

Date/Time: 10/10/2023 10:06:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:23

Air Blank	0.000	10:05
Control Test	0.079	10:06
Air Blank	0.000	10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 10/10/2023 10:08:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:23

Air Blank	0.000	10:07
Control Test	0.193	10:08
Air Blank	0.000	10:08

Agency: POLK CSO

Date/Time: 11/14/2023 09:16:14

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/14/2023 10:31

Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:16

Agency: POLK CSO

Date/Time: 11/14/2023 09:18:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 10:31

Air Blank	0.000	09:16
Control Test	0.045	09:17
Air Blank	0.000	09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 11/14/2023 09:19:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 10:31

Air Blank	0.000	09:18
Control Test	0.078	09:19
Air Blank	0.000	09:19

Agency: POLK CSO

Date/Time: 11/14/2023 09:20:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 10:31

Air Blank		09:20
Air Blank	0.000	09:20

Agency: POLK CSO

Date/Time: 11/14/2023 09:23:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 10:31

Air Blank	0.000	09:22
Control Test	0.192	09:23
Air Blank	0.000	09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 12/05/2023 08:53:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/05/2023 10:13

Air Blank	0.000	08:53
Control Test	0.080	08:53
Air Blank	0.000	08:53

Agency: POLK CSO

Date/Time: 12/05/2023 08:57:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:13

Air Blank	0.000	08:55
Control Test	0.000	08:56
Air Blank	0.000	08:57

Agency: POLK CSO

Date/Time: 12/05/2023 08:59:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:13

Air Blank	0.000	08:58
Control Test	0.048	08:58
Air Blank	0.000	08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 12/05/2023 09:01:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:13

Air Blank	0.000	09:00
Control Test	0.192	09:01
Air Blank	0.000	09:01

Agency: POLK CSO

Date/Time: 12/05/2023 09:03:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:13

Air Blank	0.000	09:02
Control Test	0.079	09:02
Air Blank	0.000	09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency: JACKSONVILLE SO

Date/Time: 07/05/2023 01:20:30

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/18/2023 14:45

Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency: FL HIGHWAY PATROL

Date/Time: 05/29/2023 23:08:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2023 00:21

Air Blank	0.000	23:07
Control Test	0.046	23:08
Air Blank	0.000	23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency: FL HIGHWAY PATROL

Date/Time: 10/30/2023 16:31:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/30/2023 17:47

Air Blank	0.000	16:30
Control Test	0.000	16:31
Air Blank	0.000	16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency: FL HIGHWAY PATROL

Date/Time: 07/05/2023 09:22:46

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/05/2023 11:28

Air Blank	0.000	09:21
Control Test	0.082	09:22
Air Blank	0.000	09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 12/28/2023 09:08:13

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/28/2023 09:56

Air Blank	0.000	09:07
Control Test	0.082	09:07
Air Blank	0.000	09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 01/10/2023 09:03:21

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/10/2023 12:22

Air Blank	0.000	09:02
Control Test	0.081	09:02
Air Blank	0.000	09:03

Agency: POLK COUNTY SO

Date/Time: 01/10/2023 09:05:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:22

Air Blank	0.000	09:04
Control Test	0.000	09:04
Air Blank	0.000	09:05

Agency: POLK COUNTY SO

Date/Time: 01/10/2023 09:07:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:22

Air Blank	0.000	09:05
Control Test	0.080	09:06
Air Blank	0.000	09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 01/10/2023 09:09:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:22

Air Blank	0.000	09:08
Control Test	0.047	09:08
Air Blank	0.000	09:09

Agency: POLK COUNTY SO

Date/Time: 01/10/2023 09:13:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/10/2023 12:22

Air Blank	0.000	09:12
Control Test	0.201	09:13
Air Blank	0.000	09:13

Agency: POLK COUNTY SO

Date/Time: 04/11/2023 09:48:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/11/2023 10:58

Air Blank	0.000	09:47
Control Test	0.049	09:47
Air Blank	0.000	09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 04/11/2023 09:50:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/11/2023 10:58

Air Blank	0.000	09:49
Control Test	0.081	09:50
Air Blank	0.000	09:50

Agency: POLK COUNTY SO

Date/Time: 04/11/2023 09:52:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/11/2023 10:59

Air Blank	0.000	09:51
Control Test	0.197	09:52
Air Blank	0.000	09:52

Agency: POLK COUNTY SO

Date/Time: 04/11/2023 09:54:08

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/11/2023 10:59

Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 05/18/2023 09:02:38

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/18/2023 12:04

Air Blank	0.000	09:01
Control Test	0.081	09:02
Air Blank	0.000	09:02

Agency: POLK COUNTY SO

Date/Time: 05/18/2023 09:04:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 12:04

Air Blank	0.000	09:03
Control Test	0.049	09:04
Air Blank	0.000	09:04

Agency: POLK COUNTY SO

Date/Time: 05/18/2023 09:06:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 12:04

Air Blank	0.000	09:05
Control Test	0.082	09:06
Air Blank	0.000	09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 05/18/2023 09:08:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/18/2023 12:04

Air Blank	0.000	09:07
Control Test	0.197	09:08
Air Blank	0.000	09:08

Agency: POLK COUNTY SO

Date/Time: 06/13/2023 09:36:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/13/2023 11:51

Air Blank	0.000	09:35
Control Test	0.080	09:35
Air Blank	0.000	09:36

Agency: POLK COUNTY SO

Date/Time: 06/13/2023 09:38:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/13/2023 11:51

Air Blank	0.000	09:37
Control Test	0.080	09:37
Air Blank	0.000	09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 06/13/2023 09:40:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/13/2023 11:51

Air Blank	0.000	09:38
Control Test	0.047	09:39
Air Blank	0.000	09:40

Agency: POLK COUNTY SO

Date/Time: 06/13/2023 09:42:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/13/2023 11:51

Air Blank	0.000	09:40
Control Test	0.198	09:41
Air Blank	0.000	09:42

Agency: POLK COUNTY SO

Date/Time: 07/13/2023 09:31:12

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/13/2023 10:53

Air Blank	0.000	09:30
Control Test	0.080	09:30
Air Blank	0.000	09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 07/13/2023 09:33:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2023 10:53

Air Blank	0.000	09:31
Control Test	0.081	09:32
Air Blank	0.000	09:33

Agency: POLK COUNTY SO

Date/Time: 07/13/2023 09:35:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2023 10:53

Air Blank	0.000	09:33
Control Test	0.048	09:34
Air Blank	0.000	09:35

Agency: POLK COUNTY SO

Date/Time: 07/13/2023 09:37:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/13/2023 10:53

Air Blank	0.000	09:35
Control Test	0.199	09:36
Air Blank	0.000	09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 08/15/2023 09:04:44

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/15/2023 10:19

Air Blank	0.000	09:03
Control Test	0.080	09:04
Air Blank	0.000	09:04

Agency: POLK COUNTY SO

Date/Time: 08/15/2023 09:06:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/15/2023 10:19

Air Blank	0.000	09:05
Control Test	0.081	09:06
Air Blank	0.000	09:06

Agency: POLK COUNTY SO

Date/Time: 08/15/2023 09:08:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/15/2023 10:19

Air Blank	0.000	09:07
Control Test	0.049	09:08
Air Blank	0.000	09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 08/15/2023 09:10:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/15/2023 10:19

Air Blank	0.000	09:09
Control Test	0.197	09:10
Air Blank	0.000	09:10

Agency: POLK COUNTY SO

Date/Time: 09/12/2023 10:00:02

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/12/2023 11:24

Air Blank	0.000	09:59
Control Test	0.080	09:59
Air Blank	0.000	10:00

Agency: POLK COUNTY SO

Date/Time: 09/12/2023 10:02:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2023 11:24

Air Blank	0.000	10:00
Control Test	0.048	10:01
Air Blank	0.000	10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 09/12/2023 10:04:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2023 11:24

Air Blank	0.000	10:03
Control Test	0.083	10:03
Air Blank	0.000	10:04

Agency: POLK COUNTY SO

Date/Time: 09/12/2023 10:06:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/12/2023 11:24

Air Blank	0.000	10:05
Control Test	0.199	10:05
Air Blank	0.000	10:06

Agency: POLK COUNTY SO

Date/Time: 10/10/2023 09:15:36

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/10/2023 11:08

Air Blank	0.000	09:14
Control Test	0.081	09:15
Air Blank	0.000	09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 10/10/2023 09:17:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:08

Air Blank	0.000	09:16
Control Test	0.048	09:17
Air Blank	0.000	09:17

Agency: POLK COUNTY SO

Date/Time: 10/10/2023 09:19:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:09

Air Blank	0.000	09:18
Control Test	0.198	09:19
Air Blank	0.000	09:19

Agency: POLK COUNTY SO

Date/Time: 10/10/2023 09:21:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/10/2023 11:09

Air Blank	0.000	09:20
Control Test	0.081	09:21
Air Blank	0.000	09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 11/14/2023 09:06:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/14/2023 10:35

Air Blank	0.000	09:05
Control Test	0.080	09:05
Air Blank	0.000	09:06

Agency: POLK COUNTY SO

Date/Time: 11/14/2023 09:08:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 10:35

Air Blank	0.000	09:07
Control Test	0.049	09:07
Air Blank	0.000	09:08

Agency: POLK COUNTY SO

Date/Time: 11/14/2023 09:10:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 10:35

Air Blank	0.000	09:09
Control Test	0.081	09:09
Air Blank	0.000	09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 11/14/2023 09:12:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 10:35

Air Blank	0.000	09:11
Control Test	0.199	09:11
Air Blank	0.000	09:12

Agency: POLK COUNTY SO

Date/Time: 12/05/2023 08:54:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/05/2023 10:20

Air Blank	0.000	08:53
Control Test	0.081	08:53
Air Blank	0.000	08:54

Agency: POLK COUNTY SO

Date/Time: 12/05/2023 08:56:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:20

Air Blank	0.000	08:55
Control Test	0.000	08:56
Air Blank	0.000	08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 12/05/2023 08:58:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:20

Air Blank	0.000	08:57
Control Test	0.049	08:57
Air Blank	0.000	08:58

Agency: POLK COUNTY SO

Date/Time: 12/05/2023 08:59:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:21

Air Blank		08:58
Air Blank		08:59

Agency: POLK COUNTY SO

Date/Time: 12/05/2023 09:01:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:21

Air Blank	0.000	09:00
Control Test	0.081	09:01
Air Blank	0.000	09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 12/05/2023 09:03:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/05/2023 10:21

Air Blank	0.000	09:02
Control Test	0.199	09:03
Air Blank	0.000	09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency: VOLUSIA COUNTY S.O.

Date/Time: 05/31/2023 14:06:25

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/13/2023 11:48

Air Blank	0.000	14:05
Control Test	0.078	14:05
Air Blank	0.000	14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 15:27:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:51

Air Blank	0.000	15:26
Control Test	0.000	15:26
Air Blank	0.000	15:27

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 15:29:13

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/20/2023 16:51

Air Blank	0.000	15:28
Control Test	0.081	15:28
Air Blank	0.000	15:29

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 15:49:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:51

Air Blank	0.000	15:48
Control Test	0.000	15:48
Air Blank	0.000	15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 15:50:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:51

Air Blank	0.000	15:49
Control Test	0.000	15:50
Air Blank	0.000	15:50

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 15:52:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:52

Air Blank	0.000	15:51
Control Test	0.050	15:52
Air Blank	0.000	15:52

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 15:55:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:52

Air Blank	0.000	15:54
Control Test	0.000	15:54
Air Blank	0.000	15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 15:58:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:52

Air Blank	0.000	15:57
Control Test	0.000	15:58
Air Blank	0.000	15:58

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 16:01:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:52

Air Blank	0.000	16:00
Control Test	0.000	16:01
Air Blank	0.000	16:01

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 16:03:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:52

Air Blank	0.000	16:02
Control Test	0.157	16:03
Air Blank	0.000	16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency: FLAGLER COUNTY SO

Date/Time: 02/20/2023 16:06:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2023 16:52

Air Blank	0.000	16:05
Control Test	0.196	16:06
Air Blank	0.000	16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 01/31/2023 13:20:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/03/2023 19:12

Air Blank	0.000	13:18
Control Test	0.051	13:19
Air Blank	0.000	13:20

Agency: GROVELAND PD

Date/Time: 01/31/2023 13:22:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/03/2023 19:13

Air Blank	0.000	13:20
Control Test	0.080	13:21
Air Blank	0.000	13:22

Agency: GROVELAND PD

Date/Time: 01/31/2023 13:24:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/03/2023 19:13

Air Blank	0.000	13:22
Control Test	0.199	13:23
Air Blank	0.000	13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 02/22/2023 10:23:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/22/2023 12:18

Air Blank	0.000	10:22
Control Test	0.050	10:22
Air Blank	0.000	10:23

Agency: GROVELAND PD

Date/Time: 02/22/2023 10:25:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/22/2023 12:18

Air Blank	0.000	10:24
Control Test	0.079	10:25
Air Blank	0.000	10:25

Agency: GROVELAND PD

Date/Time: 02/22/2023 10:27:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/22/2023 12:18

Air Blank	0.000	10:26
Control Test	0.199	10:27
Air Blank	0.000	10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 03/22/2023 11:51:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/22/2023 13:04

Air Blank	0.000	11:50
Control Test	0.050	11:51
Air Blank	0.000	11:51

Agency: GROVELAND PD

Date/Time: 03/22/2023 11:53:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/22/2023 13:32

Air Blank	0.000	11:52
Control Test	0.079	11:52
Air Blank	0.000	11:53

Agency: GROVELAND PD

Date/Time: 03/22/2023 11:55:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/22/2023 13:32

Air Blank	0.000	11:54
Control Test	0.198	11:54
Air Blank	0.000	11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 04/19/2023 12:12:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2023 13:27

Air Blank	0.000	12:11
Control Test	0.200	12:11
Air Blank	0.000	12:12

Agency: GROVELAND PD

Date/Time: 04/19/2023 12:14:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2023 13:27

Air Blank	0.000	12:13
Control Test	0.080	12:13
Air Blank	0.000	12:14

Agency: GROVELAND PD

Date/Time: 04/19/2023 12:16:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/19/2023 13:27

Air Blank	0.000	12:14
Control Test	0.050	12:15
Air Blank	0.000	12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 05/17/2023 11:21:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/17/2023 12:58

Air Blank	0.000	11:20
Control Test	0.050	11:21
Air Blank	0.000	11:21

Agency: GROVELAND PD

Date/Time: 05/17/2023 11:23:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/17/2023 12:58

Air Blank	0.000	11:22
Control Test	0.079	11:23
Air Blank	0.000	11:23

Agency: GROVELAND PD

Date/Time: 05/17/2023 11:25:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/17/2023 12:58

Air Blank	0.000	11:24
Control Test	0.199	11:25
Air Blank	0.000	11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 06/21/2023 15:01:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/21/2023 15:50

Air Blank	0.000	15:00
Control Test	0.048	15:01
Air Blank	0.000	15:01

Agency: GROVELAND PD

Date/Time: 06/21/2023 15:04:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/21/2023 15:50

Air Blank	0.000	15:02
Control Test	0.078	15:03
Air Blank	0.000	15:04

Agency: GROVELAND PD

Date/Time: 06/21/2023 15:06:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/21/2023 15:50

Air Blank	0.000	15:05
Control Test	0.198	15:06
Air Blank	0.000	15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 07/21/2023 15:51:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2023 18:09

Air Blank	0.000	15:49
Control Test	0.200	15:50
Air Blank	0.000	15:51

Agency: GROVELAND PD

Date/Time: 07/21/2023 15:53:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2023 18:09

Air Blank	0.000	15:52
Control Test	0.078	15:52
Air Blank	0.000	15:53

Agency: GROVELAND PD

Date/Time: 07/21/2023 15:55:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2023 18:09

Air Blank	0.000	15:53
Control Test	0.049	15:54
Air Blank	0.000	15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 08/18/2023 12:26:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/18/2023 13:10

Air Blank	0.000	12:25
Control Test	0.050	12:26
Air Blank	0.000	12:26

Agency: GROVELAND PD

Date/Time: 08/18/2023 12:28:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/18/2023 13:10

Air Blank	0.000	12:27
Control Test	0.079	12:28
Air Blank	0.000	12:28

Agency: GROVELAND PD

Date/Time: 08/18/2023 12:30:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/18/2023 13:10

Air Blank	0.000	12:29
Control Test	0.199	12:30
Air Blank	0.000	12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 09/25/2023 15:28:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 16:08

Air Blank	0.000	15:26
Control Test	0.080	15:27
Air Blank	0.000	15:28

Agency: GROVELAND PD

Date/Time: 10/18/2023 14:09:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/18/2023 15:12

Air Blank	0.000	14:08
Control Test	0.199	14:08
Air Blank	0.000	14:09

Agency: GROVELAND PD

Date/Time: 10/18/2023 14:11:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/18/2023 15:12

Air Blank	0.000	14:10
Control Test	0.078	14:10
Air Blank	0.000	14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 10/18/2023 14:13:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/18/2023 15:12

Air Blank	0.000	14:11
Control Test	0.049	14:12
Air Blank	0.000	14:13

Agency: GROVELAND PD

Date/Time: 12/04/2023 12:53:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 13:55

Air Blank	0.000	12:51
Control Test	0.199	12:52
Air Blank	0.000	12:53

Agency: GROVELAND PD

Date/Time: 12/04/2023 12:55:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 13:55

Air Blank	0.000	12:53
Control Test	0.078	12:54
Air Blank	0.000	12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 12/04/2023 12:56:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 13:55

Air Blank		12:55
Air Blank	0.000	12:56

Agency: GROVELAND PD

Date/Time: 12/04/2023 12:57:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 13:55

Air Blank	0.000	12:56
Control Test	0.050	12:57
Air Blank	0.000	12:57

Agency: GROVELAND PD

Date/Time: 12/04/2023 13:01:51

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/04/2023 13:55

Air Blank	0.000	13:01
Control Test	0.000	13:01
Air Blank	0.000	13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 12/04/2023 13:03:11

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/04/2023 13:55

Air Blank	0.000	13:02
Control Test	0.080	13:02
Air Blank	0.000	13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 04/06/2023 17:14:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/07/2023 13:01

Air Blank	0.000	17:13
Control Test	0.199	17:13
Air Blank	0.000	17:14

Agency: MARTIN COUNTY SO

Date/Time: 04/06/2023 17:18:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/07/2023 13:01

Air Blank	0.000	17:17
Control Test	0.078	17:17
Air Blank	0.000	17:18

Agency: MARTIN COUNTY SO

Date/Time: 04/06/2023 17:21:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/07/2023 13:01

Air Blank	0.000	17:19
Control Test	0.049	17:20
Air Blank	0.000	17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 04/06/2023 17:24:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/07/2023 13:01

Air Blank	0.000	17:23
Control Test	0.000	17:24
Air Blank	0.000	17:24

Agency: MARTIN COUNTY SO

Date/Time: 04/06/2023 17:26:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/07/2023 13:01

Air Blank	0.000	17:25
Control Test	0.000	17:26
Air Blank	0.000	17:26

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 21:48:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/02/2023 14:57

Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:00:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:09

Air Blank	0.000	20:59
Control Test	0.000	20:59
Air Blank	0.000	21:00

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:02:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:09

Air Blank	0.000	21:01
Control Test	0.199	21:01
Air Blank	0.000	21:02

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:04:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:09

Air Blank	0.000	21:03
Control Test	0.078	21:04
Air Blank	0.000	21:04

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:06:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:09

Air Blank	0.000	21:05
Control Test	0.012	21:06
Air Blank	0.000	21:06

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:09:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:09

Air Blank	0.000	21:08
Control Test	0.047	21:09
Air Blank	0.000	21:09

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:11:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:09

Air Blank	0.000	21:10
Control Test	0.048	21:10
Air Blank	0.000	21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:18:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:10

Air Blank	0.000	21:17
Control Test	0.049	21:17
Air Blank	0.000	21:18

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:20:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:10

Air Blank	0.000	21:19
Control Test	0.000	21:19
Air Blank	0.000	21:20

Agency: MARTIN COUNTY SO

Date/Time: 10/19/2023 21:22:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/19/2023 23:10

Air Blank	0.000	21:21
Control Test	0.000	21:22
Air Blank	0.000	21:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency: PANAMA CITY PD

Date/Time: 02/24/2023 15:57:28

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/24/2023 17:25

Air Blank	0.000	15:56
Control Test	0.077	15:56
Air Blank	0.000	15:57

Agency: PANAMA CITY PD

Date/Time: 02/27/2023 14:32:10

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/10/2023 16:29

Air Blank	0.000	14:31
Control Test	0.076	14:31
Air Blank	0.000	14:32

Agency: PANAMA CITY PD

Date/Time: 02/27/2023 14:34:30

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 04/10/2023 16:29

Air Blank	0.000	14:33
ITP Check	0.080	14:34
Air Blank	0.000	14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency: PANAMA CITY PD

Date/Time: 04/10/2023 15:43:56

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/27/2023 15:21

Air Blank	0.000	15:43
Control Test	0.080	15:43
Air Blank	0.000	15:43

Agency: PANAMA CITY PD

Date/Time: 07/11/2023 20:57:07

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/15/2023 12:33

Air Blank	0.000	20:56
Control Test	0.000	20:56
Air Blank	0.000	20:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 01/16/2023 14:42:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:37

Air Blank	0.000	14:40
Control Test	0.000	14:41
Air Blank	0.000	14:42

Agency: AUBURNDALE PD

Date/Time: 01/16/2023 14:44:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:37

Air Blank	0.000	14:43
Control Test	0.048	14:44
Air Blank	0.000	14:44

Agency: AUBURNDALE PD

Date/Time: 01/16/2023 14:47:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:37

Air Blank	0.000	14:45
Control Test	0.076	14:46
Air Blank	0.000	14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 01/16/2023 14:49:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:37

Air Blank	0.000	14:48
Control Test	0.197	14:49
Air Blank	0.000	14:49

Agency: AUBURNDALE PD

Date/Time: 02/07/2023 20:43:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2023 21:58

Air Blank	0.000	20:42
Control Test	0.000	20:43
Air Blank	0.000	20:43

Agency: AUBURNDALE PD

Date/Time: 02/07/2023 20:46:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2023 21:58

Air Blank	0.000	20:44
Control Test	0.077	20:45
Air Blank	0.000	20:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 02/07/2023 20:48:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2023 21:58

Air Blank	0.000	20:47
Control Test	0.047	20:47
Air Blank	0.000	20:48

Agency: AUBURNDALE PD

Date/Time: 02/07/2023 20:50:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2023 21:58

Air Blank	0.000	20:49
Control Test	0.196	20:50
Air Blank	0.000	20:50

Agency: AUBURNDALE PD

Date/Time: 03/22/2023 14:31:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/23/2023 12:38

Air Blank	0.000	14:30
Control Test	0.048	14:31
Air Blank	0.000	14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 03/22/2023 14:33:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/23/2023 12:38

Air Blank	0.000	14:32
Control Test	0.079	14:33
Air Blank	0.000	14:33

Agency: AUBURNDALE PD

Date/Time: 03/22/2023 14:36:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/23/2023 12:38

Air Blank	0.000	14:34
Control Test	0.196	14:35
Air Blank	0.000	14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency: WEST PALM BEACH PD

Date/Time: 06/23/2023 12:11:22

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/23/2023 12:53

Air Blank	0.000	12:10
Control Test	0.000	12:10
Air Blank	0.000	12:11

Agency: WEST PALM BEACH PD

Date/Time: 06/23/2023 12:13:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/23/2023 12:53

Air Blank	0.000	12:11
Control Test	0.000	12:12
Air Blank	0.000	12:13

Agency: WEST PALM BEACH PD

Date/Time: 09/09/2023 10:33:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/24/2023 10:37

Air Blank	0.000	10:32
Control Test	0.078	10:33
Air Blank	0.000	10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency: WEST PALM BEACH PD

Date/Time: 09/09/2023 10:36:25

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/24/2023 10:37

Air Blank	0.000	10:35
Control Test	0.079	10:35
Air Blank	0.000	10:36

Agency: WEST PALM BEACH PD

Date/Time: 09/09/2023 10:39:22

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 10/24/2023 10:37

Air Blank	0.000	10:38
ITP Check	0.079	10:38
Air Blank	0.000	10:39

Agency: WEST PALM BEACH PD

Date/Time: 09/13/2023 13:58:39

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 10/24/2023 10:37

Air Blank	0.000	13:57
Control Test	0.079	13:58
Air Blank	0.000	13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency: WEST PALM BEACH PD

Date/Time: 09/13/2023 14:01:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/24/2023 10:37

Air Blank	0.000	13:59
Control Test	0.000	14:00
Air Blank	0.000	14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency: DAYTONA BEACH
SHORES

Date/Time: 08/17/2023 00:31:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/17/2023 01:32

Air Blank	0.000	00:30
Control Test	0.048	00:30
Air Blank	0.000	00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

Agency: FORT LAUDERDALE PD

Date/Time: 12/26/2023 21:13:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/27/2023 20:39

Air Blank	0.000	21:12
Control Test	0.000	21:13
Air Blank	0.000	21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency: ROCKLEDGE P.D.

Date/Time: 04/11/2023 11:36:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/11/2023 14:01

Air Blank	0.000	11:36
Control Test	0.000	11:36
Air Blank	0.000	11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency: INDIAN HARBOUR PD

Date/Time: 06/19/2023 18:30:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/19/2023 19:52

Air Blank	0.000	18:28
Control Test	0.000	18:29
Air Blank	0.000	18:30

Agency: INDIAN HARBOUR PD

Date/Time: 06/19/2023 18:34:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/19/2023 19:52

Air Blank	0.000	18:33
Control Test	0.197	18:34
Air Blank	0.000	18:34

Agency: INDIAN HARBOUR PD

Date/Time: 06/19/2023 18:46:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/19/2023 19:52

Air Blank	0.000	18:44
Control Test	0.079	18:45
Air Blank	0.000	18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency: INDIAN HARBOUR PD

Date/Time: 06/19/2023 18:49:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/19/2023 19:52

Air Blank	0.000	18:48
Control Test	0.049	18:48
Air Blank	0.000	18:49

Agency: INDIAN HARBOUR PD

Date/Time: 07/31/2023 11:17:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/31/2023 13:43

Air Blank	0.000	11:16
Control Test	0.000	11:17
Air Blank	0.000	11:17

Agency: INDIAN HARBOUR PD

Date/Time: 08/29/2023 08:24:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/29/2023 09:29

Air Blank	0.000	08:23
Control Test	0.000	08:24
Air Blank	0.000	08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency: INDIAN HARBOUR PD

Date/Time: 11/01/2023 10:37:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/01/2023 11:43

Air Blank	0.000	10:36
Control Test	0.000	10:37
Air Blank	0.000	10:37

Agency: INDIAN HARBOUR PD

Date/Time: 11/24/2023 20:37:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/24/2023 21:35

Air Blank	0.000	20:35
Control Test	0.000	20:36
Air Blank	0.000	20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

Agency: PINELLAS COUNTY SO

Date/Time: 07/12/2023 20:22:05

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/16/2023 12:27

Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

Agency: MIAMI-DADE PD

Date/Time: 09/16/2023 03:16:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/05/2023 03:56

Air Blank	0.000	03:15
Control Test	0.051	03:16
Air Blank	0.000	03:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 03/09/2023 02:06:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 03:37

Air Blank	0.000	02:05
Control Test	0.000	02:06
Air Blank	0.000	02:06

Agency: MIAMI-DADE PD

Date/Time: 03/09/2023 02:08:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 03:37

Air Blank	0.000	02:07
Control Test	0.052	02:08
Air Blank	0.000	02:08

Agency: MIAMI-DADE PD

Date/Time: 03/09/2023 02:11:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 03:37

Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 03/09/2023 02:13:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/09/2023 03:37

Air Blank	0.000	02:12
Control Test	0.201	02:12
Air Blank	0.000	02:13

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 02:57:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	02:56
Control Test	0.000	02:56
Air Blank	0.000	02:57

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 02:59:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	02:58
Control Test	0.051	02:59
Air Blank	0.000	02:59

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 03:06:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	03:05
Control Test	0.200	03:05
Air Blank	0.000	03:06

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 03:08:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	03:07
Control Test	0.049	03:08
Air Blank	0.000	03:08

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 03:20:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	03:19
Control Test	0.082	03:20
Air Blank	0.000	03:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 03:39:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	03:38
Control Test	0.158	03:38
Air Blank	0.000	03:39

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 03:53:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	03:52
Control Test	0.203	03:53
Air Blank	0.000	03:53

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 05:10:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	05:09
Control Test	0.051	05:09
Air Blank	0.000	05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 05/12/2023 05:24:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:20

Air Blank	0.000	05:22
Control Test	0.043	05:23
Air Blank	0.000	05:24

Agency: MIAMI-DADE PD

Date/Time: 07/13/2023 23:44:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/14/2023 01:30

Air Blank	0.000	23:42
Control Test	0.000	23:43
Air Blank	0.000	23:44

Agency: MIAMI-DADE PD

Date/Time: 07/13/2023 23:46:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/14/2023 01:30

Air Blank	0.000	23:45
Control Test	0.000	23:45
Air Blank	0.000	23:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 07/13/2023 23:48:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/14/2023 01:30

Air Blank	0.000	23:47
Control Test	0.052	23:48
Air Blank	0.000	23:48

Agency: MIAMI-DADE PD

Date/Time: 07/13/2023 23:50:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/14/2023 01:30

Air Blank	0.000	23:49
Control Test	0.082	23:49
Air Blank	0.000	23:50

Agency: MIAMI-DADE PD

Date/Time: 07/13/2023 23:52:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/14/2023 01:30

Air Blank	0.000	23:51
Control Test	0.202	23:51
Air Blank	0.000	23:52

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 08/23/2023 23:39:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2023 02:03

Air Blank	0.000	23:38
Control Test	0.000	23:38
Air Blank	0.000	23:39

Agency: MIAMI-DADE PD

Date/Time: 08/23/2023 23:41:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2023 02:03

Air Blank	0.000	23:40
Control Test	0.007	23:40
Air Blank	0.000	23:41

Agency: MIAMI-DADE PD

Date/Time: 08/23/2023 23:43:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2023 02:03

Air Blank	0.000	23:42
Control Test	0.051	23:43
Air Blank	0.000	23:43

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 08/23/2023 23:45:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2023 02:03

Air Blank	0.000	23:44
Control Test	0.066	23:45
Air Blank	0.000	23:45

Agency: MIAMI-DADE PD

Date/Time: 08/23/2023 23:47:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2023 02:03

Air Blank	0.000	23:46
Control Test	0.062	23:47
Air Blank	0.000	23:47

Agency: MIAMI-DADE PD

Date/Time: 08/23/2023 23:50:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2023 02:03

Air Blank	0.000	23:49
Control Test	0.062	23:50
Air Blank	0.000	23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 08/24/2023 00:07:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2023 02:03

Air Blank	0.000	00:06
Control Test	0.081	00:07
Air Blank	0.000	00:07

Agency: MIAMI-DADE PD

Date/Time: 08/24/2023 00:10:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/24/2023 02:03

Air Blank	0.000	00:08
Control Test	0.202	00:09
Air Blank	0.000	00:10

Agency: MIAMI-DADE PD

Date/Time: 09/01/2023 01:22:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/01/2023 03:29

Air Blank	0.000	01:21
Control Test	0.000	01:21
Air Blank	0.000	01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 09/01/2023 01:24:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/01/2023 03:29

Air Blank	0.000	01:23
Control Test	0.007	01:23
Air Blank	0.000	01:24

Agency: MIAMI-DADE PD

Date/Time: 09/01/2023 01:26:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/01/2023 03:29

Air Blank	0.000	01:25
Control Test	0.051	01:25
Air Blank	0.000	01:26

Agency: MIAMI-DADE PD

Date/Time: 09/01/2023 01:28:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/01/2023 03:29

Air Blank	0.000	01:27
Control Test	0.082	01:27
Air Blank	0.000	01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 09/01/2023 01:30:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/01/2023 03:29

Air Blank	0.000	01:29
Control Test	0.202	01:29
Air Blank	0.000	01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

Agency: JACKSONVILLE SO

Date/Time: 01/11/2023 07:20:07

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/12/2023 15:31

Air Blank	0.000	07:19
Control Test	0.078	07:19
Air Blank	0.000	07:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

Agency: NASSAU COUNTY SO

Date/Time: 05/29/2023 19:02:03

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:11

Air Blank	0.000	19:01
Control Test	0.078	19:01
Air Blank	0.000	19:02

Agency: NASSAU COUNTY SO

Date/Time: 05/29/2023 20:21:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:11

Air Blank	0.000	20:20
Control Test	0.078	20:20
Air Blank		20:21

Agency: NASSAU COUNTY SO

Date/Time: 05/30/2023 13:36:13

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/27/2023 00:11

Air Blank	0.000	13:35
Control Test	0.078	13:35
Air Blank	0.000	13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

Agency: NASSAU COUNTY SO

Date/Time: 05/30/2023 13:47:05

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/27/2023 00:11

Air Blank	0.000	13:46
ITP Check	0.079	13:46
Air Blank	0.000	13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency: BRADFORD COUNTY SO

Date/Time: 04/28/2023 09:58:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/02/2023 16:55

Air Blank	0.000	09:57
Control Test	0.048	09:58
Air Blank	0.000	09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY
SO

Date/Time: 06/08/2023 22:38:14

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/05/2023 18:52

Air Blank		22:37
Air Blank	0.000	22:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 02/24/2023 23:44:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2023 15:05

Air Blank	0.000	23:43
Control Test	0.157	23:44
Air Blank	0.000	23:44

Agency: MARIANNA PD

Date/Time: 03/24/2023 15:58:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/24/2023 13:23

Air Blank	0.000	15:57
Control Test	0.078	15:58
Air Blank	0.000	15:58

Agency: MARIANNA PD

Date/Time: 03/24/2023 16:01:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/24/2023 13:23

Air Blank	0.000	15:59
Control Test	0.046	16:00
Air Blank	0.000	16:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 03/24/2023 16:03:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/24/2023 13:23

Air Blank	0.000	16:02
Control Test	0.192	16:02
Air Blank	0.000	16:03

Agency: MARIANNA PD

Date/Time: 04/30/2023 20:04:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 11:53

Air Blank	0.000	20:03
Control Test	0.048	20:04
Air Blank	0.000	20:04

Agency: MARIANNA PD

Date/Time: 04/30/2023 20:07:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 11:53

Air Blank	0.000	20:05
Control Test	0.077	20:06
Air Blank	0.000	20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 04/30/2023 20:09:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 11:53

Air Blank	0.000	20:08
Control Test	0.194	20:08
Air Blank	0.000	20:09

Agency: MARIANNA PD

Date/Time: 05/01/2023 01:51:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 11:53

Air Blank	0.000	01:50
Control Test	0.048	01:51
Air Blank	0.000	01:51

Agency: MARIANNA PD

Date/Time: 05/01/2023 01:53:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 11:53

Air Blank	0.000	01:52
Control Test	0.078	01:53
Air Blank	0.000	01:53

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 05/01/2023 01:55:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 11:53

Air Blank	0.000	01:54
Control Test	0.194	01:55
Air Blank	0.000	01:55

Agency: MARIANNA PD

Date/Time: 07/27/2023 11:03:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/27/2023 14:11

Air Blank	0.000	11:01
Control Test	0.000	11:02
Air Blank	0.000	11:03

Agency: MARIANNA PD

Date/Time: 08/28/2023 00:34:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/05/2023 09:39

Air Blank	0.000	00:33
Control Test	0.000	00:34
Air Blank	0.000	00:34

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 08/28/2023 00:37:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/05/2023 09:39

Air Blank	0.000	00:35
Control Test	0.000	00:36
Air Blank	0.000	00:37

Agency: MARIANNA PD

Date/Time: 08/28/2023 00:43:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/05/2023 09:39

Air Blank	0.000	00:42
Control Test	0.000	00:43
Air Blank	0.000	00:43

Agency: MARIANNA PD

Date/Time: 08/28/2023 00:53:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/05/2023 09:39

Air Blank	0.000	00:52
Control Test	0.000	00:52
Air Blank	0.000	00:53

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 08/28/2023 00:54:53

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/05/2023 09:39

Air Blank	0.000	00:54
Control Test	0.000	00:54
Air Blank	0.000	00:54

Agency: MARIANNA PD

Date/Time: 08/28/2023 01:26:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/05/2023 09:39

Air Blank	0.000	01:24
Control Test	0.000	01:25
Air Blank	0.000	01:26

Agency: MARIANNA PD

Date/Time: 08/31/2023 19:47:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/05/2023 09:39

Air Blank	0.000	19:46
Control Test	0.000	19:47
Air Blank	0.000	19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

Agency: MARIANNA PD

Date/Time: 08/31/2023 20:12:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/05/2023 09:39

Air Blank	0.000	20:11
Control Test	0.000	20:11
Air Blank	0.000	20:12

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

Agency: NICEVILLE PD

Date/Time: 06/08/2023 10:17:09

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/15/2023 13:52

Air Blank	0.000	10:16
Control Test	0.078	10:16
Air Blank	0.000	10:17

Agency: NICEVILLE PD

Date/Time: 06/08/2023 10:18:41

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/15/2023 13:52

Air Blank	0.000	10:17
Control Test	0.078	10:18
Air Blank	0.000	10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

Agency: SARASOTA COUNTY SO

Date/Time: 01/29/2023 13:57:05

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/01/2023 12:40

Air Blank	0.000	13:56
ITP Check	0.080	13:56
Air Blank	0.000	13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 07:59:00

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	07:58
Control Test	0.080	07:58
Air Blank	0.000	07:59

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 08:00:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	07:59
Control Test	0.000	08:00
Air Blank	0.000	08:00

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 08:03:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	08:02
Control Test	0.000	08:02
Air Blank	0.000	08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 08:07:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	08:06
Control Test	0.046	08:06
Air Blank	0.000	08:07

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 08:10:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	08:09
Control Test	0.076	08:10
Air Blank	0.000	08:10

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 08:12:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	08:11
Control Test	0.152	08:12
Air Blank	0.000	08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 08:16:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	08:15
Control Test	0.000	08:16
Air Blank	0.000	08:16

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 09:29:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	09:28
Control Test	0.150	09:29
Air Blank	0.000	09:29

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 09:32:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	09:30
Control Test	0.184	09:31
Air Blank	0.000	09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 09:40:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	09:39
Control Test	0.193	09:40
Air Blank	0.000	09:40

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 09:53:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	09:52
Control Test	0.000	09:53
Air Blank	0.000	09:53

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 09:56:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:48

Air Blank	0.000	09:55
Control Test	0.000	09:56
Air Blank	0.000	09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 10:00:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	09:59
Control Test	0.000	10:00
Air Blank	0.000	10:00

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 10:02:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	10:01
Control Test	0.000	10:02
Air Blank	0.000	10:02

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 10:07:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	10:06
Control Test	0.000	10:06
Air Blank	0.000	10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 10:09:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	10:08
Control Test	0.000	10:08
Air Blank	0.000	10:09

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 10:14:30

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	10:13
Control Test	0.079	10:14
Air Blank	0.000	10:14

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 10:16:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	10:15
Control Test	0.000	10:15
Air Blank	0.000	10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 12:57:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	12:55
Control Test	0.000	12:56
Air Blank	0.000	12:57

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 12:59:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	12:58
Control Test	0.192	12:58
Air Blank	0.000	12:59

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 13:01:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	13:00
Control Test	0.000	13:01
Air Blank	0.000	13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 13:04:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	13:03
Control Test	0.000	13:04
Air Blank	0.000	13:04

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 13:14:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	13:12
Control Test	0.000	13:13
Air Blank	0.000	13:14

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 13:35:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:49

Air Blank	0.000	13:33
Control Test	0.194	13:34
Air Blank	0.000	13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 13:38:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	13:37
Control Test	0.188	13:38
Air Blank	0.000	13:38

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 13:44:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	13:43
Control Test	0.192	13:44
Air Blank	0.000	13:44

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 13:46:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	13:45
Control Test	0.191	13:45
Air Blank	0.000	13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/19/2023 13:48:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	13:46
Control Test	0.192	13:47
Air Blank	0.000	13:48

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/30/2023 18:11:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	18:10
Control Test	0.174	18:10
Air Blank	0.000	18:11

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/30/2023 18:14:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	18:13
Control Test	0.044	18:14
Air Blank	0.000	18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/30/2023 18:23:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	18:21
Control Test	0.052	18:22
Air Blank	0.000	18:23

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/30/2023 18:25:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	18:24
Control Test	0.184	18:24
Air Blank	0.000	18:25

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/30/2023 18:26:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank		18:25
Air Blank		18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 04/30/2023 18:28:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 14:50

Air Blank	0.000	18:27
Control Test	0.196	18:27
Air Blank	0.000	18:28

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 07:08:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/23/2023 09:03

Air Blank	0.000	07:08
Control Test	0.080	07:08
Air Blank	0.000	07:08

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 07:10:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/23/2023 09:03

Air Blank	0.000	07:09
Control Test	0.000	07:10
Air Blank	0.000	07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 07:12:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/23/2023 09:03

Air Blank	0.000	07:11
Control Test	0.049	07:12
Air Blank	0.000	07:12

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 07:15:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/23/2023 09:03

Air Blank	0.000	07:13
Control Test	0.000	07:14
Air Blank	0.000	07:15

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 07:16:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/23/2023 09:03

Air Blank		07:15
Air Blank		07:16

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 07:18:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/23/2023 09:03

Air Blank	0.000	07:17
Control Test	0.075	07:17
Air Blank	0.000	07:18

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 07:21:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/23/2023 09:04

Air Blank	0.000	07:19
Control Test	0.191	07:20
Air Blank	0.000	07:21

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 08:11:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/23/2023 09:04

Air Blank	0.000	08:09
Control Test	0.199	08:10
Air Blank	0.000	08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 07/23/2023 08:18:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/23/2023 09:04

Air Blank	0.000	08:16
Control Test	0.199	08:17
Air Blank	0.000	08:18

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 19:00:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2023 19:57

Air Blank	0.000	18:59
Control Test	0.000	18:59
Air Blank	0.000	19:00

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 19:02:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2023 19:57

Air Blank	0.000	19:01
Control Test	0.049	19:02
Air Blank	0.000	19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 19:04:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2023 19:57

Air Blank	0.000	19:03
Control Test	0.049	19:03
Air Blank	0.000	19:04

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 19:06:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2023 19:57

Air Blank	0.000	19:05
Control Test	0.078	19:06
Air Blank	0.000	19:06

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 19:08:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2023 19:57

Air Blank	0.000	19:07
Control Test	0.079	19:07
Air Blank	0.000	19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 19:10:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2023 19:57

Air Blank	0.000	19:09
Control Test	0.197	19:10
Air Blank	0.000	19:10

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 19:12:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2023 19:57

Air Blank	0.000	19:11
Control Test	0.199	19:11
Air Blank	0.000	19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency: PINELLAS COUNTY SO

Date/Time: 03/02/2023 03:43:04

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/22/2023 14:20

Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 21:45:56

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/02/2023 14:47

Air Blank	0.000	21:45
Control Test	0.077	21:45
Air Blank	0.000	21:45

Agency: MARTIN COUNTY SO

Date/Time: 11/16/2023 19:11:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/16/2023 20:57

Air Blank	0.000	19:10
Control Test	0.200	19:11
Air Blank	0.000	19:11

Agency: MARTIN COUNTY SO

Date/Time: 11/16/2023 19:13:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/16/2023 20:57

Air Blank	0.000	19:12
Control Test	0.078	19:13
Air Blank	0.000	19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 11/16/2023 19:15:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/16/2023 20:57

Air Blank	0.000	19:14
Control Test	0.049	19:15
Air Blank	0.000	19:15

Agency: MARTIN COUNTY SO

Date/Time: 11/16/2023 19:17:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/16/2023 20:57

Air Blank	0.000	19:16
Control Test	0.000	19:17
Air Blank	0.000	19:17

Agency: MARTIN COUNTY SO

Date/Time: 11/16/2023 19:19:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/16/2023 20:57

Air Blank	0.000	19:18
Control Test	0.000	19:19
Air Blank	0.000	19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 12/14/2023 22:13:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/14/2023 23:54

Air Blank	0.000	22:12
Control Test	0.198	22:13
Air Blank	0.000	22:13

Agency: MARTIN COUNTY SO

Date/Time: 12/14/2023 22:15:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/14/2023 23:54

Air Blank	0.000	22:14
Control Test	0.078	22:15
Air Blank	0.000	22:15

Agency: MARTIN COUNTY SO

Date/Time: 12/14/2023 22:17:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/14/2023 23:54

Air Blank	0.000	22:16
Control Test	0.049	22:17
Air Blank	0.000	22:17

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 12/14/2023 22:19:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/14/2023 23:54

Air Blank	0.000	22:18
Control Test	0.000	22:19
Air Blank	0.000	22:19

Agency: MARTIN COUNTY SO

Date/Time: 12/14/2023 22:21:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/14/2023 23:54

Air Blank	0.000	22:20
Control Test	0.000	22:21
Air Blank	0.000	22:21

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency: ST PETERSBURG PD

Date/Time: 07/14/2023 13:44:04

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/28/2023 12:13

Air Blank	0.000	13:43
Control Test	0.081	13:43
Air Blank	0.000	13:44

Agency: ST PETERSBURG PD

Date/Time: 07/14/2023 13:45:32

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 07/28/2023 12:13

Air Blank	0.000	13:44
ITP Check	0.081	13:45
Air Blank	0.000	13:45

Agency: ST PETERSBURG PD

Date/Time: 07/14/2023 13:46:51

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/28/2023 12:13

Air Blank	0.000	13:46
Control Test	0.081	13:46
Air Blank	0.000	13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 04/20/2023 17:52:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/20/2023 20:41

Air Blank	0.000	17:51
Control Test	0.000	17:52
Air Blank	0.000	17:52

Agency: PINELLAS PARK P.D.

Date/Time: 04/20/2023 17:55:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/20/2023 20:41

Air Blank	0.000	17:54
Control Test	0.000	17:54
Air Blank	0.000	17:55

Agency: PINELLAS PARK P.D.

Date/Time: 04/20/2023 17:58:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/20/2023 20:41

Air Blank	0.000	17:57
Control Test	0.046	17:58
Air Blank	0.000	17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 04/20/2023 18:00:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/20/2023 20:42

Air Blank	0.000	17:59
Control Test	0.076	18:00
Air Blank	0.000	18:00

Agency: PINELLAS PARK P.D.

Date/Time: 04/20/2023 18:02:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/20/2023 20:42

Air Blank	0.000	18:01
Control Test	0.190	18:02
Air Blank	0.000	18:02

Agency: PINELLAS PARK P.D.

Date/Time: 05/12/2023 16:38:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/12/2023 18:57

Air Blank	0.000	16:37
Control Test	0.000	16:38
Air Blank	0.000	16:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 05/12/2023 16:41:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/12/2023 18:57

Air Blank	0.000	16:40
Control Test	0.000	16:40
Air Blank	0.000	16:41

Agency: PINELLAS PARK P.D.

Date/Time: 05/12/2023 16:44:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/12/2023 18:57

Air Blank	0.000	16:42
Control Test	0.045	16:43
Air Blank	0.000	16:44

Agency: PINELLAS PARK P.D.

Date/Time: 05/12/2023 16:47:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/12/2023 18:57

Air Blank	0.000	16:46
Control Test	0.189	16:47
Air Blank	0.000	16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 05/12/2023 17:08:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/12/2023 18:57

Air Blank	0.000	17:07
Control Test	0.196	17:07
Air Blank	0.000	17:08

Agency: PINELLAS PARK P.D.

Date/Time: 05/12/2023 17:10:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/12/2023 18:57

Air Blank	0.000	17:09
Control Test	0.080	17:09
Air Blank	0.000	17:10

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 16:51:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:18

Air Blank	0.000	16:50
Control Test	0.000	16:51
Air Blank	0.000	16:51

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 16:53:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:18

Air Blank	0.000	16:52
Control Test	0.000	16:53
Air Blank	0.000	16:53

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 16:56:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:18

Air Blank	0.000	16:55
Control Test	0.055	16:55
Air Blank	0.000	16:56

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 16:58:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:18

Air Blank	0.000	16:57
Control Test	0.075	16:58
Air Blank	0.000	16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 17:01:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:19

Air Blank	0.000	17:00
Control Test	0.000	17:00
Air Blank	0.000	17:01

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 17:03:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:19

Air Blank	0.000	17:02
Control Test	0.000	17:03
Air Blank	0.000	17:03

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 17:05:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:19

Air Blank	0.000	17:04
Control Test	0.051	17:05
Air Blank	0.000	17:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 17:09:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:19

Air Blank	0.000	17:08
Control Test	0.189	17:09
Air Blank	0.000	17:09

Agency: PINELLAS PARK P.D.

Date/Time: 06/08/2023 17:21:45

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/12/2023 11:19

Air Blank	0.000	17:20
Control Test	0.194	17:21
Air Blank	0.000	17:21

Agency: PINELLAS PARK P.D.

Date/Time: 07/21/2023 18:32:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2023 19:46

Air Blank	0.000	18:31
Control Test	0.000	18:32
Air Blank	0.000	18:32

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 07/21/2023 18:34:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2023 19:46

Air Blank	0.000	18:33
Control Test	0.000	18:34
Air Blank	0.000	18:34

Agency: PINELLAS PARK P.D.

Date/Time: 07/21/2023 18:38:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2023 19:46

Air Blank	0.000	18:36
Control Test	0.046	18:37
Air Blank	0.000	18:38

Agency: PINELLAS PARK P.D.

Date/Time: 07/21/2023 18:40:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2023 19:46

Air Blank	0.000	18:38
Control Test	0.075	18:39
Air Blank	0.000	18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 07/21/2023 18:42:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/21/2023 19:46

Air Blank	0.000	18:41
Control Test	0.193	18:41
Air Blank	0.000	18:42

Agency: PINELLAS PARK P.D.

Date/Time: 08/22/2023 12:38:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/22/2023 13:45

Air Blank	0.000	12:37
Control Test	0.000	12:38
Air Blank	0.000	12:38

Agency: PINELLAS PARK P.D.

Date/Time: 08/22/2023 12:41:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/22/2023 13:45

Air Blank	0.000	12:39
Control Test	0.000	12:40
Air Blank	0.000	12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 08/22/2023 12:43:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/22/2023 13:45

Air Blank	0.000	12:42
Control Test	0.046	12:42
Air Blank	0.000	12:43

Agency: PINELLAS PARK P.D.

Date/Time: 08/22/2023 12:45:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/22/2023 13:45

Air Blank	0.000	12:44
Control Test	0.075	12:44
Air Blank	0.000	12:45

Agency: PINELLAS PARK P.D.

Date/Time: 08/22/2023 12:47:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/22/2023 13:45

Air Blank	0.000	12:46
Control Test	0.194	12:47
Air Blank	0.000	12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 08/22/2023 12:49:43

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/22/2023 13:45

Air Blank	0.000	12:48
Control Test	0.079	12:49
Air Blank	0.000	12:49

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 06:51:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	06:50
Control Test	0.000	06:51
Air Blank	0.000	06:51

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 06:53:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	06:52
Control Test	0.000	06:53
Air Blank	0.000	06:53

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 06:56:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	06:54
Control Test	0.045	06:55
Air Blank	0.000	06:56

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 06:58:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	06:57
Control Test	0.072	06:58
Air Blank	0.000	06:58

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 07:01:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	06:59
Control Test	0.187	07:00
Air Blank	0.000	07:01

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 07:18:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	07:17
Control Test	0.103	07:18
Air Blank	0.000	07:18

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 07:27:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	07:25
Control Test	0.187	07:26
Air Blank	0.000	07:27

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 07:29:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	07:28
Control Test	0.082	07:28
Air Blank	0.000	07:29

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 09/25/2023 07:31:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 08:18

Air Blank	0.000	07:30
Control Test	0.196	07:30
Air Blank	0.000	07:31

Agency: PINELLAS PARK P.D.

Date/Time: 10/23/2023 11:51:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/23/2023 12:49

Air Blank	0.000	11:50
Control Test	0.000	11:50
Air Blank	0.000	11:51

Agency: PINELLAS PARK P.D.

Date/Time: 10/23/2023 11:53:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/23/2023 12:49

Air Blank	0.000	11:52
Control Test	0.000	11:52
Air Blank	0.000	11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 10/23/2023 11:55:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/23/2023 12:49

Air Blank	0.000	11:54
Control Test	0.047	11:54
Air Blank	0.000	11:55

Agency: PINELLAS PARK P.D.

Date/Time: 10/23/2023 11:57:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/23/2023 12:49

Air Blank	0.000	11:56
Control Test	0.076	11:56
Air Blank	0.000	11:57

Agency: PINELLAS PARK P.D.

Date/Time: 10/23/2023 11:59:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/23/2023 12:49

Air Blank	0.000	11:58
Control Test	0.195	11:58
Air Blank	0.000	11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 11/14/2023 07:24:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 09:11

Air Blank	0.000	07:23
Control Test	0.000	07:24
Air Blank	0.000	07:24

Agency: PINELLAS PARK P.D.

Date/Time: 11/14/2023 07:26:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 09:11

Air Blank	0.000	07:25
Control Test	0.000	07:26
Air Blank	0.000	07:26

Agency: PINELLAS PARK P.D.

Date/Time: 11/14/2023 07:29:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 09:11

Air Blank	0.000	07:27
Control Test	0.047	07:28
Air Blank	0.000	07:29

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 11/14/2023 07:31:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 09:11

Air Blank	0.000	07:30
Control Test	0.075	07:30
Air Blank	0.000	07:31

Agency: PINELLAS PARK P.D.

Date/Time: 11/14/2023 07:39:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 09:11

Air Blank	0.000	07:38
Control Test	0.194	07:39
Air Blank	0.000	07:39

Agency: PINELLAS PARK P.D.

Date/Time: 11/14/2023 07:42:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/14/2023 09:11

Air Blank	0.000	07:41
Control Test	0.195	07:42
Air Blank	0.000	07:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 12/04/2023 10:33:49

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 11:30

Air Blank	0.000	10:32
Control Test	0.000	10:33
Air Blank	0.000	10:33

Agency: PINELLAS PARK P.D.

Date/Time: 12/04/2023 10:36:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 11:30

Air Blank	0.000	10:35
Control Test	0.000	10:35
Air Blank	0.000	10:36

Agency: PINELLAS PARK P.D.

Date/Time: 12/04/2023 10:38:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 11:30

Air Blank	0.000	10:37
Control Test	0.047	10:38
Air Blank	0.000	10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 12/04/2023 10:41:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 11:30

Air Blank	0.000	10:40
Control Test	0.075	10:41
Air Blank	0.000	10:41

Agency: PINELLAS PARK P.D.

Date/Time: 12/04/2023 10:43:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/04/2023 11:30

Air Blank	0.000	10:42
Control Test	0.193	10:43
Air Blank	0.000	10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 21:25:31

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/24/2023 22:16

Air Blank	0.000	21:24
Control Test	0.079	21:25
Air Blank	0.000	21:25

Agency: CHARLOTTE COUNTY
SO

Date/Time: 09/24/2023 21:29:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/24/2023 22:16

Air Blank	0.000	21:28
Control Test	0.000	21:29
Air Blank	0.000	21:29

Agency: CHARLOTTE COUNTY
SO

Date/Time: 11/19/2023 23:07:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/20/2023 00:03

Air Blank	0.000	23:06
Control Test	0.199	23:07
Air Blank	0.000	23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY
SO

Date/Time: 11/19/2023 23:10:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/20/2023 00:03

Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:10

Agency: CHARLOTTE COUNTY
SO

Date/Time: 11/19/2023 23:12:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/20/2023 00:03

Air Blank	0.000	23:11
Control Test	0.049	23:12
Air Blank	0.000	23:12

Agency: CHARLOTTE COUNTY
SO

Date/Time: 11/19/2023 23:15:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/20/2023 00:03

Air Blank	0.000	23:14
Control Test	0.000	23:14
Air Blank	0.000	23:15

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY
SO

Date/Time: 11/19/2023 23:18:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/20/2023 00:03

Air Blank	0.000	23:17
Control Test	0.000	23:18
Air Blank	0.000	23:18

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency: JACKSONVILLE SO

Date/Time: 09/15/2023 09:18:51

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/20/2023 12:10

Air Blank	0.000	09:18
Control Test	0.079	09:18
Air Blank	0.000	09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 04/19/2023 13:11:33

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/20/2023 09:47

Air Blank	0.000	13:10
Control Test	0.080	13:11
Air Blank	0.000	13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency: TALLAHASSEE PD

Date/Time: 11/07/2023 07:44:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/07/2023 08:51

Air Blank	0.000	07:42
Control Test	0.048	07:43
Air Blank	0.000	07:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 02/07/2023 17:58:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2023 19:02

Air Blank	0.000	17:57
Control Test	0.073	17:57
Air Blank	0.000	17:58

Agency: PASCO COUNTY SO

Date/Time: 09/16/2023 10:03:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/16/2023 11:01

Air Blank	0.000	10:02
Control Test	0.075	10:03
Air Blank	0.000	10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency: FL HIGHWAY PATROL

Date/Time: 06/30/2023 18:19:46

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/01/2023 17:25

Air Blank	0.000	18:18
Control Test	0.080	18:19
Air Blank	0.000	18:19

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency: FL HIGHWAY PATROL

Date/Time: 02/21/2023 08:10:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/21/2023 11:31

Air Blank	0.000	08:09
Control Test	0.197	08:10
Air Blank	0.000	08:10

Agency: FL HIGHWAY PATROL

Date/Time: 04/21/2023 12:33:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/21/2023 16:49

Air Blank	0.000	12:32
Control Test	0.000	12:32
Air Blank	0.000	12:33

Agency: FL HIGHWAY PATROL

Date/Time: 04/21/2023 12:36:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/21/2023 16:49

Air Blank	0.000	12:35
Control Test	0.000	12:36
Air Blank	0.000	12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency: FL HIGHWAY PATROL

Date/Time: 04/21/2023 12:42:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/21/2023 16:49

Air Blank	0.000	12:41
Control Test	0.000	12:42
Air Blank	0.000	12:42

Agency: FL HIGHWAY PATROL

Date/Time: 08/28/2023 08:55:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 08/28/2023 14:02

Air Blank	0.000	08:54
Control Test	0.000	08:55
Air Blank	0.000	08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency: FORT WALTON BEACH
PD

Date/Time: 09/29/2023 08:47:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/29/2023 10:29

Air Blank	0.000	08:46
Control Test	0.046	08:46
Air Blank	0.000	08:47

Agency: FORT WALTON BEACH
PD

Date/Time: 10/25/2023 08:46:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/25/2023 10:35

Air Blank	0.000	08:45
Control Test	0.047	08:46
Air Blank	0.000	08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency: HERNANDO COUNTY SO

Date/Time: 09/29/2023 18:14:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/29/2023 19:29

Air Blank	0.000	18:13
Control Test	0.000	18:14
Air Blank	0.000	18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency: ST PETERSBURG PD

Date/Time: 12/28/2023 16:05:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2024 08:21

Air Blank	0.000	16:04
Control Test	0.199	16:04
Air Blank	0.000	16:05

Agency: ST PETERSBURG PD

Date/Time: 12/28/2023 16:07:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2024 08:21

Air Blank	0.000	16:06
Control Test	0.000	16:06
Air Blank	0.000	16:07

Agency: ST PETERSBURG PD

Date/Time: 12/28/2023 16:09:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/03/2024 08:21

Air Blank	0.000	16:08
Control Test	0.077	16:08
Air Blank	0.000	16:09

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency: PORT ORANGE PD

Date/Time: 01/16/2023 13:37:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/09/2023 10:08

Air Blank	0.000	13:37
Control Test	0.083	13:37
Air Blank	0.000	13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency: FERNANDINA BEACH PD

Date/Time: 03/13/2023 07:04:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/13/2023 09:01

Air Blank	0.000	07:02
Control Test	0.078	07:03
Air Blank	0.000	07:04

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency: MIAMI DADE PD

Date/Time: 06/09/2023 02:42:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/09/2023 07:47

Air Blank	0.000	02:41
Control Test	0.000	02:41
Air Blank	0.000	02:42

Agency: MIAMI DADE PD

Date/Time: 06/09/2023 02:44:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/09/2023 07:47

Air Blank	0.000	02:43
Control Test	0.000	02:43
Air Blank	0.000	02:44

Agency: MIAMI DADE PD

Date/Time: 06/09/2023 02:46:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/09/2023 07:47

Air Blank	0.000	02:45
Control Test	0.050	02:46
Air Blank	0.000	02:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency: MIAMI DADE PD

Date/Time: 06/09/2023 02:50:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/09/2023 07:47

Air Blank	0.000	02:49
Control Test	0.078	02:50
Air Blank	0.000	02:50

Agency: MIAMI DADE PD

Date/Time: 06/09/2023 02:53:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/09/2023 07:47

Air Blank	0.000	02:52
Control Test	0.196	02:52
Air Blank	0.000	02:53

Agency: MIAMI DADE PD

Date/Time: 07/05/2023 23:03:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/06/2023 02:34

Air Blank	0.000	23:02
Control Test	0.000	23:02
Air Blank	0.000	23:03

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency: MIAMI DADE PD

Date/Time: 07/05/2023 23:05:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/06/2023 02:34

Air Blank	0.000	23:04
Control Test	0.000	23:05
Air Blank	0.000	23:05

Agency: MIAMI DADE PD

Date/Time: 07/06/2023 00:08:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/06/2023 02:34

Air Blank	0.000	00:07
Control Test	0.000	00:08
Air Blank	0.000	00:08

Agency: MIAMI DADE PD

Date/Time: 07/06/2023 00:11:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/06/2023 02:34

Air Blank	0.000	00:09
Control Test	0.079	00:10
Air Blank	0.000	00:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency: MIAMI DADE PD

Date/Time: 07/06/2023 00:13:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/06/2023 02:34

Air Blank	0.000	00:11
Control Test	0.198	00:12
Air Blank	0.000	00:13

Agency: MIAMI DADE PD

Date/Time: 07/06/2023 00:15:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/06/2023 02:34

Air Blank	0.000	00:14
Control Test	0.000	00:14
Air Blank	0.000	00:15

Agency: MIAMI DADE PD

Date/Time: 07/06/2023 00:35:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/06/2023 02:34

Air Blank	0.000	00:33
Control Test	0.050	00:34
Air Blank	0.000	00:35

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency: MIAMI DADE PD

Date/Time: 12/01/2023 02:25:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/01/2023 04:07

Air Blank	0.000	02:24
Control Test	0.000	02:25
Air Blank	0.000	02:25

Agency: MIAMI DADE PD

Date/Time: 12/01/2023 02:27:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/01/2023 04:07

Air Blank	0.000	02:26
Control Test	0.050	02:27
Air Blank	0.000	02:27

Agency: MIAMI DADE PD

Date/Time: 12/01/2023 02:29:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/01/2023 04:07

Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency: MIAMI DADE PD

Date/Time: 12/01/2023 02:31:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 12/01/2023 04:07

Air Blank	0.000	02:30
Control Test	0.199	02:31
Air Blank	0.000	02:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency: MIAMI DADE PD

Date/Time: 05/27/2023 02:34:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	02:33
Control Test	0.000	02:34
Air Blank	0.000	02:34

Agency: MIAMI DADE PD

Date/Time: 05/27/2023 02:39:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	02:37
Control Test	0.000	02:38
Air Blank	0.000	02:39

Agency: MIAMI DADE PD

Date/Time: 05/27/2023 02:51:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	02:50
Control Test	0.000	02:51
Air Blank	0.000	02:51

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency: MIAMI DADE PD

Date/Time: 05/27/2023 03:07:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	03:06
Control Test	0.000	03:06
Air Blank	0.000	03:07

Agency: MIAMI DADE PD

Date/Time: 05/27/2023 03:20:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	03:18
Control Test	0.050	03:19
Air Blank	0.000	03:20

Agency: MIAMI DADE PD

Date/Time: 05/27/2023 03:23:55

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency: MIAMI DADE PD

Date/Time: 06/02/2023 01:31:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	01:29
Control Test	0.000	01:30
Air Blank	0.000	01:31

Agency: MIAMI DADE PD

Date/Time: 06/02/2023 01:55:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	01:54
Control Test	0.000	01:54
Air Blank	0.000	01:55

Agency: MIAMI DADE PD

Date/Time: 06/02/2023 01:57:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	01:56
Control Test	0.050	01:56
Air Blank	0.000	01:57

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency: MIAMI DADE PD

Date/Time: 06/02/2023 02:00:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00

Agency: MIAMI DADE PD

Date/Time: 06/02/2023 02:03:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	02:02
Control Test	0.199	02:03
Air Blank	0.000	02:03

Agency: MIAMI DADE PD

Date/Time: 06/02/2023 02:05:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/02/2023 08:23

Air Blank	0.000	02:04
Control Test	0.000	02:05
Air Blank	0.000	02:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency: MIAMI DADE PD

Date/Time: 12/16/2023 23:06:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:26

Air Blank	0.000	23:04
Control Test	0.048	23:05
Air Blank	0.000	23:06

Agency: MIAMI DADE PD

Date/Time: 12/16/2023 23:08:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:26

Air Blank	0.000	23:07
Control Test	0.046	23:08
Air Blank	0.000	23:08

Agency: MIAMI DADE PD

Date/Time: 12/16/2023 23:25:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:26

Air Blank	0.000	23:24
Control Test	0.049	23:24
Air Blank	0.000	23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency: MIAMI DADE PD

Date/Time: 12/16/2023 23:27:40

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:26

Air Blank	0.000	23:26
Control Test	0.048	23:27
Air Blank	0.000	23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency: MIAMI DADE PD

Date/Time: 03/30/2023 22:46:34

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/06/2023 07:38

Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46

Agency: MIAMI DADE PD

Date/Time: 04/06/2023 07:49:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/16/2023 10:39

Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:49

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

Agency: MARTIN COUNTY SO

Date/Time: 05/04/2023 21:38:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2023 23:17

Air Blank	0.000	21:36
Control Test	0.198	21:37
Air Blank	0.000	21:38

Agency: MARTIN COUNTY SO

Date/Time: 05/04/2023 21:40:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2023 23:17

Air Blank	0.000	21:38
Control Test	0.078	21:39
Air Blank	0.000	21:40

Agency: MARTIN COUNTY SO

Date/Time: 05/04/2023 21:42:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2023 23:17

Air Blank	0.000	21:40
Control Test	0.049	21:41
Air Blank	0.000	21:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

Agency: MARTIN COUNTY SO

Date/Time: 05/04/2023 21:44:21

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2023 23:17

Air Blank	0.000	21:43
Control Test	0.000	21:43
Air Blank	0.000	21:44

Agency: MARTIN COUNTY SO

Date/Time: 05/04/2023 21:46:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/04/2023 23:17

Air Blank	0.000	21:45
Control Test	0.000	21:45
Air Blank	0.000	21:46

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:44:29

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/02/2023 14:50

Air Blank	0.000	16:43
Control Test	0.080	16:43
Air Blank	0.000	16:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

Agency: MARTIN COUNTY SO

Date/Time: 03/09/2023 13:37:16

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 03/09/2023 14:49

Air Blank	0.000	13:36
Control Test	0.079	13:36
Air Blank	0.000	13:37

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 16:37:04

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/02/2023 14:44

Air Blank	0.000	16:36
Control Test	0.079	16:36
Air Blank	0.000	16:37

Agency: MARTIN COUNTY SO

Date/Time: 07/26/2023 19:19:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2023 22:22

Air Blank	0.000	19:17
Control Test	0.198	19:18
Air Blank	0.000	19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

Agency: MARTIN COUNTY SO

Date/Time: 07/26/2023 19:21:23

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2023 22:22

Air Blank	0.000	19:20
Control Test	0.079	19:20
Air Blank	0.000	19:21

Agency: MARTIN COUNTY SO

Date/Time: 07/26/2023 19:23:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2023 22:22

Air Blank	0.000	19:22
Control Test	0.050	19:23
Air Blank	0.000	19:23

Agency: MARTIN COUNTY SO

Date/Time: 07/26/2023 19:25:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2023 22:22

Air Blank	0.000	19:24
Control Test	0.000	19:24
Air Blank	0.000	19:25

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

Agency: MARTIN COUNTY SO

Date/Time: 07/26/2023 19:27:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/26/2023 22:22

Air Blank	0.000	19:26
Control Test	0.000	19:27
Air Blank	0.000	19:27

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 01/16/2023 13:51:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:20

Air Blank	0.000	13:50
Control Test	0.000	13:51
Air Blank	0.000	13:51

Agency: MARTIN COUNTY SO

Date/Time: 01/16/2023 13:53:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:20

Air Blank	0.000	13:52
Control Test	0.000	13:53
Air Blank	0.000	13:53

Agency: MARTIN COUNTY SO

Date/Time: 01/16/2023 13:56:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:20

Air Blank	0.000	13:54
Control Test	0.049	13:55
Air Blank	0.000	13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency: MARTIN COUNTY SO

Date/Time: 01/16/2023 13:59:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:20

Air Blank	0.000	13:57
Control Test	0.080	13:58
Air Blank	0.000	13:59

Agency: MARTIN COUNTY SO

Date/Time: 01/16/2023 14:01:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/16/2023 15:20

Air Blank	0.000	14:00
Control Test	0.201	14:01
Air Blank	0.000	14:01

Agency: MARTIN COUNTY SO

Date/Time: 06/01/2023 15:11:48

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/01/2023 18:18

Air Blank	0.000	15:10
Control Test	0.081	15:11
Air Blank	0.000	15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

Agency: BREVARD COUNTY S.O.

Date/Time: 11/15/2023 14:35:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/30/2023 08:38

Air Blank	0.000	14:34
Control Test	0.081	14:34
Air Blank	0.000	14:35

Agency: BREVARD COUNTY S.O.

Date/Time: 11/15/2023 14:37:02

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/30/2023 08:38

Air Blank	0.000	14:36
Control Test	0.079	14:36
Air Blank	0.000	14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

Agency: HENDRY COUNTY SO

Date/Time: 08/17/2023 17:27:22

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/17/2023 17:56

Air Blank		17:26
Air Blank	0.000	17:27

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

Agency: MANATEE COUNTY SO

Date/Time: 08/03/2023 16:37:13

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/14/2023 10:03

Air Blank	0.000	16:36
Control Test	0.079	16:36
Air Blank	0.000	16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 16:32:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:58

Air Blank	0.000	16:31
Control Test	0.047	16:31
Air Blank	0.000	16:32

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:04:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:58

Air Blank	0.000	17:03
Control Test	0.167	17:03
Air Blank	0.000	17:04

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:06:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:58

Air Blank	0.000	17:04
Control Test	0.172	17:05
Air Blank	0.000	17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:08:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:58

Air Blank	0.000	17:07
Control Test	0.063	17:07
Air Blank	0.000	17:08

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:10:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:59

Air Blank	0.000	17:08
Control Test	0.064	17:09
Air Blank	0.000	17:10

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:11:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:59

Air Blank		17:10
Air Blank	0.000	17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:13:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:59

Air Blank	0.000	17:11
Control Test	0.077	17:12
Air Blank	0.000	17:13

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:21:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:59

Air Blank	0.000	17:20
Control Test	0.000	17:20
Air Blank	0.000	17:21

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:24:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:59

Air Blank	0.000	17:23
Control Test	0.000	17:23
Air Blank	0.000	17:24

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:28:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:59

Air Blank		17:28
Air Blank	0.000	17:28

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:31:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2023 18:59

Air Blank	0.000	17:29
Control Test	0.048	17:30
Air Blank	0.000	17:31

Agency: FL HIGHWAY PATROL

Date/Time: 01/25/2023 17:50:27

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/25/2023 18:59

Air Blank	0.000	17:49
Control Test	0.080	17:49
Air Blank	0.000	17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

Agency: FL HIGHWAY PATROL

Date/Time: 03/22/2023 18:03:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/28/2023 12:41

Air Blank	0.000	18:02
Control Test	0.194	18:02
Air Blank	0.000	18:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

Agency: MIAMI PD

Date/Time: 11/06/2023 15:27:13

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 11/27/2023 09:36

Air Blank	0.000	15:26
ITP Check	0.078	15:26
Air Blank	0.000	15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

Agency: MIAMI PD

Date/Time: 11/06/2023 15:29:42

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/27/2023 09:18

Air Blank	0.000	15:28
Control Test	0.079	15:29
Air Blank	0.000	15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

Agency: MONROE COUNTY S.O.

Date/Time: 07/31/2023 10:31:45

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/02/2023 14:31

Air Blank	0.000	10:30
Control Test	0.080	10:31
Air Blank	0.000	10:31

Agency: MONROE COUNTY S.O.

Date/Time: 10/30/2023 11:38:30

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/09/2023 08:51

Air Blank	0.000	11:37
Control Test	0.080	11:37
Air Blank	0.000	11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

Agency: MANATEE COUNTY SO

Date/Time: 07/20/2023 15:26:10

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/20/2023 16:17

Air Blank	0.000	15:25
Control Test	0.086	15:25
Air Blank	0.000	15:26

Agency: MANATEE COUNTY SO

Date/Time: 07/20/2023 15:29:56

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/20/2023 16:17

Air Blank	0.000	15:29
Control Test	0.117	15:29
Air Blank	0.000	15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

Agency: FLAGLER COUNTY SO

Date/Time: 10/30/2023 14:33:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/30/2023 15:47

Air Blank	0.000	14:32
Control Test	0.048	14:33
Air Blank	0.000	14:33

Agency: FLAGLER COUNTY SO

Date/Time: 10/30/2023 14:40:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/30/2023 15:47

Air Blank	0.000	14:38
Control Test	0.079	14:39
Air Blank	0.000	14:40

Agency: FLAGLER COUNTY SO

Date/Time: 10/30/2023 14:42:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/30/2023 15:47

Air Blank	0.000	14:41
Control Test	0.199	14:41
Air Blank	0.000	14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

Agency: MANATEE COUNTY SO

Date/Time: 07/01/2023 14:07:52

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/20/2023 12:37

Air Blank	0.000	14:07
Control Test	0.079	14:07
Air Blank	0.000	14:07

Agency: MANATEE COUNTY SO

Date/Time: 07/01/2023 14:10:21

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/20/2023 12:37

Air Blank	0.000	14:09
Control Test	0.079	14:09
Air Blank	0.000	14:10

Agency: MANATEE COUNTY SO

Date/Time: 07/01/2023 14:17:22

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/20/2023 12:37

Air Blank	0.000	14:16
Control Test	0.079	14:16
Air Blank	0.000	14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

Agency: FL HIGHWAY PATROL

Date/Time: 04/28/2023 16:38:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/03/2023 00:47

Air Blank	0.000	16:37
Control Test	0.000	16:38
Air Blank	0.000	16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

Agency: MANATEE COUNTY SO

Date/Time: 11/16/2023 15:45:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 12/04/2023 13:43

Air Blank	0.000	15:44
Control Test	0.078	15:44
Air Blank	0.000	15:45

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

Agency: MONROE COUNTY SO

Date/Time: 06/07/2023 10:13:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2023 09:53

Air Blank	0.000	10:12
Control Test	0.047	10:13
Air Blank	0.000	10:13

Agency: MONROE COUNTY SO

Date/Time: 06/07/2023 10:16:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2023 09:53

Air Blank	0.000	10:14
Control Test	0.077	10:15
Air Blank	0.000	10:16

Agency: MONROE COUNTY SO

Date/Time: 06/07/2023 10:18:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/08/2023 09:53

Air Blank	0.000	10:17
Control Test	0.194	10:17
Air Blank	0.000	10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

Agency: FHP

Date/Time: 05/16/2023 09:22:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/16/2023 14:29

Air Blank	0.000	09:20
Control Test	0.000	09:21
Air Blank	0.000	09:22

Agency: FHP

Date/Time: 05/16/2023 09:49:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/16/2023 14:29

Air Blank	0.000	09:48
Control Test	0.187	09:49
Air Blank	0.000	09:49

Agency: FHP

Date/Time: 05/16/2023 09:56:47

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/16/2023 14:29

Air Blank	0.000	09:55
Control Test	0.187	09:56
Air Blank	0.000	09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

Agency: FHP

Date/Time: 05/25/2023 09:18:50

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/14/2023 13:33

Air Blank	0.000	09:18
Control Test	0.076	09:18
Air Blank	0.000	09:18

Agency: FHP

Date/Time: 05/25/2023 09:30:11

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 06/14/2023 13:33

Air Blank	0.000	09:29
ITP Check	0.079	09:29
Air Blank	0.000	09:30

Agency: FHP

Date/Time: 07/19/2023 09:34:18

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 07/19/2023 15:33

Air Blank	0.000	09:33
Control Test	0.079	09:33
Air Blank	0.000	09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

Agency: FHP

Date/Time: 10/24/2023 08:09:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 10/24/2023 10:16

Air Blank	0.000	08:08
Control Test	0.000	08:08
Air Blank	0.000	08:09

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

Agency: TALLAHASSEE PD

Date/Time: 03/29/2023 09:42:08

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/25/2023 09:46

Air Blank	0.000	09:41
Control Test	0.084	09:41
Air Blank	0.000	09:42

Agency: TALLAHASSEE PD

Date/Time: 03/29/2023 09:44:04

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 04/25/2023 09:46

Air Blank	0.000	09:43
ITP Check	0.080	09:43
Air Blank	0.000	09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

Agency: BROWARD COUNTY SO

Date/Time: 06/07/2023 10:05:18

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 06/07/2023 11:50

Air Blank	0.000	10:04
Control Test	0.080	10:04
Air Blank	0.000	10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

Agency: FORT LAUDERDALE PD

Date/Time: 03/27/2023 20:32:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/10/2023 15:04

Air Blank	0.000	20:31
Control Test	0.048	20:31
Air Blank	0.000	20:32

Agency: FORT LAUDERDALE PD

Date/Time: 03/27/2023 20:34:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/10/2023 15:04

Air Blank		20:33
Air Blank	0.000	20:34

Agency: FORT LAUDERDALE PD

Date/Time: 04/25/2023 19:34:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 04/25/2023 20:45

Air Blank	0.000	19:33
		19:33
Air Blank	0.000	19:34

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

Agency: FORT LAUDERDALE PD

Date/Time: 11/28/2023 20:55:50

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/28/2023 22:06

Air Blank	0.000	20:55
Control Test	0.000	20:55
Air Blank	0.000	20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

Agency: BROWARD COUNTY S.O.

Date/Time: 01/04/2023 09:13:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2023 10:07

Air Blank	0.000	09:12
Control Test	0.200	09:13
Air Blank	0.000	09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

Agency: HILLSBOROUGH CO SO

Date/Time: 02/07/2023 09:48:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2023 10:35

Air Blank	0.000	09:47
Control Test	0.075	09:47
Air Blank	0.000	09:48

Agency: HILLSBOROUGH CO SO

Date/Time: 02/07/2023 09:55:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/07/2023 10:35

Air Blank	0.000	09:54
Control Test	0.076	09:55
Air Blank	0.000	09:55

Agency: HILLSBOROUGH CO SO

Date/Time: 03/22/2023 11:31:42

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/10/2023 10:16

Air Blank	0.000	11:30
Control Test	0.080	11:31
Air Blank	0.000	11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

Agency: FFWCC

Date/Time: 11/01/2023 13:35:40

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/01/2023 15:10

Air Blank	0.000	13:34
Control Test	0.081	13:35
Air Blank	0.000	13:35

Agency: FFWCC

Date/Time: 11/01/2023 13:33:48

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 11/01/2023 15:10

Air Blank	0.000	13:32
Control Test	0.081	13:33
Air Blank	0.000	13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

Agency: MONROE COUNTY SO

Date/Time: 04/04/2023 12:24:52

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 04/04/2023 13:06

Air Blank	0.000	12:24
Control Test	0.080	12:24
Air Blank	0.000	12:24

Agency: MONROE COUNTY SO

Date/Time: 09/14/2023 11:01:12

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/14/2023 11:54

Air Blank	0.000	11:00
Control Test	0.081	11:00
Air Blank	0.000	11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 09:48:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:17

Air Blank	0.000	09:47
Control Test	0.076	09:48
Air Blank	0.000	09:48

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 09:52:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:17

Air Blank	0.000	09:51
Control Test	0.076	09:51
Air Blank	0.000	09:52

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 09:56:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:17

Air Blank	0.000	09:54
Control Test	0.045	09:55
Air Blank	0.000	09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 09:59:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:17

Air Blank	0.000	09:58
Control Test	0.052	09:59
Air Blank	0.000	09:59

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:01:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:17

Air Blank	0.000	10:00
Control Test	0.057	10:01
Air Blank	0.000	10:01

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:20:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:17

Air Blank	0.000	10:19
Control Test	0.046	10:19
Air Blank	0.000	10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:22:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:18

Air Blank	0.000	10:20
Control Test	0.047	10:21
Air Blank	0.000	10:22

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:26:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:18

Air Blank	0.000	10:24
Control Test	0.077	10:25
Air Blank	0.000	10:26

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:28:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:18

Air Blank	0.000	10:27
Control Test	0.078	10:27
Air Blank	0.000	10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

Agency: MONROE COUNTY SO

Date/Time: 03/07/2023 10:30:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/07/2023 11:18

Air Blank	0.000	10:29
Control Test	0.047	10:29
Air Blank	0.000	10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

Agency: MARION COUNTY SO

Date/Time: 02/20/2023 14:49:24

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/20/2023 15:56

Air Blank	0.000	14:48
Control Test	0.000	14:48
Air Blank	0.000	14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

Agency: GROVELAND PD

Date/Time: 08/28/2023 12:26:12

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/28/2023 13:51

Air Blank	0.000	12:25
Control Test	0.077	12:25
Air Blank	0.000	12:26

Agency: GROVELAND PD

Date/Time: 08/28/2023 12:27:32

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 08/28/2023 13:51

Air Blank	0.000	12:26
Control Test	0.078	12:27
Air Blank	0.000	12:27

Agency: GROVELAND PD

Date/Time: 09/25/2023 14:40:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 15:26

Air Blank	0.000	14:39
Control Test	0.048	14:40
Air Blank	0.000	14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

Agency: GROVELAND PD

Date/Time: 09/25/2023 14:42:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 15:26

Air Blank	0.000	14:41
Control Test	0.077	14:42
Air Blank	0.000	14:42

Agency: GROVELAND PD

Date/Time: 09/25/2023 14:44:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 15:26

Air Blank	0.000	14:43
Control Test	0.197	14:44
Air Blank	0.000	14:44

Agency: GROVELAND PD

Date/Time: 11/13/2023 14:23:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/13/2023 15:28

Air Blank	0.000	14:22
Control Test	0.047	14:23
Air Blank	0.000	14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

Agency: GROVELAND PD

Date/Time: 11/13/2023 14:25:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/13/2023 15:28

Air Blank	0.000	14:24
Control Test	0.077	14:24
Air Blank	0.000	14:25

Agency: GROVELAND PD

Date/Time: 11/13/2023 14:26:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 11/13/2023 15:28

Air Blank	0.000	14:25
Control Test	0.195	14:26
Air Blank	0.000	14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

Agency: ST JOHNS SO

Date/Time: 11/18/2023 13:50:31

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 12/06/2023 00:34

Air Blank	0.000	13:49
ITP Check	0.100	13:50
Air Blank	0.000	13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

Agency: HOMESTEAD PD

Date/Time: 03/22/2023 13:32:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/22/2023 14:42

Air Blank	0.000	13:30
Control Test	0.197	13:31
Air Blank	0.000	13:32

Agency: HOMESTEAD PD

Date/Time: 05/30/2023 13:31:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 05/30/2023 14:32

Air Blank	0.000	13:30
Control Test	0.048	13:30
Air Blank	0.000	13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

Agency: FHP

Date/Time: 03/30/2023 13:45:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/31/2023 14:59

Air Blank	0.000	13:43
Control Test	0.078	13:44
Air Blank	0.000	13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

Agency: FHP

Date/Time: 09/25/2023 07:20:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 09/25/2023 10:11

Air Blank	0.000	07:19
Control Test	0.000	07:20
Air Blank	0.000	07:20

Agency: FHP

Date/Time: 09/25/2023 07:22:58

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 09/25/2023 10:11

Air Blank	0.000	07:22
Control Test	0.083	07:22
Air Blank	0.000	07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

Agency: FHP

Date/Time: 01/25/2023 17:30:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2023 08:46

Air Blank	0.000	17:29
Control Test	0.000	17:30
Air Blank	0.000	17:30

Agency: FHP

Date/Time: 01/25/2023 17:33:17

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2023 08:47

Air Blank	0.000	17:32
Control Test	0.000	17:32
Air Blank	0.000	17:33

Agency: FHP

Date/Time: 01/25/2023 17:36:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/27/2023 08:47

Air Blank	0.000	17:35
Control Test	0.049	17:35
Air Blank	0.000	17:36

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

Agency: FHP

Date/Time: 01/25/2023 17:53:27

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/27/2023 08:47

Air Blank	0.000	17:52
Control Test	0.082	17:52
Air Blank	0.000	17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

Agency: FHP

Date/Time: 06/14/2023 08:50:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 06/14/2023 13:28

Air Blank	0.000	08:49
Control Test	0.000	08:50
Air Blank	0.000	08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

Agency: HILLSBOROUGH CO SO

Date/Time: 02/09/2023 16:21:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/10/2023 14:49

Air Blank	0.000	16:20
Control Test	0.195	16:21
Air Blank	0.000	16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

Agency: FFWCC

Date/Time: 05/01/2023 13:32:17

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 05/15/2023 16:33

Air Blank	0.000	13:31
Control Test	0.078	13:31
Air Blank	0.000	13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:15:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/28/2023 02:58

Air Blank	0.000	21:14
Control Test	0.000	21:15
Air Blank	0.000	21:15

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:18:22

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/28/2023 02:58

Air Blank	0.000	21:17
Control Test	0.000	21:17
Air Blank	0.000	21:18

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:40:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/28/2023 02:58

Air Blank	0.000	21:39
Control Test	0.000	21:40
Air Blank	0.000	21:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:42:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:18

Air Blank	0.000	04:41
Control Test	0.075	04:41
Air Blank	0.000	04:42

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:44:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:18

Air Blank	0.000	04:42
Control Test	0.195	04:43
Air Blank	0.000	04:44

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:46:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:18

Air Blank	0.000	04:45
Control Test	0.000	04:46
Air Blank	0.000	04:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:47:31

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:18

Air Blank		04:47
Air Blank	0.000	04:47

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:50:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:18

Air Blank	0.000	04:48
Control Test	0.046	04:49
Air Blank	0.000	04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/18/2023 18:36:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2023 19:25

Air Blank	0.000	18:34
Control Test	0.196	18:35
Air Blank	0.000	18:36

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/18/2023 18:38:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2023 19:25

Air Blank	0.000	18:37
Control Test	0.076	18:38
Air Blank	0.000	18:38

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/18/2023 18:40:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2023 19:25

Air Blank	0.000	18:39
Control Test	0.048	18:40
Air Blank	0.000	18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/18/2023 18:43:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2023 19:25

Air Blank	0.000	18:41
Control Test	0.000	18:42
Air Blank	0.000	18:43

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:08:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 01:41

Air Blank	0.000	21:07
Control Test	0.197	21:07
Air Blank	0.000	21:08

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:10:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 01:41

Air Blank	0.000	21:09
Control Test	0.076	21:10
Air Blank	0.000	21:10

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:12:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 01:41

Air Blank	0.000	21:11
Control Test	0.048	21:12
Air Blank	0.000	21:12

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:14:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 01:41

Air Blank	0.000	21:13
Control Test	0.000	21:14
Air Blank	0.000	21:14

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:16:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 01:41

Air Blank	0.000	21:15
Control Test	0.000	21:15
Air Blank	0.000	21:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:25:34

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 01:41

Air Blank	0.000	21:24
Control Test	0.000	21:25
Air Blank	0.000	21:25

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/23/2023 21:41:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/24/2023 01:41

Air Blank	0.000	21:40
Control Test	0.000	21:40
Air Blank	0.000	21:41

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:38:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:58

Air Blank	0.000	04:37
Control Test	0.197	04:38
Air Blank	0.000	04:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:41:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:58

Air Blank	0.000	04:40
Control Test	0.076	04:40
Air Blank	0.000	04:41

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:43:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:58

Air Blank	0.000	04:42
Control Test	0.046	04:42
Air Blank	0.000	04:43

Agency: HIGHLANDS COUNTY SO

Date/Time: 07/10/2023 04:45:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 07/10/2023 06:58

Air Blank	0.000	04:44
Control Test	0.000	04:44
Air Blank	0.000	04:45

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

Agency: ST JOHNS COUNTY SO

Date/Time: 12/08/2023 20:52:52

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/01/2024 21:08

Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:52

Agency: ST JOHNS COUNTY SO

Date/Time: 12/11/2023 08:54:19

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/01/2024 21:08

Air Blank	0.000	08:53
Control Test	0.080	08:53
Air Blank	0.000	08:54

Agency: ST JOHNS COUNTY SO

Date/Time: 12/11/2023 08:56:00

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/01/2024 21:08

Air Blank	0.000	08:55
Control Test	0.080	08:55
Air Blank	0.000	08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

Agency: MIAMI PD

Date/Time: 01/15/2023 18:43:34

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/18/2023 12:44

Air Blank	0.000	18:42
Control Test	0.082	18:43
Air Blank	0.000	18:43
